

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>CHEMTECH</u>			Contract:	<u>WOOD06</u>		
Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1725</u>	SAS No.:	<u>Q1725</u>	SDG NO.:	<u>Q1725</u>
Lab File ID:	<u>VL042228.D</u>			Date Analyzed:	<u>04/07/2025</u>		
Instrument ID:	<u>MSVOA_L</u>			Time Analyzed:	<u>09:31</u>		
GC Column:	RTX-1	ID:	0.32 (mm)	Heated Purge:	(Y/N)	N	

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	190462	2.80	513778	3.98	464477	8.90
UPPER LIMIT	266647	3.13	719289	4.31	650268	9.23
LOWER LIMIT	114277	2.47	308267	3.65	278686	8.57
EPA SAMPLE NO.						
SSSG-B1-2-04012025	198158	2.80	527291	3.98	473671	8.90
SSSG-B10-4-04012025	200550	2.80	538469	3.98	487966	8.90
SSSG-B10-4-04012025DL	203982	2.80	532301	3.98	458285	8.90
SSSG-B10-4-04012025DL2	196689	2.81	503068	3.99	430861	8.91
SSSG-DUP-2-04012025DL	190938	2.80	485801	3.99	425989	8.90
AA-1-03312025	202620	2.80	551417	3.98	479781	8.90
AA-1-03312025DUP	201241	2.80	543996	3.98	475092	8.90
SSSG-B14-3-04012025	191155	2.80	493918	3.99	441320	8.91
SSSG-B14-3-04012025DL	192644	2.80	502203	3.99	443302	8.91
SSSG-B10-3-03312025	194534	2.80	495845	3.99	437126	8.90
SSSG-B10-3-03312025DL	194329	2.80	493793	3.98	430982	8.90
SSSG-B10-1-03312025-SPLIT	197758	2.81	506596	3.99	465539	8.91
SSSG-B10-1-03312025-SPLITDL	193361	2.80	501287	3.98	435889	8.90
SSSG-B10-2-03312025	202914	2.80	515559	3.98	442488	8.90
SSSG-B10-2-03312025DL	194708	2.80	494678	3.99	424907	8.91
VL0407ABL01	195732	2.80	527491	3.98	463741	8.90
VL0407ABS01	197822	2.80	531347	3.98	473981	8.90

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: WOOD06
 Lab Code: CHEM Case No.: Q1725 SAS No.: Q1725 SDG NO.: Q1725
 Lab File ID: VL042228.D Date Analyzed: 04/07/2025
 Instrument ID: MSVOA_L Time Analyzed: 09:31
 GC Column: RTX-1 ID: 0.32 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	190462	2.80	513778	3.98	464477	8.90
UPPER LIMIT	266647	3.13	719289	4.31	650268	9.23
LOWER LIMIT	114277	2.47	308267	3.65	278686	8.57
EPA SAMPLE NO.						

IS1 = Bromochloromethane
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +40% of internal standard area
 AREA LOWER LIMIT = -40% of internal standard area
 RT UPPER LIMIT = +0.33 minutes of internal standard RT
 RT LOWER LIMIT = -0.33 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.