

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: WOOD06
Lab Code: CHEM Case No.: Q1725 SAS No.: Q1725 SDG NO.: Q1725
Lab File ID: VL042253.D Date Analyzed: 04/08/2025
Instrument ID: MSVOA_L Time Analyzed: 08:39
GC Column: RTX-1 ID: 0.32 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	192748	2.80	496271	3.98	446954	8.90
UPPER LIMIT	269847	3.13	694779	4.31	625736	9.23
LOWER LIMIT	115649	2.47	297763	3.65	268172	8.57
EPA SAMPLE NO.						
SSSG-DUP-2-04012025	196935	2.80	510940	3.98	446395	8.90
VL0408ABL01	194254	2.80	507155	3.98	433594	8.90
VL0408ABS01	190420	2.80	494678	3.98	447066	8.90

IS1 = Bromochloromethane
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +40% of internal standard area
AREA LOWER LIMIT = -40% of internal standard area
RT UPPER LIMIT = +0.33 minutes of internal standard RT
RT LOWER LIMIT = -0.33 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.