

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>CHEMTECH</u>			Contract:	<u>WOOD06</u>		
Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1725</u>	SAS No.:	<u>Q1725</u>	SDG NO.:	<u>Q1725</u>
Lab File ID:	<u>VL042228.D</u>			Date Analyzed:	<u>04/07/2025</u>		
Instrument ID:	<u>MSVOA_L</u>			Time Analyzed:	<u>09:31</u>		
GC Column:	RTX-1	ID:	0.32 (mm)	Heated Purge:	(Y/N)	N	

	IS4 AREA #	RT #				
12 HOUR STD	0	0				
UPPER LIMIT	0					
LOWER LIMIT	0					
EPA SAMPLE NO.						
SSSG-B1-2-04012025	0	0.00				
SSSG-B10-4-04012025	0	0.00				
SSSG-B10-4-04012025DL	0	0.00				
SSSG-B10-4-04012025DL2	0	0.00				
SSSG-DUP-2-04012025DL	0	0.00				
AA-1-03312025	0	0.00				
AA-1-03312025DUP	0	0.00				
SSSG-B14-3-04012025	0	0.00				
SSSG-B14-3-04012025DL	0	0.00				
SSSG-B10-3-03312025	0	0.00				
SSSG-B10-3-03312025DL	0	0.00				
SSSG-B10-1-03312025-SPLIT	0	0.00				
SSSG-B10-1-03312025-SPLITDL	0	0.00				
SSSG-B10-2-03312025	0	0.00				
SSSG-B10-2-03312025DL	0	0.00				
VL0407ABL01	0	0.00				
VL0407ABS01	0	0.00				

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: WOOD06
 Lab Code: CHEM Case No.: Q1725 SAS No.: Q1725 SDG NO.: Q1725
 Lab File ID: VL042228.D Date Analyzed: 04/07/2025
 Instrument ID: MSVOA_L Time Analyzed: 09:31
 GC Column: RTX-1 ID: 0.32 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	0	0				
UPPER LIMIT	0					
LOWER LIMIT	0					
EPA SAMPLE NO.						

IS4 =

AREA UPPER LIMIT = +40% of internal standard area

AREA LOWER LIMIT = -40% of internal standard area

RT UPPER LIMIT = +0.33 minutes of internal standard RT

RT LOWER LIMIT = -0.33 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.