



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_L\Data\VL041725\  
Data File : VL042348.D  
Acq On : 17 Apr 2025 07:50  
Operator : SY/MD  
Sample : BFB  
Misc : 400mL/MSVOA\_L  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_L  
ClientSampleId :  
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA\_L\methods\VL041725AIR.M  
Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA\_L Fri Aug 26 06:05:16 2022  
Last Update : Fri Apr 18 02:14:00 2025



AutoFind: Scans 3177, 3178, 3179; Background Corrected with Scan 3162

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 50             | 95              | 8               | 40              | 23.1         | 42933      | PASS                |
| 75             | 95              | 30              | 66              | 57.0         | 105781     | PASS                |
| 95             | 95              | 100             | 100             | 100.0        | 185564     | PASS                |
| 96             | 95              | 5               | 9               | 6.8          | 12627      | PASS                |
| 173            | 174             | 0.00            | 2               | 0.0          | 0          | PASS                |
| 174            | 95              | 50              | 120             | 59.5         | 110360     | PASS                |
| 175            | 174             | 4               | 9               | 7.6          | 8420       | PASS                |
| 176            | 174             | 93              | 101             | 96.9         | 106933     | PASS                |
| 177            | 176             | 5               | 9               | 6.3          | 6744       | PASS                |

Average of 10.371 to 10.377 min.: VL042348.D\data.ms

BFB

Modified:subtracted

| m/z   | abund. | m/z   | abund. | m/z   | abund. | m/z   | abund. |
|-------|--------|-------|--------|-------|--------|-------|--------|
| 33.10 | 400    | 45.00 | 1996   | 55.00 | 78     | 67.20 | 32     |
| 34.00 | 54     | 45.90 | 94     | 56.05 | 2900   | 68.05 | 2203   |
| 36.05 | 2214   | 46.20 | 34     | 57.05 | 5754   | 69.00 | 2101   |
| 37.10 | 11145  | 47.10 | 2100   | 57.95 | 265    | 70.00 | 130    |
| 38.10 | 9197   | 48.10 | 1438   | 60.05 | 1919   | 70.20 | 209    |
| 39.10 | 4289   | 49.05 | 9045   | 61.05 | 9155   | 72.05 | 1141   |
| 39.95 | 158    | 50.10 | 42933  | 62.05 | 9788   | 73.05 | 8240   |
| 41.05 | 513    | 51.10 | 13188  | 63.10 | 7593   | 74.10 | 34064  |
| 41.80 | 79     | 52.00 | 718    | 63.95 | 632    | 75.10 | 105781 |
| 42.95 | 504    | 52.95 | 9      | 66.10 | 115    | 76.10 | 9240   |
| 44.00 | 1207   | 54.15 | 112    | 66.90 | 196    | 77.05 | 979    |

Average of 10.371 to 10.377 min.: VL042348.D\data.ms

BFB

Modified:subtracted

| m/z   | abund. | m/z    | abund. | m/z    | abund. | m/z    | abund. |
|-------|--------|--------|--------|--------|--------|--------|--------|
| 77.90 | 494    | 90.85  | 447    | 105.00 | 95     | 115.95 | 528    |
| 78.90 | 4344   | 91.10  | 262    | 105.85 | 680    | 116.95 | 712    |
| 79.95 | 1296   | 92.05  | 5876   | 107.00 | 225    | 118.00 | 575    |
| 80.95 | 4648   | 93.10  | 8069   | 110.10 | 136    | 119.00 | 881    |
| 82.00 | 1107   | 94.10  | 22619  | 110.80 | 37     | 120.10 | 70     |
| 83.00 | 249    | 95.05  | 185564 | 111.10 | 8      | 120.90 | 30     |
| 84.40 | 55     | 96.10  | 12627  | 111.95 | 156    | 121.30 | 47     |
| 85.00 | 227    | 97.05  | 457    | 112.75 | 77     | 122.20 | 21     |
| 85.85 | 131    | 98.05  | 5      | 113.10 | 144    | 122.95 | 44     |
| 87.05 | 5767   | 102.90 | 78     | 114.80 | 135    | 123.50 | 25     |
| 88.05 | 5057   | 104.00 | 767    | 115.10 | 50     | 124.10 | 89     |

Average of 10.371 to 10.377 min.: VL042348.D\data.ms

BFB

Modified:subtracted

| m/z    | abund. | m/z    | abund. | m/z    | abund. | m/z    | abund. |
|--------|--------|--------|--------|--------|--------|--------|--------|
| 124.90 | 71     | 137.05 | 364    | 146.15 | 158    | 153.90 | 93     |
| 125.75 | 72     | 139.65 | 74     | 146.65 | 58     | 154.20 | 27     |
| 126.90 | 80     | 140.00 | 38     | 147.10 | 68     | 154.95 | 399    |
| 127.80 | 297    | 141.00 | 1612   | 147.70 | 38     | 156.90 | 228    |
| 128.00 | 275    | 141.80 | 139    | 148.00 | 288    | 158.95 | 203    |
| 128.85 | 366    | 142.10 | 34     | 148.95 | 115    | 160.80 | 26     |
| 130.00 | 693    | 142.70 | 434    | 149.80 | 37     | 161.05 | 186    |
| 130.80 | 55     | 143.00 | 1656   | 149.95 | 79     | 168.90 | 83     |
| 130.95 | 285    | 143.95 | 200    | 151.90 | 60     | 170.50 | 119    |
| 134.70 | 94     | 144.95 | 161    | 152.75 | 91     | 171.30 | 19     |
| 135.00 | 292    | 145.80 | 67     | 153.05 | 144    | 171.90 | 90     |

Average of 10.371 to 10.377 min.: VL042348.D\data.ms

BFB

Modified:subtracted

| m/z    | abund. | m/z | abund. | m/z | abund. | m/z | abund. |
|--------|--------|-----|--------|-----|--------|-----|--------|
| 172.30 | 132    |     |        |     |        |     |        |
| 174.00 | 110360 |     |        |     |        |     |        |
| 175.05 | 8420   |     |        |     |        |     |        |
| 176.00 | 106933 |     |        |     |        |     |        |
| 177.05 | 6744   |     |        |     |        |     |        |
| 178.05 | 298    |     |        |     |        |     |        |
| 194.20 | 17     |     |        |     |        |     |        |
| 214.20 | 17     |     |        |     |        |     |        |
| 216.30 | 17     |     |        |     |        |     |        |
| 254.90 | 17     |     |        |     |        |     |        |

Instrument :

MSVOA\_L

ClientSampleId :

BFB