

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	Q1726	Analytical Method:	SWTO-15
Client:	Woodard & Curran	Datafile :	VL042385.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VL0418ABS01	Dichlorodifluoromethane	10	9.30	ppbv	93			40 (70)	160 (130)	
	Ethanol	10	2.30	ppbv	23		*	70 (70)	130 (130)	
	Chloromethane	10	8.60	ppbv	86			40 (70)	160 (130)	
	Vinyl Chloride	10	7.60	ppbv	76			70 (70)	130 (130)	
	Bromomethane	10	8.50	ppbv	85			40 (70)	160 (130)	
	Chloroethane	10	8.30	ppbv	83			40 (70)	160 (130)	
	Tetrahydrofuran	10	6.90	ppbv	69		*	70 (70)	130 (130)	
	Trichlorofluoromethane	10	9.50	ppbv	95			40 (70)	160 (130)	
	1,1,2-Trichlorotrifluoroethane	10	8.90	ppbv	89			70 (70)	130 (130)	
	Dichlorotetrafluoroethane	10	9.10	ppbv	91			70 (70)	130 (130)	
	Bromoethene	10	9.10	ppbv	91			70 (70)	130 (130)	
	tert-Butyl Alcohol	10	8.20	ppbv	82			70 (70)	130 (130)	
	Heptane	10	6.80	ppbv	68		*	70 (70)	130 (130)	
	Isopropyl Alcohol	10	8.20	ppbv	82			70 (70)	130 (130)	
	1,1-Dichloroethene	10	8.70	ppbv	87			70 (70)	130 (130)	
	Acetone	10	7.30	ppbv	73			40 (70)	160 (130)	
	Carbon disulfide	10	8.70	ppbv	87			40 (70)	160 (130)	
	Methyl tert-butyl Ether	10	8.30	ppbv	83			70 (70)	130 (130)	
	Methylene Chloride	10	7.40	ppbv	74			70 (70)	130 (130)	
	trans-1,2-Dichloroethene	10	8.70	ppbv	87			70 (70)	130 (130)	
	1,1-Dichloroethane	10	8.90	ppbv	89			70 (70)	130 (130)	
	Cyclohexane	10	8.10	ppbv	81			70 (70)	130 (130)	
	2-Butanone	10	7.20	ppbv	72			40 (70)	160 (130)	
	Carbon Tetrachloride	10	8.60	ppbv	86			70 (70)	130 (130)	
	cis-1,2-Dichloroethene	10	8.60	ppbv	86			70 (70)	130 (130)	
	Chloroform	10	9.10	ppbv	91			70 (70)	130 (130)	
	1,1,1-Trichloroethane	10	8.90	ppbv	89			70 (70)	130 (130)	
	2,2,4-Trimethylpentane	10	7.20	ppbv	72			70 (70)	130 (130)	
	Benzene	10	7.70	ppbv	77			70 (70)	130 (130)	
	1,2-Dichloroethane	10	8.70	ppbv	87			70 (70)	130 (130)	
	Trichloroethene	10	7.60	ppbv	76			70 (70)	130 (130)	
	1,2-Dichloropropane	10	7.20	ppbv	72			70 (70)	130 (130)	
	Bromodichloromethane	10	8.50	ppbv	85			70 (70)	130 (130)	
	4-Methyl-2-Pentanone	10	6.40	ppbv	64			40 (70)	160 (130)	
	Toluene	10	7.60	ppbv	76			70 (70)	130 (130)	
	t-1,3-Dichloropropene	10	7.60	ppbv	76			70 (70)	130 (130)	
	cis-1,3-Dichloropropene	10	7.50	ppbv	75			70 (70)	130 (130)	
	1,1,2-Trichloroethane	10	8.20	ppbv	82			70 (70)	130 (130)	
	Dibromochloromethane	10	8.80	ppbv	88			70 (70)	130 (130)	
	1,2-Dibromoethane	10	8.20	ppbv	82			70 (70)	130 (130)	
	Tetrachloroethene	10	7.70	ppbv	77			70 (70)	130 (130)	
	Chlorobenzene	10	8.40	ppbv	84			70 (70)	130 (130)	
	Ethyl Benzene	10	8.10	ppbv	81			70 (70)	130 (130)	
	m/p-Xylene	20	16.4	ppbv	82			70 (70)	130 (130)	
	o-Xylene	10	8.20	ppbv	82			70 (70)	130 (130)	
	Styrene	10	8.20	ppbv	82			70 (70)	130 (130)	
	Bromoform	10	8.80	ppbv	88			70 (70)	130 (130)	
	1,1,2,2-Tetrachloroethane	10	7.30	ppbv	73			70 (70)	130 (130)	

() = LABORATORY INHOUSE LIMIT

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q1726</u>	Analytical Method:	<u>SWTO-15</u>
Client:	<u>Woodard & Curran</u>	Datafile :	<u>VL042385.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VL0418ABS01	2-Chlorotoluene	10	8.20	ppbv	82			70 (70)	130 (130)	
	1,3,5-Trimethylbenzene	10	8.10	ppbv	81			70 (70)	130 (130)	
	1,2,4-Trimethylbenzene	10	7.70	ppbv	77			70 (70)	130 (130)	
	1,3-Dichlorobenzene	10	8.60	ppbv	86			70 (70)	130 (130)	
	1,4-Dichlorobenzene	10	8.50	ppbv	85			70 (70)	130 (130)	
	1,2-Dichlorobenzene	10	8.40	ppbv	84			70 (70)	130 (130)	
	1,2,4-Trichlorobenzene	10	8.40	ppbv	84			70 (70)	130 (130)	
	Hexachloro-1,3-butadiene	10	8.60	ppbv	86			70 (70)	130 (130)	
	Naphthalene	10	7.70	ppbv	77			40 (70)	160 (130)	
	1,3-Butadiene	10	7.60	ppbv	76			70 (70)	130 (130)	
	4-Ethyltoluene	10	8.30	ppbv	83			70 (70)	130 (130)	
	Hexane	10	7.60	ppbv	76			70 (70)	130 (130)	
	Allyl Chloride	10	8.00	ppbv	80			70 (70)	130 (130)	
	1,4-Dioxane	10	6.80	ppbv	68			40 (70)	160 (130)	
	Methyl methacrylate	10	6.90	ppbv	69		*	70 (70)	130 (130)	