

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: WOOD06
 Lab Code: CHEM Case No.: Q1726 SAS No.: Q1726 SDG NO.: Q1726
 Lab File ID: VL042349.D Date Analyzed: 04/17/2025
 Instrument ID: MSVOA_L Time Analyzed: 09:44
 GC Column: RTX-1 ID: 0.32 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	125455	2.79	366544	3.96	338126	8.88
UPPER LIMIT	175637	3.12	513162	4.29	473376	9.21
LOWER LIMIT	75273	2.46	219926	3.63	202876	8.55
EPA SAMPLE NO.						
IA-DUP-3-03312025	122216	2.79	344823	3.96	303405	8.88
IA-DUP-3-03312025DL	100356	2.79	288404	3.96	249449	8.88
IA-B2-6-03312025	117602	2.79	335422	3.96	296955	8.88
IA-B2-6-03312025DL	99418	2.78	272592	3.96	237645	8.88
IA-B2-2-03312025	119191	2.79	333003	3.96	297493	8.88
IA-B2-2-03312025DL	92120	2.79	265690	3.96	228865	8.88
IA-B2-5-03312025	115766	2.79	331154	3.96	288378	8.88
IA-B2-5-03312025DL	92518	2.78	260556	3.96	221578	8.88
IA-B2-1-03312025	113751	2.79	331268	3.96	291013	8.88
IA-B2-1-03312025DL	90034	2.79	252514	3.96	219443	8.88
IA-B2-3-03312025	113154	2.79	322735	3.96	284478	8.88
IA-B2-3-03312025DL	87593	2.78	246915	3.95	211054	8.88
IA-B2-4-03312025	110448	2.79	317011	3.96	275197	8.88
IA-B2-4-03312025DL	107602	2.78	303479	3.95	266154	8.88
VL0417ABL01	113566	2.79	328716	3.97	297633	8.89
VL0417ABS01	122375	2.79	352201	3.97	321521	8.89

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: WOOD06
 Lab Code: CHEM Case No.: Q1726 SAS No.: Q1726 SDG NO.: Q1726
 Lab File ID: VL042349.D Date Analyzed: 04/17/2025
 Instrument ID: MSVOA_L Time Analyzed: 09:44
 GC Column: RTX-1 ID: 0.32 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	125455	2.79	366544	3.96	338126	8.88
UPPER LIMIT	175637	3.12	513162	4.29	473376	9.21
LOWER LIMIT	75273	2.46	219926	3.63	202876	8.55
EPA SAMPLE NO.						

IS1 = Bromochloromethane
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +40% of internal standard area
 AREA LOWER LIMIT = -40% of internal standard area
 RT UPPER LIMIT = +0.33 minutes of internal standard RT
 RT LOWER LIMIT = -0.33 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.