

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: WOOD06
Lab Code: CHEM Case No.: Q1726 SAS No.: Q1726 SDG NO.: Q1726
Lab File ID: VL042383.D Date Analyzed: 04/18/2025
Instrument ID: MSVOA_L Time Analyzed: 10:23
GC Column: RTX-1 ID: 0.32 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	135964	2.78	427541	3.95	379215	8.88
UPPER LIMIT	190350	3.11	598557	4.28	530901	9.21
LOWER LIMIT	81578.4	2.45	256525	3.62	227529	8.55
EPA SAMPLE NO.						
IA-DUP-3-03312025DL2	133179	2.78	411086	3.95	357783	8.88
IA-B2-6-03312025DL2	132986	2.78	409322	3.95	354583	8.88
IA-B2-2-03312025DL2	127324	2.79	395389	3.96	342875	8.89
IA-B2-5-03312025DL2	134501	2.79	412843	3.96	353048	8.89
IA-B2-1-03312025DL2	125703	2.79	390416	3.96	327579	8.88
IA-B2-3-03312025DL2	121646	2.79	379829	3.97	325040	8.89
IA-B2-3-03312025DL2DUP	123948	2.79	373006	3.97	318921	8.89
IA-B2-4-03312025DL2	137576	2.79	424572	3.96	368674	8.88
VL0418ABL01	137116	2.78	425910	3.96	379163	8.88
VL0418ABS01	137596	2.78	421834	3.95	371546	8.88

IS1 = Bromochloromethane
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +40% of internal standard area
AREA LOWER LIMIT = -40% of internal standard area
RT UPPER LIMIT = +0.33 minutes of internal standard RT
RT LOWER LIMIT = -0.33 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.