



Lab Name:	<u>CHEMTECH</u>			Contract:	<u>WOOD06</u>		
Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1728</u>	SAS No.:	<u>Q1728</u>	SDG NO.:	<u>Q1728</u>
Lab File ID:	<u>VL042253.D</u>			Date Analyzed:	<u>04/08/2025</u>		
Instrument ID:	<u>MSVOA_L</u>			Time Analyzed:	<u>08:39</u>		
GC Column:	RTX-1	ID:	0.32 (mm)	Heated Purge:	(Y/N)	N	

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	192748	2.80	496271	3.98	446954	8.90
UPPER LIMIT	269847	3.13	694779	4.31	625736	9.23
LOWER LIMIT	115649	2.47	297763	3.65	268172	8.57
EPA SAMPLE NO.						
SSSG-B14-2-04012025	164142	2.79	421001	3.97	366383	8.89
SSSG-B14-2-04012025DL	162199	2.80	442538	3.98	374779	8.90
SSSG-B1-5-04012025	158831	2.79	406344	3.96	351845	8.89
SSSG-B2-2-04022025	156706	2.79	394605	3.96	369232	8.89
SSSG-B2-7-04022025	154351	2.79	392096	3.97	376954	8.89
SSSG-B2-7-04022025DL	147116	2.79	371349	3.97	326259	8.89
SSSG-B2-6-04022025	153705	2.79	388271	3.96	381682	8.88
SSSG-B2-6-04022025DL	151544	2.79	379950	3.96	331294	8.88
SSSG-B2-5-04022025	155754	2.80	446160	3.98	402493	8.90
SSSG-B2-5-04022025DL	158005	2.80	439373	3.98	391568	8.90
SSSG-B1-1-04012025	160928	2.80	398547	3.98	347731	8.89
SSSG-B2-4-04022025	164494	2.80	443030	3.98	423016	8.90
SSSG-B2-4-04022025DL	152932	2.79	379227	3.97	368372	8.89
SSSG-B2-1-04022025	157346	2.80	411449	3.98	406705	8.90
SSSG-B2-1-04022025DUP	158259	2.79	390082	3.96	405082	8.89
SSSG-B14-4-04022025	155290	2.79	388359	3.96	341830	8.88
SSSG-B1-4-04012025	156372	2.79	398763	3.97	359554	8.89
SSSG-B1-3-04012025	160295	2.79	399383	3.97	355820	8.89
VL0408ABL01	194254	2.80	507155	3.98	433594	8.90
VL0408ABS01	190420	2.80	494678	3.98	447066	8.90

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: WOOD06
 Lab Code: CHEM Case No.: Q1728 SAS No.: Q1728 SDG NO.: Q1728
 Lab File ID: VL042253.D Date Analyzed: 04/08/2025
 Instrument ID: MSVOA_L Time Analyzed: 08:39
 GC Column: RTX-1 ID: 0.32 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	192748	2.80	496271	3.98	446954	8.90
UPPER LIMIT	269847	3.13	694779	4.31	625736	9.23
LOWER LIMIT	115649	2.47	297763	3.65	268172	8.57
EPA SAMPLE NO.						

IS1 = Bromochloromethane
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +40% of internal standard area
 AREA LOWER LIMIT = -40% of internal standard area
 RT UPPER LIMIT = +0.33 minutes of internal standard RT
 RT LOWER LIMIT = -0.33 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.