

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1729 SAS No.: Q1729 SDG NO.: Q1729

Continuing Calib Date: 04/08/2025 Initial Calibration Date(s): 03/17/2025 03/17/2025

Continuing Calib Time: 10:20 Initial Calibration Time(s): 12:56 22:25

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	4.50	4.49	4.42	4.56	-0.01
Aroclor-1016-2	(2)	4.52	4.50	4.43	4.57	-0.02
Aroclor-1016-3	(3)	4.58	4.56	4.49	4.63	-0.01
Aroclor-1016-4	(4)	4.66	4.65	4.58	4.72	-0.01
Aroclor-1016-5	(5)	4.95	4.94	4.87	5.01	-0.01
Aroclor-1260-1	(1)	6.04	6.03	5.96	6.10	-0.01
Aroclor-1260-2	(2)	6.29	6.28	6.21	6.35	-0.01
Aroclor-1260-3	(3)	6.65	6.63	6.56	6.70	-0.01
Aroclor-1260-4	(4)	6.86	6.85	6.78	6.92	-0.01
Aroclor-1260-5	(5)	7.16	7.15	7.08	7.22	-0.01
TCX		3.42	3.41	3.36	3.46	0.00
DCB		8.52	8.50	8.40	8.60	-0.02

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1729 SAS No.: Q1729 SDG NO.: Q1729

Continuing Calib Date: 04/08/2025 Initial Calibration Date(s): 03/17/2025 03/17/2025

Continuing Calib Time: 10:20 Initial Calibration Time(s): 12:56 22:25

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	3.79	3.79	3.72	3.86	0.00
Aroclor-1016-2	(2)	3.80	3.80	3.73	3.87	0.00
Aroclor-1016-3	(3)	3.96	3.96	3.89	4.03	0.00
Aroclor-1016-4	(4)	4.01	4.01	3.94	4.08	0.00
Aroclor-1016-5	(5)	4.21	4.21	4.14	4.28	0.01
Aroclor-1260-1	(1)	5.20	5.20	5.13	5.27	0.00
Aroclor-1260-2	(2)	5.39	5.39	5.32	5.46	0.00
Aroclor-1260-3	(3)	5.53	5.53	5.46	5.60	0.00
Aroclor-1260-4	(4)	5.99	5.99	5.92	6.06	0.00
Aroclor-1260-5	(5)	6.24	6.24	6.17	6.31	0.00
TCX		2.78	2.78	2.73	2.83	0.00
DCB		7.55	7.55	7.45	7.65	0.00