

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1729 SAS No.: Q1729 SDG NO.: Q1729

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/17/2025 03/17/2025

Client Sample No.: CCAL01 Date Analyzed: 04/08/2025

Lab Sample No.: AR1660CCC400 Data File : PQ070129.D Time Analyzed: 10:20

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.498	4.386	4.586	230024400.000	205557000.000	-10.6
Aroclor-1016-2	4.516	4.404	4.604	387085800.000	330588000.000	-14.6
Aroclor-1016-3	4.575	4.463	4.663	252199800.000	262613000.000	4.1
Aroclor-1016-4	4.664	4.551	4.751	199753500.000	195212000.000	-2.3
Aroclor-1016-5	4.948	4.836	5.036	196065600.000	202282000.000	3.2
Aroclor-1260-1	6.037	5.926	6.126	342744100.000	365885000.000	6.8
Aroclor-1260-2	6.294	6.181	6.381	375376700.000	340035000.000	-9.4
Aroclor-1260-3	6.645	6.533	6.733	301270200.000	307704000.000	2.1
Aroclor-1260-4	6.859	6.746	6.946	326357800.000	328194000.000	0.6
Aroclor-1260-5	7.164	7.052	7.252	583624600.000	520209000.000	-10.9
DCB	8.518	8.403	8.603	3797495000.000	3380240000.000	-11.0
TCX	3.415	3.305	3.505	8939766000.000	9937660000.000	11.2

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1729 SAS No.: Q1729 SDG NO.: Q1729

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/17/2025 03/17/2025

Client Sample No.: CCAL01 Date Analyzed: 04/08/2025

Lab Sample No.: AR1660CCC400 Data File : PQ070129.D Time Analyzed: 10:20

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	3.786	3.687	3.887	223585300.000	196870000.000	-11.9
Aroclor-1016-2	3.801	3.702	3.902	365003900.000	427993000.000	17.3
Aroclor-1016-3	3.962	3.863	4.063	206649600.000	218776000.000	5.9
Aroclor-1016-4	4.009	3.911	4.111	189703100.000	220552000.000	16.3
Aroclor-1016-5	4.205	4.106	4.306	222979400.000	245560000.000	10.1
Aroclor-1260-1	5.200	5.100	5.300	400599300.000	441281000.000	10.2
Aroclor-1260-2	5.389	5.289	5.489	466852700.000	469377000.000	0.5
Aroclor-1260-3	5.531	5.431	5.631	444827800.000	460376000.000	3.5
Aroclor-1260-4	5.992	5.892	6.092	392153300.000	406057000.000	3.5
Aroclor-1260-5	6.236	6.136	6.336	828820000.000	792075000.000	-4.4
DCB	7.550	7.449	7.649	5861421000.000	6094740000.000	4.0
TCX	2.782	2.683	2.883	8065420000.000	9359340000.000	16.1