

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1729 SAS No.: Q1729 SDG NO.: Q1729

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/17/2025 03/17/2025

Client Sample No.: CCAL02 Date Analyzed: 04/08/2025

Lab Sample No.: AR1660CCC400 Data File : PQ070133.D Time Analyzed: 11:23

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.486	4.386	4.586	230024400.000	241024000.000	4.8
Aroclor-1016-2	4.504	4.404	4.604	387085800.000	409685000.000	5.8
Aroclor-1016-3	4.562	4.463	4.663	252199800.000	276408000.000	9.6
Aroclor-1016-4	4.651	4.551	4.751	199753500.000	216041000.000	8.2
Aroclor-1016-5	4.936	4.836	5.036	196065600.000	214693000.000	9.5
Aroclor-1260-1	6.026	5.926	6.126	342744100.000	375431000.000	9.5
Aroclor-1260-2	6.282	6.181	6.381	375376700.000	392611000.000	4.6
Aroclor-1260-3	6.633	6.533	6.733	301270200.000	326687000.000	8.4
Aroclor-1260-4	6.846	6.746	6.946	326357800.000	353657000.000	8.4
Aroclor-1260-5	7.152	7.052	7.252	583624600.000	603578000.000	3.4
DCB	8.504	8.403	8.603	3797495000.000	3907300000.000	2.9
TCX	3.405	3.305	3.505	8939766000.000	9968610000.000	11.5

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1729 SAS No.: Q1729 SDG NO.: Q1729

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/17/2025 03/17/2025

Client Sample No.: CCAL02 Date Analyzed: 04/08/2025

Lab Sample No.: AR1660CCC400 Data File : PQ070133.D Time Analyzed: 11:23

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	3.786	3.687	3.887	223585300.000	246358000.000	10.2
Aroclor-1016-2	3.800	3.702	3.902	365003900.000	388243000.000	6.4
Aroclor-1016-3	3.962	3.863	4.063	206649600.000	235638000.000	14.0
Aroclor-1016-4	4.009	3.911	4.111	189703100.000	221509000.000	16.8
Aroclor-1016-5	4.205	4.106	4.306	222979400.000	258516000.000	15.9
Aroclor-1260-1	5.199	5.100	5.300	400599300.000	466045000.000	16.3
Aroclor-1260-2	5.388	5.289	5.489	466852700.000	532649000.000	14.1
Aroclor-1260-3	5.530	5.431	5.631	444827800.000	506659000.000	13.9
Aroclor-1260-4	5.991	5.892	6.092	392153300.000	443980000.000	13.2
Aroclor-1260-5	6.235	6.136	6.336	828820000.000	917305000.000	10.7
DCB	7.548	7.449	7.649	5861421000.000	6535480000.000	11.5
TCX	2.783	2.683	2.883	8065420000.000	9501880000.000	17.8