

Surrogate Summary

SDG No.: Q1729

Client: Chemtech Consulting Group

Analytical Method: SFAM_PCB

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PQ070047.D	PIBLK-PQ070047.D	TCX	1	20	0.18	91		30	150
		DCB	1	40	0.38	94		30	150
		TCX	2	20	0.19	97		30	150
		DCB	2	40	0.39	98		30	150
I.BLK-PQ070128.D	PIBLK-PQ070128.D	TCX	1	20	0.19	97		30	150
		DCB	1	40	0.37	92		30	150
		TCX	2	20	0.22	109		30	150
		DCB	2	40	0.40	100		30	150
Q1729-13	PCB-GPC2-BLANK	TCX	1	20	0	0	*	30	150
		DCB	1	40	0	0	*	30	150
		TCX	2	20	0	0	*	30	150
		DCB	2	40	0	0	*	30	150
Q1729-14	PCB-GPC2-BLANK-SPIKE	TCX	1	20	0	0	*	30	150
		DCB	1	40	0	0	*	30	150
		TCX	2	20	0	0	*	30	150
		DCB	2	40	0	0	*	30	150
I.BLK-PQ070132.D	PIBLK-PQ070132.D	TCX	1	20	0.19	94		30	150
		DCB	1	40	0.36	91		30	150
		TCX	2	20	0.21	107		30	150
		DCB	2	40	0.40	100		30	150