

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1729 SAS No.: Q1729 SDG NO.: Q1729

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/21/2025 03/21/2025

Client Sample No.: CCAL01 Date Analyzed: 04/08/2025

Lab Sample No.: INDA367 Data File : PD088060.D Time Analyzed: 11:09

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
4,4'-DDD	6.710	6.640	6.780	2577180.000	2464230.000	-4.4
4,4'-DDT	7.026	6.956	7.096	2701560.000	2553010.000	-5.5
alpha-BHC	4.002	3.952	4.052	4110160.000	4046070.000	-1.6
DCB	9.080	8.980	9.180	3351440.000	3149370.000	-6.0
Dieldrin	6.352	6.282	6.422	3672580.000	3568680.000	-2.8
Endosulfan I	6.078	6.008	6.148	3344420.000	3240440.000	-3.1
Endrin	6.579	6.509	6.649	3065760.000	2981150.000	-2.8
gamma-BHC (Lindane)	4.333	4.283	4.383	3987070.000	3914400.000	-1.8
Heptachlor	4.934	4.884	4.984	4090350.000	3874640.000	-5.3
Methoxychlor	7.498	7.428	7.568	1489960.000	1419130.000	-4.8
TCX	3.553	3.503	3.603	1972170.000	2007160.000	1.8

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1729 SAS No.: Q1729 SDG NO.: Q1729

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/21/2025 03/21/2025

Client Sample No.: CCAL01 Date Analyzed: 04/08/2025

Lab Sample No.: INDA367 Data File : PD088060.D Time Analyzed: 11:09

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.936	5.866	6.006	16337800.000	15678900.000	-4.0
4,4'-DDT	6.190	6.120	6.260	16751800.000	15984600.000	-4.6
alpha-BHC	3.397	3.347	3.447	24236900.000	23482100.000	-3.1
DCB	8.081	7.981	8.181	17755800.000	16135700.000	-9.1
Dieldrin	5.519	5.449	5.589	20235100.000	19484100.000	-3.7
Endosulfan I	5.253	5.183	5.323	17350700.000	18041200.000	4.0
Endrin	5.795	5.725	5.865	18150200.000	17005200.000	-6.3
gamma-BHC (Lindane)	3.734	3.684	3.784	22254300.000	21638400.000	-2.8
Heptachlor	4.088	4.038	4.138	22826800.000	22183800.000	-2.8
Methoxychlor	6.761	6.691	6.831	7450680.000	7124550.000	-4.4
TCX	2.883	2.833	2.933	15371300.000	14763400.000	-4.0