

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: CHEM02
 Lab Code: CHEM Case No.: Q1729 SAS No.: Q1729 SDG No.: Q1729
 Instrument ID: MSVOA_X Calibration Date/Time: 04/09/2025 10:35
 Lab File ID: VX045664.D Init. Calib. Date(s): 04/01/2025 04/01/2025
 Heated Purge: (Y/N) N Init. Calib. Time(s): 17:06 19:02
 GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
t-1,3-Dichloropropene	50.000	49.05	1.903999	20