

Initial and Continuing Calibration Verification

Client: Nobis Group

SDG No.: Q1730

Project: Raymark Superfund Site

RunNo.: LB135346

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Cyanide	mg/L	0.098	0.099	99	90-110	04/08/2025
Sample ID: CCV1 Cyanide	mg/L	0.24	0.25	96	90-110	04/08/2025
Sample ID: CCV2 Cyanide	mg/L	0.24	0.25	96	90-110	04/08/2025
Sample ID: CCV3 Cyanide	mg/L	0.24	0.25	96	90-110	04/08/2025