



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Nobis Group SDG No.: Q1730
Contract: NOBI03 Lab Code: CHEM Case No.: Q1730 SAS No.: Q1730
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.21	0.2	106	40 - 160	CV	04/08/2025	09:52	LB135337
CRI01	Aluminum	91.7	100	92	40 - 160	P	04/14/2025	14:09	LB135424
	Antimony	51.7	50.0	103	40 - 160	P	04/14/2025	14:09	LB135424
	Arsenic	22.3	20.0	112	40 - 160	P	04/14/2025	14:09	LB135424
	Barium	95.6	100	96	40 - 160	P	04/14/2025	14:09	LB135424
	Beryllium	5.76	6.0	96	40 - 160	P	04/14/2025	14:09	LB135424
	Cadmium	6.08	6.0	101	40 - 160	P	04/14/2025	14:09	LB135424
	Calcium	1970	2000	98	40 - 160	P	04/14/2025	14:09	LB135424
	Chromium	9.67	10.0	97	40 - 160	P	04/14/2025	14:09	LB135424
	Cobalt	29.3	30.0	98	40 - 160	P	04/14/2025	14:09	LB135424
	Copper	22.7	20.0	113	40 - 160	P	04/14/2025	14:09	LB135424
	Iron	102	100	102	40 - 160	P	04/14/2025	14:09	LB135424
	Lead	12.2	12.0	102	40 - 160	P	04/14/2025	14:09	LB135424
	Magnesium	2060	2000	103	40 - 160	P	04/14/2025	14:09	LB135424
	Manganese	20.4	20.0	102	40 - 160	P	04/14/2025	14:09	LB135424
	Nickel	39.4	40.0	98	40 - 160	P	04/14/2025	14:09	LB135424
	Potassium	1930	2000	97	40 - 160	P	04/14/2025	14:09	LB135424
	Selenium	19.6	20.0	98	40 - 160	P	04/14/2025	14:09	LB135424
	Silver	9.88	10.0	99	40 - 160	P	04/14/2025	14:09	LB135424
	Sodium	1840	2000	92	40 - 160	P	04/14/2025	14:09	LB135424
	Thallium	41.8	40.0	104	40 - 160	P	04/14/2025	14:09	LB135424
	Vanadium	40.7	40.0	102	40 - 160	P	04/14/2025	14:09	LB135424
	Zinc	41.0	40.0	102	40 - 160	P	04/14/2025	14:09	LB135424