



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	<u>Nobis Group</u>				SDG No.:	<u>Q1730</u>				
Contract:	<u>NOBI03</u>	Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1730</u>	SAS No.:	<u>Q1730</u>			
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB83	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	04/10/2025	13:03	LB135365

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Nobis Group **SDG No.:** Q1730
Contract: NOBI03 **Lab Code:** CHEM **Case No.:** Q1730 **SAS No.:** Q1730

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB53	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	04/10/2025	13:08	LB135365
CCB54	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	04/10/2025	13:41	LB135365
CCB55	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	04/10/2025	14:08	LB135365
CCB56	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	04/10/2025	14:38	LB135365
CCB57	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	04/10/2025	15:08	LB135365
CCB58	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	04/10/2025	15:28	LB135365

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Nobis Group **SDG No.:** Q1730
Contract: NOBI03 **Lab Code:** CHEM **Case No.:** Q1730 **SAS No.:** Q1730

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Antimony	2.00	+/-2.00	U	0.25	2.00	P	04/10/2025	14:20	LB135405
	Arsenic	1.00	+/-1.00	U	0.25	1.00	P	04/10/2025	14:20	LB135405
	Barium	10.0	+/-10.0	U	1.25	10.0	P	04/10/2025	14:20	LB135405
	Beryllium	1.00	+/-1.00	U	0.75	1.00	P	04/10/2025	14:20	LB135405
	Cadmium	1.00	+/-1.00	U	0.50	1.00	P	04/10/2025	14:20	LB135405
	Chromium	2.00	+/-2.00	U	0.75	2.00	P	04/10/2025	14:20	LB135405
	Copper	2.00	+/-2.00	U	1.50	2.00	P	04/10/2025	14:20	LB135405
	Lead	1.00	+/-1.00	U	0.75	1.00	P	04/10/2025	14:20	LB135405
	Nickel	1.00	+/-1.00	U	0.75	1.00	P	04/10/2025	14:20	LB135405
	Selenium	5.00	+/-5.00	U	4.50	5.00	P	04/10/2025	14:20	LB135405
	Silver	1.00	+/-1.00	U	0.50	1.00	P	04/10/2025	14:20	LB135405
	Thallium	1.00	+/-1.00	U	0.50	1.00	P	04/10/2025	14:20	LB135405
	Vanadium	5.00	+/-5.00	U	0.25	5.00	P	04/10/2025	14:20	LB135405
	Zinc	5.00	+/-5.00	U	1.50	5.00	P	04/10/2025	14:20	LB135405

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Nobis Group</u>		SDG No.: <u>Q1730</u>								
Contract: <u>NOBI03</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1730</u>	SAS No.: <u>Q1730</u>							
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Antimony	2.00	+/-2.00	U	0.25	2.00	P	04/10/2025	14:58	LB135405
	Arsenic	1.00	+/-1.00	U	0.25	1.00	P	04/10/2025	14:58	LB135405
	Barium	10.0	+/-10.0	U	1.25	10.0	P	04/10/2025	14:58	LB135405
	Beryllium	1.00	+/-1.00	U	0.75	1.00	P	04/10/2025	14:58	LB135405
	Cadmium	1.00	+/-1.00	U	0.50	1.00	P	04/10/2025	14:58	LB135405
	Chromium	2.00	+/-2.00	U	0.75	2.00	P	04/10/2025	14:58	LB135405
	Copper	2.00	+/-2.00	U	1.50	2.00	P	04/10/2025	14:58	LB135405
	Lead	1.00	+/-1.00	U	0.75	1.00	P	04/10/2025	14:58	LB135405
	Nickel	1.00	+/-1.00	U	0.75	1.00	P	04/10/2025	14:58	LB135405
	Selenium	5.00	+/-5.00	U	4.50	5.00	P	04/10/2025	14:58	LB135405
	Silver	1.00	+/-1.00	U	0.50	1.00	P	04/10/2025	14:58	LB135405
	Thallium	1.00	+/-1.00	U	0.50	1.00	P	04/10/2025	14:58	LB135405
	Vanadium	5.00	+/-5.00	U	0.25	5.00	P	04/10/2025	14:58	LB135405
	Zinc	5.00	+/-5.00	U	1.50	5.00	P	04/10/2025	14:58	LB135405
CCB02	Antimony	2.00	+/-2.00	U	0.25	2.00	P	04/10/2025	16:14	LB135405
	Arsenic	1.00	+/-1.00	U	0.25	1.00	P	04/10/2025	16:14	LB135405
	Barium	10.0	+/-10.0	U	1.25	10.0	P	04/10/2025	16:14	LB135405
	Beryllium	1.00	+/-1.00	U	0.75	1.00	P	04/10/2025	16:14	LB135405
	Cadmium	1.00	+/-1.00	U	0.50	1.00	P	04/10/2025	16:14	LB135405
	Chromium	2.00	+/-2.00	U	0.75	2.00	P	04/10/2025	16:14	LB135405
	Copper	2.00	+/-2.00	U	1.50	2.00	P	04/10/2025	16:14	LB135405
	Lead	1.00	+/-1.00	U	0.75	1.00	P	04/10/2025	16:14	LB135405
	Nickel	1.00	+/-1.00	U	0.75	1.00	P	04/10/2025	16:14	LB135405
	Selenium	5.00	+/-5.00	U	4.50	5.00	P	04/10/2025	16:14	LB135405
	Silver	1.00	+/-1.00	U	0.50	1.00	P	04/10/2025	16:14	LB135405
	Thallium	1.00	+/-1.00	U	0.50	1.00	P	04/10/2025	16:14	LB135405
	Vanadium	5.00	+/-5.00	U	0.25	5.00	P	04/10/2025	16:14	LB135405
	Zinc	5.00	+/-5.00	U	1.50	5.00	P	04/10/2025	16:14	LB135405
CCB03	Antimony	2.00	+/-2.00	U	0.25	2.00	P	04/10/2025	16:49	LB135405
	Arsenic	1.00	+/-1.00	U	0.25	1.00	P	04/10/2025	16:49	LB135405
	Barium	10.0	+/-10.0	U	1.25	10.0	P	04/10/2025	16:49	LB135405
	Beryllium	1.00	+/-1.00	U	0.75	1.00	P	04/10/2025	16:49	LB135405
	Cadmium	1.00	+/-1.00	U	0.50	1.00	P	04/10/2025	16:49	LB135405
	Chromium	2.00	+/-2.00	U	0.75	2.00	P	04/10/2025	16:49	LB135405
	Copper	2.00	+/-2.00	U	1.50	2.00	P	04/10/2025	16:49	LB135405
	Lead	1.00	+/-1.00	U	0.75	1.00	P	04/10/2025	16:49	LB135405
	Nickel	1.00	+/-1.00	U	0.75	1.00	P	04/10/2025	16:49	LB135405
	Selenium	5.00	+/-5.00	U	4.50	5.00	P	04/10/2025	16:49	LB135405
	Silver	1.00	+/-1.00	U	0.50	1.00	P	04/10/2025	16:49	LB135405

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Nobis Group **SDG No.:** Q1730
Contract: NOBI03 **Lab Code:** CHEM **Case No.:** Q1730 **SAS No.:** Q1730

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB03	Thallium	1.00	+/-1.00	U	0.50	1.00	P	04/10/2025	16:49	LB135405
	Vanadium	5.00	+/-5.00	U	0.25	5.00	P	04/10/2025	16:49	LB135405
	Zinc	5.00	+/-5.00	U	1.50	5.00	P	04/10/2025	16:49	LB135405
CCB04	Antimony	2.00	+/-2.00	U	0.25	2.00	P	04/10/2025	17:23	LB135405
	Arsenic	1.00	+/-1.00	U	0.25	1.00	P	04/10/2025	17:23	LB135405
	Barium	10.0	+/-10.0	U	1.25	10.0	P	04/10/2025	17:23	LB135405
	Beryllium	1.00	+/-1.00	U	0.75	1.00	P	04/10/2025	17:23	LB135405
	Cadmium	1.00	+/-1.00	U	0.50	1.00	P	04/10/2025	17:23	LB135405
	Chromium	2.00	+/-2.00	U	0.75	2.00	P	04/10/2025	17:23	LB135405
	Copper	2.00	+/-2.00	U	1.50	2.00	P	04/10/2025	17:23	LB135405
	Lead	1.00	+/-1.00	U	0.75	1.00	P	04/10/2025	17:23	LB135405
	Nickel	1.00	+/-1.00	U	0.75	1.00	P	04/10/2025	17:23	LB135405
	Selenium	5.00	+/-5.00	U	4.50	5.00	P	04/10/2025	17:23	LB135405
	Silver	1.00	+/-1.00	U	0.50	1.00	P	04/10/2025	17:23	LB135405
	Thallium	1.00	+/-1.00	U	0.50	1.00	P	04/10/2025	17:23	LB135405
	Vanadium	5.00	+/-5.00	U	0.25	5.00	P	04/10/2025	17:23	LB135405
	Zinc	5.00	+/-5.00	U	1.50	5.00	P	04/10/2025	17:23	LB135405

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Nobis Group</u>		SDG No.: <u>Q1730</u>								
Contract: <u>NOBI03</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1730</u>			SAS No.: <u>Q1730</u>					
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Antimony	2.00	+/-2.00	U	0.25	2.00	P	04/14/2025	16:12	LB135426
	Arsenic	1.00	+/-1.00	U	0.25	1.00	P	04/14/2025	16:12	LB135426
	Barium	10.0	+/-10.0	U	1.25	10.0	P	04/14/2025	16:12	LB135426
	Beryllium	1.00	+/-1.00	U	0.75	1.00	P	04/14/2025	16:12	LB135426
	Cadmium	1.00	+/-1.00	U	0.50	1.00	P	04/14/2025	16:12	LB135426
	Chromium	2.00	+/-2.00	U	0.75	2.00	P	04/14/2025	16:12	LB135426
	Copper	2.00	+/-2.00	U	1.50	2.00	P	04/14/2025	16:12	LB135426
	Lead	1.00	+/-1.00	U	0.75	1.00	P	04/14/2025	16:12	LB135426
	Nickel	1.00	+/-1.00	U	0.75	1.00	P	04/14/2025	16:12	LB135426
	Selenium	5.00	+/-5.00	U	4.50	5.00	P	04/14/2025	16:12	LB135426
	Silver	1.00	+/-1.00	U	0.50	1.00	P	04/14/2025	16:12	LB135426
	Thallium	1.00	+/-1.00	U	0.50	1.00	P	04/14/2025	16:12	LB135426
	Vanadium	5.00	+/-5.00	U	0.25	5.00	P	04/14/2025	16:12	LB135426
	Zinc	5.00	+/-5.00	U	1.50	5.00	P	04/14/2025	16:12	LB135426

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Nobis Group</u>		SDG No.: <u>Q1730</u>								
Contract: <u>NOBI03</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1730</u>	SAS No.: <u>Q1730</u>							
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Antimony	2.00	+/-2.00	U	0.25	2.00	P	04/14/2025	16:44	LB135426
	Arsenic	1.00	+/-1.00	U	0.25	1.00	P	04/14/2025	16:44	LB135426
	Barium	10.0	+/-10.0	U	1.25	10.0	P	04/14/2025	16:44	LB135426
	Beryllium	1.00	+/-1.00	U	0.75	1.00	P	04/14/2025	16:44	LB135426
	Cadmium	1.00	+/-1.00	U	0.50	1.00	P	04/14/2025	16:44	LB135426
	Chromium	2.00	+/-2.00	U	0.75	2.00	P	04/14/2025	16:44	LB135426
	Copper	2.00	+/-2.00	U	1.50	2.00	P	04/14/2025	16:44	LB135426
	Lead	1.00	+/-1.00	U	0.75	1.00	P	04/14/2025	16:44	LB135426
	Nickel	1.00	+/-1.00	U	0.75	1.00	P	04/14/2025	16:44	LB135426
	Selenium	5.00	+/-5.00	U	4.50	5.00	P	04/14/2025	16:44	LB135426
	Silver	1.00	+/-1.00	U	0.50	1.00	P	04/14/2025	16:44	LB135426
	Thallium	0.060	+/-1.00	J	0.50	1.00	P	04/14/2025	16:44	LB135426
	Vanadium	5.00	+/-5.00	U	0.25	5.00	P	04/14/2025	16:44	LB135426
	Zinc	5.00	+/-5.00	U	1.50	5.00	P	04/14/2025	16:44	LB135426
CCB02	Antimony	2.00	+/-2.00	U	0.25	2.00	P	04/14/2025	17:20	LB135426
	Arsenic	1.00	+/-1.00	U	0.25	1.00	P	04/14/2025	17:20	LB135426
	Barium	10.0	+/-10.0	U	1.25	10.0	P	04/14/2025	17:20	LB135426
	Beryllium	1.00	+/-1.00	U	0.75	1.00	P	04/14/2025	17:20	LB135426
	Cadmium	1.00	+/-1.00	U	0.50	1.00	P	04/14/2025	17:20	LB135426
	Chromium	2.00	+/-2.00	U	0.75	2.00	P	04/14/2025	17:20	LB135426
	Copper	2.00	+/-2.00	U	1.50	2.00	P	04/14/2025	17:20	LB135426
	Lead	1.00	+/-1.00	U	0.75	1.00	P	04/14/2025	17:20	LB135426
	Nickel	1.00	+/-1.00	U	0.75	1.00	P	04/14/2025	17:20	LB135426
	Selenium	5.00	+/-5.00	U	4.50	5.00	P	04/14/2025	17:20	LB135426
	Silver	1.00	+/-1.00	U	0.50	1.00	P	04/14/2025	17:20	LB135426
	Thallium	1.00	+/-1.00	U	0.50	1.00	P	04/14/2025	17:20	LB135426
	Vanadium	5.00	+/-5.00	U	0.25	5.00	P	04/14/2025	17:20	LB135426
	Zinc	5.00	+/-5.00	U	1.50	5.00	P	04/14/2025	17:20	LB135426
CCB03	Antimony	2.00	+/-2.00	U	0.25	2.00	P	04/14/2025	17:39	LB135426
	Arsenic	1.00	+/-1.00	U	0.25	1.00	P	04/14/2025	17:39	LB135426
	Barium	10.0	+/-10.0	U	1.25	10.0	P	04/14/2025	17:39	LB135426
	Beryllium	1.00	+/-1.00	U	0.75	1.00	P	04/14/2025	17:39	LB135426
	Cadmium	1.00	+/-1.00	U	0.50	1.00	P	04/14/2025	17:39	LB135426
	Chromium	2.00	+/-2.00	U	0.75	2.00	P	04/14/2025	17:39	LB135426
	Copper	2.00	+/-2.00	U	1.50	2.00	P	04/14/2025	17:39	LB135426
	Lead	1.00	+/-1.00	U	0.75	1.00	P	04/14/2025	17:39	LB135426
	Nickel	1.00	+/-1.00	U	0.75	1.00	P	04/14/2025	17:39	LB135426
	Selenium	5.00	+/-5.00	U	4.50	5.00	P	04/14/2025	17:39	LB135426
	Silver	1.00	+/-1.00	U	0.50	1.00	P	04/14/2025	17:39	LB135426

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	<u>Nobis Group</u>				SDG No.:	<u>Q1730</u>				
Contract:	<u>NOBI03</u>	Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1730</u>	SAS No.:	<u>Q1730</u>			
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB03	Thallium	1.00	+/-1.00	U	0.50	1.00	P	04/14/2025	17:39	LB135426
	Vanadium	5.00	+/-5.00	U	0.25	5.00	P	04/14/2025	17:39	LB135426
	Zinc	5.00	+/-5.00	U	1.50	5.00	P	04/14/2025	17:39	LB135426