

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Nobis Group

SDG No.: Q1730

Instrument: CV1

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	LOD ug/L	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB167540BL		WATER		Batch Number:		PB167540		Prep Date:	04/10/2025	
	Mercury	0.20	<0.20	U	0.16	0.20	CV	04/10/2025	14:24	LB135365
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	LOD ug/L	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB167540TB		WATER		Batch Number:		PB167540		Prep Date:	04/10/2025	
	Mercury	0.20	<0.20	U	0.16	0.20	CV	04/10/2025	15:15	LB135365

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Nobis Group

SDG No.: Q1730

Instrument: P8

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	LOD ug/L	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB167502BL	WATER			Batch Number:		PB167502		Prep Date:	04/08/2025	
	Antimony	2.00	<2.00	U	0.25	2.00	P	04/10/2025	15:17	LB135405
	Arsenic	1.00	<1.00	U	0.25	1.00	P	04/10/2025	15:17	LB135405
	Barium	10.0	<10.0	U	1.25	10.0	P	04/10/2025	15:17	LB135405
	Beryllium	1.00	<1.00	U	0.75	1.00	P	04/10/2025	15:17	LB135405
	Cadmium	1.00	<1.00	U	0.50	1.00	P	04/10/2025	15:17	LB135405
	Chromium	2.00	<2.00	U	0.75	2.00	P	04/10/2025	15:17	LB135405
	Copper	2.00	<2.00	U	1.50	2.00	P	04/10/2025	15:17	LB135405
	Lead	1.00	<1.00	U	0.75	1.00	P	04/10/2025	15:17	LB135405
	Nickel	1.00	<1.00	U	0.75	1.00	P	04/10/2025	15:17	LB135405
	Selenium	5.00	<5.00	U	4.50	5.00	P	04/10/2025	15:17	LB135405
	Silver	1.00	<1.00	U	0.50	1.00	P	04/10/2025	15:17	LB135405
	Thallium	1.00	<1.00	U	0.50	1.00	P	04/10/2025	15:17	LB135405
	Vanadium	5.00	<5.00	U	0.25	5.00	P	04/10/2025	15:17	LB135405
	Zinc	5.00	<5.00	U	1.50	5.00	P	04/10/2025	15:17	LB135405
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	LOD ug/L	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB167502TB	WATER			Batch Number:		PB167502		Prep Date:	04/08/2025	
	Antimony	2.00	<2.00	U	0.25	2.00	P	04/10/2025	15:33	LB135405
	Arsenic	1.00	<1.00	U	0.25	1.00	P	04/10/2025	15:33	LB135405
	Barium	10.0	<10.0	U	1.25	10.0	P	04/10/2025	15:33	LB135405
	Beryllium	1.00	<1.00	U	0.75	1.00	P	04/10/2025	15:33	LB135405
	Cadmium	1.00	<1.00	U	0.50	1.00	P	04/10/2025	15:33	LB135405
	Chromium	2.00	<2.00	U	0.75	2.00	P	04/10/2025	15:33	LB135405
	Copper	2.00	<2.00	U	1.50	2.00	P	04/10/2025	15:33	LB135405
	Lead	1.00	<1.00	U	0.75	1.00	P	04/10/2025	15:33	LB135405
	Nickel	1.00	<1.00	U	0.75	1.00	P	04/10/2025	15:33	LB135405
	Selenium	5.00	<5.00	U	4.50	5.00	P	04/10/2025	15:33	LB135405
	Silver	1.00	<1.00	U	0.50	1.00	P	04/10/2025	15:33	LB135405
	Thallium	1.00	<1.00	U	0.50	1.00	P	04/10/2025	15:33	LB135405
	Vanadium	5.00	<5.00	U	0.25	5.00	P	04/10/2025	15:33	LB135405
	Zinc	5.00	<5.00	U	1.50	5.00	P	04/10/2025	15:33	LB135405