

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Nobis Group **SDG No.:** Q1730
Contract: NOBI03 **Lab Code:** CHEM **Case No.:** Q1730 **SAS No.:** Q1730

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167502BS							
Antimony	ug/L	500	514		103	85 - 117	P
Arsenic	ug/L	500	507		101	84 - 116	P
Barium	ug/L	2500	2560		102	86 - 114	P
Beryllium	ug/L	500	518		104	83 - 121	P
Cadmium	ug/L	500	522		104	87 - 115	P
Chromium	ug/L	500	503		101	85 - 116	P
Copper	ug/L	5000	5050		101	85 - 118	P
Lead	ug/L	2500	2470		99	88 - 115	P
Nickel	ug/L	500	501		100	85 - 117	P
Selenium	ug/L	500	493		99	80 - 120	P
Silver	ug/L	500	528		106	85 - 116	P
Thallium	ug/L	500	490		98	82 - 116	P
Vanadium	ug/L	500	507		101	86 - 115	P
Zinc	ug/L	5000	4970		99	83 - 119	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Nobis Group **SDG No.:** Q1730
Contract: NOBI03 **Lab Code:** CHEM **Case No.:** Q1730 **SAS No.:** Q1730

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167540BS Mercury	ug/L	4.0	4.05		101	82 - 119	CV