

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Nobis Group
Lab Code: CHEM Case no.: Q1730
Instrument ID: P8
Run Number: LB135405

Contract: NOBI03
Sas No.: Q1730 SDG No.: Q1730
Start Date : 04/10/2025
End Date : 04/10/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Non-Collision Cell									
			Element 165Ho	Q	Element 209Bi	Q	Element	Q	Element	Q	Element	Q
S0	S0	1155	100		100							
S2	S2	1158	100		100							
S3	S3	1205	101		100							
S4	S4	1208	102		100							
S5	S5	1211	101		99							
S6	S6	1214	99		94							
S7	S7	1217	98		90							
S8	S8	1219	90		78							
ICV01	ICV01	1358	101		100							
LLICV	LLICV	1416	97		97							
ICB01	ICB01	1420	99		98							
ICSA01	ICSA01	1430	96		89							
ICSAB01	ICSAB01	1451	98		91							
CCV01	CCV01	1454	92		84							
CCB01	CCB01	1458	100		100							
CRI	CRI	1502	101		100							
ZZZZZZ	ZZZZZZ	1507										
ZZZZZZ	ZZZZZZ	1512										
PB167502BL	PB167502BL	1517	100		100							
PB167502BS	PB167502BS	1522	99		93							
PB167502TB	PB167502TB	1533	101		101							
Q1730-02	OU4-VSL-15-C	1555	102		96							
Q1730-04	OU4-VSL-16-C	1559	102		95							
Q1730-06	OU4-VSL-17-C	1602	103		97							
Q1730-08	OU4-PCS-TC-2	1605	104		97							
CCV02	CCV02	1611	95		86							
CCB02	CCB02	1614	106		103							
Q1730-10	OU4-PCS-TC-2	1619	101		96							
Q1730-12	OU4-PCS-TC-2	1623	105		99							
Q1730-14	OU4-PCS-TC-2	1626	104		100							
Q1730-16	OU4-PCS-TC-2	1629	108		102							
Q1730-18	OU4-PCS-TC-2	1632	107		102							
Q1730-20	OU4-CF-15-04	1636	106		102							
Q1730-20DUP	OU4-CF-15-04	1639	106		101							

Internal Standard %RI Limit: 30 -120

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Nobis Group

Contract: NOBI03

Lab Code: CHEM Case no.: Q1730

Sas No.: Q1730 SDG No.: Q1730

Instrument ID: P8

Start Date : 04/10/2025

Run Number: LB135405

End Date : 04/10/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Non-Collision Cell									
			Element 165Ho	Q	Element 209Bi	Q	Element	Q	Element	Q	Element	Q
Q1730-20L	OU4-CF-15-04	1642	105		103							
CCV03	CCV03	1646	97		87							
CCB03	CCB03	1649	106		106							
Q1730-20MS	OU4-CF-15-04	1653	106		100							
Q1730-20MSD	OU4-CF-15-04	1656	103		98							
Q1730-20A	OU4-CF-15-04	1659	104		98							
Q1730-02	OU4-VSL-15-01	1702	104		100							
Q1730-04	OU4-VSL-16-01	1705	104		99							
Q1730-06	OU4-VSL-17-01	1708	106		102							
Q1730-12	OU4-PCS-TC-2	1711	106		100							
Q1730-16	OU4-PCS-TC-2	1715	104		100							
CCV04	CCV04	1719	98		88							
CCB04	CCB04	1723	106		105							

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Nobis Group

Contract: NOBI03

Lab Code: CHEM Case no.: Q1730

Sas No.: Q1730 SDG No.: Q1730

Instrument ID: P8

Start Date : 04/10/2025

Run Number: LB135405

End Date : 04/10/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Collision Cell									
			Element 209Bi	Q	Element	Q	Element	Q	Element	Q	Element	Q
S0	S0	1155	100									
S2	S2	1158	100									
S3	S3	1205	100									
S4	S4	1208	98									
S5	S5	1211	95									
S6	S6	1214	92									
S7	S7	1217	88									
S8	S8	1219	79									
ICV01	ICV01	1358	101									
LLICV	LLICV	1416	102									
ICB01	ICB01	1420	101									
ICSA01	ICSA01	1430	89									
ICSAB01	ICSAB01	1451	90									
CCV01	CCV01	1454	85									
CCB01	CCB01	1458	100									
CRI	CRI	1502	101									
ZZZZZZ	ZZZZZZ	1507										
ZZZZZZ	ZZZZZZ	1512										
PB167502BL	PB167502BL	1517	100									
PB167502BS	PB167502BS	1522	92									
PB167502TB	PB167502TB	1533	100									
Q1730-02	OU4-VSL-15-C	1555	95									
Q1730-04	OU4-VSL-16-C	1559	94									
Q1730-06	OU4-VSL-17-C	1602	96									
Q1730-08	OU4-PCS-TC-2	1605	95									
CCV02	CCV02	1611	86									
CCB02	CCB02	1614	102									
Q1730-10	OU4-PCS-TC-2	1619	95									
Q1730-12	OU4-PCS-TC-2	1623	95									
Q1730-14	OU4-PCS-TC-2	1626	98									
Q1730-16	OU4-PCS-TC-2	1629	101									
Q1730-18	OU4-PCS-TC-2	1632	97									
Q1730-20	OU4-CF-15-04	1636	98									
Q1730-20DUP	OU4-CF-15-04	1639	96									

Internal Standard %RI Limit: 30 -120

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Nobis Group

Contract: NOBI03

Lab Code: CHEM Case no.: Q1730

Sas No.: Q1730 SDG No.: Q1730

Instrument ID: P8

Start Date : 04/10/2025

Run Number: LB135405

End Date : 04/10/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Collision Cell									
			Element 209Bi	Q	Element	Q	Element	Q	Element	Q	Element	Q
Q1730-20L	OU4-CF-15-04	1642	99									
CCV03	CCV03	1646	85									
CCB03	CCB03	1649	101									
Q1730-20MS	OU4-CF-15-04	1653	95									
Q1730-20MSD	OU4-CF-15-04	1656	95									
Q1730-20A	OU4-CF-15-04	1659	95									
Q1730-02	OU4-VSL-15-01	1702	99									
Q1730-04	OU4-VSL-16-01	1705	99									
Q1730-06	OU4-VSL-17-01	1708	99									
Q1730-12	OU4-PCS-TC-2	1711	99									
Q1730-16	OU4-PCS-TC-2	1715	100									
CCV04	CCV04	1719	85									
CCB04	CCB04	1723	103									

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Nobis Group

Contract: NOBI03

Lab Code: CHEM Case no.: Q1730

Sas No.: Q1730 SDG No.: Q1730

Instrument ID: P8

Start Date : 04/14/2025

Run Number: LB135426

End Date : 04/14/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Non-Collision Cell									
			Element 165Ho	Q	Element 209Bi	Q	Element	Q	Element	Q	Element	Q
S0	S0	1505	100		100							
S2	S2	1512	99		100							
S3	S3	1515	100		99							
S4	S4	1518	100		99							
S5	S5	1521	99		98							
S6	S6	1524	98		98							
S7	S7	1527	99		92							
S8	S8	1530	94		82							
ICV01	ICV01	1556	104		105							
LLICV	LLICV	1609	103		104							
ICB01	ICB01	1612	104		104							
ICSA01	ICSA01	1623	99		93							
ICSAB01	ICSAB01	1638	104		95							
CCV01	CCV01	1641	99		87							
CCB01	CCB01	1644	105		104							
CRI	CRI	1647	104		103							
Q1730-06	OU4-VSL-17-C	1651	106		103							
Q1730-08	OU4-PCS-TC-2	1655	104		104							
Q1730-10	OU4-PCS-TC-2	1658	106		103							
Q1730-18	OU4-PCS-TC-2	1701	106		100							
Q1730-20	OU4-CF-15-04	1704	106		102							
Q1730-20DUP	OU4-CF-15-04	1708	95		89							
Q1730-20L	OU4-CF-15-04	1711	104		102							
CCV02	CCV02	1716	95		85							
CCB02	CCB02	1720	102		100							
Q1730-20MS	OU4-CF-15-04	1724	106		102							
Q1730-20MSD	OU4-CF-15-04	1727	105		100							
Q1730-20A	OU4-CF-15-04	1730	104		102							
CCV03	CCV03	1735	97		87							
CCB03	CCB03	1739	106		105							

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Nobis Group

Contract: NOBI03

Lab Code: CHEM Case no.: Q1730

Sas No.: Q1730 SDG No.: Q1730

Instrument ID: P8

Start Date : 04/14/2025

Run Number: LB135426

End Date : 04/14/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Collision Cell									
			Element 209Bi	Q	Element	Q	Element	Q	Element	Q	Element	Q
S0	S0	1505	100									
S2	S2	1512	101									
S3	S3	1515	98									
S4	S4	1518	96									
S5	S5	1521	94									
S6	S6	1524	91									
S7	S7	1527	87									
S8	S8	1530	80									
ICV01	ICV01	1556	103									
LLICV	LLICV	1609	103									
ICB01	ICB01	1612	102									
ICSA01	ICSA01	1623	92									
ICSAB01	ICSAB01	1638	92									
CCV01	CCV01	1641	87									
CCB01	CCB01	1644	104									
CRI	CRI	1647	102									
Q1730-06	OU4-VSL-17-C	1651	103									
Q1730-08	OU4-PCS-TC-2	1655	99									
Q1730-10	OU4-PCS-TC-2	1658	100									
Q1730-18	OU4-PCS-TC-2	1701	98									
Q1730-20	OU4-CF-15-04	1704	99									
Q1730-20DUP	OU4-CF-15-04	1708	97									
Q1730-20L	OU4-CF-15-04	1711	100									
CCV02	CCV02	1716	85									
CCB02	CCB02	1720	102									
Q1730-20MS	OU4-CF-15-04	1724	98									
Q1730-20MSD	OU4-CF-15-04	1727	99									
Q1730-20A	OU4-CF-15-04	1730	98									
CCV03	CCV03	1735	86									
CCB03	CCB03	1739	103									