

RETENTION TIMES OF INITIAL CALIBRATION

Contract: NOBI03

Lab Code: CHEM Case No.: Q1730 SAS No.: Q1730 SDG NO.: Q1730

Instrument ID: ECD_S Calibration Date(s): 04/02/2025 04/02/2025

Calibration Times: 17:32 20:44

GC Column: RTX-CLP ID: 0.32 (mm)

LAB FILE ID:	RT 200 = <u>PS029657.D</u>	RT 500 = <u>PS029658.D</u>
	RT 750 = <u>PS029659.D</u>	RT 1000 = <u>PS029660.D</u>
		RT 1500 = <u>PS029661.D</u>

COMPOUND	RT 200	RT 500	RT 750	RT 1000	RT 1500	MEAN RT	RT WINDOW	
							FROM	TO
2,4,5-T	9.17	9.17	9.17	9.17	9.17	9.17	9.07	9.27
2,4,5-TP(Silvex)	8.89	8.89	8.89	8.89	8.89	8.89	8.79	8.99
2,4-D	8.04	8.04	8.04	8.04	8.04	8.04	7.94	8.14
2,4-DB	9.73	9.73	9.73	9.73	9.73	9.73	9.63	9.83
2,4-DCAA	6.96	6.96	6.96	6.96	6.96	6.96	6.86	7.06
Dalapon	2.46	2.46	2.46	2.46	2.46	2.46	2.36	2.56
DICAMBA	7.14	7.14	7.14	7.14	7.14	7.14	7.04	7.24
DICHLORPROP	7.82	7.82	7.82	7.82	7.82	7.82	7.72	7.92
Dinoseb	10.90	10.90	10.90	10.90	10.89	10.90	10.80	11.00



Contract:	<u>NOBI03</u>						
Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1730</u>	SAS No.:	<u>Q1730</u>	SDG NO.:	<u>Q1730</u>
Instrument ID:	<u>ECD_S</u>	Calibration Date(s):		<u>04/02/2025</u>	<u>04/02/2025</u>		
		Calibration Times:		<u>17:32</u>	<u>20:44</u>		

LAB FILE ID:	RT 200 =	<u>PS029657.D</u>	RT 500 =	<u>PS029658.D</u>
RT 750 =	<u>PS029659.D</u>	RT 1000 =	<u>PS029660.D</u>	RT 1500 =
			<u>PS029661.D</u>	

[illegible]