

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1730 SAS No.: Q1730 SDG NO.: Q1730

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 04/02/2025 04/02/2025

Client Sample No.: CCAL02 Date Analyzed: 04/08/2025

Lab Sample No.: HSTDCCC750 Data File : PS029739.D Time Analyzed: 21:27

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.157	9.070	9.270	737.660	712.500	3.5
2,4,5-TP(Silvex)	8.876	8.788	8.988	724.930	712.500	1.7
2,4-D	8.031	7.942	8.142	692.190	705.000	-1.8
2,4-DB	9.717	9.629	9.829	780.120	712.500	9.5
2,4-DCAA	6.953	6.863	7.063	740.580	750.000	-1.3
Dalapon	2.460	2.363	2.563	659.060	682.500	-3.4
DICAMBA	7.130	7.040	7.240	717.020	705.000	1.7
DICHLORPROP	7.811	7.722	7.922	709.750	705.000	0.7
Dinoseb	10.881	10.795	10.995	705.590	705.000	0.1

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1730 SAS No.: Q1730 SDG NO.: Q1730

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/02/2025 04/02/2025

Client Sample No.: CCAL02 Date Analyzed: 04/08/2025

Lab Sample No.: HSTDCCC750 Data File : PS029739.D Time Analyzed: 21:27

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.946	9.857	10.057	740.610	712.500	3.9
2,4,5-TP(Silvex)	9.542	9.451	9.651	728.880	712.500	2.3
2,4-D	8.663	8.571	8.771	674.960	705.000	-4.3
2,4-DB	10.504	10.414	10.614	669.740	712.500	-6.0
2,4-DCAA	7.469	7.375	7.575	730.420	750.000	-2.6
Dalapon	2.531	2.433	2.633	611.880	682.500	-10.3
DICAMBA	7.656	7.562	7.762	706.430	705.000	0.2
DICHLORPROP	8.350	8.256	8.456	712.620	705.000	1.1
Dinoseb	10.875	10.787	10.987	688.400	705.000	-2.4