

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1730 SAS No.: Q1730 SDG NO.: Q1730

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 04/02/2025 04/02/2025

Client Sample No.: CCAL03 Date Analyzed: 04/09/2025

Lab Sample No.: HSTDCCC750 Data File : PS029750.D Time Analyzed: 02:39

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.154	9.070	9.270	727.850	712.500	2.2
2,4,5-TP(Silvex)	8.873	8.788	8.988	709.180	712.500	-0.5
2,4-D	8.028	7.942	8.142	682.610	705.000	-3.2
2,4-DB	9.713	9.629	9.829	758.050	712.500	6.4
2,4-DCAA	6.950	6.863	7.063	730.060	750.000	-2.7
Dalapon	2.460	2.363	2.563	644.420	682.500	-5.6
DICAMBA	7.128	7.040	7.240	710.570	705.000	0.8
DICHLORPROP	7.809	7.722	7.922	696.280	705.000	-1.2
Dinoseb	10.878	10.795	10.995	696.900	705.000	-1.1

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1730 SAS No.: Q1730 SDG NO.: Q1730

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/02/2025 04/02/2025

Client Sample No.: CCAL03 Date Analyzed: 04/09/2025

Lab Sample No.: HSTDCCC750 Data File : PS029750.D Time Analyzed: 02:39

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.944	9.857	10.057	720.370	712.500	1.1
2,4,5-TP(Silvex)	9.539	9.451	9.651	709.890	712.500	-0.4
2,4-D	8.660	8.571	8.771	658.990	705.000	-6.5
2,4-DB	10.500	10.414	10.614	670.930	712.500	-5.8
2,4-DCAA	7.467	7.375	7.575	718.080	750.000	-4.3
Dalapon	2.530	2.433	2.633	599.100	682.500	-12.2
DICAMBA	7.654	7.562	7.762	690.000	705.000	-2.1
DICHLORPROP	8.347	8.256	8.456	688.180	705.000	-2.4
Dinoseb	10.872	10.787	10.987	679.230	705.000	-3.7