

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1730 SAS No.: Q1730 SDG NO.: Q1730

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 04/02/2025 04/02/2025

Client Sample No.: CCAL04 Date Analyzed: 04/09/2025

Lab Sample No.: HSTDCCC750 Data File : PS029753.D Time Analyzed: 11:33

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.152	9.070	9.270	762.220	712.500	7.0
2,4,5-TP(Silvex)	8.870	8.788	8.988	746.700	712.500	4.8
2,4-D	8.026	7.942	8.142	715.020	705.000	1.4
2,4-DB	9.710	9.629	9.829	806.200	712.500	13.2
2,4-DCAA	6.948	6.863	7.063	753.720	750.000	0.5
Dalapon	2.459	2.363	2.563	662.470	682.500	-2.9
DICAMBA	7.126	7.040	7.240	748.480	705.000	6.2
DICHLORPROP	7.806	7.722	7.922	728.070	705.000	3.3
Dinoseb	10.874	10.795	10.995	719.380	705.000	2.0

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1730 SAS No.: Q1730 SDG NO.: Q1730

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/02/2025 04/02/2025

Client Sample No.: CCAL04 Date Analyzed: 04/09/2025

Lab Sample No.: HSTDCCC750 Data File : PS029753.D Time Analyzed: 11:33

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.941	9.857	10.057	755.590	712.500	6.0
2,4,5-TP(Silvex)	9.537	9.451	9.651	745.880	712.500	4.7
2,4-D	8.658	8.571	8.771	691.050	705.000	-2.0
2,4-DB	10.498	10.414	10.614	708.720	712.500	-0.5
2,4-DCAA	7.465	7.375	7.575	750.790	750.000	0.1
Dalapon	2.529	2.433	2.633	623.640	682.500	-8.6
DICAMBA	7.652	7.562	7.762	724.520	705.000	2.8
DICHLORPROP	8.345	8.256	8.456	719.760	705.000	2.1
Dinoseb	10.870	10.787	10.987	708.680	705.000	0.5