

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1730 SAS No.: Q1730 SDG NO.: Q1730

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 04/02/2025 04/02/2025

Client Sample No.: CCAL05 Date Analyzed: 04/09/2025

Lab Sample No.: HSTDCCC750 Data File : PS029764.D Time Analyzed: 20:34

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.144	9.070	9.270	774.100	712.500	8.6
2,4,5-TP(Silvex)	8.863	8.788	8.988	735.080	712.500	3.2
2,4-D	8.020	7.942	8.142	723.340	705.000	2.6
2,4-DB	9.702	9.629	9.829	809.890	712.500	13.7
2,4-DCAA	6.943	6.863	7.063	740.470	750.000	-1.3
Dalapon	2.459	2.363	2.563	669.240	682.500	-1.9
DICAMBA	7.120	7.040	7.240	753.600	705.000	6.9
DICHLORPROP	7.800	7.722	7.922	734.710	705.000	4.2
Dinoseb	10.866	10.795	10.995	725.530	705.000	2.9

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1730 SAS No.: Q1730 SDG NO.: Q1730

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/02/2025 04/02/2025

Client Sample No.: CCAL05 Date Analyzed: 04/09/2025

Lab Sample No.: HSTDCCC750 Data File : PS029764.D Time Analyzed: 20:34

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.942	9.857	10.057	776.110	712.500	8.9
2,4,5-TP(Silvex)	9.538	9.451	9.651	765.850	712.500	7.5
2,4-D	8.659	8.571	8.771	705.850	705.000	0.1
2,4-DB	10.499	10.414	10.614	715.490	712.500	0.4
2,4-DCAA	7.467	7.375	7.575	764.180	750.000	1.9
Dalapon	2.530	2.433	2.633	618.550	682.500	-9.4
DICAMBA	7.653	7.562	7.762	742.830	705.000	5.4
DICHLORPROP	8.346	8.256	8.456	741.610	705.000	5.2
Dinoseb	10.871	10.787	10.987	725.420	705.000	2.9