

### CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1730 SAS No.: Q1730 SDG NO.: Q1730

Continuing Calib Date: 04/07/2025 Initial Calibration Date(s): 03/18/2025 03/18/2025

Continuing Calib Time: 10:23 Initial Calibration Time(s): 14:03 22:15

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1248-1       | (1) | 4.78       | 4.79      | 4.69      | 4.89 | 0.01       |
| Aroclor-1248-2       | (2) | 5.02       | 5.02      | 4.92      | 5.12 | 0.00       |
| Aroclor-1248-3       | (3) | 5.24       | 5.24      | 5.14      | 5.34 | 0.00       |
| Aroclor-1248-4       | (4) | 5.59       | 5.59      | 5.49      | 5.69 | 0.00       |
| Aroclor-1248-5       | (5) | 5.63       | 5.63      | 5.53      | 5.73 | 0.00       |
| Tetrachloro-m-xylene |     | 3.69       | 3.69      | 3.59      | 3.79 | 0.00       |
| Decachlorobiphenyl   |     | 8.74       | 8.74      | 8.64      | 8.84 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1730 SAS No.: Q1730 SDG NO.: Q1730

Continuing Calib Date: 04/07/2025 Initial Calibration Date(s): 03/18/2025 03/18/2025

Continuing Calib Time: 10:23 Initial Calibration Time(s): 14:03 22:15

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1248-1       | (1) | 4.77       | 4.77      | 4.67      | 4.87 | 0.00       |
| Aroclor-1248-2       | (2) | 5.01       | 5.01      | 4.91      | 5.11 | 0.00       |
| Aroclor-1248-3       | (3) | 5.05       | 5.05      | 4.95      | 5.15 | 0.00       |
| Aroclor-1248-4       | (4) | 5.22       | 5.22      | 5.12      | 5.32 | 0.00       |
| Aroclor-1248-5       | (5) | 5.61       | 5.61      | 5.51      | 5.71 | 0.00       |
| Tetrachloro-m-xylene |     | 3.69       | 3.69      | 3.59      | 3.79 | 0.00       |
| Decachlorobiphenyl   |     | 8.69       | 8.70      | 8.60      | 8.80 | 0.01       |