

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1730 SAS No.: Q1730 SDG NO.: Q1730

Continuing Calib Date: 04/08/2025 Initial Calibration Date(s): 03/18/2025 03/18/2025

Continuing Calib Time: 12:58 Initial Calibration Time(s): 14:03 22:15

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.78	4.79	4.69	4.89	0.01
Aroclor-1016-2	(2)	4.80	4.80	4.70	4.90	0.00
Aroclor-1016-3	(3)	4.86	4.86	4.76	4.96	0.00
Aroclor-1016-4	(4)	4.98	4.98	4.88	5.08	0.00
Aroclor-1016-5	(5)	5.24	5.24	5.14	5.34	0.00
Aroclor-1260-1	(1)	6.28	6.28	6.18	6.38	0.00
Aroclor-1260-2	(2)	6.47	6.47	6.37	6.57	0.00
Aroclor-1260-3	(3)	6.83	6.84	6.74	6.94	0.01
Aroclor-1260-4	(4)	7.09	7.10	7.00	7.20	0.01
Aroclor-1260-5	(5)	7.34	7.34	7.24	7.44	0.00
Tetrachloro-m-xylene		3.69	3.69	3.59	3.79	0.00
Decachlorobiphenyl		8.74	8.75	8.65	8.85	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1730 SAS No.: Q1730 SDG NO.: Q1730

Continuing Calib Date: 04/08/2025 Initial Calibration Date(s): 03/18/2025 03/18/2025

Continuing Calib Time: 12:58 Initial Calibration Time(s): 14:03 22:15

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.77	4.77	4.67	4.87	0.00
Aroclor-1016-2	(2)	4.79	4.79	4.69	4.89	0.00
Aroclor-1016-3	(3)	4.96	4.97	4.87	5.07	0.01
Aroclor-1016-4	(4)	5.01	5.01	4.91	5.11	0.00
Aroclor-1016-5	(5)	5.22	5.22	5.12	5.32	0.00
Aroclor-1260-1	(1)	6.25	6.26	6.16	6.36	0.01
Aroclor-1260-2	(2)	6.44	6.44	6.34	6.54	0.00
Aroclor-1260-3	(3)	6.59	6.60	6.50	6.70	0.01
Aroclor-1260-4	(4)	7.06	7.07	6.97	7.17	0.01
Aroclor-1260-5	(5)	7.30	7.31	7.21	7.41	0.01
Tetrachloro-m-xylene		3.69	3.69	3.59	3.79	0.00
Decachlorobiphenyl		8.69	8.70	8.60	8.80	0.01