

### CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1730 SAS No.: Q1730 SDG NO.: Q1730

Continuing Calib Date: 04/08/2025 Initial Calibration Date(s): 03/27/2025 03/27/2025

Continuing Calib Time: 21:17 Initial Calibration Time(s): 10:30 17:53

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|-------|------------|
|                      |     |            |           | FROM      | TO    |            |
| Aroclor-1242-1       | (1) | 5.67       | 5.66      | 5.56      | 5.76  | -0.01      |
| Aroclor-1242-2       | (2) | 5.69       | 5.69      | 5.59      | 5.79  | 0.00       |
| Aroclor-1242-3       | (3) | 5.76       | 5.75      | 5.65      | 5.85  | -0.01      |
| Aroclor-1242-4       | (4) | 5.85       | 5.85      | 5.75      | 5.95  | 0.00       |
| Aroclor-1242-5       | (5) | 6.59       | 6.58      | 6.48      | 6.68  | -0.01      |
| Tetrachloro-m-xylene |     | 4.52       | 4.51      | 4.41      | 4.61  | -0.01      |
| Decachlorobiphenyl   |     | 10.24      | 10.23     | 10.13     | 10.33 | -0.01      |

### CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1730 SAS No.: Q1730 SDG NO.: Q1730

Continuing Calib Date: 04/08/2025 Initial Calibration Date(s): 03/27/2025 03/27/2025

Continuing Calib Time: 21:17 Initial Calibration Time(s): 10:30 17:53

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1242-1       | (1) | 4.90       | 4.90      | 4.80      | 5.00 | 0.00       |
| Aroclor-1242-2       | (2) | 4.92       | 4.92      | 4.82      | 5.02 | 0.00       |
| Aroclor-1242-3       | (3) | 5.10       | 5.10      | 5.00      | 5.20 | 0.00       |
| Aroclor-1242-4       | (4) | 5.18       | 5.18      | 5.08      | 5.28 | 0.00       |
| Aroclor-1242-5       | (5) | 5.71       | 5.71      | 5.61      | 5.81 | 0.01       |
| Tetrachloro-m-xylene |     | 3.82       | 3.82      | 3.72      | 3.92 | 0.00       |
| Decachlorobiphenyl   |     | 8.86       | 8.86      | 8.76      | 8.96 | 0.00       |