

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1730 SAS No.: Q1730 SDG NO.: Q1730

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/18/2025 03/18/2025

Client Sample No.: CCAL01 Date Analyzed: 04/07/2025

Lab Sample No.: AR1660CCC500 Data File : PO110263.D Time Analyzed: 09:46

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.784	4.685	4.885	533.050	500.000	6.6
Aroclor-1016-2	4.803	4.704	4.904	532.950	500.000	6.6
Aroclor-1016-3	4.860	4.761	4.961	534.220	500.000	6.8
Aroclor-1016-4	4.980	4.881	5.081	533.580	500.000	6.7
Aroclor-1016-5	5.238	5.139	5.339	534.560	500.000	6.9
Aroclor-1260-1	6.279	6.180	6.380	540.840	500.000	8.2
Aroclor-1260-2	6.467	6.369	6.569	507.690	500.000	1.5
Aroclor-1260-3	6.835	6.737	6.937	498.120	500.000	-0.4
Aroclor-1260-4	7.096	6.998	7.198	486.870	500.000	-2.6
Aroclor-1260-5	7.337	7.239	7.439	480.250	500.000	-4.0
Decachlorobiphenyl	8.742	8.645	8.845	57.530	50.000	15.1
Tetrachloro-m-xylene	3.692	3.592	3.792	53.720	50.000	7.4

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1730 SAS No.: Q1730 SDG NO.: Q1730

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/18/2025 03/18/2025

Client Sample No.: CCAL01 Date Analyzed: 04/07/2025

Lab Sample No.: AR1660CCC500 Data File : PO110263.D Time Analyzed: 09:46

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.769	4.672	4.872	509.410	500.000	1.9
Aroclor-1016-2	4.788	4.691	4.891	506.900	500.000	1.4
Aroclor-1016-3	4.964	4.867	5.067	511.350	500.000	2.3
Aroclor-1016-4	5.005	4.909	5.109	504.600	500.000	0.9
Aroclor-1016-5	5.219	5.122	5.322	510.930	500.000	2.2
Aroclor-1260-1	6.250	6.155	6.355	495.180	500.000	-1.0
Aroclor-1260-2	6.438	6.342	6.542	487.390	500.000	-2.5
Aroclor-1260-3	6.591	6.495	6.695	473.050	500.000	-5.4
Aroclor-1260-4	7.062	6.967	7.167	467.970	500.000	-6.4
Aroclor-1260-5	7.302	7.207	7.407	455.220	500.000	-9.0
Decachlorobiphenyl	8.690	8.596	8.796	48.910	50.000	-2.2
Tetrachloro-m-xylene	3.687	3.590	3.790	52.660	50.000	5.3