

### CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1730 SAS No.: Q1730 SDG NO.: Q1730

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/18/2025 03/18/2025

Client Sample No.: CCAL02 Date Analyzed: 04/07/2025

Lab Sample No.: AR1660CCC500 Data File : PO110278.D Time Analyzed: 17:14

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|----------------------|-------|-----------|-------|--------------------|-------------------|-------|
|                      |       | FROM      | TO    |                    |                   |       |
| Aroclor-1016-1       | 4.782 | 4.685     | 4.885 | 511.900            | 500.000           | 2.4   |
| Aroclor-1016-2       | 4.802 | 4.704     | 4.904 | 513.800            | 500.000           | 2.8   |
| Aroclor-1016-3       | 4.858 | 4.761     | 4.961 | 513.560            | 500.000           | 2.7   |
| Aroclor-1016-4       | 4.979 | 4.881     | 5.081 | 505.500            | 500.000           | 1.1   |
| Aroclor-1016-5       | 5.236 | 5.139     | 5.339 | 483.610            | 500.000           | -3.3  |
| Aroclor-1260-1       | 6.276 | 6.180     | 6.380 | 516.970            | 500.000           | 3.4   |
| Aroclor-1260-2       | 6.466 | 6.369     | 6.569 | 486.490            | 500.000           | -2.7  |
| Aroclor-1260-3       | 6.834 | 6.737     | 6.937 | 450.780            | 500.000           | -9.8  |
| Aroclor-1260-4       | 7.094 | 6.998     | 7.198 | 442.100            | 500.000           | -11.6 |
| Aroclor-1260-5       | 7.336 | 7.239     | 7.439 | 429.820            | 500.000           | -14.0 |
| Decachlorobiphenyl   | 8.741 | 8.645     | 8.845 | 52.440             | 50.000            | 4.9   |
| Tetrachloro-m-xylene | 3.691 | 3.592     | 3.792 | 52.070             | 50.000            | 4.1   |

### CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1730 SAS No.: Q1730 SDG NO.: Q1730

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/18/2025 03/18/2025

Client Sample No.: CCAL02 Date Analyzed: 04/07/2025

Lab Sample No.: AR1660CCC500 Data File : PO110278.D Time Analyzed: 17:14

| COMPOUND             | RT    | RT WINDOW<br>FROM TO |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|----------------------|-------|----------------------|-------|--------------------|-------------------|-------|
| Aroclor-1016-1       | 4.769 | 4.672                | 4.872 | 501.230            | 500.000           | 0.2   |
| Aroclor-1016-2       | 4.788 | 4.691                | 4.891 | 505.970            | 500.000           | 1.2   |
| Aroclor-1016-3       | 4.964 | 4.867                | 5.067 | 506.650            | 500.000           | 1.3   |
| Aroclor-1016-4       | 5.006 | 4.909                | 5.109 | 501.000            | 500.000           | 0.2   |
| Aroclor-1016-5       | 5.218 | 5.122                | 5.322 | 500.710            | 500.000           | 0.1   |
| Aroclor-1260-1       | 6.250 | 6.155                | 6.355 | 459.750            | 500.000           | -8.1  |
| Aroclor-1260-2       | 6.438 | 6.342                | 6.542 | 450.220            | 500.000           | -10.0 |
| Aroclor-1260-3       | 6.591 | 6.495                | 6.695 | 435.550            | 500.000           | -12.9 |
| Aroclor-1260-4       | 7.063 | 6.967                | 7.167 | 432.750            | 500.000           | -13.5 |
| Aroclor-1260-5       | 7.303 | 7.207                | 7.407 | 419.150            | 500.000           | -16.2 |
| Decachlorobiphenyl   | 8.691 | 8.596                | 8.796 | 46.440             | 50.000            | -7.1  |
| Tetrachloro-m-xylene | 3.688 | 3.590                | 3.790 | 52.230             | 50.000            | 4.5   |