

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1730 SAS No.: Q1730 SDG NO.: Q1730

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/18/2025 03/18/2025

Client Sample No.: CCAL03 Date Analyzed: 04/07/2025

Lab Sample No.: AR1660CCC500 Data File : PO110286.D Time Analyzed: 21:27

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.783	4.685	4.885	528.320	500.000	5.7
Aroclor-1016-2	4.802	4.704	4.904	526.450	500.000	5.3
Aroclor-1016-3	4.858	4.761	4.961	529.370	500.000	5.9
Aroclor-1016-4	4.979	4.881	5.081	518.990	500.000	3.8
Aroclor-1016-5	5.236	5.139	5.339	496.090	500.000	-0.8
Aroclor-1260-1	6.277	6.180	6.380	550.160	500.000	10.0
Aroclor-1260-2	6.466	6.369	6.569	508.400	500.000	1.7
Aroclor-1260-3	6.834	6.737	6.937	481.740	500.000	-3.7
Aroclor-1260-4	7.094	6.998	7.198	471.490	500.000	-5.7
Aroclor-1260-5	7.336	7.239	7.439	462.920	500.000	-7.4
Decachlorobiphenyl	8.740	8.645	8.845	54.690	50.000	9.4
Tetrachloro-m-xylene	3.691	3.592	3.792	53.480	50.000	7.0

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1730 SAS No.: Q1730 SDG NO.: Q1730

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/18/2025 03/18/2025

Client Sample No.: CCAL03 Date Analyzed: 04/07/2025

Lab Sample No.: AR1660CCC500 Data File : PO110286.D Time Analyzed: 21:27

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.769	4.672	4.872	512.420	500.000	2.5
Aroclor-1016-2	4.788	4.691	4.891	514.070	500.000	2.8
Aroclor-1016-3	4.964	4.867	5.067	515.160	500.000	3.0
Aroclor-1016-4	5.006	4.909	5.109	469.250	500.000	-6.2
Aroclor-1016-5	5.218	5.122	5.322	514.370	500.000	2.9
Aroclor-1260-1	6.251	6.155	6.355	483.370	500.000	-3.3
Aroclor-1260-2	6.439	6.342	6.542	477.180	500.000	-4.6
Aroclor-1260-3	6.591	6.495	6.695	462.160	500.000	-7.6
Aroclor-1260-4	7.063	6.967	7.167	461.970	500.000	-7.6
Aroclor-1260-5	7.303	7.207	7.407	444.510	500.000	-11.1
Decachlorobiphenyl	8.691	8.596	8.796	48.770	50.000	-2.5
Tetrachloro-m-xylene	3.688	3.590	3.790	53.150	50.000	6.3