

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1730 SAS No.: Q1730 SDG NO.: Q1730

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/18/2025 03/18/2025

Client Sample No.: CCAL04 Date Analyzed: 04/08/2025

Lab Sample No.: AR1660CCC500 Data File : PO110292.D Time Analyzed: 09:08

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.783	4.685	4.885	490.750	500.000	-1.9
Aroclor-1016-2	4.802	4.704	4.904	490.770	500.000	-1.8
Aroclor-1016-3	4.858	4.761	4.961	494.230	500.000	-1.2
Aroclor-1016-4	4.979	4.881	5.081	493.910	500.000	-1.2
Aroclor-1016-5	5.236	5.139	5.339	541.310	500.000	8.3
Aroclor-1260-1	6.277	6.180	6.380	505.880	500.000	1.2
Aroclor-1260-2	6.467	6.369	6.569	476.430	500.000	-4.7
Aroclor-1260-3	6.834	6.737	6.937	473.120	500.000	-5.4
Aroclor-1260-4	7.094	6.998	7.198	466.380	500.000	-6.7
Aroclor-1260-5	7.337	7.239	7.439	469.070	500.000	-6.2
Decachlorobiphenyl	8.742	8.645	8.845	53.700	50.000	7.4
Tetrachloro-m-xylene	3.690	3.592	3.792	48.950	50.000	-2.1

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1730 SAS No.: Q1730 SDG NO.: Q1730

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/18/2025 03/18/2025

Client Sample No.: CCAL04 Date Analyzed: 04/08/2025

Lab Sample No.: AR1660CCC500 Data File : PO110292.D Time Analyzed: 09:08

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.768	4.672	4.872	460.700	500.000	-7.9
Aroclor-1016-2	4.787	4.691	4.891	458.320	500.000	-8.3
Aroclor-1016-3	4.963	4.867	5.067	460.180	500.000	-8.0
Aroclor-1016-4	5.004	4.909	5.109	424.320	500.000	-15.1
Aroclor-1016-5	5.218	5.122	5.322	488.430	500.000	-2.3
Aroclor-1260-1	6.250	6.155	6.355	451.100	500.000	-9.8
Aroclor-1260-2	6.438	6.342	6.542	446.060	500.000	-10.8
Aroclor-1260-3	6.590	6.495	6.695	427.140	500.000	-14.6
Aroclor-1260-4	7.062	6.967	7.167	425.210	500.000	-15.0
Aroclor-1260-5	7.303	7.207	7.407	416.030	500.000	-16.8
Decachlorobiphenyl	8.691	8.596	8.796	44.590	50.000	-10.8
Tetrachloro-m-xylene	3.687	3.590	3.790	47.020	50.000	-6.0