

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1730 SAS No.: Q1730 SDG NO.: Q1730

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/18/2025 03/18/2025

Client Sample No.: CCAL05 Date Analyzed: 04/08/2025

Lab Sample No.: AR1660CCC500 Data File : PO110302.D Time Analyzed: 12:58

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.783	4.685	4.885	478.010	500.000	-4.4
Aroclor-1016-2	4.802	4.704	4.904	481.930	500.000	-3.6
Aroclor-1016-3	4.858	4.761	4.961	488.150	500.000	-2.4
Aroclor-1016-4	4.979	4.881	5.081	483.730	500.000	-3.3
Aroclor-1016-5	5.237	5.139	5.339	484.580	500.000	-3.1
Aroclor-1260-1	6.277	6.180	6.380	484.710	500.000	-3.1
Aroclor-1260-2	6.466	6.369	6.569	466.260	500.000	-6.7
Aroclor-1260-3	6.834	6.737	6.937	478.180	500.000	-4.4
Aroclor-1260-4	7.094	6.998	7.198	469.910	500.000	-6.0
Aroclor-1260-5	7.336	7.239	7.439	460.630	500.000	-7.9
Decachlorobiphenyl	8.739	8.645	8.845	56.540	50.000	13.1
Tetrachloro-m-xylene	3.691	3.592	3.792	47.860	50.000	-4.3

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1730 SAS No.: Q1730 SDG NO.: Q1730

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/18/2025 03/18/2025

Client Sample No.: CCAL05 Date Analyzed: 04/08/2025

Lab Sample No.: AR1660CCC500 Data File : PO110302.D Time Analyzed: 12:58

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.769	4.672	4.872	459.150	500.000	-8.2
Aroclor-1016-2	4.788	4.691	4.891	461.630	500.000	-7.7
Aroclor-1016-3	4.963	4.867	5.067	464.240	500.000	-7.2
Aroclor-1016-4	5.005	4.909	5.109	462.280	500.000	-7.5
Aroclor-1016-5	5.218	5.122	5.322	459.630	500.000	-8.1
Aroclor-1260-1	6.250	6.155	6.355	447.090	500.000	-10.6
Aroclor-1260-2	6.438	6.342	6.542	439.930	500.000	-12.0
Aroclor-1260-3	6.591	6.495	6.695	426.390	500.000	-14.7
Aroclor-1260-4	7.063	6.967	7.167	425.830	500.000	-14.8
Aroclor-1260-5	7.303	7.207	7.407	416.620	500.000	-16.7
Decachlorobiphenyl	8.692	8.596	8.796	45.310	50.000	-9.4
Tetrachloro-m-xylene	3.688	3.590	3.790	47.270	50.000	-5.5