284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Cover Page

Order ID : Q1730

Project ID : Raymark Superfund Site

Client : Nobis Group

Lab Sample Number	Client Sample Number
Q1730-01	OU4-VSL-15-040325
Q1730-02	OU4-VSL-15-040325
Q1730-03	OU4-VSL-16-040325
Q1730-04	OU4-VSL-16-040325
Q1730-05	OU4-VSL-17-040325
Q1730-06	OU4-VSL-17-040325
Q1730-07	OU4-PCS-TC-21-040325
Q1730-08	OU4-PCS-TC-21-040325
Q1730-09	OU4-PCS-TC-22-040325
Q1730-10	OU4-PCS-TC-22-040325
Q1730-11	OU4-PCS-TC-23-040325
Q1730-12	OU4-PCS-TC-23-040325
Q1730-13	OU4-PCS-TC-24-040325
Q1730-14	OU4-PCS-TC-24-040325
Q1730-15	OU4-PCS-TC-25-040325
Q1730-16	OU4-PCS-TC-25-040325
Q1730-17	OU4-PCS-TC-26-040325
Q1730-18	OU4-PCS-TC-26-040325
Q1730-19	OU4-CF-15-040325
Q1730-20	OU4-CF-15-040325

I certify that the data package is in compliance with the terms and conditions of the contract,both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 4/10/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

Nobis Group
Project Name: Raymark Superfund Site
Project # N/A
Chemtech Project # Q1730
Test Name: PCB

A. Number of Samples and Date of Receipt:
20 Solid samples were received on 04/04/2025.

B. Parameters
According to the Chain of Custody document, the following analyses were requested: Cyanide, Herbicide Group1, Mercury, Metals ICP-TAL, METALS-TAL, PCB, Pesticide-TCL, SPLP Extraction, SPLP Mercury, SPLP MetalGroup3, SVOCMS Group3 and VOCMS Group3. This data package contains results for PCB.

C. Analytical Techniques:
The analyses were performed on instrument GCECD_P. The front column is ZB-MR1 which is 30 meters, 0.32 mm ID, 0.5 um df, Catalogue # 7HM-G016-17. The rear column is ZB-MR2 which is 30 meters, 0.32 mm ID, 0.25 µm; Catalogue # 7HM-G017-11.The analyses were performed on instrument GCECD_O. The front column is ZB-MR1 which is 30 meters, 0.32 mm ID, 0.5 um df, Catalogue # 7HM-G016-17. The rear column is ZB-MR2 which is 30 meters, 0.32 mm ID, 0.25 µm; Catalogue # 7HM-G017-11.The analysis of PCBs was based on method 8082A and extraction was done based on method 3541.

D. QA/ QC Samples:
The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria except for OU4-CF-15-040325 [Tetrachloro-m-xylene(2) - 138%]AS per method one surrogate allowed to fail to meet the criteria per column. No further corrective action was taken.

The Retention Times were acceptable for all samples.
The MS recoveries met the requirements for all compounds.
The MSD recoveries met the acceptable requirements.
The RPD met criteria .
The Blank Spike met requirements for all samples .
The Blank analysis did not indicate the presence of lab contamination.
The Initial Calibration met the requirements .
The Continuous Calibration met the requirements .

284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

E. Additional Comments:
The not QT review data is reported in the Miscellaneous.
The soil samples results are based on a dry weight basis.

F. Manual Integration Comments:
Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

284 Sheffield Street, Mountainside NJ 07092 Tel. 908-789-8900 Fax: 908-789-8922

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following " Results Qualifiers" are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. "10 U". This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value**.** This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This is flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as "12 B".
E	Indicates the analyte 's concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a "P".
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

GC ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: Q1730 MATRIX: Solid

METHOD: 8082A/3541

		NA	NO	YES
1.	Chromatograms Labeled/Compounds Identified.			✓
2.	Standard Summary Submitted.			✓
3.	Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis, 12 HOURS IF 8000 SERIES METHOD. The Initial Calibration met the requirements . The Continuous Calibration met the requirements .			✓
4.	Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
5.	Surrogate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable ranges. The Surrogate recoveries met the acceptable criteria except for OU4-CF-15-040325 [Tetrachloro-m-xylene(2) - 138%]AS per method one surrogate allowed to fail to meet the criteria per column. No further corrective action was taken.		✓	
6.	Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range. The MS recoveries met the requirements for all compounds. The MSD recoveries met the acceptable requirements. The Blank Spike met requirements for all samples . The RPD met criteria .		✓	
7.	Retention Time Shift Meet Criteria (if applicable) Comments:			✓
8.	Extraction Holding Time Met If not met, list number of days exceeded for each sample:			✓

284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

GC ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY (CONTINUED)

		NA	NO	YES
9.	Analysis Holding Time Met			✓
	If not met, list those compounds and their recoveries which fall outside the acceptable range.			

ADDITIONAL COMMENTS:

The not QT review data is reported in the Miscellaneous.

The soil samples results are based on a dry weight basis.

_____ _____

QA REVIEW Date

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: **Q1730**

	Completed
For thorough review, the report must have the following:	
GENERAL:	
Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)	✓
Check chain-of-custody for proper relinquish/return of samples	✓
Is the chain of custody signed and complete	✓
Check internal chain-of-custody for proper relinquish/return of samples /sample extracts	✓
Collect information for each project id from server. Were all requirements followed	✓
COVER PAGE:	
Do numbers of samples correspond to the number of samples in the Chain of Custody on login page	✓
Do lab numbers and client Ids on cover page agree with the Chain of Custody	✓
CHAIN OF CUSTODY:	
Do requested analyses on Chain of Custody agree with form I results	✓
Do requested analyses on Chain of Custody agree with the log-in page	✓
Were the correct method log-in for analysis according to the Analytical Request and Chain of Castody	✓
Were the samples received within hold time	✓
Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle	✓
ANALYTICAL:	
Was method requirement followed?	✓
Was client requirement followed?	✓
Does the case narrative summarize all QC failure?	✓
All runlogs and manual integration are reviewed for requirements	✓
All manual calculations and /or hand notations verified	✓

QA Review Signature: **PATEL VAISHALI** **Date:** **04/10/2025**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

LAB CHRONICLE

OrderID:	Q1730	**OrderDate:**	4/4/2025 10:51:00 AM
Client:	Nobis Group	**Project:**	Raymark Superfund Site
Contact:	Adam Roy	**Location:**	L31,VOA Ref. #2 Soil

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1730-01	**OU4-VSL-15-040325**	**SOIL**			**04/03/25**			**04/04/25**
			PCB	8082A		04/07/25	04/07/25	
Q1730-03	**OU4-VSL-16-040325**	**SOIL**			**04/03/25**			**04/04/25**
			PCB	8082A		04/07/25	04/07/25	
Q1730-05	**OU4-VSL-17-040325**	**SOIL**			**04/03/25**			**04/04/25**
			PCB	8082A		04/07/25	04/08/25	
Q1730-07	**OU4-PCS-TC-21-040325**	**SOIL**			**04/03/25**			**04/04/25**
			PCB	8082A		04/07/25	04/08/25	
Q1730-09	**OU4-PCS-TC-22-040325**	**SOIL**			**04/03/25**			**04/04/25**
			PCB	8082A		04/07/25	04/08/25	
Q1730-11	**OU4-PCS-TC-23-040325**	**SOIL**			**04/03/25**			**04/04/25**
			PCB	8082A		04/07/25	04/08/25	
Q1730-13	**OU4-PCS-TC-24-040325**	**SOIL**			**04/03/25**			**04/04/25**
			PCB	8082A		04/07/25	04/08/25	
Q1730-15	**OU4-PCS-TC-25-040325**	**SOIL**			**04/03/25**			**04/04/25**
			PCB	8082A		04/07/25	04/08/25	
Q1730-17	**OU4-PCS-TC-26-040325**	**SOIL**			**04/03/25**			**04/04/25**
			PCB	8082A		04/07/25	04/08/25	
Q1730-19	**OU4-CF-15-040325**	**SOIL**			**04/03/25**			**04/04/25**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

LAB CHRONICLE

PCB	8082A	04/07/25	04/08/25

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q1730 **Order ID:** Q1730

Client: Nobis Group **Project ID:** Raymark Superfund Site

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	LOD	RDL	Units

Client ID :

Total Concentration: 0.000



QC SUMMARY

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Surrogate Summary

SDG No.: Q1730

Client: Nobis Group

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits Low	Limits High
I.BLK-PO109971.D	PIBLK-PO109971.D	Tetrachloro-m-xylene	1	20	18.5	93		60	140
		Decachlorobiphenyl	1	20	20.0	100		60	140
		Tetrachloro-m-xylene	2	20	18.6	93		60	140
		Decachlorobiphenyl	2	20	21.3	107		60	140
I.BLK-PO110267.D	PIBLK-PO110267.D	Tetrachloro-m-xylene	1	20	19.5	97		60	140
		Decachlorobiphenyl	1	20	18.4	92		60	140
		Tetrachloro-m-xylene	2	20	18.8	94		60	140
		Decachlorobiphenyl	2	20	15.9	79		60	140
Q1730-01	OU4-VSL-15-040325	Tetrachloro-m-xylene	1	20	20.2	101		44	130
		Decachlorobiphenyl	1	20	17.6	88		60	125
		Tetrachloro-m-xylene	2	20	19.4	97		44	130
		Decachlorobiphenyl	2	20	15.9	79		60	125
Q1730-03	OU4-VSL-16-040325	Tetrachloro-m-xylene	1	20	22.4	112		44	130
		Decachlorobiphenyl	1	20	19.4	97		60	125
		Tetrachloro-m-xylene	2	20	21.5	108		44	130
		Decachlorobiphenyl	2	20	16.8	84		60	125
Q1730-03MS	OU4-VSL-16-040325MS	Tetrachloro-m-xylene	1	20	22.8	114		44	130
		Decachlorobiphenyl	1	20	20.9	104		60	125
		Tetrachloro-m-xylene	2	20	21.8	109		44	130
		Decachlorobiphenyl	2	20	18.0	90		60	125
Q1730-03MSD	OU4-VSL-16-040325MSD	Tetrachloro-m-xylene	1	20	21.6	108		44	130
		Decachlorobiphenyl	1	20	19.8	99		60	125
		Tetrachloro-m-xylene	2	20	20.8	104		44	130
		Decachlorobiphenyl	2	20	17.3	86		60	125
I.BLK-PO110282.D	PIBLK-PO110282.D	Tetrachloro-m-xylene	1	20	19.3	97		60	140
		Decachlorobiphenyl	1	20	17.2	86		60	140
		Tetrachloro-m-xylene	2	20	18.8	94		60	140
		Decachlorobiphenyl	2	20	16.2	81		60	140
PB167469BL	PB167469BL	Tetrachloro-m-xylene	1	20	19.0	95		44	130
		Decachlorobiphenyl	1	20	17.3	86		60	125
		Tetrachloro-m-xylene	2	20	18.3	91		44	130
		Decachlorobiphenyl	2	20	16.0	80		60	125
PB167469BS	PB167469BS	Tetrachloro-m-xylene	1	20	19.9	99		44	130
		Decachlorobiphenyl	1	20	19.1	95		60	125
		Tetrachloro-m-xylene	2	20	18.9	95		44	130
		Decachlorobiphenyl	2	20	17.3	86		60	125
I.BLK-PO110290.D	PIBLK-PO110290.D	Tetrachloro-m-xylene	1	20	19.5	98		60	140
		Decachlorobiphenyl	1	20	17.7	89		60	140
		Tetrachloro-m-xylene	2	20	18.9	94		60	140
		Decachlorobiphenyl	2	20	16.2	81		60	140
I.BLK-PO110296.D	PIBLK-PO110296.D	Tetrachloro-m-xylene	1	20	18.7	93		60	140

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Surrogate Summary

SDG No.: Q1730

Client: Nobis Group

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits Low	Limits High
I.BLK-PO110296.D	PIBLK-PO110296.D	Decachlorobiphenyl	1	20	17.8	89		60	140
		Tetrachloro-m-xylene	2	20	17.7	89		60	140
		Decachlorobiphenyl	2	20	15.4	77		60	140
Q1730-05	OU4-VSL-17-040325	Tetrachloro-m-xylene	1	20	23.2	116		44	130
		Decachlorobiphenyl	1	20	21.1	105		60	125
		Tetrachloro-m-xylene	2	20	22.1	110		44	130
		Decachlorobiphenyl	2	20	18.1	91		60	125
Q1730-07	OU4-PCS-TC-21-040325	Tetrachloro-m-xylene	1	20	20.3	102		44	130
		Decachlorobiphenyl	1	20	17.1	85		60	125
		Tetrachloro-m-xylene	2	20	19.2	96		44	130
		Decachlorobiphenyl	2	20	14.1	71		60	125
Q1730-09	OU4-PCS-TC-22-040325	Tetrachloro-m-xylene	1	20	20.7	104		44	130
		Decachlorobiphenyl	1	20	17.9	90		60	125
		Tetrachloro-m-xylene	2	20	19.7	98		44	130
		Decachlorobiphenyl	2	20	14.8	74		60	125
Q1730-11	OU4-PCS-TC-23-040325	Tetrachloro-m-xylene	1	20	20.9	104		44	130
		Decachlorobiphenyl	1	20	18.0	90		60	125
		Tetrachloro-m-xylene	2	20	19.9	99		44	130
		Decachlorobiphenyl	2	20	14.8	74		60	125
Q1730-13	OU4-PCS-TC-24-040325	Tetrachloro-m-xylene	1	20	20.5	102		44	130
		Decachlorobiphenyl	1	20	18.9	94		60	125
		Tetrachloro-m-xylene	2	20	19.4	97		44	130
		Decachlorobiphenyl	2	20	15.7	78		60	125
I.BLK-PO110306.D	PIBLK-PO110306.D	Tetrachloro-m-xylene	1	20	18.2	91		60	140
		Decachlorobiphenyl	1	20	18.4	92		60	140
		Tetrachloro-m-xylene	2	20	17.4	87		60	140
		Decachlorobiphenyl	2	20	15.3	77		60	140
I.BLK-PP070916.D	PIBLK-PP070916.D	Tetrachloro-m-xylene	1	20	18.8	94		60	140
		Decachlorobiphenyl	1	20	19.4	97		60	140
		Tetrachloro-m-xylene	2	20	18.1	91		60	140
		Decachlorobiphenyl	2	20	23.4	117		60	140
I.BLK-PP071140.D	PIBLK-PP071140.D	Tetrachloro-m-xylene	1	20	22.8	114		60	140
		Decachlorobiphenyl	1	20	23.9	119		60	140
		Tetrachloro-m-xylene	2	20	23.1	116		60	140
		Decachlorobiphenyl	2	20	21.5	108		60	140
Q1730-15	OU4-PCS-TC-25-040325	Tetrachloro-m-xylene	1	20	24.7	123		44	130
		Decachlorobiphenyl	1	20	23.3	116		60	125
		Tetrachloro-m-xylene	2	20	24.8	124		44	130
		Decachlorobiphenyl	2	20	20.5	102		60	125
Q1730-17	OU4-PCS-TC-26-040325	Tetrachloro-m-xylene	1	20	25.1	125		44	130
		Decachlorobiphenyl	1	20	24.9	124		60	125

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Surrogate Summary

SDG No.: Q1730

Client: Nobis Group

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits Low	Limits High
Q1730-17	OU4-PCS-TC-26-040325	Tetrachloro-m-xylene	2	20	24.8	124		44	130
		Decachlorobiphenyl	2	20	22.0	110		60	125
Q1730-19	OU4-CF-15-040325	Tetrachloro-m-xylene	1	20	25.8	129		44	130
		Decachlorobiphenyl	1	20	24.1	121		60	125
		Tetrachloro-m-xylene	2	20	27.5	138	*	44	130
		Decachlorobiphenyl	2	20	21.7	108		60	125
I.BLK-PP071155.D	PIBLK-PP071155.D	Tetrachloro-m-xylene	1	20	20.6	103		60	140
		Decachlorobiphenyl	1	20	21.0	105		60	140
		Tetrachloro-m-xylene	2	20	21.1	105		60	140
		Decachlorobiphenyl	2	20	18.4	92		60	140

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1730

Client: Nobis Group

Analytical Method: 8082A **DataFile :** PO110271.D

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Limits Low	Limits High	Limits RPD
Client Sample ID:	**OU4-VSL-16-040325MS**											
Q1730-03MS	AR1016	176.1	0	170	ug/kg	97				47	134	
	AR1260	176.1	0	157	ug/kg	89				53	140	

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1730

Client: Nobis Group

Analytical Method: 8082A **DataFile :** PO110272.D

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Limits Low	Limits High	Limits RPD
Client Sample ID:	**OU4-VSL-16-040325MSD**											
Q1730-03MSD	AR1016	176.2	0	164	ug/kg	93		4		47	134	20
	AR1260	176.2	0	153	ug/kg	87		2		53	140	20

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1730

Client: Nobis Group

Analytical Method: 8082A **Datafile :** PO110284.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD Qual	Limits Low	Limits High	RPD
PB167469BS	AR1016	166.6	141	ug/kg	85				47	134	
	AR1260	166.6	134	ug/kg	80				53	140	

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

4C

PESTICIDE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB167469BL

Lab Name: CHEMTECH Contract: NOBI03

Lab Code: CHEM Case No.: Q1730 SAS No.: Q1730 SDG NO.: Q1730

Lab Sample ID: PB167469BL Lab File ID: PO110283.D

Matrix: (soil/water) Solid Extraction: (Type) SOXH

Sulfur Cleanup: (Y/N) N Date Extracted: 04/07/2025

Date Analyzed (1): 04/07/2025 Date Analyzed (2): 04/07/2025

Time Analyzed (1): 19:22 Time Analyzed (2): 19:22

Instrument ID (1): ECD_O Instrument ID (2): ECD_O

GC Column (1): ZB-MR1 ID: 0.32 (mm) GC Column (2): ZB-MR2 ID: 0.32 (mm)

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED 1	DATE ANALYZED 2
OU4-VSL-15-040325	Q1730-01	PO110269.D	04/07/2025	04/07/2025
OU4-VSL-16-040325	Q1730-03	PO110270.D	04/07/2025	04/07/2025
OU4-VSL-16-040325MS	Q1730-03MS	PO110271.D	04/07/2025	04/07/2025
OU4-VSL-16-040325MSD	Q1730-03MSD	PO110272.D	04/07/2025	04/07/2025
PB167469BS	PB167469BS	PO110284.D	04/07/2025	04/07/2025
OU4-VSL-17-040325	Q1730-05	PO110297.D	04/08/2025	04/08/2025
OU4-PCS-TC-21-040325	Q1730-07	PO110298.D	04/08/2025	04/08/2025
OU4-PCS-TC-22-040325	Q1730-09	PO110299.D	04/08/2025	04/08/2025
OU4-PCS-TC-23-040325	Q1730-11	PO110300.D	04/08/2025	04/08/2025
OU4-PCS-TC-24-040325	Q1730-13	PO110301.D	04/08/2025	04/08/2025
OU4-PCS-TC-25-040325	Q1730-15	PP071142.D	04/08/2025	04/08/2025
OU4-PCS-TC-26-040325	Q1730-17	PP071143.D	04/08/2025	04/08/2025
OU4-CF-15-040325	Q1730-19	PP071144.D	04/08/2025	04/08/2025

COMMENTS:



SAMPLE DATA

Raw Data: **PO110269.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Nobis Group	Date Collected:	04/03/25
Project:	Raymark Superfund Site	Date Received:	04/04/25
Client Sample ID:	OU4-VSL-15-040325	SDG No.:	Q1730
Lab Sample ID:	Q1730-01	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	96.9 Decanted:
Sample Wt/Vol:	30.07 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110269.D	1	04/07/25 08:20	04/07/25 13:38	PB167469

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units(Dry Weight)
TARGETS							
12674-11-2	Aroclor-1016	8.50	U	4.10	8.50	17.5	ug/kg
11104-28-2	Aroclor-1221	13.4	U	4.10	13.4	17.5	ug/kg
11141-16-5	Aroclor-1232	8.50	U	3.80	8.50	17.5	ug/kg
53469-21-9	Aroclor-1242	8.50	U	4.10	8.50	17.5	ug/kg
12672-29-6	Aroclor-1248	13.4	U	6.10	13.4	17.5	ug/kg
11097-69-1	Aroclor-1254	8.50	U	3.30	8.50	17.5	ug/kg
37324-23-5	Aroclor-1262	13.4	U	5.20	13.4	17.5	ug/kg
11100-14-4	Aroclor-1268	8.50	U	3.70	8.50	17.5	ug/kg
11096-82-5	Aroclor-1260	8.50	U	3.30	8.50	17.5	ug/kg
SURROGATES							
877-09-8	Tetrachloro-m-xylene	20.2		44 - 130		101%	SPK: 20
2051-24-3	Decachlorobiphenyl	17.6		60 - 125		88%	SPK: 20

Comments:

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Quantitation Report (QT Reviewed)

Sample Results: **PO110269.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040725\
Data File : PO110269.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Apr 2025 13:38
Operator : YP/AJ
Sample : Q1730-01
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
OU4-VSL-15-040325

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 07 14:27:09 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.690	3.687	183.9E6	101.8E6	20.206	19.404
2)	SA Decachlor...	8.741	8.690	135.7E6	38509502	17.644	15.873
	Target Compounds						

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Sample Results: **PO110269.D**


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040725\
Data File : PO110269.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Apr 2025 13:38
Operator : YP/AJ
Sample : Q1730-01
Misc :
ALS Vial : 8 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
OU4-VSL-15-040325

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 07 14:27:09 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.690 min
Delta R.T.: -0.002 min
Response: 183908751
Conc: 20.21 ng/ml

Instrument :
ECD_O
ClientSampleId :
OU4-VSL-15-040325




#1 Tetrachloro-m-xylene

R.T.: 3.687 min
Delta R.T.: -0.003 min
Response: 101756501
Conc: 19.40 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.741 min
Delta R.T.: -0.005 min
Response: 135656538
Conc: 17.64 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.690 min
Delta R.T.: -0.006 min
Response: 38509502
Conc: 15.87 ng/ml

Raw Data: **PO110270.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Nobis Group	Date Collected:	04/03/25
Project:	Raymark Superfund Site	Date Received:	04/04/25
Client Sample ID:	OU4-VSL-16-040325	SDG No.:	Q1730
Lab Sample ID:	Q1730-03	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	94.5 Decanted:
Sample Wt/Vol:	30.03 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110270.D	1	04/07/25 08:20	04/07/25 13:56	PB167469

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units(Dry Weight)
TARGETS							
12674-11-2	Aroclor-1016	8.80	U	4.20	8.80	18.0	ug/kg
11104-28-2	Aroclor-1221	13.7	U	4.30	13.7	18.0	ug/kg
11141-16-5	Aroclor-1232	8.80	U	3.90	8.80	18.0	ug/kg
53469-21-9	Aroclor-1242	8.80	U	4.20	8.80	18.0	ug/kg
12672-29-6	Aroclor-1248	13.7	U	6.30	13.7	18.0	ug/kg
11097-69-1	Aroclor-1254	8.80	U	3.40	8.80	18.0	ug/kg
37324-23-5	Aroclor-1262	13.7	U	5.30	13.7	18.0	ug/kg
11100-14-4	Aroclor-1268	8.80	U	3.80	8.80	18.0	ug/kg
11096-82-5	Aroclor-1260	8.80	U	3.40	8.80	18.0	ug/kg
SURROGATES							
877-09-8	Tetrachloro-m-xylene	22.4		44 - 130		112%	SPK: 20
2051-24-3	Decachlorobiphenyl	19.4		60 - 125		97%	SPK: 20

Comments:

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Quantitation Report (QT Reviewed)

Sample Results: **PO110270.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040725\
Data File : PO110270.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Apr 2025 13:56
Operator : YP/AJ
Sample : Q1730-03
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
OU4-VSL-16-040325

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 07 14:27:50 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.690	3.687	203.6E6	112.9E6	22.365	21.523
2)	SA Decachlor...	8.740	8.690	149.1E6	40812607	19.387	16.822
	Target Compounds						


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040725\
Data File : PO110270.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Apr 2025 13:56
Operator : YP/AJ
Sample : Q1730-03
Misc :
ALS Vial : 9 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
OU4-VSL-16-040325

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 07 14:27:50 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.690 min
Delta R.T.: -0.002 min
Response: 203553270
Conc: 22.36 ng/ml

Instrument :
ECD_O
ClientSampleId :
OU4-VSL-16-040325




#1 Tetrachloro-m-xylene

R.T.: 3.687 min
Delta R.T.: -0.002 min
Response: 112868485
Conc: 21.52 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.740 min
Delta R.T.: -0.006 min
Response: 149058190
Conc: 19.39 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.690 min
Delta R.T.: -0.006 min
Response: 40812607
Conc: 16.82 ng/ml

Raw Data: **PO110297.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Nobis Group	Date Collected:	04/03/25
Project:	Raymark Superfund Site	Date Received:	04/04/25
Client Sample ID:	OU4-VSL-17-040325	SDG No.:	Q1730
Lab Sample ID:	Q1730-05	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	92.4 Decanted:
Sample Wt/Vol:	30.08 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110297.D	1	04/07/25 08:20	04/08/25 10:39	PB167469

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units(Dry Weight)
TARGETS							
12674-11-2	Aroclor-1016	9.00	U	4.30	9.00	18.3	ug/kg
11104-28-2	Aroclor-1221	14.0	U	4.30	14.0	18.3	ug/kg
11141-16-5	Aroclor-1232	9.00	U	4.00	9.00	18.3	ug/kg
53469-21-9	Aroclor-1242	9.00	U	4.30	9.00	18.3	ug/kg
12672-29-6	Aroclor-1248	14.0	U	6.40	14.0	18.3	ug/kg
11097-69-1	Aroclor-1254	9.00	U	3.50	9.00	18.3	ug/kg
37324-23-5	Aroclor-1262	14.0	U	5.40	14.0	18.3	ug/kg
11100-14-4	Aroclor-1268	9.00	U	3.90	9.00	18.3	ug/kg
11096-82-5	Aroclor-1260	9.00	U	3.50	9.00	18.3	ug/kg
SURROGATES							
877-09-8	Tetrachloro-m-xylene	23.2		44 - 130		116%	SPK: 20
2051-24-3	Decachlorobiphenyl	21.1		60 - 125		105%	SPK: 20

Comments:

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Quantitation Report (QT Reviewed)

Sample Results: **PO110297.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040825\
Data File : PO110297.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Apr 2025 10:39
Operator : YP/AJ
Sample : Q1730-05
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
OU4-VSL-17-040325

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 11:13:12 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.691	3.688	211.3E6	115.8E6	23.218	22.077
2)	SA Decachlor...	8.740	8.691	162.1E6	43918378	21.086	18.102
	Target Compounds						


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040825\
Data File : PO110297.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Apr 2025 10:39
Operator : YP/AJ
Sample : Q1730-05
Misc :
ALS Vial : 7 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
OU4-VSL-17-040325

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 11:13:12 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.691 min
Delta R.T.: -0.001 min
Response: 211317711
Conc: 23.22 ng/ml

Instrument :
ECD_O
ClientSampleId :
OU4-VSL-17-040325




#1 Tetrachloro-m-xylene

R.T.: 3.688 min
Delta R.T.: -0.002 min
Response: 115771341
Conc: 22.08 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.740 min
Delta R.T.: -0.005 min
Response: 162122100
Conc: 21.09 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.691 min
Delta R.T.: -0.004 min
Response: 43918378
Conc: 18.10 ng/ml

Raw Data: **PO110298.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Nobis Group	Date Collected:	04/03/25
Project:	Raymark Superfund Site	Date Received:	04/04/25
Client Sample ID:	OU4-PCS-TC-21-040325	SDG No.:	Q1730
Lab Sample ID:	Q1730-07	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	89 Decanted:
Sample Wt/Vol:	30.01 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110298.D	1	04/07/25 08:20	04/08/25 10:58	PB167469

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units(Dry Weight)
TARGETS							
12674-11-2	Aroclor-1016	9.30	U	4.40	9.30	19.1	ug/kg
11104-28-2	Aroclor-1221	14.6	U	4.50	14.6	19.1	ug/kg
11141-16-5	Aroclor-1232	9.30	U	4.20	9.30	19.1	ug/kg
53469-21-9	Aroclor-1242	9.30	U	4.50	9.30	19.1	ug/kg
12672-29-6	Aroclor-1248	14.6	U	6.60	14.6	19.1	ug/kg
11097-69-1	Aroclor-1254	9.30	U	3.60	9.30	19.1	ug/kg
37324-23-5	Aroclor-1262	14.6	U	5.60	14.6	19.1	ug/kg
11100-14-4	Aroclor-1268	9.30	U	4.00	9.30	19.1	ug/kg
11096-82-5	Aroclor-1260	9.30	U	3.60	9.30	19.1	ug/kg
SURROGATES							
877-09-8	Tetrachloro-m-xylene	20.3		44 - 130		102%	SPK: 20
2051-24-3	Decachlorobiphenyl	17.1		60 - 125		85%	SPK: 20

Comments:

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Quantitation Report (QT Reviewed)

Sample Results: **PO110298.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040825\
Data File : PO110298.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Apr 2025 10:58
Operator : YP/AJ
Sample : Q1730-07
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
OU4-PCS-TC-21-040325

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 11:14:05 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.690	3.687	184.9E6	100.8E6	20.313	19.219
2)	SA Decachlor...	8.741	8.692	131.1E6	34240978	17.046	14.114
	Target Compounds						


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040825\
Data File : PO110298.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Apr 2025 10:58
Operator : YP/AJ
Sample : Q1730-07
Misc :
ALS Vial : 8 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
OU4-PCS-TC-21-040325

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 11:14:05 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.690 min
Delta R.T.: -0.002 min
Response: 184875104
Conc: 20.31 ng/ml

Instrument :
ECD_O
ClientSampleId :
OU4-PCS-TC-21-040325




#1 Tetrachloro-m-xylene

R.T.: 3.687 min
Delta R.T.: -0.002 min
Response: 100787167
Conc: 19.22 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.741 min
Delta R.T.: -0.004 min
Response: 131063538
Conc: 17.05 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.692 min
Delta R.T.: -0.004 min
Response: 34240978
Conc: 14.11 ng/ml

Raw Data: **PO110299.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Nobis Group	Date Collected:	04/03/25
Project:	Raymark Superfund Site	Date Received:	04/04/25
Client Sample ID:	OU4-PCS-TC-22-040325	SDG No.:	Q1730
Lab Sample ID:	Q1730-09	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	90.8 Decanted:
Sample Wt/Vol:	30.02 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110299.D	1	04/07/25 08:20	04/08/25 11:16	PB167469

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units(Dry Weight)
TARGETS							
12674-11-2	Aroclor-1016	9.10	U	4.30	9.10	18.7	ug/kg
11104-28-2	Aroclor-1221	14.3	U	4.40	14.3	18.7	ug/kg
11141-16-5	Aroclor-1232	9.10	U	4.10	9.10	18.7	ug/kg
53469-21-9	Aroclor-1242	9.10	U	4.40	9.10	18.7	ug/kg
12672-29-6	Aroclor-1248	14.3	U	6.50	14.3	18.7	ug/kg
11097-69-1	Aroclor-1254	9.10	U	3.50	9.10	18.7	ug/kg
37324-23-5	Aroclor-1262	14.3	U	5.50	14.3	18.7	ug/kg
11100-14-4	Aroclor-1268	9.10	U	4.00	9.10	18.7	ug/kg
11096-82-5	Aroclor-1260	9.10	U	3.60	9.10	18.7	ug/kg
SURROGATES							
877-09-8	Tetrachloro-m-xylene	20.7		44 - 130		104%	SPK: 20
2051-24-3	Decachlorobiphenyl	17.9		60 - 125		90%	SPK: 20

Comments:

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Quantitation Report (QT Reviewed)

Sample Results: **PO110299.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040825\
Data File : PO110299.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Apr 2025 11:16
Operator : YP/AJ
Sample : Q1730-09
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
OU4-PCS-TC-22-040325

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 11:32:32 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.690	3.688	188.8E6	103.3E6	20.743	19.690
2)	SA Decachlor...	8.741	8.692	138.0E6	35929484	17.943	14.810
	Target Compounds						


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040825\
Data File : PO110299.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Apr 2025 11:16
Operator : YP/AJ
Sample : Q1730-09
Misc :
ALS Vial : 9 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
OU4-PCS-TC-22-040325

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 11:32:32 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.690 min
Delta R.T.: -0.002 min
Response: 188790826
Conc: 20.74 ng/ml

Instrument :
ECD_O
ClientSampleId :
OU4-PCS-TC-22-040325




#1 Tetrachloro-m-xylene

R.T.: 3.688 min
Delta R.T.: -0.002 min
Response: 103254489
Conc: 19.69 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.741 min
Delta R.T.: -0.005 min
Response: 137955475
Conc: 17.94 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.692 min
Delta R.T.: -0.004 min
Response: 35929484
Conc: 14.81 ng/ml

Raw Data: **PO110300.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Nobis Group	Date Collected:	04/03/25
Project:	Raymark Superfund Site	Date Received:	04/04/25
Client Sample ID:	OU4-PCS-TC-23-040325	SDG No.:	Q1730
Lab Sample ID:	Q1730-11	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	90.9 Decanted:
Sample Wt/Vol:	30.05 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110300.D	1	04/07/25 08:20	04/08/25 11:35	PB167469

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units(Dry Weight)
TARGETS							
12674-11-2	Aroclor-1016	9.10	U	4.30	9.10	18.7	ug/kg
11104-28-2	Aroclor-1221	14.3	U	4.40	14.3	18.7	ug/kg
11141-16-5	Aroclor-1232	9.10	U	4.10	9.10	18.7	ug/kg
53469-21-9	Aroclor-1242	9.10	U	4.40	9.10	18.7	ug/kg
12672-29-6	Aroclor-1248	14.3	U	6.50	14.3	18.7	ug/kg
11097-69-1	Aroclor-1254	9.10	U	3.50	9.10	18.7	ug/kg
37324-23-5	Aroclor-1262	14.3	U	5.50	14.3	18.7	ug/kg
11100-14-4	Aroclor-1268	9.10	U	4.00	9.10	18.7	ug/kg
11096-82-5	Aroclor-1260	9.10	U	3.50	9.10	18.7	ug/kg
SURROGATES							
877-09-8	Tetrachloro-m-xylene	20.9		44 - 130		104%	SPK: 20
2051-24-3	Decachlorobiphenyl	18.0		60 - 125		90%	SPK: 20

Comments:

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Quantitation Report (QT Reviewed)

Sample Results: **PO110300.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040825\
Data File : PO110300.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Apr 2025 11:35
Operator : YP/AJ
Sample : Q1730-11
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
OU4-PCS-TC-23-040325

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 11:51:20 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.691	3.687	189.8E6	104.3E6	20.849	19.893
2)	SA Decachlor...	8.740	8.692	138.3E6	35798912	17.988	14.756
	Target Compounds						

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Sample Results: **PO110300.D**


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040825\
Data File : PO110300.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Apr 2025 11:35
Operator : YP/AJ
Sample : Q1730-11
Misc :
ALS Vial : 10 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
OU4-PCS-TC-23-040325

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 11:51:20 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.691 min
Delta R.T.: -0.001 min
Response: 189761354
Conc: 20.85 ng/ml

Instrument :
ECD_O
ClientSampleId :
OU4-PCS-TC-23-040325




#1 Tetrachloro-m-xylene

R.T.: 3.687 min
Delta R.T.: -0.002 min
Response: 104320886
Conc: 19.89 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.740 min
Delta R.T.: -0.006 min
Response: 138303216
Conc: 17.99 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.692 min
Delta R.T.: -0.004 min
Response: 35798912
Conc: 14.76 ng/ml

Raw Data: **PO110301.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Nobis Group	Date Collected:	04/03/25
Project:	Raymark Superfund Site	Date Received:	04/04/25
Client Sample ID:	OU4-PCS-TC-24-040325	SDG No.:	Q1730
Lab Sample ID:	Q1730-13	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	91.3 Decanted:
Sample Wt/Vol:	30.08 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110301.D	1	04/07/25 08:20	04/08/25 11:53	PB167469

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units(Dry Weight)
TARGETS							
12674-11-2	Aroclor-1016	9.10	U	4.30	9.10	18.6	ug/kg
11104-28-2	Aroclor-1221	14.2	U	4.40	14.2	18.6	ug/kg
11141-16-5	Aroclor-1232	9.10	U	4.10	9.10	18.6	ug/kg
53469-21-9	Aroclor-1242	9.10	U	4.40	9.10	18.6	ug/kg
12672-29-6	Aroclor-1248	14.2	U	6.50	14.2	18.6	ug/kg
11097-69-1	Aroclor-1254	9.10	U	3.50	9.10	18.6	ug/kg
37324-23-5	Aroclor-1262	14.2	U	5.50	14.2	18.6	ug/kg
11100-14-4	Aroclor-1268	9.10	U	3.90	9.10	18.6	ug/kg
11096-82-5	Aroclor-1260	9.10	U	3.50	9.10	18.6	ug/kg
SURROGATES							
877-09-8	Tetrachloro-m-xylene	20.5		44 - 130		102%	SPK: 20
2051-24-3	Decachlorobiphenyl	18.9		60 - 125		94%	SPK: 20

Comments:

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Quantitation Report (QT Reviewed)

Sample Results: **PO110301.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040825\
Data File : PO110301.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Apr 2025 11:53
Operator : YP/AJ
Sample : Q1730-13
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
OU4-PCS-TC-24-040325

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 12:09:53 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.690	3.687	186.4E6	101.7E6	20.482	19.389
2)	SA Decachlor...	8.741	8.691	145.2E6	37966376	18.886	15.649
	Target Compounds						


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040825\
Data File : PO110301.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Apr 2025 11:53
Operator : YP/AJ
Sample : Q1730-13
Misc :
ALS Vial : 11 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
OU4-PCS-TC-24-040325

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 12:09:53 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.690 min
Delta R.T.: -0.002 min
Response: 186415453
Conc: 20.48 ng/ml

Instrument :
ECD_O
ClientSampleId :
OU4-PCS-TC-24-040325




#1 Tetrachloro-m-xylene

R.T.: 3.687 min
Delta R.T.: -0.002 min
Response: 101677050
Conc: 19.39 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.741 min
Delta R.T.: -0.004 min
Response: 145207235
Conc: 18.89 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.691 min
Delta R.T.: -0.005 min
Response: 37966376
Conc: 15.65 ng/ml

Raw Data: **PP071142.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Nobis Group	Date Collected:	04/03/25
Project:	Raymark Superfund Site	Date Received:	04/04/25
Client Sample ID:	OU4-PCS-TC-25-040325	SDG No.:	Q1730
Lab Sample ID:	Q1730-15	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	92.5 Decanted:
Sample Wt/Vol:	30.06 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071142.D	1	04/07/25 08:20	04/08/25 11:39	PB167469

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units(Dry Weight)
TARGETS							
12674-11-2	Aroclor-1016	9.00	U	4.30	9.00	18.3	ug/kg
11104-28-2	Aroclor-1221	14.0	U	4.30	14.0	18.3	ug/kg
11141-16-5	Aroclor-1232	9.00	U	4.00	9.00	18.3	ug/kg
53469-21-9	Aroclor-1242	9.00	U	4.30	9.00	18.3	ug/kg
12672-29-6	Aroclor-1248	14.0	U	6.40	14.0	18.3	ug/kg
11097-69-1	Aroclor-1254	9.00	U	3.50	9.00	18.3	ug/kg
37324-23-5	Aroclor-1262	14.0	U	5.40	14.0	18.3	ug/kg
11100-14-4	Aroclor-1268	9.00	U	3.90	9.00	18.3	ug/kg
11096-82-5	Aroclor-1260	9.00	U	3.50	9.00	18.3	ug/kg
SURROGATES							
877-09-8	Tetrachloro-m-xylene	24.8		44 - 130		124%	SPK: 20
2051-24-3	Decachlorobiphenyl	23.3		60 - 125		116%	SPK: 20

Comments:

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Quantitation Report (QT Reviewed)

Sample Results: **PP071142.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040825\
Data File : PP071142.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Apr 2025 11:39
Operator : YP\AJ
Sample : Q1730-15
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
OU4-PCS-TC-25-040325

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 12:00:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 28 03:13:46 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.514	3.815	36065391	25688268	24.676	24.780
2)	SA Decachlor...	10.242	8.859	24468736	16145301	23.246	20.484
	Target Compounds						


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040825\
Data File : PP071142.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Apr 2025 11:39
Operator : YP\AJ
Sample : Q1730-15
Misc :
ALS Vial : 8 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
OU4-PCS-TC-25-040325

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 12:00:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 28 03:13:46 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.514 min
Delta R.T.: 0.002 min
Response: 36065391
Conc: 24.68 ng/ml

Instrument :
ECD_P
ClientSampleId :
OU4-PCS-TC-25-040325




#1 Tetrachloro-m-xylene

R.T.: 3.815 min
Delta R.T.: 0.000 min
Response: 25688268
Conc: 24.78 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.242 min
Delta R.T.: 0.016 min
Response: 24468736
Conc: 23.25 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.859 min
Delta R.T.: 0.000 min
Response: 16145301
Conc: 20.48 ng/ml

Raw Data: **PP071143.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Nobis Group	Date Collected:	04/03/25
Project:	Raymark Superfund Site	Date Received:	04/04/25
Client Sample ID:	OU4-PCS-TC-26-040325	SDG No.:	Q1730
Lab Sample ID:	Q1730-17	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	92.1 Decanted:
Sample Wt/Vol:	30.03 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071143.D	1	04/07/25 08:20	04/08/25 11:56	PB167469

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units(Dry Weight)
TARGETS							
12674-11-2	Aroclor-1016	9.00	U	4.30	9.00	18.4	ug/kg
11104-28-2	Aroclor-1221	14.1	U	4.40	14.1	18.4	ug/kg
11141-16-5	Aroclor-1232	9.00	U	4.00	9.00	18.4	ug/kg
53469-21-9	Aroclor-1242	9.00	U	4.30	9.00	18.4	ug/kg
12672-29-6	Aroclor-1248	14.1	U	6.40	14.1	18.4	ug/kg
11097-69-1	Aroclor-1254	9.00	U	3.50	9.00	18.4	ug/kg
37324-23-5	Aroclor-1262	14.1	U	5.40	14.1	18.4	ug/kg
11100-14-4	Aroclor-1268	9.00	U	3.90	9.00	18.4	ug/kg
11096-82-5	Aroclor-1260	9.00	U	3.50	9.00	18.4	ug/kg
SURROGATES							
877-09-8	Tetrachloro-m-xylene	25.1		44 - 130		125%	SPK: 20
2051-24-3	Decachlorobiphenyl	24.9		60 - 125		124%	SPK: 20

Comments:

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Quantitation Report (QT Reviewed)

Sample Results: **PP071143.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040825\
Data File : PP071143.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Apr 2025 11:56
Operator : YP\AJ
Sample : Q1730-17
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
OU4-PCS-TC-26-040325

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 12:13:50 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 28 03:13:46 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.518	3.815	36660054	25694077	25.083	24.786
2)	SA Decachlor...	10.246	8.859	26192698	17327330	24.883	21.984
	Target Compounds						

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Sample Results: **PP071143.D**


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040825\
Data File : PP071143.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Apr 2025 11:56
Operator : YP\AJ
Sample : Q1730-17
Misc :
ALS Vial : 9 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
OU4-PCS-TC-26-040325

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 12:13:50 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 28 03:13:46 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.518 min
Delta R.T.: 0.006 min
Response: 36660054
Conc: 25.08 ng/ml

Instrument :
ECD_P
ClientSampleId :
OU4-PCS-TC-26-040325




#1 Tetrachloro-m-xylene

R.T.: 3.815 min
Delta R.T.: 0.000 min
Response: 25694077
Conc: 24.79 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.246 min
Delta R.T.: 0.021 min
Response: 26192698
Conc: 24.88 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.859 min
Delta R.T.: 0.000 min
Response: 17327330
Conc: 21.98 ng/ml

Raw Data: **PP071144.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Nobis Group	Date Collected:	04/03/25
Project:	Raymark Superfund Site	Date Received:	04/04/25
Client Sample ID:	OU4-CF-15-040325	SDG No.:	Q1730
Lab Sample ID:	Q1730-19	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	93.9 Decanted:
Sample Wt/Vol:	30.02 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071144.D	1	04/07/25 08:20	04/08/25 12:12	PB167469

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units(Dry Weight)
TARGETS							
12674-11-2	Aroclor-1016	8.80	U	4.20	8.80	18.1	ug/kg
11104-28-2	Aroclor-1221	13.8	U	4.30	13.8	18.1	ug/kg
11141-16-5	Aroclor-1232	8.80	U	4.00	8.80	18.1	ug/kg
53469-21-9	Aroclor-1242	8.80	U	4.30	8.80	18.1	ug/kg
12672-29-6	Aroclor-1248	13.8	U	6.30	13.8	18.1	ug/kg
11097-69-1	Aroclor-1254	8.80	U	3.40	8.80	18.1	ug/kg
37324-23-5	Aroclor-1262	13.8	U	5.30	13.8	18.1	ug/kg
11100-14-4	Aroclor-1268	8.80	U	3.80	8.80	18.1	ug/kg
11096-82-5	Aroclor-1260	8.80	U	3.40	8.80	18.1	ug/kg
SURROGATES							
877-09-8	Tetrachloro-m-xylene	27.5	*	44 - 130		138%	SPK: 20
2051-24-3	Decachlorobiphenyl	24.1		60 - 125		121%	SPK: 20

Comments:

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Quantitation Report (QT Reviewed)

Sample Results: **PP071144.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040825\
Data File : PP071144.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Apr 2025 12:12
Operator : YP\AJ
Sample : Q1730-19
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
OU4-CF-15-040325

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 12:26:41 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 28 03:13:46 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.516	3.815	37706422	28536768	25.798	27.528
2)	SA Decachlor...	10.242	8.859	25401796	17071245	24.132	21.659
	Target Compounds						

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Sample Results: **PP071144.D**


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040825\
Data File : PP071144.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Apr 2025 12:12
Operator : YP\AJ
Sample : Q1730-19
Misc :
ALS Vial : 10 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
OU4-CF-15-040325

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 12:26:41 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 28 03:13:46 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.516 min
Delta R.T.: 0.004 min
Response: 37706422
Conc: 25.80 ng/ml

Instrument :
ECD_P
ClientSampleId :
OU4-CF-15-040325




#1 Tetrachloro-m-xylene

R.T.: 3.815 min
Delta R.T.: 0.000 min
Response: 28536768
Conc: 27.53 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.242 min
Delta R.T.: 0.017 min
Response: 25401796
Conc: 24.13 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.859 min
Delta R.T.: 0.000 min
Response: 17071245
Conc: 21.66 ng/ml

CALIBRATION SUMMARY

Raw Data

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

RETENTION TIMES OF INITIAL CALIBRATION

Contract: NOBI03

Lab Code: CHEM **Case No.:** Q1730 **SAS No.:** Q1730 **SDG NO.:** Q1730

Instrument ID: ECD_O **Calibration Date(s):** 03/18/2025 03/18/2025

Calibration Times: 14:03 22:15

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID: RT 1000 = PO109972.D RT 750 = PO109973.D

RT 500 = PO109974.D RT 250 = PO109975.D RT 050 = PO109976.D

COMPOUND		RT 1000	RT 750	RT 500	RT 250	RT 050	MEAN RT	RT WINDOW FROM	RT WINDOW TO
Aroclor-1016-1	(1)	4.79	4.79	4.79	4.79	4.78	4.79	4.69	4.89
Aroclor-1016-2	(2)	4.81	4.81	4.80	4.80	4.80	4.80	4.70	4.90
Aroclor-1016-3	(3)	4.86	4.86	4.86	4.86	4.86	4.86	4.76	4.96
Aroclor-1016-4	(4)	4.98	4.98	4.98	4.98	4.98	4.98	4.88	5.08
Aroclor-1016-5	(5)	5.24	5.24	5.24	5.24	5.24	5.24	5.14	5.34
Aroclor-1260-1	(1)	6.28	6.28	6.28	6.28	6.28	6.28	6.18	6.38
Aroclor-1260-2	(2)	6.47	6.47	6.47	6.47	6.47	6.47	6.37	6.57
Aroclor-1260-3	(3)	6.84	6.84	6.84	6.84	6.84	6.84	6.74	6.94
Aroclor-1260-4	(4)	7.10	7.10	7.10	7.10	7.10	7.10	7.00	7.20
Aroclor-1260-5	(5)	7.34	7.34	7.34	7.34	7.34	7.34	7.24	7.44
Decachlorobiphenyl		8.75	8.75	8.75	8.75	8.74	8.75	8.65	8.85
Tetrachloro-m-xylene		3.69	3.69	3.69	3.69	3.69	3.69	3.59	3.79
Aroclor-1242-1	(1)	4.79	4.79	4.79	4.79	4.78	4.79	4.69	4.89
Aroclor-1242-2	(2)	4.80	4.80	4.81	4.80	4.80	4.80	4.70	4.90
Aroclor-1242-3	(3)	4.86	4.86	4.86	4.86	4.86	4.86	4.76	4.96
Aroclor-1242-4	(4)	4.98	4.98	4.98	4.98	4.98	4.98	4.88	5.08
Aroclor-1242-5	(5)	5.64	5.63	5.64	5.63	5.63	5.63	5.53	5.73
Decachlorobiphenyl		8.74	8.74	8.74	8.74	8.74	8.74	8.64	8.84
Tetrachloro-m-xylene		3.69	3.69	3.69	3.69	3.69	3.69	3.59	3.79
Aroclor-1248-1	(1)	4.78	4.79	4.79	4.78	4.78	4.78	4.68	4.88
Aroclor-1248-2	(2)	5.02	5.02	5.02	5.02	5.02	5.02	4.92	5.12
Aroclor-1248-3	(3)	5.24	5.24	5.24	5.24	5.24	5.24	5.14	5.34
Aroclor-1248-4	(4)	5.59	5.59	5.59	5.59	5.59	5.59	5.49	5.69
Aroclor-1248-5	(5)	5.63	5.64	5.63	5.63	5.63	5.63	5.53	5.73
Decachlorobiphenyl		8.74	8.74	8.74	8.74	8.74	8.74	8.64	8.84
Tetrachloro-m-xylene		3.69	3.69	3.69	3.69	3.69	3.69	3.59	3.79
Aroclor-1254-1	(1)	5.59	5.59	5.59	5.59	5.59	5.59	5.49	5.69
Aroclor-1254-2	(2)	5.74	5.74	5.74	5.74	5.74	5.74	5.64	5.84
Aroclor-1254-3	(3)	6.15	6.15	6.15	6.15	6.15	6.15	6.05	6.25
Aroclor-1254-4	(4)	6.38	6.38	6.38	6.38	6.38	6.38	6.28	6.48
Aroclor-1254-5	(5)	6.80	6.80	6.80	6.80	6.80	6.80	6.70	6.90
Decachlorobiphenyl		8.74	8.74	8.74	8.74	8.74	8.74	8.64	8.84
Tetrachloro-m-xylene		3.69	3.69	3.69	3.69	3.69	3.69	3.59	3.79
Aroclor-1268-1	(1)	7.62	7.62	7.62	7.62	7.62	7.62	7.52	7.72
Aroclor-1268-2	(2)	7.69	7.69	7.69	7.69	7.69	7.69	7.59	7.79
Aroclor-1268-3	(3)	7.90	7.90	7.90	7.90	7.90	7.90	7.80	8.00
Aroclor-1268-4	(4)	8.18	8.18	8.19	8.18	8.18	8.18	8.08	8.28
Aroclor-1268-5	(5)	8.48	8.48	8.48	8.48	8.48	8.48	8.38	8.58

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

RETENTION TIMES OF INITIAL CALIBRATION

Decachlorobiphenyl	8.74	8.74	8.74	8.74	8.74	8.74	8.64	8.84
Tetrachloro-m-xylene	3.69	3.69	3.69	3.69	3.69	3.69	3.59	3.79

Raw Data

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

RETENTION TIMES OF INITIAL CALIBRATION

Contract: NOBI03

Lab Code: CHEM **Case No.:** Q1730 **SAS No.:** Q1730 **SDG NO.:** Q1730

Instrument ID: ECD_O **Calibration Date(s):** 03/18/2025 03/18/2025

Calibration Times: 14:03 22:15

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID: RT 1000 = PO109972.D RT 750 = PO109973.D

RT 500 = PO109974.D RT 250 = PO109975.D RT 050 = PO109976.D

COMPOUND		RT 1000	RT 750	RT 500	RT 250	RT 050	MEAN RT	RT WINDOW FROM	RT WINDOW TO
Aroclor-1016-1	(1)	4.77	4.77	4.77	4.77	4.77	4.77	4.67	4.87
Aroclor-1016-2	(2)	4.79	4.79	4.79	4.79	4.79	4.79	4.69	4.89
Aroclor-1016-3	(3)	4.97	4.97	4.97	4.97	4.97	4.97	4.87	5.07
Aroclor-1016-4	(4)	5.01	5.01	5.01	5.01	5.01	5.01	4.91	5.11
Aroclor-1016-5	(5)	5.22	5.22	5.22	5.22	5.22	5.22	5.12	5.32
Aroclor-1260-1	(1)	6.25	6.26	6.26	6.25	6.25	6.25	6.15	6.35
Aroclor-1260-2	(2)	6.44	6.44	6.44	6.44	6.44	6.44	6.34	6.54
Aroclor-1260-3	(3)	6.60	6.60	6.60	6.59	6.59	6.59	6.49	6.69
Aroclor-1260-4	(4)	7.07	7.07	7.07	7.07	7.07	7.07	6.97	7.17
Aroclor-1260-5	(5)	7.31	7.31	7.31	7.31	7.31	7.31	7.21	7.41
Decachlorobiphenyl		8.70	8.70	8.70	8.70	8.70	8.70	8.60	8.80
Tetrachloro-m-xylene		3.69	3.69	3.69	3.69	3.69	3.69	3.59	3.79
Aroclor-1242-1	(1)	4.77	4.77	4.77	4.77	4.77	4.77	4.67	4.87
Aroclor-1242-2	(2)	4.79	4.79	4.79	4.79	4.79	4.79	4.69	4.89
Aroclor-1242-3	(3)	4.97	4.97	4.97	4.97	4.97	4.97	4.87	5.07
Aroclor-1242-4	(4)	5.05	5.05	5.05	5.05	5.05	5.05	4.95	5.15
Aroclor-1242-5	(5)	5.57	5.57	5.57	5.57	5.57	5.57	5.47	5.67
Decachlorobiphenyl		8.70	8.70	8.70	8.70	8.70	8.70	8.60	8.80
Tetrachloro-m-xylene		3.69	3.69	3.69	3.69	3.69	3.69	3.59	3.79
Aroclor-1248-1	(1)	4.77	4.77	4.77	4.77	4.77	4.77	4.67	4.87
Aroclor-1248-2	(2)	5.01	5.01	5.01	5.01	5.01	5.01	4.91	5.11
Aroclor-1248-3	(3)	5.05	5.05	5.05	5.05	5.05	5.05	4.95	5.15
Aroclor-1248-4	(4)	5.22	5.22	5.22	5.22	5.22	5.22	5.12	5.32
Aroclor-1248-5	(5)	5.61	5.61	5.61	5.61	5.61	5.61	5.51	5.71
Decachlorobiphenyl		8.70	8.70	8.70	8.70	8.70	8.70	8.60	8.80
Tetrachloro-m-xylene		3.69	3.69	3.69	3.69	3.69	3.69	3.59	3.79
Aroclor-1254-1	(1)	5.57	5.57	5.57	5.57	5.57	5.57	5.47	5.67
Aroclor-1254-2	(2)	5.72	5.72	5.72	5.72	5.72	5.72	5.62	5.82
Aroclor-1254-3	(3)	6.12	6.12	6.12	6.12	6.12	6.12	6.02	6.22
Aroclor-1254-4	(4)	6.35	6.35	6.35	6.35	6.35	6.35	6.25	6.45
Aroclor-1254-5	(5)	6.77	6.77	6.77	6.77	6.77	6.77	6.67	6.87
Decachlorobiphenyl		8.69	8.70	8.70	8.70	8.69	8.70	8.60	8.80
Tetrachloro-m-xylene		3.69	3.69	3.69	3.69	3.69	3.69	3.59	3.79
Aroclor-1268-1	(1)	7.59	7.59	7.59	7.59	7.59	7.59	7.49	7.69
Aroclor-1268-2	(2)	7.66	7.66	7.65	7.65	7.65	7.65	7.55	7.75
Aroclor-1268-3	(3)	7.86	7.86	7.86	7.86	7.86	7.86	7.76	7.96
Aroclor-1268-4	(4)	8.15	8.15	8.15	8.15	8.15	8.15	8.05	8.25
Aroclor-1268-5	(5)	8.44	8.44	8.44	8.44	8.44	8.44	8.34	8.54

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

RETENTION TIMES OF INITIAL CALIBRATION

Decachlorobiphenyl	8.70	8.70	8.69	8.70	8.70	8.70	8.60	8.80
Tetrachloro-m-xylene	3.69	3.69	3.69	3.69	3.69	3.69	3.59	3.79

Raw Data

Alliance TECHNICAL GROUP

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: NOBI03

Lab Code: CHEM **Case No.:** Q1730 **SAS No.:** Q1730 **SDG NO.:** Q1730

Instrument ID: ECD_O **Calibration Date(s):** 03/18/2025 03/18/2025

Calibration Times: 14:03 22:15

GC Column: ZB-MR1 **ID:** 0.32 **(mm)**

LAB FILE ID: **CF 1000 =** PO109972.D **CF 750 =** PO109973.D

CF 500 = PO109974.D **CF 250 =** PO109975.D **CF 050 =** PO109976.D

COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	CF	% RSD
Aroclor-1016-1	(1)	318793340	329109113	342176988	354587888	338053060	336544078	4
Aroclor-1016-2	(2)	446388679	463709840	475172436	492289896	440868120	463685794	5
Aroclor-1016-3	(3)	297123739	311570976	324283686	341714956	328237360	320586143	5
Aroclor-1016-4	(4)	237983440	248526219	257445742	268101480	251389660	252689308	4
Aroclor-1016-5	(5)	250501585	263674729	272727188	289617476	297814400	274867076	7
Aroclor-1260-1	(1)	448474074	467586512	475904282	500944916	463845060	471350969	4
Aroclor-1260-2	(2)	549756400	573035712	574538170	604900316	675365020	595519124	8
Aroclor-1260-3	(3)	466244522	483474499	486507370	506233648	484047880	485301584	3
Aroclor-1260-4	(4)	395577993	413728377	429812132	453058452	428662340	424167859	5
Aroclor-1260-5	(5)	1037246832	1067855560	1076605366	1088369780	959152240	1045845956	5
Decachlorobiphenyl		7314268770	7585202853	7715836840	8064386520	7763558800	7688650757	4
Tetrachloro-m-xylene		9191618080	9409704013	9556350980	9350200040	7999627400	9101500103	7
Aroclor-1242-1	(1)	257021206	271242037	274350776	287750984	246670040	267407009	6
Aroclor-1242-2	(2)	361656206	376061644	380771426	397296944	325229300	368203104	7
Aroclor-1242-3	(3)	239272274	256595597	260500336	277304732	256081460	257950880	5
Aroclor-1242-4	(4)	191269111	202702439	206227546	215710620	184561000	200094143	6
Aroclor-1242-5	(5)	200342595	212233400	219728410	230305716	192670380	211056100	7
Decachlorobiphenyl		7073021470	7288055093	7382383840	7783557400	6481660800	7201735721	7
Tetrachloro-m-xylene		8702773030	8995385800	8954862760	9053765120	6655083400	8472374022	12
Aroclor-1248-1	(1)	203222777	210833551	219941524	231478528	236936100	220482496	6
Aroclor-1248-2	(2)	271717932	285234192	301067598	320199420	327717760	301187380	8
Aroclor-1248-3	(3)	341855080	357243465	374409124	395224264	397211640	373188715	6
Aroclor-1248-4	(4)	484718685	503724613	523274664	543824872	584555460	528019659	7
Aroclor-1248-5	(5)	341874536	354831647	369725290	387414496	410903120	372949818	7
Decachlorobiphenyl		7279802030	7532134600	7794766540	8111831880	8262032800	7796113570	5
Tetrachloro-m-xylene		8896906280	9107530813	9275989780	9374608920	8757655400	9082538239	3
Aroclor-1254-1	(1)	527461301	541865743	559314886	571016624	626589000	565249511	7
Aroclor-1254-2	(2)	454907170	469007857	483702014	470938584	527635000	481238125	6
Aroclor-1254-3	(3)	753626169	772675721	789277352	760170300	789680960	773086100	2
Aroclor-1254-4	(4)	465418236	477798716	486446390	499614064	471424900	480140461	3
Aroclor-1254-5	(5)	652046620	666447925	680525202	682124008	569584060	650145563	7
Decachlorobiphenyl		7506577390	7722603000	7867257780	8042146160	8253331200	7878383106	4
Tetrachloro-m-xylene		9124017490	9264222013	9387241720	8732477200	8689133000	9039418285	3
Aroclor-1268-1	(1)	1369838059	1374439380	1401378006	1431422640	1423576940	1400131005	2

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	1266296617	1262736967	1279632688	1321583800	1270503560	1280150726	2
Aroclor-1268-3	(3)	1019701747	1019547960	1038648586	1056461448	1003450380	1027562024	2
Aroclor-1268-4	(4)	432675859	421354135	447439086	473337560	435173000	441995928	4
Aroclor-1268-5	(5)	3202130253	3192211076	3228608174	3247761892	3036317900	3181405859	3
Decachlorobiphenyl		13640006290	13735597093	14120230440	14566780400	14594697800	14131462405	3
Tetrachloro-m-xylene		9016621250	8994153573	9277514520	9319153840	8874982800	9096485197	2

Raw Data

Alliance TECHNICAL GROUP

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: NOBI03

Lab Code: CHEM **Case No.:** Q1730 **SAS No.:** Q1730 **SDG NO.:** Q1730

Instrument ID: ECD_O **Calibration Date(s):** 03/18/2025 03/18/2025

Calibration Times: 14:03 22:15

GC Column: ZB-MR2 **ID:** 0.32 **(mm)**

LAB FILE ID: **CF 1000 =** PO109972.D **CF 750 =** PO109973.D

CF 500 = PO109974.D **CF 250 =** PO109975.D **CF 050 =** PO109976.D

COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	CF	% RSD
Aroclor-1016-1	(1)	173731434	178393712	185954312	194985500	183209860	183254964	4
Aroclor-1016-2	(2)	252919767	260842944	268062016	276298580	249399140	261504489	4
Aroclor-1016-3	(3)	134029267	138651783	144331354	150165168	141715700	141778654	4
Aroclor-1016-4	(4)	107671732	112396073	118651742	125867636	125061600	117929757	7
Aroclor-1016-5	(5)	141734749	147685869	153486032	162580972	159105760	152918676	6
Aroclor-1260-1	(1)	240074381	247747439	258832620	273383796	257170920	255441831	5
Aroclor-1260-2	(2)	281757793	288999419	301276956	315867360	313526020	300285510	5
Aroclor-1260-3	(3)	267611935	278954948	288530668	305847736	303966000	288982257	6
Aroclor-1260-4	(4)	196048044	205344145	218294872	229292036	226412760	215078371	7
Aroclor-1260-5	(5)	481983300	481761991	513782868	520791536	495161780	498696295	4
Decachlorobiphenyl		2314898720	2277821173	2432747880	2620033760	2485044400	2426109187	6
Tetrachloro-m-xylene		5279380900	5378354400	5480250420	5516817960	4565680000	5244096736	7
Aroclor-1242-1	(1)	141331278	146781083	150759024	159963196	132242600	146215436	7
Aroclor-1242-2	(2)	205951481	214434265	215763772	224841412	179186640	208035514	8
Aroclor-1242-3	(3)	108828803	114341291	116166340	122887996	95840060	111612898	9
Aroclor-1242-4	(4)	107629937	113953347	117437436	126339948	114521460	115976426	6
Aroclor-1242-5	(5)	132486572	139043136	144664964	152408060	117607760	137242098	10
Decachlorobiphenyl		2177773130	2225931947	2270761420	2435729160	2067519400	2235543011	6
Tetrachloro-m-xylene		5049304150	5177204613	5163887520	5207194080	3529903800	4825498833	15
Aroclor-1248-1	(1)	111262575	115666484	120322434	126703700	133730220	121537083	7
Aroclor-1248-2	(2)	153166873	160144767	168518512	179835124	180628540	168458763	7
Aroclor-1248-3	(3)	163639212	170966695	179105274	191715488	200520960	181189526	8
Aroclor-1248-4	(4)	192034431	200649257	208567738	223057368	227978540	210457467	7
Aroclor-1248-5	(5)	189825244	196639567	203671792	215836044	233387660	207872061	8
Decachlorobiphenyl		2147083620	2251559707	2344655820	2535991200	2605882000	2377034469	8
Tetrachloro-m-xylene		5118943080	5218637120	5295153740	5366945840	5070595200	5214054996	2
Aroclor-1254-1	(1)	289813236	298739165	307722880	314204860	352998300	312695688	8
Aroclor-1254-2	(2)	250736607	259627313	268490268	271062664	307765180	271536406	8
Aroclor-1254-3	(3)	402159669	413409125	421207530	415558372	424838600	415434659	2
Aroclor-1254-4	(4)	231846868	237861840	244387952	246184552	252577700	242571782	3
Aroclor-1254-5	(5)	329301884	337898680	348664586	361298960	302328380	335898498	7
Decachlorobiphenyl		2196807900	2276377227	2418572320	2516603360	2645427200	2410757601	7
Tetrachloro-m-xylene		5207556450	5288727027	5346917000	5051956440	5061201200	5191271623	3
Aroclor-1268-1	(1)	560759375	568986040	578205714	598591620	620126960	585333942	4

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	519206129	525348819	532636024	548739284	562258340	537637719	3
Aroclor-1268-3	(3)	397648794	399556300	412917038	427809808	440374680	415661324	4
Aroclor-1268-4	(4)	153792414	154678659	162023954	168698684	166024000	161043542	4
Aroclor-1268-5	(5)	963659212	966574348	987141016	1016062428	1031847640	993056929	3
Decachlorobiphenyl		3902633900	3983566053	4144258760	4413490680	4552839600	4199357799	7
Tetrachloro-m-xylene		5158364110	5158184947	5281917420	5320423880	5129069800	5209592031	2

Raw Data

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: NOBI03

Lab Code: CHEM **Case No.:** Q1730 **SAS No.:** Q1730 **SDG NO.:** Q1730

Instrument ID: ECD_O **Date(s) Analyzed:** 03/18/2025 03/18/2025

GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW FROM	RT WINDOW TO	CALIBRATION FACTOR
Aroclor-1221	500	1	3.91	3.81	4.01	115828000
		2	3.99	3.89	4.09	86499800
		3	4.07	3.97	4.17	261696000
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.07	3.97	4.17	215744000
		2	4.56	4.46	4.66	114786000
		3	4.81	4.71	4.91	211780000
		4	4.98	4.88	5.08	113561000
		5	5.02	4.92	5.12	80454000
Aroclor-1262	500	1	6.84	6.74	6.94	688120000
		2	7.34	7.24	7.44	1191970000
		3	7.62	7.52	7.72	470630000
		4	7.69	7.59	7.79	874576000
		5	8.18	8.08	8.28	394442000

Raw Data

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: NOBI03

Lab Code: CHEM **Case No.:** Q1730 **SAS No.:** Q1730 **SDG NO.:** Q1730

Instrument ID: ECD_O **Date(s) Analyzed:** 03/18/2025 03/18/2025

GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW FROM	RT WINDOW TO	CALIBRATION FACTOR
Aroclor-1221	500	1	3.90	3.80	4.00	66882200
		2	3.99	3.89	4.09	49381600
		3	4.06	3.96	4.16	151684000
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.06	3.96	4.16	124093000
		2	4.79	4.69	4.89	120349000
		3	4.97	4.87	5.07	63656000
		4	5.05	4.95	5.15	59341000
		5	5.22	5.12	5.32	63217800
Aroclor-1262	500	1	6.81	6.71	6.91	344312000
		2	7.31	7.21	7.41	537220000
		3	7.59	7.49	7.69	194051000
		4	7.65	7.55	7.75	361786000
		5	8.15	8.05	8.25	141162000

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109972.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 14:03
Operator : YP/AJ
Sample : AR1660ICC1000
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 18 16:33:58 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 18 16:30:31 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.693	3.690	919.2E6	527.9E6	98.055	98.133
2)	SA Decachlor...	8.745	8.696	731.4E6	231.5E6	97.328	97.518
	Target Compounds						
3)	L1 AR-1016-1	4.786	4.772	318.8E6	173.7E6	964.622	966.018
4)	L1 AR-1016-2	4.805	4.791	446.4E6	252.9E6	968.766	970.935
5)	L1 AR-1016-3	4.862	4.967	297.1E6	134.0E6	956.293	962.990
6)	L1 AR-1016-4	4.982	5.009	238.0E6	107.7E6	960.716	951.485
7)	L1 AR-1016-5	5.240	5.222	250.5E6	141.7E6	957.522	960.195
31)	L7 AR-1260-1	6.281	6.254	448.5E6	240.1E6	970.326	962.401
32)	L7 AR-1260-2	6.470	6.442	549.8E6	281.8E6	977.958	966.521
33)	L7 AR-1260-3	6.838	6.595	466.2E6	267.6E6	978.732	962.386
34)	L7 AR-1260-4	7.098	7.066	395.6E6	196.0E6	958.524	946.308
35)	L7 AR-1260-5	7.340	7.307	1037.2E6	482.0E6	981.381	968.065

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109972.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 14:03
Operator : YP/AJ
Sample : AR1660ICC1000
Misc :
ALS Vial : 3 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
AR1660ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 18 16:33:58 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 18 16:30:31 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.693 min
Delta R.T.: 0.000 min
Response: 919161808
Conc: 98.05 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC1000




#1 Tetrachloro-m-xylene

R.T.: 3.690 min
Delta R.T.: 0.000 min
Response: 527938090
Conc: 98.13 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.745 min
Delta R.T.: 0.000 min
Response: 731426877
Conc: 97.33 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.696 min
Delta R.T.: 0.000 min
Response: 231489872
Conc: 97.52 ng/ml



#3 AR-1016-1


R.T.: 4.786 min
Delta R.T.: 0.000 min
Response: 318793340
Conc: 964.62 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC1000




#3 AR-1016-1

R.T.: 4.772 min
Delta R.T.: 0.000 min
Response: 173731434
Conc: 966.02 ng/ml




#4 AR-1016-2

R.T.: 4.805 min
Delta R.T.: 0.000 min
Response: 446388679
Conc: 968.77 ng/ml



#4 AR-1016-2


R.T.: 4.791 min
Delta R.T.: 0.000 min
Response: 252919767
Conc: 970.94 ng/ml



#5 AR-1016-3


R.T.: 4.862 min
Delta R.T.: 0.000 min
Response: 297123739
Conc: 956.29 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC1000




#5 AR-1016-3

R.T.: 4.967 min
Delta R.T.: 0.000 min
Response: 134029267
Conc: 962.99 ng/ml




#6 AR-1016-4

R.T.: 4.982 min
Delta R.T.: 0.000 min
Response: 237983440
Conc: 960.72 ng/ml



#6 AR-1016-4


R.T.: 5.009 min
Delta R.T.: 0.000 min
Response: 107671732
Conc: 951.49 ng/ml



#7 AR-1016-5


R.T.: 5.240 min
Delta R.T.: 0.000 min
Response: 250501585
Conc: 957.52 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC1000




#7 AR-1016-5

R.T.: 5.222 min
Delta R.T.: 0.000 min
Response: 141734749
Conc: 960.19 ng/ml




#31 AR-1260-1

R.T.: 6.281 min
Delta R.T.: 0.000 min
Response: 448474074
Conc: 970.33 ng/ml



#31 AR-1260-1


R.T.: 6.254 min
Delta R.T.: 0.000 min
Response: 240074381
Conc: 962.40 ng/ml



#32 AR-1260-2


R.T.: 6.470 min
Delta R.T.: 0.000 min
Response: 549756400
Conc: 977.96 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC1000




#32 AR-1260-2

R.T.: 6.442 min
Delta R.T.: 0.000 min
Response: 281757793
Conc: 966.52 ng/ml




#33 AR-1260-3

R.T.: 6.838 min
Delta R.T.: 0.000 min
Response: 466244522
Conc: 978.73 ng/ml



#33 AR-1260-3


R.T.: 6.595 min
Delta R.T.: 0.000 min
Response: 267611935
Conc: 962.39 ng/ml



#34 AR-1260-4


R.T.: 7.098 min
Delta R.T.: 0.000 min
Response: 395577993
Conc: 958.52 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC1000




#34 AR-1260-4

R.T.: 7.066 min
Delta R.T.: 0.000 min
Response: 196048044
Conc: 946.31 ng/ml



#35 AR-1260-5

R.T.: 7.340 min
Delta R.T.: 0.000 min
Response: 1037246832
Conc: 981.38 ng/ml



#35 AR-1260-5

R.T.: 7.307 min
Delta R.T.: 0.000 min
Response: 481983300
Conc: 968.07 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109973.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 14:21
Operator : YP/AJ
Sample : AR1660ICC750
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660ICC750

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/19/2025
Supervised By :mohammad ahmed 03/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 18 16:36:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 18 16:30:31 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.693	3.690	705.7E6	403.4E6	75.190	74.986
2)	SA Decachlor...	8.745	8.696	568.9E6	170.8E6	75.465	72.950
	Target Compounds						
3)	L1 AR-1016-1	4.785	4.772	246.8E6	133.8E6	747.915	745.960
4)	L1 AR-1016-2	4.805	4.791	347.8E6	195.6E6	753.172	750.675
5)	L1 AR-1016-3	4.862	4.967	233.7E6	104.0E6	751.394	748.099
6)	L1 AR-1016-4	4.982	5.009	186.4E6	84297055	751.636	746.609
7)	L1 AR-1016-5	5.239	5.222	197.8E6	110.8E6	753.927	750.256
31)	L7 AR-1260-1	6.280	6.255	350.7E6	185.8E6	755.816	746.573
32)	L7 AR-1260-2	6.470	6.442	429.8E6	216.7E6	759.623	745.669
33)	L7 AR-1260-3	6.837	6.595	362.6E6	209.2E6	757.414	751.587
34)	L7 AR-1260-4	7.098	7.067	310.3E6	154.0E6	751.251	745.577
35)	L7 AR-1260-5	7.340	7.307	800.9E6	361.3E6	755.153	735.355m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109973.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 14:21
Operator : YP/AJ
Sample : AR1660ICC750
Misc :
ALS Vial : 4 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
AR1660ICC750


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 03/19/2025
Supervised By :mohammad ahmed 03/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 18 16:36:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 18 16:30:31 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene

R.T.: 3.693 min
Delta R.T.: 0.000 min
Response: 705727801
Conc: 75.19 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC750


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/19/2025
Supervised By :mohammad ahmed 03/24/2025




#1 Tetrachloro-m-xylene

R.T.: 3.690 min
Delta R.T.: 0.000 min
Response: 403376580
Conc: 74.99 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.745 min
Delta R.T.: 0.000 min
Response: 568890214
Conc: 75.47 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.696 min
Delta R.T.: 0.000 min
Response: 170836588
Conc: 72.95 ng/ml




#3 AR-1016-1

R.T.: 4.785 min
Delta R.T.: 0.000 min
Response: 246831835
Conc: 747.92 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC750


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/19/2025
Supervised By :mohammad ahmed 03/24/2025




#3 AR-1016-1

R.T.: 4.772 min
Delta R.T.: 0.000 min
Response: 133795284
Conc: 745.96 ng/ml




#4 AR-1016-2

R.T.: 4.805 min
Delta R.T.: 0.000 min
Response: 347782380
Conc: 753.17 ng/ml



#4 AR-1016-2

R.T.: 4.791 min
Delta R.T.: 0.000 min
Response: 195632208
Conc: 750.68 ng/ml




#5 AR-1016-3

R.T.: 4.862 min
Delta R.T.: 0.000 min
Response: 233678232
Conc: 751.39 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC750


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/19/2025
Supervised By :mohammad ahmed 03/24/2025




#5 AR-1016-3

R.T.: 4.967 min
Delta R.T.: 0.000 min
Response: 103988837
Conc: 748.10 ng/ml




#6 AR-1016-4

R.T.: 4.982 min
Delta R.T.: 0.000 min
Response: 186394664
Conc: 751.64 ng/ml



#6 AR-1016-4

R.T.: 5.009 min
Delta R.T.: 0.000 min
Response: 84297055
Conc: 746.61 ng/ml




#7 AR-1016-5

R.T.: 5.239 min
Delta R.T.: 0.000 min
Response: 197756047
Conc: 753.93 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC750


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/19/2025
Supervised By :mohammad ahmed 03/24/2025




#7 AR-1016-5

R.T.: 5.222 min
Delta R.T.: 0.000 min
Response: 110764402
Conc: 750.26 ng/ml




#31 AR-1260-1

R.T.: 6.280 min
Delta R.T.: 0.000 min
Response: 350689884
Conc: 755.82 ng/ml



#31 AR-1260-1

R.T.: 6.255 min
Delta R.T.: 0.000 min
Response: 185810579
Conc: 746.57 ng/ml




#32 AR-1260-2

R.T.: 6.470 min
Delta R.T.: 0.000 min
Response: 429776784
Conc: 759.62 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC750


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/19/2025
Supervised By :mohammad ahmed 03/24/2025




#32 AR-1260-2

R.T.: 6.442 min
Delta R.T.: 0.000 min
Response: 216749564
Conc: 745.67 ng/ml




#33 AR-1260-3

R.T.: 6.837 min
Delta R.T.: 0.000 min
Response: 362605874
Conc: 757.41 ng/ml



#33 AR-1260-3

R.T.: 6.595 min
Delta R.T.: 0.000 min
Response: 209216211
Conc: 751.59 ng/ml




#34 AR-1260-4

R.T.: 7.098 min
Delta R.T.: 0.000 min
Response: 310296283
Conc: 751.25 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC750


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/19/2025
Supervised By :mohammad ahmed 03/24/2025




#34 AR-1260-4

R.T.: 7.067 min
Delta R.T.: 0.000 min
Response: 154008109
Conc: 745.58 ng/ml



#35 AR-1260-5

R.T.: 7.340 min
Delta R.T.: 0.000 min
Response: 800891670
Conc: 755.15 ng/ml



#35 AR-1260-5

R.T.: 7.307 min
Delta R.T.: 0.000 min
Response: 361321493
Conc: 735.35 ng/ml m

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109974.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 14:39
Operator : YP/AJ
Sample : AR1660ICC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 18 16:30:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 18 16:30:31 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.692	3.690	477.8E6	274.0E6	50.000	50.000
2)	SA Decachlor...	8.745	8.696	385.8E6	121.6E6	50.000	50.000
	Target Compounds						
3)	L1 AR-1016-1	4.785	4.772	171.1E6	92977156	500.000	500.000
4)	L1 AR-1016-2	4.804	4.791	237.6E6	134.0E6	500.000	500.000
5)	L1 AR-1016-3	4.861	4.967	162.1E6	72165677	500.000	500.000
6)	L1 AR-1016-4	4.981	5.009	128.7E6	59325871	500.000	500.000
7)	L1 AR-1016-5	5.239	5.222	136.4E6	76743016	500.000	500.000
31)	L7 AR-1260-1	6.280	6.255	238.0E6	129.4E6	500.000	500.000
32)	L7 AR-1260-2	6.469	6.442	287.3E6	150.6E6	500.000	500.000
33)	L7 AR-1260-3	6.837	6.595	243.3E6	144.3E6	500.000	500.000
34)	L7 AR-1260-4	7.098	7.067	214.9E6	109.1E6	500.000	500.000
35)	L7 AR-1260-5	7.339	7.307	538.3E6	256.9E6	500.000	500.000


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109974.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 14:39
Operator : YP/AJ
Sample : AR1660ICC500
Misc :
ALS Vial : 5 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
AR1660ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 18 16:30:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 18 16:30:31 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.692 min
Delta R.T.: 0.000 min
Response: 477817549
Conc: 50.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.690 min
Delta R.T.: 0.000 min
Response: 274012521
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.745 min
Delta R.T.: 0.000 min
Response: 385791842
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.696 min
Delta R.T.: 0.000 min
Response: 121637394
Conc: 50.00 ng/ml



#3 AR-1016-1


R.T.: 4.785 min
Delta R.T.: 0.000 min
Response: 171088494
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#3 AR-1016-1

R.T.: 4.772 min
Delta R.T.: 0.000 min
Response: 92977156
Conc: 500.00 ng/ml




#4 AR-1016-2

R.T.: 4.804 min
Delta R.T.: 0.000 min
Response: 237586218
Conc: 500.00 ng/ml



#4 AR-1016-2


R.T.: 4.791 min
Delta R.T.: 0.000 min
Response: 134031008
Conc: 500.00 ng/ml



#5 AR-1016-3


R.T.: 4.861 min
Delta R.T.: 0.000 min
Response: 162141843
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#5 AR-1016-3

R.T.: 4.967 min
Delta R.T.: 0.000 min
Response: 72165677
Conc: 500.00 ng/ml




#6 AR-1016-4

R.T.: 4.981 min
Delta R.T.: 0.000 min
Response: 128722871
Conc: 500.00 ng/ml



#6 AR-1016-4


R.T.: 5.009 min
Delta R.T.: 0.000 min
Response: 59325871
Conc: 500.00 ng/ml



#7 AR-1016-5


R.T.: 5.239 min
Delta R.T.: 0.000 min
Response: 136363594
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#7 AR-1016-5

R.T.: 5.222 min
Delta R.T.: 0.000 min
Response: 76743016
Conc: 500.00 ng/ml




#31 AR-1260-1

R.T.: 6.280 min
Delta R.T.: 0.000 min
Response: 237952141
Conc: 500.00 ng/ml



#31 AR-1260-1


R.T.: 6.255 min
Delta R.T.: 0.000 min
Response: 129416310
Conc: 500.00 ng/ml



#32 AR-1260-2


R.T.: 6.469 min
Delta R.T.: 0.000 min
Response: 287269085
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#32 AR-1260-2

R.T.: 6.442 min
Delta R.T.: 0.000 min
Response: 150638478
Conc: 500.00 ng/ml




#33 AR-1260-3

R.T.: 6.837 min
Delta R.T.: 0.000 min
Response: 243253685
Conc: 500.00 ng/ml



#33 AR-1260-3


R.T.: 6.595 min
Delta R.T.: 0.000 min
Response: 144265334
Conc: 500.00 ng/ml



#34 AR-1260-4


R.T.: 7.098 min
Delta R.T.: 0.000 min
Response: 214906066
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#34 AR-1260-4

R.T.: 7.067 min
Delta R.T.: 0.000 min
Response: 109147436
Conc: 500.00 ng/ml



#35 AR-1260-5

R.T.: 7.339 min
Delta R.T.: 0.000 min
Response: 538302683
Conc: 500.00 ng/ml



#35 AR-1260-5

R.T.: 7.307 min
Delta R.T.: 0.000 min
Response: 256891434
Conc: 500.00 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109975.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 14:58
Operator : YP/AJ
Sample : AR1660ICC250
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 18 16:39:25 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 18 16:30:31 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound		RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds							
1)	SA	Tetrachlo...	3.692	3.690	233.8E6	137.9E6	24.929	25.476
2)	SA	Decachlor...	8.745	8.696	201.6E6	65500844	26.286	27.163
	Target Compounds							
3)	L1	AR-1016-1	4.785	4.772	88646972	48746375	263.699	265.987
4)	L1	AR-1016-2	4.804	4.791	123.1E6	69074645	262.197	261.121
5)	L1	AR-1016-3	4.861	4.967	85428739	37541292	268.076	264.759
6)	L1	AR-1016-4	4.981	5.009	67025370	31466909	264.908	270.924
7)	L1	AR-1016-5	5.239	5.222	72404369	40645243	269.031	268.512
31)	L7	AR-1260-1	6.280	6.254	125.2E6	68345949	264.643	268.013
32)	L7	AR-1260-2	6.469	6.441	151.2E6	78966840	262.745	265.904
33)	L7	AR-1260-3	6.837	6.594	126.6E6	76461934	260.615	268.065
34)	L7	AR-1260-4	7.097	7.066	113.3E6	57323009	267.737	270.080
35)	L7	AR-1260-5	7.339	7.307	272.1E6	130.2E6	254.883	260.615

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109975.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 14:58
Operator : YP/AJ
Sample : AR1660ICC250
Misc :
ALS Vial : 6 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
AR1660ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 18 16:39:25 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 18 16:30:31 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.692 min
Delta R.T.: 0.000 min
Response: 233755001
Conc: 24.93 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.690 min
Delta R.T.: 0.000 min
Response: 137920449
Conc: 25.48 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.745 min
Delta R.T.: 0.000 min
Response: 201609663
Conc: 26.29 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.696 min
Delta R.T.: 0.000 min
Response: 65500844
Conc: 27.16 ng/ml



#3 AR-1016-1


R.T.: 4.785 min
Delta R.T.: 0.000 min
Response: 88646972
Conc: 263.70 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC250




#3 AR-1016-1

R.T.: 4.772 min
Delta R.T.: 0.000 min
Response: 48746375
Conc: 265.99 ng/ml




#4 AR-1016-2

R.T.: 4.804 min
Delta R.T.: 0.000 min
Response: 123072474
Conc: 262.20 ng/ml



#4 AR-1016-2


R.T.: 4.791 min
Delta R.T.: 0.000 min
Response: 69074645
Conc: 261.12 ng/ml



#5 AR-1016-3


R.T.: 4.861 min
Delta R.T.: 0.000 min
Response: 85428739
Conc: 268.08 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC250




#5 AR-1016-3

R.T.: 4.967 min
Delta R.T.: 0.000 min
Response: 37541292
Conc: 264.76 ng/ml




#6 AR-1016-4

R.T.: 4.981 min
Delta R.T.: 0.000 min
Response: 67025370
Conc: 264.91 ng/ml



#6 AR-1016-4


R.T.: 5.009 min
Delta R.T.: 0.000 min
Response: 31466909
Conc: 270.92 ng/ml



#7 AR-1016-5


R.T.: 5.239 min
Delta R.T.: 0.000 min
Response: 72404369
Conc: 269.03 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC250




#7 AR-1016-5

R.T.: 5.222 min
Delta R.T.: 0.000 min
Response: 40645243
Conc: 268.51 ng/ml




#31 AR-1260-1

R.T.: 6.280 min
Delta R.T.: 0.000 min
Response: 125236229
Conc: 264.64 ng/ml



#31 AR-1260-1


R.T.: 6.254 min
Delta R.T.: 0.000 min
Response: 68345949
Conc: 268.01 ng/ml



#32 AR-1260-2


R.T.: 6.469 min
Delta R.T.: 0.000 min
Response: 151225079
Conc: 262.75 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC250




#32 AR-1260-2

R.T.: 6.441 min
Delta R.T.: 0.000 min
Response: 78966840
Conc: 265.90 ng/ml




#33 AR-1260-3

R.T.: 6.837 min
Delta R.T.: 0.000 min
Response: 126558412
Conc: 260.61 ng/ml



#33 AR-1260-3


R.T.: 6.594 min
Delta R.T.: 0.000 min
Response: 76461934
Conc: 268.07 ng/ml



#34 AR-1260-4


R.T.: 7.097 min
Delta R.T.: 0.000 min
Response: 113264613
Conc: 267.74 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC250




#34 AR-1260-4

R.T.: 7.066 min
Delta R.T.: 0.000 min
Response: 57323009
Conc: 270.08 ng/ml



#35 AR-1260-5

R.T.: 7.339 min
Delta R.T.: 0.000 min
Response: 272092445
Conc: 254.88 ng/ml



#35 AR-1260-5

R.T.: 7.307 min
Delta R.T.: 0.000 min
Response: 130197884
Conc: 260.61 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109976.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 15:16
Operator : YP/AJ
Sample : AR1660ICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/19/2025
Supervised By :mohammad ahmed 03/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 18 16:42:31 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 18 16:30:31 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.691	3.688	39998137	22828400	4.395	4.353
2)	SA Decachlor...	8.744	8.695	38817794	12425222	5.049	5.121
	Target Compounds						
3)	L1 AR-1016-1	4.784	4.771	16902653	9160493	50.224	49.988
4)	L1 AR-1016-2	4.804	4.790	22043406	12469957	47.540	47.685
5)	L1 AR-1016-3	4.860	4.965	16411868	7085785	51.193	49.978
6)	L1 AR-1016-4	4.981	5.008	12569483	6253080	49.743	53.024
7)	L1 AR-1016-5	5.238	5.220	14890720	7955288	54.174	52.023
31)	L7 AR-1260-1	6.278	6.252	23192253	12858546	48.956m	50.252m
32)	L7 AR-1260-2	6.468	6.441	33768251	15676301	56.704	52.205
33)	L7 AR-1260-3	6.836	6.593	24202394	15198300	49.871	53.072m
34)	L7 AR-1260-4	7.096	7.065	21433117	11320638	50.530	52.635
35)	L7 AR-1260-5	7.339	7.306	47957612	24758089	45.855	49.646

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109976.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 15:16
Operator : YP/AJ
Sample : AR1660ICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
AR1660ICC050


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 03/19/2025
Supervised By :mohammad ahmed 03/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 18 16:42:31 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 18 16:30:31 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene

R.T.: 3.691 min
Delta R.T.: 0.000 min
Response: 39998137
Conc: 4.39 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC050


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/19/2025
Supervised By :mohammad ahmed 03/24/2025




#1 Tetrachloro-m-xylene

R.T.: 3.688 min
Delta R.T.: 0.000 min
Response: 22828400
Conc: 4.35 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.744 min
Delta R.T.: 0.000 min
Response: 38817794
Conc: 5.05 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.695 min
Delta R.T.: 0.000 min
Response: 12425222
Conc: 5.12 ng/ml




#3 AR-1016-1

R.T.: 4.784 min
Delta R.T.: 0.000 min
Response: 16902653
Conc: 50.22 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC050


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/19/2025
Supervised By :mohammad ahmed 03/24/2025




#3 AR-1016-1

R.T.: 4.771 min
Delta R.T.: 0.000 min
Response: 9160493
Conc: 49.99 ng/ml




#4 AR-1016-2

R.T.: 4.804 min
Delta R.T.: 0.000 min
Response: 22043406
Conc: 47.54 ng/ml



#4 AR-1016-2

R.T.: 4.790 min
Delta R.T.: 0.000 min
Response: 12469957
Conc: 47.69 ng/ml




#5 AR-1016-3

R.T.: 4.860 min
Delta R.T.: 0.000 min
Response: 16411868
Conc: 51.19 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC050


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/19/2025
Supervised By :mohammad ahmed 03/24/2025




#5 AR-1016-3

R.T.: 4.965 min
Delta R.T.: 0.000 min
Response: 7085785
Conc: 49.98 ng/ml




#6 AR-1016-4

R.T.: 4.981 min
Delta R.T.: 0.000 min
Response: 12569483
Conc: 49.74 ng/ml



#6 AR-1016-4

R.T.: 5.008 min
Delta R.T.: 0.000 min
Response: 6253080
Conc: 53.02 ng/ml




#7 AR-1016-5

R.T.: 5.238 min
Delta R.T.: 0.000 min
Response: 14890720
Conc: 54.17 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC050


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/19/2025
Supervised By :mohammad ahmed 03/24/2025




#7 AR-1016-5

R.T.: 5.220 min
Delta R.T.: 0.000 min
Response: 7955288
Conc: 52.02 ng/ml




#31 AR-1260-1

R.T.: 6.278 min
Delta R.T.: 0.000 min
Response: 23192253
Conc: 48.96 ng/ml m



#31 AR-1260-1

R.T.: 6.252 min
Delta R.T.: 0.000 min
Response: 12858546
Conc: 50.25 ng/ml m




#32 AR-1260-2

R.T.: 6.468 min
Delta R.T.: 0.000 min
Response: 33768251
Conc: 56.70 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC050


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/19/2025
Supervised By :mohammad ahmed 03/24/2025




#32 AR-1260-2

R.T.: 6.441 min
Delta R.T.: 0.000 min
Response: 15676301
Conc: 52.20 ng/ml




#33 AR-1260-3

R.T.: 6.836 min
Delta R.T.: 0.000 min
Response: 24202394
Conc: 49.87 ng/ml



#33 AR-1260-3

R.T.: 6.593 min
Delta R.T.: 0.000 min
Response: 15198300
Conc: 53.07 ng/ml m




#34 AR-1260-4

R.T.: 7.096 min
Delta R.T.: 0.000 min
Response: 21433117
Conc: 50.53 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC050


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/19/2025
Supervised By :mohammad ahmed 03/24/2025




#34 AR-1260-4

R.T.: 7.065 min
Delta R.T.: 0.000 min
Response: 11320638
Conc: 52.63 ng/ml



#35 AR-1260-5

R.T.: 7.339 min
Delta R.T.: 0.000 min
Response: 47957612
Conc: 45.86 ng/ml



#35 AR-1260-5

R.T.: 7.306 min
Delta R.T.: 0.000 min
Response: 24758089
Conc: 49.65 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109977.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 15:34
Operator : YP/AJ
Sample : AR1221ICC500
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1221ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 18 16:27:37 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 18 16:27:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.692	3.690	428.4E6	261.3E6	50.000	50.000
2)	SA Decachlor...	8.743	8.696	367.0E6	116.0E6	50.000	50.000
	Target Compounds						
8)	L2 AR-1221-1	3.906	3.901	57914123	33441082	500.000	500.000
9)	L2 AR-1221-2	3.992	3.986	43249856	24690825	500.000	500.000
10)	L2 AR-1221-3	4.069	4.062	130.8E6	75841971	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109977.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 15:34
Operator : YP/AJ
Sample : AR1221ICC500
Misc :
ALS Vial : 8 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
AR1221ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 18 16:27:37 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 18 16:27:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.692 min
Delta R.T.: 0.000 min
Response: 428446750
Conc: 50.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1221ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.690 min
Delta R.T.: 0.000 min
Response: 261250272
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.743 min
Delta R.T.: 0.000 min
Response: 367001269
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.696 min
Delta R.T.: 0.000 min
Response: 115987092
Conc: 50.00 ng/ml



#8 AR-1221-1


R.T.: 3.906 min
Delta R.T.: 0.000 min
Response: 57914123
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1221ICC500




#8 AR-1221-1

R.T.: 3.901 min
Delta R.T.: 0.000 min
Response: 33441082
Conc: 500.00 ng/ml




#9 AR-1221-2

R.T.: 3.992 min
Delta R.T.: 0.000 min
Response: 43249856
Conc: 500.00 ng/ml



#9 AR-1221-2


R.T.: 3.986 min
Delta R.T.: 0.000 min
Response: 24690825
Conc: 500.00 ng/ml



#10 AR-1221-3

R.T.: 4.069 min
Delta R.T.: 0.000 min
Response: 130847955
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1221ICC500



#10 AR-1221-3

R.T.: 4.062 min
Delta R.T.: 0.000 min
Response: 75841971
Conc: 500.00 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109978.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 15:53
Operator : YP/AJ
Sample : AR1232ICC500
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1232ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 18 16:25:04 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 18 16:24:54 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.692	3.689	438.9E6	262.6E6	50.000	50.000
2)	SA Decachlor...	8.744	8.696	372.1E6	116.2E6	50.000	50.000
	Target Compounds						
11)	L3 AR-1232-1	4.068	4.062	107.9E6	62046675	500.000	500.000
12)	L3 AR-1232-2	4.563	4.791	57393041	60174277	500.000	500.000
13)	L3 AR-1232-3	4.805	4.966	105.9E6	31828000	500.000	500.000
14)	L3 AR-1232-4	4.981	5.051	56780516	29670451	500.000	500.000
15)	L3 AR-1232-5	5.024	5.221	40226994	31608944	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109978.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 15:53
Operator : YP/AJ
Sample : AR1232ICC500
Misc :
ALS Vial : 9 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
AR1232ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 18 16:25:04 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 18 16:24:54 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.692 min
Delta R.T.: 0.000 min
Response: 438922587
Conc: 50.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1232ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.689 min
Delta R.T.: 0.000 min
Response: 262556927
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.744 min
Delta R.T.: 0.000 min
Response: 372070287
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.696 min
Delta R.T.: 0.000 min
Response: 116151811
Conc: 50.00 ng/ml



#11 AR-1232-1


R.T.: 4.068 min
Delta R.T.: 0.000 min
Response: 107872261
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1232ICC500




#11 AR-1232-1

R.T.: 4.062 min
Delta R.T.: 0.000 min
Response: 62046675
Conc: 500.00 ng/ml




#12 AR-1232-2

R.T.: 4.563 min
Delta R.T.: 0.000 min
Response: 57393041
Conc: 500.00 ng/ml



#12 AR-1232-2


R.T.: 4.791 min
Delta R.T.: 0.000 min
Response: 60174277
Conc: 500.00 ng/ml



#13 AR-1232-3


R.T.: 4.805 min
Delta R.T.: 0.000 min
Response: 105889863
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1232ICC500




#13 AR-1232-3

R.T.: 4.966 min
Delta R.T.: 0.000 min
Response: 31828000
Conc: 500.00 ng/ml




#14 AR-1232-4

R.T.: 4.981 min
Delta R.T.: 0.000 min
Response: 56780516
Conc: 500.00 ng/ml



#14 AR-1232-4


R.T.: 5.051 min
Delta R.T.: 0.000 min
Response: 29670451
Conc: 500.00 ng/ml



#15 AR-1232-5

R.T.: 5.024 min
Delta R.T.: 0.000 min
Response: 40226994
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1232ICC500



#15 AR-1232-5

R.T.: 5.221 min
Delta R.T.: 0.000 min
Response: 31608944
Conc: 500.00 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109979.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 16:11
Operator : YP/AJ
Sample : AR1242ICC1000
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 18 17:01:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 18 16:58:44 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.692	3.689	870.3E6	504.9E6	98.572	98.878
2)	SA Decachlor...	8.744	8.697	707.3E6	217.8E6	97.860	97.910
	Target Compounds						
16)	L4 AR-1242-1	4.785	4.772	257.0E6	141.3E6	967.387	967.723
17)	L4 AR-1242-2	4.804	4.791	361.7E6	206.0E6	974.253	976.732
18)	L4 AR-1242-3	4.861	4.966	239.3E6	108.8E6	957.525	967.388
19)	L4 AR-1242-4	4.981	5.051	191.3E6	107.6E6	962.368	956.424
20)	L4 AR-1242-5	5.635	5.572	200.3E6	132.5E6	953.851	956.059

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109979.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 16:11
Operator : YP/AJ
Sample : AR1242ICC1000
Misc :
ALS Vial : 10 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
AR1242ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 18 17:01:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 18 16:58:44 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.692 min
Delta R.T.: 0.000 min
Response: 870277303
Conc: 98.57 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC1000




#1 Tetrachloro-m-xylene

R.T.: 3.689 min
Delta R.T.: 0.000 min
Response: 504930415
Conc: 98.88 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.744 min
Delta R.T.: 0.000 min
Response: 707302147
Conc: 97.86 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.697 min
Delta R.T.: 0.000 min
Response: 217777313
Conc: 97.91 ng/ml



#16 AR-1242-1


R.T.: 4.785 min
Delta R.T.: 0.000 min
Response: 257021206
Conc: 967.39 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC1000




#16 AR-1242-1

R.T.: 4.772 min
Delta R.T.: 0.000 min
Response: 141331278
Conc: 967.72 ng/ml




#17 AR-1242-2

R.T.: 4.804 min
Delta R.T.: 0.000 min
Response: 361656206
Conc: 974.25 ng/ml



#17 AR-1242-2


R.T.: 4.791 min
Delta R.T.: 0.000 min
Response: 205951481
Conc: 976.73 ng/ml



#18 AR-1242-3


R.T.: 4.861 min
Delta R.T.: 0.000 min
Response: 239272274
Conc: 957.52 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC1000




#18 AR-1242-3

R.T.: 4.966 min
Delta R.T.: 0.000 min
Response: 108828803
Conc: 967.39 ng/ml




#19 AR-1242-4

R.T.: 4.981 min
Delta R.T.: 0.000 min
Response: 191269111
Conc: 962.37 ng/ml



#19 AR-1242-4


R.T.: 5.051 min
Delta R.T.: 0.000 min
Response: 107629937
Conc: 956.42 ng/ml



#20 AR-1242-5

R.T.: 5.635 min
Delta R.T.: 0.000 min
Response: 200342595
Conc: 953.85 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC1000



#20 AR-1242-5

R.T.: 5.572 min
Delta R.T.: 0.000 min
Response: 132486572
Conc: 956.06 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109980.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 16:30
Operator : YP/AJ
Sample : AR1242ICC750
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 18 17:04:28 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 18 16:58:44 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.691	3.689	674.7E6	388.3E6	75.937	75.688
2)	SA Decachlor...	8.744	8.696	546.6E6	166.9E6	75.416	75.037
	Target Compounds						
16)	L4 AR-1242-1	4.785	4.771	203.4E6	110.1E6	760.384	752.515
17)	L4 AR-1242-2	4.804	4.790	282.0E6	160.8E6	756.501	758.433
18)	L4 AR-1242-3	4.860	4.966	192.4E6	85755968	763.306	758.150
19)	L4 AR-1242-4	4.981	5.050	152.0E6	85465010	759.882	756.281
20)	L4 AR-1242-5	5.634	5.572	159.2E6	104.3E6	755.214	751.684

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109980.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 16:30
Operator : YP/AJ
Sample : AR1242ICC750
Misc :
ALS Vial : 11 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
AR1242ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 18 17:04:28 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 18 16:58:44 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.691 min
Delta R.T.: 0.000 min
Response: 674653935
Conc: 75.94 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC750




#1 Tetrachloro-m-xylene

R.T.: 3.689 min
Delta R.T.: 0.000 min
Response: 388290346
Conc: 75.69 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.744 min
Delta R.T.: 0.000 min
Response: 546604132
Conc: 75.42 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.696 min
Delta R.T.: 0.000 min
Response: 166944896
Conc: 75.04 ng/ml



#16 AR-1242-1


R.T.: 4.785 min
Delta R.T.: 0.000 min
Response: 203431528
Conc: 760.38 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC750




#16 AR-1242-1

R.T.: 4.771 min
Delta R.T.: 0.000 min
Response: 110085812
Conc: 752.52 ng/ml




#17 AR-1242-2

R.T.: 4.804 min
Delta R.T.: 0.000 min
Response: 282046233
Conc: 756.50 ng/ml



#17 AR-1242-2


R.T.: 4.790 min
Delta R.T.: 0.000 min
Response: 160825699
Conc: 758.43 ng/ml



#18 AR-1242-3


R.T.: 4.860 min
Delta R.T.: 0.000 min
Response: 192446698
Conc: 763.31 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC750




#18 AR-1242-3

R.T.: 4.966 min
Delta R.T.: 0.000 min
Response: 85755968
Conc: 758.15 ng/ml




#19 AR-1242-4

R.T.: 4.981 min
Delta R.T.: 0.000 min
Response: 152026829
Conc: 759.88 ng/ml



#19 AR-1242-4


R.T.: 5.050 min
Delta R.T.: 0.000 min
Response: 85465010
Conc: 756.28 ng/ml



#20 AR-1242-5

R.T.: 5.634 min
Delta R.T.: 0.000 min
Response: 159175050
Conc: 755.21 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC750



#20 AR-1242-5

R.T.: 5.572 min
Delta R.T.: 0.000 min
Response: 104282352
Conc: 751.68 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109981.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 16:48
Operator : YP/AJ
Sample : AR1242ICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 18 16:59:00 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 18 16:58:44 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.692	3.690	447.7E6	258.2E6	50.000	50.000
2)	SA Decachlor...	8.744	8.695	369.1E6	113.5E6	50.000	50.000
	Target Compounds						
16)	L4 AR-1242-1	4.785	4.772	137.2E6	75379512	500.000	500.000
17)	L4 AR-1242-2	4.805	4.791	190.4E6	107.9E6	500.000	500.000
18)	L4 AR-1242-3	4.861	4.967	130.3E6	58083170	500.000	500.000
19)	L4 AR-1242-4	4.982	5.051	103.1E6	58718718	500.000	500.000
20)	L4 AR-1242-5	5.635	5.573	109.9E6	72332482	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109981.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 16:48
Operator : YP/AJ
Sample : AR1242ICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
AR1242ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 18 16:59:00 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 18 16:58:44 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.692 min
Delta R.T.: 0.000 min
Response: 447743138
Conc: 50.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.690 min
Delta R.T.: 0.000 min
Response: 258194376
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.744 min
Delta R.T.: 0.000 min
Response: 369119192
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.695 min
Delta R.T.: 0.000 min
Response: 113538071
Conc: 50.00 ng/ml



#16 AR-1242-1


R.T.: 4.785 min
Delta R.T.: 0.000 min
Response: 137175388
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC500




#16 AR-1242-1

R.T.: 4.772 min
Delta R.T.: 0.000 min
Response: 75379512
Conc: 500.00 ng/ml




#17 AR-1242-2

R.T.: 4.805 min
Delta R.T.: 0.000 min
Response: 190385713
Conc: 500.00 ng/ml



#17 AR-1242-2


R.T.: 4.791 min
Delta R.T.: 0.000 min
Response: 107881886
Conc: 500.00 ng/ml



#18 AR-1242-3


R.T.: 4.861 min
Delta R.T.: 0.000 min
Response: 130250168
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC500




#18 AR-1242-3

R.T.: 4.967 min
Delta R.T.: 0.000 min
Response: 58083170
Conc: 500.00 ng/ml




#19 AR-1242-4

R.T.: 4.982 min
Delta R.T.: 0.000 min
Response: 103113773
Conc: 500.00 ng/ml



#19 AR-1242-4


R.T.: 5.051 min
Delta R.T.: 0.000 min
Response: 58718718
Conc: 500.00 ng/ml



#20 AR-1242-5

R.T.: 5.635 min
Delta R.T.: 0.000 min
Response: 109864205
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC500



#20 AR-1242-5

R.T.: 5.573 min
Delta R.T.: 0.000 min
Response: 72332482
Conc: 500.00 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109982.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 17:06
Operator : YP/AJ
Sample : AR1242ICC250
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 18 17:16:15 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 18 17:16:03 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.692	3.689	226.3E6	130.2E6	25.356	25.281
2)	SA Decachlor...	8.744	8.695	194.6E6	60893229	26.361	26.736
	Target Compounds						
16)	L4 AR-1242-1	4.785	4.772	71937746	39990799	263.903	267.124
17)	L4 AR-1242-2	4.804	4.791	99324236	56210353	262.106	261.143
18)	L4 AR-1242-3	4.860	4.967	69326183	30721999	268.271	265.862
19)	L4 AR-1242-4	4.981	5.051	53927655	31584987	264.381	271.488
20)	L4 AR-1242-5	5.634	5.572	57576429	38102015	266.987	268.040

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109982.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 17:06
Operator : YP/AJ
Sample : AR1242ICC250
Misc :
ALS Vial : 13 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
AR1242ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 18 17:16:15 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 18 17:16:03 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.692 min
Delta R.T.: 0.000 min
Response: 226344128
Conc: 25.36 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.689 min
Delta R.T.: 0.000 min
Response: 130179852
Conc: 25.28 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.744 min
Delta R.T.: 0.000 min
Response: 194588935
Conc: 26.36 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.695 min
Delta R.T.: 0.000 min
Response: 60893229
Conc: 26.74 ng/ml



#16 AR-1242-1


R.T.: 4.785 min
Delta R.T.: 0.000 min
Response: 71937746
Conc: 263.90 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC250




#16 AR-1242-1

R.T.: 4.772 min
Delta R.T.: 0.000 min
Response: 39990799
Conc: 267.12 ng/ml




#17 AR-1242-2

R.T.: 4.804 min
Delta R.T.: 0.000 min
Response: 99324236
Conc: 262.11 ng/ml



#17 AR-1242-2


R.T.: 4.791 min
Delta R.T.: 0.000 min
Response: 56210353
Conc: 261.14 ng/ml



#18 AR-1242-3


R.T.: 4.860 min
Delta R.T.: 0.000 min
Response: 69326183
Conc: 268.27 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC250




#18 AR-1242-3

R.T.: 4.967 min
Delta R.T.: 0.000 min
Response: 30721999
Conc: 265.86 ng/ml




#19 AR-1242-4

R.T.: 4.981 min
Delta R.T.: 0.000 min
Response: 53927655
Conc: 264.38 ng/ml



#19 AR-1242-4


R.T.: 5.051 min
Delta R.T.: 0.000 min
Response: 31584987
Conc: 271.49 ng/ml



#20 AR-1242-5

R.T.: 5.634 min
Delta R.T.: 0.000 min
Response: 57576429
Conc: 266.99 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC250



#20 AR-1242-5

R.T.: 5.572 min
Delta R.T.: 0.000 min
Response: 38102015
Conc: 268.04 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109983.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 17:25
Operator : YP/AJ
Sample : AR1242ICC050
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/19/2025
Supervised By :mohammad ahmed 03/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 18 17:34:54 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 18 17:34:29 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.691	3.688	33275417	17649519	3.978m	3.680m
2)	SA Decachlor...	8.744	8.696	32408304	10337597	4.500	4.624
	Target Compounds						
16)	L4 AR-1242-1	4.784	4.772	12333502	6612130	46.123	45.222
17)	L4 AR-1242-2	4.804	4.791	16261465	8959332	44.164	43.066
18)	L4 AR-1242-3	4.859	4.966	12804073	4792003	49.638	42.926m
19)	L4 AR-1242-4	4.980	5.051	9228050	5726073	46.259m	49.373
20)	L4 AR-1242-5	5.633	5.572	9633519	5880388	45.772m	42.708m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109983.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 17:25
Operator : YP/AJ
Sample : AR1242ICC050
Misc :
ALS Vial : 14 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
AR1242ICC050


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 03/19/2025
Supervised By :mohammad ahmed 03/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 18 17:34:54 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 18 17:34:29 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene

R.T.: 3.691 min
Delta R.T.: 0.000 min
Response: 33275417
Conc: 3.98 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC050


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/19/2025
Supervised By :mohammad ahmed 03/24/2025




#1 Tetrachloro-m-xylene

R.T.: 3.688 min
Delta R.T.: 0.000 min
Response: 17649519
Conc: 3.68 ng/ml m




#2 Decachlorobiphenyl

R.T.: 8.744 min
Delta R.T.: 0.000 min
Response: 32408304
Conc: 4.50 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.696 min
Delta R.T.: 0.000 min
Response: 10337597
Conc: 4.62 ng/ml




#16 AR-1242-1

R.T.: 4.784 min
Delta R.T.: 0.000 min
Response: 12333502
Conc: 46.12 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC050


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/19/2025
Supervised By :mohammad ahmed 03/24/2025




#16 AR-1242-1

R.T.: 4.772 min
Delta R.T.: 0.000 min
Response: 6612130
Conc: 45.22 ng/ml




#17 AR-1242-2

R.T.: 4.804 min
Delta R.T.: 0.000 min
Response: 16261465
Conc: 44.16 ng/ml



#17 AR-1242-2

R.T.: 4.791 min
Delta R.T.: 0.000 min
Response: 8959332
Conc: 43.07 ng/ml




#18 AR-1242-3

R.T.: 4.859 min
Delta R.T.: 0.000 min
Response: 12804073
Conc: 49.64 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC050


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/19/2025
Supervised By :mohammad ahmed 03/24/2025




#18 AR-1242-3

R.T.: 4.966 min
Delta R.T.: 0.000 min
Response: 4792003
Conc: 42.93 ng/ml m




#19 AR-1242-4

R.T.: 4.980 min
Delta R.T.: 0.000 min
Response: 9228050
Conc: 46.26 ng/ml m



#19 AR-1242-4

R.T.: 5.051 min
Delta R.T.: 0.000 min
Response: 5726073
Conc: 49.37 ng/ml




#20 AR-1242-5

R.T.: 5.633 min
Delta R.T.: 0.000 min
Response: 9633519
Conc: 45.77 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/19/2025
Supervised By :mohammad ahmed 03/24/2025



#20 AR-1242-5

R.T.: 5.572 min
Delta R.T.: 0.000 min
Response: 5880388
Conc: 42.71 ng/ml m

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109984.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 17:42
Operator : YP/AJ
Sample : AR1248ICC1000
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 01:40:07 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 01:38:33 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.690	3.689	889.7E6	511.9E6	95.913	96.672
2)	SA Decachlor...	8.742	8.696	728.0E6	214.7E6	93.393	91.574
	Target Compounds						
21)	L5 AR-1248-1	4.783	4.771	203.2E6	111.3E6	923.985	924.703
22)	L5 AR-1248-2	5.022	5.008	271.7E6	153.2E6	902.515	908.902
23)	L5 AR-1248-3	5.237	5.050	341.9E6	163.6E6	913.052	913.648
24)	L5 AR-1248-4	5.591	5.221	484.7E6	192.0E6	926.318	920.729
25)	L5 AR-1248-5	5.633	5.614	341.9E6	189.8E6	924.672	932.015


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109984.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 17:42
Operator : YP/AJ
Sample : AR1248ICC1000
Misc :
ALS Vial : 15 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
AR1248ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 01:40:07 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 01:38:33 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.690 min
Delta R.T.: -0.002 min
Response: 889690628
Conc: 95.91 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC1000




#1 Tetrachloro-m-xylene

R.T.: 3.689 min
Delta R.T.: 0.000 min
Response: 511894308
Conc: 96.67 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.742 min
Delta R.T.: -0.002 min
Response: 727980203
Conc: 93.39 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.696 min
Delta R.T.: 0.000 min
Response: 214708362
Conc: 91.57 ng/ml



#21 AR-1248-1


R.T.: 4.783 min
Delta R.T.: -0.002 min
Response: 203222777
Conc: 923.99 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC1000




#21 AR-1248-1

R.T.: 4.771 min
Delta R.T.: 0.000 min
Response: 111262575
Conc: 924.70 ng/ml




#22 AR-1248-2

R.T.: 5.022 min
Delta R.T.: -0.001 min
Response: 271717932
Conc: 902.51 ng/ml



#22 AR-1248-2


R.T.: 5.008 min
Delta R.T.: 0.000 min
Response: 153166873
Conc: 908.90 ng/ml



#23 AR-1248-3


R.T.: 5.237 min
Delta R.T.: -0.002 min
Response: 341855080
Conc: 913.05 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC1000




#23 AR-1248-3

R.T.: 5.050 min
Delta R.T.: 0.000 min
Response: 163639212
Conc: 913.65 ng/ml




#24 AR-1248-4

R.T.: 5.591 min
Delta R.T.: -0.001 min
Response: 484718685
Conc: 926.32 ng/ml



#24 AR-1248-4


R.T.: 5.221 min
Delta R.T.: 0.000 min
Response: 192034431
Conc: 920.73 ng/ml



#25 AR-1248-5

R.T.: 5.633 min
Delta R.T.: -0.001 min
Response: 341874536
Conc: 924.67 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC1000



#25 AR-1248-5

R.T.: 5.614 min
Delta R.T.: 0.000 min
Response: 189825244
Conc: 932.02 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109985.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 18:00
Operator : YP/AJ
Sample : AR1248ICC750
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 01:40:18 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 01:38:33 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.693	3.690	683.1E6	391.4E6	73.638	73.916
2)	SA Decachlor...	8.743	8.696	564.9E6	168.9E6	72.473	72.022
	Target Compounds						
21)	L5 AR-1248-1	4.785	4.772	158.1E6	86749863	718.942	720.978
22)	L5 AR-1248-2	5.024	5.009	213.9E6	120.1E6	710.557	712.732
23)	L5 AR-1248-3	5.239	5.051	267.9E6	128.2E6	715.615	715.920
24)	L5 AR-1248-4	5.593	5.222	377.8E6	150.5E6	721.979	721.526
25)	L5 AR-1248-5	5.635	5.614	266.1E6	147.5E6	719.788	724.105

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109985.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 18:00
Operator : YP/AJ
Sample : AR1248ICC750
Misc :
ALS Vial : 16 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
AR1248ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 01:40:18 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 01:38:33 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.693 min
Delta R.T.: 0.000 min
Response: 683064811
Conc: 73.64 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC750




#1 Tetrachloro-m-xylene

R.T.: 3.690 min
Delta R.T.: 0.000 min
Response: 391397784
Conc: 73.92 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.743 min
Delta R.T.: 0.000 min
Response: 564910095
Conc: 72.47 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.696 min
Delta R.T.: 0.000 min
Response: 168866978
Conc: 72.02 ng/ml



#21 AR-1248-1


R.T.: 4.785 min
Delta R.T.: 0.000 min
Response: 158125163
Conc: 718.94 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC750




#21 AR-1248-1

R.T.: 4.772 min
Delta R.T.: 0.000 min
Response: 86749863
Conc: 720.98 ng/ml




#22 AR-1248-2

R.T.: 5.024 min
Delta R.T.: 0.000 min
Response: 213925644
Conc: 710.56 ng/ml



#22 AR-1248-2


R.T.: 5.009 min
Delta R.T.: 0.000 min
Response: 120108575
Conc: 712.73 ng/ml



#23 AR-1248-3


R.T.: 5.239 min
Delta R.T.: 0.000 min
Response: 267932599
Conc: 715.61 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC750




#23 AR-1248-3

R.T.: 5.051 min
Delta R.T.: 0.000 min
Response: 128225021
Conc: 715.92 ng/ml




#24 AR-1248-4

R.T.: 5.593 min
Delta R.T.: 0.000 min
Response: 377793460
Conc: 721.98 ng/ml



#24 AR-1248-4


R.T.: 5.222 min
Delta R.T.: 0.000 min
Response: 150486943
Conc: 721.53 ng/ml



#25 AR-1248-5

R.T.: 5.635 min
Delta R.T.: 0.000 min
Response: 266123735
Conc: 719.79 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC750



#25 AR-1248-5

R.T.: 5.614 min
Delta R.T.: 0.001 min
Response: 147479675
Conc: 724.10 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109986.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 18:19
Operator : YP/AJ
Sample : AR1248ICC500
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 01:40:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 01:38:33 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.692	3.690	463.8E6	264.8E6	50.000	50.000
2)	SA Decachlor...	8.744	8.695	389.7E6	117.2E6	50.000	50.000
	Target Compounds						
21)	L5 AR-1248-1	4.785	4.771	110.0E6	60161217	500.000	500.000
22)	L5 AR-1248-2	5.023	5.008	150.5E6	84259256	500.000	500.000
23)	L5 AR-1248-3	5.239	5.051	187.2E6	89552637	500.000	500.000
24)	L5 AR-1248-4	5.593	5.221	261.6E6	104.3E6	500.000	500.000
25)	L5 AR-1248-5	5.634	5.613	184.9E6	101.8E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109986.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 18:19
Operator : YP/AJ
Sample : AR1248ICC500
Misc :
ALS Vial : 17 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
AR1248ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 01:40:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 01:38:33 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.692 min
Delta R.T.: 0.000 min
Response: 463799489
Conc: 50.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.690 min
Delta R.T.: 0.000 min
Response: 264757687
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.744 min
Delta R.T.: 0.000 min
Response: 389738327
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.695 min
Delta R.T.: 0.000 min
Response: 117232791
Conc: 50.00 ng/ml



#21 AR-1248-1


R.T.: 4.785 min
Delta R.T.: 0.000 min
Response: 109970762
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC500




#21 AR-1248-1

R.T.: 4.771 min
Delta R.T.: 0.000 min
Response: 60161217
Conc: 500.00 ng/ml




#22 AR-1248-2

R.T.: 5.023 min
Delta R.T.: 0.000 min
Response: 150533799
Conc: 500.00 ng/ml



#22 AR-1248-2


R.T.: 5.008 min
Delta R.T.: 0.000 min
Response: 84259256
Conc: 500.00 ng/ml



#23 AR-1248-3


R.T.: 5.239 min
Delta R.T.: 0.000 min
Response: 187204562
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC500




#23 AR-1248-3

R.T.: 5.051 min
Delta R.T.: 0.000 min
Response: 89552637
Conc: 500.00 ng/ml




#24 AR-1248-4

R.T.: 5.593 min
Delta R.T.: 0.000 min
Response: 261637332
Conc: 500.00 ng/ml



#24 AR-1248-4


R.T.: 5.221 min
Delta R.T.: 0.000 min
Response: 104283869
Conc: 500.00 ng/ml



#25 AR-1248-5

R.T.: 5.634 min
Delta R.T.: 0.000 min
Response: 184862645
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC500



#25 AR-1248-5

R.T.: 5.613 min
Delta R.T.: 0.000 min
Response: 101835896
Conc: 500.00 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109987.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 18:37
Operator : YP/AJ
Sample : AR1248ICC250
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 01:40:44 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 01:38:33 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.691	3.689	234.4E6	134.2E6	25.266	25.339
2)	SA Decachlor...	8.744	8.695	202.8E6	63399780	26.017	27.040
	Target Compounds						
21)	L5 AR-1248-1	4.784	4.771	57869632	31675925	263.114	263.259
22)	L5 AR-1248-2	5.023	5.008	80049855	44958781	265.887	266.788
23)	L5 AR-1248-3	5.238	5.050	98806066	47928872	263.899	267.602
24)	L5 AR-1248-4	5.592	5.221	136.0E6	55764342	259.818	267.368
25)	L5 AR-1248-5	5.634	5.613	96853624	53959011	261.961	264.931

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109987.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 18:37
Operator : YP/AJ
Sample : AR1248ICC250
Misc :
ALS Vial : 18 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
AR1248ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 01:40:44 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 01:38:33 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.691 min
Delta R.T.: 0.000 min
Response: 234365223
Conc: 25.27 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.689 min
Delta R.T.: 0.000 min
Response: 134173646
Conc: 25.34 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.744 min
Delta R.T.: 0.000 min
Response: 202795797
Conc: 26.02 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.695 min
Delta R.T.: 0.000 min
Response: 63399780
Conc: 27.04 ng/ml



#21 AR-1248-1


R.T.: 4.784 min
Delta R.T.: 0.000 min
Response: 57869632
Conc: 263.11 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC250




#21 AR-1248-1

R.T.: 4.771 min
Delta R.T.: 0.000 min
Response: 31675925
Conc: 263.26 ng/ml




#22 AR-1248-2

R.T.: 5.023 min
Delta R.T.: 0.000 min
Response: 80049855
Conc: 265.89 ng/ml



#22 AR-1248-2


R.T.: 5.008 min
Delta R.T.: 0.000 min
Response: 44958781
Conc: 266.79 ng/ml



#23 AR-1248-3


R.T.: 5.238 min
Delta R.T.: 0.000 min
Response: 98806066
Conc: 263.90 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC250




#23 AR-1248-3

R.T.: 5.050 min
Delta R.T.: 0.000 min
Response: 47928872
Conc: 267.60 ng/ml




#24 AR-1248-4

R.T.: 5.592 min
Delta R.T.: 0.000 min
Response: 135956218
Conc: 259.82 ng/ml



#24 AR-1248-4


R.T.: 5.221 min
Delta R.T.: 0.000 min
Response: 55764342
Conc: 267.37 ng/ml



#25 AR-1248-5

R.T.: 5.634 min
Delta R.T.: 0.000 min
Response: 96853624
Conc: 261.96 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC250



#25 AR-1248-5

R.T.: 5.613 min
Delta R.T.: 0.000 min
Response: 53959011
Conc: 264.93 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109988.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 18:54
Operator : YP/AJ
Sample : AR1248ICC050
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/19/2025
Supervised By :mohammad ahmed 03/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 01:40:58 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 01:38:33 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.690	3.689	43788277	25352976	4.721	4.788
2)	SA Decachlor...	8.742	8.695	41310164	13029410	5.300	5.557
	Target Compounds						
21)	L5 AR-1248-1	4.782	4.771	11846805	6686511	53.863	55.572
22)	L5 AR-1248-2	5.021	5.008	16385888	9031427	54.426	53.593
23)	L5 AR-1248-3	5.235	5.050	19860582	10026048	53.045m	55.979
24)	L5 AR-1248-4	5.591	5.221	29227773	11398927	55.856	54.653m
25)	L5 AR-1248-5	5.632	5.613	20545156	11669383	55.569	57.295

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109988.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 18:54
Operator : YP/AJ
Sample : AR1248ICC050
Misc :
ALS Vial : 19 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
AR1248ICC050


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 03/19/2025
Supervised By :mohammad ahmed 03/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 01:40:58 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 01:38:33 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene

R.T.: 3.690 min
Delta R.T.: -0.002 min
Response: 43788277
Conc: 4.72 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC050


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/19/2025
Supervised By :mohammad ahmed 03/24/2025




#1 Tetrachloro-m-xylene

R.T.: 3.689 min
Delta R.T.: 0.000 min
Response: 25352976
Conc: 4.79 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.742 min
Delta R.T.: -0.002 min
Response: 41310164
Conc: 5.30 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.695 min
Delta R.T.: 0.000 min
Response: 13029410
Conc: 5.56 ng/ml




#21 AR-1248-1

R.T.: 4.782 min
Delta R.T.: -0.003 min
Response: 11846805
Conc: 53.86 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC050


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/19/2025
Supervised By :mohammad ahmed 03/24/2025




#21 AR-1248-1

R.T.: 4.771 min
Delta R.T.: 0.000 min
Response: 6686511
Conc: 55.57 ng/ml




#22 AR-1248-2

R.T.: 5.021 min
Delta R.T.: -0.002 min
Response: 16385888
Conc: 54.43 ng/ml



#22 AR-1248-2

R.T.: 5.008 min
Delta R.T.: 0.000 min
Response: 9031427
Conc: 53.59 ng/ml




#23 AR-1248-3

R.T.: 5.235 min
Delta R.T.: -0.003 min
Response: 19860582
Conc: 53.05 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC050


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/19/2025
Supervised By :mohammad ahmed 03/24/2025




#23 AR-1248-3

R.T.: 5.050 min
Delta R.T.: 0.000 min
Response: 10026048
Conc: 55.98 ng/ml




#24 AR-1248-4

R.T.: 5.591 min
Delta R.T.: -0.002 min
Response: 29227773
Conc: 55.86 ng/ml



#24 AR-1248-4

R.T.: 5.221 min
Delta R.T.: 0.000 min
Response: 11398927
Conc: 54.65 ng/ml m




#25 AR-1248-5

R.T.: 5.632 min
Delta R.T.: -0.002 min
Response: 20545156
Conc: 55.57 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/19/2025
Supervised By :mohammad ahmed 03/24/2025



#25 AR-1248-5

R.T.: 5.613 min
Delta R.T.: 0.000 min
Response: 11669383
Conc: 57.30 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109989.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 19:13
Operator : YP/AJ
Sample : AR1254ICC1000
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 01:49:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 01:48:40 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.692	3.690	912.4E6	520.8E6	97.196	97.394
2)	SA Decachlor...	8.744	8.694	750.7E6	219.7E6	95.415	90.831
	Target Compounds						
26)	L6 AR-1254-1	5.593	5.573	527.5E6	289.8E6	943.049	941.799
27)	L6 AR-1254-2	5.742	5.721	454.9E6	250.7E6	940.470	933.876
28)	L6 AR-1254-3	6.147	6.123	753.6E6	402.2E6	954.831	954.778
29)	L6 AR-1254-4	6.376	6.351	465.4E6	231.8E6	956.772	948.684
30)	L6 AR-1254-5	6.797	6.768	652.0E6	329.3E6	958.152	944.466


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109989.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 19:13
Operator : YP/AJ
Sample : AR1254ICC1000
Misc :
ALS Vial : 20 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
AR1254ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 01:49:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 01:48:40 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.692 min
Delta R.T.: 0.000 min
Response: 912401749
Conc: 97.20 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC1000




#1 Tetrachloro-m-xylene

R.T.: 3.690 min
Delta R.T.: 0.000 min
Response: 520755645
Conc: 97.39 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.744 min
Delta R.T.: 0.000 min
Response: 750657739
Conc: 95.42 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.694 min
Delta R.T.: -0.001 min
Response: 219680790
Conc: 90.83 ng/ml



#26 AR-1254-1


R.T.: 5.593 min
Delta R.T.: 0.000 min
Response: 527461301
Conc: 943.05 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC1000




#26 AR-1254-1

R.T.: 5.573 min
Delta R.T.: 0.000 min
Response: 289813236
Conc: 941.80 ng/ml




#27 AR-1254-2

R.T.: 5.742 min
Delta R.T.: 0.000 min
Response: 454907170
Conc: 940.47 ng/ml



#27 AR-1254-2


R.T.: 5.721 min
Delta R.T.: 0.000 min
Response: 250736607
Conc: 933.88 ng/ml



#28 AR-1254-3


R.T.: 6.147 min
Delta R.T.: 0.000 min
Response: 753626169
Conc: 954.83 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC1000




#28 AR-1254-3

R.T.: 6.123 min
Delta R.T.: 0.000 min
Response: 402159669
Conc: 954.78 ng/ml




#29 AR-1254-4

R.T.: 6.376 min
Delta R.T.: -0.002 min
Response: 465418236
Conc: 956.77 ng/ml



#29 AR-1254-4


R.T.: 6.351 min
Delta R.T.: 0.000 min
Response: 231846868
Conc: 948.68 ng/ml



#30 AR-1254-5

R.T.: 6.797 min
Delta R.T.: 0.000 min
Response: 652046620
Conc: 958.15 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC1000



#30 AR-1254-5

R.T.: 6.768 min
Delta R.T.: 0.000 min
Response: 329301884
Conc: 944.47 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109990.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 19:31
Operator : YP/AJ
Sample : AR1254ICC750
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 01:49:53 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 01:48:40 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.691	3.689	694.8E6	396.7E6	74.017	74.184
2)	SA Decachlor...	8.742	8.695	579.2E6	170.7E6	73.621	70.591
	Target Compounds						
26)	L6 AR-1254-1	5.593	5.572	406.4E6	224.1E6	726.602	728.104
27)	L6 AR-1254-2	5.742	5.720	351.8E6	194.7E6	727.216	725.242
28)	L6 AR-1254-3	6.148	6.122	579.5E6	310.1E6	734.225	736.114
29)	L6 AR-1254-4	6.376	6.350	358.3E6	178.4E6	736.667	729.972
30)	L6 AR-1254-5	6.796	6.768	499.8E6	253.4E6	734.486	726.842


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109990.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 19:31
Operator : YP/AJ
Sample : AR1254ICC750
Misc :
ALS Vial : 21 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
AR1254ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 01:49:53 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 01:48:40 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.691 min
Delta R.T.: 0.000 min
Response: 694816651
Conc: 74.02 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC750




#1 Tetrachloro-m-xylene

R.T.: 3.689 min
Delta R.T.: 0.000 min
Response: 396654527
Conc: 74.18 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.742 min
Delta R.T.: -0.001 min
Response: 579195225
Conc: 73.62 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.695 min
Delta R.T.: 0.000 min
Response: 170728292
Conc: 70.59 ng/ml



#26 AR-1254-1


R.T.: 5.593 min
Delta R.T.: 0.000 min
Response: 406399307
Conc: 726.60 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC750




#26 AR-1254-1

R.T.: 5.572 min
Delta R.T.: 0.000 min
Response: 224054374
Conc: 728.10 ng/ml




#27 AR-1254-2

R.T.: 5.742 min
Delta R.T.: -0.001 min
Response: 351755893
Conc: 727.22 ng/ml



#27 AR-1254-2


R.T.: 5.720 min
Delta R.T.: 0.000 min
Response: 194720485
Conc: 725.24 ng/ml



#28 AR-1254-3


R.T.: 6.148 min
Delta R.T.: 0.000 min
Response: 579506791
Conc: 734.22 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC750




#28 AR-1254-3

R.T.: 6.122 min
Delta R.T.: 0.000 min
Response: 310056844
Conc: 736.11 ng/ml




#29 AR-1254-4

R.T.: 6.376 min
Delta R.T.: -0.001 min
Response: 358349037
Conc: 736.67 ng/ml



#29 AR-1254-4


R.T.: 6.350 min
Delta R.T.: 0.000 min
Response: 178396380
Conc: 729.97 ng/ml



#30 AR-1254-5

R.T.: 6.796 min
Delta R.T.: 0.000 min
Response: 499835944
Conc: 734.49 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC750



#30 AR-1254-5

R.T.: 6.768 min
Delta R.T.: 0.000 min
Response: 253424010
Conc: 726.84 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109991.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 19:49
Operator : YP/AJ
Sample : AR1254ICC500
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 01:52:43 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 01:48:40 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.692	3.689	469.4E6	267.3E6	50.000	50.000
2)	SA Decachlor...	8.744	8.695	393.4E6	120.9E6	50.000	50.000
	Target Compounds						
26)	L6 AR-1254-1	5.594	5.573	279.7E6	153.9E6	500.000	500.000
27)	L6 AR-1254-2	5.743	5.720	241.9E6	134.2E6	500.000	500.000
28)	L6 AR-1254-3	6.148	6.123	394.6E6	210.6E6	500.000	500.000
29)	L6 AR-1254-4	6.378	6.351	243.2E6	122.2E6	500.000	500.000
30)	L6 AR-1254-5	6.797	6.768	340.3E6	174.3E6	500.000	500.000


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109991.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 19:49
Operator : YP/AJ
Sample : AR1254ICC500
Misc :
ALS Vial : 22 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
AR1254ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 01:52:43 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 01:48:40 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.692 min
Delta R.T.: 0.000 min
Response: 469362086
Conc: 50.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.689 min
Delta R.T.: 0.000 min
Response: 267345850
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.744 min
Delta R.T.: 0.000 min
Response: 393362889
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.695 min
Delta R.T.: 0.000 min
Response: 120928616
Conc: 50.00 ng/ml



#26 AR-1254-1


R.T.: 5.594 min
Delta R.T.: 0.000 min
Response: 279657443
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC500




#26 AR-1254-1

R.T.: 5.573 min
Delta R.T.: 0.000 min
Response: 153861440
Conc: 500.00 ng/ml




#27 AR-1254-2

R.T.: 5.743 min
Delta R.T.: 0.000 min
Response: 241851007
Conc: 500.00 ng/ml



#27 AR-1254-2


R.T.: 5.720 min
Delta R.T.: 0.000 min
Response: 134245134
Conc: 500.00 ng/ml



#28 AR-1254-3


R.T.: 6.148 min
Delta R.T.: 0.000 min
Response: 394638676
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC500




#28 AR-1254-3

R.T.: 6.123 min
Delta R.T.: 0.000 min
Response: 210603765
Conc: 500.00 ng/ml




#29 AR-1254-4

R.T.: 6.378 min
Delta R.T.: 0.000 min
Response: 243223195
Conc: 500.00 ng/ml



#29 AR-1254-4


R.T.: 6.351 min
Delta R.T.: 0.000 min
Response: 122193976
Conc: 500.00 ng/ml



#30 AR-1254-5

R.T.: 6.797 min
Delta R.T.: 0.000 min
Response: 340262601
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC500



#30 AR-1254-5

R.T.: 6.768 min
Delta R.T.: 0.000 min
Response: 174332293
Conc: 500.00 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109992.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 20:07
Operator : YP/AJ
Sample : AR1254ICC250
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 01:50:17 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 01:48:40 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.690	3.689	218.3E6	126.3E6	23.256	23.621
2)	SA Decachlor...	8.743	8.695	201.1E6	62915084	25.556	26.013
	Target Compounds						
26)	L6 AR-1254-1	5.592	5.572	142.8E6	78551215	255.230	255.266
27)	L6 AR-1254-2	5.741	5.720	117.7E6	67765666	243.403	252.395
28)	L6 AR-1254-3	6.146	6.123	190.0E6	103.9E6	240.780	246.647
29)	L6 AR-1254-4	6.376	6.351	124.9E6	61546138	256.767	251.838
30)	L6 AR-1254-5	6.795	6.768	170.5E6	90324740	250.587	259.059

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109992.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 20:07
Operator : YP/AJ
Sample : AR1254ICC250
Misc :
ALS Vial : 23 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
AR1254ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 01:50:17 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 01:48:40 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.690 min
Delta R.T.: -0.002 min
Response: 218311930
Conc: 23.26 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.689 min
Delta R.T.: 0.000 min
Response: 126298911
Conc: 23.62 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.743 min
Delta R.T.: 0.000 min
Response: 201053654
Conc: 25.56 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.695 min
Delta R.T.: 0.000 min
Response: 62915084
Conc: 26.01 ng/ml



#26 AR-1254-1


R.T.: 5.592 min
Delta R.T.: -0.002 min
Response: 142754156
Conc: 255.23 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC250




#26 AR-1254-1

R.T.: 5.572 min
Delta R.T.: 0.000 min
Response: 78551215
Conc: 255.27 ng/ml




#27 AR-1254-2

R.T.: 5.741 min
Delta R.T.: -0.002 min
Response: 117734646
Conc: 243.40 ng/ml



#27 AR-1254-2


R.T.: 5.720 min
Delta R.T.: 0.000 min
Response: 67765666
Conc: 252.40 ng/ml



#28 AR-1254-3


R.T.: 6.146 min
Delta R.T.: -0.002 min
Response: 190042575
Conc: 240.78 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC250




#28 AR-1254-3

R.T.: 6.123 min
Delta R.T.: 0.000 min
Response: 103889593
Conc: 246.65 ng/ml




#29 AR-1254-4

R.T.: 6.376 min
Delta R.T.: -0.002 min
Response: 124903516
Conc: 256.77 ng/ml



#29 AR-1254-4


R.T.: 6.351 min
Delta R.T.: 0.000 min
Response: 61546138
Conc: 251.84 ng/ml



#30 AR-1254-5

R.T.: 6.795 min
Delta R.T.: -0.002 min
Response: 170531002
Conc: 250.59 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC250



#30 AR-1254-5

R.T.: 6.768 min
Delta R.T.: 0.000 min
Response: 90324740
Conc: 259.06 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109993.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 20:25
Operator : YP/AJ
Sample : AR1254ICC050
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/19/2025
Supervised By :mohammad ahmed 03/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 01:50:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 01:48:40 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.692	3.689	43445665	25306006	4.628	4.733
2)	SA Decachlor...	8.743	8.694	41266656	13227136	5.245	5.469
	Target Compounds						
26)	L6 AR-1254-1	5.593	5.573	31329450	17649915	56.014	57.357
27)	L6 AR-1254-2	5.742	5.720	26381750	15388259	54.541	57.314
28)	L6 AR-1254-3	6.148	6.123	39484048	21241930	50.026	50.431
29)	L6 AR-1254-4	6.377	6.350	23571245	12628885	48.456	51.676
30)	L6 AR-1254-5	6.797	6.768	28479203	15116419	41.849	43.355m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109993.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 20:25
Operator : YP/AJ
Sample : AR1254ICC050
Misc :
ALS Vial : 24 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
AR1254ICC050


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 03/19/2025
Supervised By :mohammad ahmed 03/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 01:50:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 01:48:40 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene

R.T.: 3.692 min
Delta R.T.: 0.000 min
Response: 43445665
Conc: 4.63 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC050


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/19/2025
Supervised By :mohammad ahmed 03/24/2025




#1 Tetrachloro-m-xylene

R.T.: 3.689 min
Delta R.T.: 0.000 min
Response: 25306006
Conc: 4.73 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.743 min
Delta R.T.: 0.000 min
Response: 41266656
Conc: 5.25 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.694 min
Delta R.T.: -0.002 min
Response: 13227136
Conc: 5.47 ng/ml




#26 AR-1254-1

R.T.: 5.593 min
Delta R.T.: 0.000 min
Response: 31329450
Conc: 56.01 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC050


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/19/2025
Supervised By :mohammad ahmed 03/24/2025




#26 AR-1254-1

R.T.: 5.573 min
Delta R.T.: 0.000 min
Response: 17649915
Conc: 57.36 ng/ml




#27 AR-1254-2

R.T.: 5.742 min
Delta R.T.: 0.000 min
Response: 26381750
Conc: 54.54 ng/ml



#27 AR-1254-2

R.T.: 5.720 min
Delta R.T.: 0.000 min
Response: 15388259
Conc: 57.31 ng/ml




#28 AR-1254-3

R.T.: 6.148 min
Delta R.T.: 0.000 min
Response: 39484048
Conc: 50.03 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC050


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/19/2025
Supervised By :mohammad ahmed 03/24/2025




#28 AR-1254-3

R.T.: 6.123 min
Delta R.T.: 0.000 min
Response: 21241930
Conc: 50.43 ng/ml




#29 AR-1254-4

R.T.: 6.377 min
Delta R.T.: 0.000 min
Response: 23571245
Conc: 48.46 ng/ml



#29 AR-1254-4

R.T.: 6.350 min
Delta R.T.: 0.000 min
Response: 12628885
Conc: 51.68 ng/ml




#30 AR-1254-5

R.T.: 6.797 min
Delta R.T.: 0.000 min
Response: 28479203
Conc: 41.85 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/19/2025
Supervised By :mohammad ahmed 03/24/2025



#30 AR-1254-5

R.T.: 6.768 min
Delta R.T.: 0.000 min
Response: 15116419
Conc: 43.36 ng/ml m

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109994.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 20:43
Operator : YP/AJ
Sample : AR1262ICC500
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1262ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 01:59:58 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 01:59:23 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.692	3.689	443.4E6	253.7E6	50.000	50.000
2)	SA Decachlor...	8.744	8.695	383.2E6	113.8E6	50.000	50.000
	Target Compounds						
36)	L8 AR-1262-1	6.836	6.807	344.1E6	172.2E6	500.000	500.000
37)	L8 AR-1262-2	7.338	7.306	596.0E6	268.6E6	500.000	500.000
38)	L8 AR-1262-3	7.623	7.590	235.3E6	97025324	500.000	500.000
39)	L8 AR-1262-4	7.687	7.654	437.3E6	180.9E6	500.000	500.000
40)	L8 AR-1262-5	8.184	8.146	197.2E6	70580972	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109994.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 20:43
Operator : YP/AJ
Sample : AR1262ICC500
Misc :
ALS Vial : 25 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
AR1262ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 01:59:58 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 01:59:23 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.692 min
Delta R.T.: 0.000 min
Response: 443392340
Conc: 50.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1262ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.689 min
Delta R.T.: 0.000 min
Response: 253657981
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.744 min
Delta R.T.: 0.000 min
Response: 383169446
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.695 min
Delta R.T.: 0.000 min
Response: 113848015
Conc: 50.00 ng/ml



#36 AR-1262-1


R.T.: 6.836 min
Delta R.T.: 0.000 min
Response: 344060290
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1262ICC500




#36 AR-1262-1

R.T.: 6.807 min
Delta R.T.: 0.000 min
Response: 172156478
Conc: 500.00 ng/ml




#37 AR-1262-2

R.T.: 7.338 min
Delta R.T.: 0.000 min
Response: 595986678
Conc: 500.00 ng/ml



#37 AR-1262-2


R.T.: 7.306 min
Delta R.T.: 0.000 min
Response: 268610130
Conc: 500.00 ng/ml



#38 AR-1262-3


R.T.: 7.623 min
Delta R.T.: 0.000 min
Response: 235314860
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1262ICC500




#38 AR-1262-3

R.T.: 7.590 min
Delta R.T.: 0.000 min
Response: 97025324
Conc: 500.00 ng/ml




#39 AR-1262-4

R.T.: 7.687 min
Delta R.T.: 0.000 min
Response: 437287698
Conc: 500.00 ng/ml



#39 AR-1262-4


R.T.: 7.654 min
Delta R.T.: 0.000 min
Response: 180893327
Conc: 500.00 ng/ml



#40 AR-1262-5

R.T.: 8.184 min
Delta R.T.: 0.000 min
Response: 197221208
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1262ICC500



#40 AR-1262-5

R.T.: 8.146 min
Delta R.T.: 0.000 min
Response: 70580972
Conc: 500.00 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109995.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 21:02
Operator : YP/AJ
Sample : AR1268ICC1000
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1268ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 02:06:17 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:04:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.692	3.689	901.7E6	515.8E6	97.188	97.661
2)	SA Decachlor...	8.743	8.696	1364.0E6	390.3E6	96.599	94.170
	Target Compounds						
41)	L9 AR-1268-1	7.623	7.590	1369.8E6	560.8E6	976.135	969.827
42)	L9 AR-1268-2	7.688	7.655	1266.3E6	519.2E6	989.578	974.786
43)	L9 AR-1268-3	7.896	7.862	1019.7E6	397.6E6	981.758	963.023
44)	L9 AR-1268-4	8.183	8.147	432.7E6	153.8E6	967.005	949.196
45)	L9 AR-1268-5	8.481	8.440	3202.1E6	963.7E6	991.799	976.212

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109995.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 21:02
Operator : YP/AJ
Sample : AR1268ICC1000
Misc :
ALS Vial : 26 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
AR1268ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 02:06:17 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:04:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.692 min
Delta R.T.: 0.000 min
Response: 901662125
Conc: 97.19 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC1000




#1 Tetrachloro-m-xylene

R.T.: 3.689 min
Delta R.T.: 0.000 min
Response: 515836411
Conc: 97.66 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.743 min
Delta R.T.: 0.000 min
Response: 1364000629
Conc: 96.60 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.696 min
Delta R.T.: 0.002 min
Response: 390263390
Conc: 94.17 ng/ml



#41 AR-1268-1


R.T.: 7.623 min
Delta R.T.: 0.000 min
Response: 1369838059
Conc: 976.13 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC1000




#41 AR-1268-1

R.T.: 7.590 min
Delta R.T.: 0.001 min
Response: 560759375
Conc: 969.83 ng/ml




#42 AR-1268-2

R.T.: 7.688 min
Delta R.T.: 0.000 min
Response: 1266296617
Conc: 989.58 ng/ml



#42 AR-1268-2


R.T.: 7.655 min
Delta R.T.: 0.002 min
Response: 519206129
Conc: 974.79 ng/ml



#43 AR-1268-3


R.T.: 7.896 min
Delta R.T.: 0.000 min
Response: 1019701747
Conc: 981.76 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC1000




#43 AR-1268-3

R.T.: 7.862 min
Delta R.T.: 0.001 min
Response: 397648794
Conc: 963.02 ng/ml




#44 AR-1268-4

R.T.: 8.183 min
Delta R.T.: -0.001 min
Response: 432675859
Conc: 967.01 ng/ml



#44 AR-1268-4


R.T.: 8.147 min
Delta R.T.: 0.002 min
Response: 153792414
Conc: 949.20 ng/ml



#45 AR-1268-5

R.T.: 8.481 min
Delta R.T.: -0.001 min
Response: 3202130253
Conc: 991.80 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC1000



#45 AR-1268-5

R.T.: 8.440 min
Delta R.T.: 0.001 min
Response: 963659212
Conc: 976.21 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109996.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 21:20
Operator : YP/AJ
Sample : AR1268ICC750
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1268ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 02:06:27 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:04:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1) SA	Tetrachlo...	3.692	3.689	674.6E6	386.9E6	72.709	73.243
2) SA	Decachlor...	8.743	8.696	1030.2E6	298.8E6	72.957	72.092
	Target Compounds						
41) L9	AR-1268-1	7.624	7.590	1030.8E6	426.7E6	734.560	738.041
42) L9	AR-1268-2	7.688	7.655	947.1E6	394.0E6	740.097	739.739
43) L9	AR-1268-3	7.897	7.862	764.7E6	299.7E6	736.208	725.732
44) L9	AR-1268-4	8.184	8.147	316.0E6	116.0E6	706.276	715.999
45) L9	AR-1268-5	8.482	8.440	2394.2E6	724.9E6	741.545	734.374


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109996.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 21:20
Operator : YP/AJ
Sample : AR1268ICC750
Misc :
ALS Vial : 27 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
AR1268ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 02:06:27 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:04:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.692 min
Delta R.T.: 0.000 min
Response: 674561518
Conc: 72.71 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC750




#1 Tetrachloro-m-xylene

R.T.: 3.689 min
Delta R.T.: 0.000 min
Response: 386863871
Conc: 73.24 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.743 min
Delta R.T.: 0.000 min
Response: 1030169782
Conc: 72.96 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.696 min
Delta R.T.: 0.002 min
Response: 298767454
Conc: 72.09 ng/ml



#41 AR-1268-1


R.T.: 7.624 min
Delta R.T.: 0.002 min
Response: 1030829535
Conc: 734.56 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC750




#41 AR-1268-1

R.T.: 7.590 min
Delta R.T.: 0.001 min
Response: 426739530
Conc: 738.04 ng/ml




#42 AR-1268-2

R.T.: 7.688 min
Delta R.T.: 0.000 min
Response: 947052725
Conc: 740.10 ng/ml



#42 AR-1268-2


R.T.: 7.655 min
Delta R.T.: 0.002 min
Response: 394011614
Conc: 739.74 ng/ml



#43 AR-1268-3


R.T.: 7.897 min
Delta R.T.: 0.000 min
Response: 764660970
Conc: 736.21 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC750




#43 AR-1268-3

R.T.: 7.862 min
Delta R.T.: 0.002 min
Response: 299667225
Conc: 725.73 ng/ml




#44 AR-1268-4

R.T.: 8.184 min
Delta R.T.: 0.000 min
Response: 316015601
Conc: 706.28 ng/ml



#44 AR-1268-4


R.T.: 8.147 min
Delta R.T.: 0.002 min
Response: 116008994
Conc: 716.00 ng/ml



#45 AR-1268-5

R.T.: 8.482 min
Delta R.T.: 0.000 min
Response: 2394158307
Conc: 741.55 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC750



#45 AR-1268-5

R.T.: 8.440 min
Delta R.T.: 0.000 min
Response: 724930761
Conc: 734.37 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109997.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 21:39
Operator : YP/AJ
Sample : AR1268ICC500
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1268ICC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/19/2025
Supervised By :mohammad ahmed 03/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 02:06:41 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:04:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.692	3.689	463.9E6	264.1E6	50.000	50.000
2)	SA Decachlor...	8.743	8.694	706.0E6	207.2E6	50.000	50.000
	Target Compounds						
41)	L9 AR-1268-1	7.622	7.589	700.7E6	289.1E6	499.305m	500.000
42)	L9 AR-1268-2	7.688	7.653	639.8E6	266.3E6	500.000	500.000
43)	L9 AR-1268-3	7.896	7.861	519.3E6	206.5E6	500.000	500.000
44)	L9 AR-1268-4	8.185	8.145	223.7E6	81011977	500.000	500.000
45)	L9 AR-1268-5	8.482	8.439	1614.3E6	493.6E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109997.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 21:39
Operator : YP/AJ
Sample : AR1268ICC500
Misc :
ALS Vial : 28 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
AR1268ICC500


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 03/19/2025
Supervised By :mohammad ahmed 03/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 02:06:41 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:04:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene

R.T.: 3.692 min
Delta R.T.: 0.000 min
Response: 463875726
Conc: 50.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC500


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/19/2025
Supervised By :mohammad ahmed 03/24/2025




#1 Tetrachloro-m-xylene

R.T.: 3.689 min
Delta R.T.: 0.000 min
Response: 264095871
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.743 min
Delta R.T.: 0.000 min
Response: 706011522
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.694 min
Delta R.T.: 0.000 min
Response: 207212938
Conc: 50.00 ng/ml




#41 AR-1268-1

R.T.: 7.622 min
Delta R.T.: 0.000 min
Response: 700689003
Conc: 499.31 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC500


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/19/2025
Supervised By :mohammad ahmed 03/24/2025




#41 AR-1268-1

R.T.: 7.589 min
Delta R.T.: 0.000 min
Response: 289102857
Conc: 500.00 ng/ml




#42 AR-1268-2

R.T.: 7.688 min
Delta R.T.: 0.000 min
Response: 639816344
Conc: 500.00 ng/ml



#42 AR-1268-2

R.T.: 7.653 min
Delta R.T.: 0.000 min
Response: 266318012
Conc: 500.00 ng/ml




#43 AR-1268-3

R.T.: 7.896 min
Delta R.T.: 0.000 min
Response: 519324293
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC500


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/19/2025
Supervised By :mohammad ahmed 03/24/2025




#43 AR-1268-3

R.T.: 7.861 min
Delta R.T.: 0.000 min
Response: 206458519
Conc: 500.00 ng/ml




#44 AR-1268-4

R.T.: 8.185 min
Delta R.T.: 0.000 min
Response: 223719543
Conc: 500.00 ng/ml



#44 AR-1268-4

R.T.: 8.145 min
Delta R.T.: 0.000 min
Response: 81011977
Conc: 500.00 ng/ml




#45 AR-1268-5

R.T.: 8.482 min
Delta R.T.: 0.000 min
Response: 1614304087
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/19/2025
Supervised By :mohammad ahmed 03/24/2025



#45 AR-1268-5

R.T.: 8.439 min
Delta R.T.: 0.000 min
Response: 493570508
Conc: 500.00 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109998.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 21:57
Operator : YP/AJ
Sample : AR1268ICC250
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1268ICC250

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/19/2025
Supervised By :mohammad ahmed 03/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 02:06:56 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:04:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.692	3.690	233.0E6	133.0E6	25.112	25.182
2)	SA Decachlor...	8.743	8.695	364.2E6	110.3E6	25.791	26.624
	Target Compounds						
41)	L9 AR-1268-1	7.623	7.590	357.9E6	149.6E6	255.005m	258.814
42)	L9 AR-1268-2	7.688	7.654	330.4E6	137.2E6	258.196	257.558
43)	L9 AR-1268-3	7.897	7.861	264.1E6	107.0E6	254.288	259.017
44)	L9 AR-1268-4	8.184	8.146	118.3E6	42174671	264.470	260.299
45)	L9 AR-1268-5	8.482	8.439	811.9E6	254.0E6	251.483	257.325

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109998.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 21:57
Operator : YP/AJ
Sample : AR1268ICC250
Misc :
ALS Vial : 29 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
AR1268ICC250


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 03/19/2025
Supervised By :mohammad ahmed 03/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 02:06:56 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:04:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene

R.T.: 3.692 min
Delta R.T.: 0.000 min
Response: 232978846
Conc: 25.11 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC250


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/19/2025
Supervised By :mohammad ahmed 03/24/2025




#1 Tetrachloro-m-xylene

R.T.: 3.690 min
Delta R.T.: 0.000 min
Response: 133010597
Conc: 25.18 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.743 min
Delta R.T.: 0.000 min
Response: 364169510
Conc: 25.79 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.695 min
Delta R.T.: 0.001 min
Response: 110337267
Conc: 26.62 ng/ml




#41 AR-1268-1

R.T.: 7.623 min
Delta R.T.: 0.000 min
Response: 357855660
Conc: 255.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC250


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/19/2025
Supervised By :mohammad ahmed 03/24/2025




#41 AR-1268-1

R.T.: 7.590 min
Delta R.T.: 0.000 min
Response: 149647905
Conc: 258.81 ng/ml




#42 AR-1268-2

R.T.: 7.688 min
Delta R.T.: 0.000 min
Response: 330395950
Conc: 258.20 ng/ml



#42 AR-1268-2

R.T.: 7.654 min
Delta R.T.: 0.001 min
Response: 137184821
Conc: 257.56 ng/ml




#43 AR-1268-3

R.T.: 7.897 min
Delta R.T.: 0.000 min
Response: 264115362
Conc: 254.29 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC250


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/19/2025
Supervised By :mohammad ahmed 03/24/2025




#43 AR-1268-3

R.T.: 7.861 min
Delta R.T.: 0.000 min
Response: 106952452
Conc: 259.02 ng/ml




#44 AR-1268-4

R.T.: 8.184 min
Delta R.T.: 0.000 min
Response: 118334390
Conc: 264.47 ng/ml



#44 AR-1268-4

R.T.: 8.146 min
Delta R.T.: 0.000 min
Response: 42174671
Conc: 260.30 ng/ml




#45 AR-1268-5

R.T.: 8.482 min
Delta R.T.: 0.000 min
Response: 811940473
Conc: 251.48 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC250

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/19/2025
Supervised By :mohammad ahmed 03/24/2025



#45 AR-1268-5

R.T.: 8.439 min
Delta R.T.: 0.000 min
Response: 254015607
Conc: 257.32 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109999.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 22:15
Operator : YP/AJ
Sample : AR1268ICC050
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1268ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/19/2025
Supervised By :mohammad ahmed 03/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 02:07:11 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:04:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.691	3.688	44374914	25645349	4.783m	4.855m
2)	SA Decachlor...	8.743	8.695	72973489	22764198	5.168	5.493
	Target Compounds						
41)	L9 AR-1268-1	7.622	7.590	71178847	31006348	50.721m	53.625
42)	L9 AR-1268-2	7.688	7.654	63525178	28112917	49.643	52.781
43)	L9 AR-1268-3	7.896	7.861	50172519	22018734	48.306	53.325
44)	L9 AR-1268-4	8.183	8.146	21758650	8301200	48.629m	51.234
45)	L9 AR-1268-5	8.481	8.438	151.8E6	51592382	47.022	52.264

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109999.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 22:15
Operator : YP/AJ
Sample : AR1268ICC050
Misc :
ALS Vial : 30 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
AR1268ICC050


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 03/19/2025
Supervised By :mohammad ahmed 03/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 02:07:11 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:04:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene

R.T.: 3.691 min
Delta R.T.: -0.001 min
Response: 44374914
Conc: 4.78 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC050


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/19/2025
Supervised By :mohammad ahmed 03/24/2025




#1 Tetrachloro-m-xylene

R.T.: 3.688 min
Delta R.T.: -0.001 min
Response: 25645349
Conc: 4.86 ng/ml m




#2 Decachlorobiphenyl

R.T.: 8.743 min
Delta R.T.: 0.000 min
Response: 72973489
Conc: 5.17 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.695 min
Delta R.T.: 0.001 min
Response: 22764198
Conc: 5.49 ng/ml




#41 AR-1268-1

R.T.: 7.622 min
Delta R.T.: 0.000 min
Response: 71178847
Conc: 50.72 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC050


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/19/2025
Supervised By :mohammad ahmed 03/24/2025




#41 AR-1268-1

R.T.: 7.590 min
Delta R.T.: 0.000 min
Response: 31006348
Conc: 53.63 ng/ml




#42 AR-1268-2

R.T.: 7.688 min
Delta R.T.: 0.000 min
Response: 63525178
Conc: 49.64 ng/ml



#42 AR-1268-2

R.T.: 7.654 min
Delta R.T.: 0.001 min
Response: 28112917
Conc: 52.78 ng/ml




#43 AR-1268-3

R.T.: 7.896 min
Delta R.T.: 0.000 min
Response: 50172519
Conc: 48.31 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC050


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/19/2025
Supervised By :mohammad ahmed 03/24/2025




#43 AR-1268-3

R.T.: 7.861 min
Delta R.T.: 0.000 min
Response: 22018734
Conc: 53.32 ng/ml




#44 AR-1268-4

R.T.: 8.183 min
Delta R.T.: -0.001 min
Response: 21758650
Conc: 48.63 ng/ml m



#44 AR-1268-4

R.T.: 8.146 min
Delta R.T.: 0.000 min
Response: 8301200
Conc: 51.23 ng/ml




#45 AR-1268-5

R.T.: 8.481 min
Delta R.T.: 0.000 min
Response: 151815895
Conc: 47.02 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/19/2025
Supervised By :mohammad ahmed 03/24/2025



#45 AR-1268-5

R.T.: 8.438 min
Delta R.T.: 0.000 min
Response: 51592382
Conc: 52.26 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO110000.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 22:34
Operator : YP/AJ
Sample : PO031925ICV500
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO031925

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 02:41:09 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:39:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.692	3.689	445.5E6	261.7E6	48.943	49.897
2)	SA Decachlor...	8.744	8.696	386.1E6	117.6E6	50.211	48.467
	Target Compounds						
3)	L1 AR-1016-1	4.784	4.771	165.3E6	89338991	491.286	487.512
4)	L1 AR-1016-2	4.804	4.790	229.3E6	128.0E6	494.517	489.369
5)	L1 AR-1016-3	4.860	4.966	158.0E6	69159098	492.753	487.796
6)	L1 AR-1016-4	4.981	5.008	124.5E6	57510192	492.750	487.665
7)	L1 AR-1016-5	5.238	5.221	133.3E6	74393726	484.896	486.492
31)	L7 AR-1260-1	6.280	6.253	235.0E6	125.1E6	498.624	489.686
32)	L7 AR-1260-2	6.469	6.441	287.0E6	145.2E6	481.955	483.545
33)	L7 AR-1260-3	6.837	6.594	243.8E6	137.4E6	502.299	475.363
34)	L7 AR-1260-4	7.097	7.065	212.4E6	103.8E6	500.751	482.426
35)	L7 AR-1260-5	7.339	7.306	531.0E6	241.3E6	507.683	483.788


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO110000.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 22:34
Operator : YP/AJ
Sample : PO031925ICV500
Misc :
ALS Vial : 31 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
ICVPO031925

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 02:41:09 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:39:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.692 min
Delta R.T.: 0.000 min
Response: 445455679
Conc: 48.94 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO031925




#1 Tetrachloro-m-xylene

R.T.: 3.689 min
Delta R.T.: 0.000 min
Response: 261663576
Conc: 49.90 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.744 min
Delta R.T.: -0.001 min
Response: 386058371
Conc: 50.21 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.696 min
Delta R.T.: 0.000 min
Response: 117585250
Conc: 48.47 ng/ml



#3 AR-1016-1


R.T.: 4.784 min
Delta R.T.: 0.000 min
Response: 165339415
Conc: 491.29 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO031925




#3 AR-1016-1

R.T.: 4.771 min
Delta R.T.: 0.000 min
Response: 89338991
Conc: 487.51 ng/ml




#4 AR-1016-2

R.T.: 4.804 min
Delta R.T.: 0.000 min
Response: 229300386
Conc: 494.52 ng/ml



#4 AR-1016-2


R.T.: 4.790 min
Delta R.T.: 0.000 min
Response: 127972100
Conc: 489.37 ng/ml



#5 AR-1016-3


R.T.: 4.860 min
Delta R.T.: 0.000 min
Response: 157969845
Conc: 492.75 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO031925




#5 AR-1016-3

R.T.: 4.966 min
Delta R.T.: 0.000 min
Response: 69159098
Conc: 487.80 ng/ml




#6 AR-1016-4

R.T.: 4.981 min
Delta R.T.: 0.000 min
Response: 124512743
Conc: 492.75 ng/ml



#6 AR-1016-4


R.T.: 5.008 min
Delta R.T.: 0.000 min
Response: 57510192
Conc: 487.66 ng/ml



#7 AR-1016-5


R.T.: 5.238 min
Delta R.T.: -0.001 min
Response: 133281979
Conc: 484.90 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO031925




#7 AR-1016-5

R.T.: 5.221 min
Delta R.T.: -0.001 min
Response: 74393726
Conc: 486.49 ng/ml




#31 AR-1260-1

R.T.: 6.280 min
Delta R.T.: 0.000 min
Response: 235026886
Conc: 498.62 ng/ml



#31 AR-1260-1


R.T.: 6.253 min
Delta R.T.: -0.002 min
Response: 125086324
Conc: 489.69 ng/ml



#32 AR-1260-2


R.T.: 6.469 min
Delta R.T.: 0.000 min
Response: 287013411
Conc: 481.95 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO031925




#32 AR-1260-2

R.T.: 6.441 min
Delta R.T.: -0.001 min
Response: 145201446
Conc: 483.54 ng/ml




#33 AR-1260-3

R.T.: 6.837 min
Delta R.T.: 0.000 min
Response: 243766592
Conc: 502.30 ng/ml



#33 AR-1260-3


R.T.: 6.594 min
Delta R.T.: -0.001 min
Response: 137371501
Conc: 475.36 ng/ml



#34 AR-1260-4


R.T.: 7.097 min
Delta R.T.: 0.000 min
Response: 212402639
Conc: 500.75 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO031925




#34 AR-1260-4

R.T.: 7.065 min
Delta R.T.: -0.001 min
Response: 103759489
Conc: 482.43 ng/ml



#35 AR-1260-5

R.T.: 7.339 min
Delta R.T.: 0.000 min
Response: 530958243
Conc: 507.68 ng/ml



#35 AR-1260-5

R.T.: 7.306 min
Delta R.T.: -0.001 min
Response: 241263068
Conc: 483.79 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO110001.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 23:47
Operator : YP/AJ
Sample : AR1242ICV500
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO031925AR1242

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 02:45:06 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:43:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.692	3.689	445.5E6	254.1E6	52.578	52.663
2)	SA Decachlor...	8.744	8.695	378.1E6	113.8E6	52.496	50.899
	Target Compounds						
16)	L4 AR-1242-1	4.785	4.771	138.5E6	75308322	517.896	515.050
17)	L4 AR-1242-2	4.804	4.791	190.3E6	107.0E6	516.964	514.473
18)	L4 AR-1242-3	4.861	4.966	132.7E6	58059678	514.485	520.188
19)	L4 AR-1242-4	4.981	5.050	104.1E6	58816332	520.236	507.140
20)	L4 AR-1242-5	5.634	5.572	109.1E6	71078329	516.760	517.905

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO110001.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 23:47
Operator : YP/AJ
Sample : AR1242ICV500
Misc :
ALS Vial : 32 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
ICVPO031925AR1242

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 02:45:06 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:43:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.692 min
Delta R.T.: 0.000 min
Response: 445457667
Conc: 52.58 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO031925AR1242




#1 Tetrachloro-m-xylene

R.T.: 3.689 min
Delta R.T.: 0.000 min
Response: 254125961
Conc: 52.66 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.744 min
Delta R.T.: 0.000 min
Response: 378060054
Conc: 52.50 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.695 min
Delta R.T.: 0.000 min
Response: 113786439
Conc: 50.90 ng/ml



#16 AR-1242-1


R.T.: 4.785 min
Delta R.T.: 0.000 min
Response: 138489149
Conc: 517.90 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO031925AR1242




#16 AR-1242-1

R.T.: 4.771 min
Delta R.T.: 0.000 min
Response: 75308322
Conc: 515.05 ng/ml




#17 AR-1242-2

R.T.: 4.804 min
Delta R.T.: 0.000 min
Response: 190347656
Conc: 516.96 ng/ml



#17 AR-1242-2


R.T.: 4.791 min
Delta R.T.: 0.000 min
Response: 107028573
Conc: 514.47 ng/ml



#18 AR-1242-3


R.T.: 4.861 min
Delta R.T.: 0.000 min
Response: 132711818
Conc: 514.48 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO031925AR1242




#18 AR-1242-3

R.T.: 4.966 min
Delta R.T.: -0.001 min
Response: 58059678
Conc: 520.19 ng/ml




#19 AR-1242-4

R.T.: 4.981 min
Delta R.T.: 0.000 min
Response: 104096142
Conc: 520.24 ng/ml



#19 AR-1242-4


R.T.: 5.050 min
Delta R.T.: -0.001 min
Response: 58816332
Conc: 507.14 ng/ml



#20 AR-1242-5

R.T.: 5.634 min
Delta R.T.: 0.000 min
Response: 109065446
Conc: 516.76 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO031925AR1242



#20 AR-1242-5

R.T.: 5.572 min
Delta R.T.: 0.000 min
Response: 71078329
Conc: 517.90 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO110002.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Mar 2025 00:05
Operator : YP/AJ
Sample : AR1248ICV500
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO031925AR1248

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 01:41:11 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 01:38:33 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.692	3.689	453.1E6	257.7E6	48.851	48.659
2)	SA Decachlor...	8.743	8.695	387.8E6	116.2E6	49.748	49.559
	Target Compounds						
21)	L5 AR-1248-1	4.784	4.771	108.4E6	58618795	492.939	487.181
22)	L5 AR-1248-2	5.023	5.008	147.5E6	82104463	489.895	487.213
23)	L5 AR-1248-3	5.238	5.050	184.5E6	87652547	492.701	489.391
24)	L5 AR-1248-4	5.592	5.221	258.0E6	102.0E6	493.069	489.250
25)	L5 AR-1248-5	5.634	5.613	182.7E6	99745645	494.045	489.737

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO110002.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Mar 2025 00:05
Operator : YP/AJ
Sample : AR1248ICV500
Misc :
ALS Vial : 33 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
ICVPO031925AR1248

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 01:41:11 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 01:38:33 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.692 min
Delta R.T.: 0.000 min
Response: 453137446
Conc: 48.85 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO031925AR1248




#1 Tetrachloro-m-xylene

R.T.: 3.689 min
Delta R.T.: 0.000 min
Response: 257654315
Conc: 48.66 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.743 min
Delta R.T.: -0.001 min
Response: 387771183
Conc: 49.75 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.695 min
Delta R.T.: 0.000 min
Response: 116199328
Conc: 49.56 ng/ml



#21 AR-1248-1


R.T.: 4.784 min
Delta R.T.: 0.000 min
Response: 108417858
Conc: 492.94 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO031925AR1248




#21 AR-1248-1

R.T.: 4.771 min
Delta R.T.: 0.000 min
Response: 58618795
Conc: 487.18 ng/ml




#22 AR-1248-2

R.T.: 5.023 min
Delta R.T.: 0.000 min
Response: 147491637
Conc: 489.90 ng/ml



#22 AR-1248-2


R.T.: 5.008 min
Delta R.T.: 0.000 min
Response: 82104463
Conc: 487.21 ng/ml



#23 AR-1248-3


R.T.: 5.238 min
Delta R.T.: 0.000 min
Response: 184471566
Conc: 492.70 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO031925AR1248




#23 AR-1248-3

R.T.: 5.050 min
Delta R.T.: 0.000 min
Response: 87652547
Conc: 489.39 ng/ml




#24 AR-1248-4

R.T.: 5.592 min
Delta R.T.: 0.000 min
Response: 258010469
Conc: 493.07 ng/ml



#24 AR-1248-4


R.T.: 5.221 min
Delta R.T.: 0.000 min
Response: 102041710
Conc: 489.25 ng/ml



#25 AR-1248-5

R.T.: 5.634 min
Delta R.T.: 0.000 min
Response: 182660838
Conc: 494.04 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO031925AR1248



#25 AR-1248-5

R.T.: 5.613 min
Delta R.T.: 0.000 min
Response: 99745645
Conc: 489.74 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO110003.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Mar 2025 00:24
Operator : YP/AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO031925AR1254

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 01:50:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 01:48:40 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.692	3.689	450.3E6	255.9E6	47.966	47.857
2)	SA Decachlor...	8.743	8.695	386.1E6	115.4E6	49.078	47.699
	Target Compounds						
26)	L6 AR-1254-1	5.593	5.572	270.9E6	148.9E6	484.350	483.813
27)	L6 AR-1254-2	5.743	5.720	234.3E6	129.7E6	484.326	483.086
28)	L6 AR-1254-3	6.148	6.123	381.8E6	203.6E6	483.781	483.489
29)	L6 AR-1254-4	6.377	6.350	236.7E6	117.7E6	486.513	481.783
30)	L6 AR-1254-5	6.797	6.768	332.2E6	169.1E6	488.082	485.104


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO110003.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Mar 2025 00:24
Operator : YP/AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 34 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
ICVPO031925AR1254

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 01:50:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 01:48:40 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.692 min
Delta R.T.: 0.000 min
Response: 450272597
Conc: 47.97 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO031925AR1254




#1 Tetrachloro-m-xylene

R.T.: 3.689 min
Delta R.T.: 0.000 min
Response: 255886625
Conc: 47.86 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.743 min
Delta R.T.: 0.000 min
Response: 386111719
Conc: 49.08 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.695 min
Delta R.T.: 0.000 min
Response: 115363596
Conc: 47.70 ng/ml



#26 AR-1254-1


R.T.: 5.593 min
Delta R.T.: 0.000 min
Response: 270904261
Conc: 484.35 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO031925AR1254




#26 AR-1254-1

R.T.: 5.572 min
Delta R.T.: 0.000 min
Response: 148880437
Conc: 483.81 ng/ml




#27 AR-1254-2

R.T.: 5.743 min
Delta R.T.: 0.000 min
Response: 234269646
Conc: 484.33 ng/ml



#27 AR-1254-2


R.T.: 5.720 min
Delta R.T.: 0.000 min
Response: 129704001
Conc: 483.09 ng/ml



#28 AR-1254-3


R.T.: 6.148 min
Delta R.T.: 0.000 min
Response: 381837309
Conc: 483.78 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO031925AR1254




#28 AR-1254-3

R.T.: 6.123 min
Delta R.T.: 0.000 min
Response: 203649298
Conc: 483.49 ng/ml




#29 AR-1254-4

R.T.: 6.377 min
Delta R.T.: 0.000 min
Response: 236662409
Conc: 486.51 ng/ml



#29 AR-1254-4


R.T.: 6.350 min
Delta R.T.: 0.000 min
Response: 117742050
Conc: 481.78 ng/ml



#30 AR-1254-5

R.T.: 6.797 min
Delta R.T.: 0.000 min
Response: 332152185
Conc: 488.08 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO031925AR1254



#30 AR-1254-5

R.T.: 6.768 min
Delta R.T.: 0.000 min
Response: 169138710
Conc: 485.10 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO110004.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Mar 2025 00:42
Operator : YP/AJ
Sample : AR1268ICV500
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO031925AR1268

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 02:07:25 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:04:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.692	3.690	450.4E6	255.9E6	48.546	48.447
2)	SA Decachlor...	8.744	8.696	690.0E6	201.4E6	48.863	48.609
	Target Compounds						
41)	L9 AR-1268-1	7.624	7.590	683.9E6	280.8E6	487.341	485.631
42)	L9 AR-1268-2	7.689	7.654	622.6E6	259.1E6	486.560	486.478
43)	L9 AR-1268-3	7.897	7.862	509.4E6	200.7E6	490.473	486.052
44)	L9 AR-1268-4	8.184	8.146	217.2E6	79710818	485.391	491.969
45)	L9 AR-1268-5	8.482	8.440	1574.6E6	478.9E6	487.695	485.135


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO110004.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Mar 2025 00:42
Operator : YP/AJ
Sample : AR1268ICV500
Misc :
ALS Vial : 35 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
ICVPO031925AR1268

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 02:07:25 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:04:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.692 min
Delta R.T.: 0.000 min
Response: 450382059
Conc: 48.55 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO031925AR1268




#1 Tetrachloro-m-xylene

R.T.: 3.690 min
Delta R.T.: 0.000 min
Response: 255894244
Conc: 48.45 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.744 min
Delta R.T.: 0.000 min
Response: 689957964
Conc: 48.86 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.696 min
Delta R.T.: 0.002 min
Response: 201448574
Conc: 48.61 ng/ml



#41 AR-1268-1


R.T.: 7.624 min
Delta R.T.: 0.002 min
Response: 683899311
Conc: 487.34 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO031925AR1268




#41 AR-1268-1

R.T.: 7.590 min
Delta R.T.: 0.001 min
Response: 280794579
Conc: 485.63 ng/ml




#42 AR-1268-2

R.T.: 7.689 min
Delta R.T.: 0.000 min
Response: 622617772
Conc: 486.56 ng/ml



#42 AR-1268-2


R.T.: 7.654 min
Delta R.T.: 0.001 min
Response: 259115661
Conc: 486.48 ng/ml



#43 AR-1268-3


R.T.: 7.897 min
Delta R.T.: 0.000 min
Response: 509428626
Conc: 490.47 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO031925AR1268




#43 AR-1268-3

R.T.: 7.862 min
Delta R.T.: 0.000 min
Response: 200699301
Conc: 486.05 ng/ml




#44 AR-1268-4

R.T.: 8.184 min
Delta R.T.: 0.000 min
Response: 217182692
Conc: 485.39 ng/ml



#44 AR-1268-4


R.T.: 8.146 min
Delta R.T.: 0.000 min
Response: 79710818
Conc: 491.97 ng/ml



#45 AR-1268-5

R.T.: 8.482 min
Delta R.T.: 0.000 min
Response: 1574576275
Conc: 487.70 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO031925AR1268



#45 AR-1268-5

R.T.: 8.440 min
Delta R.T.: 0.000 min
Response: 478896315
Conc: 485.13 ng/ml

Raw Data

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

RETENTION TIMES OF INITIAL CALIBRATION

Contract: NOBI03

Lab Code: CHEM **Case No.:** Q1730 **SAS No.:** Q1730 **SDG NO.:** Q1730

Instrument ID: ECD_P **Calibration Date(s):** 03/27/2025 03/27/2025

Calibration Times: 10:30 17:53

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID: RT 1000 = PP070917.D RT 750 = PP070918.D

RT 500 = PP070919.D RT 250 = PP070920.D RT 050 = PP070921.D

COMPOUND		RT 1000	RT 750	RT 500	RT 250	RT 050	MEAN RT	RT WINDOW FROM	RT WINDOW TO
Aroclor-1016-1	(1)	5.67	5.67	5.66	5.67	5.67	5.67	5.57	5.77
Aroclor-1016-2	(2)	5.69	5.69	5.69	5.69	5.69	5.69	5.59	5.79
Aroclor-1016-3	(3)	5.75	5.75	5.75	5.75	5.75	5.75	5.65	5.85
Aroclor-1016-4	(4)	5.85	5.85	5.85	5.85	5.85	5.85	5.75	5.95
Aroclor-1016-5	(5)	6.14	6.14	6.14	6.14	6.14	6.14	6.04	6.24
Aroclor-1260-1	(1)	7.26	7.26	7.26	7.26	7.26	7.26	7.16	7.36
Aroclor-1260-2	(2)	7.51	7.51	7.51	7.51	7.51	7.51	7.41	7.61
Aroclor-1260-3	(3)	7.87	7.87	7.87	7.87	7.87	7.87	7.77	7.97
Aroclor-1260-4	(4)	8.10	8.10	8.09	8.09	8.09	8.09	7.99	8.19
Aroclor-1260-5	(5)	8.41	8.42	8.41	8.41	8.41	8.41	8.31	8.51
Decachlorobiphenyl		10.23	10.23	10.23	10.23	10.23	10.23	10.13	10.33
Tetrachloro-m-xylene		4.51	4.52	4.51	4.52	4.51	4.51	4.41	4.61
Aroclor-1242-1	(1)	5.67	5.67	5.66	5.67	5.66	5.67	5.57	5.77
Aroclor-1242-2	(2)	5.69	5.69	5.69	5.69	5.69	5.69	5.59	5.79
Aroclor-1242-3	(3)	5.75	5.75	5.75	5.75	5.75	5.75	5.65	5.85
Aroclor-1242-4	(4)	5.85	5.85	5.85	5.85	5.84	5.85	5.75	5.95
Aroclor-1242-5	(5)	6.58	6.58	6.58	6.58	6.57	6.58	6.48	6.68
Decachlorobiphenyl		10.23	10.23	10.23	10.23	10.23	10.23	10.13	10.33
Tetrachloro-m-xylene		4.51	4.51	4.51	4.51	4.51	4.51	4.41	4.61
Aroclor-1248-1	(1)	5.67	5.67	5.67	5.67	5.67	5.67	5.57	5.77
Aroclor-1248-2	(2)	5.94	5.94	5.94	5.94	5.94	5.94	5.84	6.04
Aroclor-1248-3	(3)	6.14	6.14	6.14	6.14	6.14	6.14	6.04	6.24
Aroclor-1248-4	(4)	6.54	6.54	6.54	6.54	6.54	6.54	6.44	6.64
Aroclor-1248-5	(5)	6.58	6.58	6.58	6.58	6.58	6.58	6.48	6.68
Decachlorobiphenyl		10.23	10.23	10.23	10.23	10.23	10.23	10.13	10.33
Tetrachloro-m-xylene		4.51	4.51	4.51	4.52	4.51	4.51	4.41	4.61
Aroclor-1254-1	(1)	6.52	6.52	6.52	6.52	6.52	6.52	6.42	6.62
Aroclor-1254-2	(2)	6.73	6.73	6.73	6.73	6.73	6.73	6.63	6.83
Aroclor-1254-3	(3)	7.10	7.09	7.10	7.10	7.10	7.10	7.00	7.20
Aroclor-1254-4	(4)	7.38	7.38	7.38	7.38	7.38	7.38	7.28	7.48
Aroclor-1254-5	(5)	7.79	7.79	7.79	7.79	7.79	7.79	7.69	7.89
Decachlorobiphenyl		10.23	10.23	10.23	10.23	10.23	10.23	10.13	10.33
Tetrachloro-m-xylene		4.51	4.51	4.51	4.51	4.52	4.51	4.41	4.61
Aroclor-1268-1	(1)	8.73	8.73	8.73	8.73	8.73	8.73	8.63	8.83
Aroclor-1268-2	(2)	8.82	8.82	8.82	8.82	8.82	8.82	8.72	8.92
Aroclor-1268-3	(3)	9.06	9.06	9.05	9.06	9.05	9.06	8.96	9.16
Aroclor-1268-4	(4)	9.47	9.47	9.47	9.47	9.47	9.47	9.37	9.57
Aroclor-1268-5	(5)	9.89	9.89	9.89	9.89	9.89	9.89	9.79	9.99

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

RETENTION TIMES OF INITIAL CALIBRATION

Decachlorobiphenyl	10.23	10.23	10.23	10.23	10.23	10.23	10.13	10.33
Tetrachloro-m-xylene	4.52	4.51	4.51	4.52	4.51	4.51	4.41	4.61

Raw Data

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

RETENTION TIMES OF INITIAL CALIBRATION

Contract: NOBI03

Lab Code: CHEM **Case No.:** Q1730 **SAS No.:** Q1730 **SDG NO.:** Q1730

Instrument ID: ECD_P **Calibration Date(s):** 03/27/2025 03/27/2025

Calibration Times: 10:30 17:53

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID: RT 1000 = PP070917.D RT 750 = PP070918.D

RT 500 = PP070919.D RT 250 = PP070920.D RT 050 = PP070921.D

COMPOUND		RT 1000	RT 750	RT 500	RT 250	RT 050	MEAN RT	RT WINDOW FROM	RT WINDOW TO
Aroclor-1016-1	(1)	4.90	4.90	4.90	4.90	4.90	4.90	4.80	5.00
Aroclor-1016-2	(2)	4.92	4.92	4.92	4.92	4.92	4.92	4.82	5.02
Aroclor-1016-3	(3)	5.10	5.10	5.10	5.10	5.10	5.10	5.00	5.20
Aroclor-1016-4	(4)	5.14	5.14	5.14	5.14	5.14	5.14	5.04	5.24
Aroclor-1016-5	(5)	5.36	5.36	5.36	5.36	5.36	5.36	5.26	5.46
Aroclor-1260-1	(1)	6.39	6.39	6.39	6.39	6.39	6.39	6.29	6.49
Aroclor-1260-2	(2)	6.58	6.58	6.58	6.58	6.58	6.58	6.48	6.68
Aroclor-1260-3	(3)	6.73	6.73	6.73	6.73	6.73	6.73	6.63	6.83
Aroclor-1260-4	(4)	7.20	7.20	7.21	7.21	7.21	7.20	7.10	7.30
Aroclor-1260-5	(5)	7.45	7.44	7.45	7.45	7.45	7.45	7.35	7.55
Decachlorobiphenyl		8.86	8.86	8.86	8.86	8.86	8.86	8.76	8.96
Tetrachloro-m-xylene		3.82	3.82	3.82	3.82	3.82	3.82	3.72	3.92
Aroclor-1242-1	(1)	4.90	4.90	4.90	4.90	4.90	4.90	4.80	5.00
Aroclor-1242-2	(2)	4.92	4.92	4.92	4.92	4.92	4.92	4.82	5.02
Aroclor-1242-3	(3)	5.10	5.10	5.10	5.10	5.10	5.10	5.00	5.20
Aroclor-1242-4	(4)	5.18	5.18	5.18	5.18	5.18	5.18	5.08	5.28
Aroclor-1242-5	(5)	5.71	5.71	5.71	5.71	5.71	5.71	5.61	5.81
Decachlorobiphenyl		8.86	8.86	8.86	8.86	8.86	8.86	8.76	8.96
Tetrachloro-m-xylene		3.82	3.82	3.82	3.82	3.82	3.82	3.72	3.92
Aroclor-1248-1	(1)	4.90	4.90	4.90	4.90	4.90	4.90	4.80	5.00
Aroclor-1248-2	(2)	5.14	5.14	5.14	5.14	5.14	5.14	5.04	5.24
Aroclor-1248-3	(3)	5.18	5.18	5.18	5.18	5.18	5.18	5.08	5.28
Aroclor-1248-4	(4)	5.36	5.36	5.36	5.36	5.35	5.36	5.26	5.46
Aroclor-1248-5	(5)	5.75	5.75	5.75	5.75	5.75	5.75	5.65	5.85
Decachlorobiphenyl		8.86	8.86	8.86	8.86	8.86	8.86	8.76	8.96
Tetrachloro-m-xylene		3.82	3.82	3.82	3.82	3.82	3.82	3.72	3.92
Aroclor-1254-1	(1)	5.71	5.71	5.71	5.71	5.71	5.71	5.61	5.81
Aroclor-1254-2	(2)	5.86	5.86	5.86	5.86	5.86	5.86	5.76	5.96
Aroclor-1254-3	(3)	6.26	6.26	6.26	6.26	6.26	6.26	6.16	6.36
Aroclor-1254-4	(4)	6.49	6.49	6.49	6.49	6.49	6.49	6.39	6.59
Aroclor-1254-5	(5)	6.91	6.91	6.91	6.91	6.91	6.91	6.81	7.01
Decachlorobiphenyl		8.86	8.86	8.86	8.86	8.86	8.86	8.76	8.96
Tetrachloro-m-xylene		3.82	3.82	3.82	3.82	3.82	3.82	3.72	3.92
Aroclor-1268-1	(1)	7.73	7.73	7.73	7.73	7.73	7.73	7.63	7.83
Aroclor-1268-2	(2)	7.80	7.80	7.79	7.79	7.79	7.79	7.69	7.89
Aroclor-1268-3	(3)	8.00	8.00	8.00	8.00	8.00	8.00	7.90	8.10
Aroclor-1268-4	(4)	8.30	8.29	8.29	8.30	8.29	8.29	8.19	8.39
Aroclor-1268-5	(5)	8.60	8.60	8.60	8.60	8.60	8.60	8.50	8.70

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

RETENTION TIMES OF INITIAL CALIBRATION

Decachlorobiphenyl	8.86	8.86	8.86	8.86	8.86	8.86	8.76	8.96
Tetrachloro-m-xylene	3.82	3.82	3.82	3.82	3.82	3.82	3.72	3.92

Raw Data

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: NOBI03

Lab Code: CHEM **Case No.:** Q1730 **SAS No.:** Q1730 **SDG NO.:** Q1730

Instrument ID: ECD_P

Calibration Date(s): 03/27/2025 03/27/2025

Calibration Times: 10:30 17:53

GC Column: ZB-MR1 **ID:** 0.32 **(mm)**

LAB FILE ID: **CF 1000 =** PP070917.D **CF 750 =** PP070918.D

CF 500 = PP070919.D **CF 250 =** PP070920.D **CF 050 =** PP070921.D

COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	CF	% RSD
Aroclor-1016-1	(1)	47478014	49171687	50934060	54324668	52347880	50851262	5
Aroclor-1016-2	(2)	72483549	74568483	78811172	80433884	83971080	78053634	6
Aroclor-1016-3	(3)	43541298	44043953	46815864	48849960	52621940	47174603	7
Aroclor-1016-4	(4)	36661280	36602855	39014928	39212148	32287440	36755730	7
Aroclor-1016-5	(5)	32204227	33487635	35040682	36610940	32217200	33912137	5
Aroclor-1260-1	(1)	60625379	62238407	66823960	72729028	73347280	67152811	8
Aroclor-1260-2	(2)	90805769	95497569	97808710	102858984	93710780	96136362	4
Aroclor-1260-3	(3)	75946924	78957439	82970652	86114216	75601760	79918198	5
Aroclor-1260-4	(4)	67190656	75020355	81451428	78145636	91876440	78736903	11
Aroclor-1260-5	(5)	154592706	156108101	159960800	166857288	156911780	158886135	3
Decachlorobiphenyl		1019047430	1065730867	1116449000	1115108760	946747800	1052616771	7
Tetrachloro-m-xylene		1424837400	1461405720	1519248320	1595716440	1306667000	1461574976	7
Aroclor-1242-1	(1)	40990074	42746281	43432670	43984636	48732080	43977148	6
Aroclor-1242-2	(2)	64704598	63862937	67024876	68850388	57138840	64316328	7
Aroclor-1242-3	(3)	39349869	38048504	39702972	40627540	47468380	41039453	9
Aroclor-1242-4	(4)	30814559	31466187	32277086	32194644	25864220	30523339	8
Aroclor-1242-5	(5)	36006298	34981232	37334376	38560396	35669740	36510408	4
Decachlorobiphenyl		1036298130	1063489027	1106293120	1153452680	1056566000	1083219791	4
Tetrachloro-m-xylene		1443947330	1507747640	1537397120	1567165920	1326698400	1476591282	6
Aroclor-1248-1	(1)	31141042	33379292	33781538	45988760	37146900	36287506	15
Aroclor-1248-2	(2)	42399779	44979081	46192256	47122176	43999640	44938586	4
Aroclor-1248-3	(3)	48055172	50777464	51825052	50697180	51832600	50637494	3
Aroclor-1248-4	(4)	59878410	63471275	67039892	73429036	56919340	64147591	9
Aroclor-1248-5	(5)	56652876	60504408	64286140	68519384	54268440	60846250	9
Decachlorobiphenyl		1072017300	1124462173	1205136680	1167091680	1109917800	1135725127	5
Tetrachloro-m-xylene		1462181050	1544091280	1549039360	1559485200	1464932600	1515945898	3
Aroclor-1254-1	(1)	60157506	63067431	65471886	66806416	66447920	64390232	4
Aroclor-1254-2	(2)	92838442	97047215	101802392	105556104	108208740	101090579	6
Aroclor-1254-3	(3)	92334650	98055471	99416554	103841140	104991280	99727819	5
Aroclor-1254-4	(4)	84238161	87780547	86802510	88788836	85373700	86596751	2
Aroclor-1254-5	(5)	74433164	79038301	80854650	83958584	83470140	80350968	5
Decachlorobiphenyl		1095432410	1144692320	1168417220	1194330360	1035516600	1127677782	6
Tetrachloro-m-xylene		1501288360	1551275560	1575746420	1626904120	1301718600	1511386612	8
Aroclor-1268-1	(1)	226575551	237927453	239660032	247892192	225482800	235507606	4

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	190380276	198934901	200232346	204948924	182079900	195315269	4
Aroclor-1268-3	(3)	163648333	171737784	173955356	177246576	167979860	170913582	3
Aroclor-1268-4	(4)	70377008	73665913	73237284	74802952	61652780	70747187	7
Aroclor-1268-5	(5)	459553484	481598519	482514296	498456744	458756600	476175929	3
Decachlorobiphenyl		1956984350	2055290840	2095781900	2171413960	2007035400	2057301290	4
Tetrachloro-m-xylene		1523485290	1594639080	1607574540	1645012520	1440121400	1562166566	5

Raw Data

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: NOBI03

Lab Code: CHEM **Case No.:** Q1730 **SAS No.:** Q1730 **SDG NO.:** Q1730

Instrument ID: ECD_P **Calibration Date(s):** 03/27/2025 03/27/2025

Calibration Times: 10:30 17:53

GC Column: ZB-MR2 **ID:** 0.32 (mm)

LAB FILE ID: **CF 1000 =** PP070917.D **CF 750 =** PP070918.D

CF 500 = PP070919.D **CF 250 =** PP070920.D **CF 050 =** PP070921.D

COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	CF	% RSD
Aroclor-1016-1	(1)	34208953	37661787	37955886	45294856	45319420	40088180	12
Aroclor-1016-2	(2)	49610574	53393839	54380778	65503040	66437080	57865062	12
Aroclor-1016-3	(3)	27000462	29732755	29849058	35508136	32526120	30923306	10
Aroclor-1016-4	(4)	21229955	23465415	23793788	28635320	24655080	24355912	10
Aroclor-1016-5	(5)	26939544	29893109	30154374	37292184	37326660	32321174	14
Aroclor-1260-1	(1)	45249076	46519276	48240522	56226688	59796020	51206316	12
Aroclor-1260-2	(2)	58475471	61260192	62904038	69183788	67145740	63793846	6
Aroclor-1260-3	(3)	50913622	52312629	53920638	58739540	56547700	54486826	5
Aroclor-1260-4	(4)	42366044	44395528	47488628	49955128	54926140	47826294	10
Aroclor-1260-5	(5)	113066934	116639771	125312810	127263128	140268800	124510289	8
Decachlorobiphenyl		793547960	846236520	865016160	753569280	682611000	788196184	9
Tetrachloro-m-xylene		926488500	1023245573	1036748080	1124812280	1071924800	1036643847	7
Aroclor-1242-1	(1)	30391529	31772724	34649408	38698720	35466820	34195840	9
Aroclor-1242-2	(2)	43879042	45075739	49434800	55503780	52395240	49257720	9
Aroclor-1242-3	(3)	24103124	24630729	26618958	29833472	28035100	26644277	8
Aroclor-1242-4	(4)	23086538	23159225	25581538	28618736	31445140	26378235	13
Aroclor-1242-5	(5)	28647762	29235153	31558738	32962404	32387440	30958299	6
Decachlorobiphenyl		794686010	779465320	775544600	851493360	677466800	775731218	8
Tetrachloro-m-xylene		994474400	1020857307	1119824860	1120535680	1106593200	1072457089	6
Aroclor-1248-1	(1)	23297586	26737011	28501250	29207540	27656520	27079981	8
Aroclor-1248-2	(2)	31071420	35918237	38274460	38702620	39645560	36722459	9
Aroclor-1248-3	(3)	32592570	37701093	39964120	40477712	42561380	38659375	9
Aroclor-1248-4	(4)	38114246	43411732	47181776	48244244	48753000	45141000	9
Aroclor-1248-5	(5)	37031094	41115012	43925318	46634268	44688040	42678746	8
Decachlorobiphenyl		763432400	822192813	827097300	806313080	732328400	790272799	5
Tetrachloro-m-xylene		969522160	1045815880	1138159760	1084071480	1169729800	1081459816	7
Aroclor-1254-1	(1)	58928147	61799583	66082486	73787844	69490960	66017804	8
Aroclor-1254-2	(2)	50145967	53786207	56793554	64034708	60730300	57098147	9
Aroclor-1254-3	(3)	77448562	83642240	87764778	99090376	86260880	86841367	9
Aroclor-1254-4	(4)	49998215	54614948	55966900	64174348	55747980	56100478	9
Aroclor-1254-5	(5)	65979008	71358709	71825896	82425644	69163880	72150627	8
Decachlorobiphenyl		783483810	796181933	910832540	850283840	759277600	820011945	7
Tetrachloro-m-xylene		1042595600	1091536387	1108578240	1195164040	1067355200	1101045893	5
Aroclor-1268-1	(1)	165356134	175541800	174753666	176934688	164423820	171402022	3

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	137801505	147739099	145758868	146778256	138336660	143282878	3
Aroclor-1268-3	(3)	120132746	127507232	124391202	127247744	116467620	123149309	4
Aroclor-1268-4	(4)	51096977	54505272	53344994	55689972	49428660	52813175	5
Aroclor-1268-5	(5)	357060145	376226307	370949858	366544744	338563340	361868879	4
Decachlorobiphenyl		1484820890	1579158933	1563503220	1620729960	1432317200	1536106041	5
Tetrachloro-m-xylene		1051289840	1118865693	1118283520	1183542080	1260888000	1146573827	7

Raw Data

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: NOBI03

Lab Code: CHEM **Case No.:** Q1730 **SAS No.:** Q1730 **SDG NO.:** Q1730

Instrument ID: ECD_P **Date(s) Analyzed:** 03/27/2025 03/27/2025

GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW FROM	RT WINDOW TO	CALIBRATION FACTOR
Aroclor-1221	500	1	4.72	4.62	4.82	17353700
		2	4.80	4.70	4.90	12952900
		3	4.88	4.78	4.98	41717200
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.88	4.78	4.98	33288800
		2	5.41	5.31	5.51	15910300
		3	5.69	5.59	5.79	34270400
		4	5.85	5.75	5.95	15707600
		5	5.94	5.84	6.04	10640400
Aroclor-1262	500	1	8.09	7.99	8.19	106300000
		2	8.41	8.31	8.51	198383000
		3	8.73	8.63	8.83	131742000
		4	8.82	8.72	8.92	95614800
		5	9.47	9.37	9.57	65066400

Raw Data

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: NOBI03

Lab Code: CHEM **Case No.:** Q1730 **SAS No.:** Q1730 **SDG NO.:** Q1730

Instrument ID: ECD_P **Date(s) Analyzed:** 03/27/2025 03/27/2025

GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW FROM	RT WINDOW TO	CALIBRATION FACTOR
Aroclor-1221	500	1	4.03	3.93	4.13	15243600
		2	4.11	4.01	4.21	11436300
		3	4.19	4.09	4.29	34089000
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.19	4.09	4.29	27986800
		2	4.92	4.82	5.02	28061200
		3	5.10	5.00	5.20	14889600
		4	5.19	5.09	5.29	12804400
		5	5.36	5.26	5.46	14174800
Aroclor-1262	500	1	6.95	6.85	7.05	85563600
		2	7.21	7.11	7.31	68464400
		3	7.73	7.63	7.83	60723600
		4	7.79	7.69	7.89	99406800
		5	8.29	8.19	8.39	46871600

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070917.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 10:30
Operator : YP\AJ
Sample : AR1660ICC1000
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 11:43:00 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 11:40:16 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.513	3.815	142.5E6	92648850	93.786	89.365
2)	SA Decachlor...	10.227	8.858	101.9E6	79354796	91.276	91.738
	Target Compounds						
3)	L1 AR-1016-1	5.665	4.903	47478014	34208953	932.147	901.282
4)	L1 AR-1016-2	5.687	4.921	72483549	49610574	919.712	912.281
5)	L1 AR-1016-3	5.749	5.098	43541298	27000462	930.054	904.567
6)	L1 AR-1016-4	5.846	5.140	36661280	21229955	939.673	892.248
7)	L1 AR-1016-5	6.140	5.355	32204227	26939544	919.053	893.388
31)	L7 AR-1260-1	7.258	6.391	60625379	45249076	907.240	937.989
32)	L7 AR-1260-2	7.512	6.579	90805769	58475471	928.402	929.598
33)	L7 AR-1260-3	7.870	6.732	75946924	50913622	915.347	944.233
34)	L7 AR-1260-4	8.095	7.204	67190656	42366044	824.917	892.130
35)	L7 AR-1260-5	8.413	7.445	154.6E6	113.1E6	966.441	902.278

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070917.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 10:30
Operator : YP\AJ
Sample : AR1660ICC1000
Misc :
ALS Vial : 3 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 11:43:00 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 11:40:16 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.513 min
Delta R.T.: 0.001 min
Response: 142483740
Conc: 93.79 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000




#1 Tetrachloro-m-xylene

R.T.: 3.815 min
Delta R.T.: 0.000 min
Response: 92648850
Conc: 89.36 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.227 min
Delta R.T.: 0.002 min
Response: 101904743
Conc: 91.28 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.858 min
Delta R.T.: 0.000 min
Response: 79354796
Conc: 91.74 ng/ml



#3 AR-1016-1


R.T.: 5.665 min
Delta R.T.: 0.002 min
Response: 47478014
Conc: 932.15 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000




#3 AR-1016-1

R.T.: 4.903 min
Delta R.T.: 0.000 min
Response: 34208953
Conc: 901.28 ng/ml




#4 AR-1016-2

R.T.: 5.687 min
Delta R.T.: 0.002 min
Response: 72483549
Conc: 919.71 ng/ml



#4 AR-1016-2


R.T.: 4.921 min
Delta R.T.: 0.000 min
Response: 49610574
Conc: 912.28 ng/ml



#5 AR-1016-3


R.T.: 5.749 min
Delta R.T.: 0.002 min
Response: 43541298
Conc: 930.05 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000




#5 AR-1016-3

R.T.: 5.098 min
Delta R.T.: 0.000 min
Response: 27000462
Conc: 904.57 ng/ml




#6 AR-1016-4

R.T.: 5.846 min
Delta R.T.: 0.001 min
Response: 36661280
Conc: 939.67 ng/ml



#6 AR-1016-4


R.T.: 5.140 min
Delta R.T.: 0.000 min
Response: 21229955
Conc: 892.25 ng/ml



#7 AR-1016-5


R.T.: 6.140 min
Delta R.T.: 0.002 min
Response: 32204227
Conc: 919.05 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000




#7 AR-1016-5

R.T.: 5.355 min
Delta R.T.: 0.000 min
Response: 26939544
Conc: 893.39 ng/ml




#31 AR-1260-1

R.T.: 7.258 min
Delta R.T.: 0.002 min
Response: 60625379
Conc: 907.24 ng/ml



#31 AR-1260-1


R.T.: 6.391 min
Delta R.T.: 0.000 min
Response: 45249076
Conc: 937.99 ng/ml



#32 AR-1260-2


R.T.: 7.512 min
Delta R.T.: 0.002 min
Response: 90805769
Conc: 928.40 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000




#32 AR-1260-2

R.T.: 6.579 min
Delta R.T.: 0.000 min
Response: 58475471
Conc: 929.60 ng/ml




#33 AR-1260-3

R.T.: 7.870 min
Delta R.T.: 0.002 min
Response: 75946924
Conc: 915.35 ng/ml



#33 AR-1260-3


R.T.: 6.732 min
Delta R.T.: -0.001 min
Response: 50913622
Conc: 944.23 ng/ml



#34 AR-1260-4


R.T.: 8.095 min
Delta R.T.: 0.003 min
Response: 67190656
Conc: 824.92 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000




#34 AR-1260-4

R.T.: 7.204 min
Delta R.T.: 0.000 min
Response: 42366044
Conc: 892.13 ng/ml



#35 AR-1260-5

R.T.: 8.413 min
Delta R.T.: 0.002 min
Response: 154592706
Conc: 966.44 ng/ml



#35 AR-1260-5

R.T.: 7.445 min
Delta R.T.: 0.000 min
Response: 113066934
Conc: 902.28 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070918.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 10:47
Operator : YP\AJ
Sample : AR1660ICC750
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 11:43:22 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 11:40:16 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1) SA	Tetrachlo...	4.516	3.816	109.6E6	76743418	72.145	74.023
2) SA	Decachlor...	10.232	8.858	79929815	63467739	71.593	73.372
	Target Compounds						
3) L1	AR-1016-1	5.668	4.903	36878765	28246340	724.049	744.189
4) L1	AR-1016-2	5.690	4.922	55926362	40045379	709.625	736.388
5) L1	AR-1016-3	5.752	5.098	33032965	22299566	705.593	747.078
6) L1	AR-1016-4	5.850	5.140	27452141	17599061	703.632	739.649
7) L1	AR-1016-5	6.142	5.355	25115726	22419832	716.759	743.502
31) L7	AR-1260-1	7.261	6.391	46678805	34889457	698.534	723.240
32) L7	AR-1260-2	7.514	6.578	71623177	45945144	732.278	730.401
33) L7	AR-1260-3	7.873	6.732	59218079	39234472	713.723	727.634
34) L7	AR-1260-4	8.097	7.203	56265266	33296646	690.783	701.150
35) L7	AR-1260-5	8.417	7.444	117.1E6	87479828	731.936	698.092

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070918.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 10:47
Operator : YP\AJ
Sample : AR1660ICC750
Misc :
ALS Vial : 4 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1660ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 11:43:22 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 11:40:16 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.516 min
Delta R.T.: 0.004 min
Response: 109605429
Conc: 72.14 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750




#1 Tetrachloro-m-xylene

R.T.: 3.816 min
Delta R.T.: 0.000 min
Response: 76743418
Conc: 74.02 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.232 min
Delta R.T.: 0.007 min
Response: 79929815
Conc: 71.59 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.858 min
Delta R.T.: -0.001 min
Response: 63467739
Conc: 73.37 ng/ml



#3 AR-1016-1


R.T.: 5.668 min
Delta R.T.: 0.005 min
Response: 36878765
Conc: 724.05 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750




#3 AR-1016-1

R.T.: 4.903 min
Delta R.T.: 0.000 min
Response: 28246340
Conc: 744.19 ng/ml




#4 AR-1016-2

R.T.: 5.690 min
Delta R.T.: 0.005 min
Response: 55926362
Conc: 709.62 ng/ml



#4 AR-1016-2


R.T.: 4.922 min
Delta R.T.: 0.000 min
Response: 40045379
Conc: 736.39 ng/ml



#5 AR-1016-3


R.T.: 5.752 min
Delta R.T.: 0.005 min
Response: 33032965
Conc: 705.59 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750




#5 AR-1016-3

R.T.: 5.098 min
Delta R.T.: 0.000 min
Response: 22299566
Conc: 747.08 ng/ml




#6 AR-1016-4

R.T.: 5.850 min
Delta R.T.: 0.005 min
Response: 27452141
Conc: 703.63 ng/ml



#6 AR-1016-4


R.T.: 5.140 min
Delta R.T.: 0.000 min
Response: 17599061
Conc: 739.65 ng/ml



#7 AR-1016-5


R.T.: 6.142 min
Delta R.T.: 0.004 min
Response: 25115726
Conc: 716.76 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750




#7 AR-1016-5

R.T.: 5.355 min
Delta R.T.: 0.000 min
Response: 22419832
Conc: 743.50 ng/ml




#31 AR-1260-1

R.T.: 7.261 min
Delta R.T.: 0.005 min
Response: 46678805
Conc: 698.53 ng/ml



#31 AR-1260-1


R.T.: 6.391 min
Delta R.T.: 0.000 min
Response: 34889457
Conc: 723.24 ng/ml



#32 AR-1260-2


R.T.: 7.514 min
Delta R.T.: 0.005 min
Response: 71623177
Conc: 732.28 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750




#32 AR-1260-2

R.T.: 6.578 min
Delta R.T.: -0.001 min
Response: 45945144
Conc: 730.40 ng/ml




#33 AR-1260-3

R.T.: 7.873 min
Delta R.T.: 0.005 min
Response: 59218079
Conc: 713.72 ng/ml



#33 AR-1260-3


R.T.: 6.732 min
Delta R.T.: -0.001 min
Response: 39234472
Conc: 727.63 ng/ml



#34 AR-1260-4


R.T.: 8.097 min
Delta R.T.: 0.005 min
Response: 56265266
Conc: 690.78 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750




#34 AR-1260-4

R.T.: 7.203 min
Delta R.T.: -0.001 min
Response: 33296646
Conc: 701.15 ng/ml



#35 AR-1260-5

R.T.: 8.417 min
Delta R.T.: 0.005 min
Response: 117081076
Conc: 731.94 ng/ml



#35 AR-1260-5

R.T.: 7.444 min
Delta R.T.: -0.002 min
Response: 87479828
Conc: 698.09 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070919.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 11:03
Operator : YP\AJ
Sample : AR1660ICC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 11:43:35 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 11:40:16 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1) SA	Tetrachlo...	4.512	3.816	75962416	51837404	50.000	50.000
2) SA	Decachlor...	10.225	8.859	55822450	43250808	50.000	50.000
	Target Compounds						
3) L1	AR-1016-1	5.663	4.903	25467030	18977943	500.000	500.000
4) L1	AR-1016-2	5.685	4.922	39405586	27190389	500.000	500.000
5) L1	AR-1016-3	5.748	5.099	23407932	14924529	500.000	500.000
6) L1	AR-1016-4	5.845	5.141	19507464	11896894	500.000	500.000
7) L1	AR-1016-5	6.138	5.355	17520341	15077187	500.000	500.000
31) L7	AR-1260-1	7.256	6.392	33411980	24120261	500.000	500.000
32) L7	AR-1260-2	7.510	6.580	48904355	31452019	500.000	500.000
33) L7	AR-1260-3	7.868	6.734	41485326	26960319	500.000	500.000
34) L7	AR-1260-4	8.092	7.205	40725714	23744314	500.000	500.000
35) L7	AR-1260-5	8.411	7.446	79980400	62656405	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070919.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 11:03
Operator : YP\AJ
Sample : AR1660ICC500
Misc :
ALS Vial : 5 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1660ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 11:43:35 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 11:40:16 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.512 min
Delta R.T.: 0.000 min
Response: 75962416
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.816 min
Delta R.T.: 0.000 min
Response: 51837404
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.225 min
Delta R.T.: 0.000 min
Response: 55822450
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.859 min
Delta R.T.: 0.000 min
Response: 43250808
Conc: 50.00 ng/ml



#3 AR-1016-1


R.T.: 5.663 min
Delta R.T.: 0.000 min
Response: 25467030
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500




#3 AR-1016-1

R.T.: 4.903 min
Delta R.T.: 0.000 min
Response: 18977943
Conc: 500.00 ng/ml




#4 AR-1016-2

R.T.: 5.685 min
Delta R.T.: 0.000 min
Response: 39405586
Conc: 500.00 ng/ml



#4 AR-1016-2


R.T.: 4.922 min
Delta R.T.: 0.000 min
Response: 27190389
Conc: 500.00 ng/ml



#5 AR-1016-3


R.T.: 5.748 min
Delta R.T.: 0.000 min
Response: 23407932
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500




#5 AR-1016-3

R.T.: 5.099 min
Delta R.T.: 0.000 min
Response: 14924529
Conc: 500.00 ng/ml




#6 AR-1016-4

R.T.: 5.845 min
Delta R.T.: 0.000 min
Response: 19507464
Conc: 500.00 ng/ml



#6 AR-1016-4


R.T.: 5.141 min
Delta R.T.: 0.000 min
Response: 11896894
Conc: 500.00 ng/ml



#7 AR-1016-5


R.T.: 6.138 min
Delta R.T.: 0.000 min
Response: 17520341
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500




#7 AR-1016-5

R.T.: 5.355 min
Delta R.T.: 0.000 min
Response: 15077187
Conc: 500.00 ng/ml




#31 AR-1260-1

R.T.: 7.256 min
Delta R.T.: 0.000 min
Response: 33411980
Conc: 500.00 ng/ml



#31 AR-1260-1


R.T.: 6.392 min
Delta R.T.: 0.000 min
Response: 24120261
Conc: 500.00 ng/ml



#32 AR-1260-2


R.T.: 7.510 min
Delta R.T.: 0.000 min
Response: 48904355
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500




#32 AR-1260-2

R.T.: 6.580 min
Delta R.T.: 0.000 min
Response: 31452019
Conc: 500.00 ng/ml




#33 AR-1260-3

R.T.: 7.868 min
Delta R.T.: 0.000 min
Response: 41485326
Conc: 500.00 ng/ml



#33 AR-1260-3


R.T.: 6.734 min
Delta R.T.: 0.000 min
Response: 26960319
Conc: 500.00 ng/ml



#34 AR-1260-4


R.T.: 8.092 min
Delta R.T.: 0.000 min
Response: 40725714
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500




#34 AR-1260-4

R.T.: 7.205 min
Delta R.T.: 0.000 min
Response: 23744314
Conc: 500.00 ng/ml



#35 AR-1260-5

R.T.: 8.411 min
Delta R.T.: 0.000 min
Response: 79980400
Conc: 500.00 ng/ml



#35 AR-1260-5

R.T.: 7.446 min
Delta R.T.: 0.000 min
Response: 62656405
Conc: 500.00 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070920.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 11:19
Operator : YP\AJ
Sample : AR1660ICC250
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 12:11:10 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 11:40:16 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1) SA	Tetrachlo...	4.515	3.816	39892911	28120307	26.258	27.124
2) SA	Decachlor...	10.229	8.859	27877719	18839232	24.970	21.779
	Target Compounds						
3) L1	AR-1016-1	5.667	4.903	13581167	11323714	266.642	298.339
4) L1	AR-1016-2	5.689	4.922	20108471	16375760	255.147	301.131
5) L1	AR-1016-3	5.751	5.099	12212490	8877034	260.862	297.397
6) L1	AR-1016-4	5.848	5.141	9803037	7158830	251.264	300.870
7) L1	AR-1016-5	6.141	5.355	9152735	9323046	261.203	309.177
31) L7	AR-1260-1	7.260	6.391	18182257	14056672	272.092	291.387
32) L7	AR-1260-2	7.513	6.579	25714746	17295947	262.909	274.958
33) L7	AR-1260-3	7.870	6.733	21528554	14684885	259.472m	272.343
34) L7	AR-1260-4	8.094	7.205	19536409	12488782	239.853m	262.985
35) L7	AR-1260-5	8.414	7.446	41714322	31815782	260.778m	253.891

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070920.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 11:19
Operator : YP\AJ
Sample : AR1660ICC250
Misc :
ALS Vial : 6 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1660ICC250


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 12:11:10 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 11:40:16 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene

R.T.: 4.515 min
Delta R.T.: 0.003 min
Response: 39892911
Conc: 26.26 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025




#1 Tetrachloro-m-xylene

R.T.: 3.816 min
Delta R.T.: 0.000 min
Response: 28120307
Conc: 27.12 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.229 min
Delta R.T.: 0.004 min
Response: 27877719
Conc: 24.97 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.859 min
Delta R.T.: 0.000 min
Response: 18839232
Conc: 21.78 ng/ml




#3 AR-1016-1

R.T.: 5.667 min
Delta R.T.: 0.004 min
Response: 13581167
Conc: 266.64 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025




#3 AR-1016-1

R.T.: 4.903 min
Delta R.T.: 0.000 min
Response: 11323714
Conc: 298.34 ng/ml




#4 AR-1016-2

R.T.: 5.689 min
Delta R.T.: 0.004 min
Response: 20108471
Conc: 255.15 ng/ml



#4 AR-1016-2

R.T.: 4.922 min
Delta R.T.: 0.000 min
Response: 16375760
Conc: 301.13 ng/ml




#5 AR-1016-3

R.T.: 5.751 min
Delta R.T.: 0.004 min
Response: 12212490
Conc: 260.86 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025




#5 AR-1016-3

R.T.: 5.099 min
Delta R.T.: 0.000 min
Response: 8877034
Conc: 297.40 ng/ml




#6 AR-1016-4

R.T.: 5.848 min
Delta R.T.: 0.003 min
Response: 9803037
Conc: 251.26 ng/ml



#6 AR-1016-4

R.T.: 5.141 min
Delta R.T.: 0.000 min
Response: 7158830
Conc: 300.87 ng/ml




#7 AR-1016-5

R.T.: 6.141 min
Delta R.T.: 0.003 min
Response: 9152735
Conc: 261.20 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025




#7 AR-1016-5

R.T.: 5.355 min
Delta R.T.: 0.000 min
Response: 9323046
Conc: 309.18 ng/ml




#31 AR-1260-1

R.T.: 7.260 min
Delta R.T.: 0.004 min
Response: 18182257
Conc: 272.09 ng/ml



#31 AR-1260-1

R.T.: 6.391 min
Delta R.T.: -0.001 min
Response: 14056672
Conc: 291.39 ng/ml




#32 AR-1260-2

R.T.: 7.513 min
Delta R.T.: 0.004 mi
Response: 25714746
Conc: 262.91 ng/m

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025




#32 AR-1260-2

R.T.: 6.579 min
Delta R.T.: 0.000 min
Response: 17295947
Conc: 274.96 ng/ml




#33 AR-1260-3

R.T.: 7.870 min
Delta R.T.: 0.002 min
Response: 21528554
Conc: 259.47 ng/ml m



#33 AR-1260-3

R.T.: 6.733 min
Delta R.T.: 0.000 min
Response: 14684885
Conc: 272.34 ng/ml




#34 AR-1260-4

R.T.: 8.094 min
Delta R.T.: 0.002 mi
Response: 19536409
Conc: 239.85 ng/m

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025




#34 AR-1260-4

R.T.: 7.205 min
Delta R.T.: 0.000 min
Response: 12488782
Conc: 262.98 ng/ml



#35 AR-1260-5

R.T.: 8.414 min
Delta R.T.: 0.002 min
Response: 41714322
Conc: 260.78 ng/ml m



#35 AR-1260-5

R.T.: 7.446 min
Delta R.T.: 0.000 min
Response: 31815782
Conc: 253.89 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070921.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 11:35
Operator : YP\AJ
Sample : AR1660ICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 12:09:28 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 11:40:16 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml	
	System Monitoring Compounds							
1)	SA Tetrachlo...	4.513	3.816	6533335	5359624	4.300	5.170	
2)	SA Decachlor...	10.230	8.858	4733739	3413055	4.240	3.946	
	Target Compounds							
3)	L1 AR-1016-1	5.666	4.903	2617394	2265971	51.388	59.700	
4)	L1 AR-1016-2	5.688	4.922	4198554	3321854	53.274	61.085	
5)	L1 AR-1016-3	5.751	5.099	2631097	1626306	56.201	54.484	
6)	L1 AR-1016-4	5.848	5.140	1614372	1232754	41.378	51.810	#
7)	L1 AR-1016-5	6.139	5.356	1610860	1866333	45.971m	61.893	#
31)	L7 AR-1260-1	7.257	6.391	3667364	2989801	54.881m	61.977	
32)	L7 AR-1260-2	7.512	6.579	4685539	3357287	47.905	53.372	
33)	L7 AR-1260-3	7.869	6.732	3780088	2827385	45.559m	52.436	
34)	L7 AR-1260-4	8.094	7.205	4593822	2746307	56.400m	57.831	
35)	L7 AR-1260-5	8.412	7.445	7845589	7013440	49.047m	55.967	

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070921.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 11:35
Operator : YP\AJ
Sample : AR1660ICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1660ICC050


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 12:09:28 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 11:40:16 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene

R.T.: 4.513 min
Delta R.T.: 0.001 min
Response: 6533335
Conc: 4.30 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025




#1 Tetrachloro-m-xylene

R.T.: 3.816 min
Delta R.T.: 0.000 min
Response: 5359624
Conc: 5.17 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.230 min
Delta R.T.: 0.005 min
Response: 4733739
Conc: 4.24 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.858 min
Delta R.T.: 0.000 min
Response: 3413055
Conc: 3.95 ng/ml




#3 AR-1016-1

R.T.: 5.666 min
Delta R.T.: 0.002 mi
Response: 2617394
Conc: 51.39 ng/m

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025




#3 AR-1016-1

R.T.: 4.903 min
Delta R.T.: 0.000 min
Response: 2265971
Conc: 59.70 ng/ml




#4 AR-1016-2

R.T.: 5.688 min
Delta R.T.: 0.002 min
Response: 4198554
Conc: 53.27 ng/ml



#4 AR-1016-2

R.T.: 4.922 min
Delta R.T.: 0.000 min
Response: 3321854
Conc: 61.09 ng/ml




#5 AR-1016-3

R.T.: 5.751 min
Delta R.T.: 0.003 min
Response: 2631097
Conc: 56.20 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025




#5 AR-1016-3

R.T.: 5.099 min
Delta R.T.: 0.000 min
Response: 1626306
Conc: 54.48 ng/ml




#6 AR-1016-4

R.T.: 5.848 min
Delta R.T.: 0.003 min
Response: 1614372
Conc: 41.38 ng/ml



#6 AR-1016-4

R.T.: 5.140 min
Delta R.T.: 0.000 min
Response: 1232754
Conc: 51.81 ng/ml




#7 AR-1016-5

R.T.: 6.139 min
Delta R.T.: 0.000 mi
Response: 1610860
Conc: 45.97 ng/m

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025




#7 AR-1016-5

R.T.: 5.356 min
Delta R.T.: 0.000 min
Response: 1866333
Conc: 61.89 ng/ml




#31 AR-1260-1

R.T.: 7.257 min
Delta R.T.: 0.001 min
Response: 3667364
Conc: 54.88 ng/ml m



#31 AR-1260-1

R.T.: 6.391 min
Delta R.T.: 0.000 min
Response: 2989801
Conc: 61.98 ng/ml




#32 AR-1260-2

R.T.: 7.512 min
Delta R.T.: 0.002 mi
Response: 4685539
Conc: 47.91 ng/m

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025




#32 AR-1260-2

R.T.: 6.579 min
Delta R.T.: 0.000 min
Response: 3357287
Conc: 53.37 ng/ml




#33 AR-1260-3

R.T.: 7.869 min
Delta R.T.: 0.002 min
Response: 3780088
Conc: 45.56 ng/ml m



#33 AR-1260-3

R.T.: 6.732 min
Delta R.T.: -0.001 min
Response: 2827385
Conc: 52.44 ng/ml




#34 AR-1260-4

R.T.: 8.094 min
Delta R.T.: 0.002 mi
Response: 4593822
Conc: 56.40 ng/m

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025




#34 AR-1260-4

R.T.: 7.205 min
Delta R.T.: 0.000 min
Response: 2746307
Conc: 57.83 ng/ml



#35 AR-1260-5

R.T.: 8.412 min
Delta R.T.: 0.000 min
Response: 7845589
Conc: 49.05 ng/ml m



#35 AR-1260-5

R.T.: 7.445 min
Delta R.T.: 0.000 min
Response: 7013440
Conc: 55.97 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070922.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 11:52
Operator : YP\AJ
Sample : AR1221ICC500
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 12:33:19 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 12:32:55 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.515	3.817	75324924	52315210	50.000	50.000
2)	SA Decachlor...	10.229	8.859	52098024	39006992	50.000	50.000
	Target Compounds						
8)	L2 AR-1221-1	4.716	4.027	8676827	7621792	500.000	500.000
9)	L2 AR-1221-2	4.802	4.114	6476464	5718144	500.000	500.000
10)	L2 AR-1221-3	4.878	4.190	20858618	17044480	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070922.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 11:52
Operator : YP\AJ
Sample : AR1221ICC500
Misc :
ALS Vial : 8 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1221ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 12:33:19 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 12:32:55 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.515 min
Delta R.T.: 0.000 min
Response: 75324924
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.817 min
Delta R.T.: 0.000 min
Response: 52315210
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.229 min
Delta R.T.: 0.000 min
Response: 52098024
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.859 min
Delta R.T.: 0.000 min
Response: 39006992
Conc: 50.00 ng/ml



#8 AR-1221-1


R.T.: 4.716 min
Delta R.T.: 0.000 mi
Response: 8676827
Conc: 500.00 ng/m

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500




#8 AR-1221-1

R.T.: 4.027 min
Delta R.T.: 0.000 min
Response: 7621792
Conc: 500.00 ng/ml




#9 AR-1221-2

R.T.: 4.802 min
Delta R.T.: 0.000 min
Response: 6476464
Conc: 500.00 ng/ml



#9 AR-1221-2


R.T.: 4.114 min
Delta R.T.: 0.000 min
Response: 5718144
Conc: 500.00 ng/ml



#10 AR-1221-3

R.T.: 4.878 min
Delta R.T.: 0.000 min
Response: 20858618
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500



#10 AR-1221-3

R.T.: 4.190 min
Delta R.T.: 0.000 min
Response: 17044480
Conc: 500.00 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070923.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 12:11
Operator : YP\AJ
Sample : AR1232ICC500
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1232ICC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 12:39:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 12:38:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.520	3.817	74435949	51438566	50.000	50.000
2)	SA Decachlor...	10.234	8.860	46115316	40517792	51.013m	50.000
	Target Compounds						
11)	L3 AR-1232-1	4.883	4.191	16644355	13993352	500.000	500.000
12)	L3 AR-1232-2	5.408	4.923	7955169	14030567	500.000	500.000
13)	L3 AR-1232-3	5.694	5.100	17135200	7444793	500.000	500.000
14)	L3 AR-1232-4	5.854	5.185	7853790	6402183	500.000	500.000
15)	L3 AR-1232-5	5.943	5.357	5320177	7087391	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070923.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 12:11
Operator : YP\AJ
Sample : AR1232ICC500
Misc :
ALS Vial : 9 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1232ICC500


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 12:39:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 12:38:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene

R.T.: 4.520 min
Delta R.T.: 0.000 mi
Response: 74435949
Conc: 50.00 ng/m

Instrument :
ECD_P
ClientSampleId :
AR1232ICC500


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025




#1 Tetrachloro-m-xylene

R.T.: 3.817 min
Delta R.T.: 0.000 min
Response: 51438566
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.234 min
Delta R.T.: 0.000 min
Response: 46115316
Conc: 51.01 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.860 min
Delta R.T.: 0.000 min
Response: 40517792
Conc: 50.00 ng/ml




#11 AR-1232-1

R.T.: 4.883 min
Delta R.T.: 0.000 min
Response: 16644355
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1232ICC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025




#11 AR-1232-1

R.T.: 4.191 min
Delta R.T.: 0.000 min
Response: 13993352
Conc: 500.00 ng/ml




#12 AR-1232-2

R.T.: 5.408 min
Delta R.T.: 0.000 min
Response: 7955169
Conc: 500.00 ng/ml



#12 AR-1232-2

R.T.: 4.923 min
Delta R.T.: 0.000 min
Response: 14030567
Conc: 500.00 ng/ml




#13 AR-1232-3

R.T.: 5.694 min
Delta R.T.: 0.000 min
Response: 17135200
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1232ICC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025




#13 AR-1232-3

R.T.: 5.100 min
Delta R.T.: 0.000 min
Response: 7444793
Conc: 500.00 ng/ml




#14 AR-1232-4

R.T.: 5.854 min
Delta R.T.: 0.000 min
Response: 7853790
Conc: 500.00 ng/ml



#14 AR-1232-4

R.T.: 5.185 min
Delta R.T.: 0.000 min
Response: 6402183
Conc: 500.00 ng/ml




#15 AR-1232-5

R.T.: 5.943 min
Delta R.T.: 0.000 min
Response: 5320177
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1232ICC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025



#15 AR-1232-5

R.T.: 5.357 min
Delta R.T.: 0.000 min
Response: 7087391
Conc: 500.00 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070924.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 12:28
Operator : YP\AJ
Sample : AR1242ICC1000
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 13:28:07 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 13:24:34 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.513	3.816	144.4E6	99447440	93.922	88.806
2)	SA Decachlor...	10.227	8.858	103.6E6	79468601	93.673	102.468
	Target Compounds						
16)	L4 AR-1242-1	5.666	4.903	40990074	30391529	943.761	877.115
17)	L4 AR-1242-2	5.687	4.922	64704598	43879042	965.382	887.614
18)	L4 AR-1242-3	5.749	5.099	39349869	24103124	991.106	905.487
19)	L4 AR-1242-4	5.846	5.183	30814559	23086538	954.688	902.469
20)	L4 AR-1242-5	6.577	5.707	36006298	28647762	964.427	907.760

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070924.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 12:28
Operator : YP\AJ
Sample : AR1242ICC1000
Misc :
ALS Vial : 10 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1242ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 13:28:07 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 13:24:34 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.513 min
Delta R.T.: 0.002 min
Response: 144394733
Conc: 93.92 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC1000




#1 Tetrachloro-m-xylene

R.T.: 3.816 min
Delta R.T.: 0.000 min
Response: 99447440
Conc: 88.81 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.227 min
Delta R.T.: 0.002 min
Response: 103629813
Conc: 93.67 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.858 min
Delta R.T.: 0.000 min
Response: 79468601
Conc: 102.47 ng/ml



#16 AR-1242-1


R.T.: 5.666 min
Delta R.T.: 0.002 min
Response: 40990074
Conc: 943.76 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC1000




#16 AR-1242-1

R.T.: 4.903 min
Delta R.T.: 0.001 min
Response: 30391529
Conc: 877.12 ng/ml




#17 AR-1242-2

R.T.: 5.687 min
Delta R.T.: 0.002 min
Response: 64704598
Conc: 965.38 ng/ml



#17 AR-1242-2


R.T.: 4.922 min
Delta R.T.: 0.001 min
Response: 43879042
Conc: 887.61 ng/ml



#18 AR-1242-3


R.T.: 5.749 min
Delta R.T.: 0.000 min
Response: 39349869
Conc: 991.11 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC1000




#18 AR-1242-3

R.T.: 5.099 min
Delta R.T.: 0.000 min
Response: 24103124
Conc: 905.49 ng/ml




#19 AR-1242-4

R.T.: 5.846 min
Delta R.T.: 0.001 min
Response: 30814559
Conc: 954.69 ng/ml



#19 AR-1242-4


R.T.: 5.183 min
Delta R.T.: 0.000 min
Response: 23086538
Conc: 902.47 ng/ml



#20 AR-1242-5

R.T.: 6.577 min
Delta R.T.: 0.001 mi
Response: 36006298
Conc: 964.43 ng/m

Instrument :
ECD_P
ClientSampleId :
AR1242ICC1000



#20 AR-1242-5

R.T.: 5.707 min
Delta R.T.: 0.000 min
Response: 28647762
Conc: 907.76 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070925.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 12:44
Operator : YP\AJ
Sample : AR1242ICC750
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 13:28:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 13:24:34 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.513	3.816	113.1E6	76564298	73.554	68.372
2)	SA Decachlor...	10.229	8.859	79761677	58459899	72.098	75.379
	Target Compounds						
16)	L4 AR-1242-1	5.666	4.903	32059711	23829543	738.147	687.733
17)	L4 AR-1242-2	5.687	4.921	47897203	33806804	714.618	683.867
18)	L4 AR-1242-3	5.750	5.099	28536378	18473047	718.747	693.981
19)	L4 AR-1242-4	5.847	5.183	23599640	17369419	731.158	678.983
20)	L4 AR-1242-5	6.577	5.707	26235924	21926365	702.728	694.780

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070925.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 12:44
Operator : YP\AJ
Sample : AR1242ICC750
Misc :
ALS Vial : 11 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1242ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 13:28:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 13:24:34 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.513 min
Delta R.T.: 0.002 min
Response: 113081073
Conc: 73.55 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC750




#1 Tetrachloro-m-xylene

R.T.: 3.816 min
Delta R.T.: 0.000 min
Response: 76564298
Conc: 68.37 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.229 min
Delta R.T.: 0.003 min
Response: 79761677
Conc: 72.10 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.859 min
Delta R.T.: 0.000 min
Response: 58459899
Conc: 75.38 ng/ml



#16 AR-1242-1


R.T.: 5.666 min
Delta R.T.: 0.002 min
Response: 32059711
Conc: 738.15 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC750




#16 AR-1242-1

R.T.: 4.903 min
Delta R.T.: 0.000 min
Response: 23829543
Conc: 687.73 ng/ml




#17 AR-1242-2

R.T.: 5.687 min
Delta R.T.: 0.001 min
Response: 47897203
Conc: 714.62 ng/ml



#17 AR-1242-2


R.T.: 4.921 min
Delta R.T.: 0.000 min
Response: 33806804
Conc: 683.87 ng/ml



#18 AR-1242-3


R.T.: 5.750 min
Delta R.T.: 0.003 min
Response: 28536378
Conc: 718.75 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC750




#18 AR-1242-3

R.T.: 5.099 min
Delta R.T.: 0.000 min
Response: 18473047
Conc: 693.98 ng/ml




#19 AR-1242-4

R.T.: 5.847 min
Delta R.T.: 0.002 min
Response: 23599640
Conc: 731.16 ng/ml



#19 AR-1242-4


R.T.: 5.183 min
Delta R.T.: 0.000 min
Response: 17369419
Conc: 678.98 ng/ml



#20 AR-1242-5

R.T.: 6.577 min
Delta R.T.: 0.002 mi
Response: 26235924
Conc: 702.73 ng/m

Instrument :
ECD_P
ClientSampleId :
AR1242ICC750



#20 AR-1242-5

R.T.: 5.707 min
Delta R.T.: 0.000 min
Response: 21926365
Conc: 694.78 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070926.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 13:00
Operator : YP\AJ
Sample : AR1242ICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 13:28:40 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 13:24:34 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.511	3.815	76869856	55991243	50.000	50.000
2)	SA Decachlor...	10.225	8.858	55314656	38777230	50.000	50.000
	Target Compounds						
16)	L4 AR-1242-1	5.664	4.902	21716335	17324704	500.000	500.000
17)	L4 AR-1242-2	5.686	4.921	33512438	24717400	500.000	500.000
18)	L4 AR-1242-3	5.748	5.099	19851486	13309479	500.000	500.000
19)	L4 AR-1242-4	5.845	5.182	16138543	12790769	500.000	500.000
20)	L4 AR-1242-5	6.576	5.707	18667188	15779369	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070926.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 13:00
Operator : YP\AJ
Sample : AR1242ICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1242ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 13:28:40 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 13:24:34 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.511 min
Delta R.T.: 0.000 min
Response: 76869856
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.815 min
Delta R.T.: 0.000 min
Response: 55991243
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.225 min
Delta R.T.: 0.000 min
Response: 55314656
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.858 min
Delta R.T.: 0.000 min
Response: 38777230
Conc: 50.00 ng/ml



#16 AR-1242-1


R.T.: 5.664 min
Delta R.T.: 0.000 min
Response: 21716335
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC500




#16 AR-1242-1

R.T.: 4.902 min
Delta R.T.: 0.000 min
Response: 17324704
Conc: 500.00 ng/ml




#17 AR-1242-2

R.T.: 5.686 min
Delta R.T.: 0.000 min
Response: 33512438
Conc: 500.00 ng/ml



#17 AR-1242-2


R.T.: 4.921 min
Delta R.T.: 0.000 min
Response: 24717400
Conc: 500.00 ng/ml



#18 AR-1242-3


R.T.: 5.748 min
Delta R.T.: 0.000 min
Response: 19851486
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC500




#18 AR-1242-3

R.T.: 5.099 min
Delta R.T.: 0.000 min
Response: 13309479
Conc: 500.00 ng/ml




#19 AR-1242-4

R.T.: 5.845 min
Delta R.T.: 0.000 min
Response: 16138543
Conc: 500.00 ng/ml



#19 AR-1242-4


R.T.: 5.182 min
Delta R.T.: 0.000 min
Response: 12790769
Conc: 500.00 ng/ml



#20 AR-1242-5

R.T.: 6.576 min
Delta R.T.: 0.000 min
Response: 18667188
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC500



#20 AR-1242-5

R.T.: 5.707 min
Delta R.T.: 0.000 min
Response: 15779369
Conc: 500.00 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070927.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 13:16
Operator : YP\AJ
Sample : AR1242ICC250
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 13:36:45 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 13:24:34 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.514	3.816	39179148	28013392	25.484	25.016
2)	SA Decachlor...	10.230	8.858	28836317	21287334	26.066	27.448
	Target Compounds						
16)	L4 AR-1242-1	5.666	4.903	10996159	9674680	253.177	279.216
17)	L4 AR-1242-2	5.688	4.922	17212597	13875945	256.809	280.692
18)	L4 AR-1242-3	5.750	5.099	10156885	7458368	255.822	280.190
19)	L4 AR-1242-4	5.847	5.183	8048661	7154684	249.361	279.682
20)	L4 AR-1242-5	6.578	5.708	9640099	8240601	258.210	261.119

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070927.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 13:16
Operator : YP\AJ
Sample : AR1242ICC250
Misc :
ALS Vial : 13 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1242ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 13:36:45 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 13:24:34 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.514 min
Delta R.T.: 0.003 min
Response: 39179148
Conc: 25.48 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.816 min
Delta R.T.: 0.000 min
Response: 28013392
Conc: 25.02 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.230 min
Delta R.T.: 0.004 min
Response: 28836317
Conc: 26.07 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.858 min
Delta R.T.: 0.000 min
Response: 21287334
Conc: 27.45 ng/ml



#16 AR-1242-1


R.T.: 5.666 min
Delta R.T.: 0.002 min
Response: 10996159
Conc: 253.18 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250




#16 AR-1242-1

R.T.: 4.903 min
Delta R.T.: 0.000 min
Response: 9674680
Conc: 279.22 ng/ml




#17 AR-1242-2

R.T.: 5.688 min
Delta R.T.: 0.002 min
Response: 17212597
Conc: 256.81 ng/ml



#17 AR-1242-2


R.T.: 4.922 min
Delta R.T.: 0.000 min
Response: 13875945
Conc: 280.69 ng/ml



#18 AR-1242-3


R.T.: 5.750 min
Delta R.T.: 0.002 min
Response: 10156885
Conc: 255.82 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250




#18 AR-1242-3

R.T.: 5.099 min
Delta R.T.: 0.000 min
Response: 7458368
Conc: 280.19 ng/ml




#19 AR-1242-4

R.T.: 5.847 min
Delta R.T.: 0.002 min
Response: 8048661
Conc: 249.36 ng/ml



#19 AR-1242-4


R.T.: 5.183 min
Delta R.T.: 0.000 min
Response: 7154684
Conc: 279.68 ng/ml



#20 AR-1242-5

R.T.: 6.578 min
Delta R.T.: 0.002 mi
Response: 9640099
Conc: 258.21 ng/m

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250



#20 AR-1242-5

R.T.: 5.708 min
Delta R.T.: 0.000 min
Response: 8240601
Conc: 261.12 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070928.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 13:33
Operator : YP\AJ
Sample : AR1242ICC050
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 13:57:16 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 13:24:34 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.512	3.815	6633492	5532966	4.273	4.939
2)	SA Decachlor...	10.226	8.857	5282830	3387334	4.676	4.164
	Target Compounds						
16)	L4 AR-1242-1	5.663	4.902	2436604	1773341	55.746m	48.354
17)	L4 AR-1242-2	5.686	4.921	2856942	2619762	42.052m	49.929m
18)	L4 AR-1242-3	5.748	5.098	2373419	1401755	59.091m	49.661
19)	L4 AR-1242-4	5.843	5.181	1293211	1572257	40.117m	58.017m#
20)	L4 AR-1242-5	6.574	5.705	1783487	1619372	46.999m	50.197m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070928.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 13:33
Operator : YP\AJ
Sample : AR1242ICC050
Misc :
ALS Vial : 14 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1242ICC050


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 13:57:16 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 13:24:34 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene

R.T.: 4.512 min
Delta R.T.: 0.001 min
Response: 6633492
Conc: 4.27 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC050


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025




#1 Tetrachloro-m-xylene

R.T.: 3.815 min
Delta R.T.: 0.000 min
Response: 5532966
Conc: 4.94 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.226 min
Delta R.T.: 0.000 min
Response: 5282830
Conc: 4.68 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.857 min
Delta R.T.: -0.001 min
Response: 3387334
Conc: 4.16 ng/ml




#16 AR-1242-1

R.T.: 5.663 min
Delta R.T.: -0.002 mi
Response: 2436604
Conc: 55.75 ng/m

Instrument :
ECD_P
ClientSampleId :
AR1242ICC050


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025




#16 AR-1242-1

R.T.: 4.902 min
Delta R.T.: 0.000 min
Response: 1773341
Conc: 48.35 ng/ml




#17 AR-1242-2

R.T.: 5.686 min
Delta R.T.: 0.000 min
Response: 2856942
Conc: 42.05 ng/ml m



#17 AR-1242-2

R.T.: 4.921 min
Delta R.T.: 0.000 min
Response: 2619762
Conc: 49.93 ng/ml m




#18 AR-1242-3

R.T.: 5.748 min
Delta R.T.: 0.000 min
Response: 2373419
Conc: 59.09 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC050


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025




#18 AR-1242-3

R.T.: 5.098 min
Delta R.T.: 0.000 min
Response: 1401755
Conc: 49.66 ng/ml




#19 AR-1242-4

R.T.: 5.843 min
Delta R.T.: -0.002 min
Response: 1293211
Conc: 40.12 ng/ml m



#19 AR-1242-4

R.T.: 5.181 min
Delta R.T.: -0.002 min
Response: 1572257
Conc: 58.02 ng/ml m




#20 AR-1242-5

R.T.: 6.574 min
Delta R.T.: -0.002 mi
Response: 1783487
Conc: 47.00 ng/m

Instrument :
ECD_P
ClientSampleId :
AR1242ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025



#20 AR-1242-5

R.T.: 5.705 min
Delta R.T.: -0.001 min
Response: 1619372
Conc: 50.20 ng/ml m

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070929.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 13:49
Operator : YP\AJ
Sample : AR1248ICC1000
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 14:45:43 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 14:45:17 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.514	3.816	146.2E6	96952216	94.393	85.183
2)	SA Decachlor...	10.229	8.859	107.2E6	76343240	88.954	92.303
	Target Compounds						
21)	L5 AR-1248-1	5.666	4.903	31141042	23297586	921.836	817.423
22)	L5 AR-1248-2	5.938	5.141	42399779	31071420	917.898	811.806
23)	L5 AR-1248-3	6.140	5.183	48055172	32592570	927.258	815.546
24)	L5 AR-1248-4	6.539	5.355	59878410	38114246	893.176	807.817
25)	L5 AR-1248-5	6.579	5.749	56652876	37031094	881.261	843.047

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070929.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 13:49
Operator : YP\AJ
Sample : AR1248ICC1000
Misc :
ALS Vial : 15 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 14:45:43 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 14:45:17 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.514 min
Delta R.T.: 0.000 min
Response: 146218105
Conc: 94.39 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000




#1 Tetrachloro-m-xylene

R.T.: 3.816 min
Delta R.T.: 0.000 min
Response: 96952216
Conc: 85.18 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.229 min
Delta R.T.: 0.002 min
Response: 107201730
Conc: 88.95 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.859 min
Delta R.T.: 0.000 min
Response: 76343240
Conc: 92.30 ng/ml



#21 AR-1248-1


R.T.: 5.666 min
Delta R.T.: 0.000 mi
Response: 31141042
Conc: 921.84 ng/m

Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000




#21 AR-1248-1

R.T.: 4.903 min
Delta R.T.: 0.000 min
Response: 23297586
Conc: 817.42 ng/ml




#22 AR-1248-2

R.T.: 5.938 min
Delta R.T.: 0.000 min
Response: 42399779
Conc: 917.90 ng/ml



#22 AR-1248-2


R.T.: 5.141 min
Delta R.T.: 0.000 min
Response: 31071420
Conc: 811.81 ng/ml



#23 AR-1248-3


R.T.: 6.140 min
Delta R.T.: 0.000 min
Response: 48055172
Conc: 927.26 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000




#23 AR-1248-3

R.T.: 5.183 min
Delta R.T.: 0.000 min
Response: 32592570
Conc: 815.55 ng/ml




#24 AR-1248-4

R.T.: 6.539 min
Delta R.T.: 0.000 min
Response: 59878410
Conc: 893.18 ng/ml



#24 AR-1248-4


R.T.: 5.355 min
Delta R.T.: 0.000 min
Response: 38114246
Conc: 807.82 ng/ml



#25 AR-1248-5

R.T.: 6.579 min
Delta R.T.: 0.000 min
Response: 56652876
Conc: 881.26 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000



#25 AR-1248-5

R.T.: 5.749 min
Delta R.T.: 0.000 min
Response: 37031094
Conc: 843.05 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070930.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 14:05
Operator : YP\AJ
Sample : AR1248ICC750
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 14:46:03 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 14:45:17 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1) SA	Tetrachlo...	4.514	3.816	115.8E6	78436191	74.760	68.915
2) SA	Decachlor...	10.229	8.860	84334663	61664461	69.979	74.555
	Target Compounds						
21) L5	AR-1248-1	5.666	4.903	25034469	20052758	741.070	703.575
22) L5	AR-1248-2	5.938	5.141	33734311	26938678	730.302	703.829
23) L5	AR-1248-3	6.140	5.183	38083098	28275820	734.840	707.530
24) L5	AR-1248-4	6.539	5.356	47603456	32558799	710.077	690.072
25) L5	AR-1248-5	6.578	5.749	45378306	30836259	705.880	702.016

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070930.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 14:05
Operator : YP\AJ
Sample : AR1248ICC750
Misc :
ALS Vial : 16 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1248ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 14:46:03 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 14:45:17 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.514 min
Delta R.T.: 0.000 min
Response: 115806846
Conc: 74.76 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC750




#1 Tetrachloro-m-xylene

R.T.: 3.816 min
Delta R.T.: 0.000 min
Response: 78436191
Conc: 68.91 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.229 min
Delta R.T.: 0.003 min
Response: 84334663
Conc: 69.98 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.860 min
Delta R.T.: 0.000 min
Response: 61664461
Conc: 74.56 ng/ml



#21 AR-1248-1


R.T.: 5.666 min
Delta R.T.: 0.000 min
Response: 25034469
Conc: 741.07 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC750




#21 AR-1248-1

R.T.: 4.903 min
Delta R.T.: 0.000 min
Response: 20052758
Conc: 703.57 ng/ml




#22 AR-1248-2

R.T.: 5.938 min
Delta R.T.: 0.000 min
Response: 33734311
Conc: 730.30 ng/ml



#22 AR-1248-2


R.T.: 5.141 min
Delta R.T.: 0.000 min
Response: 26938678
Conc: 703.83 ng/ml



#23 AR-1248-3


R.T.: 6.140 min
Delta R.T.: 0.000 min
Response: 38083098
Conc: 734.84 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC750




#23 AR-1248-3

R.T.: 5.183 min
Delta R.T.: 0.000 min
Response: 28275820
Conc: 707.53 ng/ml




#24 AR-1248-4

R.T.: 6.539 min
Delta R.T.: 0.000 min
Response: 47603456
Conc: 710.08 ng/ml



#24 AR-1248-4


R.T.: 5.356 min
Delta R.T.: 0.000 min
Response: 32558799
Conc: 690.07 ng/ml



#25 AR-1248-5

R.T.: 6.578 min
Delta R.T.: 0.000 min
Response: 45378306
Conc: 705.88 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC750



#25 AR-1248-5

R.T.: 5.749 min
Delta R.T.: 0.000 min
Response: 30836259
Conc: 702.02 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070931.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 14:22
Operator : YP\AJ
Sample : AR1248ICC500
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 14:46:15 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 14:45:17 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.513	3.816	77451968	56907988	50.000	50.000
2)	SA Decachlor...	10.226	8.859	60256834	41354865	50.000	50.000
	Target Compounds						
21)	L5 AR-1248-1	5.665	4.903	16890769	14250625	500.000	500.000
22)	L5 AR-1248-2	5.937	5.140	23096128	19137230	500.000	500.000
23)	L5 AR-1248-3	6.140	5.183	25912526	19982060	500.000	500.000
24)	L5 AR-1248-4	6.539	5.355	33519946	23590888	500.000	500.000
25)	L5 AR-1248-5	6.578	5.748	32143070	21962659	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070931.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 14:22
Operator : YP\AJ
Sample : AR1248ICC500
Misc :
ALS Vial : 17 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1248ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 14:46:15 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 14:45:17 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.513 min
Delta R.T.: 0.000 min
Response: 77451968
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.816 min
Delta R.T.: 0.000 min
Response: 56907988
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.226 min
Delta R.T.: 0.000 min
Response: 60256834
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.859 min
Delta R.T.: 0.000 min
Response: 41354865
Conc: 50.00 ng/ml



#21 AR-1248-1


R.T.: 5.665 min
Delta R.T.: 0.000 min
Response: 16890769
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC500




#21 AR-1248-1

R.T.: 4.903 min
Delta R.T.: 0.000 min
Response: 14250625
Conc: 500.00 ng/ml




#22 AR-1248-2

R.T.: 5.937 min
Delta R.T.: 0.000 min
Response: 23096128
Conc: 500.00 ng/ml



#22 AR-1248-2


R.T.: 5.140 min
Delta R.T.: 0.000 min
Response: 19137230
Conc: 500.00 ng/ml



#23 AR-1248-3


R.T.: 6.140 min
Delta R.T.: 0.000 min
Response: 25912526
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC500




#23 AR-1248-3

R.T.: 5.183 min
Delta R.T.: 0.000 min
Response: 19982060
Conc: 500.00 ng/ml




#24 AR-1248-4

R.T.: 6.539 min
Delta R.T.: 0.000 min
Response: 33519946
Conc: 500.00 ng/ml



#24 AR-1248-4


R.T.: 5.355 min
Delta R.T.: 0.000 min
Response: 23590888
Conc: 500.00 ng/ml



#25 AR-1248-5

R.T.: 6.578 min
Delta R.T.: 0.000 min
Response: 32143070
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC500



#25 AR-1248-5

R.T.: 5.748 min
Delta R.T.: 0.000 min
Response: 21962659
Conc: 500.00 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070932.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 14:38
Operator : YP\AJ
Sample : AR1248ICC250
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 15:07:19 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 15:04:54 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.517	3.817	38987130	27101787	25.676	25.783
2)	SA Decachlor...	10.232	8.860	29177292	20157827	25.732	25.064
	Target Compounds						
21)	L5 AR-1248-1	5.672	4.904	11497190	7301885	350.874	278.926
22)	L5 AR-1248-2	5.941	5.141	11780544	9675655	264.590	275.754
23)	L5 AR-1248-3	6.144	5.184	12674295	10119428	252.379	275.339
24)	L5 AR-1248-4	6.543	5.356	18357259	12061061	289.258	281.127
25)	L5 AR-1248-5	6.582	5.750	17129846	11658567	283.226	286.518

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070932.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 14:38
Operator : YP\AJ
Sample : AR1248ICC250
Misc :
ALS Vial : 18 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1248ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 15:07:19 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 15:04:54 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.517 min
Delta R.T.: 0.004 min
Response: 38987130
Conc: 25.68 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.817 min
Delta R.T.: 0.000 min
Response: 27101787
Conc: 25.78 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.232 min
Delta R.T.: 0.006 min
Response: 29177292
Conc: 25.73 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.860 min
Delta R.T.: 0.000 min
Response: 20157827
Conc: 25.06 ng/ml



#21 AR-1248-1


R.T.: 5.672 min
Delta R.T.: 0.007 min
Response: 11497190
Conc: 350.87 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC250




#21 AR-1248-1

R.T.: 4.904 min
Delta R.T.: 0.000 min
Response: 7301885
Conc: 278.93 ng/ml




#22 AR-1248-2

R.T.: 5.941 min
Delta R.T.: 0.004 min
Response: 11780544
Conc: 264.59 ng/ml



#22 AR-1248-2


R.T.: 5.141 min
Delta R.T.: 0.001 min
Response: 9675655
Conc: 275.75 ng/ml



#23 AR-1248-3


R.T.: 6.144 min
Delta R.T.: 0.004 min
Response: 12674295
Conc: 252.38 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC250




#23 AR-1248-3

R.T.: 5.184 min
Delta R.T.: 0.000 min
Response: 10119428
Conc: 275.34 ng/ml




#24 AR-1248-4

R.T.: 6.543 min
Delta R.T.: 0.004 min
Response: 18357259
Conc: 289.26 ng/ml



#24 AR-1248-4


R.T.: 5.356 min
Delta R.T.: 0.002 min
Response: 12061061
Conc: 281.13 ng/ml



#25 AR-1248-5

R.T.: 6.582 min
Delta R.T.: 0.004 min
Response: 17129846
Conc: 283.23 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC250



#25 AR-1248-5

R.T.: 5.750 min
Delta R.T.: 0.002 min
Response: 11658567
Conc: 286.52 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070933.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 14:55
Operator : YP\AJ
Sample : AR1248ICC050
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 15:16:04 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 15:04:54 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.513	3.815	7324663	5848649	4.791	5.521
2)	SA Decachlor...	10.227	8.859	5549589	3661642	4.859	4.550
	Target Compounds						
21)	L5 AR-1248-1	5.665	4.902	1857345	1382826	51.489	51.338
22)	L5 AR-1248-2	5.935	5.140	2199982	1982278	48.701m	55.076
23)	L5 AR-1248-3	6.137	5.182	2591630	2128069	51.484m	56.472
24)	L5 AR-1248-4	6.536	5.354	2845967	2437650	43.150m	55.103 #
25)	L5 AR-1248-5	6.576	5.748	2713422	2234402	43.421m	52.978

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070933.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 14:55
Operator : YP\AJ
Sample : AR1248ICC050
Misc :
ALS Vial : 19 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1248ICC050


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 15:16:04 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 15:04:54 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene

R.T.: 4.513 min
Delta R.T.: 0.000 min
Response: 7324663
Conc: 4.79 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC050


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025




#1 Tetrachloro-m-xylene

R.T.: 3.815 min
Delta R.T.: 0.000 min
Response: 5848649
Conc: 5.52 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.227 min
Delta R.T.: 0.000 min
Response: 5549589
Conc: 4.86 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.859 min
Delta R.T.: 0.000 min
Response: 3661642
Conc: 4.55 ng/ml




#21 AR-1248-1

R.T.: 5.665 min
Delta R.T.: 0.000 mi
Response: 1857345
Conc: 51.49 ng/m

Instrument :
ECD_P
ClientSampleId :
AR1248ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025




#21 AR-1248-1

R.T.: 4.902 min
Delta R.T.: 0.000 min
Response: 1382826
Conc: 51.34 ng/ml




#22 AR-1248-2

R.T.: 5.935 min
Delta R.T.: -0.002 min
Response: 2199982
Conc: 48.70 ng/ml m



#22 AR-1248-2

R.T.: 5.140 min
Delta R.T.: 0.000 min
Response: 1982278
Conc: 55.08 ng/ml




#23 AR-1248-3

R.T.: 6.137 min
Delta R.T.: -0.002 mi
Response: 2591630
Conc: 51.48 ng/m

Instrument :
ECD_P
ClientSampleId :
AR1248ICC050


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025




#23 AR-1248-3

R.T.: 5.182 min
Delta R.T.: 0.000 min
Response: 2128069
Conc: 56.47 ng/ml




#24 AR-1248-4

R.T.: 6.536 min
Delta R.T.: -0.003 min
Response: 2845967
Conc: 43.15 ng/ml m



#24 AR-1248-4

R.T.: 5.354 min
Delta R.T.: 0.000 min
Response: 2437650
Conc: 55.10 ng/ml




#25 AR-1248-5

R.T.: 6.576 min
Delta R.T.: -0.001 min
Response: 2713422
Conc: 43.42 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025



#25 AR-1248-5

R.T.: 5.748 min
Delta R.T.: 0.000 min
Response: 2234402
Conc: 52.98 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070934.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 15:11
Operator : YP\AJ
Sample : AR1254ICC1000
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 16:19:31 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 16:07:11 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.514	3.816	150.1E6	104.3E6	95.275	94.048
2)	SA Decachlor...	10.231	8.860	109.5E6	78348381	93.754	86.018
	Target Compounds						
26)	L6 AR-1254-1	6.516	5.709	60157506	58928147	918.830	891.736
27)	L6 AR-1254-2	6.732	5.857	92838442	50145967	911.948	882.952
28)	L6 AR-1254-3	7.095	6.261	92334650	77448562	928.765	882.456
29)	L6 AR-1254-4	7.379	6.490	84238161	49998215	970.458	893.353
30)	L6 AR-1254-5	7.792	6.908	74433164	65979008	920.580m	918.596

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070934.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 15:11
Operator : YP\AJ
Sample : AR1254ICC1000
Misc :
ALS Vial : 20 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 16:19:31 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 16:07:11 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene

R.T.: 4.514 min
Delta R.T.: 0.000 min
Response: 150128836
Conc: 95.27 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025




#1 Tetrachloro-m-xylene

R.T.: 3.816 min
Delta R.T.: 0.000 min
Response: 104259560
Conc: 94.05 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.231 min
Delta R.T.: 0.001 min
Response: 109543241
Conc: 93.75 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.860 min
Delta R.T.: 0.000 min
Response: 78348381
Conc: 86.02 ng/ml




#26 AR-1254-1

R.T.: 6.516 min
Delta R.T.: 0.000 min
Response: 60157506
Conc: 918.83 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025




#26 AR-1254-1

R.T.: 5.709 min
Delta R.T.: 0.000 min
Response: 58928147
Conc: 891.74 ng/ml




#27 AR-1254-2

R.T.: 6.732 min
Delta R.T.: 0.000 min
Response: 92838442
Conc: 911.95 ng/ml



#27 AR-1254-2

R.T.: 5.857 min
Delta R.T.: 0.000 min
Response: 50145967
Conc: 882.95 ng/ml




#28 AR-1254-3

R.T.: 7.095 min
Delta R.T.: 0.000 min
Response: 92334650
Conc: 928.77 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025




#28 AR-1254-3

R.T.: 6.261 min
Delta R.T.: 0.001 min
Response: 77448562
Conc: 882.46 ng/ml




#29 AR-1254-4

R.T.: 7.379 min
Delta R.T.: 0.002 min
Response: 84238161
Conc: 970.46 ng/ml



#29 AR-1254-4

R.T.: 6.490 min
Delta R.T.: 0.000 min
Response: 49998215
Conc: 893.35 ng/ml




#30 AR-1254-5

R.T.: 7.792 min
Delta R.T.: 0.000 min
Response: 74433164
Conc: 920.58 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025



#30 AR-1254-5

R.T.: 6.908 min
Delta R.T.: 0.000 min
Response: 65979008
Conc: 918.60 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070935.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 15:27
Operator : YP\AJ
Sample : AR1254ICC750
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 16:19:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 16:07:11 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.514	3.816	116.3E6	81865229	73.835	73.847
2)	SA Decachlor...	10.228	8.859	85851924	59713645	73.477	65.559
	Target Compounds						
26)	L6 AR-1254-1	6.515	5.708	47300573	46349687	722.456	701.391
27)	L6 AR-1254-2	6.731	5.856	72785411	40339655	714.968	710.286
28)	L6 AR-1254-3	7.094	6.260	73541603	62731680	739.732	714.771
29)	L6 AR-1254-4	7.377	6.490	65835410	40961211	758.451	731.883
30)	L6 AR-1254-5	7.792	6.907	59278726	53519032	733.152m	745.122

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070935.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 15:27
Operator : YP\AJ
Sample : AR1254ICC750
Misc :
ALS Vial : 21 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1254ICC750


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 16:19:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 16:07:11 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene

R.T.: 4.514 min
Delta R.T.: 0.000 min
Response: 116345667
Conc: 73.84 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025




#1 Tetrachloro-m-xylene

R.T.: 3.816 min
Delta R.T.: 0.000 min
Response: 81865229
Conc: 73.85 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.228 min
Delta R.T.: -0.002 min
Response: 85851924
Conc: 73.48 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.859 min
Delta R.T.: 0.000 min
Response: 59713645
Conc: 65.56 ng/ml




#26 AR-1254-1

R.T.: 6.515 min
Delta R.T.: 0.000 min
Response: 47300573
Conc: 722.46 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025




#26 AR-1254-1

R.T.: 5.708 min
Delta R.T.: 0.000 min
Response: 46349687
Conc: 701.39 ng/ml




#27 AR-1254-2

R.T.: 6.731 min
Delta R.T.: 0.000 min
Response: 72785411
Conc: 714.97 ng/ml



#27 AR-1254-2

R.T.: 5.856 min
Delta R.T.: 0.000 min
Response: 40339655
Conc: 710.29 ng/ml




#28 AR-1254-3

R.T.: 7.094 min
Delta R.T.: 0.000 min
Response: 73541603
Conc: 739.73 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025




#28 AR-1254-3

R.T.: 6.260 min
Delta R.T.: 0.000 min
Response: 62731680
Conc: 714.77 ng/ml




#29 AR-1254-4

R.T.: 7.377 min
Delta R.T.: 0.000 min
Response: 65835410
Conc: 758.45 ng/ml



#29 AR-1254-4

R.T.: 6.490 min
Delta R.T.: 0.000 min
Response: 40961211
Conc: 731.88 ng/ml




#30 AR-1254-5

R.T.: 7.792 min
Delta R.T.: -0.001 mi
Response: 59278726
Conc: 733.15 ng/m

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025



#30 AR-1254-5

R.T.: 6.907 min
Delta R.T.: 0.000 min
Response: 53519032
Conc: 745.12 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070936.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 15:43
Operator : YP\AJ
Sample : AR1254ICC500
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 16:19:54 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 16:07:11 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.514	3.816	78787321	55428912	50.000	50.000
2)	SA Decachlor...	10.230	8.859	58420861	45541627	50.000	50.000
	Target Compounds						
26)	L6 AR-1254-1	6.515	5.708	32735943	33041243	500.000	500.000
27)	L6 AR-1254-2	6.732	5.857	50901196	28396777	500.000	500.000
28)	L6 AR-1254-3	7.095	6.260	49708277	43882389	500.000	500.000
29)	L6 AR-1254-4	7.377	6.489	43401255	27983450	500.000	500.000
30)	L6 AR-1254-5	7.793	6.907	40427325	35912948	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070936.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 15:43
Operator : YP\AJ
Sample : AR1254ICC500
Misc :
ALS Vial : 22 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1254ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 16:19:54 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 16:07:11 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.514 min
Delta R.T.: 0.000 min
Response: 78787321
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.816 min
Delta R.T.: 0.000 min
Response: 55428912
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.230 min
Delta R.T.: 0.000 min
Response: 58420861
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.859 min
Delta R.T.: 0.000 min
Response: 45541627
Conc: 50.00 ng/ml



#26 AR-1254-1


R.T.: 6.515 min
Delta R.T.: 0.000 min
Response: 32735943
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500




#26 AR-1254-1

R.T.: 5.708 min
Delta R.T.: 0.000 min
Response: 33041243
Conc: 500.00 ng/ml




#27 AR-1254-2

R.T.: 6.732 min
Delta R.T.: 0.000 min
Response: 50901196
Conc: 500.00 ng/ml



#27 AR-1254-2


R.T.: 5.857 min
Delta R.T.: 0.000 min
Response: 28396777
Conc: 500.00 ng/ml



#28 AR-1254-3


R.T.: 7.095 min
Delta R.T.: 0.000 min
Response: 49708277
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500




#28 AR-1254-3

R.T.: 6.260 min
Delta R.T.: 0.000 min
Response: 43882389
Conc: 500.00 ng/ml




#29 AR-1254-4

R.T.: 7.377 min
Delta R.T.: 0.000 min
Response: 43401255
Conc: 500.00 ng/ml



#29 AR-1254-4


R.T.: 6.489 min
Delta R.T.: 0.000 min
Response: 27983450
Conc: 500.00 ng/ml



#30 AR-1254-5

R.T.: 7.793 min
Delta R.T.: 0.000 min
Response: 40427325
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500



#30 AR-1254-5

R.T.: 6.907 min
Delta R.T.: 0.000 min
Response: 35912948
Conc: 500.00 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070937.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 16:00
Operator : YP\AJ
Sample : AR1254ICC250
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 16:20:08 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 16:07:11 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.514	3.816	40672603	29879101	25.812	26.953
2)	SA Decachlor...	10.230	8.859	29858259	21257096	25.554	23.338
	Target Compounds						
26)	L6 AR-1254-1	6.516	5.708	16701604	18446961	255.096	279.151
27)	L6 AR-1254-2	6.732	5.856	26389026	16008677	259.218	281.875
28)	L6 AR-1254-3	7.095	6.260	25960285	24772594	261.126	282.261
29)	L6 AR-1254-4	7.377	6.489	22197209	16043587	255.721	286.662
30)	L6 AR-1254-5	7.791	6.907	20989646	20606411	259.597m	286.894

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070937.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 16:00
Operator : YP\AJ
Sample : AR1254ICC250
Misc :
ALS Vial : 23 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1254ICC250


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 16:20:08 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 16:07:11 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene

R.T.: 4.514 min
Delta R.T.: 0.000 min
Response: 40672603
Conc: 25.81 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025




#1 Tetrachloro-m-xylene

R.T.: 3.816 min
Delta R.T.: 0.000 min
Response: 29879101
Conc: 26.95 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.230 min
Delta R.T.: 0.000 min
Response: 29858259
Conc: 25.55 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.859 min
Delta R.T.: 0.000 min
Response: 21257096
Conc: 23.34 ng/ml




#26 AR-1254-1

R.T.: 6.516 min
Delta R.T.: 0.000 min
Response: 16701604
Conc: 255.10 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025




#26 AR-1254-1

R.T.: 5.708 min
Delta R.T.: 0.000 min
Response: 18446961
Conc: 279.15 ng/ml




#27 AR-1254-2

R.T.: 6.732 min
Delta R.T.: 0.000 min
Response: 26389026
Conc: 259.22 ng/ml



#27 AR-1254-2

R.T.: 5.856 min
Delta R.T.: 0.000 min
Response: 16008677
Conc: 281.87 ng/ml




#28 AR-1254-3

R.T.: 7.095 min
Delta R.T.: 0.000 min
Response: 25960285
Conc: 261.13 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025




#28 AR-1254-3

R.T.: 6.260 min
Delta R.T.: 0.000 min
Response: 24772594
Conc: 282.26 ng/ml




#29 AR-1254-4

R.T.: 7.377 min
Delta R.T.: 0.000 min
Response: 22197209
Conc: 255.72 ng/ml



#29 AR-1254-4

R.T.: 6.489 min
Delta R.T.: 0.000 min
Response: 16043587
Conc: 286.66 ng/ml




#30 AR-1254-5

R.T.: 7.791 min
Delta R.T.: -0.001 min
Response: 20989646
Conc: 259.60 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025



#30 AR-1254-5

R.T.: 6.907 min
Delta R.T.: 0.000 min
Response: 20606411
Conc: 286.89 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070938.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 16:16
Operator : YP\AJ
Sample : AR1254ICC050
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 16:35:00 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 16:07:11 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.515	3.816	6508593	5336776	4.107	4.715
2)	SA Decachlor...	10.232	8.861	5177583	3796388	4.429	4.454
	Target Compounds						
26)	L6 AR-1254-1	6.516	5.708	3322396	3474548	51.023	51.687
27)	L6 AR-1254-2	6.733	5.857	5410437	3036515	53.321	52.169
28)	L6 AR-1254-3	7.096	6.261	5249564	4313044	52.267	47.835
29)	L6 AR-1254-4	7.378	6.490	4268685	2787399	48.623	47.851
30)	L6 AR-1254-5	7.794	6.907	4173507	3458194	51.345	45.985

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070938.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 16:16
Operator : YP\AJ
Sample : AR1254ICC050
Misc :
ALS Vial : 24 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1254ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 16:35:00 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 16:07:11 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.515 min
Delta R.T.: 0.000 min
Response: 6508593
Conc: 4.11 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC050




#1 Tetrachloro-m-xylene

R.T.: 3.816 min
Delta R.T.: 0.000 min
Response: 5336776
Conc: 4.72 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.232 min
Delta R.T.: 0.002 min
Response: 5177583
Conc: 4.43 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.861 min
Delta R.T.: 0.001 min
Response: 3796388
Conc: 4.45 ng/ml



#26 AR-1254-1


R.T.: 6.516 min
Delta R.T.: 0.000 min
Response: 3322396
Conc: 51.02 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC050




#26 AR-1254-1

R.T.: 5.708 min
Delta R.T.: 0.000 min
Response: 3474548
Conc: 51.69 ng/ml




#27 AR-1254-2

R.T.: 6.733 min
Delta R.T.: 0.001 min
Response: 5410437
Conc: 53.32 ng/ml



#27 AR-1254-2


R.T.: 5.857 min
Delta R.T.: 0.000 min
Response: 3036515
Conc: 52.17 ng/ml



#28 AR-1254-3


R.T.: 7.096 min
Delta R.T.: 0.001 min
Response: 5249564
Conc: 52.27 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC050




#28 AR-1254-3

R.T.: 6.261 min
Delta R.T.: 0.000 min
Response: 4313044
Conc: 47.83 ng/ml




#29 AR-1254-4

R.T.: 7.378 min
Delta R.T.: 0.001 min
Response: 4268685
Conc: 48.62 ng/ml



#29 AR-1254-4


R.T.: 6.490 min
Delta R.T.: 0.000 min
Response: 2787399
Conc: 47.85 ng/ml



#30 AR-1254-5

R.T.: 7.794 min
Delta R.T.: 0.002 mi
Response: 4173507
Conc: 51.34 ng/m

Instrument :
ECD_P
ClientSampleId :
AR1254ICC050



#30 AR-1254-5

R.T.: 6.907 min
Delta R.T.: 0.000 min
Response: 3458194
Conc: 45.98 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070939.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 16:32
Operator : YP\AJ
Sample : AR1262ICC500
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1262ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 17:04:37 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 17:04:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.512	3.816	78302128	53762587	50.000	50.000
2)	SA Decachlor...	10.226	8.859	57984668	42902548	50.000	50.000
	Target Compounds						
36)	L8 AR-1262-1	8.093	6.945	53150122	42781773	500.000	500.000
37)	L8 AR-1262-2	8.412	7.205	99191436	34232243	500.000	500.000
38)	L8 AR-1262-3	8.729	7.729	65870896	30361758	500.000	500.000
39)	L8 AR-1262-4	8.817	7.793	47807443	49703360	500.000	500.000
40)	L8 AR-1262-5	9.469	8.294	32533206	23435811	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070939.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 16:32
Operator : YP\AJ
Sample : AR1262ICC500
Misc :
ALS Vial : 25 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1262ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 17:04:37 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 17:04:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.512 min
Delta R.T.: 0.000 min
Response: 78302128
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1262ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.816 min
Delta R.T.: 0.000 min
Response: 53762587
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.226 min
Delta R.T.: 0.000 min
Response: 57984668
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.859 min
Delta R.T.: 0.000 min
Response: 42902548
Conc: 50.00 ng/ml



#36 AR-1262-1


R.T.: 8.093 min
Delta R.T.: 0.000 min
Response: 53150122
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1262ICC500




#36 AR-1262-1

R.T.: 6.945 min
Delta R.T.: 0.000 min
Response: 42781773
Conc: 500.00 ng/ml




#37 AR-1262-2

R.T.: 8.412 min
Delta R.T.: 0.000 min
Response: 99191436
Conc: 500.00 ng/ml



#37 AR-1262-2


R.T.: 7.205 min
Delta R.T.: 0.000 min
Response: 34232243
Conc: 500.00 ng/ml



#38 AR-1262-3


R.T.: 8.729 min
Delta R.T.: 0.000 min
Response: 65870896
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1262ICC500




#38 AR-1262-3

R.T.: 7.729 min
Delta R.T.: 0.000 min
Response: 30361758
Conc: 500.00 ng/ml




#39 AR-1262-4

R.T.: 8.817 min
Delta R.T.: 0.000 min
Response: 47807443
Conc: 500.00 ng/ml



#39 AR-1262-4


R.T.: 7.793 min
Delta R.T.: 0.000 min
Response: 49703360
Conc: 500.00 ng/ml



#40 AR-1262-5

R.T.: 9.469 min
Delta R.T.: 0.000 mi
Response: 32533206
Conc: 500.00 ng/m

Instrument :
ECD_P
ClientSampleId :
AR1262ICC500



#40 AR-1262-5

R.T.: 8.294 min
Delta R.T.: 0.000 min
Response: 23435811
Conc: 500.00 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070940.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 16:48
Operator : YP\AJ
Sample : AR1268ICC1000
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC1000

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 17:58:03 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 17:56:10 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.516	3.817	152.3E6	105.1E6	94.769	94.009
2)	SA Decachlor...	10.232	8.860	195.7E6	148.5E6	93.377	94.968
	Target Compounds						
41)	L9 AR-1268-1	8.730	7.729	226.6E6	165.4E6	945.404	946.509m
42)	L9 AR-1268-2	8.824	7.795	190.4E6	137.8E6	950.797	945.407
43)	L9 AR-1268-3	9.056	8.001	163.6E6	120.1E6	940.749	965.766
44)	L9 AR-1268-4	9.474	8.295	70377008	51096977	960.945	957.859
45)	L9 AR-1268-5	9.891	8.596	459.6E6	357.1E6	952.414	962.556

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070940.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 16:48
Operator : YP\AJ
Sample : AR1268ICC1000
Misc :
ALS Vial : 26 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1268ICC1000


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 17:58:03 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 17:56:10 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene

R.T.: 4.516 min
Delta R.T.: 0.003 min
Response: 152348529
Conc: 94.77 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC1000


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025




#1 Tetrachloro-m-xylene

R.T.: 3.817 min
Delta R.T.: 0.000 min
Response: 105128984
Conc: 94.01 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.232 min
Delta R.T.: 0.003 min
Response: 195698435
Conc: 93.38 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.860 min
Delta R.T.: 0.000 min
Response: 148482089
Conc: 94.97 ng/ml




#41 AR-1268-1

R.T.: 8.730 min
Delta R.T.: 0.001 mi
Response: 226575551
Conc: 945.40 ng/m

Instrument :
ECD_P
ClientSampleId :
AR1268ICC1000


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025




#41 AR-1268-1

R.T.: 7.729 min
Delta R.T.: 0.001 min
Response: 165356134
Conc: 946.51 ng/ml m




#42 AR-1268-2

R.T.: 8.824 min
Delta R.T.: 0.003 min
Response: 190380276
Conc: 950.80 ng/ml



#42 AR-1268-2

R.T.: 7.795 min
Delta R.T.: 0.001 min
Response: 137801505
Conc: 945.41 ng/ml




#43 AR-1268-3

R.T.: 9.056 min
Delta R.T.: 0.003 min
Response: 163648333
Conc: 940.75 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC1000


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025




#43 AR-1268-3

R.T.: 8.001 min
Delta R.T.: 0.001 min
Response: 120132746
Conc: 965.77 ng/ml




#44 AR-1268-4

R.T.: 9.474 min
Delta R.T.: 0.002 min
Response: 70377008
Conc: 960.95 ng/ml



#44 AR-1268-4

R.T.: 8.295 min
Delta R.T.: 0.002 min
Response: 51096977
Conc: 957.86 ng/ml




#45 AR-1268-5

R.T.: 9.891 min
Delta R.T.: 0.003 min
Response: 459553484
Conc: 952.41 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC1000

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025



#45 AR-1268-5

R.T.: 8.596 min
Delta R.T.: 0.000 min
Response: 357060145
Conc: 962.56 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070941.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 17:05
Operator : YP\AJ
Sample : AR1268ICC750
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC750

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 17:58:14 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 17:56:10 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.514	3.817	119.6E6	83914927	74.397	75.039
2)	SA Decachlor...	10.230	8.861	154.1E6	118.4E6	73.551	75.751
	Target Compounds						
41)	L9 AR-1268-1	8.730	7.729	178.4E6	131.7E6	744.578	753.609m
42)	L9 AR-1268-2	8.823	7.795	149.2E6	110.8E6	745.140	760.189
43)	L9 AR-1268-3	9.055	8.000	128.8E6	95630424	740.439	768.788
44)	L9 AR-1268-4	9.473	8.294	55249435	40878954	754.389	766.313
45)	L9 AR-1268-5	9.890	8.596	361.2E6	282.2E6	748.577	760.668

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070941.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 17:05
Operator : YP\AJ
Sample : AR1268ICC750
Misc :
ALS Vial : 27 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1268ICC750


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 17:58:14 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 17:56:10 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene

R.T.: 4.514 min
Delta R.T.: 0.001 min
Response: 119597931
Conc: 74.40 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC750


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025




#1 Tetrachloro-m-xylene

R.T.: 3.817 min
Delta R.T.: 0.000 min
Response: 83914927
Conc: 75.04 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.230 min
Delta R.T.: 0.000 min
Response: 154146813
Conc: 73.55 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.861 min
Delta R.T.: 0.000 min
Response: 118436920
Conc: 75.75 ng/ml




#41 AR-1268-1

R.T.: 8.730 min
Delta R.T.: 0.001 mi
Response: 178445590
Conc: 744.58 ng/m

Instrument :
ECD_P
ClientSampleId :
AR1268ICC750


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025




#41 AR-1268-1

R.T.: 7.729 min
Delta R.T.: 0.001 min
Response: 131656350
Conc: 753.61 ng/ml m




#42 AR-1268-2

R.T.: 8.823 min
Delta R.T.: 0.002 min
Response: 149201176
Conc: 745.14 ng/ml



#42 AR-1268-2

R.T.: 7.795 min
Delta R.T.: 0.001 min
Response: 110804324
Conc: 760.19 ng/ml




#43 AR-1268-3

R.T.: 9.055 min
Delta R.T.: 0.002 min
Response: 128803338
Conc: 740.44 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC750


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025




#43 AR-1268-3

R.T.: 8.000 min
Delta R.T.: 0.000 min
Response: 95630424
Conc: 768.79 ng/ml




#44 AR-1268-4

R.T.: 9.473 min
Delta R.T.: 0.001 min
Response: 55249435
Conc: 754.39 ng/ml



#44 AR-1268-4

R.T.: 8.294 min
Delta R.T.: 0.000 min
Response: 40878954
Conc: 766.31 ng/ml




#45 AR-1268-5

R.T.: 9.890 min
Delta R.T.: 0.002 mi
Response: 361198889
Conc: 748.58 ng/m

Instrument :
ECD_P
ClientSampleId :
AR1268ICC750

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025



#45 AR-1268-5

R.T.: 8.596 min
Delta R.T.: 0.000 min
Response: 282169730
Conc: 760.67 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070942.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 17:21
Operator : YP\AJ
Sample : AR1268ICC500
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 17:58:28 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 17:56:10 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.513	3.816	80378727	55914176	50.000	50.000
2)	SA Decachlor...	10.230	8.860	104.8E6	78175161	50.000	50.000
	Target Compounds						
41)	L9 AR-1268-1	8.729	7.728	119.8E6	87376833	500.000	500.150m
42)	L9 AR-1268-2	8.821	7.793	100.1E6	72879434	500.000	500.000
43)	L9 AR-1268-3	9.053	8.000	86977678	62195601	500.000	500.000
44)	L9 AR-1268-4	9.472	8.293	36618642	26672497	500.000	500.000
45)	L9 AR-1268-5	9.888	8.596	241.3E6	185.5E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070942.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 17:21
Operator : YP\AJ
Sample : AR1268ICC500
Misc :
ALS Vial : 28 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1268ICC500


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 17:58:28 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 17:56:10 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene

R.T.: 4.513 min
Delta R.T.: 0.000 min
Response: 80378727
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC500


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025




#1 Tetrachloro-m-xylene

R.T.: 3.816 min
Delta R.T.: 0.000 min
Response: 55914176
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.230 min
Delta R.T.: 0.000 min
Response: 104789095
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.860 min
Delta R.T.: 0.000 min
Response: 78175161
Conc: 50.00 ng/ml




#41 AR-1268-1

R.T.: 8.729 min
Delta R.T.: 0.000 mi
Response: 119830016
Conc: 500.00 ng/m

Instrument :
ECD_P
ClientSampleId :
AR1268ICC500


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025




#41 AR-1268-1

R.T.: 7.728 min
Delta R.T.: 0.000 min
Response: 87376833
Conc: 500.15 ng/ml m




#42 AR-1268-2

R.T.: 8.821 min
Delta R.T.: 0.000 min
Response: 100116173
Conc: 500.00 ng/ml



#42 AR-1268-2

R.T.: 7.793 min
Delta R.T.: 0.000 min
Response: 72879434
Conc: 500.00 ng/ml




#43 AR-1268-3

R.T.: 9.053 min
Delta R.T.: 0.000 min
Response: 86977678
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC500


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025




#43 AR-1268-3

R.T.: 8.000 min
Delta R.T.: 0.000 min
Response: 62195601
Conc: 500.00 ng/ml




#44 AR-1268-4

R.T.: 9.472 min
Delta R.T.: 0.000 min
Response: 36618642
Conc: 500.00 ng/ml



#44 AR-1268-4

R.T.: 8.293 min
Delta R.T.: 0.000 min
Response: 26672497
Conc: 500.00 ng/ml




#45 AR-1268-5

R.T.: 9.888 min
Delta R.T.: 0.000 min
Response: 241257148
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025



#45 AR-1268-5

R.T.: 8.596 min
Delta R.T.: 0.000 min
Response: 185474929
Conc: 500.00 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070943.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 17:37
Operator : YP\AJ
Sample : AR1268ICC250
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 17:58:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 17:56:10 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.515	3.817	41125313	29588552	25.582	26.459
2)	SA Decachlor...	10.232	8.859	54285349	40518249	25.902	25.915
	Target Compounds						
41)	L9 AR-1268-1	8.730	7.729	61973048	44233672	258.587	253.196m
42)	L9 AR-1268-2	8.824	7.794	51237231	36694564	255.889	251.748
43)	L9 AR-1268-3	9.055	8.000	44311644	31811936	254.730	255.741
44)	L9 AR-1268-4	9.473	8.295	18700738	13922493	255.345	260.990
45)	L9 AR-1268-5	9.890	8.596	124.6E6	91636186	258.260	247.031

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070943.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 17:37
Operator : YP\AJ
Sample : AR1268ICC250
Misc :
ALS Vial : 29 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1268ICC250


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 17:58:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 17:56:10 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene

R.T.: 4.515 min
Delta R.T.: 0.002 mi
Response: 41125313
Conc: 25.58 ng/m

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025




#1 Tetrachloro-m-xylene

R.T.: 3.817 min
Delta R.T.: 0.000 min
Response: 29588552
Conc: 26.46 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.232 min
Delta R.T.: 0.003 min
Response: 54285349
Conc: 25.90 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.859 min
Delta R.T.: 0.000 min
Response: 40518249
Conc: 25.92 ng/ml




#41 AR-1268-1

R.T.: 8.730 min
Delta R.T.: 0.002 mi
Response: 61973048
Conc: 258.59 ng/m

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025




#41 AR-1268-1

R.T.: 7.729 min
Delta R.T.: 0.000 min
Response: 44233672
Conc: 253.20 ng/ml m




#42 AR-1268-2

R.T.: 8.824 min
Delta R.T.: 0.002 min
Response: 51237231
Conc: 255.89 ng/ml



#42 AR-1268-2

R.T.: 7.794 min
Delta R.T.: 0.000 min
Response: 36694564
Conc: 251.75 ng/ml




#43 AR-1268-3

R.T.: 9.055 min
Delta R.T.: 0.002 mi
Response: 44311644
Conc: 254.73 ng/m

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025




#43 AR-1268-3

R.T.: 8.000 min
Delta R.T.: 0.000 min
Response: 31811936
Conc: 255.74 ng/ml




#44 AR-1268-4

R.T.: 9.473 min
Delta R.T.: 0.002 min
Response: 18700738
Conc: 255.34 ng/ml



#44 AR-1268-4

R.T.: 8.295 min
Delta R.T.: 0.002 min
Response: 13922493
Conc: 260.99 ng/ml




#45 AR-1268-5

R.T.: 9.890 min
Delta R.T.: 0.002 min
Response: 124614186
Conc: 258.26 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025



#45 AR-1268-5

R.T.: 8.596 min
Delta R.T.: 0.000 min
Response: 91636186
Conc: 247.03 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070944.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 17:53
Operator : YP\AJ
Sample : AR1268ICC050
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 18:13:54 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 17:56:10 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.514	3.816	7200607	6304440	4.457m	5.529
2)	SA Decachlor...	10.230	8.859	10035177	7161586	4.762	4.510
	Target Compounds						
41)	L9 AR-1268-1	8.729	7.728	11274140	8221191	46.621	46.780m
42)	L9 AR-1268-2	8.822	7.794	9103995	6916833	45.210	47.131
43)	L9 AR-1268-3	9.054	7.999	8398993	5823381	48.183	46.078
44)	L9 AR-1268-4	9.471	8.293	3082639	2471433	41.712	45.336m
45)	L9 AR-1268-5	9.889	8.596	22937830	16928167	47.050	45.599

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070944.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 17:53
Operator : YP\AJ
Sample : AR1268ICC050
Misc :
ALS Vial : 30 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1268ICC050


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 18:13:54 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 17:56:10 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene

R.T.: 4.514 min
Delta R.T.: 0.000 min
Response: 7200607
Conc: 4.46 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025




#1 Tetrachloro-m-xylene

R.T.: 3.816 min
Delta R.T.: 0.000 min
Response: 6304440
Conc: 5.53 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.230 min
Delta R.T.: 0.000 min
Response: 10035177
Conc: 4.76 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.859 min
Delta R.T.: -0.001 min
Response: 7161586
Conc: 4.51 ng/ml




#41 AR-1268-1

R.T.: 8.729 min
Delta R.T.: 0.000 mi
Response: 11274140
Conc: 46.62 ng/m

Instrument :
ECD_P
ClientSampleId :
AR1268ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025




#41 AR-1268-1

R.T.: 7.728 min
Delta R.T.: 0.000 min
Response: 8221191
Conc: 46.78 ng/ml m




#42 AR-1268-2

R.T.: 8.822 min
Delta R.T.: 0.000 min
Response: 9103995
Conc: 45.21 ng/ml



#42 AR-1268-2

R.T.: 7.794 min
Delta R.T.: 0.000 min
Response: 6916833
Conc: 47.13 ng/ml




#43 AR-1268-3

R.T.: 9.054 min
Delta R.T.: 0.002 mi
Response: 8398993
Conc: 48.18 ng/m

Instrument :
ECD_P
ClientSampleId :
AR1268ICC050


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025




#43 AR-1268-3

R.T.: 7.999 min
Delta R.T.: 0.000 min
Response: 5823381
Conc: 46.08 ng/ml




#44 AR-1268-4

R.T.: 9.471 min
Delta R.T.: 0.000 min
Response: 3082639
Conc: 41.71 ng/ml



#44 AR-1268-4

R.T.: 8.293 min
Delta R.T.: 0.000 min
Response: 2471433
Conc: 45.34 ng/ml m




#45 AR-1268-5

R.T.: 9.889 min
Delta R.T.: 0.000 mi
Response: 22937830
Conc: 47.05 ng/m

Instrument :
ECD_P
ClientSampleId :
AR1268ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025



#45 AR-1268-5

R.T.: 8.596 min
Delta R.T.: 0.000 min
Response: 16928167
Conc: 45.60 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070945.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 18:26
Operator : YP\AJ
Sample : PP032725ICV500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP032725

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 18:44:38 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 18:19:30 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1) SA	Tetrachlo...	4.513	3.817	83053328	55050464	56.825	53.105
2) SA	Decachlor...	10.229	8.860	60452615	44643396	57.431	56.640
	Target Compounds						
3) L1	AR-1016-1	5.666	4.904	27342824	21140954	537.702	527.361
4) L1	AR-1016-2	5.688	4.923	41229999	30414011	528.227	525.602
5) L1	AR-1016-3	5.750	5.100	25160452	16688586	533.347	539.677
6) L1	AR-1016-4	5.847	5.141	20613627	13786307	560.828	566.035
7) L1	AR-1016-5	6.140	5.356	19310102	17090774	569.416	528.779
31) L7	AR-1260-1	7.259	6.392	37489891	28161151	558.277	549.955
32) L7	AR-1260-2	7.512	6.580	54825330	35718929	570.287	559.912
33) L7	AR-1260-3	7.871	6.734	45118892	30750809	564.563	564.371
34) L7	AR-1260-4	8.095	7.206	41972897	25791995	533.078	539.285
35) L7	AR-1260-5	8.414	7.447	87345749	64063663	549.738	514.525


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070945.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 18:26
Operator : YP\AJ
Sample : PP032725ICV500
Misc :
ALS Vial : 3 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
ICVPP032725

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 18:44:38 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 18:19:30 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.513 min
Delta R.T.: 0.001 min
Response: 83053328
Conc: 56.82 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP032725




#1 Tetrachloro-m-xylene

R.T.: 3.817 min
Delta R.T.: 0.000 min
Response: 55050464
Conc: 53.10 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.229 min
Delta R.T.: 0.003 min
Response: 60452615
Conc: 57.43 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.860 min
Delta R.T.: 0.000 min
Response: 44643396
Conc: 56.64 ng/ml



#3 AR-1016-1


R.T.: 5.666 min
Delta R.T.: 0.002 min
Response: 27342824
Conc: 537.70 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP032725




#3 AR-1016-1

R.T.: 4.904 min
Delta R.T.: 0.001 min
Response: 21140954
Conc: 527.36 ng/ml




#4 AR-1016-2

R.T.: 5.688 min
Delta R.T.: 0.003 min
Response: 41229999
Conc: 528.23 ng/ml



#4 AR-1016-2


R.T.: 4.923 min
Delta R.T.: 0.001 min
Response: 30414011
Conc: 525.60 ng/ml



#5 AR-1016-3


R.T.: 5.750 min
Delta R.T.: 0.003 min
Response: 25160452
Conc: 533.35 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP032725




#5 AR-1016-3

R.T.: 5.100 min
Delta R.T.: 0.000 min
Response: 16688586
Conc: 539.68 ng/ml




#6 AR-1016-4

R.T.: 5.847 min
Delta R.T.: 0.002 min
Response: 20613627
Conc: 560.83 ng/ml



#6 AR-1016-4


R.T.: 5.141 min
Delta R.T.: 0.000 min
Response: 13786307
Conc: 566.04 ng/ml



#7 AR-1016-5


R.T.: 6.140 min
Delta R.T.: 0.002 min
Response: 19310102
Conc: 569.42 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP032725




#7 AR-1016-5

R.T.: 5.356 min
Delta R.T.: 0.000 min
Response: 17090774
Conc: 528.78 ng/ml




#31 AR-1260-1

R.T.: 7.259 min
Delta R.T.: 0.003 min
Response: 37489891
Conc: 558.28 ng/ml



#31 AR-1260-1


R.T.: 6.392 min
Delta R.T.: 0.000 min
Response: 28161151
Conc: 549.95 ng/ml



#32 AR-1260-2


R.T.: 7.512 min
Delta R.T.: 0.002 min
Response: 54825330
Conc: 570.29 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP032725




#32 AR-1260-2

R.T.: 6.580 min
Delta R.T.: 0.000 min
Response: 35718929
Conc: 559.91 ng/ml




#33 AR-1260-3

R.T.: 7.871 min
Delta R.T.: 0.003 min
Response: 45118892
Conc: 564.56 ng/ml



#33 AR-1260-3


R.T.: 6.734 min
Delta R.T.: 0.000 min
Response: 30750809
Conc: 564.37 ng/ml



#34 AR-1260-4


R.T.: 8.095 min
Delta R.T.: 0.003 min
Response: 41972897
Conc: 533.08 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP032725




#34 AR-1260-4

R.T.: 7.206 min
Delta R.T.: 0.000 min
Response: 25791995
Conc: 539.28 ng/ml



#35 AR-1260-5

R.T.: 8.414 min
Delta R.T.: 0.003 min
Response: 87345749
Conc: 549.74 ng/ml



#35 AR-1260-5

R.T.: 7.447 min
Delta R.T.: 0.000 min
Response: 64063663
Conc: 514.53 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070946.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 18:42
Operator : YP\AJ
Sample : AR1242ICV500
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP032725AR1242

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 19:00:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 18:58:04 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.515	3.816	78786041	55016883	53.357	51.300
2)	SA Decachlor...	10.230	8.859	57489647	39612860	53.073	51.065
	Target Compounds						
16)	L4 AR-1242-1	5.667	4.904	21998453	17657296	500.225	516.358
17)	L4 AR-1242-2	5.689	4.922	33746674	24829266	524.698	504.069
18)	L4 AR-1242-3	5.752	5.099	20359598	13889595	496.098	521.298
19)	L4 AR-1242-4	5.849	5.183	16580558	13666228	543.209	518.087
20)	L4 AR-1242-5	6.580	5.708	19203834	15992906	525.982	516.595

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070946.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 18:42
Operator : YP\AJ
Sample : AR1242ICV500
Misc :
ALS Vial : 32 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
ICVPP032725AR1242

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 19:00:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 18:58:04 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.515 min
Delta R.T.: 0.004 min
Response: 78786041
Conc: 53.36 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP032725AR1242




#1 Tetrachloro-m-xylene

R.T.: 3.816 min
Delta R.T.: 0.000 min
Response: 55016883
Conc: 51.30 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.230 min
Delta R.T.: 0.005 min
Response: 57489647
Conc: 53.07 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.859 min
Delta R.T.: 0.001 min
Response: 39612860
Conc: 51.07 ng/ml



#16 AR-1242-1


R.T.: 5.667 min
Delta R.T.: 0.003 mi
Response: 21998453
Conc: 500.22 ng/m

Instrument :
ECD_P
ClientSampleId :
ICVPP032725AR1242




#16 AR-1242-1

R.T.: 4.904 min
Delta R.T.: 0.001 min
Response: 17657296
Conc: 516.36 ng/ml




#17 AR-1242-2

R.T.: 5.689 min
Delta R.T.: 0.004 min
Response: 33746674
Conc: 524.70 ng/ml



#17 AR-1242-2


R.T.: 4.922 min
Delta R.T.: 0.000 min
Response: 24829266
Conc: 504.07 ng/ml



#18 AR-1242-3


R.T.: 5.752 min
Delta R.T.: 0.004 min
Response: 20359598
Conc: 496.10 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP032725AR1242




#18 AR-1242-3

R.T.: 5.099 min
Delta R.T.: 0.000 min
Response: 13889595
Conc: 521.30 ng/ml




#19 AR-1242-4

R.T.: 5.849 min
Delta R.T.: 0.004 min
Response: 16580558
Conc: 543.21 ng/ml



#19 AR-1242-4


R.T.: 5.183 min
Delta R.T.: 0.000 min
Response: 13666228
Conc: 518.09 ng/ml



#20 AR-1242-5

R.T.: 6.580 min
Delta R.T.: 0.004 min
Response: 19203834
Conc: 525.98 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP032725AR1242



#20 AR-1242-5

R.T.: 5.708 min
Delta R.T.: 0.000 min
Response: 15992906
Conc: 516.60 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070947.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 19:15
Operator : YP\AJ
Sample : AR1248ICV500
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP032725AR1248

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 19:34:19 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 19:07:15 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.514	3.817	80868018	57291573	53.345	52.976
2)	SA Decachlor...	10.229	8.860	60625729	42784323	53.381	54.139
	Target Compounds						
21)	L5 AR-1248-1	5.666	4.904	17556753	15279017	483.824	564.218
22)	L5 AR-1248-2	5.937	5.141	24046455	20552003	535.096	559.658
23)	L5 AR-1248-3	6.140	5.183	26863901	21355068	530.514	552.390
24)	L5 AR-1248-4	6.539	5.356	33592146	24182903	523.670	535.719
25)	L5 AR-1248-5	6.578	5.750	32500443	22380961	534.140	524.405

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070947.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 19:15
Operator : YP\AJ
Sample : AR1248ICV500
Misc :
ALS Vial : 33 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
ICVPP032725AR1248

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 19:34:19 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 19:07:15 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.514 min
Delta R.T.: 0.000 min
Response: 80868018
Conc: 53.34 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP032725AR1248




#1 Tetrachloro-m-xylene

R.T.: 3.817 min
Delta R.T.: 0.001 min
Response: 57291573
Conc: 52.98 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.229 min
Delta R.T.: 0.002 min
Response: 60625729
Conc: 53.38 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.860 min
Delta R.T.: 0.001 min
Response: 42784323
Conc: 54.14 ng/ml



#21 AR-1248-1


R.T.: 5.666 min
Delta R.T.: 0.000 min
Response: 17556753
Conc: 483.82 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP032725AR1248




#21 AR-1248-1

R.T.: 4.904 min
Delta R.T.: 0.000 min
Response: 15279017
Conc: 564.22 ng/ml




#22 AR-1248-2

R.T.: 5.937 min
Delta R.T.: 0.000 min
Response: 24046455
Conc: 535.10 ng/ml



#22 AR-1248-2


R.T.: 5.141 min
Delta R.T.: 0.001 min
Response: 20552003
Conc: 559.66 ng/ml



#23 AR-1248-3


R.T.: 6.140 min
Delta R.T.: 0.000 mi
Response: 26863901
Conc: 530.51 ng/m

Instrument :
ECD_P
ClientSampleId :
ICVPP032725AR1248




#23 AR-1248-3

R.T.: 5.183 min
Delta R.T.: 0.000 min
Response: 21355068
Conc: 552.39 ng/ml




#24 AR-1248-4

R.T.: 6.539 min
Delta R.T.: 0.000 min
Response: 33592146
Conc: 523.67 ng/ml



#24 AR-1248-4


R.T.: 5.356 min
Delta R.T.: 0.000 min
Response: 24182903
Conc: 535.72 ng/ml



#25 AR-1248-5

R.T.: 6.578 min
Delta R.T.: 0.000 mi
Response: 32500443
Conc: 534.14 ng/m

Instrument :
ECD_P
ClientSampleId :
ICVPP032725AR1248



#25 AR-1248-5

R.T.: 5.750 min
Delta R.T.: 0.002 min
Response: 22380961
Conc: 524.41 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070948.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 19:47
Operator : YP\AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP032725AR1254

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 28 01:39:10 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 28 01:37:11 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.513	3.816	80813408	55809150	53.470	50.687
2)	SA Decachlor...	10.228	8.859	60442284	44113864	53.599	53.797
	Target Compounds						
26)	L6 AR-1254-1	6.514	5.708	34026950	33372474	528.449	505.507
27)	L6 AR-1254-2	6.730	5.856	52171420	28060451	516.086	491.442
28)	L6 AR-1254-3	7.093	6.260	52059367	42608413	522.014	490.647
29)	L6 AR-1254-4	7.375	6.489	45346150	27407702	523.647	488.547
30)	L6 AR-1254-5	7.790	6.907	41706053	36336253	519.049m	503.617

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070948.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 19:47
Operator : YP\AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 34 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
ICVPP032725AR1254


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 28 01:39:10 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 28 01:37:11 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene

R.T.: 4.513 min
Delta R.T.: -0.001 min
Response: 80813408
Conc: 53.47 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP032725AR1254


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025




#1 Tetrachloro-m-xylene

R.T.: 3.816 min
Delta R.T.: 0.000 min
Response: 55809150
Conc: 50.69 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.228 min
Delta R.T.: -0.002 min
Response: 60442284
Conc: 53.60 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.859 min
Delta R.T.: 0.000 min
Response: 44113864
Conc: 53.80 ng/ml




#26 AR-1254-1

R.T.: 6.514 min
Delta R.T.: -0.001 min
Response: 34026950
Conc: 528.45 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP032725AR1254


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025




#26 AR-1254-1

R.T.: 5.708 min
Delta R.T.: 0.000 min
Response: 33372474
Conc: 505.51 ng/ml




#27 AR-1254-2

R.T.: 6.730 min
Delta R.T.: -0.002 min
Response: 52171420
Conc: 516.09 ng/ml



#27 AR-1254-2

R.T.: 5.856 min
Delta R.T.: 0.000 min
Response: 28060451
Conc: 491.44 ng/ml




#28 AR-1254-3

R.T.: 7.093 min
Delta R.T.: -0.001 min
Response: 52059367
Conc: 522.01 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP032725AR1254


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025




#28 AR-1254-3

R.T.: 6.260 min
Delta R.T.: 0.000 min
Response: 42608413
Conc: 490.65 ng/ml




#29 AR-1254-4

R.T.: 7.375 min
Delta R.T.: -0.001 min
Response: 45346150
Conc: 523.65 ng/ml



#29 AR-1254-4

R.T.: 6.489 min
Delta R.T.: 0.000 min
Response: 27407702
Conc: 488.55 ng/ml




#30 AR-1254-5

R.T.: 7.790 min
Delta R.T.: -0.003 min
Response: 41706053
Conc: 519.05 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP032725AR1254

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/28/2025
Supervised By :mohammad ahmed 03/31/2025



#30 AR-1254-5

R.T.: 6.907 min
Delta R.T.: 0.000 min
Response: 36336253
Conc: 503.62 ng/ml

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070949.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 20:20
Operator : YP\AJ
Sample : AR1268ICV500
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP032725AR1268

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 28 01:48:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 28 01:43:54 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.516	3.816	81260121	54086649	52.018	47.172
2)	SA Decachlor...	10.231	8.858	106.7E6	75635416	51.843	49.238
	Target Compounds						
41)	L9 AR-1268-1	8.729	7.728	122.6E6	85915063	520.500	501.249
42)	L9 AR-1268-2	8.822	7.793	102.3E6	71467631	523.804	498.787
43)	L9 AR-1268-3	9.054	7.999	88458310	61199335	517.562	496.952
44)	L9 AR-1268-4	9.472	8.292	37948087	26301297	536.390	498.006
45)	L9 AR-1268-5	9.889	8.595	246.0E6	175.8E6	516.662	485.835

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070949.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 20:20
Operator : YP\AJ
Sample : AR1268ICV500
Misc :
ALS Vial : 35 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
ICVPP032725AR1268

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 28 01:48:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 28 01:43:54 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.516 min
Delta R.T.: 0.003 min
Response: 81260121
Conc: 52.02 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP032725AR1268




#1 Tetrachloro-m-xylene

R.T.: 3.816 min
Delta R.T.: 0.000 min
Response: 54086649
Conc: 47.17 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.231 min
Delta R.T.: 0.000 min
Response: 106656431
Conc: 51.84 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.858 min
Delta R.T.: -0.002 min
Response: 75635416
Conc: 49.24 ng/ml



#41 AR-1268-1


R.T.: 8.729 min
Delta R.T.: 0.000 min
Response: 122581794
Conc: 520.50 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP032725AR1268




#41 AR-1268-1

R.T.: 7.728 min
Delta R.T.: 0.000 min
Response: 85915063
Conc: 501.25 ng/ml




#42 AR-1268-2

R.T.: 8.822 min
Delta R.T.: 0.000 min
Response: 102307003
Conc: 523.80 ng/ml



#42 AR-1268-2


R.T.: 7.793 min
Delta R.T.: 0.000 min
Response: 71467631
Conc: 498.79 ng/ml



#43 AR-1268-3


R.T.: 9.054 min
Delta R.T.: 0.000 mi
Response: 88458310
Conc: 517.56 ng/m

Instrument :
ECD_P
ClientSampleId :
ICVPP032725AR1268




#43 AR-1268-3

R.T.: 7.999 min
Delta R.T.: -0.001 min
Response: 61199335
Conc: 496.95 ng/ml




#44 AR-1268-4

R.T.: 9.472 min
Delta R.T.: 0.000 min
Response: 37948087
Conc: 536.39 ng/ml



#44 AR-1268-4


R.T.: 8.292 min
Delta R.T.: 0.000 min
Response: 26301297
Conc: 498.01 ng/ml



#45 AR-1268-5

R.T.: 9.889 min
Delta R.T.: 0.000 mi
Response: 246021928
Conc: 516.66 ng/m

Instrument :
ECD_P
ClientSampleId :
ICVPP032725AR1268



#45 AR-1268-5

R.T.: 8.595 min
Delta R.T.: -0.001 min
Response: 175808654
Conc: 485.84 ng/ml

Raw Data: **PO110263.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM **Case No.:** Q1730 **SAS No.:** Q1730 **SDG NO.:** Q1730

Continuing Calib Date: 04/07/2025 **Initial Calibration Date(s):** 03/18/2025 03/18/2025

Continuing Calib Time: 09:46 **Initial Calibration Time(s):** 14:03 22:15

GC Column: ZB-MR1 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	RT WINDOW TO	DIFF RT
Aroclor-1016-1	(1)	4.78	4.79	4.69	4.89	0.01
Aroclor-1016-2	(2)	4.80	4.80	4.70	4.90	0.00
Aroclor-1016-3	(3)	4.86	4.86	4.76	4.96	0.00
Aroclor-1016-4	(4)	4.98	4.98	4.88	5.08	0.00
Aroclor-1016-5	(5)	5.24	5.24	5.14	5.34	0.00
Aroclor-1260-1	(1)	6.28	6.28	6.18	6.38	0.00
Aroclor-1260-2	(2)	6.47	6.47	6.37	6.57	0.00
Aroclor-1260-3	(3)	6.84	6.84	6.74	6.94	0.00
Aroclor-1260-4	(4)	7.10	7.10	7.00	7.20	0.00
Aroclor-1260-5	(5)	7.34	7.34	7.24	7.44	0.00
Tetrachloro-m-xylene		3.69	3.69	3.59	3.79	0.00
Decachlorobiphenyl		8.74	8.75	8.65	8.85	0.01

Raw Data: **PO110263.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM **Case No.:** Q1730 **SAS No.:** Q1730 **SDG NO.:** Q1730

Continuing Calib Date: 04/07/2025 **Initial Calibration Date(s):** 03/18/2025 03/18/2025

Continuing Calib Time: 09:46 **Initial Calibration Time(s):** 14:03 22:15

GC Column: ZB-MR2 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	RT WINDOW TO	DIFF RT
Aroclor-1016-1	(1)	4.77	4.77	4.67	4.87	0.00
Aroclor-1016-2	(2)	4.79	4.79	4.69	4.89	0.00
Aroclor-1016-3	(3)	4.96	4.97	4.87	5.07	0.01
Aroclor-1016-4	(4)	5.01	5.01	4.91	5.11	0.00
Aroclor-1016-5	(5)	5.22	5.22	5.12	5.32	0.00
Aroclor-1260-1	(1)	6.25	6.26	6.16	6.36	0.01
Aroclor-1260-2	(2)	6.44	6.44	6.34	6.54	0.00
Aroclor-1260-3	(3)	6.59	6.60	6.50	6.70	0.01
Aroclor-1260-4	(4)	7.06	7.07	6.97	7.17	0.01
Aroclor-1260-5	(5)	7.30	7.31	7.21	7.41	0.01
Tetrachloro-m-xylene		3.69	3.69	3.59	3.79	0.00
Decachlorobiphenyl		8.69	8.70	8.60	8.80	0.01

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM **Case No.:** Q1730 **SAS No.:** Q1730 **SDG NO.:** Q1730

GC Column: ZB-MR1 **ID:** 0.32 **(mm)** **Initi. Calib. Date(s):** 03/18/2025 03/18/2025

Client Sample No.: CCAL01 **Date Analyzed:** 04/07/2025

Lab Sample No.: AR1660CCC500 **Data File :** PO110263.D **Time Analyzed:** 09:46

COMPOUND	RT	RT WINDOW FROM	RT WINDOW TO	CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.784	4.685	4.885	533.050	500.000	6.6
Aroclor-1016-2	4.803	4.704	4.904	532.950	500.000	6.6
Aroclor-1016-3	4.860	4.761	4.961	534.220	500.000	6.8
Aroclor-1016-4	4.980	4.881	5.081	533.580	500.000	6.7
Aroclor-1016-5	5.238	5.139	5.339	534.560	500.000	6.9
Aroclor-1260-1	6.279	6.180	6.380	540.840	500.000	8.2
Aroclor-1260-2	6.467	6.369	6.569	507.690	500.000	1.5
Aroclor-1260-3	6.835	6.737	6.937	498.120	500.000	-0.4
Aroclor-1260-4	7.096	6.998	7.198	486.870	500.000	-2.6
Aroclor-1260-5	7.337	7.239	7.439	480.250	500.000	-4.0
Decachlorobiphenyl	8.742	8.645	8.845	57.530	50.000	15.1
Tetrachloro-m-xylene	3.692	3.592	3.792	53.720	50.000	7.4

Alliance TECHNICAL GROUP

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM **Case No.:** Q1730 **SAS No.:** Q1730 **SDG NO.:** Q1730

GC Column: ZB-MR2 **ID:** 0.32 **(mm)** **Initi. Calib. Date(s):** 03/18/2025 03/18/2025

Client Sample No.: CCAL01 **Date Analyzed:** 04/07/2025

Lab Sample No.: AR1660CCC500 **Data File :** PO110263.D **Time Analyzed:** 09:46

COMPOUND	RT	RT WINDOW FROM	RT WINDOW TO	CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.769	4.672	4.872	509.410	500.000	1.9
Aroclor-1016-2	4.788	4.691	4.891	506.900	500.000	1.4
Aroclor-1016-3	4.964	4.867	5.067	511.350	500.000	2.3
Aroclor-1016-4	5.005	4.909	5.109	504.600	500.000	0.9
Aroclor-1016-5	5.219	5.122	5.322	510.930	500.000	2.2
Aroclor-1260-1	6.250	6.155	6.355	495.180	500.000	-1.0
Aroclor-1260-2	6.438	6.342	6.542	487.390	500.000	-2.5
Aroclor-1260-3	6.591	6.495	6.695	473.050	500.000	-5.4
Aroclor-1260-4	7.062	6.967	7.167	467.970	500.000	-6.4
Aroclor-1260-5	7.302	7.207	7.407	455.220	500.000	-9.0
Decachlorobiphenyl	8.690	8.596	8.796	48.910	50.000	-2.2
Tetrachloro-m-xylene	3.687	3.590	3.790	52.660	50.000	5.3

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040725\
Data File : PO110263.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Apr 2025 09:46
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 07 12:29:19 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.692	3.687	489.0E6	276.2E6	53.724	52.660
2)	SA Decachlor...	8.742	8.690	442.3E6	118.7E6	57.530	48.907
	Target Compounds						
3)	L1 AR-1016-1	4.784	4.769	179.4E6	93351070	533.053	509.405
4)	L1 AR-1016-2	4.803	4.788	247.1E6	132.6E6	532.949	506.901
5)	L1 AR-1016-3	4.860	4.964	171.3E6	72498019	534.221	511.346
6)	L1 AR-1016-4	4.980	5.005	134.8E6	59506839	533.582	504.596
7)	L1 AR-1016-5	5.238	5.219	146.9E6	78129978	534.558	510.925
31)	L7 AR-1260-1	6.279	6.250	254.9E6	126.5E6	540.836	495.179
32)	L7 AR-1260-2	6.467	6.438	302.3E6	146.4E6	507.691	487.393
33)	L7 AR-1260-3	6.835	6.591	241.7E6	136.7E6	498.121	473.054
34)	L7 AR-1260-4	7.096	7.062	206.5E6	100.7E6	486.872	467.969
35)	L7 AR-1260-5	7.337	7.302	502.3E6	227.0E6	480.254	455.222

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Sample Results: **PO110263.D**


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040725\
Data File : PO110263.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Apr 2025 09:46
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 07 12:29:19 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.692 min
Delta R.T.: 0.000 min
Response: 488973455
Conc: 53.72 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.687 min
Delta R.T.: -0.002 min
Response: 276154148
Conc: 52.66 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.742 min
Delta R.T.: -0.004 min
Response: 442329479
Conc: 57.53 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.690 min
Delta R.T.: -0.006 min
Response: 118652560
Conc: 48.91 ng/ml



#3 AR-1016-1


R.T.: 4.784 min
Delta R.T.: -0.001 min
Response: 179395675
Conc: 533.05 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.769 min
Delta R.T.: -0.003 min
Response: 93351070
Conc: 509.41 ng/ml




#4 AR-1016-2

R.T.: 4.803 min
Delta R.T.: -0.001 min
Response: 247120791
Conc: 532.95 ng/ml



#4 AR-1016-2


R.T.: 4.788 min
Delta R.T.: -0.003 min
Response: 132556824
Conc: 506.90 ng/ml



#5 AR-1016-3


R.T.: 4.860 min
Delta R.T.: -0.001 min
Response: 171263953
Conc: 534.22 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.964 min
Delta R.T.: -0.003 min
Response: 72498019
Conc: 511.35 ng/ml




#6 AR-1016-4

R.T.: 4.980 min
Delta R.T.: -0.001 min
Response: 134830395
Conc: 533.58 ng/ml



#6 AR-1016-4


R.T.: 5.005 min
Delta R.T.: -0.004 min
Response: 59506839
Conc: 504.60 ng/ml



#7 AR-1016-5


R.T.: 5.238 min
Delta R.T.: -0.001 min
Response: 146932389
Conc: 534.56 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.219 min
Delta R.T.: -0.004 min
Response: 78129978
Conc: 510.93 ng/ml




#31 AR-1260-1

R.T.: 6.279 min
Delta R.T.: 0.000 min
Response: 254923351
Conc: 540.84 ng/ml



#31 AR-1260-1


R.T.: 6.250 min
Delta R.T.: -0.005 min
Response: 126489343
Conc: 495.18 ng/ml



#32 AR-1260-2


R.T.: 6.467 min
Delta R.T.: -0.002 min
Response: 302339417
Conc: 507.69 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.438 min
Delta R.T.: -0.005 min
Response: 146357167
Conc: 487.39 ng/ml




#33 AR-1260-3

R.T.: 6.835 min
Delta R.T.: -0.002 min
Response: 241738750
Conc: 498.12 ng/ml



#33 AR-1260-3


R.T.: 6.591 min
Delta R.T.: -0.004 min
Response: 136704083
Conc: 473.05 ng/ml



#34 AR-1260-4


R.T.: 7.096 min
Delta R.T.: -0.002 min
Response: 206515387
Conc: 486.87 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.062 min
Delta R.T.: -0.005 min
Response: 100650011
Conc: 467.97 ng/ml



#35 AR-1260-5

R.T.: 7.337 min
Delta R.T.: -0.003 min
Response: 502271562
Conc: 480.25 ng/ml



#35 AR-1260-5

R.T.: 7.302 min
Delta R.T.: -0.005 min
Response: 227017361
Conc: 455.22 ng/ml

Raw Data: **PO110278.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM **Case No.:** Q1730 **SAS No.:** Q1730 **SDG NO.:** Q1730

Continuing Calib Date: 04/07/2025 **Initial Calibration Date(s):** 03/18/2025 03/18/2025

Continuing Calib Time: 17:14 **Initial Calibration Time(s):** 14:03 22:15

GC Column: ZB-MR1 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	RT WINDOW TO	DIFF RT
Aroclor-1016-1	(1)	4.78	4.79	4.69	4.89	0.01
Aroclor-1016-2	(2)	4.80	4.80	4.70	4.90	0.00
Aroclor-1016-3	(3)	4.86	4.86	4.76	4.96	0.00
Aroclor-1016-4	(4)	4.98	4.98	4.88	5.08	0.00
Aroclor-1016-5	(5)	5.24	5.24	5.14	5.34	0.00
Aroclor-1260-1	(1)	6.28	6.28	6.18	6.38	0.00
Aroclor-1260-2	(2)	6.47	6.47	6.37	6.57	0.00
Aroclor-1260-3	(3)	6.83	6.84	6.74	6.94	0.01
Aroclor-1260-4	(4)	7.09	7.10	7.00	7.20	0.01
Aroclor-1260-5	(5)	7.34	7.34	7.24	7.44	0.00
Tetrachloro-m-xylene		3.69	3.69	3.59	3.79	0.00
Decachlorobiphenyl		8.74	8.75	8.65	8.85	0.01

Raw Data: **PO110278.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM **Case No.:** Q1730 **SAS No.:** Q1730 **SDG NO.:** Q1730

Continuing Calib Date: 04/07/2025 **Initial Calibration Date(s):** 03/18/2025 03/18/2025

Continuing Calib Time: 17:14 **Initial Calibration Time(s):** 14:03 22:15

GC Column: ZB-MR2 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	TO	DIFF RT
Aroclor-1016-1	(1)	4.77	4.77	4.67	4.87	0.00
Aroclor-1016-2	(2)	4.79	4.79	4.69	4.89	0.00
Aroclor-1016-3	(3)	4.96	4.97	4.87	5.07	0.01
Aroclor-1016-4	(4)	5.01	5.01	4.91	5.11	0.00
Aroclor-1016-5	(5)	5.22	5.22	5.12	5.32	0.00
Aroclor-1260-1	(1)	6.25	6.26	6.16	6.36	0.01
Aroclor-1260-2	(2)	6.44	6.44	6.34	6.54	0.00
Aroclor-1260-3	(3)	6.59	6.60	6.50	6.70	0.01
Aroclor-1260-4	(4)	7.06	7.07	6.97	7.17	0.01
Aroclor-1260-5	(5)	7.30	7.31	7.21	7.41	0.01
Tetrachloro-m-xylene		3.69	3.69	3.59	3.79	0.00
Decachlorobiphenyl		8.69	8.70	8.60	8.80	0.01

Alliance TECHNICAL GROUP

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM **Case No.:** Q1730 **SAS No.:** Q1730 **SDG NO.:** Q1730

GC Column: ZB-MR1 **ID:** 0.32 **(mm)** **Initi. Calib. Date(s):** 03/18/2025 03/18/2025

Client Sample No.: CCAL02 **Date Analyzed:** 04/07/2025

Lab Sample No.: AR1660CCC500 **Data File :** PO110278.D **Time Analyzed:** 17:14

COMPOUND	RT	RT WINDOW FROM	RT WINDOW TO	CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.782	4.685	4.885	511.900	500.000	2.4
Aroclor-1016-2	4.802	4.704	4.904	513.800	500.000	2.8
Aroclor-1016-3	4.858	4.761	4.961	513.560	500.000	2.7
Aroclor-1016-4	4.979	4.881	5.081	505.500	500.000	1.1
Aroclor-1016-5	5.236	5.139	5.339	483.610	500.000	-3.3
Aroclor-1260-1	6.276	6.180	6.380	516.970	500.000	3.4
Aroclor-1260-2	6.466	6.369	6.569	486.490	500.000	-2.7
Aroclor-1260-3	6.834	6.737	6.937	450.780	500.000	-9.8
Aroclor-1260-4	7.094	6.998	7.198	442.100	500.000	-11.6
Aroclor-1260-5	7.336	7.239	7.439	429.820	500.000	-14.0
Decachlorobiphenyl	8.741	8.645	8.845	52.440	50.000	4.9
Tetrachloro-m-xylene	3.691	3.592	3.792	52.070	50.000	4.1

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM **Case No.:** Q1730 **SAS No.:** Q1730 **SDG NO.:** Q1730

GC Column: ZB-MR2 **ID:** 0.32 **(mm)** **Initi. Calib. Date(s):** 03/18/2025 03/18/2025

Client Sample No.: CCAL02 **Date Analyzed:** 04/07/2025

Lab Sample No.: AR1660CCC500 **Data File :** PO110278.D **Time Analyzed:** 17:14

COMPOUND	RT	RT WINDOW FROM	RT WINDOW TO	CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.769	4.672	4.872	501.230	500.000	0.2
Aroclor-1016-2	4.788	4.691	4.891	505.970	500.000	1.2
Aroclor-1016-3	4.964	4.867	5.067	506.650	500.000	1.3
Aroclor-1016-4	5.006	4.909	5.109	501.000	500.000	0.2
Aroclor-1016-5	5.218	5.122	5.322	500.710	500.000	0.1
Aroclor-1260-1	6.250	6.155	6.355	459.750	500.000	-8.1
Aroclor-1260-2	6.438	6.342	6.542	450.220	500.000	-10.0
Aroclor-1260-3	6.591	6.495	6.695	435.550	500.000	-12.9
Aroclor-1260-4	7.063	6.967	7.167	432.750	500.000	-13.5
Aroclor-1260-5	7.303	7.207	7.407	419.150	500.000	-16.2
Decachlorobiphenyl	8.691	8.596	8.796	46.440	50.000	-7.1
Tetrachloro-m-xylene	3.688	3.590	3.790	52.230	50.000	4.5

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040725\
Data File : PO110278.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Apr 2025 17:14
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 07 17:33:50 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.691	3.688	473.9E6	273.9E6	52.066	52.232
2)	SA Decachlor...	8.741	8.691	403.2E6	112.7E6	52.443	46.444
	Target Compounds						
3)	L1 AR-1016-1	4.782	4.769	172.3E6	91853593	511.903	501.234
4)	L1 AR-1016-2	4.802	4.788	238.2E6	132.3E6	513.796	505.969
5)	L1 AR-1016-3	4.858	4.964	164.6E6	71832300	513.560	506.651
6)	L1 AR-1016-4	4.979	5.006	127.7E6	59082308	505.504	500.996
7)	L1 AR-1016-5	5.236	5.218	132.9E6	76567933	483.609	500.710
31)	L7 AR-1260-1	6.276	6.250	243.7E6	117.4E6	516.972	459.747
32)	L7 AR-1260-2	6.466	6.438	289.7E6	135.2E6	486.489	450.225
33)	L7 AR-1260-3	6.834	6.591	218.8E6	125.9E6	450.781	435.554
34)	L7 AR-1260-4	7.094	7.063	187.5E6	93074613	442.104	432.747
35)	L7 AR-1260-5	7.336	7.303	449.5E6	209.0E6	429.819	419.151


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040725\
Data File : PO110278.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Apr 2025 17:14
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 07 17:33:50 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.691 min
Delta R.T.: -0.001 min
Response: 473883144
Conc: 52.07 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.688 min
Delta R.T.: -0.002 min
Response: 273908061
Conc: 52.23 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.741 min
Delta R.T.: -0.004 min
Response: 403219619
Conc: 52.44 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.691 min
Delta R.T.: -0.005 min
Response: 112677778
Conc: 46.44 ng/ml



#3 AR-1016-1


R.T.: 4.782 min
Delta R.T.: -0.003 mi
Response: 172277908
Conc: 511.90 ng/m

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.769 min
Delta R.T.: -0.003 min
Response: 91853593
Conc: 501.23 ng/ml




#4 AR-1016-2

R.T.: 4.802 min
Delta R.T.: -0.003 min
Response: 238239899
Conc: 513.80 ng/ml



#4 AR-1016-2


R.T.: 4.788 min
Delta R.T.: -0.003 min
Response: 132313036
Conc: 505.97 ng/ml



#5 AR-1016-3


R.T.: 4.858 min
Delta R.T.: -0.002 min
Response: 164640078
Conc: 513.56 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.964 min
Delta R.T.: -0.003 min
Response: 71832300
Conc: 506.65 ng/ml




#6 AR-1016-4

R.T.: 4.979 min
Delta R.T.: -0.003 min
Response: 127735566
Conc: 505.50 ng/ml



#6 AR-1016-4


R.T.: 5.006 min
Delta R.T.: -0.003 min
Response: 59082308
Conc: 501.00 ng/ml



#7 AR-1016-5


R.T.: 5.236 min
Delta R.T.: -0.003 min
Response: 132928200
Conc: 483.61 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.218 min
Delta R.T.: -0.004 min
Response: 76567933
Conc: 500.71 ng/ml




#31 AR-1260-1

R.T.: 6.276 min
Delta R.T.: -0.003 min
Response: 243675158
Conc: 516.97 ng/ml



#31 AR-1260-1


R.T.: 6.250 min
Delta R.T.: -0.005 min
Response: 117438638
Conc: 459.75 ng/ml



#32 AR-1260-2


R.T.: 6.466 min
Delta R.T.: -0.003 min
Response: 289713640
Conc: 486.49 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.438 min
Delta R.T.: -0.005 min
Response: 135195944
Conc: 450.22 ng/ml




#33 AR-1260-3

R.T.: 6.834 min
Delta R.T.: -0.004 min
Response: 218764620
Conc: 450.78 ng/ml



#33 AR-1260-3


R.T.: 6.591 min
Delta R.T.: -0.004 min
Response: 125867255
Conc: 435.55 ng/ml



#34 AR-1260-4


R.T.: 7.094 min
Delta R.T.: -0.004 min
Response: 187526171
Conc: 442.10 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.063 min
Delta R.T.: -0.004 min
Response: 93074613
Conc: 432.75 ng/ml



#35 AR-1260-5

R.T.: 7.336 min
Delta R.T.: -0.003 min
Response: 449524286
Conc: 429.82 ng/ml



#35 AR-1260-5

R.T.: 7.303 min
Delta R.T.: -0.004 min
Response: 209028955
Conc: 419.15 ng/ml

Raw Data: **PO110286.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM **Case No.:** Q1730 **SAS No.:** Q1730 **SDG NO.:** Q1730

Continuing Calib Date: 04/07/2025 **Initial Calibration Date(s):** 03/18/2025 03/18/2025

Continuing Calib Time: 21:27 **Initial Calibration Time(s):** 14:03 22:15

GC Column: ZB-MR1 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	RT WINDOW TO	DIFF RT
Aroclor-1016-1	(1)	4.78	4.79	4.69	4.89	0.01
Aroclor-1016-2	(2)	4.80	4.80	4.70	4.90	0.00
Aroclor-1016-3	(3)	4.86	4.86	4.76	4.96	0.00
Aroclor-1016-4	(4)	4.98	4.98	4.88	5.08	0.00
Aroclor-1016-5	(5)	5.24	5.24	5.14	5.34	0.00
Aroclor-1260-1	(1)	6.28	6.28	6.18	6.38	0.00
Aroclor-1260-2	(2)	6.47	6.47	6.37	6.57	0.00
Aroclor-1260-3	(3)	6.83	6.84	6.74	6.94	0.01
Aroclor-1260-4	(4)	7.09	7.10	7.00	7.20	0.01
Aroclor-1260-5	(5)	7.34	7.34	7.24	7.44	0.00
Tetrachloro-m-xylene		3.69	3.69	3.59	3.79	0.00
Decachlorobiphenyl		8.74	8.75	8.65	8.85	0.01

Raw Data: **PO110286.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM **Case No.:** Q1730 **SAS No.:** Q1730 **SDG NO.:** Q1730

Continuing Calib Date: 04/07/2025 **Initial Calibration Date(s):** 03/18/2025 03/18/2025

Continuing Calib Time: 21:27 **Initial Calibration Time(s):** 14:03 22:15

GC Column: ZB-MR2 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	RT WINDOW TO	DIFF RT
Aroclor-1016-1	(1)	4.77	4.77	4.67	4.87	0.00
Aroclor-1016-2	(2)	4.79	4.79	4.69	4.89	0.00
Aroclor-1016-3	(3)	4.96	4.97	4.87	5.07	0.01
Aroclor-1016-4	(4)	5.01	5.01	4.91	5.11	0.00
Aroclor-1016-5	(5)	5.22	5.22	5.12	5.32	0.00
Aroclor-1260-1	(1)	6.25	6.26	6.16	6.36	0.01
Aroclor-1260-2	(2)	6.44	6.44	6.34	6.54	0.00
Aroclor-1260-3	(3)	6.59	6.60	6.50	6.70	0.01
Aroclor-1260-4	(4)	7.06	7.07	6.97	7.17	0.01
Aroclor-1260-5	(5)	7.30	7.31	7.21	7.41	0.01
Tetrachloro-m-xylene		3.69	3.69	3.59	3.79	0.00
Decachlorobiphenyl		8.69	8.70	8.60	8.80	0.01

Alliance TECHNICAL GROUP

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM **Case No.:** Q1730 **SAS No.:** Q1730 **SDG NO.:** Q1730

GC Column: ZB-MR1 **ID:** 0.32 **(mm)** **Initi. Calib. Date(s):** 03/18/2025 03/18/2025

Client Sample No.: CCAL03 **Date Analyzed:** 04/07/2025

Lab Sample No.: AR1660CCC500 **Data File :** PO110286.D **Time Analyzed:** 21:27

COMPOUND	RT	RT WINDOW FROM	RT WINDOW TO	CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.783	4.685	4.885	528.320	500.000	5.7
Aroclor-1016-2	4.802	4.704	4.904	526.450	500.000	5.3
Aroclor-1016-3	4.858	4.761	4.961	529.370	500.000	5.9
Aroclor-1016-4	4.979	4.881	5.081	518.990	500.000	3.8
Aroclor-1016-5	5.236	5.139	5.339	496.090	500.000	-0.8
Aroclor-1260-1	6.277	6.180	6.380	550.160	500.000	10.0
Aroclor-1260-2	6.466	6.369	6.569	508.400	500.000	1.7
Aroclor-1260-3	6.834	6.737	6.937	481.740	500.000	-3.7
Aroclor-1260-4	7.094	6.998	7.198	471.490	500.000	-5.7
Aroclor-1260-5	7.336	7.239	7.439	462.920	500.000	-7.4
Decachlorobiphenyl	8.740	8.645	8.845	54.690	50.000	9.4
Tetrachloro-m-xylene	3.691	3.592	3.792	53.480	50.000	7.0

Alliance TECHNICAL GROUP

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM **Case No.:** Q1730 **SAS No.:** Q1730 **SDG NO.:** Q1730

GC Column: ZB-MR2 **ID:** 0.32 **(mm)** **Initi. Calib. Date(s):** 03/18/2025 03/18/2025

Client Sample No.: CCAL03 **Date Analyzed:** 04/07/2025

Lab Sample No.: AR1660CCC500 **Data File :** PO110286.D **Time Analyzed:** 21:27

COMPOUND	RT	RT WINDOW FROM	RT WINDOW TO	CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.769	4.672	4.872	512.420	500.000	2.5
Aroclor-1016-2	4.788	4.691	4.891	514.070	500.000	2.8
Aroclor-1016-3	4.964	4.867	5.067	515.160	500.000	3.0
Aroclor-1016-4	5.006	4.909	5.109	469.250	500.000	-6.2
Aroclor-1016-5	5.218	5.122	5.322	514.370	500.000	2.9
Aroclor-1260-1	6.251	6.155	6.355	483.370	500.000	-3.3
Aroclor-1260-2	6.439	6.342	6.542	477.180	500.000	-4.6
Aroclor-1260-3	6.591	6.495	6.695	462.160	500.000	-7.6
Aroclor-1260-4	7.063	6.967	7.167	461.970	500.000	-7.6
Aroclor-1260-5	7.303	7.207	7.407	444.510	500.000	-11.1
Decachlorobiphenyl	8.691	8.596	8.796	48.770	50.000	-2.5
Tetrachloro-m-xylene	3.688	3.590	3.790	53.150	50.000	6.3

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040725\
Data File : PO110286.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Apr 2025 21:27
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 04/08/2025
Supervised By :mohammad ahmed 04/09/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 01:36:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.691	3.688	486.8E6	278.7E6	53.484	53.152
2)	SA Decachlor...	8.740	8.691	420.5E6	118.3E6	54.690	48.766
	Target Compounds						
3)	L1 AR-1016-1	4.783	4.769	177.8E6	93902896	528.316	512.417
4)	L1 AR-1016-2	4.802	4.788	244.1E6	134.4E6	526.454	514.067
5)	L1 AR-1016-3	4.858	4.964	169.7E6	73038464	529.369	515.158
6)	L1 AR-1016-4	4.979	5.006	131.1E6	55338718	518.987	469.252
7)	L1 AR-1016-5	5.236	5.218	136.4E6	78657229	496.086m	514.373m
31)	L7 AR-1260-1	6.277	6.251	259.3E6	123.5E6	550.162	483.373
32)	L7 AR-1260-2	6.466	6.439	302.8E6	143.3E6	508.402	477.179
33)	L7 AR-1260-3	6.834	6.591	233.8E6	133.6E6	481.744	462.164
34)	L7 AR-1260-4	7.094	7.063	200.0E6	99360275	471.486	461.972
35)	L7 AR-1260-5	7.336	7.303	484.1E6	221.7E6	462.918	444.506

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040725\
Data File : PO110286.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Apr 2025 21:27
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
AR1660CCC500


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 04/08/2025
Supervised By :mohammad ahmed 04/09/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 01:36:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene

R.T.: 3.691 min
Delta R.T.: -0.001 min
Response: 486781502
Conc: 53.48 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/08/2025
Supervised By :mohammad ahmed 04/09/2025




#1 Tetrachloro-m-xylene

R.T.: 3.688 min
Delta R.T.: -0.002 min
Response: 278731990
Conc: 53.15 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.740 min
Delta R.T.: -0.005 min
Response: 420491355
Conc: 54.69 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.691 min
Delta R.T.: -0.004 min
Response: 118311884
Conc: 48.77 ng/ml




#3 AR-1016-1

R.T.: 4.783 min
Delta R.T.: -0.002 mi
Response: 177801623
Conc: 528.32 ng/m

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/08/2025
Supervised By :mohammad ahmed 04/09/2025




#3 AR-1016-1

R.T.: 4.769 min
Delta R.T.: -0.003 min
Response: 93902896
Conc: 512.42 ng/ml




#4 AR-1016-2

R.T.: 4.802 min
Delta R.T.: -0.003 min
Response: 244109403
Conc: 526.45 ng/ml



#4 AR-1016-2

R.T.: 4.788 min
Delta R.T.: -0.003 min
Response: 134430740
Conc: 514.07 ng/ml




#5 AR-1016-3

R.T.: 4.858 min
Delta R.T.: -0.002 min
Response: 169708224
Conc: 529.37 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 04/08/2025
Supervised By :mohammad ahmed 04/09/2025




#5 AR-1016-3

R.T.: 4.964 min
Delta R.T.: -0.003 min
Response: 73038464
Conc: 515.16 ng/ml




#6 AR-1016-4

R.T.: 4.979 min
Delta R.T.: -0.002 min
Response: 131142472
Conc: 518.99 ng/ml



#6 AR-1016-4

R.T.: 5.006 min
Delta R.T.: -0.003 min
Response: 55338718
Conc: 469.25 ng/ml




#7 AR-1016-5

R.T.: 5.236 min
Delta R.T.: -0.003 min
Response: 136357657
Conc: 496.09 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 04/08/2025
Supervised By :mohammad ahmed 04/09/2025




#7 AR-1016-5

R.T.: 5.218 min
Delta R.T.: -0.004 min
Response: 78657229
Conc: 514.37 ng/ml m




#31 AR-1260-1

R.T.: 6.277 min
Delta R.T.: -0.003 min
Response: 259319565
Conc: 550.16 ng/ml



#31 AR-1260-1

R.T.: 6.251 min
Delta R.T.: -0.004 min
Response: 123473695
Conc: 483.37 ng/ml




#32 AR-1260-2

R.T.: 6.466 min
Delta R.T.: -0.003 min
Response: 302763163
Conc: 508.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/08/2025
Supervised By :mohammad ahmed 04/09/2025




#32 AR-1260-2

R.T.: 6.439 min
Delta R.T.: -0.004 min
Response: 143289838
Conc: 477.18 ng/ml




#33 AR-1260-3

R.T.: 6.834 min
Delta R.T.: -0.003 min
Response: 233791087
Conc: 481.74 ng/ml



#33 AR-1260-3

R.T.: 6.591 min
Delta R.T.: -0.004 min
Response: 133557256
Conc: 462.16 ng/ml




#34 AR-1260-4

R.T.: 7.094 min
Delta R.T.: -0.004 min
Response: 199989246
Conc: 471.49 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 04/08/2025
Supervised By :mohammad ahmed 04/09/2025




#34 AR-1260-4

R.T.: 7.063 min
Delta R.T.: -0.004 min
Response: 99360275
Conc: 461.97 ng/ml



#35 AR-1260-5

R.T.: 7.336 min
Delta R.T.: -0.004 min
Response: 484140405
Conc: 462.92 ng/ml



#35 AR-1260-5

R.T.: 7.303 min
Delta R.T.: -0.004 min
Response: 221673295
Conc: 444.51 ng/ml

Raw Data: **PO110292.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM **Case No.:** Q1730 **SAS No.:** Q1730 **SDG NO.:** Q1730

Continuing Calib Date: 04/08/2025 **Initial Calibration Date(s):** 03/18/2025 03/18/2025

Continuing Calib Time: 09:08 **Initial Calibration Time(s):** 14:03 22:15

GC Column: ZB-MR1 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	RT WINDOW TO	DIFF RT
Aroclor-1016-1	(1)	4.78	4.79	4.69	4.89	0.01
Aroclor-1016-2	(2)	4.80	4.80	4.70	4.90	0.00
Aroclor-1016-3	(3)	4.86	4.86	4.76	4.96	0.00
Aroclor-1016-4	(4)	4.98	4.98	4.88	5.08	0.00
Aroclor-1016-5	(5)	5.24	5.24	5.14	5.34	0.00
Aroclor-1260-1	(1)	6.28	6.28	6.18	6.38	0.00
Aroclor-1260-2	(2)	6.47	6.47	6.37	6.57	0.00
Aroclor-1260-3	(3)	6.83	6.84	6.74	6.94	0.01
Aroclor-1260-4	(4)	7.09	7.10	7.00	7.20	0.01
Aroclor-1260-5	(5)	7.34	7.34	7.24	7.44	0.00
Tetrachloro-m-xylene		3.69	3.69	3.59	3.79	0.00
Decachlorobiphenyl		8.74	8.75	8.65	8.85	0.01

Raw Data: **PO110292.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM **Case No.:** Q1730 **SAS No.:** Q1730 **SDG NO.:** Q1730

Continuing Calib Date: 04/08/2025 **Initial Calibration Date(s):** 03/18/2025 03/18/2025

Continuing Calib Time: 09:08 **Initial Calibration Time(s):** 14:03 22:15

GC Column: ZB-MR2 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	RT WINDOW TO	DIFF RT
Aroclor-1016-1	(1)	4.77	4.77	4.67	4.87	0.00
Aroclor-1016-2	(2)	4.79	4.79	4.69	4.89	0.00
Aroclor-1016-3	(3)	4.96	4.97	4.87	5.07	0.01
Aroclor-1016-4	(4)	5.00	5.01	4.91	5.11	0.01
Aroclor-1016-5	(5)	5.22	5.22	5.12	5.32	0.00
Aroclor-1260-1	(1)	6.25	6.26	6.16	6.36	0.01
Aroclor-1260-2	(2)	6.44	6.44	6.34	6.54	0.00
Aroclor-1260-3	(3)	6.59	6.60	6.50	6.70	0.01
Aroclor-1260-4	(4)	7.06	7.07	6.97	7.17	0.01
Aroclor-1260-5	(5)	7.30	7.31	7.21	7.41	0.01
Tetrachloro-m-xylene		3.69	3.69	3.59	3.79	0.00
Decachlorobiphenyl		8.69	8.70	8.60	8.80	0.01

Raw Data: **PO110292.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM **Case No.:** Q1730 **SAS No.:** Q1730 **SDG NO.:** Q1730

GC Column: ZB-MR1 **ID:** 0.32 **(mm)** **Initi. Calib. Date(s):** 03/18/2025 03/18/2025

Client Sample No.: CCAL04 **Date Analyzed:** 04/08/2025

Lab Sample No.: AR1660CCC500 **Data File :** PO110292.D **Time Analyzed:** 09:08

COMPOUND	RT	RT WINDOW FROM	RT WINDOW TO	CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.783	4.685	4.885	490.750	500.000	-1.9
Aroclor-1016-2	4.802	4.704	4.904	490.770	500.000	-1.8
Aroclor-1016-3	4.858	4.761	4.961	494.230	500.000	-1.2
Aroclor-1016-4	4.979	4.881	5.081	493.910	500.000	-1.2
Aroclor-1016-5	5.236	5.139	5.339	541.310	500.000	8.3
Aroclor-1260-1	6.277	6.180	6.380	505.880	500.000	1.2
Aroclor-1260-2	6.467	6.369	6.569	476.430	500.000	-4.7
Aroclor-1260-3	6.834	6.737	6.937	473.120	500.000	-5.4
Aroclor-1260-4	7.094	6.998	7.198	466.380	500.000	-6.7
Aroclor-1260-5	7.337	7.239	7.439	469.070	500.000	-6.2
Decachlorobiphenyl	8.742	8.645	8.845	53.700	50.000	7.4
Tetrachloro-m-xylene	3.690	3.592	3.792	48.950	50.000	-2.1

Raw Data: **PO110292.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM **Case No.:** Q1730 **SAS No.:** Q1730 **SDG NO.:** Q1730

GC Column: ZB-MR2 **ID:** 0.32 **(mm)** **Initi. Calib. Date(s):** 03/18/2025 03/18/2025

Client Sample No.: CCAL04 **Date Analyzed:** 04/08/2025

Lab Sample No.: AR1660CCC500 **Data File :** PO110292.D **Time Analyzed:** 09:08

COMPOUND	RT	RT WINDOW FROM	RT WINDOW TO	CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.768	4.672	4.872	460.700	500.000	-7.9
Aroclor-1016-2	4.787	4.691	4.891	458.320	500.000	-8.3
Aroclor-1016-3	4.963	4.867	5.067	460.180	500.000	-8.0
Aroclor-1016-4	5.004	4.909	5.109	424.320	500.000	-15.1
Aroclor-1016-5	5.218	5.122	5.322	488.430	500.000	-2.3
Aroclor-1260-1	6.250	6.155	6.355	451.100	500.000	-9.8
Aroclor-1260-2	6.438	6.342	6.542	446.060	500.000	-10.8
Aroclor-1260-3	6.590	6.495	6.695	427.140	500.000	-14.6
Aroclor-1260-4	7.062	6.967	7.167	425.210	500.000	-15.0
Aroclor-1260-5	7.303	7.207	7.407	416.030	500.000	-16.8
Decachlorobiphenyl	8.691	8.596	8.796	44.590	50.000	-10.8
Tetrachloro-m-xylene	3.687	3.590	3.790	47.020	50.000	-6.0

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040825\
Data File : PO110292.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Apr 2025 09:08
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 10:41:56 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1) SA	Tetrachlo...	3.690	3.687	445.6E6	246.6E6	48.954	47.018
2) SA	Decachlor...	8.742	8.691	412.9E6	108.2E6	53.697	44.591
	Target Compounds						
3) L1	AR-1016-1	4.783	4.768	165.2E6	84424795	490.747	460.696
4) L1	AR-1016-2	4.802	4.787	227.6E6	119.9E6	490.775	458.320
5) L1	AR-1016-3	4.858	4.963	158.4E6	65243266	494.235	460.177
6) L1	AR-1016-4	4.979	5.004	124.8E6	50039864	493.910	424.319
7) L1	AR-1016-5	5.236	5.218	148.8E6	74690506	541.314	488.433
31) L7	AR-1260-1	6.277	6.250	238.4E6	115.2E6	505.880	451.099
32) L7	AR-1260-2	6.467	6.438	283.7E6	133.9E6	476.430	446.059
33) L7	AR-1260-3	6.834	6.590	229.6E6	123.4E6	473.116	427.141
34) L7	AR-1260-4	7.094	7.062	197.8E6	91454142	466.378	425.213
35) L7	AR-1260-5	7.337	7.303	490.6E6	207.5E6	469.072	416.025


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040825\
Data File : PO110292.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Apr 2025 09:08
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 10:41:56 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.690 min
Delta R.T.: -0.002 min
Response: 445550594
Conc: 48.95 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.687 min
Delta R.T.: -0.003 min
Response: 246565424
Conc: 47.02 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.742 min
Delta R.T.: -0.004 min
Response: 412860287
Conc: 53.70 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.691 min
Delta R.T.: -0.005 min
Response: 108181543
Conc: 44.59 ng/ml



#3 AR-1016-1


R.T.: 4.783 min
Delta R.T.: -0.002 min
Response: 165158163
Conc: 490.75 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.768 min
Delta R.T.: -0.004 min
Response: 84424795
Conc: 460.70 ng/ml




#4 AR-1016-2

R.T.: 4.802 min
Delta R.T.: -0.002 min
Response: 227565206
Conc: 490.77 ng/ml



#4 AR-1016-2


R.T.: 4.787 min
Delta R.T.: -0.004 min
Response: 119852669
Conc: 458.32 ng/ml



#5 AR-1016-3


R.T.: 4.858 min
Delta R.T.: -0.003 min
Response: 158444813
Conc: 494.23 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.963 min
Delta R.T.: -0.004 min
Response: 65243266
Conc: 460.18 ng/ml




#6 AR-1016-4

R.T.: 4.979 min
Delta R.T.: -0.002 min
Response: 124805818
Conc: 493.91 ng/ml



#6 AR-1016-4


R.T.: 5.004 min
Delta R.T.: -0.004 min
Response: 50039864
Conc: 424.32 ng/ml



#7 AR-1016-5


R.T.: 5.236 min
Delta R.T.: -0.003 min
Response: 148789423
Conc: 541.31 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.218 min
Delta R.T.: -0.004 min
Response: 74690506
Conc: 488.43 ng/ml




#31 AR-1260-1

R.T.: 6.277 min
Delta R.T.: -0.003 min
Response: 238446806
Conc: 505.88 ng/ml



#31 AR-1260-1


R.T.: 6.250 min
Delta R.T.: -0.005 min
Response: 115229432
Conc: 451.10 ng/ml



#32 AR-1260-2


R.T.: 6.467 min
Delta R.T.: -0.002 min
Response: 283723218
Conc: 476.43 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.438 min
Delta R.T.: -0.005 min
Response: 133944941
Conc: 446.06 ng/ml




#33 AR-1260-3

R.T.: 6.834 min
Delta R.T.: -0.003 min
Response: 229604002
Conc: 473.12 ng/ml



#33 AR-1260-3


R.T.: 6.590 min
Delta R.T.: -0.005 min
Response: 123436061
Conc: 427.14 ng/ml



#34 AR-1260-4


R.T.: 7.094 min
Delta R.T.: -0.004 min
Response: 197822722
Conc: 466.38 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.062 min
Delta R.T.: -0.005 min
Response: 91454142
Conc: 425.21 ng/ml



#35 AR-1260-5

R.T.: 7.337 min
Delta R.T.: -0.003 min
Response: 490577287
Conc: 469.07 ng/ml



#35 AR-1260-5

R.T.: 7.303 min
Delta R.T.: -0.004 min
Response: 207470142
Conc: 416.03 ng/ml

Raw Data: **PO110302.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM **Case No.:** Q1730 **SAS No.:** Q1730 **SDG NO.:** Q1730

Continuing Calib Date: 04/08/2025 **Initial Calibration Date(s):** 03/18/2025 03/18/2025

Continuing Calib Time: 12:58 **Initial Calibration Time(s):** 14:03 22:15

GC Column: ZB-MR1 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	RT WINDOW TO	DIFF RT
Aroclor-1016-1	(1)	4.78	4.79	4.69	4.89	0.01
Aroclor-1016-2	(2)	4.80	4.80	4.70	4.90	0.00
Aroclor-1016-3	(3)	4.86	4.86	4.76	4.96	0.00
Aroclor-1016-4	(4)	4.98	4.98	4.88	5.08	0.00
Aroclor-1016-5	(5)	5.24	5.24	5.14	5.34	0.00
Aroclor-1260-1	(1)	6.28	6.28	6.18	6.38	0.00
Aroclor-1260-2	(2)	6.47	6.47	6.37	6.57	0.00
Aroclor-1260-3	(3)	6.83	6.84	6.74	6.94	0.01
Aroclor-1260-4	(4)	7.09	7.10	7.00	7.20	0.01
Aroclor-1260-5	(5)	7.34	7.34	7.24	7.44	0.00
Tetrachloro-m-xylene		3.69	3.69	3.59	3.79	0.00
Decachlorobiphenyl		8.74	8.75	8.65	8.85	0.01

Raw Data: **PO110302.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM **Case No.:** Q1730 **SAS No.:** Q1730 **SDG NO.:** Q1730

Continuing Calib Date: 04/08/2025 **Initial Calibration Date(s):** 03/18/2025 03/18/2025

Continuing Calib Time: 12:58 **Initial Calibration Time(s):** 14:03 22:15

GC Column: ZB-MR2 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	TO	DIFF RT
Aroclor-1016-1	(1)	4.77	4.77	4.67	4.87	0.00
Aroclor-1016-2	(2)	4.79	4.79	4.69	4.89	0.00
Aroclor-1016-3	(3)	4.96	4.97	4.87	5.07	0.01
Aroclor-1016-4	(4)	5.01	5.01	4.91	5.11	0.00
Aroclor-1016-5	(5)	5.22	5.22	5.12	5.32	0.00
Aroclor-1260-1	(1)	6.25	6.26	6.16	6.36	0.01
Aroclor-1260-2	(2)	6.44	6.44	6.34	6.54	0.00
Aroclor-1260-3	(3)	6.59	6.60	6.50	6.70	0.01
Aroclor-1260-4	(4)	7.06	7.07	6.97	7.17	0.01
Aroclor-1260-5	(5)	7.30	7.31	7.21	7.41	0.01
Tetrachloro-m-xylene		3.69	3.69	3.59	3.79	0.00
Decachlorobiphenyl		8.69	8.70	8.60	8.80	0.01

Raw Data: **PO110302.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM **Case No.:** Q1730 **SAS No.:** Q1730 **SDG NO.:** Q1730

GC Column: ZB-MR1 **ID:** 0.32 **(mm)** **Initi. Calib. Date(s):** 03/18/2025 03/18/2025

Client Sample No.: CCAL05 **Date Analyzed:** 04/08/2025

Lab Sample No.: AR1660CCC500 **Data File :** PO110302.D **Time Analyzed:** 12:58

COMPOUND	RT	RT WINDOW FROM	RT WINDOW TO	CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.783	4.685	4.885	478.010	500.000	-4.4
Aroclor-1016-2	4.802	4.704	4.904	481.930	500.000	-3.6
Aroclor-1016-3	4.858	4.761	4.961	488.150	500.000	-2.4
Aroclor-1016-4	4.979	4.881	5.081	483.730	500.000	-3.3
Aroclor-1016-5	5.237	5.139	5.339	484.580	500.000	-3.1
Aroclor-1260-1	6.277	6.180	6.380	484.710	500.000	-3.1
Aroclor-1260-2	6.466	6.369	6.569	466.260	500.000	-6.7
Aroclor-1260-3	6.834	6.737	6.937	478.180	500.000	-4.4
Aroclor-1260-4	7.094	6.998	7.198	469.910	500.000	-6.0
Aroclor-1260-5	7.336	7.239	7.439	460.630	500.000	-7.9
Decachlorobiphenyl	8.739	8.645	8.845	56.540	50.000	13.1
Tetrachloro-m-xylene	3.691	3.592	3.792	47.860	50.000	-4.3

Raw Data: **PO110302.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM **Case No.:** Q1730 **SAS No.:** Q1730 **SDG NO.:** Q1730

GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 03/18/2025 03/18/2025

Client Sample No.: CCAL05 **Date Analyzed:** 04/08/2025

Lab Sample No.: AR1660CCC500 **Data File :** PO110302.D **Time Analyzed:** 12:58

COMPOUND	RT	RT WINDOW FROM	RT WINDOW TO	CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.769	4.672	4.872	459.150	500.000	-8.2
Aroclor-1016-2	4.788	4.691	4.891	461.630	500.000	-7.7
Aroclor-1016-3	4.963	4.867	5.067	464.240	500.000	-7.2
Aroclor-1016-4	5.005	4.909	5.109	462.280	500.000	-7.5
Aroclor-1016-5	5.218	5.122	5.322	459.630	500.000	-8.1
Aroclor-1260-1	6.250	6.155	6.355	447.090	500.000	-10.6
Aroclor-1260-2	6.438	6.342	6.542	439.930	500.000	-12.0
Aroclor-1260-3	6.591	6.495	6.695	426.390	500.000	-14.7
Aroclor-1260-4	7.063	6.967	7.167	425.830	500.000	-14.8
Aroclor-1260-5	7.303	7.207	7.407	416.620	500.000	-16.7
Decachlorobiphenyl	8.692	8.596	8.796	45.310	50.000	-9.4
Tetrachloro-m-xylene	3.688	3.590	3.790	47.270	50.000	-5.5

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040825\
Data File : PO110302.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Apr 2025 12:58
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 13:41:57 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.691	3.688	435.6E6	247.9E6	47.860	47.272
2)	SA Decachlor...	8.739	8.692	434.7E6	109.9E6	56.541	45.308
	Target Compounds						
3)	L1 AR-1016-1	4.783	4.769	160.9E6	84140877	478.015	459.147
4)	L1 AR-1016-2	4.802	4.788	223.5E6	120.7E6	481.932	461.633
5)	L1 AR-1016-3	4.858	4.963	156.5E6	65819182	488.147	464.239
6)	L1 AR-1016-4	4.979	5.005	122.2E6	54516660	483.727	462.281
7)	L1 AR-1016-5	5.237	5.218	133.2E6	70285787	484.577	459.629
31)	L7 AR-1260-1	6.277	6.250	228.5E6	114.2E6	484.705	447.094
32)	L7 AR-1260-2	6.466	6.438	277.7E6	132.1E6	466.258	439.928
33)	L7 AR-1260-3	6.834	6.591	232.1E6	123.2E6	478.181	426.386
34)	L7 AR-1260-4	7.094	7.063	199.3E6	91586397	469.907	425.828
35)	L7 AR-1260-5	7.336	7.303	481.7E6	207.8E6	460.625	416.615


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040825\
Data File : PO110302.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Apr 2025 12:58
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 13:41:57 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.691 min
Delta R.T.: -0.001 min
Response: 435598852
Conc: 47.86 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.688 min
Delta R.T.: -0.002 min
Response: 247897432
Conc: 47.27 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.739 min
Delta R.T.: -0.006 min
Response: 434722533
Conc: 56.54 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.692 min
Delta R.T.: -0.004 min
Response: 109921655
Conc: 45.31 ng/ml



#3 AR-1016-1


R.T.: 4.783 min
Delta R.T.: -0.002 min
Response: 160873001
Conc: 478.01 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.769 min
Delta R.T.: -0.003 min
Response: 84140877
Conc: 459.15 ng/ml




#4 AR-1016-2

R.T.: 4.802 min
Delta R.T.: -0.003 min
Response: 223464953
Conc: 481.93 ng/ml



#4 AR-1016-2


R.T.: 4.788 min
Delta R.T.: -0.003 min
Response: 120719056
Conc: 461.63 ng/ml



#5 AR-1016-3


R.T.: 4.858 min
Delta R.T.: -0.003 min
Response: 156493158
Conc: 488.15 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.963 min
Delta R.T.: -0.004 min
Response: 65819182
Conc: 464.24 ng/ml




#6 AR-1016-4

R.T.: 4.979 min
Delta R.T.: -0.002 min
Response: 122232752
Conc: 483.73 ng/ml



#6 AR-1016-4


R.T.: 5.005 min
Delta R.T.: -0.004 min
Response: 54516660
Conc: 462.28 ng/ml



#7 AR-1016-5


R.T.: 5.237 min
Delta R.T.: -0.002 min
Response: 133194313
Conc: 484.58 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.218 min
Delta R.T.: -0.004 min
Response: 70285787
Conc: 459.63 ng/ml




#31 AR-1260-1

R.T.: 6.277 min
Delta R.T.: -0.003 min
Response: 228466177
Conc: 484.71 ng/ml



#31 AR-1260-1


R.T.: 6.250 min
Delta R.T.: -0.005 min
Response: 114206424
Conc: 447.09 ng/ml



#32 AR-1260-2


R.T.: 6.466 min
Delta R.T.: -0.003 min
Response: 277665605
Conc: 466.26 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.438 min
Delta R.T.: -0.005 min
Response: 132103988
Conc: 439.93 ng/ml




#33 AR-1260-3

R.T.: 6.834 min
Delta R.T.: -0.004 min
Response: 232062227
Conc: 478.18 ng/ml



#33 AR-1260-3


R.T.: 6.591 min
Delta R.T.: -0.004 min
Response: 123217994
Conc: 426.39 ng/ml



#34 AR-1260-4


R.T.: 7.094 min
Delta R.T.: -0.003 mi
Response: 199319657
Conc: 469.91 ng/m

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.063 min
Delta R.T.: -0.004 min
Response: 91586397
Conc: 425.83 ng/ml



#35 AR-1260-5

R.T.: 7.336 min
Delta R.T.: -0.004 min
Response: 481743144
Conc: 460.63 ng/ml



#35 AR-1260-5

R.T.: 7.303 min
Delta R.T.: -0.004 min
Response: 207764421
Conc: 416.62 ng/ml

Raw Data: **PP071136.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM **Case No.:** Q1730 **SAS No.:** Q1730 **SDG NO.:** Q1730

Continuing Calib Date: 04/08/2025 **Initial Calibration Date(s):** 03/27/2025 03/27/2025

Continuing Calib Time: 09:37 **Initial Calibration Time(s):** 10:30 17:53

GC Column: ZB-MR1 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	RT WINDOW TO	DIFF RT
Aroclor-1016-1	(1)	5.68	5.66	5.56	5.76	-0.02
Aroclor-1016-2	(2)	5.70	5.69	5.59	5.79	-0.01
Aroclor-1016-3	(3)	5.76	5.75	5.65	5.85	-0.01
Aroclor-1016-4	(4)	5.86	5.85	5.75	5.95	-0.01
Aroclor-1016-5	(5)	6.15	6.14	6.04	6.24	-0.01
Aroclor-1260-1	(1)	7.27	7.26	7.16	7.36	-0.01
Aroclor-1260-2	(2)	7.53	7.51	7.41	7.61	-0.02
Aroclor-1260-3	(3)	7.89	7.87	7.77	7.97	-0.02
Aroclor-1260-4	(4)	8.11	8.09	7.99	8.19	-0.02
Aroclor-1260-5	(5)	8.43	8.41	8.31	8.51	-0.02
Tetrachloro-m-xylene		4.52	4.51	4.41	4.61	-0.01
Decachlorobiphenyl		10.26	10.23	10.13	10.33	-0.03

Raw Data: **PP071136.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM **Case No.:** Q1730 **SAS No.:** Q1730 **SDG NO.:** Q1730

Continuing Calib Date: 04/08/2025 **Initial Calibration Date(s):** 03/27/2025 03/27/2025

Continuing Calib Time: 09:37 **Initial Calibration Time(s):** 10:30 17:53

GC Column: ZB-MR2 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	TO	DIFF RT
Aroclor-1016-1	(1)	4.90	4.90	4.80	5.00	0.00
Aroclor-1016-2	(2)	4.92	4.92	4.82	5.02	0.00
Aroclor-1016-3	(3)	5.10	5.10	5.00	5.20	0.00
Aroclor-1016-4	(4)	5.14	5.14	5.04	5.24	0.00
Aroclor-1016-5	(5)	5.36	5.36	5.26	5.46	0.00
Aroclor-1260-1	(1)	6.39	6.39	6.29	6.49	0.00
Aroclor-1260-2	(2)	6.58	6.58	6.48	6.68	0.00
Aroclor-1260-3	(3)	6.74	6.73	6.63	6.83	-0.01
Aroclor-1260-4	(4)	7.21	7.21	7.11	7.31	0.00
Aroclor-1260-5	(5)	7.45	7.45	7.35	7.55	0.00
Tetrachloro-m-xylene		3.82	3.82	3.72	3.92	0.00
Decachlorobiphenyl		8.86	8.86	8.76	8.96	0.00

Raw Data: **PP071136.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM **Case No.:** Q1730 **SAS No.:** Q1730 **SDG NO.:** Q1730

GC Column: ZB-MR1 **ID:** 0.32 **(mm)** **Initi. Calib. Date(s):** 03/27/2025 03/27/2025

Client Sample No.: CCAL06 **Date Analyzed:** 04/08/2025

Lab Sample No.: AR1660CCC500 **Data File :** PP071136.D **Time Analyzed:** 09:37

COMPOUND	RT	RT WINDOW FROM	RT WINDOW TO	CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.679	5.563	5.763	504.230	500.000	0.8
Aroclor-1016-2	5.700	5.585	5.785	472.510	500.000	-5.5
Aroclor-1016-3	5.763	5.648	5.848	488.820	500.000	-2.2
Aroclor-1016-4	5.861	5.745	5.945	525.740	500.000	5.1
Aroclor-1016-5	6.154	6.038	6.238	539.860	500.000	8.0
Aroclor-1260-1	7.273	7.156	7.356	518.030	500.000	3.6
Aroclor-1260-2	7.527	7.410	7.610	536.780	500.000	7.4
Aroclor-1260-3	7.885	7.768	7.968	503.360	500.000	0.7
Aroclor-1260-4	8.109	7.992	8.192	501.370	500.000	0.3
Aroclor-1260-5	8.431	8.311	8.511	506.740	500.000	1.3
Decachlorobiphenyl	10.255	10.125	10.325	51.060	50.000	2.1
Tetrachloro-m-xylene	4.524	4.412	4.612	51.340	50.000	2.7

Raw Data: **PP071136.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM **Case No.:** Q1730 **SAS No.:** Q1730 **SDG NO.:** Q1730

GC Column: ZB-MR2 **ID:** 0.32 **(mm)** **Initi. Calib. Date(s):** 03/27/2025 03/27/2025

Client Sample No.: CCAL06 **Date Analyzed:** 04/08/2025

Lab Sample No.: AR1660CCC500 **Data File :** PP071136.D **Time Analyzed:** 09:37

COMPOUND	RT	RT WINDOW FROM	RT WINDOW TO	CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.904	4.803	5.003	491.030	500.000	-1.8
Aroclor-1016-2	4.923	4.822	5.022	483.510	500.000	-3.3
Aroclor-1016-3	5.101	4.999	5.199	522.710	500.000	4.5
Aroclor-1016-4	5.142	5.041	5.241	542.480	500.000	8.5
Aroclor-1016-5	5.357	5.255	5.455	530.130	500.000	6.0
Aroclor-1260-1	6.393	6.292	6.492	547.620	500.000	9.5
Aroclor-1260-2	6.582	6.480	6.680	534.460	500.000	6.9
Aroclor-1260-3	6.735	6.634	6.834	546.870	500.000	9.4
Aroclor-1260-4	7.207	7.105	7.305	489.720	500.000	-2.1
Aroclor-1260-5	7.448	7.346	7.546	462.070	500.000	-7.6
Decachlorobiphenyl	8.862	8.759	8.959	43.790	50.000	-12.4
Tetrachloro-m-xylene	3.817	3.716	3.916	47.940	50.000	-4.1

Quantitation Report (QT Reviewed)

Sample Results: **PP071136.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040825\
Data File : PP071136.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Apr 2025 09:37
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 04/09/2025
Supervised By :mohammad ahmed 04/10/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 10:57:00 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 28 03:13:46 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.524	3.817	75031562	49693990	51.336	47.937
2)	SA Decachlor...	10.255	8.862	53747529	34514110	51.061	43.789
	Target Compounds						
3)	L1 AR-1016-1	5.679	4.904	25640835	19684563	504.232	491.032
4)	L1 AR-1016-2	5.700	4.923	36881275	27978272	472.512m	483.509
5)	L1 AR-1016-3	5.763	5.101	23060044	16163902	488.823	522.709
6)	L1 AR-1016-4	5.861	5.142	19323906	13212574	525.739	542.479
7)	L1 AR-1016-5	6.154	5.357	18307730	17134433	539.858	530.130
31)	L7 AR-1260-1	7.273	6.393	34786874	28041460	518.026	547.617
32)	L7 AR-1260-2	7.527	6.582	51604417	34095263	536.784	534.460
33)	L7 AR-1260-3	7.885	6.735	40227656	29796945	503.360	546.865
34)	L7 AR-1260-4	8.109	7.207	39475978	23421548	501.366m	489.721
35)	L7 AR-1260-5	8.431	7.448	80513512	57532352	506.737	462.069

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040825\
Data File : PP071136.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Apr 2025 09:37
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1660CCC500


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 04/09/2025
Supervised By :mohammad ahmed 04/10/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 10:57:00 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 28 03:13:46 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene

R.T.: 4.524 min
Delta R.T.: 0.012 min
Response: 75031562
Conc: 51.34 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 04/09/2025
Supervised By :mohammad ahmed 04/10/2025




#1 Tetrachloro-m-xylene

R.T.: 3.817 min
Delta R.T.: 0.000 min
Response: 49693990
Conc: 47.94 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.255 min
Delta R.T.: 0.030 min
Response: 53747529
Conc: 51.06 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.862 min
Delta R.T.: 0.003 min
Response: 34514110
Conc: 43.79 ng/ml




#3 AR-1016-1

R.T.: 5.679 min
Delta R.T.: 0.016 mi
Response: 25640835
Conc: 504.23 ng/m

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/09/2025
Supervised By :mohammad ahmed 04/10/2025




#3 AR-1016-1

R.T.: 4.904 min
Delta R.T.: 0.001 min
Response: 19684563
Conc: 491.03 ng/ml




#4 AR-1016-2

R.T.: 5.700 min
Delta R.T.: 0.015 min
Response: 36881275
Conc: 472.51 ng/ml m



#4 AR-1016-2

R.T.: 4.923 min
Delta R.T.: 0.001 min
Response: 27978272
Conc: 483.51 ng/ml




#5 AR-1016-3

R.T.: 5.763 min
Delta R.T.: 0.016 mi
Response: 23060044
Conc: 488.82 ng/m

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/09/2025
Supervised By :mohammad ahmed 04/10/2025




#5 AR-1016-3

R.T.: 5.101 min
Delta R.T.: 0.001 min
Response: 16163902
Conc: 522.71 ng/ml




#6 AR-1016-4

R.T.: 5.861 min
Delta R.T.: 0.015 min
Response: 19323906
Conc: 525.74 ng/ml



#6 AR-1016-4

R.T.: 5.142 min
Delta R.T.: 0.001 min
Response: 13212574
Conc: 542.48 ng/ml




#7 AR-1016-5

R.T.: 6.154 min
Delta R.T.: 0.016 min
Response: 18307730
Conc: 539.86 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/09/2025
Supervised By :mohammad ahmed 04/10/2025




#7 AR-1016-5

R.T.: 5.357 min
Delta R.T.: 0.002 min
Response: 17134433
Conc: 530.13 ng/ml




#31 AR-1260-1

R.T.: 7.273 min
Delta R.T.: 0.017 min
Response: 34786874
Conc: 518.03 ng/ml



#31 AR-1260-1

R.T.: 6.393 min
Delta R.T.: 0.001 min
Response: 28041460
Conc: 547.62 ng/ml




#32 AR-1260-2

R.T.: 7.527 min
Delta R.T.: 0.017 min
Response: 51604417
Conc: 536.78 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/09/2025
Supervised By :mohammad ahmed 04/10/2025




#32 AR-1260-2

R.T.: 6.582 min
Delta R.T.: 0.002 min
Response: 34095263
Conc: 534.46 ng/ml




#33 AR-1260-3

R.T.: 7.885 min
Delta R.T.: 0.018 min
Response: 40227656
Conc: 503.36 ng/ml



#33 AR-1260-3

R.T.: 6.735 min
Delta R.T.: 0.001 min
Response: 29796945
Conc: 546.87 ng/ml




#34 AR-1260-4

R.T.: 8.109 min
Delta R.T.: 0.018 min
Response: 39475978
Conc: 501.37 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/09/2025
Supervised By :mohammad ahmed 04/10/2025




#34 AR-1260-4

R.T.: 7.207 min
Delta R.T.: 0.002 min
Response: 23421548
Conc: 489.72 ng/ml



#35 AR-1260-5

R.T.: 8.431 min
Delta R.T.: 0.020 min
Response: 80513512
Conc: 506.74 ng/ml



#35 AR-1260-5

R.T.: 7.448 min
Delta R.T.: 0.002 min
Response: 57532352
Conc: 462.07 ng/ml

Raw Data: **PP071151.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM **Case No.:** Q1730 **SAS No.:** Q1730 **SDG NO.:** Q1730

Continuing Calib Date: 04/08/2025 **Initial Calibration Date(s):** 03/27/2025 03/27/2025

Continuing Calib Time: 15:29 **Initial Calibration Time(s):** 10:30 17:53

GC Column: ZB-MR1 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	RT WINDOW TO	DIFF RT
Aroclor-1016-1	(1)	5.67	5.66	5.56	5.76	-0.01
Aroclor-1016-2	(2)	5.69	5.69	5.59	5.79	0.00
Aroclor-1016-3	(3)	5.76	5.75	5.65	5.85	-0.01
Aroclor-1016-4	(4)	5.85	5.85	5.75	5.95	0.00
Aroclor-1016-5	(5)	6.15	6.14	6.04	6.24	-0.01
Aroclor-1260-1	(1)	7.27	7.26	7.16	7.36	0.00
Aroclor-1260-2	(2)	7.52	7.51	7.41	7.61	-0.01
Aroclor-1260-3	(3)	7.88	7.87	7.77	7.97	-0.01
Aroclor-1260-4	(4)	8.10	8.09	7.99	8.19	-0.01
Aroclor-1260-5	(5)	8.42	8.41	8.31	8.51	-0.01
Tetrachloro-m-xylene		4.52	4.51	4.41	4.61	-0.01
Decachlorobiphenyl		10.24	10.23	10.13	10.33	-0.01

Raw Data: **PP071151.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM **Case No.:** Q1730 **SAS No.:** Q1730 **SDG NO.:** Q1730

Continuing Calib Date: 04/08/2025 **Initial Calibration Date(s):** 03/27/2025 03/27/2025

Continuing Calib Time: 15:29 **Initial Calibration Time(s):** 10:30 17:53

GC Column: ZB-MR2 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	RT WINDOW TO	DIFF RT
Aroclor-1016-1	(1)	4.90	4.90	4.80	5.00	0.00
Aroclor-1016-2	(2)	4.92	4.92	4.82	5.02	0.00
Aroclor-1016-3	(3)	5.10	5.10	5.00	5.20	0.00
Aroclor-1016-4	(4)	5.14	5.14	5.04	5.24	0.00
Aroclor-1016-5	(5)	5.36	5.36	5.26	5.46	0.00
Aroclor-1260-1	(1)	6.39	6.39	6.29	6.49	0.00
Aroclor-1260-2	(2)	6.58	6.58	6.48	6.68	0.00
Aroclor-1260-3	(3)	6.73	6.73	6.63	6.83	0.00
Aroclor-1260-4	(4)	7.20	7.21	7.11	7.31	0.01
Aroclor-1260-5	(5)	7.45	7.45	7.35	7.55	0.00
Tetrachloro-m-xylene		3.82	3.82	3.72	3.92	0.00
Decachlorobiphenyl		8.86	8.86	8.76	8.96	0.00

Raw Data: **PP071151.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM **Case No.:** Q1730 **SAS No.:** Q1730 **SDG NO.:** Q1730

GC Column: ZB-MR1 **ID:** 0.32 **(mm)** **Initi. Calib. Date(s):** 03/27/2025 03/27/2025

Client Sample No.: CCAL07 **Date Analyzed:** 04/08/2025

Lab Sample No.: AR1660CCC500 **Data File :** PP071151.D **Time Analyzed:** 15:29

COMPOUND	RT	RT WINDOW FROM	RT WINDOW TO	CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.671	5.563	5.763	484.620	500.000	-3.1
Aroclor-1016-2	5.692	5.585	5.785	482.920	500.000	-3.4
Aroclor-1016-3	5.755	5.648	5.848	470.770	500.000	-5.8
Aroclor-1016-4	5.852	5.745	5.945	505.800	500.000	1.2
Aroclor-1016-5	6.146	6.038	6.238	510.370	500.000	2.1
Aroclor-1260-1	7.265	7.156	7.356	508.140	500.000	1.6
Aroclor-1260-2	7.519	7.410	7.610	516.880	500.000	3.4
Aroclor-1260-3	7.879	7.768	7.968	504.190	500.000	0.8
Aroclor-1260-4	8.103	7.992	8.192	470.830	500.000	-5.8
Aroclor-1260-5	8.424	8.311	8.511	486.490	500.000	-2.7
Decachlorobiphenyl	10.243	10.125	10.325	49.190	50.000	-1.6
Tetrachloro-m-xylene	4.516	4.412	4.612	50.490	50.000	1.0

Raw Data: **PP071151.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM **Case No.:** Q1730 **SAS No.:** Q1730 **SDG NO.:** Q1730

GC Column: ZB-MR2 **ID:** 0.32 **(mm)** **Initi. Calib. Date(s):** 03/27/2025 03/27/2025

Client Sample No.: CCAL07 **Date Analyzed:** 04/08/2025

Lab Sample No.: AR1660CCC500 **Data File :** PP071151.D **Time Analyzed:** 15:29

COMPOUND	RT	RT WINDOW FROM	RT WINDOW TO	CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.902	4.803	5.003	520.460	500.000	4.1
Aroclor-1016-2	4.921	4.822	5.022	507.900	500.000	1.6
Aroclor-1016-3	5.098	4.999	5.199	543.550	500.000	8.7
Aroclor-1016-4	5.140	5.041	5.241	556.220	500.000	11.2
Aroclor-1016-5	5.355	5.255	5.455	535.660	500.000	7.1
Aroclor-1260-1	6.390	6.292	6.492	521.040	500.000	4.2
Aroclor-1260-2	6.579	6.480	6.680	514.470	500.000	2.9
Aroclor-1260-3	6.732	6.634	6.834	529.510	500.000	5.9
Aroclor-1260-4	7.204	7.105	7.305	494.430	500.000	-1.1
Aroclor-1260-5	7.446	7.346	7.546	448.730	500.000	-10.3
Decachlorobiphenyl	8.859	8.759	8.959	40.800	50.000	-18.4
Tetrachloro-m-xylene	3.816	3.716	3.916	47.690	50.000	-4.6

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040825\
Data File : PP071151.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Apr 2025 15:29
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 15:40:28 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 28 03:13:46 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1) SA	Tetrachlo...	4.516	3.816	73793328	49435615	50.489	47.688
2) SA	Decachlor...	10.243	8.859	51773119	32161650	49.185	40.804
	Target Compounds						
3) L1	AR-1016-1	5.671	4.902	24643519	20864142	484.620	520.456
4) L1	AR-1016-2	5.692	4.921	37693781	29389805	482.922	507.902
5) L1	AR-1016-3	5.755	5.098	22208174	16808495	470.765	543.554
6) L1	AR-1016-4	5.852	5.140	18590980	13547316	505.798	556.223
7) L1	AR-1016-5	6.146	5.355	17307697	17313298	510.369	535.664
31) L7	AR-1260-1	7.265	6.390	34122856	26680779	508.137	521.045
32) L7	AR-1260-2	7.519	6.579	49691070	32819802	516.881	514.467
33) L7	AR-1260-3	7.879	6.732	40293738	28851540	504.187	529.514
34) L7	AR-1260-4	8.103	7.204	37071987	23646683	470.834	494.429
35) L7	AR-1260-5	8.424	7.446	77297206	55871530	486.494	448.730


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040825\
Data File : PP071151.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Apr 2025 15:29
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 15:40:28 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 28 03:13:46 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.516 min
Delta R.T.: 0.004 min
Response: 73793328
Conc: 50.49 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.816 min
Delta R.T.: 0.000 min
Response: 49435615
Conc: 47.69 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.243 min
Delta R.T.: 0.018 min
Response: 51773119
Conc: 49.19 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.859 min
Delta R.T.: 0.000 min
Response: 32161650
Conc: 40.80 ng/ml



#3 AR-1016-1


R.T.: 5.671 min
Delta R.T.: 0.007 min
Response: 24643519
Conc: 484.62 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.902 min
Delta R.T.: 0.000 min
Response: 20864142
Conc: 520.46 ng/ml




#4 AR-1016-2

R.T.: 5.692 min
Delta R.T.: 0.007 min
Response: 37693781
Conc: 482.92 ng/ml



#4 AR-1016-2


R.T.: 4.921 min
Delta R.T.: 0.000 min
Response: 29389805
Conc: 507.90 ng/ml



#5 AR-1016-3


R.T.: 5.755 min
Delta R.T.: 0.007 min
Response: 22208174
Conc: 470.77 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 5.098 min
Delta R.T.: 0.000 min
Response: 16808495
Conc: 543.55 ng/ml




#6 AR-1016-4

R.T.: 5.852 min
Delta R.T.: 0.007 min
Response: 18590980
Conc: 505.80 ng/ml



#6 AR-1016-4


R.T.: 5.140 min
Delta R.T.: 0.000 min
Response: 13547316
Conc: 556.22 ng/ml



#7 AR-1016-5


R.T.: 6.146 min
Delta R.T.: 0.008 min
Response: 17307697
Conc: 510.37 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.355 min
Delta R.T.: 0.000 min
Response: 17313298
Conc: 535.66 ng/ml




#31 AR-1260-1

R.T.: 7.265 min
Delta R.T.: 0.009 min
Response: 34122856
Conc: 508.14 ng/ml



#31 AR-1260-1


R.T.: 6.390 min
Delta R.T.: -0.002 min
Response: 26680779
Conc: 521.04 ng/ml



#32 AR-1260-2


R.T.: 7.519 min
Delta R.T.: 0.010 min
Response: 49691070
Conc: 516.88 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.579 min
Delta R.T.: 0.000 min
Response: 32819802
Conc: 514.47 ng/ml




#33 AR-1260-3

R.T.: 7.879 min
Delta R.T.: 0.011 min
Response: 40293738
Conc: 504.19 ng/ml



#33 AR-1260-3


R.T.: 6.732 min
Delta R.T.: -0.002 min
Response: 28851540
Conc: 529.51 ng/ml



#34 AR-1260-4


R.T.: 8.103 min
Delta R.T.: 0.012 min
Response: 37071987
Conc: 470.83 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.204 min
Delta R.T.: 0.000 min
Response: 23646683
Conc: 494.43 ng/ml



#35 AR-1260-5

R.T.: 8.424 min
Delta R.T.: 0.013 min
Response: 77297206
Conc: 486.49 ng/ml



#35 AR-1260-5

R.T.: 7.446 min
Delta R.T.: 0.000 min
Response: 55871530
Conc: 448.73 ng/ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Analytical Sequence

Client: Nobis Group **SDG No.:** Q1730

Project: Raymark Superfund Site **Instrument ID:** ECD_O

GC Column: ZB-MR1 **ID:** 0.32 (mm) **Inst. Calib. Date(s):** 03/18/2025 03/18/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES, AND STANDARDS IS GIVEN BELOW:

EPA SAMPLE NO.	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
I.BLK	I.BLK	03/18/2025	13:44	PO109971.D	8.74	3.69
AR1660ICC1000	AR1660ICC1000	03/18/2025	14:03	PO109972.D	8.75	3.69
AR1660ICC750	AR1660ICC750	03/18/2025	14:21	PO109973.D	8.75	3.69
AR1660ICC500	AR1660ICC500	03/18/2025	14:39	PO109974.D	8.75	3.69
AR1660ICC250	AR1660ICC250	03/18/2025	14:58	PO109975.D	8.75	3.69
AR1660ICC050	AR1660ICC050	03/18/2025	15:16	PO109976.D	8.74	3.69
AR1221ICC500	AR1221ICC500	03/18/2025	15:34	PO109977.D	8.74	3.69
AR1232ICC500	AR1232ICC500	03/18/2025	15:53	PO109978.D	8.74	3.69
AR1242ICC1000	AR1242ICC1000	03/18/2025	16:11	PO109979.D	8.74	3.69
AR1242ICC750	AR1242ICC750	03/18/2025	16:30	PO109980.D	8.74	3.69
AR1242ICC500	AR1242ICC500	03/18/2025	16:48	PO109981.D	8.74	3.69
AR1242ICC250	AR1242ICC250	03/18/2025	17:06	PO109982.D	8.74	3.69
AR1242ICC050	AR1242ICC050	03/18/2025	17:25	PO109983.D	8.74	3.69
AR1248ICC1000	AR1248ICC1000	03/18/2025	17:42	PO109984.D	8.74	3.69
AR1248ICC750	AR1248ICC750	03/18/2025	18:00	PO109985.D	8.74	3.69
AR1248ICC500	AR1248ICC500	03/18/2025	18:19	PO109986.D	8.74	3.69
AR1248ICC250	AR1248ICC250	03/18/2025	18:37	PO109987.D	8.74	3.69
AR1248ICC050	AR1248ICC050	03/18/2025	18:54	PO109988.D	8.74	3.69
AR1254ICC1000	AR1254ICC1000	03/18/2025	19:13	PO109989.D	8.74	3.69
AR1254ICC750	AR1254ICC750	03/18/2025	19:31	PO109990.D	8.74	3.69
AR1254ICC500	AR1254ICC500	03/18/2025	19:49	PO109991.D	8.74	3.69
AR1254ICC250	AR1254ICC250	03/18/2025	20:07	PO109992.D	8.74	3.69
AR1254ICC050	AR1254ICC050	03/18/2025	20:25	PO109993.D	8.74	3.69
AR1262ICC500	AR1262ICC500	03/18/2025	20:43	PO109994.D	8.74	3.69
AR1268ICC1000	AR1268ICC1000	03/18/2025	21:02	PO109995.D	8.74	3.69
AR1268ICC750	AR1268ICC750	03/18/2025	21:20	PO109996.D	8.74	3.69
AR1268ICC500	AR1268ICC500	03/18/2025	21:39	PO109997.D	8.74	3.69
AR1268ICC250	AR1268ICC250	03/18/2025	21:57	PO109998.D	8.74	3.69
AR1268ICC050	AR1268ICC050	03/18/2025	22:15	PO109999.D	8.74	3.69
AR1660CCC500	AR1660CCC500	04/07/2025	09:46	PO110263.D	8.74	3.69
I.BLK	I.BLK	04/07/2025	11:36	PO110267.D	8.74	3.69
OU4-VSL-15-040325	Q1730-01	04/07/2025	13:38	PO110269.D	8.74	3.69
OU4-VSL-16-040325	Q1730-03	04/07/2025	13:56	PO110270.D	8.74	3.69
OU4-VSL-16-040325MS	Q1730-03MS	04/07/2025	14:15	PO110271.D	8.74	3.69
OU4-VSL-16-040325MSD	Q1730-03MSD	04/07/2025	14:33	PO110272.D	8.74	3.69
AR1660CCC500	AR1660CCC500	04/07/2025	17:14	PO110278.D	8.74	3.69
I.BLK	I.BLK	04/07/2025	19:04	PO110282.D	8.74	3.69
PB167469BL	PB167469BL	04/07/2025	19:22	PO110283.D	8.74	3.69
PB167469BS	PB167469BS	04/07/2025	19:41	PO110284.D	8.74	3.69
AR1660CCC500	AR1660CCC500	04/07/2025	21:27	PO110286.D	8.74	3.69
I.BLK	I.BLK	04/07/2025	23:17	PO110290.D	8.74	3.69
AR1660CCC500	AR1660CCC500	04/08/2025	09:08	PO110292.D	8.74	3.69

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Analytical Sequence

I.BLK	I.BLK	04/08/2025	10:21	PO110296.D	8.74	3.69
OU4-VSL-17-040325	Q1730-05	04/08/2025	10:39	PO110297.D	8.74	3.69
OU4-PCS-TC-21-040325	Q1730-07	04/08/2025	10:58	PO110298.D	8.74	3.69
OU4-PCS-TC-22-040325	Q1730-09	04/08/2025	11:16	PO110299.D	8.74	3.69
OU4-PCS-TC-23-040325	Q1730-11	04/08/2025	11:35	PO110300.D	8.74	3.69
OU4-PCS-TC-24-040325	Q1730-13	04/08/2025	11:53	PO110301.D	8.74	3.69
AR1660CCC500	AR1660CCC500	04/08/2025	12:58	PO110302.D	8.74	3.69
I.BLK	I.BLK	04/08/2025	14:12	PO110306.D	8.74	3.69
I.BLK	I.BLK	03/27/2025	10:14	PP070916.D	10.23	4.51
AR1660ICC1000	AR1660ICC1000	03/27/2025	10:30	PP070917.D	10.23	4.51
AR1660ICC750	AR1660ICC750	03/27/2025	10:47	PP070918.D	10.23	4.52
AR1660ICC500	AR1660ICC500	03/27/2025	11:03	PP070919.D	10.23	4.51
AR1660ICC250	AR1660ICC250	03/27/2025	11:19	PP070920.D	10.23	4.52
AR1660ICC050	AR1660ICC050	03/27/2025	11:35	PP070921.D	10.23	4.51
AR1221ICC500	AR1221ICC500	03/27/2025	11:52	PP070922.D	10.23	4.52
AR1232ICC500	AR1232ICC500	03/27/2025	12:11	PP070923.D	10.23	4.52
AR1242ICC1000	AR1242ICC1000	03/27/2025	12:28	PP070924.D	10.23	4.51
AR1242ICC750	AR1242ICC750	03/27/2025	12:44	PP070925.D	10.23	4.51
AR1242ICC500	AR1242ICC500	03/27/2025	13:00	PP070926.D	10.23	4.51
AR1242ICC250	AR1242ICC250	03/27/2025	13:16	PP070927.D	10.23	4.51
AR1242ICC050	AR1242ICC050	03/27/2025	13:33	PP070928.D	10.23	4.51
AR1248ICC1000	AR1248ICC1000	03/27/2025	13:49	PP070929.D	10.23	4.51
AR1248ICC750	AR1248ICC750	03/27/2025	14:05	PP070930.D	10.23	4.51
AR1248ICC500	AR1248ICC500	03/27/2025	14:22	PP070931.D	10.23	4.51
AR1248ICC250	AR1248ICC250	03/27/2025	14:38	PP070932.D	10.23	4.52
AR1248ICC050	AR1248ICC050	03/27/2025	14:55	PP070933.D	10.23	4.51
AR1254ICC1000	AR1254ICC1000	03/27/2025	15:11	PP070934.D	10.23	4.51
AR1254ICC750	AR1254ICC750	03/27/2025	15:27	PP070935.D	10.23	4.51
AR1254ICC500	AR1254ICC500	03/27/2025	15:43	PP070936.D	10.23	4.51
AR1254ICC250	AR1254ICC250	03/27/2025	16:00	PP070937.D	10.23	4.51
AR1254ICC050	AR1254ICC050	03/27/2025	16:16	PP070938.D	10.23	4.52
AR1262ICC500	AR1262ICC500	03/27/2025	16:32	PP070939.D	10.23	4.51
AR1268ICC1000	AR1268ICC1000	03/27/2025	16:48	PP070940.D	10.23	4.52
AR1268ICC750	AR1268ICC750	03/27/2025	17:05	PP070941.D	10.23	4.51
AR1268ICC500	AR1268ICC500	03/27/2025	17:21	PP070942.D	10.23	4.51
AR1268ICC250	AR1268ICC250	03/27/2025	17:37	PP070943.D	10.23	4.52
AR1268ICC050	AR1268ICC050	03/27/2025	17:53	PP070944.D	10.23	4.51
AR1660CCC500	AR1660CCC500	04/08/2025	09:37	PP071136.D	10.26	4.52
I.BLK	I.BLK	04/08/2025	11:05	PP071140.D	10.25	4.52
OU4-PCS-TC-25-040325	Q1730-15	04/08/2025	11:39	PP071142.D	10.24	4.51
OU4-PCS-TC-26-040325	Q1730-17	04/08/2025	11:56	PP071143.D	10.25	4.52
OU4-CF-15-040325	Q1730-19	04/08/2025	12:12	PP071144.D	10.24	4.52
AR1660CCC500	AR1660CCC500	04/08/2025	15:29	PP071151.D	10.24	4.52
I.BLK	I.BLK	04/08/2025	16:51	PP071155.D	10.24	4.52

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Analytical Sequence

Client: Nobis Group **SDG No.:** Q1730

Project: Raymark Superfund Site **Instrument ID:** ECD_O

GC Column: ZB-MR2 **ID:** 0.32 (mm) **Inst. Calib. Date(s):** 03/18/2025 03/18/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES, AND STANDARDS IS GIVEN BELOW:

EPA SAMPLE NO.	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
I.BLK	I.BLK	03/18/2025	13:44	PO109971.D	8.70	3.69
AR1660ICC1000	AR1660ICC1000	03/18/2025	14:03	PO109972.D	8.70	3.69
AR1660ICC750	AR1660ICC750	03/18/2025	14:21	PO109973.D	8.70	3.69
AR1660ICC500	AR1660ICC500	03/18/2025	14:39	PO109974.D	8.70	3.69
AR1660ICC250	AR1660ICC250	03/18/2025	14:58	PO109975.D	8.70	3.69
AR1660ICC050	AR1660ICC050	03/18/2025	15:16	PO109976.D	8.70	3.69
AR1221ICC500	AR1221ICC500	03/18/2025	15:34	PO109977.D	8.70	3.69
AR1232ICC500	AR1232ICC500	03/18/2025	15:53	PO109978.D	8.70	3.69
AR1242ICC1000	AR1242ICC1000	03/18/2025	16:11	PO109979.D	8.70	3.69
AR1242ICC750	AR1242ICC750	03/18/2025	16:30	PO109980.D	8.70	3.69
AR1242ICC500	AR1242ICC500	03/18/2025	16:48	PO109981.D	8.70	3.69
AR1242ICC250	AR1242ICC250	03/18/2025	17:06	PO109982.D	8.70	3.69
AR1242ICC050	AR1242ICC050	03/18/2025	17:25	PO109983.D	8.70	3.69
AR1248ICC1000	AR1248ICC1000	03/18/2025	17:42	PO109984.D	8.70	3.69
AR1248ICC750	AR1248ICC750	03/18/2025	18:00	PO109985.D	8.70	3.69
AR1248ICC500	AR1248ICC500	03/18/2025	18:19	PO109986.D	8.70	3.69
AR1248ICC250	AR1248ICC250	03/18/2025	18:37	PO109987.D	8.70	3.69
AR1248ICC050	AR1248ICC050	03/18/2025	18:54	PO109988.D	8.70	3.69
AR1254ICC1000	AR1254ICC1000	03/18/2025	19:13	PO109989.D	8.69	3.69
AR1254ICC750	AR1254ICC750	03/18/2025	19:31	PO109990.D	8.70	3.69
AR1254ICC500	AR1254ICC500	03/18/2025	19:49	PO109991.D	8.70	3.69
AR1254ICC250	AR1254ICC250	03/18/2025	20:07	PO109992.D	8.70	3.69
AR1254ICC050	AR1254ICC050	03/18/2025	20:25	PO109993.D	8.69	3.69
AR1262ICC500	AR1262ICC500	03/18/2025	20:43	PO109994.D	8.70	3.69
AR1268ICC1000	AR1268ICC1000	03/18/2025	21:02	PO109995.D	8.70	3.69
AR1268ICC750	AR1268ICC750	03/18/2025	21:20	PO109996.D	8.70	3.69
AR1268ICC500	AR1268ICC500	03/18/2025	21:39	PO109997.D	8.69	3.69
AR1268ICC250	AR1268ICC250	03/18/2025	21:57	PO109998.D	8.70	3.69
AR1268ICC050	AR1268ICC050	03/18/2025	22:15	PO109999.D	8.70	3.69
AR1660CCC500	AR1660CCC500	04/07/2025	09:46	PO110263.D	8.69	3.69
I.BLK	I.BLK	04/07/2025	11:36	PO110267.D	8.69	3.69
OU4-VSL-15-040325	Q1730-01	04/07/2025	13:38	PO110269.D	8.69	3.69
OU4-VSL-16-040325	Q1730-03	04/07/2025	13:56	PO110270.D	8.69	3.69
OU4-VSL-16-040325MS	Q1730-03MS	04/07/2025	14:15	PO110271.D	8.69	3.69
OU4-VSL-16-040325MSD	Q1730-03MSD	04/07/2025	14:33	PO110272.D	8.69	3.69
AR1660CCC500	AR1660CCC500	04/07/2025	17:14	PO110278.D	8.69	3.69
I.BLK	I.BLK	04/07/2025	19:04	PO110282.D	8.69	3.69
PB167469BL	PB167469BL	04/07/2025	19:22	PO110283.D	8.69	3.69
PB167469BS	PB167469BS	04/07/2025	19:41	PO110284.D	8.69	3.69
AR1660CCC500	AR1660CCC500	04/07/2025	21:27	PO110286.D	8.69	3.69
I.BLK	I.BLK	04/07/2025	23:17	PO110290.D	8.69	3.69
AR1660CCC500	AR1660CCC500	04/08/2025	09:08	PO110292.D	8.69	3.69

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Analytical Sequence

I.BLK	I.BLK	04/08/2025	10:21	PO110296.D	8.69	3.69
OU4-VSL-17-040325	Q1730-05	04/08/2025	10:39	PO110297.D	8.69	3.69
OU4-PCS-TC-21-040325	Q1730-07	04/08/2025	10:58	PO110298.D	8.69	3.69
OU4-PCS-TC-22-040325	Q1730-09	04/08/2025	11:16	PO110299.D	8.69	3.69
OU4-PCS-TC-23-040325	Q1730-11	04/08/2025	11:35	PO110300.D	8.69	3.69
OU4-PCS-TC-24-040325	Q1730-13	04/08/2025	11:53	PO110301.D	8.69	3.69
AR1660CCC500	AR1660CCC500	04/08/2025	12:58	PO110302.D	8.69	3.69
I.BLK	I.BLK	04/08/2025	14:12	PO110306.D	8.69	3.69
I.BLK	I.BLK	03/27/2025	10:14	PP070916.D	8.86	3.82
AR1660ICC1000	AR1660ICC1000	03/27/2025	10:30	PP070917.D	8.86	3.82
AR1660ICC750	AR1660ICC750	03/27/2025	10:47	PP070918.D	8.86	3.82
AR1660ICC500	AR1660ICC500	03/27/2025	11:03	PP070919.D	8.86	3.82
AR1660ICC250	AR1660ICC250	03/27/2025	11:19	PP070920.D	8.86	3.82
AR1660ICC050	AR1660ICC050	03/27/2025	11:35	PP070921.D	8.86	3.82
AR1221ICC500	AR1221ICC500	03/27/2025	11:52	PP070922.D	8.86	3.82
AR1232ICC500	AR1232ICC500	03/27/2025	12:11	PP070923.D	8.86	3.82
AR1242ICC1000	AR1242ICC1000	03/27/2025	12:28	PP070924.D	8.86	3.82
AR1242ICC750	AR1242ICC750	03/27/2025	12:44	PP070925.D	8.86	3.82
AR1242ICC500	AR1242ICC500	03/27/2025	13:00	PP070926.D	8.86	3.82
AR1242ICC250	AR1242ICC250	03/27/2025	13:16	PP070927.D	8.86	3.82
AR1242ICC050	AR1242ICC050	03/27/2025	13:33	PP070928.D	8.86	3.82
AR1248ICC1000	AR1248ICC1000	03/27/2025	13:49	PP070929.D	8.86	3.82
AR1248ICC750	AR1248ICC750	03/27/2025	14:05	PP070930.D	8.86	3.82
AR1248ICC500	AR1248ICC500	03/27/2025	14:22	PP070931.D	8.86	3.82
AR1248ICC250	AR1248ICC250	03/27/2025	14:38	PP070932.D	8.86	3.82
AR1248ICC050	AR1248ICC050	03/27/2025	14:55	PP070933.D	8.86	3.82
AR1254ICC1000	AR1254ICC1000	03/27/2025	15:11	PP070934.D	8.86	3.82
AR1254ICC750	AR1254ICC750	03/27/2025	15:27	PP070935.D	8.86	3.82
AR1254ICC500	AR1254ICC500	03/27/2025	15:43	PP070936.D	8.86	3.82
AR1254ICC250	AR1254ICC250	03/27/2025	16:00	PP070937.D	8.86	3.82
AR1254ICC050	AR1254ICC050	03/27/2025	16:16	PP070938.D	8.86	3.82
AR1262ICC500	AR1262ICC500	03/27/2025	16:32	PP070939.D	8.86	3.82
AR1268ICC1000	AR1268ICC1000	03/27/2025	16:48	PP070940.D	8.86	3.82
AR1268ICC750	AR1268ICC750	03/27/2025	17:05	PP070941.D	8.86	3.82
AR1268ICC500	AR1268ICC500	03/27/2025	17:21	PP070942.D	8.86	3.82
AR1268ICC250	AR1268ICC250	03/27/2025	17:37	PP070943.D	8.86	3.82
AR1268ICC050	AR1268ICC050	03/27/2025	17:53	PP070944.D	8.86	3.82
AR1660CCC500	AR1660CCC500	04/08/2025	09:37	PP071136.D	8.86	3.82
I.BLK	I.BLK	04/08/2025	11:05	PP071140.D	8.86	3.82
OU4-PCS-TC-25-040325	Q1730-15	04/08/2025	11:39	PP071142.D	8.86	3.82
OU4-PCS-TC-26-040325	Q1730-17	04/08/2025	11:56	PP071143.D	8.86	3.82
OU4-CF-15-040325	Q1730-19	04/08/2025	12:12	PP071144.D	8.86	3.82
AR1660CCC500	AR1660CCC500	04/08/2025	15:29	PP071151.D	8.86	3.82
I.BLK	I.BLK	04/08/2025	16:51	PP071155.D	8.86	3.82

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

OU4-VSL-16-040325MS

Contract: NOBI03

Lab Code: CHEM Case No.: Q1730 SAS No.: Q1730 SDG NO.: Q1730

Lab Sample ID: Q1730-03MS Date(s) Analyzed: 04/07/2025 04/07/2025

Instrument ID (1): ECD_O Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column:(2): ZB-MR2 ID: 0.32 (mm)

Data file PO110271.D

ANALYTE	COL	RT	RT WINDOW FROM	RT WINDOW TO	CONCENTRATION	MEAN CONCENTRATION	%RPD
Aroclor-1016	1	4.782	4.732	4.832	171		
	2	4.802	4.752	4.852	172		
	3	4.858	4.808	4.908	172		
COLUMN 1	4	4.978	4.928	5.028	171		
	5	5.236	5.186	5.286	163	170	
	1	4.768	4.718	4.818	166		
	2	4.787	4.737	4.837	166		
	3	4.963	4.913	5.013	166		
	4	5.005	4.955	5.055	163		
COLUMN 2	5	5.218	5.168	5.268	159	164	3.59
Aroclor-1260	1	6.276	6.226	6.326	178		
	2	6.466	6.416	6.516	169		
	3	6.833	6.783	6.883	144		
COLUMN 1	4	7.094	7.044	7.144	152		
	5	7.335	7.285	7.385	141	157	
	1	6.25	6.2	6.3	164		
	2	6.438	6.388	6.488	160		
	3	6.591	6.541	6.641	155		
	4	7.062	7.012	7.112	142		
COLUMN 2	5	7.303	7.253	7.353	133	151	3.9

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.
OU4-VSL-16-040325MSI

Contract: NOBI03

Lab Code: CHEM Case No.: Q1730 SAS No.: Q1730 SDG NO.: Q1730

Lab Sample ID: Q1730-03MSD Date(s) Analyzed: 04/07/2025 04/07/2025

Instrument ID (1): ECD_O Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column:(2): ZB-MR2 ID: 0.32 (mm)

Data file PO110272.D

ANALYTE	COL	RT	RT WINDOW FROM	RT WINDOW TO	CONCENTRATION	MEAN CONCENTRATION	%RPD
Aroclor-1016	1	4.782	4.732	4.832	163		
	2	4.801	4.751	4.851	167		
	3	4.857	4.807	4.907	167		
COLUMN 1	4	4.978	4.928	5.028	166		
	5	5.235	5.185	5.285	159	164	
	1	4.768	4.718	4.818	159		
	2	4.788	4.738	4.838	160		
	3	4.963	4.913	5.013	160		
	4	5.005	4.955	5.055	157		
COLUMN 2	5	5.217	5.167	5.267	154	158	3.73
Aroclor-1260	1	6.276	6.226	6.326	169		
	2	6.466	6.416	6.516	163		
	3	6.833	6.783	6.883	143		
COLUMN 1	4	7.093	7.043	7.143	150		
	5	7.335	7.285	7.385	138	153	
	1	6.25	6.2	6.3	159		
	2	6.438	6.388	6.488	156		
	3	6.59	6.54	6.64	150		
	4	7.061	7.011	7.111	138		
COLUMN 2	5	7.302	7.252	7.352	129	146	4.68

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PB167469BS

Contract: NOBI03

Lab Code: CHEM Case No.: Q1730 SAS No.: Q1730 SDG NO.: Q1730

Lab Sample ID: PB167469BS Date(s) Analyzed: 04/07/2025 04/07/2025

Instrument ID (1): ECD_O Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column:(2): ZB-MR2 ID: 0.32 (mm)

Data file PO110284.D

ANALYTE	COL	RT	RT WINDOW FROM	RT WINDOW TO	CONCENTRATION	MEAN CONCENTRATION	%RPD
Aroclor-1016	1	4.782	4.732	4.832	141		
	2	4.802	4.752	4.852	143		
	3	4.858	4.808	4.908	142		
COLUMN 1	4	4.978	4.928	5.028	142		
	5	5.236	5.186	5.286	138	141	
	1	4.769	4.719	4.819	136		
	2	4.788	4.738	4.838	134		
	3	4.964	4.914	5.014	136		
	4	5.005	4.955	5.055	135		
COLUMN 2	5	5.218	5.168	5.268	131	134	5.09
Aroclor-1260	1	6.277	6.227	6.327	149		
	2	6.465	6.415	6.515	141		
	3	6.833	6.783	6.883	119		
COLUMN 1	4	7.093	7.043	7.143	131		
	5	7.336	7.286	7.386	129	134	
	1	6.25	6.2	6.3	140		
	2	6.438	6.388	6.488	138		
	3	6.59	6.54	6.64	133		
	4	7.062	7.012	7.112	119		
COLUMN 2	5	7.303	7.253	7.353	113	128	4.58

QC SAMPLE DATA

Raw Data: **PO110283.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Nobis Group		Date Collected:		
Project:	Raymark Superfund Site		Date Received:		
Client Sample ID:	PB167469BL		SDG No.:	Q1730	
Lab Sample ID:	PB167469BL		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	100	Decanted:
Sample Wt/Vol:	30.01 Units:	g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110283.D	1	04/07/25 08:20	04/07/25 19:22	PB167469

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units(Dry Weight)
TARGETS							
12674-11-2	Aroclor-1016	8.30	U	3.90	8.30	17.0	ug/kg
11104-28-2	Aroclor-1221	13.0	U	4.00	13.0	17.0	ug/kg
11141-16-5	Aroclor-1232	8.30	U	3.70	8.30	17.0	ug/kg
53469-21-9	Aroclor-1242	8.30	U	4.00	8.30	17.0	ug/kg
12672-29-6	Aroclor-1248	13.0	U	5.90	13.0	17.0	ug/kg
11097-69-1	Aroclor-1254	8.30	U	3.20	8.30	17.0	ug/kg
37324-23-5	Aroclor-1262	13.0	U	5.00	13.0	17.0	ug/kg
11100-14-4	Aroclor-1268	8.30	U	3.60	8.30	17.0	ug/kg
11096-82-5	Aroclor-1260	8.30	U	3.20	8.30	17.0	ug/kg
SURROGATES							
877-09-8	Tetrachloro-m-xylene	19.0		44 - 130		95%	SPK: 20
2051-24-3	Decachlorobiphenyl	17.3		60 - 125		86%	SPK: 20

Comments:

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Quantitation Report (QT Reviewed)

Sample Results: **PO110283.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040725\
Data File : PO110283.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Apr 2025 19:22
Operator : YP/AJ
Sample : PB167469BL
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB167469BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 01:35:40 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.691	3.688	173.1E6	95749355	19.020	18.259
2)	SA Decachlor...	8.741	8.691	133.0E6	38862706	17.297	16.019
	Target Compounds						

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Sample Results: **PO110283.D**


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040725\
Data File : PO110283.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Apr 2025 19:22
Operator : YP/AJ
Sample : PB167469BL
Misc :
ALS Vial : 17 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
PB167469BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 01:35:40 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.691 min
Delta R.T.: 0.000 min
Response: 173109300
Conc: 19.02 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB167469BL




#1 Tetrachloro-m-xylene

R.T.: 3.688 min
Delta R.T.: -0.002 min
Response: 95749355
Conc: 18.26 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.741 min
Delta R.T.: -0.005 min
Response: 132989663
Conc: 17.30 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.691 min
Delta R.T.: -0.004 min
Response: 38862706
Conc: 16.02 ng/ml

Raw Data: **PO109971.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Nobis Group	Date Collected:	03/18/25
Project:	Raymark Superfund Site	Date Received:	03/18/25
Client Sample ID:	PIBLK-PO109971.D	SDG No.:	Q1730
Lab Sample ID:	I.BLK-PO109971.D	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0 Decanted:
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	5030		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO109971.D	1		03/18/25	PO031925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
12674-11-2	Aroclor-1016	0.25	U	0.097	0.25	0.50	ug/L
11104-28-2	Aroclor-1221	0.40	U	0.13	0.40	0.50	ug/L
11141-16-5	Aroclor-1232	0.25	U	0.096	0.25	0.50	ug/L
53469-21-9	Aroclor-1242	0.25	U	0.12	0.25	0.50	ug/L
12672-29-6	Aroclor-1248	0.25	U	0.071	0.25	0.50	ug/L
11097-69-1	Aroclor-1254	0.25	U	0.094	0.25	0.50	ug/L
11096-82-5	Aroclor-1260	0.25	U	0.081	0.25	0.50	ug/L
37324-23-5	Aroclor-1262	0.40	U	0.14	0.40	0.50	ug/L
11100-14-4	Aroclor-1268	0.25	U	0.11	0.25	0.50	ug/L
SURROGATES							
877-09-8	Tetrachloro-m-xylene	18.5		60 - 140		93%	SPK: 20
2051-24-3	Decachlorobiphenyl	20.0		60 - 140		100%	SPK: 20

Comments:

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Sample Results: **PO109971.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109971.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 13:44
Operator : YP/AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 03:44:56 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.692	3.689	168.3E6	97603621	18.495	18.612
2)	SA Decachlor...	8.744	8.695	153.9E6	51783076	20.014	21.344
	Target Compounds						

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Sample Results: **PO109971.D**


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109971.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 13:44
Operator : YP/AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 03:44:56 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.692 min
Delta R.T.: 0.000 min
Response: 168336707
Conc: 18.50 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.689 min
Delta R.T.: 0.000 min
Response: 97603621
Conc: 18.61 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.744 min
Delta R.T.: -0.001 min
Response: 153882781
Conc: 20.01 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.695 min
Delta R.T.: -0.001 min
Response: 51783076
Conc: 21.34 ng/ml

Raw Data: **PO110267.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Nobis Group	Date Collected:	04/07/25
Project:	Raymark Superfund Site	Date Received:	04/07/25
Client Sample ID:	PIBLK-PO110267.D	SDG No.:	Q1730
Lab Sample ID:	I.BLK-PO110267.D	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0 Decanted:
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	5030		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110267.D	1		04/07/25	PO040725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
12674-11-2	Aroclor-1016	0.25	U	0.097	0.25	0.50	ug/L
11104-28-2	Aroclor-1221	0.40	U	0.13	0.40	0.50	ug/L
11141-16-5	Aroclor-1232	0.25	U	0.096	0.25	0.50	ug/L
53469-21-9	Aroclor-1242	0.25	U	0.12	0.25	0.50	ug/L
12672-29-6	Aroclor-1248	0.25	U	0.071	0.25	0.50	ug/L
11097-69-1	Aroclor-1254	0.25	U	0.094	0.25	0.50	ug/L
11096-82-5	Aroclor-1260	0.25	U	0.081	0.25	0.50	ug/L
37324-23-5	Aroclor-1262	0.40	U	0.14	0.40	0.50	ug/L
11100-14-4	Aroclor-1268	0.25	U	0.11	0.25	0.50	ug/L
SURROGATES							
877-09-8	Tetrachloro-m-xylene	18.8		60 - 140		94%	SPK: 20
2051-24-3	Decachlorobiphenyl	15.9		60 - 140		79%	SPK: 20

Comments:

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Quantitation Report (QT Reviewed)

Sample Results: **PO110267.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040725\
Data File : PO110267.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Apr 2025 11:36
Operator : YP/AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 07 12:31:58 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.689	3.687	177.4E6	98699299	19.494	18.821
2)	SA Decachlor...	8.737	8.690	141.7E6	38504154	18.426	15.871
	Target Compounds						

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Sample Results: **PO110267.D**


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040725\
Data File : PO110267.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Apr 2025 11:36
Operator : YP/AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 07 12:31:58 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.689 min
Delta R.T.: -0.003 min
Response: 177426113
Conc: 19.49 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.687 min
Delta R.T.: -0.003 min
Response: 98699299
Conc: 18.82 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.737 min
Delta R.T.: -0.008 min
Response: 141670453
Conc: 18.43 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.690 min
Delta R.T.: -0.006 min
Response: 38504154
Conc: 15.87 ng/ml

Raw Data: **PO110282.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Nobis Group	Date Collected:	04/07/25
Project:	Raymark Superfund Site	Date Received:	04/07/25
Client Sample ID:	PIBLK-PO110282.D	SDG No.:	Q1730
Lab Sample ID:	I.BLK-PO110282.D	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0 Decanted:
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	5030		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110282.D	1		04/07/25	PO040725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
12674-11-2	Aroclor-1016	0.25	U	0.097	0.25	0.50	ug/L
11104-28-2	Aroclor-1221	0.40	U	0.13	0.40	0.50	ug/L
11141-16-5	Aroclor-1232	0.25	U	0.096	0.25	0.50	ug/L
53469-21-9	Aroclor-1242	0.25	U	0.12	0.25	0.50	ug/L
12672-29-6	Aroclor-1248	0.25	U	0.071	0.25	0.50	ug/L
11097-69-1	Aroclor-1254	0.25	U	0.094	0.25	0.50	ug/L
11096-82-5	Aroclor-1260	0.25	U	0.081	0.25	0.50	ug/L
37324-23-5	Aroclor-1262	0.40	U	0.14	0.40	0.50	ug/L
11100-14-4	Aroclor-1268	0.25	U	0.11	0.25	0.50	ug/L
SURROGATES							
877-09-8	Tetrachloro-m-xylene	18.8		60 - 140		94%	SPK: 20
2051-24-3	Decachlorobiphenyl	16.2		60 - 140		81%	SPK: 20

Comments:

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Quantitation Report (QT Reviewed)

Sample Results: **PO110282.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040725\
Data File : PO110282.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Apr 2025 19:04
Operator : YP/AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 01:35:20 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.691	3.688	176.0E6	98785740	19.334	18.838
2)	SA Decachlor...	8.740	8.692	132.1E6	39381036	17.185	16.232
	Target Compounds						


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040725\
Data File : PO110282.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Apr 2025 19:04
Operator : YP/AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 01:35:20 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.691 min
Delta R.T.: -0.001 min
Response: 175966498
Conc: 19.33 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.688 min
Delta R.T.: -0.002 min
Response: 98785740
Conc: 18.84 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.740 min
Delta R.T.: -0.005 min
Response: 132132646
Conc: 17.19 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.692 min
Delta R.T.: -0.004 min
Response: 39381036
Conc: 16.23 ng/ml

Raw Data: **PO110290.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Nobis Group	Date Collected:	04/07/25
Project:	Raymark Superfund Site	Date Received:	04/07/25
Client Sample ID:	PIBLK-PO110290.D	SDG No.:	Q1730
Lab Sample ID:	I.BLK-PO110290.D	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0 Decanted:
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	5030		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110290.D	1		04/07/25	PO040725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
12674-11-2	Aroclor-1016	0.25	U	0.097	0.25	0.50	ug/L
11104-28-2	Aroclor-1221	0.40	U	0.13	0.40	0.50	ug/L
11141-16-5	Aroclor-1232	0.25	U	0.096	0.25	0.50	ug/L
53469-21-9	Aroclor-1242	0.25	U	0.12	0.25	0.50	ug/L
12672-29-6	Aroclor-1248	0.25	U	0.071	0.25	0.50	ug/L
11097-69-1	Aroclor-1254	0.25	U	0.094	0.25	0.50	ug/L
11096-82-5	Aroclor-1260	0.25	U	0.081	0.25	0.50	ug/L
37324-23-5	Aroclor-1262	0.40	U	0.14	0.40	0.50	ug/L
11100-14-4	Aroclor-1268	0.25	U	0.11	0.25	0.50	ug/L
SURROGATES							
877-09-8	Tetrachloro-m-xylene	18.9		60 - 140		94%	SPK: 20
2051-24-3	Decachlorobiphenyl	16.2		60 - 140		81%	SPK: 20

Comments:

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Quantitation Report (QT Reviewed)

Sample Results: **PO110290.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040725\
Data File : PO110290.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Apr 2025 23:17
Operator : YP/AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 01:38:04 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.691	3.687	177.5E6	98959783	19.499	18.871
2)	SA Decachlor...	8.742	8.692	136.3E6	39361941	17.725	16.224
	Target Compounds						

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Sample Results: **PO110290.D**


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040725\
Data File : PO110290.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Apr 2025 23:17
Operator : YP/AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 01:38:04 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.691 min
Delta R.T.: -0.001 min
Response: 177471328
Conc: 19.50 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.687 min
Delta R.T.: -0.002 min
Response: 98959783
Conc: 18.87 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.742 min
Delta R.T.: -0.004 min
Response: 136281296
Conc: 17.72 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.692 min
Delta R.T.: -0.004 min
Response: 39361941
Conc: 16.22 ng/ml

Raw Data: **PO110296.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Nobis Group	Date Collected:	04/08/25
Project:	Raymark Superfund Site	Date Received:	04/08/25
Client Sample ID:	PIBLK-PO110296.D	SDG No.:	Q1730
Lab Sample ID:	I.BLK-PO110296.D	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0 Decanted:
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	5030		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110296.D	1		04/08/25	PO040825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
12674-11-2	Aroclor-1016	0.25	U	0.097	0.25	0.50	ug/L
11104-28-2	Aroclor-1221	0.40	U	0.13	0.40	0.50	ug/L
11141-16-5	Aroclor-1232	0.25	U	0.096	0.25	0.50	ug/L
53469-21-9	Aroclor-1242	0.25	U	0.12	0.25	0.50	ug/L
12672-29-6	Aroclor-1248	0.25	U	0.071	0.25	0.50	ug/L
11097-69-1	Aroclor-1254	0.25	U	0.094	0.25	0.50	ug/L
11096-82-5	Aroclor-1260	0.25	U	0.081	0.25	0.50	ug/L
37324-23-5	Aroclor-1262	0.40	U	0.14	0.40	0.50	ug/L
11100-14-4	Aroclor-1268	0.25	U	0.11	0.25	0.50	ug/L
SURROGATES							
877-09-8	Tetrachloro-m-xylene	17.7		60 - 140		89%	SPK: 20
2051-24-3	Decachlorobiphenyl	15.4		60 - 140		77%	SPK: 20

Comments:

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Quantitation Report (QT Reviewed)

Sample Results: **PO110296.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040825\
Data File : PO110296.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Apr 2025 10:21
Operator : YP/AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 10:44:40 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.688	3.687	169.9E6	93006368	18.670	17.735
2)	SA Decachlor...	8.738	8.690	136.7E6	37456258	17.778	15.439
	Target Compounds						


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040825\
Data File : PO110296.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Apr 2025 10:21
Operator : YP/AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 10:44:40 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.688 min
Delta R.T.: -0.004 min
Response: 169928217
Conc: 18.67 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.687 min
Delta R.T.: -0.003 min
Response: 93006368
Conc: 17.74 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.738 min
Delta R.T.: -0.008 min
Response: 136688438
Conc: 17.78 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.690 min
Delta R.T.: -0.005 min
Response: 37456258
Conc: 15.44 ng/ml

Raw Data: **PO110306.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Nobis Group	Date Collected:	04/08/25
Project:	Raymark Superfund Site	Date Received:	04/08/25
Client Sample ID:	PIBLK-PO110306.D	SDG No.:	Q1730
Lab Sample ID:	I.BLK-PO110306.D	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0 Decanted:
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	5030		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110306.D	1		04/08/25	PO040825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
12674-11-2	Aroclor-1016	0.25	U	0.097	0.25	0.50	ug/L
11104-28-2	Aroclor-1221	0.40	U	0.13	0.40	0.50	ug/L
11141-16-5	Aroclor-1232	0.25	U	0.096	0.25	0.50	ug/L
53469-21-9	Aroclor-1242	0.25	U	0.12	0.25	0.50	ug/L
12672-29-6	Aroclor-1248	0.25	U	0.071	0.25	0.50	ug/L
11097-69-1	Aroclor-1254	0.25	U	0.094	0.25	0.50	ug/L
11096-82-5	Aroclor-1260	0.25	U	0.081	0.25	0.50	ug/L
37324-23-5	Aroclor-1262	0.40	U	0.14	0.40	0.50	ug/L
11100-14-4	Aroclor-1268	0.25	U	0.11	0.25	0.50	ug/L
SURROGATES							
877-09-8	Tetrachloro-m-xylene	17.4		60 - 140		87%	SPK: 20
2051-24-3	Decachlorobiphenyl	15.3		60 - 140		77%	SPK: 20

Comments:

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Quantitation Report (QT Reviewed)

Sample Results: **PO110306.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040825\
Data File : PO110306.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Apr 2025 14:12
Operator : YP/AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 14:20:54 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.690	3.687	165.6E6	91042032	18.193	17.361
2)	SA Decachlor...	8.739	8.691	141.3E6	37124521	18.373	15.302
	Target Compounds						


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040825\
Data File : PO110306.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Apr 2025 14:12
Operator : YP/AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 14:20:54 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.690 min
Delta R.T.: -0.002 min
Response: 165584494
Conc: 18.19 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.687 min
Delta R.T.: -0.002 min
Response: 91042032
Conc: 17.36 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.739 min
Delta R.T.: -0.006 min
Response: 141264866
Conc: 18.37 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.691 min
Delta R.T.: -0.005 min
Response: 37124521
Conc: 15.30 ng/ml

Raw Data: **PP070916.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Nobis Group	Date Collected:	03/27/25
Project:	Raymark Superfund Site	Date Received:	03/27/25
Client Sample ID:	PIBLK-PP070916.D	SDG No.:	Q1730
Lab Sample ID:	I.BLK-PP070916.D	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0 Decanted:
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	5030		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP070916.D	1		03/27/25	PP032725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
12674-11-2	Aroclor-1016	0.25	U	0.097	0.25	0.50	ug/L
11104-28-2	Aroclor-1221	0.40	U	0.13	0.40	0.50	ug/L
11141-16-5	Aroclor-1232	0.25	U	0.096	0.25	0.50	ug/L
53469-21-9	Aroclor-1242	0.25	U	0.12	0.25	0.50	ug/L
12672-29-6	Aroclor-1248	0.25	U	0.071	0.25	0.50	ug/L
11097-69-1	Aroclor-1254	0.25	U	0.094	0.25	0.50	ug/L
11096-82-5	Aroclor-1260	0.25	U	0.081	0.25	0.50	ug/L
37324-23-5	Aroclor-1262	0.40	U	0.14	0.40	0.50	ug/L
11100-14-4	Aroclor-1268	0.25	U	0.11	0.25	0.50	ug/L
SURROGATES							
877-09-8	Tetrachloro-m-xylene	18.1		60 - 140		91%	SPK: 20
2051-24-3	Decachlorobiphenyl	19.4		60 - 140		97%	SPK: 20

Comments:

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Quantitation Report (QT Reviewed)

Sample Results: **PP070916.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070916.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 10:14
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 28 06:35:09 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 28 03:13:46 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.514	3.816	27521844	18794888	18.830	18.131
2)	SA Decachlor...	10.229	8.858	20399611	18404317	19.380	23.350
	Target Compounds						


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070916.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 10:14
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 28 06:35:09 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 28 03:13:46 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.514 min
Delta R.T.: 0.002 mi
Response: 27521844
Conc: 18.83 ng/m

Instrument :
ECD_P
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.816 min
Delta R.T.: 0.000 min
Response: 18794888
Conc: 18.13 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.229 min
Delta R.T.: 0.003 min
Response: 20399611
Conc: 19.38 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.858 min
Delta R.T.: 0.000 min
Response: 18404317
Conc: 23.35 ng/ml

Raw Data: **PP071140.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Nobis Group	Date Collected:	04/08/25
Project:	Raymark Superfund Site	Date Received:	04/08/25
Client Sample ID:	PIBLK-PP071140.D	SDG No.:	Q1730
Lab Sample ID:	I.BLK-PP071140.D	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0 Decanted:
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	5030		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071140.D	1		04/08/25	PP040825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
12674-11-2	Aroclor-1016	0.25	U	0.097	0.25	0.50	ug/L
11104-28-2	Aroclor-1221	0.40	U	0.13	0.40	0.50	ug/L
11141-16-5	Aroclor-1232	0.25	U	0.096	0.25	0.50	ug/L
53469-21-9	Aroclor-1242	0.25	U	0.12	0.25	0.50	ug/L
12672-29-6	Aroclor-1248	0.25	U	0.071	0.25	0.50	ug/L
11097-69-1	Aroclor-1254	0.25	U	0.094	0.25	0.50	ug/L
11096-82-5	Aroclor-1260	0.25	U	0.081	0.25	0.50	ug/L
37324-23-5	Aroclor-1262	0.40	U	0.14	0.40	0.50	ug/L
11100-14-4	Aroclor-1268	0.25	U	0.11	0.25	0.50	ug/L
SURROGATES							
877-09-8	Tetrachloro-m-xylene	22.8		60 - 140		114%	SPK: 20
2051-24-3	Decachlorobiphenyl	21.5		60 - 140		108%	SPK: 20

Comments:

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Quantitation Report (QT Reviewed)

Sample Results: **PP071140.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040825\
Data File : PP071140.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Apr 2025 11:05
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 11:17:07 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 28 03:13:46 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.518	3.815	33330770	23969414	22.805	23.122
2)	SA Decachlor...	10.246	8.859	25105022	16966968	23.850	21.526
	Target Compounds						


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040825\
Data File : PP071140.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Apr 2025 11:05
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 11:17:07 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 28 03:13:46 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.518 min
Delta R.T.: 0.006 min
Response: 33330770
Conc: 22.80 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.815 min
Delta R.T.: 0.000 min
Response: 23969414
Conc: 23.12 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.246 min
Delta R.T.: 0.020 min
Response: 25105022
Conc: 23.85 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.859 min
Delta R.T.: 0.000 min
Response: 16966968
Conc: 21.53 ng/ml

Raw Data: **PP071155.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Nobis Group	Date Collected:	04/08/25
Project:	Raymark Superfund Site	Date Received:	04/08/25
Client Sample ID:	PIBLK-PP071155.D	SDG No.:	Q1730
Lab Sample ID:	I.BLK-PP071155.D	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0 Decanted:
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	5030		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071155.D	1		04/08/25	PP040825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
12674-11-2	Aroclor-1016	0.25	U	0.097	0.25	0.50	ug/L
11104-28-2	Aroclor-1221	0.40	U	0.13	0.40	0.50	ug/L
11141-16-5	Aroclor-1232	0.25	U	0.096	0.25	0.50	ug/L
53469-21-9	Aroclor-1242	0.25	U	0.12	0.25	0.50	ug/L
12672-29-6	Aroclor-1248	0.25	U	0.071	0.25	0.50	ug/L
11097-69-1	Aroclor-1254	0.25	U	0.094	0.25	0.50	ug/L
11096-82-5	Aroclor-1260	0.25	U	0.081	0.25	0.50	ug/L
37324-23-5	Aroclor-1262	0.40	U	0.14	0.40	0.50	ug/L
11100-14-4	Aroclor-1268	0.25	U	0.11	0.25	0.50	ug/L
SURROGATES							
877-09-8	Tetrachloro-m-xylene	20.6		60 - 140		103%	SPK: 20
2051-24-3	Decachlorobiphenyl	18.4		60 - 140		92%	SPK: 20

Comments:

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Quantitation Report (QT Reviewed)

Sample Results: **PP071155.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040825\
Data File : PP071155.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Apr 2025 16:51
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 17:04:32 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 28 03:13:46 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.516	3.816	30070321	21846080	20.574	21.074
2)	SA Decachlor...	10.242	8.859	22060594	14463728	20.958	18.350
	Target Compounds						


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040825\
Data File : PP071155.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Apr 2025 16:51
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 17:04:32 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 28 03:13:46 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.516 min
Delta R.T.: 0.004 min
Response: 30070321
Conc: 20.57 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.816 min
Delta R.T.: 0.000 min
Response: 21846080
Conc: 21.07 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.242 min
Delta R.T.: 0.016 min
Response: 22060594
Conc: 20.96 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.859 min
Delta R.T.: 0.000 min
Response: 14463728
Conc: 18.35 ng/ml

Raw Data: **PO110284.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Nobis Group	Date Collected:	
Project:	Raymark Superfund Site	Date Received:	
Client Sample ID:	PB167469BS	SDG No.:	Q1730
Lab Sample ID:	PB167469BS	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	100 Decanted:
Sample Wt/Vol:	30.02 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110284.D	1	04/07/25 08:20	04/07/25 19:41	PB167469

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units(Dry Weight)
TARGETS							
12674-11-2	Aroclor-1016	141		3.90	8.30	17.0	ug/kg
11104-28-2	Aroclor-1221	13.0	U	4.00	13.0	17.0	ug/kg
11141-16-5	Aroclor-1232	8.30	U	3.70	8.30	17.0	ug/kg
53469-21-9	Aroclor-1242	8.30	U	4.00	8.30	17.0	ug/kg
12672-29-6	Aroclor-1248	13.0	U	5.90	13.0	17.0	ug/kg
11097-69-1	Aroclor-1254	8.30	U	3.20	8.30	17.0	ug/kg
37324-23-5	Aroclor-1262	13.0	U	5.00	13.0	17.0	ug/kg
11100-14-4	Aroclor-1268	8.30	U	3.60	8.30	17.0	ug/kg
11096-82-5	Aroclor-1260	134		3.20	8.30	17.0	ug/kg
SURROGATES							
877-09-8	Tetrachloro-m-xylene	19.9		44 - 130		99%	SPK: 20
2051-24-3	Decachlorobiphenyl	19.1		60 - 125		95%	SPK: 20

Comments:

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040725\
Data File : PO110284.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Apr 2025 19:41
Operator : YP/AJ
Sample : PB167469BS
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB167469BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 01:36:05 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.691	3.688	180.6E6	99143578	19.847	18.906
2)	SA Decachlor...	8.740	8.691	146.7E6	41863016	19.080	17.255
	Target Compounds						
3)	L1 AR-1016-1	4.782	4.769	142.0E6	74923264	421.865	408.847
4)	L1 AR-1016-2	4.802	4.788	199.0E6	105.5E6	429.173	403.452
5)	L1 AR-1016-3	4.858	4.964	137.1E6	57705339	427.778	407.010
6)	L1 AR-1016-4	4.978	5.005	107.7E6	47781418	426.066	405.168
7)	L1 AR-1016-5	5.236	5.218	113.5E6	60210398	412.871	393.741
31)	L7 AR-1260-1	6.277	6.250	210.2E6	107.2E6	446.044	419.635
32)	L7 AR-1260-2	6.465	6.438	251.9E6	124.0E6	422.920	413.030
33)	L7 AR-1260-3	6.833	6.590	173.8E6	115.3E6	358.056	399.028
34)	L7 AR-1260-4	7.093	7.062	166.9E6	76919777	393.513	357.636
35)	L7 AR-1260-5	7.336	7.303	405.0E6	168.5E6	387.245	337.822


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040725\
Data File : PO110284.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Apr 2025 19:41
Operator : YP/AJ
Sample : PB167469BS
Misc :
ALS Vial : 18 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
PB167469BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 01:36:05 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.691 min
Delta R.T.: -0.001 min
Response: 180638940
Conc: 19.85 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB167469BS




#1 Tetrachloro-m-xylene

R.T.: 3.688 min
Delta R.T.: -0.002 min
Response: 99143578
Conc: 18.91 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.740 min
Delta R.T.: -0.006 min
Response: 146698953
Conc: 19.08 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.691 min
Delta R.T.: -0.005 min
Response: 41863016
Conc: 17.26 ng/ml



#3 AR-1016-1


R.T.: 4.782 min
Delta R.T.: -0.003 min
Response: 141976321
Conc: 421.87 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB167469BS




#3 AR-1016-1

R.T.: 4.769 min
Delta R.T.: -0.003 min
Response: 74923264
Conc: 408.85 ng/ml




#4 AR-1016-2

R.T.: 4.802 min
Delta R.T.: -0.003 min
Response: 199001347
Conc: 429.17 ng/ml



#4 AR-1016-2


R.T.: 4.788 min
Delta R.T.: -0.003 min
Response: 105504424
Conc: 403.45 ng/ml



#5 AR-1016-3


R.T.: 4.858 min
Delta R.T.: -0.003 min
Response: 137139634
Conc: 427.78 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB167469BS




#5 AR-1016-3

R.T.: 4.964 min
Delta R.T.: -0.003 min
Response: 57705339
Conc: 407.01 ng/ml




#6 AR-1016-4

R.T.: 4.978 min
Delta R.T.: -0.003 min
Response: 107662297
Conc: 426.07 ng/ml



#6 AR-1016-4


R.T.: 5.005 min
Delta R.T.: -0.004 min
Response: 47781418
Conc: 405.17 ng/ml



#7 AR-1016-5


R.T.: 5.236 min
Delta R.T.: -0.003 min
Response: 113484662
Conc: 412.87 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB167469BS




#7 AR-1016-5

R.T.: 5.218 min
Delta R.T.: -0.004 min
Response: 60210398
Conc: 393.74 ng/ml




#31 AR-1260-1

R.T.: 6.277 min
Delta R.T.: -0.003 min
Response: 210243094
Conc: 446.04 ng/ml



#31 AR-1260-1


R.T.: 6.250 min
Delta R.T.: -0.005 min
Response: 107192380
Conc: 419.64 ng/ml



#32 AR-1260-2


R.T.: 6.465 min
Delta R.T.: -0.004 mi
Response: 251857024
Conc: 422.92 ng/m

Instrument :
ECD_O
ClientSampleId :
PB167469BS




#32 AR-1260-2

R.T.: 6.438 min
Delta R.T.: -0.005 min
Response: 124026961
Conc: 413.03 ng/ml




#33 AR-1260-3

R.T.: 6.833 min
Delta R.T.: -0.004 min
Response: 173765242
Conc: 358.06 ng/ml



#33 AR-1260-3


R.T.: 6.590 min
Delta R.T.: -0.005 min
Response: 115311965
Conc: 399.03 ng/ml



#34 AR-1260-4


R.T.: 7.093 min
Delta R.T.: -0.004 min
Response: 166915626
Conc: 393.51 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB167469BS




#34 AR-1260-4

R.T.: 7.062 min
Delta R.T.: -0.005 min
Response: 76919777
Conc: 357.64 ng/ml



#35 AR-1260-5

R.T.: 7.336 min
Delta R.T.: -0.004 min
Response: 404998251
Conc: 387.24 ng/ml



#35 AR-1260-5

R.T.: 7.303 min
Delta R.T.: -0.005 min
Response: 168470492
Conc: 337.82 ng/ml

Raw Data: **PO110271.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Nobis Group	Date Collected:	04/03/25
Project:	Raymark Superfund Site	Date Received:	04/04/25
Client Sample ID:	OU4-VSL-16-040325MS	SDG No.:	Q1730
Lab Sample ID:	Q1730-03MS	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	94.5 Decanted:
Sample Wt/Vol:	30.05 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110271.D	1	04/07/25 08:20	04/07/25 14:15	PB167469

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units(Dry Weight)
TARGETS							
12674-11-2	Aroclor-1016	170		4.20	8.80	18.0	ug/kg
11104-28-2	Aroclor-1221	13.7	U	4.30	13.7	18.0	ug/kg
11141-16-5	Aroclor-1232	8.80	U	3.90	8.80	18.0	ug/kg
53469-21-9	Aroclor-1242	8.80	U	4.20	8.80	18.0	ug/kg
12672-29-6	Aroclor-1248	13.7	U	6.30	13.7	18.0	ug/kg
11097-69-1	Aroclor-1254	8.80	U	3.40	8.80	18.0	ug/kg
37324-23-5	Aroclor-1262	13.7	U	5.30	13.7	18.0	ug/kg
11100-14-4	Aroclor-1268	8.80	U	3.80	8.80	18.0	ug/kg
11096-82-5	Aroclor-1260	157		3.40	8.80	18.0	ug/kg
SURROGATES							
877-09-8	Tetrachloro-m-xylene	22.8		44 - 130		114%	SPK: 20
2051-24-3	Decachlorobiphenyl	20.9		60 - 125		104%	SPK: 20

Comments:

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Quantitation Report (QT Reviewed)

Sample Results: **PO110271.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040725\
Data File : PO110271.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Apr 2025 14:15
Operator : YP/AJ
Sample : Q1730-03MS
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
OU4-VSL-16-040325MS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 07 14:28:33 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.690	3.687	207.8E6	114.5E6	22.829	21.843
2)	SA Decachlor...	8.740	8.691	160.6E6	43661585	20.882	17.997
	Target Compounds						
3)	L1 AR-1016-1	4.782	4.768	163.4E6	86371640	485.641	471.320
4)	L1 AR-1016-2	4.802	4.787	226.9E6	123.3E6	489.395	471.405
5)	L1 AR-1016-3	4.858	4.963	156.9E6	66829797	489.339	471.367
6)	L1 AR-1016-4	4.978	5.005	122.7E6	54685236	485.623	463.710
7)	L1 AR-1016-5	5.236	5.218	127.1E6	69082264	462.402	451.758
31)	L7 AR-1260-1	6.276	6.250	237.9E6	119.0E6	504.637	465.860
32)	L7 AR-1260-2	6.466	6.438	285.8E6	136.5E6	479.876	454.662
33)	L7 AR-1260-3	6.833	6.591	199.0E6	126.8E6	409.960	438.894
34)	L7 AR-1260-4	7.094	7.062	183.3E6	86758507	432.156	403.381
35)	L7 AR-1260-5	7.335	7.303	418.2E6	188.7E6	399.895	378.330


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040725\
Data File : PO110271.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Apr 2025 14:15
Operator : YP/AJ
Sample : Q1730-03MS
Misc :
ALS Vial : 10 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
OU4-VSL-16-040325MS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 07 14:28:33 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.690 min
Delta R.T.: -0.002 min
Response: 207776497
Conc: 22.83 ng/ml

Instrument :
ECD_O
ClientSampleId :
OU4-VSL-16-040325MS




#1 Tetrachloro-m-xylene

R.T.: 3.687 min
Delta R.T.: -0.003 min
Response: 114548641
Conc: 21.84 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.740 min
Delta R.T.: -0.005 min
Response: 160555921
Conc: 20.88 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.691 min
Delta R.T.: -0.004 min
Response: 43661585
Conc: 18.00 ng/ml



#3 AR-1016-1


R.T.: 4.782 min
Delta R.T.: -0.003 min
Response: 163439503
Conc: 485.64 ng/ml

Instrument :
ECD_O
ClientSampleId :
OU4-VSL-16-040325MS




#3 AR-1016-1

R.T.: 4.768 min
Delta R.T.: -0.003 min
Response: 86371640
Conc: 471.32 ng/ml




#4 AR-1016-2

R.T.: 4.802 min
Delta R.T.: -0.003 min
Response: 226925626
Conc: 489.40 ng/ml



#4 AR-1016-2


R.T.: 4.787 min
Delta R.T.: -0.004 min
Response: 123274570
Conc: 471.41 ng/ml



#5 AR-1016-3


R.T.: 4.858 min
Delta R.T.: -0.003 min
Response: 156875297
Conc: 489.34 ng/ml

Instrument :
ECD_O
ClientSampleId :
OU4-VSL-16-040325MS




#5 AR-1016-3

R.T.: 4.963 min
Delta R.T.: -0.003 min
Response: 66829797
Conc: 471.37 ng/ml




#6 AR-1016-4

R.T.: 4.978 min
Delta R.T.: -0.003 min
Response: 122711855
Conc: 485.62 ng/ml



#6 AR-1016-4


R.T.: 5.005 min
Delta R.T.: -0.004 min
Response: 54685236
Conc: 463.71 ng/ml



#7 AR-1016-5


R.T.: 5.236 min
Delta R.T.: -0.003 min
Response: 127098994
Conc: 462.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
OU4-VSL-16-040325MS




#7 AR-1016-5

R.T.: 5.218 min
Delta R.T.: -0.004 min
Response: 69082264
Conc: 451.76 ng/ml




#31 AR-1260-1

R.T.: 6.276 min
Delta R.T.: -0.004 min
Response: 237861115
Conc: 504.64 ng/ml



#31 AR-1260-1


R.T.: 6.250 min
Delta R.T.: -0.005 min
Response: 119000242
Conc: 465.86 ng/ml



#32 AR-1260-2


R.T.: 6.466 min
Delta R.T.: -0.003 min
Response: 285775432
Conc: 479.88 ng/ml

Instrument :
ECD_O
ClientSampleId :
OU4-VSL-16-040325MS




#32 AR-1260-2

R.T.: 6.438 min
Delta R.T.: -0.005 min
Response: 136528501
Conc: 454.66 ng/ml




#33 AR-1260-3

R.T.: 6.833 min
Delta R.T.: -0.004 min
Response: 198954290
Conc: 409.96 ng/ml



#33 AR-1260-3


R.T.: 6.591 min
Delta R.T.: -0.005 min
Response: 126832599
Conc: 438.89 ng/ml



#34 AR-1260-4


R.T.: 7.094 min
Delta R.T.: -0.004 min
Response: 183306711
Conc: 432.16 ng/ml

Instrument :
ECD_O
ClientSampleId :
OU4-VSL-16-040325MS




#34 AR-1260-4

R.T.: 7.062 min
Delta R.T.: -0.005 min
Response: 86758507
Conc: 403.38 ng/ml



#35 AR-1260-5

R.T.: 7.335 min
Delta R.T.: -0.004 min
Response: 418228862
Conc: 399.90 ng/ml



#35 AR-1260-5

R.T.: 7.303 min
Delta R.T.: -0.004 min
Response: 188671824
Conc: 378.33 ng/ml

Raw Data: **PO110272.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Nobis Group	Date Collected:	04/03/25
Project:	Raymark Superfund Site	Date Received:	04/04/25
Client Sample ID:	OU4-VSL-16-040325MSD	SDG No.:	Q1730
Lab Sample ID:	Q1730-03MSD	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	94.5 Decanted:
Sample Wt/Vol:	30.02 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110272.D	1	04/07/25 08:20	04/07/25 14:33	PB167469

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units(Dry Weight)
TARGETS							
12674-11-2	Aroclor-1016	164		4.20	8.80	18.0	ug/kg
11104-28-2	Aroclor-1221	13.7	U	4.30	13.7	18.0	ug/kg
11141-16-5	Aroclor-1232	8.80	U	3.90	8.80	18.0	ug/kg
53469-21-9	Aroclor-1242	8.80	U	4.20	8.80	18.0	ug/kg
12672-29-6	Aroclor-1248	13.7	U	6.30	13.7	18.0	ug/kg
11097-69-1	Aroclor-1254	8.80	U	3.40	8.80	18.0	ug/kg
37324-23-5	Aroclor-1262	13.7	U	5.30	13.7	18.0	ug/kg
11100-14-4	Aroclor-1268	8.80	U	3.80	8.80	18.0	ug/kg
11096-82-5	Aroclor-1260	153		3.40	8.80	18.0	ug/kg
SURROGATES							
877-09-8	Tetrachloro-m-xylene	21.6		44 - 130		108%	SPK: 20
2051-24-3	Decachlorobiphenyl	19.8		60 - 125		99%	SPK: 20

Comments:

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040725\
Data File : PO110272.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Apr 2025 14:33
Operator : YP/AJ
Sample : Q1730-03MSD
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
OU4-VSL-16-040325MSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 07 14:58:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1) SA	Tetrachlo...	3.690	3.687	196.9E6	109.1E6	21.633	20.795
2) SA	Decachlor...	8.739	8.690	152.2E6	41848584	19.800	17.249
	Target Compounds						
3) L1	AR-1016-1	4.782	4.768	156.1E6	82580554	463.781	450.632
4) L1	AR-1016-2	4.801	4.788	219.3E6	118.8E6	472.954	454.443
5) L1	AR-1016-3	4.857	4.963	151.9E6	64490825	473.697	454.870
6) L1	AR-1016-4	4.978	5.005	118.8E6	52631687	470.083	446.297
7) L1	AR-1016-5	5.235	5.217	123.9E6	66656091	450.663	435.892
31) L7	AR-1260-1	6.276	6.250	225.9E6	115.3E6	479.366	451.204
32) L7	AR-1260-2	6.466	6.438	276.1E6	132.5E6	463.567	441.258
33) L7	AR-1260-3	6.833	6.590	196.5E6	122.9E6	404.805	425.130
34) L7	AR-1260-4	7.093	7.061	180.2E6	84267581	424.947	391.799
35) L7	AR-1260-5	7.335	7.302	409.9E6	182.8E6	391.944	366.611

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Sample Results: **PO110272.D**


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040725\
Data File : PO110272.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Apr 2025 14:33
Operator : YP/AJ
Sample : Q1730-03MSD
Misc :
ALS Vial : 11 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
OU4-VSL-16-040325MSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 07 14:58:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.690 min
Delta R.T.: -0.002 min
Response: 196895453
Conc: 21.63 ng/ml

Instrument :
ECD_O
ClientSampleId :
OU4-VSL-16-040325MSD




#1 Tetrachloro-m-xylene

R.T.: 3.687 min
Delta R.T.: -0.003 min
Response: 109050960
Conc: 20.79 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.739 min
Delta R.T.: -0.006 min
Response: 152232950
Conc: 19.80 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.690 min
Delta R.T.: -0.006 min
Response: 41848584
Conc: 17.25 ng/ml



#3 AR-1016-1


R.T.: 4.782 min
Delta R.T.: -0.003 min
Response: 156082810
Conc: 463.78 ng/ml

Instrument :
ECD_O
ClientSampleId :
OU4-VSL-16-040325MSD




#3 AR-1016-1

R.T.: 4.768 min
Delta R.T.: -0.004 min
Response: 82580554
Conc: 450.63 ng/ml




#4 AR-1016-2

R.T.: 4.801 min
Delta R.T.: -0.004 min
Response: 219302002
Conc: 472.95 ng/ml



#4 AR-1016-2


R.T.: 4.788 min
Delta R.T.: -0.004 min
Response: 118838816
Conc: 454.44 ng/ml



#5 AR-1016-3


R.T.: 4.857 min
Delta R.T.: -0.004 min
Response: 151860738
Conc: 473.70 ng/ml

Instrument :
ECD_O
ClientSampleId :
OU4-VSL-16-040325MSD




#5 AR-1016-3

R.T.: 4.963 min
Delta R.T.: -0.004 min
Response: 64490825
Conc: 454.87 ng/ml




#6 AR-1016-4

R.T.: 4.978 min
Delta R.T.: -0.004 min
Response: 118784963
Conc: 470.08 ng/ml



#6 AR-1016-4


R.T.: 5.005 min
Delta R.T.: -0.004 min
Response: 52631687
Conc: 446.30 ng/ml



#7 AR-1016-5


R.T.: 5.235 min
Delta R.T.: -0.004 min
Response: 123872515
Conc: 450.66 ng/ml

Instrument :
ECD_O
ClientSampleId :
OU4-VSL-16-040325MSD




#7 AR-1016-5

R.T.: 5.217 min
Delta R.T.: -0.005 min
Response: 66656091
Conc: 435.89 ng/ml




#31 AR-1260-1

R.T.: 6.276 min
Delta R.T.: -0.003 min
Response: 225949528
Conc: 479.37 ng/ml



#31 AR-1260-1


R.T.: 6.250 min
Delta R.T.: -0.005 min
Response: 115256374
Conc: 451.20 ng/ml



#32 AR-1260-2


R.T.: 6.466 min
Delta R.T.: -0.003 min
Response: 276063046
Conc: 463.57 ng/ml

Instrument :
ECD_O
ClientSampleId :
OU4-VSL-16-040325MSD




#32 AR-1260-2

R.T.: 6.438 min
Delta R.T.: -0.005 min
Response: 132503405
Conc: 441.26 ng/ml




#33 AR-1260-3

R.T.: 6.833 min
Delta R.T.: -0.004 min
Response: 196452347
Conc: 404.80 ng/ml



#33 AR-1260-3


R.T.: 6.590 min
Delta R.T.: -0.005 min
Response: 122855008
Conc: 425.13 ng/ml



#34 AR-1260-4


R.T.: 7.093 min
Delta R.T.: -0.005 min
Response: 180248679
Conc: 424.95 ng/ml

Instrument :
ECD_O
ClientSampleId :
OU4-VSL-16-040325MSD




#34 AR-1260-4

R.T.: 7.061 min
Delta R.T.: -0.006 min
Response: 84267581
Conc: 391.80 ng/ml



#35 AR-1260-5

R.T.: 7.335 min
Delta R.T.: -0.004 min
Response: 409913112
Conc: 391.94 ng/ml



#35 AR-1260-5

R.T.: 7.302 min
Delta R.T.: -0.005 min
Response: 182827616
Conc: 366.61 ng/ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	PO031925	Instrument	ECD_o

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1660ICC750	PO109973.D	AR-1260-5 #2	yogesh	3/19/2025 8:01:07 AM	mohammad	3/24/2025 3:02:46	Peak Integrated by Software
AR1660ICC050	PO109976.D	AR-1260-1	yogesh	3/19/2025 8:01:25 AM	mohammad	3/24/2025 3:02:46	Peak Integrated by Software
AR1660ICC050	PO109976.D	AR-1260-1 #2	yogesh	3/19/2025 8:01:25 AM	mohammad	3/24/2025 3:02:46	Peak Integrated by Software
AR1660ICC050	PO109976.D	AR-1260-3 #2	yogesh	3/19/2025 8:01:25 AM	mohammad	3/24/2025 3:02:46	Peak Integrated by Software
AR1242ICC050	PO109983.D	AR-1242-3 #2	yogesh	3/19/2025 8:01:09 AM	mohammad	3/24/2025 3:02:46	Peak Integrated by Software
AR1242ICC050	PO109983.D	AR-1242-4	yogesh	3/19/2025 8:01:09 AM	mohammad	3/24/2025 3:02:46	Peak Integrated by Software
AR1242ICC050	PO109983.D	AR-1242-5	yogesh	3/19/2025 8:01:09 AM	mohammad	3/24/2025 3:02:46	Peak Integrated by Software
AR1242ICC050	PO109983.D	AR-1242-5 #2	yogesh	3/19/2025 8:01:09 AM	mohammad	3/24/2025 3:02:46	Peak Integrated by Software
AR1242ICC050	PO109983.D	Tetrachloro-m-xylene	yogesh	3/19/2025 8:01:09 AM	mohammad	3/24/2025 3:02:46	Peak Integrated by Software
AR1242ICC050	PO109983.D	Tetrachloro-m-xylene #2	yogesh	3/19/2025 8:01:09 AM	mohammad	3/24/2025 3:02:46	Peak Integrated by Software
AR1248ICC050	PO109988.D	AR-1248-3	yogesh	3/19/2025 8:01:10 AM	mohammad	3/24/2025 3:02:46	Peak Integrated by Software
AR1248ICC050	PO109988.D	AR-1248-4 #2	yogesh	3/19/2025 8:01:10 AM	mohammad	3/24/2025 3:02:46	Peak Integrated by Software
AR1254ICC050	PO109993.D	AR-1254-5 #2	yogesh	3/19/2025 8:01:12 AM	mohammad	3/24/2025 3:02:46	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	PO031925	Instrument	ECD_o

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1268ICC500	PO109997.D	AR-1268-1	yogesh	3/19/2025 8:01:13 AM	mohammad	3/24/2025 3:02:46	Peak Integrated by Software
AR1268ICC250	PO109998.D	AR-1268-1	yogesh	3/19/2025 8:01:15 AM	mohammad	3/24/2025 3:02:46	Peak Integrated by Software
AR1268ICC050	PO109999.D	AR-1268-1	yogesh	3/19/2025 8:01:16 AM	mohammad	3/24/2025 3:02:46	Peak Integrated by Software
AR1268ICC050	PO109999.D	AR-1268-4	yogesh	3/19/2025 8:01:16 AM	mohammad	3/24/2025 3:02:46	Peak Integrated by Software
AR1268ICC050	PO109999.D	Tetrachloro-m-xylene	yogesh	3/19/2025 8:01:16 AM	mohammad	3/24/2025 3:02:46	Peak Integrated by Software
AR1268ICC050	PO109999.D	Tetrachloro-m-xylene #2	yogesh	3/19/2025 8:01:16 AM	mohammad	3/24/2025 3:02:46	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	PO040725	Instrument	ECD_o

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1660CCC500	PO110286.D	AR-1016-5	yogesh	4/8/2025 7:46:34 AM	mohammad	4/9/2025 2:57:57	Peak Integrated by Software
AR1660CCC500	PO110286.D	AR-1016-5 #2	yogesh	4/8/2025 7:46:34 AM	mohammad	4/9/2025 2:57:57	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	PO040825	Instrument	ECD_o

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1254CCC500	PO110295.D	AR-1254-1	yogesh	4/8/2025 2:56:17 PM	mohammad	4/10/2025 5:29:43	Peak Integrated by Software
AR1254CCC500	PO110295.D	AR-1254-2	yogesh	4/8/2025 2:56:17 PM	mohammad	4/10/2025 5:29:43	Peak Integrated by Software
AR1254CCC500	PO110295.D	AR-1254-4 #2	yogesh	4/8/2025 2:56:17 PM	mohammad	4/10/2025 5:29:43	Peak Integrated by Software

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	PP032725	Instrument	ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1660ICC250	PP070920.D	AR-1260-3	yogesh	3/28/2025 7:37:36 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1660ICC250	PP070920.D	AR-1260-4	yogesh	3/28/2025 7:37:36 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1660ICC250	PP070920.D	AR-1260-5	yogesh	3/28/2025 7:37:36 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1660ICC050	PP070921.D	AR-1016-5	yogesh	3/28/2025 7:37:38 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1660ICC050	PP070921.D	AR-1260-1	yogesh	3/28/2025 7:37:38 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1660ICC050	PP070921.D	AR-1260-3	yogesh	3/28/2025 7:37:38 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1660ICC050	PP070921.D	AR-1260-4	yogesh	3/28/2025 7:37:38 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1660ICC050	PP070921.D	AR-1260-5	yogesh	3/28/2025 7:37:38 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1232ICC500	PP070923.D	Decachlorobiphenyl	yogesh	3/28/2025 7:37:40 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1242ICC050	PP070928.D	AR-1242-1	yogesh	3/28/2025 7:37:42 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1242ICC050	PP070928.D	AR-1242-2	yogesh	3/28/2025 7:37:42 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1242ICC050	PP070928.D	AR-1242-2 #2	yogesh	3/28/2025 7:37:42 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1242ICC050	PP070928.D	AR-1242-3	yogesh	3/28/2025 7:37:42 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	PP032725	Instrument	ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1242ICC050	PP070928.D	AR-1242-4	yogesh	3/28/2025 7:37:42 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1242ICC050	PP070928.D	AR-1242-4 #2	yogesh	3/28/2025 7:37:42 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1242ICC050	PP070928.D	AR-1242-5	yogesh	3/28/2025 7:37:42 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1242ICC050	PP070928.D	AR-1242-5 #2	yogesh	3/28/2025 7:37:42 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1248ICC050	PP070933.D	AR-1248-2	yogesh	3/28/2025 7:37:44 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1248ICC050	PP070933.D	AR-1248-3	yogesh	3/28/2025 7:37:44 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1248ICC050	PP070933.D	AR-1248-4	yogesh	3/28/2025 7:37:44 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1248ICC050	PP070933.D	AR-1248-5	yogesh	3/28/2025 7:37:44 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1254ICC1000	PP070934.D	AR-1254-5	yogesh	3/28/2025 7:37:45 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1254ICC750	PP070935.D	AR-1254-5	yogesh	3/28/2025 7:37:47 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1254ICC250	PP070937.D	AR-1254-5	yogesh	3/28/2025 7:37:48 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1268ICC1000	PP070940.D	AR-1268-1 #2	yogesh	3/28/2025 7:37:50 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1268ICC750	PP070941.D	AR-1268-1 #2	yogesh	3/28/2025 7:37:51 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	PP032725	Instrument	ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1268ICC500	PP070942.D	AR-1268-1 #2	yogesh	3/28/2025 7:37:53 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1268ICC250	PP070943.D	AR-1268-1 #2	yogesh	3/28/2025 7:37:55 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1268ICC050	PP070944.D	AR-1268-1 #2	yogesh	3/28/2025 7:37:56 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1268ICC050	PP070944.D	AR-1268-4 #2	yogesh	3/28/2025 7:37:56 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1268ICC050	PP070944.D	Tetrachloro-m-xylene	yogesh	3/28/2025 7:37:56 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1254ICV500	PP070948.D	AR-1254-5	yogesh	3/28/2025 7:37:58 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	PP040825	Instrument	ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1660CCC500	PP071136.D	AR-1260-4	yogesh	4/9/2025 9:18:11 AM	mohammad	4/10/2025 5:29:57	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QCBatch ID # PO031925

Review By	yogesh	Review On	3/18/2025 3:19:21 PM		
Supervise By	mohammad	Supervise On	3/24/2025 3:02:46 AM		
SubDirectory	PO031925	HP Acquire Method		HP Processing Method	PO031925

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344 ,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP2435 9,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PO109970.D	18 Mar 2025 13:26	YP/AJ	Ok
2	I.BLK	PO109971.D	18 Mar 2025 13:44	YP/AJ	Ok
3	AR1660ICC1000	PO109972.D	18 Mar 2025 14:03	YP/AJ	Ok
4	AR1660ICC750	PO109973.D	18 Mar 2025 14:21	YP/AJ	Ok,M
5	AR1660ICC500	PO109974.D	18 Mar 2025 14:39	YP/AJ	Ok
6	AR1660ICC250	PO109975.D	18 Mar 2025 14:58	YP/AJ	Ok
7	AR1660ICC050	PO109976.D	18 Mar 2025 15:16	YP/AJ	Ok,M
8	AR1221ICC500	PO109977.D	18 Mar 2025 15:34	YP/AJ	Ok
9	AR1232ICC500	PO109978.D	18 Mar 2025 15:53	YP/AJ	Ok
10	AR1242ICC1000	PO109979.D	18 Mar 2025 16:11	YP/AJ	Ok
11	AR1242ICC750	PO109980.D	18 Mar 2025 16:30	YP/AJ	Ok
12	AR1242ICC500	PO109981.D	18 Mar 2025 16:48	YP/AJ	Ok
13	AR1242ICC250	PO109982.D	18 Mar 2025 17:06	YP/AJ	Ok
14	AR1242ICC050	PO109983.D	18 Mar 2025 17:25	YP/AJ	Ok,M
15	AR1248ICC1000	PO109984.D	18 Mar 2025 17:42	YP/AJ	Ok
16	AR1248ICC750	PO109985.D	18 Mar 2025 18:00	YP/AJ	Ok
17	AR1248ICC500	PO109986.D	18 Mar 2025 18:19	YP/AJ	Ok
18	AR1248ICC250	PO109987.D	18 Mar 2025 18:37	YP/AJ	Ok
19	AR1248ICC050	PO109988.D	18 Mar 2025 18:54	YP/AJ	Ok,M
20	AR1254ICC1000	PO109989.D	18 Mar 2025 19:13	YP/AJ	Ok
21	AR1254ICC750	PO109990.D	18 Mar 2025 19:31	YP/AJ	Ok

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QCBatch ID # PO031925

Review By	yogesh	Review On	3/18/2025 3:19:21 PM		
Supervise By	mohammad	Supervise On	3/24/2025 3:02:46 AM		
SubDirectory	PO031925	HP Acquire Method		HP Processing Method	PO031925

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344 ,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP2435 9,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

22	AR1254ICC500	PO109991.D	18 Mar 2025 19:49	YP/AJ	Ok
23	AR1254ICC250	PO109992.D	18 Mar 2025 20:07	YP/AJ	Ok
24	AR1254ICC050	PO109993.D	18 Mar 2025 20:25	YP/AJ	Ok,M
25	AR1262ICC500	PO109994.D	18 Mar 2025 20:43	YP/AJ	Ok
26	AR1268ICC1000	PO109995.D	18 Mar 2025 21:02	YP/AJ	Ok
27	AR1268ICC750	PO109996.D	18 Mar 2025 21:20	YP/AJ	Ok
28	AR1268ICC500	PO109997.D	18 Mar 2025 21:39	YP/AJ	Ok,M
29	AR1268ICC250	PO109998.D	18 Mar 2025 21:57	YP/AJ	Ok,M
30	AR1268ICC050	PO109999.D	18 Mar 2025 22:15	YP/AJ	Ok,M
31	PO031925ICV500	PO110000.D	18 Mar 2025 22:34	YP/AJ	Ok
32	AR1242ICV500	PO110001.D	18 Mar 2025 23:47	YP/AJ	Ok
33	AR1248ICV500	PO110002.D	19 Mar 2025 00:05	YP/AJ	Ok
34	AR1254ICV500	PO110003.D	19 Mar 2025 00:24	YP/AJ	Ok
35	AR1268ICV500	PO110004.D	19 Mar 2025 00:42	YP/AJ	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QCBatch ID # PO040725

Review By	yogesh	Review On	4/7/2025 11:16:09 AM		
Supervise By	mohammad	Supervise On	4/9/2025 2:57:57 AM		
SubDirectory	PO040725	HP Acquire Method		HP Processing Method	PO031925

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344 ,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP2435 9,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PO110262.D	07 Apr 2025 09:08	YP/AJ	Ok
2	AR1660CCC500	PO110263.D	07 Apr 2025 09:46	YP/AJ	Ok
3	AR1242CCC500	PO110264.D	07 Apr 2025 10:05	YP/AJ	Ok
4	AR1248CCC500	PO110265.D	07 Apr 2025 10:23	YP/AJ	Ok
5	AR1254CCC500	PO110266.D	07 Apr 2025 11:18	YP/AJ	Ok
6	I.BLK	PO110267.D	07 Apr 2025 11:36	YP/AJ	Ok
7	Q1723-04DL	PO110268.D	07 Apr 2025 12:14	YP/AJ	Ok
8	Q1730-01	PO110269.D	07 Apr 2025 13:38	YP/AJ	Ok
9	Q1730-03	PO110270.D	07 Apr 2025 13:56	YP/AJ	Ok
10	Q1730-03MS	PO110271.D	07 Apr 2025 14:15	YP/AJ	Ok
11	Q1730-03MSD	PO110272.D	07 Apr 2025 14:33	YP/AJ	Ok
12	Q1736-01	PO110273.D	07 Apr 2025 14:51	YP/AJ	Ok
13	Q1736-02	PO110274.D	07 Apr 2025 15:10	YP/AJ	Ok
14	Q1737-01	PO110275.D	07 Apr 2025 15:28	YP/AJ	Ok,M
15	Q1733-01	PO110276.D	07 Apr 2025 15:47	YP/AJ	Ok,M
16	Q1735-01	PO110277.D	07 Apr 2025 16:05	YP/AJ	Ok
17	AR1660CCC500	PO110278.D	07 Apr 2025 17:14	YP/AJ	Ok
18	AR1242CCC500	PO110279.D	07 Apr 2025 18:09	YP/AJ	Ok
19	AR1248CCC500	PO110280.D	07 Apr 2025 18:27	YP/AJ	Ok
20	AR1254CCC500	PO110281.D	07 Apr 2025 18:45	YP/AJ	Ok
21	I.BLK	PO110282.D	07 Apr 2025 19:04	YP/AJ	Ok

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QCBatch ID # PO040725

Review By	yogesh	Review On	4/7/2025 11:16:09 AM		
Supervise By	mohammad	Supervise On	4/9/2025 2:57:57 AM		
SubDirectory	PO040725	HP Acquire Method		HP Processing Method	PO031925

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344 ,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP2435 9,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

22	PB167469BL	PO110283.D	07 Apr 2025 19:22	YP/AJ	Ok
23	PB167469BS	PO110284.D	07 Apr 2025 19:41	YP/AJ	Ok
24	Q1738-01	PO110285.D	07 Apr 2025 19:59	YP/AJ	Ok
25	AR1660CCC500	PO110286.D	07 Apr 2025 21:27	YP/AJ	Ok,M
26	AR1242CCC500	PO110287.D	07 Apr 2025 22:22	YP/AJ	Ok
27	AR1248CCC500	PO110288.D	07 Apr 2025 22:40	YP/AJ	Ok
28	AR1254CCC500	PO110289.D	07 Apr 2025 22:59	YP/AJ	Ok
29	I.BLK	PO110290.D	07 Apr 2025 23:17	YP/AJ	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QCBatch ID # PO040825

Review By	yogesh	Review On	4/8/2025 10:42:21 AM		
Supervise By	mohammad	Supervise On	4/10/2025 5:29:43 AM		
SubDirectory	PO040825	HP Acquire Method		HP Processing Method	PO031925

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344 ,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP2435 9,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PO110291.D	08 Apr 2025 08:31	YP/AJ	Ok
2	AR1660CCC500	PO110292.D	08 Apr 2025 09:08	YP/AJ	Ok
3	AR1242CCC500	PO110293.D	08 Apr 2025 09:27	YP/AJ	Ok
4	AR1248CCC500	PO110294.D	08 Apr 2025 09:45	YP/AJ	Ok
5	AR1254CCC500	PO110295.D	08 Apr 2025 10:04	YP/AJ	Ok,M
6	I.BLK	PO110296.D	08 Apr 2025 10:21	YP/AJ	Ok
7	Q1730-05	PO110297.D	08 Apr 2025 10:39	YP/AJ	Ok
8	Q1730-07	PO110298.D	08 Apr 2025 10:58	YP/AJ	Ok
9	Q1730-09	PO110299.D	08 Apr 2025 11:16	YP/AJ	Ok
10	Q1730-11	PO110300.D	08 Apr 2025 11:35	YP/AJ	Ok
11	Q1730-13	PO110301.D	08 Apr 2025 11:53	YP/AJ	Ok
12	AR1660CCC500	PO110302.D	08 Apr 2025 12:58	YP/AJ	Ok
13	AR1242CCC500	PO110303.D	08 Apr 2025 13:17	YP/AJ	Ok
14	AR1248CCC500	PO110304.D	08 Apr 2025 13:35	YP/AJ	Ok
15	AR1254CCC500	PO110305.D	08 Apr 2025 13:53	YP/AJ	Ok
16	I.BLK	PO110306.D	08 Apr 2025 14:12	YP/AJ	Ok
17	PB167515BL	PO110307.D	08 Apr 2025 14:30	YP/AJ	Ok
18	PB167515BS	PO110308.D	08 Apr 2025 14:49	YP/AJ	Ok
19	PB167515BSD	PO110309.D	08 Apr 2025 15:07	YP/AJ	Ok
20	Q1735-04	PO110310.D	08 Apr 2025 15:25	YP/AJ	Ok
21	Q1746-05	PO110311.D	08 Apr 2025 15:44	YP/AJ	Ok

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QCBatch ID # PO040825

Review By	yogesh	Review On	4/8/2025 10:42:21 AM		
Supervise By	mohammad	Supervise On	4/10/2025 5:29:43 AM		
SubDirectory	PO040825	HP Acquire Method		HP Processing Method	PO031925

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344 ,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP2435 9,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

22	Q1746-06	PO110312.D	08 Apr 2025 16:02	YP/AJ	Ok
23	Q1746-07	PO110313.D	08 Apr 2025 16:20	YP/AJ	Ok
24	Q1739-01	PO110314.D	08 Apr 2025 16:39	YP/AJ	Ok,M
25	AR1660CCC500	PO110315.D	08 Apr 2025 17:54	YP/AJ	Ok
26	AR1242CCC500	PO110316.D	08 Apr 2025 18:49	YP/AJ	Ok
27	AR1248CCC500	PO110317.D	08 Apr 2025 19:08	YP/AJ	Ok
28	AR1254CCC500	PO110318.D	08 Apr 2025 19:25	YP/AJ	Ok
29	I.BLK	PO110319.D	08 Apr 2025 19:43	YP/AJ	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QCBatch ID # PP032725

Review By	yogesh	Review On	3/27/2025 12:19:10 PM		
Supervise By	mohammad	Supervise On	3/31/2025 1:17:13 AM		
SubDirectory	PP032725	HP Acquire Method		HP Processing Method	PP032725

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344 ,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP2435 9,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PP070915.D	27 Mar 2025 09:58	YP\AJ	Ok
2	I.BLK	PP070916.D	27 Mar 2025 10:14	YP\AJ	Ok
3	AR1660ICC1000	PP070917.D	27 Mar 2025 10:30	YP\AJ	Ok
4	AR1660ICC750	PP070918.D	27 Mar 2025 10:47	YP\AJ	Ok
5	AR1660ICC500	PP070919.D	27 Mar 2025 11:03	YP\AJ	Ok
6	AR1660ICC250	PP070920.D	27 Mar 2025 11:19	YP\AJ	Ok,M
7	AR1660ICC050	PP070921.D	27 Mar 2025 11:35	YP\AJ	Ok,M
8	AR1221ICC500	PP070922.D	27 Mar 2025 11:52	YP\AJ	Ok
9	AR1232ICC500	PP070923.D	27 Mar 2025 12:11	YP\AJ	Ok,M
10	AR1242ICC1000	PP070924.D	27 Mar 2025 12:28	YP\AJ	Ok
11	AR1242ICC750	PP070925.D	27 Mar 2025 12:44	YP\AJ	Ok
12	AR1242ICC500	PP070926.D	27 Mar 2025 13:00	YP\AJ	Ok
13	AR1242ICC250	PP070927.D	27 Mar 2025 13:16	YP\AJ	Ok
14	AR1242ICC050	PP070928.D	27 Mar 2025 13:33	YP\AJ	Ok,M
15	AR1248ICC1000	PP070929.D	27 Mar 2025 13:49	YP\AJ	Ok
16	AR1248ICC750	PP070930.D	27 Mar 2025 14:05	YP\AJ	Ok
17	AR1248ICC500	PP070931.D	27 Mar 2025 14:22	YP\AJ	Ok
18	AR1248ICC250	PP070932.D	27 Mar 2025 14:38	YP\AJ	Ok
19	AR1248ICC050	PP070933.D	27 Mar 2025 14:55	YP\AJ	Ok,M
20	AR1254ICC1000	PP070934.D	27 Mar 2025 15:11	YP\AJ	Ok,M
21	AR1254ICC750	PP070935.D	27 Mar 2025 15:27	YP\AJ	Ok,M

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QCBatch ID # PP032725

Review By	yogesh	Review On	3/27/2025 12:19:10 PM		
Supervise By	mohammad	Supervise On	3/31/2025 1:17:13 AM		
SubDirectory	PP032725	HP Acquire Method		HP Processing Method	PP032725

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344 ,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP2435 9,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

22	AR1254ICC500	PP070936.D	27 Mar 2025 15:43	YP\AJ	Ok
23	AR1254ICC250	PP070937.D	27 Mar 2025 16:00	YP\AJ	Ok,M
24	AR1254ICC050	PP070938.D	27 Mar 2025 16:16	YP\AJ	Ok
25	AR1262ICC500	PP070939.D	27 Mar 2025 16:32	YP\AJ	Ok
26	AR1268ICC1000	PP070940.D	27 Mar 2025 16:48	YP\AJ	Ok,M
27	AR1268ICC750	PP070941.D	27 Mar 2025 17:05	YP\AJ	Ok,M
28	AR1268ICC500	PP070942.D	27 Mar 2025 17:21	YP\AJ	Ok,M
29	AR1268ICC250	PP070943.D	27 Mar 2025 17:37	YP\AJ	Ok,M
30	AR1268ICC050	PP070944.D	27 Mar 2025 17:53	YP\AJ	Ok,M
31	PP032725ICV500	PP070945.D	27 Mar 2025 18:26	YP\AJ	Ok
32	AR1242ICV500	PP070946.D	27 Mar 2025 18:42	YP\AJ	Ok
33	AR1248ICV500	PP070947.D	27 Mar 2025 19:15	YP\AJ	Ok
34	AR1254ICV500	PP070948.D	27 Mar 2025 19:47	YP\AJ	Ok,M
35	AR1268ICV500	PP070949.D	27 Mar 2025 20:20	YP\AJ	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QCBatch ID # PP040825

Review By	yogesh	Review On	4/8/2025 3:06:55 PM		
Supervise By	mohammad	Supervise On	4/10/2025 5:29:57 AM		
SubDirectory	PP040825	HP Acquire Method		HP Processing Method	PP032725

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344 ,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP2435 9,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PP071135.D	08 Apr 2025 08:39	YP\AJ	Ok
2	AR1660CCC500	PP071136.D	08 Apr 2025 09:37	YP\AJ	Ok,M
3	AR1242CCC500	PP071137.D	08 Apr 2025 09:56	YP\AJ	Ok
4	AR1248CCC500	PP071138.D	08 Apr 2025 10:33	YP\AJ	Ok
5	AR1254CCC500	PP071139.D	08 Apr 2025 10:49	YP\AJ	Ok
6	I.BLK	PP071140.D	08 Apr 2025 11:05	YP\AJ	Ok
7	PB167469BL	PP071141.D	08 Apr 2025 11:23	YP\AJ	Not Ok
8	Q1730-15	PP071142.D	08 Apr 2025 11:39	YP\AJ	Ok
9	Q1730-17	PP071143.D	08 Apr 2025 11:56	YP\AJ	Ok
10	Q1730-19	PP071144.D	08 Apr 2025 12:12	YP\AJ	Ok
11	Q1740-01	PP071145.D	08 Apr 2025 13:08	YP\AJ	Ok
12	Q1742-01	PP071146.D	08 Apr 2025 13:25	YP\AJ	Ok,M
13	Q1742-01MS	PP071147.D	08 Apr 2025 13:41	YP\AJ	Ok,M
14	Q1742-01MSD	PP071148.D	08 Apr 2025 13:57	YP\AJ	Ok,M
15	Q1743-01	PP071149.D	08 Apr 2025 14:14	YP\AJ	Ok,M
16	Q1744-01	PP071150.D	08 Apr 2025 14:30	YP\AJ	Ok
17	AR1660CCC500	PP071151.D	08 Apr 2025 15:29	YP\AJ	Ok
18	AR1242CCC500	PP071152.D	08 Apr 2025 16:02	YP\AJ	Ok
19	AR1248CCC500	PP071153.D	08 Apr 2025 16:18	YP\AJ	Ok
20	AR1254CCC500	PP071154.D	08 Apr 2025 16:34	YP\AJ	Ok
21	I.BLK	PP071155.D	08 Apr 2025 16:51	YP\AJ	Ok

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QCBatch ID # PP040825

Review By	yogesh	Review On	4/8/2025 3:06:55 PM		
Supervise By	mohammad	Supervise On	4/10/2025 5:29:57 AM		
SubDirectory	PP040825	HP Acquire Method		HP Processing Method	PP032725

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344 ,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP2435 9,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

22	PB167507BL	PP071156.D	08 Apr 2025 17:07	YP\AJ	Ok
23	PB167507BS	PP071157.D	08 Apr 2025 17:23	YP\AJ	Ok
24	Q1744-03	PP071158.D	08 Apr 2025 17:40	YP\AJ	Ok
25	Q1745-01	PP071159.D	08 Apr 2025 17:56	YP\AJ	Ok,M
26	Q1745-09	PP071160.D	08 Apr 2025 18:12	YP\AJ	Ok,M
27	Q1746-01	PP071161.D	08 Apr 2025 18:28	YP\AJ	Ok
28	Q1746-03	PP071162.D	08 Apr 2025 18:45	YP\AJ	Ok
29	Q1747-01	PP071163.D	08 Apr 2025 19:01	YP\AJ	Ok,M
30	AR1660CCC500	PP071164.D	08 Apr 2025 20:28	YP\AJ	Ok
31	AR1242CCC500	PP071165.D	08 Apr 2025 21:17	YP\AJ	Ok
32	AR1248CCC500	PP071166.D	08 Apr 2025 21:33	YP\AJ	Ok
33	AR1254CCC500	PP071167.D	08 Apr 2025 21:49	YP\AJ	Ok
34	I.BLK	PP071168.D	08 Apr 2025 22:05	YP\AJ	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QCBatch ID # PO031925

Review By	yogesh	Review On	3/18/2025 3:19:21 PM		
Supervise By	mohammad	Supervise On	3/24/2025 3:02:46 AM		
SubDirectory	PO031925	HP Acquire Method		HP Processing Method	PO031925

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO109970.D	18 Mar 2025 13:26		YP/AJ	Ok
2	I.BLK	I.BLK	PO109971.D	18 Mar 2025 13:44		YP/AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PO109972.D	18 Mar 2025 14:03		YP/AJ	Ok
4	AR1660ICC750	AR1660ICC750	PO109973.D	18 Mar 2025 14:21		YP/AJ	Ok,M
5	AR1660ICC500	AR1660ICC500	PO109974.D	18 Mar 2025 14:39		YP/AJ	Ok
6	AR1660ICC250	AR1660ICC250	PO109975.D	18 Mar 2025 14:58		YP/AJ	Ok
7	AR1660ICC050	AR1660ICC050	PO109976.D	18 Mar 2025 15:16		YP/AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PO109977.D	18 Mar 2025 15:34		YP/AJ	Ok
9	AR1232ICC500	AR1232ICC500	PO109978.D	18 Mar 2025 15:53		YP/AJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PO109979.D	18 Mar 2025 16:11		YP/AJ	Ok
11	AR1242ICC750	AR1242ICC750	PO109980.D	18 Mar 2025 16:30		YP/AJ	Ok
12	AR1242ICC500	AR1242ICC500	PO109981.D	18 Mar 2025 16:48		YP/AJ	Ok
13	AR1242ICC250	AR1242ICC250	PO109982.D	18 Mar 2025 17:06		YP/AJ	Ok
14	AR1242ICC050	AR1242ICC050	PO109983.D	18 Mar 2025 17:25		YP/AJ	Ok,M
15	AR1248ICC1000	AR1248ICC1000	PO109984.D	18 Mar 2025 17:42		YP/AJ	Ok
16	AR1248ICC750	AR1248ICC750	PO109985.D	18 Mar 2025 18:00		YP/AJ	Ok
17	AR1248ICC500	AR1248ICC500	PO109986.D	18 Mar 2025 18:19		YP/AJ	Ok
18	AR1248ICC250	AR1248ICC250	PO109987.D	18 Mar 2025 18:37		YP/AJ	Ok

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QCBatch ID # PO031925

Review By	yogesh	Review On	3/18/2025 3:19:21 PM		
Supervise By	mohammad	Supervise On	3/24/2025 3:02:46 AM		
SubDirectory	PO031925	HP Acquire Method		HP Processing Method	PO031925

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	AR1248ICC050	AR1248ICC050	PO109988.D	18 Mar 2025 18:54		YP/AJ	Ok,M
20	AR1254ICC1000	AR1254ICC1000	PO109989.D	18 Mar 2025 19:13		YP/AJ	Ok
21	AR1254ICC750	AR1254ICC750	PO109990.D	18 Mar 2025 19:31		YP/AJ	Ok
22	AR1254ICC500	AR1254ICC500	PO109991.D	18 Mar 2025 19:49		YP/AJ	Ok
23	AR1254ICC250	AR1254ICC250	PO109992.D	18 Mar 2025 20:07		YP/AJ	Ok
24	AR1254ICC050	AR1254ICC050	PO109993.D	18 Mar 2025 20:25		YP/AJ	Ok,M
25	AR1262ICC500	AR1262ICC500	PO109994.D	18 Mar 2025 20:43		YP/AJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PO109995.D	18 Mar 2025 21:02		YP/AJ	Ok
27	AR1268ICC750	AR1268ICC750	PO109996.D	18 Mar 2025 21:20		YP/AJ	Ok
28	AR1268ICC500	AR1268ICC500	PO109997.D	18 Mar 2025 21:39		YP/AJ	Ok,M
29	AR1268ICC250	AR1268ICC250	PO109998.D	18 Mar 2025 21:57		YP/AJ	Ok,M
30	AR1268ICC050	AR1268ICC050	PO109999.D	18 Mar 2025 22:15		YP/AJ	Ok,M
31	PO031925ICV500	ICVPO031925	PO110000.D	18 Mar 2025 22:34		YP/AJ	Ok
32	AR1242ICV500	ICVPO031925AR1242	PO110001.D	18 Mar 2025 23:47		YP/AJ	Ok
33	AR1248ICV500	ICVPO031925AR1248	PO110002.D	19 Mar 2025 00:05		YP/AJ	Ok
34	AR1254ICV500	ICVPO031925AR1254	PO110003.D	19 Mar 2025 00:24		YP/AJ	Ok
35	AR1268ICV500	ICVPO031925AR1268	PO110004.D	19 Mar 2025 00:42		YP/AJ	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QCBatch ID # PO040725

Review By	yogesh	Review On	4/7/2025 11:16:09 AM		
Supervise By	mohammad	Supervise On	4/9/2025 2:57:57 AM		
SubDirectory	PO040725	HP Acquire Method		HP Processing Method	PO031925

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO110262.D	07 Apr 2025 09:08		YP/AJ	Ok
2	AR1660CCC500	AR1660CCC500	PO110263.D	07 Apr 2025 09:46		YP/AJ	Ok
3	AR1242CCC500	AR1242CCC500	PO110264.D	07 Apr 2025 10:05		YP/AJ	Ok
4	AR1248CCC500	AR1248CCC500	PO110265.D	07 Apr 2025 10:23		YP/AJ	Ok
5	AR1254CCC500	AR1254CCC500	PO110266.D	07 Apr 2025 11:18		YP/AJ	Ok
6	I.BLK	I.BLK	PO110267.D	07 Apr 2025 11:36		YP/AJ	Ok
7	Q1723-04DL	GAS-AUD-25-0037DL	PO110268.D	07 Apr 2025 12:14	AR1254 Hit	YP/AJ	Ok
8	Q1730-01	OU4-VSL-15-040325	PO110269.D	07 Apr 2025 13:38		YP/AJ	Ok
9	Q1730-03	OU4-VSL-16-040325	PO110270.D	07 Apr 2025 13:56		YP/AJ	Ok
10	Q1730-03MS	OU4-VSL-16-040325MS	PO110271.D	07 Apr 2025 14:15		YP/AJ	Ok
11	Q1730-03MSD	OU4-VSL-16-040325MS	PO110272.D	07 Apr 2025 14:33		YP/AJ	Ok
12	Q1736-01	GST1	PO110273.D	07 Apr 2025 14:51		YP/AJ	Ok
13	Q1736-02	GST2	PO110274.D	07 Apr 2025 15:10		YP/AJ	Ok
14	Q1737-01	RT3069	PO110275.D	07 Apr 2025 15:28	AR1260 hit	YP/AJ	Ok,M
15	Q1733-01	ETGI-328	PO110276.D	07 Apr 2025 15:47		YP/AJ	Ok,M
16	Q1735-01	50660-50661-50662-56	PO110277.D	07 Apr 2025 16:05		YP/AJ	Ok
17	AR1660CCC500	AR1660CCC500	PO110278.D	07 Apr 2025 17:14		YP/AJ	Ok
18	AR1242CCC500	AR1242CCC500	PO110279.D	07 Apr 2025 18:09		YP/AJ	Ok

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QCBatch ID # PO040725

Review By	yogesh	Review On	4/7/2025 11:16:09 AM		
Supervise By	mohammad	Supervise On	4/9/2025 2:57:57 AM		
SubDirectory	PO040725	HP Acquire Method		HP Processing Method	PO031925

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	AR1248CCC500	AR1248CCC500	PO110280.D	07 Apr 2025 18:27		YP/AJ	Ok
20	AR1254CCC500	AR1254CCC500	PO110281.D	07 Apr 2025 18:45		YP/AJ	Ok
21	I.BLK	I.BLK	PO110282.D	07 Apr 2025 19:04		YP/AJ	Ok
22	PB167469BL	PB167469BL	PO110283.D	07 Apr 2025 19:22		YP/AJ	Ok
23	PB167469BS	PB167469BS	PO110284.D	07 Apr 2025 19:41		YP/AJ	Ok
24	Q1738-01	OK-02-040425	PO110285.D	07 Apr 2025 19:59		YP/AJ	Ok
25	AR1660CCC500	AR1660CCC500	PO110286.D	07 Apr 2025 21:27		YP/AJ	Ok,M
26	AR1242CCC500	AR1242CCC500	PO110287.D	07 Apr 2025 22:22		YP/AJ	Ok
27	AR1248CCC500	AR1248CCC500	PO110288.D	07 Apr 2025 22:40		YP/AJ	Ok
28	AR1254CCC500	AR1254CCC500	PO110289.D	07 Apr 2025 22:59		YP/AJ	Ok
29	I.BLK	I.BLK	PO110290.D	07 Apr 2025 23:17		YP/AJ	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QCBatch ID # PO040825

Review By	yogesh	Review On	4/8/2025 10:42:21 AM		
Supervise By	mohammad	Supervise On	4/10/2025 5:29:43 AM		
SubDirectory	PO040825	HP Acquire Method		HP Processing Method	PO031925

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO110291.D	08 Apr 2025 08:31		YP/AJ	Ok
2	AR1660CCC500	AR1660CCC500	PO110292.D	08 Apr 2025 09:08		YP/AJ	Ok
3	AR1242CCC500	AR1242CCC500	PO110293.D	08 Apr 2025 09:27		YP/AJ	Ok
4	AR1248CCC500	AR1248CCC500	PO110294.D	08 Apr 2025 09:45		YP/AJ	Ok
5	AR1254CCC500	AR1254CCC500	PO110295.D	08 Apr 2025 10:04		YP/AJ	Ok,M
6	I.BLK	I.BLK	PO110296.D	08 Apr 2025 10:21		YP/AJ	Ok
7	Q1730-05	OU4-VSL-17-040325	PO110297.D	08 Apr 2025 10:39		YP/AJ	Ok
8	Q1730-07	OU4-PCS-TC-21-04032	PO110298.D	08 Apr 2025 10:58		YP/AJ	Ok
9	Q1730-09	OU4-PCS-TC-22-04032	PO110299.D	08 Apr 2025 11:16		YP/AJ	Ok
10	Q1730-11	OU4-PCS-TC-23-04032	PO110300.D	08 Apr 2025 11:35		YP/AJ	Ok
11	Q1730-13	OU4-PCS-TC-24-04032	PO110301.D	08 Apr 2025 11:53		YP/AJ	Ok
12	AR1660CCC500	AR1660CCC500	PO110302.D	08 Apr 2025 12:58		YP/AJ	Ok
13	AR1242CCC500	AR1242CCC500	PO110303.D	08 Apr 2025 13:17		YP/AJ	Ok
14	AR1248CCC500	AR1248CCC500	PO110304.D	08 Apr 2025 13:35		YP/AJ	Ok
15	AR1254CCC500	AR1254CCC500	PO110305.D	08 Apr 2025 13:53		YP/AJ	Ok
16	I.BLK	I.BLK	PO110306.D	08 Apr 2025 14:12		YP/AJ	Ok
17	PB167515BL	PB167515BL	PO110307.D	08 Apr 2025 14:30		YP/AJ	Ok
18	PB167515BS	PB167515BS	PO110308.D	08 Apr 2025 14:49		YP/AJ	Ok

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QCBatch ID # PO040825

Review By	yogesh	Review On	4/8/2025 10:42:21 AM		
Supervise By	mohammad	Supervise On	4/10/2025 5:29:43 AM		
SubDirectory	PO040825	HP Acquire Method		HP Processing Method	PO031925

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	PB167515BSD	PB167515BSD	PO110309.D	08 Apr 2025 15:07		YP/AJ	Ok
20	Q1735-04	0401-A-0401-B-COMP	PO110310.D	08 Apr 2025 15:25		YP/AJ	Ok
21	Q1746-05	B-158-GW01	PO110311.D	08 Apr 2025 15:44		YP/AJ	Ok
22	Q1746-06	B-149-GW01	PO110312.D	08 Apr 2025 16:02		YP/AJ	Ok
23	Q1746-07	EB-2025-4-7	PO110313.D	08 Apr 2025 16:20		YP/AJ	Ok
24	Q1739-01	WC-LIQUID-20250404	PO110314.D	08 Apr 2025 16:39	AR1254 Hit	YP/AJ	Ok,M
25	AR1660CCC500	AR1660CCC500	PO110315.D	08 Apr 2025 17:54		YP/AJ	Ok
26	AR1242CCC500	AR1242CCC500	PO110316.D	08 Apr 2025 18:49		YP/AJ	Ok
27	AR1248CCC500	AR1248CCC500	PO110317.D	08 Apr 2025 19:08		YP/AJ	Ok
28	AR1254CCC500	AR1254CCC500	PO110318.D	08 Apr 2025 19:25		YP/AJ	Ok
29	I.BLK	I.BLK	PO110319.D	08 Apr 2025 19:43		YP/AJ	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QCBatch ID # PP032725

Review By	yogesh	Review On	3/27/2025 12:19:10 PM		
Supervise By	mohammad	Supervise On	3/31/2025 1:17:13 AM		
SubDirectory	PP032725	HP Acquire Method		HP Processing Method	PP032725

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP070915.D	27 Mar 2025 09:58		YP\AJ	Ok
2	I.BLK	I.BLK	PP070916.D	27 Mar 2025 10:14		YP\AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PP070917.D	27 Mar 2025 10:30		YP\AJ	Ok
4	AR1660ICC750	AR1660ICC750	PP070918.D	27 Mar 2025 10:47		YP\AJ	Ok
5	AR1660ICC500	AR1660ICC500	PP070919.D	27 Mar 2025 11:03		YP\AJ	Ok
6	AR1660ICC250	AR1660ICC250	PP070920.D	27 Mar 2025 11:19		YP\AJ	Ok,M
7	AR1660ICC050	AR1660ICC050	PP070921.D	27 Mar 2025 11:35		YP\AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PP070922.D	27 Mar 2025 11:52		YP\AJ	Ok
9	AR1232ICC500	AR1232ICC500	PP070923.D	27 Mar 2025 12:11		YP\AJ	Ok,M
10	AR1242ICC1000	AR1242ICC1000	PP070924.D	27 Mar 2025 12:28		YP\AJ	Ok
11	AR1242ICC750	AR1242ICC750	PP070925.D	27 Mar 2025 12:44		YP\AJ	Ok
12	AR1242ICC500	AR1242ICC500	PP070926.D	27 Mar 2025 13:00		YP\AJ	Ok
13	AR1242ICC250	AR1242ICC250	PP070927.D	27 Mar 2025 13:16		YP\AJ	Ok
14	AR1242ICC050	AR1242ICC050	PP070928.D	27 Mar 2025 13:33		YP\AJ	Ok,M
15	AR1248ICC1000	AR1248ICC1000	PP070929.D	27 Mar 2025 13:49		YP\AJ	Ok
16	AR1248ICC750	AR1248ICC750	PP070930.D	27 Mar 2025 14:05		YP\AJ	Ok
17	AR1248ICC500	AR1248ICC500	PP070931.D	27 Mar 2025 14:22		YP\AJ	Ok
18	AR1248ICC250	AR1248ICC250	PP070932.D	27 Mar 2025 14:38		YP\AJ	Ok

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QCBatch ID # PP032725

Review By	yogesh	Review On	3/27/2025 12:19:10 PM		
Supervise By	mohammad	Supervise On	3/31/2025 1:17:13 AM		
SubDirectory	PP032725	HP Acquire Method		HP Processing Method	PP032725

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	AR1248ICC050	AR1248ICC050	PP070933.D	27 Mar 2025 14:55		YP\AJ	Ok,M
20	AR1254ICC1000	AR1254ICC1000	PP070934.D	27 Mar 2025 15:11		YP\AJ	Ok,M
21	AR1254ICC750	AR1254ICC750	PP070935.D	27 Mar 2025 15:27		YP\AJ	Ok,M
22	AR1254ICC500	AR1254ICC500	PP070936.D	27 Mar 2025 15:43		YP\AJ	Ok
23	AR1254ICC250	AR1254ICC250	PP070937.D	27 Mar 2025 16:00		YP\AJ	Ok,M
24	AR1254ICC050	AR1254ICC050	PP070938.D	27 Mar 2025 16:16		YP\AJ	Ok
25	AR1262ICC500	AR1262ICC500	PP070939.D	27 Mar 2025 16:32		YP\AJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PP070940.D	27 Mar 2025 16:48		YP\AJ	Ok,M
27	AR1268ICC750	AR1268ICC750	PP070941.D	27 Mar 2025 17:05		YP\AJ	Ok,M
28	AR1268ICC500	AR1268ICC500	PP070942.D	27 Mar 2025 17:21		YP\AJ	Ok,M
29	AR1268ICC250	AR1268ICC250	PP070943.D	27 Mar 2025 17:37		YP\AJ	Ok,M
30	AR1268ICC050	AR1268ICC050	PP070944.D	27 Mar 2025 17:53		YP\AJ	Ok,M
31	PP032725ICV500	ICVPP032725	PP070945.D	27 Mar 2025 18:26		YP\AJ	Ok
32	AR1242ICV500	ICVPP032725AR1242	PP070946.D	27 Mar 2025 18:42		YP\AJ	Ok
33	AR1248ICV500	ICVPP032725AR1248	PP070947.D	27 Mar 2025 19:15		YP\AJ	Ok
34	AR1254ICV500	ICVPP032725AR1254	PP070948.D	27 Mar 2025 19:47		YP\AJ	Ok,M
35	AR1268ICV500	ICVPP032725AR1268	PP070949.D	27 Mar 2025 20:20		YP\AJ	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QCBatch ID # PP040825

Review By	yogesh	Review On	4/8/2025 3:06:55 PM		
Supervise By	mohammad	Supervise On	4/10/2025 5:29:57 AM		
SubDirectory	PP040825	HP Acquire Method		HP Processing Method	PP032725

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP071135.D	08 Apr 2025 08:39		YP\AJ	Ok
2	AR1660CCC500	AR1660CCC500	PP071136.D	08 Apr 2025 09:37		YP\AJ	Ok,M
3	AR1242CCC500	AR1242CCC500	PP071137.D	08 Apr 2025 09:56		YP\AJ	Ok
4	AR1248CCC500	AR1248CCC500	PP071138.D	08 Apr 2025 10:33		YP\AJ	Ok
5	AR1254CCC500	AR1254CCC500	PP071139.D	08 Apr 2025 10:49		YP\AJ	Ok
6	I.BLK	I.BLK	PP071140.D	08 Apr 2025 11:05		YP\AJ	Ok
7	PB167469BL	PB167469BL	PP071141.D	08 Apr 2025 11:23	not required	YP\AJ	Not Ok
8	Q1730-15	OU4-PCS-TC-25-04032	PP071142.D	08 Apr 2025 11:39		YP\AJ	Ok
9	Q1730-17	OU4-PCS-TC-26-04032	PP071143.D	08 Apr 2025 11:56		YP\AJ	Ok
10	Q1730-19	OU4-CF-15-040325	PP071144.D	08 Apr 2025 12:12	DCB High in 2ND column	YP\AJ	Ok
11	Q1740-01	TP-20	PP071145.D	08 Apr 2025 13:08		YP\AJ	Ok
12	Q1742-01	TR-06-040725	PP071146.D	08 Apr 2025 13:25		YP\AJ	Ok,M
13	Q1742-01MS	TR-06-040725MS	PP071147.D	08 Apr 2025 13:41		YP\AJ	Ok,M
14	Q1742-01MSD	TR-06-040725MSD	PP071148.D	08 Apr 2025 13:57		YP\AJ	Ok,M
15	Q1743-01	TP-16	PP071149.D	08 Apr 2025 14:14	AR1260 Hit	YP\AJ	Ok,M
16	Q1744-01	B-158-SB01	PP071150.D	08 Apr 2025 14:30	TCMX High in 2nd column	YP\AJ	Ok
17	AR1660CCC500	AR1660CCC500	PP071151.D	08 Apr 2025 15:29		YP\AJ	Ok
18	AR1242CCC500	AR1242CCC500	PP071152.D	08 Apr 2025 16:02		YP\AJ	Ok

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QCBatch ID # PP040825

Review By	yogesh	Review On	4/8/2025 3:06:55 PM		
Supervise By	mohammad	Supervise On	4/10/2025 5:29:57 AM		
SubDirectory	PP040825	HP Acquire Method		HP Processing Method	PP032725

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	AR1248CCC500	AR1248CCC500	PP071153.D	08 Apr 2025 16:18		YP\AJ	Ok
20	AR1254CCC500	AR1254CCC500	PP071154.D	08 Apr 2025 16:34		YP\AJ	Ok
21	I.BLK	I.BLK	PP071155.D	08 Apr 2025 16:51		YP\AJ	Ok
22	PB167507BL	PB167507BL	PP071156.D	08 Apr 2025 17:07		YP\AJ	Ok
23	PB167507BS	PB167507BS	PP071157.D	08 Apr 2025 17:23		YP\AJ	Ok
24	Q1744-03	B-158-SB02	PP071158.D	08 Apr 2025 17:40		YP\AJ	Ok
25	Q1745-01	IB-6A-WC	PP071159.D	08 Apr 2025 17:56		YP\AJ	Ok,M
26	Q1745-09	IB-6.5-WC	PP071160.D	08 Apr 2025 18:12		YP\AJ	Ok,M
27	Q1746-01	B-149-SB01	PP071161.D	08 Apr 2025 18:28		YP\AJ	Ok
28	Q1746-03	B-149-SB02	PP071162.D	08 Apr 2025 18:45		YP\AJ	Ok
29	Q1747-01	ARS20-0008	PP071163.D	08 Apr 2025 19:01		YP\AJ	Ok,M
30	AR1660CCC500	AR1660CCC500	PP071164.D	08 Apr 2025 20:28		YP\AJ	Ok
31	AR1242CCC500	AR1242CCC500	PP071165.D	08 Apr 2025 21:17		YP\AJ	Ok
32	AR1248CCC500	AR1248CCC500	PP071166.D	08 Apr 2025 21:33		YP\AJ	Ok
33	AR1254CCC500	AR1254CCC500	PP071167.D	08 Apr 2025 21:49		YP\AJ	Ok
34	I.BLK	I.BLK	PP071168.D	08 Apr 2025 22:05		YP\AJ	Ok

M : Manual Integration

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 4/7/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 04/04/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:25
Out Date: 04/05/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC: LB135307

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q1729-05	SVOC-GPC-BLANK	1	1.00	1.00	2.00	2.00	100.0	
Q1729-06	PEST-GPC-BLANK	2	1.00	1.00	2.00	2.00	100.0	
Q1729-07	PEST-GPC-BLANK-SPIKE	3	1.00	1.00	2.00	2.00	100.0	
Q1729-08	PCB-GPC-BLANK	4	1.00	1.00	2.00	2.00	100.0	
Q1729-09	PCB-GPC-BLANK-SPIKE	5	1.00	1.00	2.00	2.00	100.0	
Q1729-10	SVOC-GPC2-BLANK	6	1.00	1.00	2.00	2.00	100.0	
Q1729-11	PEST-GPC2-BLANK	7	1.00	1.00	2.00	2.00	100.0	
Q1729-12	PEST-GPC2-BLANK-SPIKE	8	1.00	1.00	2.00	2.00	100.0	
Q1729-13	PCB-GPC2-BLANK	9	1.00	1.00	2.00	2.00	100.0	
Q1729-14	PCB-GPC2-BLANK-SPIKE	10	1.00	1.00	2.00	2.00	100.0	
Q1730-01	OU4-VSL-15-040325	11	1.14	10.58	11.72	11.39	96.9	
Q1730-03	OU4-VSL-16-040325	12	1.15	10.28	11.43	10.86	94.5	
Q1730-05	OU4-VSL-17-040325	13	1.16	11.32	12.48	11.62	92.4	
Q1730-07	OU4-PCS-TC-21-040325	14	1.18	10.92	12.1	10.9	89.0	
Q1730-09	OU4-PCS-TC-22-040325	15	1.16	11.79	12.95	11.87	90.8	
Q1730-11	OU4-PCS-TC-23-040325	16	1.19	11.24	12.43	11.41	90.9	
Q1730-13	OU4-PCS-TC-24-040325	17	1.18	10.42	11.6	10.69	91.3	
Q1730-15	OU4-PCS-TC-25-040325	18	1.14	11.13	12.27	11.43	92.5	
Q1730-17	OU4-PCS-TC-26-040325	19	1.15	10.84	11.99	11.13	92.1	
Q1730-19	OU4-CF-15-040325	20	1.15	10.74	11.89	11.23	93.9	
Q1732-01	TT-8	21	1.15	10.40	11.55	10.48	89.7	
Q1732-02	TT-8-EPH	22	1.15	10.33	11.48	10.34	89.0	
Q1732-03	TT-8-VOC	23	1.19	10.01	11.2	10.11	89.1	
Q1733-01	ETGI-328	24	1.18	10.19	11.37	8.37	70.6	
Q1733-02	ETGI-328-E2	25	1.18	10.29	11.47	8.27	68.9	
Q1734-01	HEH6237-1-1	26	1.00	1.00	2.00	2.00	100.0	pilc
Q1734-02	HEH6237-1-2	27	1.00	1.00	2.00	2.00	100.0	pilc
Q1734-03	STJ23-1-1	28	1.00	1.00	2.00	2.00	100.0	pilc

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 4/7/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 04/04/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:25
Out Date: 04/05/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC: LB135307

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q1734-04	STJ23-1-2	29	1.00	1.00	2.00	2.00	100.0	pilc
Q1734-05	HIA989S-1-1	30	1.00	1.00	2.00	2.00	100.0	pilc
Q1734-06	HIA989S-1-2	31	1.00	1.00	2.00	2.00	100.0	pilc
Q1734-07	HED302R-1-1	32	1.00	1.00	2.00	2.00	100.0	pilc
Q1734-08	HED302R-1-2	33	1.00	1.00	2.00	2.00	100.0	pilc
Q1734-09	HED302R-2-1	34	1.00	1.00	2.00	2.00	100.0	pilc
Q1734-10	HED302R-2-2	35	1.00	1.00	2.00	2.00	100.0	pilc
Q1735-01	50660-50661-50662-5663 -COMP	36	1.18	10.92	12.1	10.42	84.6	
Q1736-01	GST1	37	1.18	10.04	11.22	9.07	78.6	
Q1736-02	GST2	38	1.18	10.93	12.11	9.98	80.5	
Q1737-01	RT3069	39	1.19	11.58	12.77	11.34	87.7	
Q1738-01	OK-02-040425	40	1.16	10.34	11.5	10.06	86.1	
Q1738-02	OK-02-040425-E2	41	1.19	10.54	11.73	10.37	87.1	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

WORKLIST(Hardcopy Internal Chain)

VB 135307

WorkList Name : %1-040425 **WorkList ID :** 188724 **Department :** Wet-Chemistry **Date :** 04-04-2025 08:12:22

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1736-01	GST1	Solid	Percent Solids	Cool 4 deg C	GENV01	L21	04/04/2025	Chemtech -SO
Q1736-02	GST2	Solid	Percent Solids	Cool 4 deg C	GENV01	L21	04/04/2025	Chemtech -SO
Q1730-01	OU4-VSL-15-040325	Solid	Percent Solids	Cool 4 deg C	NOBI03	L31	04/03/2025	Chemtech -SO
Q1730-03	OU4-VSL-16-040325	Solid	Percent Solids	Cool 4 deg C	NOBI03	L31	04/03/2025	Chemtech -SO
Q1730-05	OU4-VSL-17-040325	Solid	Percent Solids	Cool 4 deg C	NOBI03	L31	04/03/2025	Chemtech -SO
Q1730-07	OU4-PCS-TC-21-040325	Solid	Percent Solids	Cool 4 deg C	NOBI03	L31	04/03/2025	Chemtech -SO
Q1730-09	OU4-PCS-TC-22-040325	Solid	Percent Solids	Cool 4 deg C	NOBI03	L31	04/03/2025	Chemtech -SO
Q1730-11	OU4-PCS-TC-23-040325	Solid	Percent Solids	Cool 4 deg C	NOBI03	L31	04/03/2025	Chemtech -SO
Q1730-13	OU4-PCS-TC-24-040325	Solid	Percent Solids	Cool 4 deg C	NOBI03	L31	04/03/2025	Chemtech -SO
Q1730-15	OU4-PCS-TC-25-040325	Solid	Percent Solids	Cool 4 deg C	NOBI03	L31	04/03/2025	Chemtech -SO
Q1730-17	OU4-PCS-TC-26-040325	Solid	Percent Solids	Cool 4 deg C	NOBI03	L31	04/03/2025	Chemtech -SO
Q1730-19	OU4-CF-15-040325	Solid	Percent Solids	Cool 4 deg C	NOBI03	L31	04/03/2025	Chemtech -SO
Q1729-05	SVOC-GPC-BLANK	Solid	Percent Solids	Cool 4 deg C	NOBI03	L31	04/03/2025	Chemtech -SO
Q1729-06	PEST-GPC-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	F11	04/04/2025	Chemtech -SO
Q1729-07	PEST-GPC-BLANK-SPIKE	Solid	Percent Solids	Cool 4 deg C	CHEM02	F11	04/04/2025	Chemtech -SO
Q1729-08	PCB-GPC-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	F11	04/04/2025	Chemtech -SO
Q1729-09	PCB-GPC-BLANK-SPIKE	Solid	Percent Solids	Cool 4 deg C	CHEM02	F11	04/04/2025	Chemtech -SO
Q1729-10	SVOC-GPC2-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	F11	04/04/2025	Chemtech -SO
Q1729-11	PEST-GPC2-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	F11	04/04/2025	Chemtech -SO
Q1729-12	PEST-GPC2-BLANK-SPIKE	Solid	Percent Solids	Cool 4 deg C	CHEM02	F11	04/04/2025	Chemtech -SO
Q1729-13	PCB-GPC2-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	F11	04/04/2025	Chemtech -SO

Date/Time 04/04/25 1520

Raw Sample Received by: JB CCDC

Raw Sample Relinquished by: SM

Date/Time 04/04/25 17:15

Raw Sample Received by: SM

Raw Sample Relinquished by: JB CCDC

WORKLIST(Hardcopy Internal Chain)

WP135307

WorkList Name : %1-040425 **WorkList ID :** 188724 **Department :** Wet-Chemistry **Date :** 04-04-2025 08:12:22

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1729-14	PCB-GPC2-BLANK-SPIKE	Solid	Percent Solids	Cool 4 deg C	CHEM02	F11	04/04/2025	Chemtech -SO
Q1732-01	TT-8	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	04/04/2025	Chemtech -SO
Q1732-02	TT-8-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	04/04/2025	Chemtech -SO
Q1732-03	TT-8-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	04/04/2025	Chemtech -SO
Q1733-01	ETGI-328	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	04/04/2025	Chemtech -SO
Q1733-02	ETGI-328-E2	Solid	Percent Solids	Cool 4 deg C	PSEG03	I31	04/04/2025	Chemtech -SO
Q1734-01	HEH6237-1-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	I31	04/04/2025	Chemtech -SO
Q1734-08	HED302R-1-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	04/04/2025	Chemtech -SO
Q1734-09	HED302R-2-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	04/04/2025	Chemtech -SO
Q1734-10	HED302R-2-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	04/04/2025	Chemtech -SO
Q1735-01	50660-50661-50662-5663-COM	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	04/04/2025	Chemtech -SO
Q1737-01	RT3069	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	04/04/2025	Chemtech -SO
Q1734-02	HEH6237-1-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	04/04/2025	Chemtech -SO
Q1734-03	STJ23-1-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	04/04/2025	Chemtech -SO
Q1734-04	STJ23-1-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	04/04/2025	Chemtech -SO
Q1734-05	HIA989S-1-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	04/04/2025	Chemtech -SO
Q1734-06	HIA989S-1-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	04/04/2025	Chemtech -SO
Q1734-07	HED302R-1-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	04/04/2025	Chemtech -SO
Q1738-01	OK-02-040425	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	04/04/2025	Chemtech -SO
Q1738-02	OK-02-040425-E2	Solid	Percent Solids	Cool 4 deg C	PSEG05	L31	04/04/2025	Chemtech -SO
Q1738-02	OK-02-040425-E2	Solid	Percent Solids	Cool 4 deg C	PSEG05	L31	04/04/2025	Chemtech -SO

Date/Time 04/04/25 15:20

Raw Sample Received by: [signature]

Raw Sample Relinquished by: [signature]

Date/Time 04/04/25 17:15

Raw Sample Received by: [signature]

Raw Sample Relinquished by: [signature]

Alliance TECHNICAL GROUP

EXTRACTION LOGPAGE

PB167469

SOP ID: M3541-ASE Extraction-14

Clean Up SOP #: Acid Cleanup

Matrix : Solid

Weigh By: EH

Balance check: RJ

Balance ID: EX-SC-2

pH Strip Lot#: N/A

Extraction By: RJ

Filter By: RJ

pH Meter ID: N/A

Hood ID: 3,7

Extraction Start Date : 04/07/2025

Extraction Start Time : 08:20

Extraction End Date : 04/07/2025

Extraction End Time : 11:30

Concentration By: EH

Supervisor By : rajesh

Extraction Method: [_] Seperatory Funnel [_] Continious Liquid/Liquid [_] Sonication [_] Waste Dilution [√] Soxhlet

Standared Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	1.0ML	5000 PPB	PP24328
Surrogate	1.0ML	200 PPB	PP24217
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Hexane/Acetone/1:1	N/A	EP2592
Baked Na2SO4	N/A	EP2597
Sand	N/A	E2865
Hexane	N/A	E3916
H2SO4 1:1	N/A	EP2565
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

40 ML Vial lot# 03-40 BTS721. Q1735-01 Limited volume used as sample is Oily matrix.

KD Bath ID: N/A

KD Bath Temperature: N/A

Envap ID: NEVAP-02

Envap Temperature: 40 °C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
04/04/25	RP (Ext. Lab)	Y.P. Pest + PCB.
11:35	Preparation Group	Analysis Group

Alliance TECHNICAL GROUP

EXTRACTION LOGPAGE

PrepBatch ID : PB167469

Analytical Method: M3541-ASE Extraction-14 **Concentration Date:** 04/07/2025

Sample ID	Client Sample ID	Test	g / mL	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB167469BL	ABLK469	PCB	30.01	N/A	ritesh	Evelyn	10			U7-1
PB167469BS	ALCS469	PCB	30.02	N/A	ritesh	Evelyn	10			2
Q1730-01	OU4-VSL-15-040325	PCB	30.07	N/A	ritesh	Evelyn	10	E		3
Q1730-03	OU4-VSL-16-040325	PCB	30.03	N/A	ritesh	Evelyn	10	E		4
Q1730-03MS	OU4-VSL-16-040325MS	PCB	30.05	N/A	ritesh	Evelyn	10	E		5
Q1730-03MSD	OU4-VSL-16-040325MSD	PCB	30.02	N/A	ritesh	Evelyn	10	E		6
Q1730-05	OU4-VSL-17-040325	PCB	30.08	N/A	ritesh	Evelyn	10	E		U2-1
Q1730-07	OU4-PCS-TC-21-040325	PCB	30.01	N/A	ritesh	Evelyn	10	E		2
Q1730-09	OU4-PCS-TC-22-040325	PCB	30.02	N/A	ritesh	Evelyn	10	E		3
Q1730-11	OU4-PCS-TC-23-040325	PCB	30.05	N/A	ritesh	Evelyn	10	E		4
Q1730-13	OU4-PCS-TC-24-040325	PCB	30.08	N/A	ritesh	Evelyn	10	E		5
Q1730-15	OU4-PCS-TC-25-040325	PCB	30.06	N/A	ritesh	Evelyn	10	E		6
Q1730-17	OU4-PCS-TC-26-040325	PCB	30.03	N/A	ritesh	Evelyn	10	E		U3-1
Q1730-19	OU4-CF-15-040325	PCB	30.02	N/A	ritesh	Evelyn	10	E		2
Q1732-01	TT-8	PCB	30.04	N/A	ritesh	Evelyn	10	D		3
Q1733-01	ETGI-328	PCB	30.06	N/A	ritesh	Evelyn	10	E	Small Particle	4
Q1735-01	50660-50661-50662-5663-COMP	PCB	5.06	N/A	ritesh	Evelyn	10	E	Oily Debris	5
Q1736-01	GST1	PCB	30.07	N/A	ritesh	Evelyn	10			6
Q1736-02	GST2	PCB	30.03	N/A	ritesh	Evelyn	10			U6-1
Q1737-01	RT3069	PCB	30.05	N/A	ritesh	Evelyn	10	D		2
Q1738-01	OK-02-040425	PCB	30.01	N/A	ritesh	Evelyn	10	E		3

* **Extracts relinquished on the same date as received.**

4/7/25

16742
8:25

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Q1732 **WorkList ID :** 188759 **Department :** Extraction **Date :** 04-07-2025 08:15:59

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1730-01	OU4-VSL-15-040325	Solid	Pesticide-TCL	Cool 4 deg C	NOBI03	L31	04/03/2025	8081B
Q1730-03	OU4-VSL-16-040325	Solid	Pesticide-TCL	Cool 4 deg C	NOBI03	L31	04/03/2025	8081B
Q1730-05	OU4-VSL-17-040325	Solid	Pesticide-TCL	Cool 4 deg C	NOBI03	L31	04/03/2025	8081B
Q1730-07	OU4-PCS-TC-21-040325	Solid	Pesticide-TCL	Cool 4 deg C	NOBI03	L31	04/03/2025	8081B
Q1730-09	OU4-PCS-TC-22-040325	Solid	Pesticide-TCL	Cool 4 deg C	NOBI03	L31	04/03/2025	8081B
Q1730-11	OU4-PCS-TC-23-040325	Solid	Pesticide-TCL	Cool 4 deg C	NOBI03	L31	04/03/2025	8081B
Q1730-13	OU4-PCS-TC-24-040325	Solid	Pesticide-TCL	Cool 4 deg C	NOBI03	L31	04/03/2025	8081B
Q1730-15	OU4-PCS-TC-25-040325	Solid	Pesticide-TCL	Cool 4 deg C	NOBI03	L31	04/03/2025	8081B
Q1730-17	OU4-PCS-TC-26-040325	Solid	Pesticide-TCL	Cool 4 deg C	NOBI03	L31	04/03/2025	8081B
Q1730-19	OU4-CF-15-040325	Solid	Pesticide-TCL	Cool 4 deg C	NOBI03	L31	04/03/2025	8081B
Q1732-01	TT-8	Solid	Pesticide-TCL	Cool 4 deg C	PSEG03	L31	04/04/2025	8081B
Q1733-01	ETGI-328	Solid	Pesticide-TCL	Cool 4 deg C	PSEG03	I31	04/04/2025	8081B
Q1737-01	RT3069	Solid	Pesticide-TCL	Cool 4 deg C	PSEG03	L31	04/04/2025	8081B
Q1738-01	OK-02-040425	Solid	Pesticide-TCL	Cool 4 deg C	PSEG05	L31	04/04/2025	8081B

Date/Time 04/07/25 8:14
Raw Sample Received by: RJ (Ecol 1043)
Raw Sample Relinquished by: CR SM

Date/Time 04/07/25 9:00
Raw Sample Received by: CR SM
Raw Sample Relinquished by: RJ (Ecol 1043)

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Prep Standard - Chemical Standard Summary

Order ID : Q1730

Test : PCB

Prepbatch ID : PB167469,

Sequence ID/Qc Batch ID: PO040725,PO040825,PP040825,

Standard ID :

EP2565,EP2592,EP2597,PP24217,PP24328,PP24329,PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369,PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387,

Chemical ID :

E2865,E3551,E3804,E3876,E3877,E3916,M5173,P11522,P12699,P12702,P12931,P12936,P12948,P12949,P12957,P13354,P13356,P13373,P13381,P13589,P13591,P13697,P13702,P13830,P13878,P13883,W3112,W3177,

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Extractions STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
314	1.1 H2SO4 SOLN	EP2565	11/20/2024	05/20/2025	Rajesh Parikh	None	None	RUPESHKUMAR SHAH 11/20/2024

FROM 1000.00000ml of M5173 + 1000.00000ml of W3112 = Final Quantity: 2000.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
230	1:1ACETONE/HEXANE	EP2592	02/27/2025	08/12/2025	RUPESHKUMAR SHAH	None	None	Riteshkumar Patel 02/27/2025

FROM 4000.00000ml of E3876 + 4000.00000ml of E3877 = Final Quantity: 8000.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Extractions STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3923	Baked Sodium Sulfate	EP2597	03/28/2025	07/01/2025	Rajesh Parikh	Extraction_SCALE_2 (EX-SC-2)	None	Evelyn Huang 03/28/2025

FROM 4000.00000gram of E3551 = Final Quantity: 4000.000 gram

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
465	200 PPB Pest/PCB Surrogate Spike	PP24217	03/05/2025	08/25/2025	Abdul Mirza	None	None	Yogesh Patel 03/06/2025

FROM 1.00000ml of P13354 + 999.00000ml of E3876 = Final Quantity: 1000.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3857	5000 PPB PCB SPIKE SOLUTION 2ND SOURCE	PP24328	03/17/2025	08/25/2025	Abdul Mirza	None	None	Yogesh Patel 04/02/2025

FROM 0.50000ml of P12948 + 99.50000ml of E3876 = Final Quantity: 100.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
84	Pest/PCB Surrogate Stock 20 PPM	PP24329	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 1.00000ml of P13356 + 9.00000ml of W3177 = Final Quantity: 10.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
202	AR1660 1000/100 ppb working solution 1st source	PP24330	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 0.10000ml of P13697 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
203	AR1660 750 PPB STD	PP24331	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24330 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
204	AR1660 500 PPB STD	PP24332	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24330 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
205	AR1660 250 PPB STD	PP24333	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24330 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
206	AR1660 50 PPB STD	PP24334	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24332 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
213	AR1221 1000 PPB WORKING SOLUTION	PP24335	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 0.10000ml of P13702 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1079	AR1221 750 PPB STD	PP24336	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24335 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
222	AR1221 500 PPB STD	PP24337	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24335 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1080	AR1221 250 PPB STD	PP24338	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24335 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1081	AR1221 50 PPB STD	PP24339	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24337 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
214	AR1232 1000 PPB WORKING SOLUTION	PP24340	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 0.10000ml of P13878 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1063	AR1232 750 PPB STD	PP24341	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24340 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
223	AR1232 500 PPB STD	PP24342	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24340 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1064	AR1232 250 PPB STD	PP24343	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24340 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1065	AR1232 50 PPB STD	PP24344	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24342 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
215	AR1242 1000 PPB WORKING STD	PP24345	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 0.10000ml of P12931 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1067	AR1242 750 PPB STD	PP24346	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 0.75000ml of W3177 + 0.75000ml of PP24345 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
224	AR1242 500 PPB STD	PP24347	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24345 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1068	AR1242 250 PPB STD	PP24348	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24345 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1069	AR1242 50 PPB STD	PP24349	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24347 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
216	AR1248 1000 PPB WORKING STD	PP24350	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 0.10000ml of P12936 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1075	AR1248 750 PPB STD	PP24351	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24350 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
225	AR1248 500 PPB STD	PP24352	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24350 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1076	AR1248 250 PPB STD	PP24353	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24350 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1077	AR1248 50 PPB STD	PP24354	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24352 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
217	AR1254 1000 PPB WORKING STD	PP24355	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 0.10000ml of P13830 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1071	AR1254 750 PPB STD	PP24356	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24355 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
226	AR1254 500 PPB STD	PP24357	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24355 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1072	AR1254 250 PPB STD	PP24358	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24355 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1073	AR1254 50 PPB STD	PP24359	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24357 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1529	AR1262 1000 PPB Working Solution	PP24360	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 0.10000ml of P13883 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3753	AR1262 750 PPB STD	PP24361	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24360 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1530	AR1262 500 PPB STD	PP24362	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24360 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3754	AR1262 250 PPB STD	PP24363	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24360 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3755	AR1262 50 PPB STD	PP24364	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24362 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1532	AR1268 1000 PPB Working Solution	PP24365	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 0.10000ml of P13381 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3820	AR1268 750 PPB STD	PP24366	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24365 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1533	AR1268 500 PPB STD	PP24367	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24365 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3821	AR1268 250 PPB STD	PP24368	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24365 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3822	AR1268 50 PPB STD	PP24369	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24367 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
404	AR1660 100 PPM Stock Solution 2nd Source	PP24370	03/18/2025	09/18/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 1.00000ml of P12949 + 9.00000ml of E3804 = Final Quantity: 10.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
405	AR1660 1000/100 PPB ICV STD	PP24371	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 98.50000ml of W3177 + 0.50000ml of PP24329 + 1.00000ml of PP24370 = Final Quantity: 100.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
406	AR1660 500 PPB ICV	PP24372	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24371 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3789	AR1221 1000 PPB WORKING SOL.2ND SOURCE(AGILENT)	PP24373	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 1.00000ml of P13373 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1886	AR1221 500 PPB ICV	PP24374	03/18/2025	08/12/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 0.50000ml of E3877 + 0.50000ml of W3177 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1887	AR1232 1000 PPB Working Sol. 2nd Source	PP24375	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 1.00000ml of P12699 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1888	AR1232 500 PPB ICV	PP24376	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24375 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1889	AR1242 1000 PPB Working Sol. 2nd Source	PP24377	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 1.00000ml of P13589 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1891	AR1242 500 PPB ICV	PP24378	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24377 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1890	AR1248 1000 PPB Working Sol. 2nd Source	PP24379	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 1.00000ml of P13591 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1892	AR1248 500 PPB ICV	PP24380	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24379 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1893	AR1254 1000 PPB Working Sol. 2nd Source	PP24381	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 1.00000ml of P12957 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1894	AR1254 500 PPB ICV	PP24382	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24381 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3757	AR1262 1000 PPB Working Solution second source	PP24384	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 1.00000ml of P12702 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3758	AR1262 500 PPB STD ICV	PP24385	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24384 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3817	AR1268 1000 ppb Working Soln. 2nd source	PP24386	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 1.00000ml of P11522 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3823	AR1268 500 PPB STD ICV	PP24387	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24386 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-3382-05 / Sand, Purified (cs/4x2.5kg)	0000243821	06/30/2025	04/30/2020 / RAJESH	04/28/2020 / RAJESH	E2865

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19631-100 / SODIUM SULFATE, ANHYDROUS, PEST GRADE, 1	313201	07/01/2025	01/03/2024 / Rajesh	07/20/2023 / Rajesh	E3551

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	9005-05 / Acetone Ultra (cs/4x4L)	24E0761004	11/05/2025	10/01/2024 / Rajesh	09/25/2024 / Rajesh	E3804

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H2762008	08/25/2025	02/25/2025 /	02/12/2025 / Rajesh	E3876

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	243570	08/12/2025	02/12/2025 / Rajesh	02/12/2025 / Rajesh	E3877

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	243570	10/03/2025	04/03/2025 / Rajesh	03/31/2025 / Rajesh	E3916

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9673-33 / Sulfuric Acid, Instra-Analyzed (cs/6c2.5L)	0000281827	06/02/2025	06/01/2022 /	04/05/2022 / william	M5173

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	PP-382-1 / Aroclor 1268	0006587800	09/18/2025	03/18/2025 / yogesh	02/21/2022 / Ankita	P11522

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards,Inc	91867 / Aroclor 1232 100 ug/mL	020823	09/18/2025	03/18/2025 / yogesh	08/07/2023 / Ankita	P12699

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards,Inc	x9166 / Aroclor 1262 100 ug/mL	060523	09/18/2025	03/18/2025 / yogesh	08/07/2023 / Ankita	P12702

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32009 / PCB Mix, Aroclor 1242, 1000ug/mL, Hexane, 1mL/ampul	a0203672	09/18/2025	03/18/2025 / yogesh	12/07/2023 / Ankita	P12931

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32010 / PCB Mix, Aroclor 1248, 1000ug/mL, Hexane, 1mL/ampul	a0202803	09/18/2025	03/18/2025 / yogesh	12/07/2023 / Ankita	P12936

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	20064 / Aroclor 1016/1260	022023	08/27/2025	02/27/2025 / Ankita	12/20/2023 / Yogesh	P12948

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	20064 / Aroclor 1016/1260	022023	09/18/2025	03/18/2025 / yogesh	12/20/2023 / Yogesh	P12949

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ Arochlor 1254	121823	04/03/2025	10/03/2024 / Ankita	12/20/2023 / Yogesh	P12957

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32000 / Pesticide Mix, CLP method, Pesticide Surrogate Mix, 200ug/mL, Acetone, 1mL	A0206810	09/05/2025	03/05/2025 / Abdul	04/22/2024 / Abdul	P13354

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32000 / Pesticide Mix, CLP method, Pesticide Surrogate Mix, 200ug/mL, Acetone, 1mL	A0206810	09/18/2025	03/18/2025 / yogesh	04/22/2024 / Abdul	P13356

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	PP-292-1 / Aroclor 1221	0006783205	09/18/2025	03/18/2025 / yogesh	05/02/2024 / Ankita	P13373

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32410 / PCB Stock Solution, Aroclor 1268 Std, 1mL, Hexane	A0207475	09/18/2025	03/18/2025 / yogesh	05/03/2024 / Abdul	P13381

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	PP-312-1 / Aroclor 1242	0006665550	09/18/2025	03/18/2025 / yogesh	10/14/2024 / Ankita	P13589

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	PP-342-1 / Aroclor 1248	0006726317	09/18/2025	03/18/2025 / yogesh	10/14/2024 / Ankita	P13591

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32039 / PCB Mix, Aroclor 1016/1260, 1000ug/mL, hexane, 1mL/ampul	A0210629	09/18/2025	03/18/2025 / yogesh	10/17/2024 / yogesh	P13697

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32007 / PCB Mix, Aroclor 1221, 1000ug/mL, Hexane, 1mL/ampul	A0215270	09/18/2025	03/18/2025 / yogesh	10/17/2024 / yogesh	P13702

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32011 / PCB Mix, Aroclor 1254, 1000ug/mL, Hexane, 1mL/ampul	A0217391	09/18/2025	03/18/2025 / yogesh	12/09/2024 / Ankita	P13830

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32008 / PCB Mix, Aroclor 1232, 1000ug/mL, Hexane, 1mL/ampul	A0219655	09/18/2025	03/18/2025 / yogesh	01/23/2025 / Ankita	P13878

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32409 / PCB Stock Solution, Aroclor 1262 Std, 1mL, Hexane	A0220950	09/18/2025	03/18/2025 / yogesh	01/23/2025 / Ankita	P13883

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / Iwona	07/03/2024 / Iwona	W3112

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	24G1962003	08/22/2025	02/03/2025 / jignesh	01/31/2025 / jignesh	W3177

Sand
Purified
Washed and Ignited

avantor™ | J.T.Baker®

Material No.: 3382-05
Batch No.: 0000243821
Manufactured Date: 2018/04/09
Retest Date: 2025/04/07
Revision No: 1

Certificate of Analysis

Test	Specification	Result
Substances Soluble in HCl	<= 0.16 %	0.01

For Laboratory, Research or Manufacturing Use
Meets Reagent Specifications for testing USP/NF monographs

Country of Origin: US
Packaging Site: Paris Mfg Ctr & DC

E 2865

James T Ethier

Jamie Ethier
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700
Avantor Performance Materials, LLC
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700

PRODUCTOS QUIMICOS MONTERREY, S.A. DE C.V.

MIRADOR 201, COL. MIRADOR
MONTERREY, N.L. MEXICO
CP 64070
TEL +52 81 13 52 57 57
www.pqm.com.mx

CERTIFICATE OF ANALYSIS

PRODUCT :	SODIUM SULFATE CRYSTALS ANHYDROUS		
QUALITY :	ACS (CODE RMB3375)	FORMULA :	Na_2SO_4
SPECIFICATION NUMBER :	6399	RELEASE DATE:	ABR/21/2023
LOT NUMBER :	313201		

TEST	SPECIFICATIONS	LOT VALUES
Assay (Na_2SO_4)	Min. 99.0%	99.7 %
pH of a 5% solution at 25°C	5.2 - 9.2	6.1
Insoluble matter	Max. 0.01%	0.005 %
Loss on ignition	Max. 0.5%	0.1 %
Chloride (Cl)	Max. 0.001%	<0.001 %
Nitrogen compounds (as N)	Max. 5 ppm	<5 ppm
Phosphate (PO_4)	Max. 0.001%	<0.001 %
Heavy metals (as Pb)	Max. 5 ppm	<5 ppm
Iron (Fe)	Max. 0.001%	<0.001 %
Calcium (Ca)	Max. 0.01%	0.002 %
Magnesium (Mg)	Max. 0.005%	0.001 %
Potassium (K)	Max. 0.008%	0.003 %
Extraction-concentration suitability	Passes test	Passes test
Appearance	Passes test	Passes test
Identification	Passes test	Passes test
Solubility and foreing matter	Passes test	Passes test
Retained on US Standard No. 10 sieve	Max. 1%	0.1 %
Retained on US Standard No. 60 sieve	Min. 94%	97.3 %
Through US Standard No. 60 sieve	Max. 5%	2.5 %
Through US Standard No. 100 sieve	Max. 10%	0.1 %

COMMENTS

QC: PhC Irma Belmares

If you need further details, please call our factory or contact our local distributor.

RG-02-01, Ed. 3

Recd. by RB on 7/24/23 E3551

Acetone
CMOS

avantor™ | J.T.Baker®

Material No.: 9005-05
Batch No.: 24E0761004
Manufactured Date: 2024-05-02
Retest Date: 2029-05-01
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ($(CH_3)_2CO$) (by GC, corrected for water)	≥ 99.5 %	99.8 %
Color (APHA)	≤ 10	< 5
Residue after Evaporation	≤ 5 ppm	< 1 ppm
Titrable Acid (µeq/g)	≤ 0.3	0.1
Titrable Base (µeq/g)	≤ 0.5	0.1
Water (H_2O)	≤ 0.5 %	0.1 %
Solubility in H_2O	Passes Test	Passes Test
Chloride (Cl)	≤ 0.2 ppm	< 0.2 ppm
Phosphate (PO_4)	≤ 0.05 ppm	< 0.05 ppm
Trace Impurities - Aluminum (Al)	≤ 50.0 ppb	< 5.0 ppb
Arsenic and Antimony (as As)	≤ 5.0 ppb	< 5.0 ppb
Trace Impurities - Barium (Ba)	≤ 20.0 ppb	< 1.0 ppb
Trace Impurities - Beryllium (Be)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities - Bismuth (Bi)	≤ 20.0 ppb	< 10.0 ppb
Trace Impurities - Boron (B)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities - Cadmium (Cd)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities - Calcium (Ca)	≤ 25.0 ppb	3.6 ppb
Trace Impurities - Chromium (Cr)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities - Cobalt (Co)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities - Copper (Cu)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities - Gallium (Ga)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities - Germanium (Ge)	≤ 10.0 ppb	< 10.0 ppb
Trace Impurities - Gold (Au)	≤ 20 ppb	< 5 ppb
Trace Impurities - Iron (Fe)	≤ 20.0 ppb	< 1.0 ppb
Trace Impurities - Lead (Pb)	≤ 10.0 ppb	< 10.0 ppb
Trace Impurities - Lithium (Li)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities - Magnesium (Mg)	≤ 20 ppb	< 1 ppb
Trace Impurities - Manganese (Mn)	≤ 10.0 ppb	< 1.0 ppb

>>> Continued on page 2 >>>

Recd by RP on 9/25/24

E 3804

Acetone
CMOS

avantor™ | J.T.Baker®

Material No.: 9005-05
Batch No.: 24E0761004

Test	Specification	Result
Trace Impurities – Molybdenum (Mo)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities – Nickel (Ni)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities – Niobium (Nb)	≤ 50.0 ppb	< 1.0 ppb
Trace Impurities – Potassium (K)	≤ 10.0 ppb	< 10.0 ppb
Trace Impurities – Silicon (Si)	≤ 50 ppb	< 10 ppb
Trace Impurities – Silver (Ag)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Sodium (Na)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Strontium (Sr)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Tantalum (Ta)	≤ 50.0 ppb	< 5.0 ppb
Trace Impurities – Thallium (Tl)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities – Tin (Sn)	≤ 20.0 ppb	< 10.0 ppb
Trace Impurities – Titanium (Ti)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Vanadium (V)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Zinc (Zn)	≤ 20.0 ppb	7.9 ppb
Trace Impurities – Zirconium (Zr)	≤ 10.0 ppb	< 1.0 ppb
Particle Count – 0.5 µm and greater (Rion KS42AF)	≤ 100 par/ml	8 par/ml
Particle Count – 1.0 µm and greater (Rion KS42AF)	≤ 8 par/ml	2 par/ml

>>> Continued on page 3 >>>

Acetone
CMOS

avantor™ | J.T.Baker®

Material No.: 9005-05
Batch No.: 24E0761004

Test	Specification	Result

For Microelectronic Use

Country of Origin: USA
Packaging Site: Paris Mfg Ctr & DC

Michelle Bales
Sr. Manager, Quality Assurance

Certificate of Analysis

1 Reagent Lane
Fair Lawn, NJ 07410
201.796.7100 tel
201.796.1329 fax

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System Standard ISO9001:2015 by SAI Global Certificate Number CERT – 0120633

This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Certain data have been supplied by third parties. Thermo Fisher Scientific expressly disclaims all warranties, expressed or implied, including the implied warranties of merchantability and fitness for a particular purpose. Products are for research use or further manufacturing. Not for direct administration to humans or animals. It is the responsibility of the final formulator and end user to determine suitability based upon the intended use of the end product. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained.

Catalog Number	H303	Quality Test / Release Date	11/07/2024
Lot Number	243570		
Description	HEXANES - OPTIMA		
Country of Origin	United States	Suggested Retest Date	Nov/2029
Chemical Origin	Organic - non animal		
BSE/TSE Comment	No animal products are used as starting raw material ingredients, or used in processing, including lubricants, processing aids, or any other material that might migrate to the finished product.		

N/A			
Result Name	Units	Specifications	Test Value
APPEARANCE		REPORT	Clear, colorless liquid
ASSAY (N-HEXANE)	%	>= 60	69
ASSAY (SUM C6 HYDROCARBONS)	%	>= 99.9	>99.9
COLOR	APHA	<= 5	<5
DENSITY AT 25 DEGREES C	GM/ML	Inclusive Between 0.653 - 0.673	0.669
EVAPORATION RESIDUE	ppm	<= 1	<1
FLUORESCENCE BACKGROUND	ppb	<= 1	<1
IDENTIFICATION	PASS/FAIL	= PASS TEST	PASS TEST
OPTICAL ABS AT 195 NM	ABS. UNITS	<= 1	0.74
OPTICAL ABS AT 210 NM	ABS. UNITS	<= 0.25	0.17
OPTICAL ABS AT 220 NM	ABS. UNITS	<= 0.07	0.05
OPTICAL ABS AT 254 NM	ABS. UNITS	<= 0.005	0.001
PESTICIDE RESIDUE ANALYSIS	NG/L	<= 10	<10
REFRACTIVE INDEX @ 25 DEG C		Inclusive Between 1.375 - 1.385	1.379
SUITABILITY FOR GC/MS		= PASS TEST	PASS TEST
SULFUR COMPOUNDS	%	<= 0.005	<0.005
THIOPHENE	PASS/FAIL	= PASS TEST	PASS TEST
WATER (H2O)	%	<= 0.01	<0.01
WATER-SOLUBLE TITRABLE ACID	MEQ/G	<= 0.0003	0.0001

Recd by RP on 2/12/25

E3877

Harout Sahagian - Quality Control Manager - Fair Lawn

Note: The data listed is valid for all package sizes of this lot of this product, expressed as an extension of this catalog number listed above.
If there are any questions with this certificate, please call at (800) 227-6701.
*Based on suggested storage condition.

Certificate of Analysis

ThermoFisher SCIENTIFIC

Certificate of Analysis

1 Reagent Lane
Fair Lawn, NJ 07410
201.796.7100 tel
201.796.1329 fax

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System Standard ISO9001:2015 by SAI Global Certificate Number CERT – 0120633

This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Certain data have been supplied by third parties. Thermo Fisher Scientific expressly disclaims all warranties, expressed or implied, including the implied warranties of merchantability and fitness for a particular purpose. Products are for research use or further manufacturing. Not for direct administration to humans or animals. It is the responsibility of the final formulator and end user to determine suitability based upon the intended use of the end product. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained.

Catalog Number	H303	Quality Test / Release Date	11/07/2024
Lot Number	243570		
Description	HEXANES - OPTIMA		
Country of Origin	United States	Suggested Retest Date	Nov/2029
Chemical Origin	Organic - non animal		
BSE/TSE Comment	No animal products are used as starting raw material ingredients, or used in processing, including lubricants, processing aids, or any other material that might migrate to the finished product.		

N/A

Result Name	Units	Specifications	Test Value
APPEARANCE		REPORT	Clear, colorless liquid
ASSAY (N-HEXANE)	%	>= 60	69
ASSAY (SUM C6 HYDROCARBONS)	%	>= 99.9	>99.9
COLOR	APHA	<= 5	<5
DENSITY AT 25 DEGREES C	GM/ML	Inclusive Between 0.653 - 0.673	0.669
EVAPORATION RESIDUE	ppm	<= 1	<1
FLUORESCENCE BACKGROUND	ppb	<= 1	<1
IDENTIFICATION	PASS/FAIL	= PASS TEST	PASS TEST
OPTICAL ABS AT 195 NM	ABS. UNITS	<= 1	0.74
OPTICAL ABS AT 210 NM	ABS. UNITS	<= 0.25	0.17
OPTICAL ABS AT 220 NM	ABS. UNITS	<= 0.07	0.05
OPTICAL ABS AT 254 NM	ABS. UNITS	<= 0.005	0.001
PESTICIDE RESIDUE ANALYSIS	NG/L	<= 10	<10
REFRACTIVE INDEX @ 25 DEG C		Inclusive Between 1.375 - 1.385	1.379
SUITABILITY FOR GC/MS		= PASS TEST	PASS TEST
SULFUR COMPOUNDS	%	<= 0.005	<0.005
THIOPHENE	PASS/FAIL	= PASS TEST	PASS TEST
WATER (H2O)	%	<= 0.01	<0.01
WATER-SOLUBLE TITRABLE ACID	MEQ/G	<= 0.0003	0.0001

Harout Sahagian - Quality Control Manager - Fair Lawn

Rec'd by RP on 3/31/25

E3916

Note: The data listed is valid for all package sizes of this lot of this product, expressed as an extension of this catalog number listed above.
If there are any questions with this certificate, please call at (800) 227-6701.
*Based on suggested storage condition.

Hydrochloric Acid, 36.5-38.0%
BAKER INSTRA-ANALYZED® Reagent
For Trace Metal Analysis

Material No.: 9530-33
Batch No.: 0000281827
Manufactured Date: 2021/03/30
Retest Date: 2026/03/29
Revision No: 1

Certificate of Analysis

Test	Specification	Result
ACS - Assay (as HCl) (by acid-base titrn)	36.5 - 38.0 %	37.6
ACS - Color (APHA)	<= 10	5
ACS - Residue after Ignition	<= 3 ppm	1
ACS - Specific Gravity at 60°/60°F	1.185 - 1.192	1.189
ACS - Bromide (Br)	<= 0.005 %	< 0.005
ACS - Extractable Organic Substances	<= 5 ppm	< 1
ACS - Free Chlorine (as Cl_2)	<= 0.5 ppm	< 0.5
Phosphate (PO_4)	<= 0.05 ppm	< 0.03
Sulfate (SO_4)	<= 0.5 ppm	< 0.3
Sulfite (SO_3)	<= 0.8 ppm	0.3
Ammonium (NH_4)	<= 3 ppm	< 1
Trace Impurities - Arsenic (As)	<= 0.010 ppm	< 0.003
Trace Impurities - Aluminum (Al)	<= 10.0 ppb	0.5
Arsenic and Antimony (as As)	<= 5 ppb	< 3
Trace Impurities - Barium (Ba)	<= 1.0 ppb	< 0.2
Trace Impurities - Beryllium (Be)	<= 1.0 ppb	< 0.2
Trace Impurities - Bismuth (Bi)	<= 10.0 ppb	< 1.0
Trace Impurities - Boron (B)	<= 20.0 ppb	< 5.0
Trace Impurities - Cadmium (Cd)	<= 1.0 ppb	< 0.3
Trace Impurities - Calcium (Ca)	<= 50.0 ppb	15.0
Trace Impurities - Chromium (Cr)	<= 1.0 ppb	< 0.4
Trace Impurities - Cobalt (Co)	<= 1.0 ppb	< 0.3
Trace Impurities - Copper (Cu)	<= 1.0 ppb	< 0.1
Trace Impurities - Gallium (Ga)	<= 1.0 ppb	< 0.2

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700
Avantor Performance Materials, LLC
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700

Material No.: 9530-33
Batch No.: 0000281827

Test	Specification	Result
Trace Impurities - Germanium (Ge)	<= 3.0 ppb	< 2.0
Trace Impurities - Gold (Au)	<= 4.0 ppb	3.0
Heavy Metals (as Pb)	<= 100 ppb	< 50
Trace Impurities - Iron (Fe)	<= 15.0 ppb	1.0
Trace Impurities - Lead (Pb)	<= 1.0 ppb	< 0.5
Trace Impurities - Lithium (Li)	<= 1.0 ppb	< 0.2
Trace Impurities - Magnesium (Mg)	<= 10.0 ppb	< 0.4
Trace Impurities - Manganese (Mn)	<= 1.0 ppb	< 0.4
Trace Impurities - Mercury (Hg)	<= 0.5 ppb	0.2
Trace Impurities - Molybdenum (Mo)	<= 10.0 ppb	< 5.0
Trace Impurities - Nickel (Ni)	<= 4.0 ppb	< 0.3
Trace Impurities - Niobium (Nb)	<= 1.0 ppb	< 0.2
Trace Impurities - Potassium (K)	<= 9.0 ppb	< 2.0
Trace Impurities - Selenium (Se), For Information Only	ppb	1.0
Trace Impurities - Silicon (Si)	<= 100.0 ppb	18.0
Trace Impurities - Silver (Ag)	<= 1.0 ppb	< 0.3
Trace Impurities - Sodium (Na)	<= 100.0 ppb	< 5.0
Trace Impurities - Strontium (Sr)	<= 1.0 ppb	< 0.2
Trace Impurities - Tantalum (Ta)	<= 1.0 ppb	< 0.9
Trace Impurities - Thallium (Tl)	<= 5.0 ppb	< 2.0
Trace Impurities - Tin (Sn)	<= 5.0 ppb	< 0.8
Trace Impurities - Titanium (Ti)	<= 1.0 ppb	< 0.2
Trace Impurities - Vanadium (V)	<= 1.0 ppb	< 0.2
Trace Impurities - Zinc (Zn)	<= 5.0 ppb	0.4
Trace Impurities - Zirconium (Zr)	<= 1.0 ppb	< 0.1

For Laboratory, Research or Manufacturing Use
Product Information (not specifications):
Appearance (clear, fuming liquid)
Meets ACS Specifications

Country of Origin: US
Packaging Site: Phillipsburg Mfg Ctr & DC



For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700
Avantor Performance Materials, LLC
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700

P11518 ↓ P11522 AJ 02/21/22

Agilent

Certificate of Analysis

Product Name: Aroclor 1268 Standard

Product Number: PP-382-1 **Lot Issue Date:** 09-Feb-2021

Lot Number: 0006587800 **Expiration Date:** 31-Mar-2029

Description:
This analytical reference material (RM) was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	Concentration ± Uncertainty
Aroclor 1268	011100-14-4	RM00937	100.0 ± 0.5 µg/mL

Matrix: isooctane (2,2,4-trimethylpentane)

Storage Conditions: Store at Room Temperature (15° to 30°C).

Traceability:
The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:
This RM was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Intended Use:
This RM is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Instructions for Use:
Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Hazards:
Refer to the Safety Data Sheet on www.agilent.com for information regarding this RM.

Expiration of Certification:
The certification of this RM is valid until the expiration date specified above, provided the RM is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the RM is damaged, contaminated, or otherwise modified.

Maintenance of Certification:
If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:

M Bourgeois

Monica Bourgeois
QMS Representative

ANAB ACCREDITED REFERENCE MATERIAL PRODUCER
ISO 17034 Cert No. AR-1936

RM was produced in accordance with TUV USA Inc registered ISO 9001 Quality Management System. Cert # 56 100 18560026

www.agilent.com/quality/
CSD-QA-015.1

ANAB ACCREDITED TESTING LABORATORY
ISO 17025 Cert No. AT-1937

250 Smith Street North Kingstown, Rhode Island 02852 www.agilent.com/quality

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Refere

CERTIFIED WEIGHT REPORT

Part Number:	91867
Lot Number:	020823
Description:	WP 037 - Aroclor 1232 PCB Technical Mixture
Expiration Date:	020833
Recommended Storage:	Ambient (20 °C)
Nominal Concentration (µg/mL):	100
NIST Test ID#:	6UTB

Solvent(
Aceton


5E-05 Balance Uncertainty
0.057 Flask Uncertainty

Weight(s) shown below were combined and diluted to (mL): 100.0

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g
1. Aroclor 1232	17	45-6A	100	100	0.5	0.01000

- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measure Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Comments
GC3-M1 Analysis by Melissa Stonier
Column ID SPB-608 30 meter X 0.53mm X5µm film thickness
Flow rates: Helium (carrier) = 5mL/min, Helium (make-up) = 25mL/min
Hydrogen (make-up) = 30mL/min, Air (make-up) = 350mL/min
Oven Profile: Temp 1 = 150°C (Time 1 = 4 min), Temp 2 = 290°C (Time 2 = 13.5 min)
Rate = 8°C/min, Total run time = 35 min
Injector temp. = 200°C, FID Temp. = 300°C. FID Signal = Edaq Channel 1
Standard injection =1.5µL, Range=3



RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32009 **Lot No.:** A0203672

Description : Aroclor® 1242 Standard

Aroclor® 1242 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : January 31, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

P12928
↓
P12932
AJ
12/07/23

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1242	53469-21-9	01141	----%	1,004.7 µg/mL	+/- 55.7515

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Quality Confirmation Test

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
10 ml/min.

Inj. Vol
0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Russ Bookhamer - Operations Technician I

Date Mixed: 26-Oct-2023 **Balance Serial #** B442140311

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 06-Nov-2023

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ilac-MRA A2LA ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ilac-MRA A2LA ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

P12933 ↓ P12937 AJ 12/07/23

Catalog No. : 32010 **Lot No.:** A0202803

Description : Aroclor® 1248 Standard

Aroclor® 1248 Standard 1,000µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : January 31, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1248	12672-29-6	13897600	----%	1,001.7 µg/mL	+/- 55.5850

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Quality Confirmation Test

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
10 ml/min.

Inj. Vol
0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Laith Clemente - Operations Technician I **Date Mixed:** 03-Oct-2023 **Balance Serial #** 1128360905

Jennifer Pollino - Operations Tech III - ARM QC **Date Passed:** 09-Oct-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT

Part Number:	20064
Lot Number:	022023
Description:	CLP PCB'S - Aroclor Mix
	Aroclors 1016 & 1260
Expiration Date:	022033
Recommended Storage:	Ambient (20 °C)
Nominal Concentration (µg/mL):	1000
NIST Test ID#:	6UTB

Solvent(s):	Lot#
Hexane	273615

5E-05 Balance Uncertainty

Weight(s) shown below were combined and diluted to (mL): 200.0 0.010 Flask Uncertainty


Formulated By:	Benson Chan	022023 DATE
Reviewed By:	Pedro L. Rentas	022023 DATE

P12946 ↓ P12955 — Y.P. 12/20/23

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50
1. Aroclor 1016	15	020491JC	1000	100	0.2	0.20004	0.20060	1002.8	4.0	12674-11-2	N/A	N/A
2. Aroclor 1260	21	020491JC	1000	100	0.2	0.20004	0.20081	1003.9	4.0	11096-82-5	0.5mg/m3	orl-rat 1315mg/kg

- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Comments
GC3-M1 Analysis by Melissa Stonier
Column ID SPB-608 30 meter X 0.53mm X5µm film thickness
Flow rates: Helium (carrier) = 5mL/min, Helium (make-up) = 25mL/min
Hydogen (make-up) = 30mL/min, Air (make-up) = 350mL/min
Oven Profile: Temp 1 = 150°C (Time 1 = 4 min), Temp 2 = 290°C (Time 2 = 13.5 min)
Rate = 8°C/min, Total run time = 35 min
Injector temp. = 200°C, FID Temp. = 300°C. FID Signal = Edaq Channel 1
Standard injection =1.5µL, Range=3



Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT

Part Number:	20064
Lot Number:	022023
Description:	CLP PCB'S - Aroclor Mix
	Aroclors 1016 & 1260
Expiration Date:	022033
Recommended Storage:	Ambient (20 °C)
Nominal Concentration (µg/mL):	1000
NIST Test ID#:	6UTB

Solvent(s):	Lot#
Hexane	273615

5E-05 Balance Uncertainty

Weight(s) shown below were combined and diluted to (mL): 200.0 0.010 Flask Uncertainty


Formulated By:	Benson Chan	022023 DATE
Reviewed By:	Pedro L. Rentas	022023 DATE

P12946 ↓ P12955 — Y.P. 12/20/23

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50
1. Aroclor 1016	15	020491JC	1000	100	0.2	0.20004	0.20060	1002.8	4.0	12674-11-2	N/A	N/A
2. Aroclor 1260	21	020491JC	1000	100	0.2	0.20004	0.20081	1003.9	4.0	11096-82-5	0.5mg/m3	orl-rat 1315mg/kg

- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Comments
GC3-M1 Analysis by Melissa Stonier
Column ID SPB-608 30 meter X 0.53mm X5µm film thickness
Flow rates: Helium (carrier) = 5mL/min, Helium (make-up) = 25mL/min
Hydogen (make-up) = 30mL/min, Air (make-up) = 350mL/min
Oven Profile: Temp 1 = 150°C (Time 1 = 4 min), Temp 2 = 290°C (Time 2 = 13.5 min)
Rate = 8°C/min, Total run time = 35 min
Injector temp. = 200°C, FID Temp. = 300°C. FID Signal = Edaq Channel 1
Standard injection =1.5µL, Range=3



Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT

Part Number:	99139	**Solvent(s):**	**Lot#**	
Lot Number:	121823	Iso-octane	82227	
Description:	Aroclor 1254			
Expiration Date:	121833			
Recommended Storage:	Ambient (20 °C)			
Nominal Concentration (µg/mL):	100			
NIST Test ID#:	6UTB	5E-05	Balance Uncertainty	
Volume(s) shown below were combined and diluted to (mL):		20.0	0.003	Flask Uncertainty

		121823
Formulated By:	Anthony Mahoney	DATE
		121823
Reviewed By:	Pedro L. Rentas	DATE

P12956 & P12957 Y.P. 12/20/23


Note: Aroclor 1254 is a mix of isomers.

Compound	Part Number	Lot Number	Dilution Factor	Initial Vol. (mL)	Uncertainty Pipette (mL)	Initial Conc.(µg/mL)	Final Conc.(µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50
1. Aroclor 1254	79100	121823	0.10	2.00	0.017	1003.3	100.1	1.8	11097-69-1	0.5mg/m3 (skin)	orl-rat 1295mg/kg

- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Comments
GC3-M1 Analysis by Melissa Stonier
Column ID SPB-608 30 meter X 0.53mm X5µm film thickness
Flow rates: Helium (carrier) = 5mL/min, Helium (make-up) = 25mL/min
Hydrogen (make-up) = 30mL/min, Air (make-up) = 350mL/min
Oven Profile: Temp 1 = 150°C (Time 1 = 4 min), Temp 2 = 290°C (Time 2 = 13.5 min)
Rate = 8°C/min, Total run time = 35 min
Injector temp. = 200°C, FID Temp. = 300°C. FID Signal = Edaq Channel 1
Standard injection =1.5µL, Range=3

Peak No.	Name	FID RT (min.)
1	Total Aroclor 1254	18.12



RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

P13348 ↓ P13357 (10)

04/25/2024

Catalog No. : 32000 **Lot No.:** A0206810

Description : Pesticide Surrogate Mix

Pesticide Surrogate Mix 200 µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2030 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	200.3 µg/mL	+/- 11.1143
2	Decachlorobiphenyl (BZ# 209)	2051-24-3	30638	99%	200.6 µg/mL	+/- 11.1298

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Acetone
CAS # 67-64-1
Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
hellum-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
10 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Laith Clemente - Operations Technician I

Date Mixed: 22-Jan-2024 **Balance Serial #** 1128360905

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Jan-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

P13348 (10)

↓

P13357

04/25/2025

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

P13348 ↓ P13357 (10)

04/25/2024

Catalog No. : 32000 **Lot No.:** A0206810

Description : Pesticide Surrogate Mix

Pesticide Surrogate Mix 200 µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2030 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	200.3 µg/mL	+/- 11.1143
2	Decachlorobiphenyl (BZ# 209)	2051-24-3	30638	99%	200.6 µg/mL	+/- 11.1298

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Acetone
CAS # 67-64-1
Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
hellum-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
10 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Laith Clemente - Operations Technician I

Date Mixed: 22-Jan-2024 **Balance Serial #** 1128360905

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Jan-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

P13348 (10)

↓

P13357

04/25/2025

Agilent

Trusted Answers

ISO 17034

Reference Material Certificate

Product Information Sheet

Product Name:	Aroclor 1221 Standard	**Lot Number:**	0006783205
Product Number:	PP-292-1	**Lot Issue Date:**	20-Feb-2024
Storage Conditions:	Store at Room Temperature (15° to 30°C).	**Expiration Date:**	31-Mar-2032

Component Name	Concentration		Uncertainty	CAS#	Analyte Lot
Aroclor 1221	100.3	±	0.5 µg/mL	011104-28-2	NT01017

Matrix: isooctane (2,2,4-trimethylpentane)

Description:
This document is prepared in accordance with ISO 17034 and Guide 31. This analytical reference material standard was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed above.

Traceability:
The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:
This analytical reference standard was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Instructions for Use:
Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Safety:
Refer to the Safety Data Sheet on www.agilent.com for information regarding this analytical reference material.

Intended Use:
This analytical reference standard is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Expiration of Certification:
The certification of this analytical reference standard is valid until the expiration date specified above, provided the material is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the material is damaged, contaminated, or otherwise modified.

P133f2
↓
P133f3

AJ
05106124

250 Smith Street North Kingstown, Rhode Island 02852 **www.agilent.com/quality**

Agilent

Trusted Answers

Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:

Monica Bourgeois

QMS Representative

ANAB ANSI National Accreditation Board ACCREDITED REFERENCE MATERIAL PRODUCER

ISO 17034
Cert No. AR-1936

RM was produced in accordance with the TUV/SUD registered ISO 9001:2015 Quality Management System. Cert# 951215321

www.agilent.com/quality/

CSD-QA-015.2

ISO 17025
Cert No. AT-1937

250 Smith Street North Kingstown, Rhode Island 02852 **www.agilent.com/quality**

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

ilac-MRA A2LA ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ilac-MRA A2LA ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

Certificate of Analysis

chromatographic plus

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32410 **Lot No.:** A0207475

Description : Aroclor® 1268 Standard

Aroclor® 1268 Standard 1,000 µg/mL, 1mL/ampul, Hexane

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : May 31, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1268	11100-14-4	10947000	----%	1,000.0 µg/mL	+/- 55.4925

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

P 13380 2
P 13381
1
05/6/2024

Quality Confirmation Test

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
Split ratio 500:1

Inj. Vol
0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Michael Maye - Operations Tech I

Date Mixed: 06-Feb-2024 **Balance Serial #** B442140311

Dillan Murphy - Operations Technician I

Date Passed: 09-Feb-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

P13380 (2)
↓
P13381
1

05/6/2024

ISO 17034

Agilent
Trusted Answers

Reference Material Certificate

Product Name:	Aroclor 1242 Standard	**Lot Number:**	0006665550
Product Number:	PP-312-1	**Lot Issue Date:**	08-Feb-2022
Storage Conditions:	Store at Room Temperature (15° to 30°C).	**Expiration Date:**	31-Jan-2027

	CERTIFIED VALUES				
Component Name	Concentration		Expanded Uncertainty	CAS#	Analyte Lot
Aroclor 1242	100.4	±	0.5 µg/mL	053469-21-9	NT01020

Matrix: isooctane (2,2,4-trimethylpentane)

Description:
This document is prepared in accordance with ISO 17034 and Guide 31. This analytical reference material standard was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed above.

Traceability:
The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:
This analytical reference standard was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Instructions for Use:
Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Safety:
Refer to the Safety Data Sheet on www.agilent.com for information regarding this analytical reference material.

Intended Use:
This analytical reference standard is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Expiration of Certification:
The certification of this analytical reference standard is valid until the expiration date specified above, provided the material is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the material is damaged, contaminated, or otherwise modified.

P13589
↓
P13590

AJ
10/14/24

Maintenance of Certification:
If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:

Monica Bourgeois
QMS Representative

Agilent
Trusted Answers

ISO 17034

Reference Material Certificate

Product Information Sheet

Product Name: Aroclor 1248 Standard
Product Number: PP-342-1
Storage Conditions: Store at Room Temperature (15° to 30°C).

Lot Number: 0006726317
Lot Issue Date: 27-Jan-2023
Expiration Date: 28-Feb-2031

Component Name	Concentration		Uncertainty	CAS#	Analyte Lot
Aroclor 1248	100.3	±	0.5 µg/mL	012672-29-6	NT01582

Matrix: isooctane (2,2,4-trimethylpentane)

Description:
This document is prepared in accordance with ISO 17034 and Guide 31. This analytical reference material (RM) standard was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed above. Purity values are taken from approved vendor raw material certificates.

Traceability:
The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:
This analytical reference (RM) standard was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Instructions for Use:
Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Safety:
Refer to the Safety Data Sheet on www.agilent.com for information regarding this analytical reference material.

Intended Use:
This analytical reference (RM) standard is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Expiration of Certification:
The certification of this analytical reference standard (RM) is valid until the expiration date specified above, provided the material is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the material is damaged, contaminated, or otherwise modified.

P13591
↓
P13592

AJ
10/14/2024

Agilent

Trusted Answers

Maintenance of Certification:
If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:

Monica Bourgeois
QMS Representative

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32039 **Lot No.:** A0210629

Description : Aroclor® 1016/1260 Mix

Aroclor® 1016/1260 Mix 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

P13697 ↓ P13701 } Y.P. 10/19/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1016	12674-11-2	07	----%	1,005.3 µg/mL	+/- 55.7809
2	Aroclor 1260	11096-82-5	1320657	----%	1,000.0 µg/mL	+/- 55.4850

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Quality Confirmation Test

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
10 ml/min.

Inj. Vol
0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Laith Clemente - Operations Technician I

Date Mixed: 22-Apr-2024 **Balance Serial #** B442140311

Dillan Murphy - Operations Technician I

Date Passed: 24-Apr-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

Certificate of Analysis

chromatographic plus

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32007 **Lot No.:** A0215270

Description : Aroclor® 1221 Standard

Aroclor® 1221 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : November 30, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

P13702 & P13703 } Y.P. 10/17/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1221	11104-28-2	14969200	----%	1,005.0 µg/mL	+/- 55.7700

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Quality Confirmation Test

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
10 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Michael Maye - Operations Tech I

Date Mixed: 16-Aug-2024 **Balance Serial #** 1128360905

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 20-Aug-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ilac-MRA A2LA ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ilac-MRA A2LA ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32011 **Lot No.:** A0217391

Description : Aroclor® 1254 Standard

Aroclor® 1254 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : January 31, 2031 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1254	11097-69-1	124-191-B	----%	1,004.7 µg/mL	+/- 55.7515

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

P13830
↓
P13832

AJ
12/09/24

Quality Confirmation Test

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
300 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Michael Maye - Operations Tech I **Date Mixed:** 02-Oct-2024 **Balance Serial #** C322230531

Jennifer Pollino - Operations Tech III - ARM QC **Date Passed:** 07-Oct-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32008 **Lot No.:** A0219655

Description : Aroclor® 1232 Standard

Aroclor® 1232 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : March 31, 2031 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1232	11141-16-5	15665-01	----%	1,007.0 µg/mL	+/- 55.8810

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

P13878
↓
P13880

AJ
01/28/25

Quality Confirmation Test

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
10 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Michael Maye - Operations Tech I

Date Mixed: 02-Dec-2024 **Balance Serial #** C322230531

Brittany Federinko - Operations Tech I

Date Passed: 05-Dec-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ilac-MRA A2LA ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ilac-MRA A2LA ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32409 **Lot No.:** A0220950

Description : Aroclor® 1262 Standard

Aroclor® 1262 Standard 1,000 µg/mL, 1mL/ampul, Hexane

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2031 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1262	37324-23-5	10849100	----%	1,002.0 µg/mL	+/- 55.6035

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

P138852
↓
P138853

AJ
01/28/25

Quality Confirmation Test

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
300 ml/min.

Inj. Vol
0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Tom Suckar - Mix Technician

Date Mixed: 09-Jan-2025 **Balance Serial #** C322230531

Brittany Federinko - Operations Tech I

Date Passed: 14-Jan-2025

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis

avantor™ | J.T.Baker®

W3147

W3147 update. 02/03/2025 JP

Material No.: 9262-03
Batch No.: 24G1962003
Manufactured Date: 2024-05-23
Expiration Date: 2025-08-22
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	3
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	≤ 10	1
ECD-Sensitive Impurities (as Ethylene Dibromide) - Single Impurity Peak (ng/mL)	≤ 5	1
Assay (Total Saturated C_6 Isomers) (by GC, corrected for water)	≥ 99.5 %	99.7 %
Assay (as n-Hexane) (by GC, corrected for water)	≥ 95 %	98 %
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Substances Darkened by H_2SO_4	Passes Test	Passes Test
Water (by KF, coulometric)	≤ 0.05 %	< 0.01 %

For Laboratory,Research,or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

JCroak

Jamie Croak
Director Quality Operations, Bioscience Production



SHIPPING DOCUMENTS

Q1730

Chemtech

Phone: (908) 789-8900
Fax: (908) 789-8922
284 Sheffield Street, Mountainside, NJ 07092

http://www.contestlabs.com

CHAIN OF CUSTODY RECORD

39 Spruce Street
East Longmeadow, MA 01028

Doc # 381 Rev 4_01/08/2020

Page 1 of 1

Company Name:	Nobis Group
Address:	55 Technology Dr Suite 101, Lowell, MA 01851
Phone:	978-703-6014
Project Name:	Raymark
Project Location:	Stratford, CT
Project Number:	95700
Project Manager:	Adam Roy
Con-Test Quote Name/Number:	
Invoice Recipient:	
Sampled By:	C. Odell

Requested Turnaround Time
5-Day ☐ 10-Day ☑
PFAS 10-Day (std) ☐ Due Date:

Rush-Approval Required
1-Day ☐ 3-Day ☐
2-Day ☐ 4-Day ☐

Data Delivery
Format: PDF ☑ EXCEL ☑
Other:
CLP Like Data Pkg Required: ☐ No
Email To: aroy@nobis-group.com
Fax To #:

Dissolved Metals Samples
○ Field Filtered
○ Lab to Filter

Orthophosphate Samples
○ Field Filtered
○ Lab to Filter

PCB ONLY
SOXHLET ☑
NON SOXHLET ☐

ANALYSIS REQUESTED

Con-Test Work Order#	Client Sample ID / Description	Beginning Date/Time	Ending Date/Time	COMP/GRAB	[1]Matrix Code	Conc Code	VIALS	GLASS	PLASTIC	BACTERIA	ENCORE	M/O	RCP VOCs	% Solids	PAHs	Herbicides	Pesticides	PCBs	Metals ICP + Hg - 6010	Cyanide	SPLP RCP Metals - 6020
	OU4-VSL-15-040325	4/3/25	1210	G	S		3	2	1				X	X	X	X	X	X	X	X	X
	OU4-VSL-16-040325	4/3/25	1220	G	S		3	2	1				X	X	X	X	X	X	X	X	X
	OU4-VSL-17-040325	4/3/25	1230	G	S		3	2	1				X	X	X	X	X	X	X	X	X
	OU4-PCS-TC-21-040325	4/3/25	1345	G	S		3	2	1				X	X	X	X	X	X	X	X	X
	OU4-PCS-TC-22-040325	4/3/25	1350	G	S		3	2	1				X	X	X	X	X	X	X	X	X
	OU4-PCS-TC-23-040325	4/3/25	1355	G	S		3	2	1				X	X	X	X	X	X	X	X	X
	OU4-PCS-TC-24-040325	4/3/25	1400	G	S		3	2	1				X	X	X	X	X	X	X	X	X
	OU4-PCS-TC-25-040325	4/3/25	1405	G	S		3	2	1				X	X	X	X	X	X	X	X	X
	OU4-PCS-TC-26-040325	4/3/25	1410	G	S		3	2	1				X	X	X	X	X	X	X	X	X
	OU4-CF-15-040325	4/3/25	1300	G	S		3	2	1				X	X	X	X	X	X	X	X	X

Relinquished by: (signature) — Date/Time: 4/3/25 1700 — Client Comments:
Received by: (signature) — Date/Time: 4/4/25 1045 — 2.7 ; 2.6
Relinquished by: (signature) — Date/Time:
Received by: (signature) — Date/Time:
Relinquished by: (signature) — Date/Time:
Received by: (signature) — Date/Time:
Relinquished by: (signature) — Date/Time:
Received by: (signature) — Date/Time:

Detection Limit Requirements
MA
CT
Other:

Special Requirements
☐ MA MCP Required
MCP Certification Form Required
☑ CT RCP Required
RCP Certification Form Required
☐ MA State DW Required
PWSID #

Please use the following codes to indicate possible sample concentration within the Conc Code column above:
H - High; M - Medium; L - Low; C - Clean; U - Unknown

NELAC and AIHA-LAP, LLC Accredited

Project Entity
Government ☐ Federal ☐ City ☐
Municipality ☐ 21 J ☐ Brownfield ☐
MWRA ☐ School ☐ MBTA ☐
WRTA ☐

Other
☐ Chromatogram
☐ *AIHA-LAP,LLC*

Lab Comments:

Disclaimer: Con-Test Labs is not responsible for any omitted information on the Chain of Custody. The Chain of Custody is a legal document that must be complete and accurate and is used to determine what analyses the laboratory will perform. Any missing information is not the laboratory's responsibility. Con-Test values your partnership on each project and will try to assist with missing information, but will not be held accountable.

[2] Preservation Code

Contest Use Only
Total Number Of:
VIALS_____
GLASS_____
PLASTIC_____
BACTERIA_____
ENCORE_____

Glassware in the fridge? Y / N
Glassware in freezer? Y / N
Prepackaged Cooler? Y / N
*Contest is not responsible for missing samples from prepacked coolers

[1] **Matrix Codes:**
GW = Ground Water
WW = Waste Water
DW = Drinking Water
A = Air
S = Soil
SL = Sludge
SOL = Solid
O = Other (please define)

[2] **Preservation Codes:**
I = Iced
H = HCL
M = Methanol
N = Nitric Acid
S = Sulfuric Acid
B = Sodium Bisulfate
X = Sodium Hydroxide
T = Sodium Thiosulfate
O = Other (please define)

284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488

QA Control Code: A2070148

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q1730 NOBI03
Client Name : Nobis Group
Client Contact : Adam Roy
Invoice Name : Nobis Group
Invoice Contact : Adam Roy

Order Date : 4/4/2025 10:51:00 AM
Project Name : Raymark Superfund Site
Receive DateTime : 4/4/2025 10:45:00 AM
Purchase Order :

Project Mgr : Yazmeen
Report Type : Level 4
EDD Type : EQUIS
Hard Copy Date :
Date Signoff : 4/4/2025 11:24:53 AM

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q1730-01	OU4-VSL-15-040325	Solid	04/03/2025	12:10	VOCMS Group3		8260D		10 Bus. Days
Q1730-03	OU4-VSL-16-040325	Solid	04/03/2025	12:20	VOCMS Group3		8260D		10 Bus. Days
Q1730-05	OU4-VSL-17-040325	Solid	04/03/2025	12:30	VOCMS Group3		8260D		10 Bus. Days
Q1730-07	OU4-PCS-TC-21-040325	Solid	04/03/2025	13:45	VOCMS Group3		8260D		10 Bus. Days
Q1730-09	OU4-PCS-TC-22-040325	Solid	04/03/2025	13:50	VOCMS Group3		8260D		10 Bus. Days
Q1730-11	OU4-PCS-TC-23-040325	Solid	04/03/2025	13:55	VOCMS Group3		8260D		10 Bus. Days
Q1730-13	OU4-PCS-TC-24-040325	Solid	04/03/2025	14:00	VOCMS Group3		8260D		10 Bus. Days
Q1730-15	OU4-PCS-TC-25-040325	Solid	04/03/2025	14:05	VOCMS Group3		8260D		10 Bus. Days

Alliance TECHNICAL GROUP

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q1730 NOBI03
Client Name : Nobis Group
Client Contact : Adam Roy
Invoice Name : Nobis Group
Invoice Contact : Adam Roy

Order Date : 4/4/2025 10:51:00 AM
Project Name : Raymark Superfund Site
Receive DateTime : 4/4/2025 10:45:00 AM
Purchase Order :

Project Mgr : Yazmeen
Report Type : Level 4
EDD Type : EQUIS
Hard Copy Date :
Date Signoff : 4/4/2025 11:24:53 AM

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
					VOCMS Group3		8260D	10 Bus. Days	
Q1730-17	OU4-PCS-TC-26-040325	Solid	04/03/2025	14:10					
					VOCMS Group3		8260D	10 Bus. Days	
Q1730-19	OU4-CF-15-040325	Solid	04/03/2025	13:00					
					VOCMS Group3		8260D	10 Bus. Days	

Relinguished By :
Date / Time : 4/4/25 1145

Received By :
Date / Time : 4/4/25 14:15

Storage Area : **VOA Refridgerator Room**

Quantitation Report (Not Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109973.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 14:21
Operator : YP/AJ
Sample : AR1660ICC750
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 18 16:36:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 18 16:30:31 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.693	3.690	705.7E6	403.4E6	75.190	74.986
2)	SA Decachlor...	8.745	8.696	568.9E6	170.8E6	75.465	72.950
	Target Compounds						
3)	L1 AR-1016-1	4.785	4.772	246.8E6	133.8E6	747.915	745.960
4)	L1 AR-1016-2	4.805	4.791	347.8E6	195.6E6	753.172	750.675
5)	L1 AR-1016-3	4.862	4.967	233.7E6	104.0E6	751.394	748.099
6)	L1 AR-1016-4	4.982	5.009	186.4E6	84297055	751.636	746.609
7)	L1 AR-1016-5	5.239	5.222	197.8E6	110.8E6	753.927	750.256
31)	L7 AR-1260-1	6.280	6.255	350.7E6	185.8E6	755.816	746.573
32)	L7 AR-1260-2	6.470	6.442	429.8E6	216.7E6	759.623	745.669
33)	L7 AR-1260-3	6.837	6.595	362.6E6	209.2E6	757.414	751.587
34)	L7 AR-1260-4	7.098	7.067	310.3E6	154.0E6	751.251	745.577
35)	L7 AR-1260-5	7.340	7.307	800.9E6	394.6E6	755.153	803.094

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Quantitation Report (Not Reviewed)


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109973.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 14:21
Operator : YP/AJ
Sample : AR1660ICC750
Misc :
ALS Vial : 4 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
AR1660ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 18 16:36:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 18 16:30:31 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.693 min
Delta R.T.: 0.000 min
Response: 705727801
Conc: 75.19 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC750




#1 Tetrachloro-m-xylene

R.T.: 3.690 min
Delta R.T.: 0.000 min
Response: 403376580
Conc: 74.99 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.745 min
Delta R.T.: 0.000 min
Response: 568890214
Conc: 75.47 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.696 min
Delta R.T.: 0.000 min
Response: 170836588
Conc: 72.95 ng/ml



#3 AR-1016-1


R.T.: 4.785 min
Delta R.T.: 0.000 min
Response: 246831835
Conc: 747.92 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC750




#3 AR-1016-1

R.T.: 4.772 min
Delta R.T.: 0.000 min
Response: 133795284
Conc: 745.96 ng/ml




#4 AR-1016-2

R.T.: 4.805 min
Delta R.T.: 0.000 min
Response: 347782380
Conc: 753.17 ng/ml



#4 AR-1016-2


R.T.: 4.791 min
Delta R.T.: 0.000 min
Response: 195632208
Conc: 750.68 ng/ml



#5 AR-1016-3


R.T.: 4.862 min
Delta R.T.: 0.000 min
Response: 233678232
Conc: 751.39 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC750




#5 AR-1016-3

R.T.: 4.967 min
Delta R.T.: 0.000 min
Response: 103988837
Conc: 748.10 ng/ml




#6 AR-1016-4

R.T.: 4.982 min
Delta R.T.: 0.000 min
Response: 186394664
Conc: 751.64 ng/ml



#6 AR-1016-4


R.T.: 5.009 min
Delta R.T.: 0.000 min
Response: 84297055
Conc: 746.61 ng/ml



#7 AR-1016-5


R.T.: 5.239 min
Delta R.T.: 0.000 min
Response: 197756047
Conc: 753.93 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC750




#7 AR-1016-5

R.T.: 5.222 min
Delta R.T.: 0.000 min
Response: 110764402
Conc: 750.26 ng/ml




#31 AR-1260-1

R.T.: 6.280 min
Delta R.T.: 0.000 min
Response: 350689884
Conc: 755.82 ng/ml



#31 AR-1260-1


R.T.: 6.255 min
Delta R.T.: 0.000 min
Response: 185810579
Conc: 746.57 ng/ml



#32 AR-1260-2


R.T.: 6.470 min
Delta R.T.: 0.000 min
Response: 429776784
Conc: 759.62 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC750




#32 AR-1260-2

R.T.: 6.442 min
Delta R.T.: 0.000 min
Response: 216749564
Conc: 745.67 ng/ml




#33 AR-1260-3

R.T.: 6.837 min
Delta R.T.: 0.000 min
Response: 362605874
Conc: 757.41 ng/ml



#33 AR-1260-3


R.T.: 6.595 min
Delta R.T.: 0.000 min
Response: 209216211
Conc: 751.59 ng/ml



#34 AR-1260-4


R.T.: 7.098 min
Delta R.T.: 0.000 min
Response: 310296283
Conc: 751.25 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC750




#34 AR-1260-4

R.T.: 7.067 min
Delta R.T.: 0.000 min
Response: 154008109
Conc: 745.58 ng/ml



#35 AR-1260-5

R.T.: 7.340 min
Delta R.T.: 0.000 min
Response: 800891670
Conc: 755.15 ng/ml



#35 AR-1260-5

R.T.: 7.307 min
Delta R.T.: 0.000 min
Response: 394605629
Conc: 803.09 ng/ml

Quantitation Report (Not Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109976.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 15:16
Operator : YP/AJ
Sample : AR1660ICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 18 16:42:31 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 18 16:30:31 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1) SA	Tetrachlo...	3.691	3.688	39998137	22828400	4.395	4.353
2) SA	Decachlor...	8.744	8.695	38817794	12425222	5.049	5.121
	Target Compounds						
3) L1	AR-1016-1	4.784	4.771	16902653	9160493	50.224	49.988
4) L1	AR-1016-2	4.804	4.790	22043406	12469957	47.540	47.685
5) L1	AR-1016-3	4.860	4.965	16411868	7085785	51.193	49.978
6) L1	AR-1016-4	4.981	5.008	12569483	6253080	49.743	53.024
7) L1	AR-1016-5	5.238	5.220	14890720	7955288	54.174	52.023
31) L7	AR-1260-1	6.279	6.253	27190975	14406182	57.396	56.300
32) L7	AR-1260-2	6.468	6.441	33768251	15676301	56.704	52.205
33) L7	AR-1260-3	6.836	6.594	24202394	18027606	49.871	62.952 #
34) L7	AR-1260-4	7.096	7.065	21433117	11320638	50.530	52.635
35) L7	AR-1260-5	7.339	7.306	47957612	24758089	45.855	49.646

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Quantitation Report (Not Reviewed)


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109976.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 15:16
Operator : YP/AJ
Sample : AR1660ICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
AR1660ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 18 16:42:31 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 18 16:30:31 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.691 min
Delta R.T.: 0.000 min
Response: 39998137
Conc: 4.39 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC050




#1 Tetrachloro-m-xylene

R.T.: 3.688 min
Delta R.T.: 0.000 min
Response: 22828400
Conc: 4.35 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.744 min
Delta R.T.: 0.000 min
Response: 38817794
Conc: 5.05 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.695 min
Delta R.T.: 0.000 min
Response: 12425222
Conc: 5.12 ng/ml



#3 AR-1016-1


R.T.: 4.784 min
Delta R.T.: 0.000 min
Response: 16902653
Conc: 50.22 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC050




#3 AR-1016-1

R.T.: 4.771 min
Delta R.T.: 0.000 min
Response: 9160493
Conc: 49.99 ng/ml




#4 AR-1016-2

R.T.: 4.804 min
Delta R.T.: 0.000 min
Response: 22043406
Conc: 47.54 ng/ml



#4 AR-1016-2


R.T.: 4.790 min
Delta R.T.: 0.000 min
Response: 12469957
Conc: 47.69 ng/ml



#5 AR-1016-3


R.T.: 4.860 min
Delta R.T.: 0.000 min
Response: 16411868
Conc: 51.19 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC050




#5 AR-1016-3

R.T.: 4.965 min
Delta R.T.: 0.000 min
Response: 7085785
Conc: 49.98 ng/ml




#6 AR-1016-4

R.T.: 4.981 min
Delta R.T.: 0.000 min
Response: 12569483
Conc: 49.74 ng/ml



#6 AR-1016-4


R.T.: 5.008 min
Delta R.T.: 0.000 min
Response: 6253080
Conc: 53.02 ng/ml



#7 AR-1016-5


R.T.: 5.238 min
Delta R.T.: 0.000 min
Response: 14890720
Conc: 54.17 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC050




#7 AR-1016-5

R.T.: 5.220 min
Delta R.T.: 0.000 min
Response: 7955288
Conc: 52.02 ng/ml




#31 AR-1260-1

R.T.: 6.279 min
Delta R.T.: 0.000 min
Response: 27190975
Conc: 57.40 ng/ml



#31 AR-1260-1


R.T.: 6.253 min
Delta R.T.: 0.000 min
Response: 14406182
Conc: 56.30 ng/ml



#32 AR-1260-2


R.T.: 6.468 min
Delta R.T.: 0.000 min
Response: 33768251
Conc: 56.70 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC050




#32 AR-1260-2

R.T.: 6.441 min
Delta R.T.: 0.000 min
Response: 15676301
Conc: 52.20 ng/ml




#33 AR-1260-3

R.T.: 6.836 min
Delta R.T.: 0.000 min
Response: 24202394
Conc: 49.87 ng/ml



#33 AR-1260-3


R.T.: 6.594 min
Delta R.T.: 0.000 min
Response: 18027606
Conc: 62.95 ng/ml



#34 AR-1260-4


R.T.: 7.096 min
Delta R.T.: 0.000 min
Response: 21433117
Conc: 50.53 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC050




#34 AR-1260-4

R.T.: 7.065 min
Delta R.T.: 0.000 min
Response: 11320638
Conc: 52.63 ng/ml



#35 AR-1260-5

R.T.: 7.339 min
Delta R.T.: 0.000 min
Response: 47957612
Conc: 45.86 ng/ml



#35 AR-1260-5

R.T.: 7.306 min
Delta R.T.: 0.000 min
Response: 24758089
Conc: 49.65 ng/ml

Quantitation Report (Not Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109983.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 17:25
Operator : YP/AJ
Sample : AR1242ICC050
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 18 17:34:54 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 18 17:34:29 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.692	3.689	27043743	15260251	3.233	3.182
2)	SA Decachlor...	8.744	8.696	32408304	10337597	4.500	4.624
	Target Compounds						
16)	L4 AR-1242-1	4.784	4.772	12333502	6612130	46.123	45.222
17)	L4 AR-1242-2	4.804	4.791	16261465	8959332	44.164	43.066
18)	L4 AR-1242-3	4.859	4.967	12804073	4612927	49.638	41.322
19)	L4 AR-1242-4	4.981	5.051	8657409	5726073	43.398	49.373
20)	L4 AR-1242-5	5.634	5.572	14229681	8157544	67.609	59.247


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109983.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 17:25
Operator : YP/AJ
Sample : AR1242ICC050
Misc :
ALS Vial : 14 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
AR1242ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 18 17:34:54 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 18 17:34:29 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.692 min
Delta R.T.: 0.000 min
Response: 27043743
Conc: 3.23 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC050




#1 Tetrachloro-m-xylene

R.T.: 3.689 min
Delta R.T.: 0.000 min
Response: 15260251
Conc: 3.18 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.744 min
Delta R.T.: 0.000 min
Response: 32408304
Conc: 4.50 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.696 min
Delta R.T.: 0.000 min
Response: 10337597
Conc: 4.62 ng/ml



#16 AR-1242-1


R.T.: 4.784 min
Delta R.T.: 0.000 min
Response: 12333502
Conc: 46.12 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC050




#16 AR-1242-1

R.T.: 4.772 min
Delta R.T.: 0.000 min
Response: 6612130
Conc: 45.22 ng/ml




#17 AR-1242-2

R.T.: 4.804 min
Delta R.T.: 0.000 min
Response: 16261465
Conc: 44.16 ng/ml



#17 AR-1242-2


R.T.: 4.791 min
Delta R.T.: 0.000 min
Response: 8959332
Conc: 43.07 ng/ml



#18 AR-1242-3


R.T.: 4.859 min
Delta R.T.: 0.000 min
Response: 12804073
Conc: 49.64 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC050




#18 AR-1242-3

R.T.: 4.967 min
Delta R.T.: 0.000 min
Response: 4612927
Conc: 41.32 ng/ml




#19 AR-1242-4

R.T.: 4.981 min
Delta R.T.: 0.000 min
Response: 8657409
Conc: 43.40 ng/ml



#19 AR-1242-4


R.T.: 5.051 min
Delta R.T.: 0.000 min
Response: 5726073
Conc: 49.37 ng/ml



#20 AR-1242-5

R.T.: 5.634 min
Delta R.T.: 0.000 min
Response: 14229681
Conc: 67.61 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC050



#20 AR-1242-5

R.T.: 5.572 min
Delta R.T.: 0.000 min
Response: 8157544
Conc: 59.25 ng/ml

Quantitation Report (Not Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109988.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 18:54
Operator : YP/AJ
Sample : AR1248ICC050
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 01:40:58 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 01:38:33 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.690	3.689	43788277	25352976	4.721	4.788
2)	SA Decachlor...	8.742	8.695	41310164	13029410	5.300	5.557
	Target Compounds						
21)	L5 AR-1248-1	4.782	4.771	11846805	6686511	53.863	55.572
22)	L5 AR-1248-2	5.021	5.008	16385888	9031427	54.426	53.593
23)	L5 AR-1248-3	5.236	5.050	22145957	10026048	59.149	55.979
24)	L5 AR-1248-4	5.591	5.221	29227773	12176632	55.856	58.382
25)	L5 AR-1248-5	5.632	5.613	20545156	11669383	55.569	57.295

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Quantitation Report (Not Reviewed)


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109988.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 18:54
Operator : YP/AJ
Sample : AR1248ICC050
Misc :
ALS Vial : 19 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
AR1248ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 01:40:58 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 01:38:33 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.690 min
Delta R.T.: -0.002 min
Response: 43788277
Conc: 4.72 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC050




#1 Tetrachloro-m-xylene

R.T.: 3.689 min
Delta R.T.: 0.000 min
Response: 25352976
Conc: 4.79 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.742 min
Delta R.T.: -0.002 min
Response: 41310164
Conc: 5.30 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.695 min
Delta R.T.: 0.000 min
Response: 13029410
Conc: 5.56 ng/ml



#21 AR-1248-1


R.T.: 4.782 min
Delta R.T.: -0.003 min
Response: 11846805
Conc: 53.86 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC050




#21 AR-1248-1

R.T.: 4.771 min
Delta R.T.: 0.000 min
Response: 6686511
Conc: 55.57 ng/ml




#22 AR-1248-2

R.T.: 5.021 min
Delta R.T.: -0.002 min
Response: 16385888
Conc: 54.43 ng/ml



#22 AR-1248-2


R.T.: 5.008 min
Delta R.T.: 0.000 min
Response: 9031427
Conc: 53.59 ng/ml



#23 AR-1248-3


R.T.: 5.236 min
Delta R.T.: -0.003 min
Response: 22145957
Conc: 59.15 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC050




#23 AR-1248-3

R.T.: 5.050 min
Delta R.T.: 0.000 min
Response: 10026048
Conc: 55.98 ng/ml




#24 AR-1248-4

R.T.: 5.591 min
Delta R.T.: -0.002 min
Response: 29227773
Conc: 55.86 ng/ml



#24 AR-1248-4


R.T.: 5.221 min
Delta R.T.: 0.000 min
Response: 12176632
Conc: 58.38 ng/ml



#25 AR-1248-5

R.T.: 5.632 min
Delta R.T.: -0.002 min
Response: 20545156
Conc: 55.57 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC050



#25 AR-1248-5

R.T.: 5.613 min
Delta R.T.: 0.000 min
Response: 11669383
Conc: 57.30 ng/ml

Quantitation Report (Not Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109993.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 20:25
Operator : YP/AJ
Sample : AR1254ICC050
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 01:50:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 01:48:40 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.692	3.689	43445665	25306006	4.628	4.733
2)	SA Decachlor...	8.743	8.694	41266656	13227136	5.245	5.469
	Target Compounds						
26)	L6 AR-1254-1	5.593	5.573	31329450	17649915	56.014	57.357
27)	L6 AR-1254-2	5.742	5.720	26381750	15388259	54.541	57.314
28)	L6 AR-1254-3	6.148	6.123	39484048	21241930	50.026	50.431
29)	L6 AR-1254-4	6.377	6.350	23571245	12628885	48.456	51.676
30)	L6 AR-1254-5	6.797	6.768	28479203	19977415	41.849	57.297 #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109993.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 20:25
Operator : YP/AJ
Sample : AR1254ICC050
Misc :
ALS Vial : 24 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
AR1254ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 01:50:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 01:48:40 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.692 min
Delta R.T.: 0.000 min
Response: 43445665
Conc: 4.63 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC050




#1 Tetrachloro-m-xylene

R.T.: 3.689 min
Delta R.T.: 0.000 min
Response: 25306006
Conc: 4.73 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.743 min
Delta R.T.: 0.000 min
Response: 41266656
Conc: 5.25 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.694 min
Delta R.T.: -0.002 min
Response: 13227136
Conc: 5.47 ng/ml



#26 AR-1254-1


R.T.: 5.593 min
Delta R.T.: 0.000 min
Response: 31329450
Conc: 56.01 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC050




#26 AR-1254-1

R.T.: 5.573 min
Delta R.T.: 0.000 min
Response: 17649915
Conc: 57.36 ng/ml




#27 AR-1254-2

R.T.: 5.742 min
Delta R.T.: 0.000 min
Response: 26381750
Conc: 54.54 ng/ml



#27 AR-1254-2


R.T.: 5.720 min
Delta R.T.: 0.000 min
Response: 15388259
Conc: 57.31 ng/ml



#28 AR-1254-3


R.T.: 6.148 min
Delta R.T.: 0.000 min
Response: 39484048
Conc: 50.03 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC050




#28 AR-1254-3

R.T.: 6.123 min
Delta R.T.: 0.000 min
Response: 21241930
Conc: 50.43 ng/ml




#29 AR-1254-4

R.T.: 6.377 min
Delta R.T.: 0.000 min
Response: 23571245
Conc: 48.46 ng/ml



#29 AR-1254-4


R.T.: 6.350 min
Delta R.T.: 0.000 min
Response: 12628885
Conc: 51.68 ng/ml



#30 AR-1254-5

R.T.: 6.797 min
Delta R.T.: 0.000 min
Response: 28479203
Conc: 41.85 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC050



#30 AR-1254-5

R.T.: 6.768 min
Delta R.T.: 0.000 min
Response: 19977415
Conc: 57.30 ng/ml

Quantitation Report (Not Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109997.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 21:39
Operator : YP/AJ
Sample : AR1268ICC500
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1268ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 02:06:41 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:04:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.692	3.689	463.9E6	264.1E6	50.000	50.000
2)	SA Decachlor...	8.743	8.694	706.0E6	207.2E6	50.000	50.000
	Target Compounds						
41)	L9 AR-1268-1	7.623	7.589	699.8E6	289.1E6	498.680	500.000
42)	L9 AR-1268-2	7.688	7.653	639.8E6	266.3E6	500.000	500.000
43)	L9 AR-1268-3	7.896	7.861	519.3E6	206.5E6	500.000	500.000
44)	L9 AR-1268-4	8.185	8.145	223.7E6	81011977	500.000	500.000
45)	L9 AR-1268-5	8.482	8.439	1614.3E6	493.6E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Quantitation Report (Not Reviewed)


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109997.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 21:39
Operator : YP/AJ
Sample : AR1268ICC500
Misc :
ALS Vial : 28 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
AR1268ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 02:06:41 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:04:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.692 min
Delta R.T.: 0.000 min
Response: 463875726
Conc: 50.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.689 min
Delta R.T.: 0.000 min
Response: 264095871
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.743 min
Delta R.T.: 0.000 min
Response: 706011522
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.694 min
Delta R.T.: 0.000 min
Response: 207212938
Conc: 50.00 ng/ml



#41 AR-1268-1


R.T.: 7.623 min
Delta R.T.: 0.000 min
Response: 699811647
Conc: 498.68 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC500




#41 AR-1268-1

R.T.: 7.589 min
Delta R.T.: 0.000 min
Response: 289102857
Conc: 500.00 ng/ml




#42 AR-1268-2

R.T.: 7.688 min
Delta R.T.: 0.000 min
Response: 639816344
Conc: 500.00 ng/ml



#42 AR-1268-2


R.T.: 7.653 min
Delta R.T.: 0.000 min
Response: 266318012
Conc: 500.00 ng/ml



#43 AR-1268-3


R.T.: 7.896 min
Delta R.T.: 0.000 min
Response: 519324293
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC500




#43 AR-1268-3

R.T.: 7.861 min
Delta R.T.: 0.000 min
Response: 206458519
Conc: 500.00 ng/ml




#44 AR-1268-4

R.T.: 8.185 min
Delta R.T.: 0.000 min
Response: 223719543
Conc: 500.00 ng/ml



#44 AR-1268-4


R.T.: 8.145 min
Delta R.T.: 0.000 min
Response: 81011977
Conc: 500.00 ng/ml



#45 AR-1268-5

R.T.: 8.482 min
Delta R.T.: 0.000 min
Response: 1614304087
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC500



#45 AR-1268-5

R.T.: 8.439 min
Delta R.T.: 0.000 min
Response: 493570508
Conc: 500.00 ng/ml

Quantitation Report (Not Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109998.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 21:57
Operator : YP/AJ
Sample : AR1268ICC250
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1268ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 02:06:56 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:04:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.692	3.690	233.0E6	133.0E6	25.112	25.182
2)	SA Decachlor...	8.743	8.695	364.2E6	110.3E6	25.791	26.624
	Target Compounds						
41)	L9 AR-1268-1	7.623	7.590	361.1E6	149.6E6	257.304	258.814
42)	L9 AR-1268-2	7.688	7.654	330.4E6	137.2E6	258.196	257.558
43)	L9 AR-1268-3	7.897	7.861	264.1E6	107.0E6	254.288	259.017
44)	L9 AR-1268-4	8.184	8.146	118.3E6	42174671	264.470	260.299
45)	L9 AR-1268-5	8.482	8.439	811.9E6	254.0E6	251.483	257.325

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Quantitation Report (Not Reviewed)


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109998.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 21:57
Operator : YP/AJ
Sample : AR1268ICC250
Misc :
ALS Vial : 29 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
AR1268ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 02:06:56 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:04:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.692 min
Delta R.T.: 0.000 min
Response: 232978846
Conc: 25.11 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.690 min
Delta R.T.: 0.000 min
Response: 133010597
Conc: 25.18 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.743 min
Delta R.T.: 0.000 min
Response: 364169510
Conc: 25.79 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.695 min
Delta R.T.: 0.001 min
Response: 110337267
Conc: 26.62 ng/ml



#41 AR-1268-1


R.T.: 7.623 min
Delta R.T.: 0.001 min
Response: 361082289
Conc: 257.30 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC250




#41 AR-1268-1

R.T.: 7.590 min
Delta R.T.: 0.000 min
Response: 149647905
Conc: 258.81 ng/ml




#42 AR-1268-2

R.T.: 7.688 min
Delta R.T.: 0.000 min
Response: 330395950
Conc: 258.20 ng/ml



#42 AR-1268-2


R.T.: 7.654 min
Delta R.T.: 0.001 min
Response: 137184821
Conc: 257.56 ng/ml



#43 AR-1268-3


R.T.: 7.897 min
Delta R.T.: 0.000 min
Response: 264115362
Conc: 254.29 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC250




#43 AR-1268-3

R.T.: 7.861 min
Delta R.T.: 0.000 min
Response: 106952452
Conc: 259.02 ng/ml




#44 AR-1268-4

R.T.: 8.184 min
Delta R.T.: 0.000 min
Response: 118334390
Conc: 264.47 ng/ml



#44 AR-1268-4


R.T.: 8.146 min
Delta R.T.: 0.000 min
Response: 42174671
Conc: 260.30 ng/ml



#45 AR-1268-5

R.T.: 8.482 min
Delta R.T.: 0.000 min
Response: 811940473
Conc: 251.48 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC250



#45 AR-1268-5

R.T.: 8.439 min
Delta R.T.: 0.000 min
Response: 254015607
Conc: 257.32 ng/ml

Quantitation Report (Not Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109999.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 22:15
Operator : YP/AJ
Sample : AR1268ICC050
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1268ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 02:07:11 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:04:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.691	3.689	36859656	18867712	3.973	3.572
2)	SA Decachlor...	8.743	8.695	72973489	22764198	5.168	5.493
	Target Compounds						
41)	L9 AR-1268-1	7.623	7.590	72851126	31006348	51.913	53.625
42)	L9 AR-1268-2	7.688	7.654	63525178	28112917	49.643	52.781
43)	L9 AR-1268-3	7.896	7.861	50172519	22018734	48.306	53.325
44)	L9 AR-1268-4	8.184	8.146	23615947	8301200	52.780	51.234
45)	L9 AR-1268-5	8.481	8.438	151.8E6	51592382	47.022	52.264


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031925\
Data File : PO109999.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2025 22:15
Operator : YP/AJ
Sample : AR1268ICC050
Misc :
ALS Vial : 30 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
AR1268ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 02:07:11 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:04:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.691 min
Delta R.T.: 0.000 min
Response: 36859656
Conc: 3.97 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC050




#1 Tetrachloro-m-xylene

R.T.: 3.689 min
Delta R.T.: 0.000 min
Response: 18867712
Conc: 3.57 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.743 min
Delta R.T.: 0.000 min
Response: 72973489
Conc: 5.17 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.695 min
Delta R.T.: 0.001 min
Response: 22764198
Conc: 5.49 ng/ml



#41 AR-1268-1


R.T.: 7.623 min
Delta R.T.: 0.000 min
Response: 72851126
Conc: 51.91 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC050




#41 AR-1268-1

R.T.: 7.590 min
Delta R.T.: 0.000 min
Response: 31006348
Conc: 53.63 ng/ml




#42 AR-1268-2

R.T.: 7.688 min
Delta R.T.: 0.000 min
Response: 63525178
Conc: 49.64 ng/ml



#42 AR-1268-2


R.T.: 7.654 min
Delta R.T.: 0.001 min
Response: 28112917
Conc: 52.78 ng/ml



#43 AR-1268-3


R.T.: 7.896 min
Delta R.T.: 0.000 min
Response: 50172519
Conc: 48.31 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC050




#43 AR-1268-3

R.T.: 7.861 min
Delta R.T.: 0.000 min
Response: 22018734
Conc: 53.32 ng/ml




#44 AR-1268-4

R.T.: 8.184 min
Delta R.T.: 0.000 min
Response: 23615947
Conc: 52.78 ng/ml



#44 AR-1268-4


R.T.: 8.146 min
Delta R.T.: 0.000 min
Response: 8301200
Conc: 51.23 ng/ml



#45 AR-1268-5

R.T.: 8.481 min
Delta R.T.: 0.000 min
Response: 151815895
Conc: 47.02 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC050



#45 AR-1268-5

R.T.: 8.438 min
Delta R.T.: 0.000 min
Response: 51592382
Conc: 52.26 ng/ml

Quantitation Report (Not Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040725\
Data File : PO110286.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Apr 2025 21:27
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 01:36:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound		RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml	
	System Monitoring Compounds								
1)	SA	Tetrachlo...	3.691	3.688	486.8E6	278.7E6	53.484	53.152	
2)	SA	Decachlor...	8.740	8.691	420.5E6	118.3E6	54.690	48.766	
	Target Compounds								
3)	L1	AR-1016-1	4.783	4.769	177.8E6	93902896	528.316	512.417	
4)	L1	AR-1016-2	4.802	4.788	244.1E6	134.4E6	526.454	514.067	
5)	L1	AR-1016-3	4.858	4.964	169.7E6	73038464	529.369	515.158	
6)	L1	AR-1016-4	4.979	5.006	131.1E6	55338718	518.987	469.252	
7)	L1	AR-1016-5	5.236	5.219	150.5E6	83969338	547.465	549.111	
8)	L2	AR-1221-1	3.903	3.897	19462416	11025922	168.028	164.856	
9)	L2	AR-1221-2	3.990	3.984	28148560	15829281	325.418	320.550	
10)	L2	AR-1221-3	4.066	4.059	98013268	55109401	374.531	363.317	
11)	L3	AR-1232-1	4.066	4.059	98013268	55109401	454.302	444.096	
12)	L3	AR-1232-2	4.560	4.788	132.9E6	134.4E6	1157.840	1117.012	
13)	L3	AR-1232-3	4.802	4.964	244.1E6	73038464	1152.657	1147.393	
14)	L3	AR-1232-4	4.979	5.048	131.1E6	73350252	1154.819	1236.083	
15)	L3	AR-1232-5	5.021	5.219	95769846	83969338	1190.368	1328.253	
16)	L4	AR-1242-1	4.783	4.769	177.8E6	93902896	664.910	642.223	
17)	L4	AR-1242-2	4.802	4.788	244.1E6	134.4E6	662.975	646.191	
18)	L4	AR-1242-3	4.858	4.964	169.7E6	73038464	657.909	654.391	
19)	L4	AR-1242-4	4.979	5.048	131.1E6	73350252	655.404	632.458	
20)	L4	AR-1242-5	5.632	5.570	47210124	77845170	223.685	567.211	#
21)	L5	AR-1248-1	4.783	4.769	177.8E6	93902896	806.421	772.628	
22)	L5	AR-1248-2	5.021	5.006	95769846	55338718	317.974	328.500	
23)	L5	AR-1248-3	5.236	5.048	150.5E6	73350252	403.228	404.826	
24)	L5	AR-1248-4	5.591	5.219	148.5E6	83969338	281.330	398.985	#
25)	L5	AR-1248-5	5.632	5.611	47210124	22863074	126.586	109.986	
26)	L6	AR-1254-1	5.591	5.570	148.5E6	77845170	262.801	248.949	
27)	L6	AR-1254-2	5.740	5.717	117.1E6	57497390	243.333	211.748	
28)	L6	AR-1254-3	6.161	6.136	199.8E6	94798356	258.500	228.191	
29)	L6	AR-1254-4	6.375	6.348	32159525	13331170	66.979	54.958	
30)	L6	AR-1254-5	6.794	6.765	299.0E6	152.0E6	459.925	452.452	
31)	L7	AR-1260-1	6.277	6.251	259.3E6	123.5E6	550.162	483.373	
32)	L7	AR-1260-2	6.466	6.439	302.8E6	143.3E6	508.402	477.179	
33)	L7	AR-1260-3	6.834	6.591	233.8E6	133.6E6	481.744	462.164	
34)	L7	AR-1260-4	7.094	7.063	200.0E6	99360275	471.486	461.972	
35)	L7	AR-1260-5	7.336	7.303	484.1E6	221.7E6	462.918	444.506	
36)	L8	AR-1262-1	6.834	6.804	233.8E6	118.1E6	339.753	342.880	
37)	L8	AR-1262-2	7.336	7.303	484.1E6	221.7E6	406.167	412.630	
38)	L8	AR-1262-3	7.621	7.587	112.7E6	45025623	239.417	232.030	
39)	L8	AR-1262-4	7.684	7.649	350.3E6	141.7E6	400.497	391.534	
40)	L8	AR-1262-5	8.182	8.143	121.0E6	43223300	306.815	306.197	
41)	L9	AR-1268-1	7.621	7.587	112.7E6	45025623	80.476	76.923	

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040725\
Data File : PO110286.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Apr 2025 21:27
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 01:36:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42)	L9 AR-1268-2	7.684	7.649	350.3E6	141.7E6	273.612	263.471
43)	L9 AR-1268-3	7.893	7.857	6384140	2701969	6.213	6.500
44)	L9 AR-1268-4	8.182	8.143	121.0E6	43223300	273.806	268.395
45)	L9 AR-1268-5	8.480	8.436	34068238	11602754	10.709	11.684


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040725\
Data File : PO110286.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Apr 2025 21:27
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 01:36:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.691 min
Delta R.T.: -0.001 min
Response: 486781502
Conc: 53.48 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.688 min
Delta R.T.: -0.002 min
Response: 278731990
Conc: 53.15 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.740 min
Delta R.T.: -0.005 min
Response: 420491355
Conc: 54.69 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.691 min
Delta R.T.: -0.004 min
Response: 118311884
Conc: 48.77 ng/ml



#3 AR-1016-1


R.T.: 4.783 min
Delta R.T.: -0.002 min
Response: 177801623
Conc: 528.32 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.769 min
Delta R.T.: -0.003 min
Response: 93902896
Conc: 512.42 ng/ml




#4 AR-1016-2

R.T.: 4.802 min
Delta R.T.: -0.003 min
Response: 244109403
Conc: 526.45 ng/ml



#4 AR-1016-2


R.T.: 4.788 min
Delta R.T.: -0.003 min
Response: 134430740
Conc: 514.07 ng/ml



#5 AR-1016-3


R.T.: 4.858 min
Delta R.T.: -0.002 min
Response: 169708224
Conc: 529.37 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.964 min
Delta R.T.: -0.003 min
Response: 73038464
Conc: 515.16 ng/ml




#6 AR-1016-4

R.T.: 4.979 min
Delta R.T.: -0.002 min
Response: 131142472
Conc: 518.99 ng/ml



#6 AR-1016-4


R.T.: 5.006 min
Delta R.T.: -0.003 min
Response: 55338718
Conc: 469.25 ng/ml



#7 AR-1016-5


R.T.: 5.236 min
Delta R.T.: -0.003 min
Response: 150479994
Conc: 547.46 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.219 min
Delta R.T.: -0.003 min
Response: 83969338
Conc: 549.11 ng/ml




#8 AR-1221-1

R.T.: 3.903 min
Delta R.T.: -0.003 min
Response: 19462416
Conc: 168.03 ng/ml



#8 AR-1221-1


R.T.: 3.897 min
Delta R.T.: -0.004 min
Response: 11025922
Conc: 164.86 ng/ml



#9 AR-1221-2


R.T.: 3.990 min
Delta R.T.: -0.002 min
Response: 28148560
Conc: 325.42 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#9 AR-1221-2

R.T.: 3.984 min
Delta R.T.: -0.002 min
Response: 15829281
Conc: 320.55 ng/ml




#10 AR-1221-3

R.T.: 4.066 min
Delta R.T.: -0.003 min
Response: 98013268
Conc: 374.53 ng/ml



#10 AR-1221-3


R.T.: 4.059 min
Delta R.T.: -0.003 min
Response: 55109401
Conc: 363.32 ng/ml



#11 AR-1232-1


R.T.: 4.066 min
Delta R.T.: -0.002 min
Response: 98013268
Conc: 454.30 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#11 AR-1232-1

R.T.: 4.059 min
Delta R.T.: -0.002 min
Response: 55109401
Conc: 444.10 ng/ml




#12 AR-1232-2

R.T.: 4.560 min
Delta R.T.: -0.002 min
Response: 132903875
Conc: 1157.84 ng/ml



#12 AR-1232-2


R.T.: 4.788 min
Delta R.T.: -0.002 min
Response: 134430740
Conc: 1117.01 ng/ml



#13 AR-1232-3


R.T.: 4.802 min
Delta R.T.: -0.003 min
Response: 244109403
Conc: 1152.66 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#13 AR-1232-3

R.T.: 4.964 min
Delta R.T.: -0.002 min
Response: 73038464
Conc: 1147.39 ng/ml




#14 AR-1232-4

R.T.: 4.979 min
Delta R.T.: -0.003 min
Response: 131142472
Conc: 1154.82 ng/ml



#14 AR-1232-4


R.T.: 5.048 min
Delta R.T.: -0.002 min
Response: 73350252
Conc: 1236.08 ng/ml



#15 AR-1232-5


R.T.: 5.021 min
Delta R.T.: -0.002 min
Response: 95769846
Conc: 1190.37 ng/m

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#15 AR-1232-5

R.T.: 5.219 min
Delta R.T.: -0.002 min
Response: 83969338
Conc: 1328.25 ng/ml




#16 AR-1242-1

R.T.: 4.783 min
Delta R.T.: -0.002 min
Response: 177801623
Conc: 664.91 ng/ml



#16 AR-1242-1


R.T.: 4.769 min
Delta R.T.: -0.003 min
Response: 93902896
Conc: 642.22 ng/ml



#17 AR-1242-2


R.T.: 4.802 min
Delta R.T.: -0.003 min
Response: 244109403
Conc: 662.97 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#17 AR-1242-2

R.T.: 4.788 min
Delta R.T.: -0.003 min
Response: 134430740
Conc: 646.19 ng/ml




#18 AR-1242-3

R.T.: 4.858 min
Delta R.T.: -0.003 min
Response: 169708224
Conc: 657.91 ng/ml



#18 AR-1242-3


R.T.: 4.964 min
Delta R.T.: -0.003 min
Response: 73038464
Conc: 654.39 ng/ml



#19 AR-1242-4


R.T.: 4.979 min
Delta R.T.: -0.003 min
Response: 131142472
Conc: 655.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#19 AR-1242-4

R.T.: 5.048 min
Delta R.T.: -0.003 min
Response: 73350252
Conc: 632.46 ng/ml




#20 AR-1242-5

R.T.: 5.632 min
Delta R.T.: -0.003 min
Response: 47210124
Conc: 223.69 ng/ml



#20 AR-1242-5


R.T.: 5.570 min
Delta R.T.: -0.003 min
Response: 77845170
Conc: 567.21 ng/ml



#21 AR-1248-1


R.T.: 4.783 min
Delta R.T.: -0.002 min
Response: 177801623
Conc: 806.42 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#21 AR-1248-1

R.T.: 4.769 min
Delta R.T.: -0.002 min
Response: 93902896
Conc: 772.63 ng/ml




#22 AR-1248-2

R.T.: 5.021 min
Delta R.T.: -0.002 min
Response: 95769846
Conc: 317.97 ng/ml



#22 AR-1248-2


R.T.: 5.006 min
Delta R.T.: -0.002 min
Response: 55338718
Conc: 328.50 ng/ml



#23 AR-1248-3


R.T.: 5.236 min
Delta R.T.: -0.002 min
Response: 150479994
Conc: 403.23 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#23 AR-1248-3

R.T.: 5.048 min
Delta R.T.: -0.002 min
Response: 73350252
Conc: 404.83 ng/ml




#24 AR-1248-4

R.T.: 5.591 min
Delta R.T.: -0.001 min
Response: 148547870
Conc: 281.33 ng/ml



#24 AR-1248-4


R.T.: 5.219 min
Delta R.T.: -0.002 min
Response: 83969338
Conc: 398.98 ng/ml



#25 AR-1248-5


R.T.: 5.632 min
Delta R.T.: -0.002 min
Response: 47210124
Conc: 126.59 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#25 AR-1248-5

R.T.: 5.611 min
Delta R.T.: -0.002 min
Response: 22863074
Conc: 109.99 ng/ml




#26 AR-1254-1

R.T.: 5.591 min
Delta R.T.: -0.002 min
Response: 148547870
Conc: 262.80 ng/ml



#26 AR-1254-1


R.T.: 5.570 min
Delta R.T.: -0.003 min
Response: 77845170
Conc: 248.95 ng/ml



#27 AR-1254-2


R.T.: 5.740 min
Delta R.T.: -0.003 min
Response: 117101241
Conc: 243.33 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#27 AR-1254-2

R.T.: 5.717 min
Delta R.T.: -0.003 min
Response: 57497390
Conc: 211.75 ng/ml




#28 AR-1254-3

R.T.: 6.161 min
Delta R.T.: 0.013 min
Response: 199842623
Conc: 258.50 ng/ml



#28 AR-1254-3


R.T.: 6.136 min
Delta R.T.: 0.013 min
Response: 94798356
Conc: 228.19 ng/ml



#29 AR-1254-4


R.T.: 6.375 min
Delta R.T.: -0.003 min
Response: 32159525
Conc: 66.98 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#29 AR-1254-4

R.T.: 6.348 min
Delta R.T.: -0.003 min
Response: 13331170
Conc: 54.96 ng/ml




#30 AR-1254-5

R.T.: 6.794 min
Delta R.T.: -0.003 min
Response: 299018131
Conc: 459.92 ng/ml



#30 AR-1254-5


R.T.: 6.765 min
Delta R.T.: -0.003 min
Response: 151977984
Conc: 452.45 ng/ml



#31 AR-1260-1


R.T.: 6.277 min
Delta R.T.: -0.003 min
Response: 259319565
Conc: 550.16 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#31 AR-1260-1

R.T.: 6.251 min
Delta R.T.: -0.004 min
Response: 123473695
Conc: 483.37 ng/ml




#32 AR-1260-2

R.T.: 6.466 min
Delta R.T.: -0.003 min
Response: 302763163
Conc: 508.40 ng/ml



#32 AR-1260-2


R.T.: 6.439 min
Delta R.T.: -0.004 min
Response: 143289838
Conc: 477.18 ng/ml



#33 AR-1260-3


R.T.: 6.834 min
Delta R.T.: -0.003 min
Response: 233791087
Conc: 481.74 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#33 AR-1260-3

R.T.: 6.591 min
Delta R.T.: -0.004 min
Response: 133557256
Conc: 462.16 ng/ml




#34 AR-1260-4

R.T.: 7.094 min
Delta R.T.: -0.004 min
Response: 199989246
Conc: 471.49 ng/ml



#34 AR-1260-4


R.T.: 7.063 min
Delta R.T.: -0.004 min
Response: 99360275
Conc: 461.97 ng/ml



#35 AR-1260-5


R.T.: 7.336 min
Delta R.T.: -0.004 min
Response: 484140405
Conc: 462.92 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#35 AR-1260-5

R.T.: 7.303 min
Delta R.T.: -0.004 min
Response: 221673295
Conc: 444.51 ng/ml




#36 AR-1262-1

R.T.: 6.834 min
Delta R.T.: -0.002 min
Response: 233791087
Conc: 339.75 ng/ml



#36 AR-1262-1


R.T.: 6.804 min
Delta R.T.: -0.003 min
Response: 118057901
Conc: 342.88 ng/ml



#37 AR-1262-2


R.T.: 7.336 min
Delta R.T.: -0.002 min
Response: 484140405
Conc: 406.17 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#37 AR-1262-2

R.T.: 7.303 min
Delta R.T.: -0.003 min
Response: 221673295
Conc: 412.63 ng/ml




#38 AR-1262-3

R.T.: 7.621 min
Delta R.T.: -0.002 min
Response: 112676982
Conc: 239.42 ng/ml



#38 AR-1262-3


R.T.: 7.587 min
Delta R.T.: -0.003 min
Response: 45025623
Conc: 232.03 ng/ml



#39 AR-1262-4


R.T.: 7.684 min
Delta R.T.: -0.003 min
Response: 350265107
Conc: 400.50 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#39 AR-1262-4

R.T.: 7.649 min
Delta R.T.: -0.004 min
Response: 141651763
Conc: 391.53 ng/ml




#40 AR-1262-5

R.T.: 8.182 min
Delta R.T.: -0.002 min
Response: 121020918
Conc: 306.82 ng/ml



#40 AR-1262-5


R.T.: 8.143 min
Delta R.T.: -0.003 min
Response: 43223300
Conc: 306.20 ng/ml



#41 AR-1268-1


R.T.: 7.621 min
Delta R.T.: 0.000 min
Response: 112676982
Conc: 80.48 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#41 AR-1268-1

R.T.: 7.587 min
Delta R.T.: -0.002 min
Response: 45025623
Conc: 76.92 ng/ml




#42 AR-1268-2

R.T.: 7.684 min
Delta R.T.: -0.004 min
Response: 350265107
Conc: 273.61 ng/ml



#42 AR-1268-2


R.T.: 7.649 min
Delta R.T.: -0.004 min
Response: 141651763
Conc: 263.47 ng/ml



#43 AR-1268-3


R.T.: 7.893 min
Delta R.T.: -0.003 min
Response: 6384140
Conc: 6.21 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#43 AR-1268-3

R.T.: 7.857 min
Delta R.T.: -0.004 min
Response: 2701969
Conc: 6.50 ng/ml




#44 AR-1268-4

R.T.: 8.182 min
Delta R.T.: -0.003 min
Response: 121020918
Conc: 273.81 ng/ml



#44 AR-1268-4


R.T.: 8.143 min
Delta R.T.: -0.002 min
Response: 43223300
Conc: 268.40 ng/ml



#45 AR-1268-5

R.T.: 8.480 min
Delta R.T.: -0.002 min
Response: 34068238
Conc: 10.71 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500



#45 AR-1268-5

R.T.: 8.436 min
Delta R.T.: -0.003 min
Response: 11602754
Conc: 11.68 ng/ml

Quantitation Report (Not Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070920.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 11:19
Operator : YP\AJ
Sample : AR1660ICC250
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 31 03:02:22 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 28 03:13:46 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml	
	System Monitoring Compounds							
1)	SA Tetrachlo...	4.515	3.816	39892911	28120307	27.294	27.126	
2)	SA Decachlor...	10.229	8.859	27877719	18839232	26.484	23.902	
	Target Compounds							
3)	L1 AR-1016-1	5.667	4.903	13581167	11323714	267.076	282.470	
4)	L1 AR-1016-2	5.689	4.922	20108471	16375760	257.624	282.999	
5)	L1 AR-1016-3	5.751	5.099	12212490	8877034	258.878	287.066	
6)	L1 AR-1016-4	5.848	5.141	9803037	7158830	266.708	293.926	
7)	L1 AR-1016-5	6.141	5.355	9152735	9323046	269.896	288.450	
8)	L2 AR-1221-1	4.717	4.026	1321625	1122384	76.158	73.630	
9)	L2 AR-1221-2	4.802	4.115	2090814	1735253	161.416	151.732	
10)	L2 AR-1221-3	4.877	4.190	7052777	6326573	169.061	185.590	
11)	L3 AR-1232-1	4.877	4.190	7052777	6326573	211.867	226.056	
12)	L3 AR-1232-2	5.402	4.922	9379837	16375760	589.544	583.574	
13)	L3 AR-1232-3	5.689	5.099	20108471	8877034	586.759	596.191	
14)	L3 AR-1232-4	5.848	5.183	9803037	8604427	624.096	671.992	
15)	L3 AR-1232-5	5.939	5.355	7957300	9323046	747.842	657.721	
16)	L4 AR-1242-1	5.667	4.903	13581167	11323714	308.823	331.143	
17)	L4 AR-1242-2	5.689	4.922	20108471	16375760	312.650	332.451	
18)	L4 AR-1242-3	5.751	5.099	12212490	8877034	297.579	333.169	
19)	L4 AR-1242-4	5.848	5.183	9803037	8604427	321.165	326.194	
20)	L4 AR-1242-5	6.579	5.708	4761571	8766943	130.417	283.186	#
21)	L5 AR-1248-1	5.667	4.903	13581167	11323714	374.266	418.158	
22)	L5 AR-1248-2	5.939	5.141	7957300	7158830	177.071	194.944	
23)	L5 AR-1248-3	6.141	5.183	9152735	8604427	180.750	222.570	
24)	L5 AR-1248-4	6.539	5.355	4049391	9323046	63.126	206.532	#
25)	L5 AR-1248-5	6.579	5.749	4761571	2416322	78.256	56.617	#
26)	L6 AR-1254-1	6.515	5.708	10069870	8766943	156.388	132.797	
27)	L6 AR-1254-2	6.731	5.856	10106718	6407649	99.977	112.222	
28)	L6 AR-1254-3	7.094	6.276	7072472	11091003	70.918	127.716	#
29)	L6 AR-1254-4	7.379	6.489	4686448	1684678	54.118	30.030	#
30)	L6 AR-1254-5	7.792	6.906	26750553	17034582	332.921	236.097	#
31)	L7 AR-1260-1	7.260	6.391	18182257	14056672	270.759	274.511	
32)	L7 AR-1260-2	7.513	6.579	26780079	17295947	278.563	271.122	
33)	L7 AR-1260-3	7.871	6.733	23122222	14684885	289.324	269.513	
34)	L7 AR-1260-4	8.095	7.205	23709378	12488782	301.122	261.128	
35)	L7 AR-1260-5	8.415	7.446	43743345	31815782	275.313	255.527	
36)	L8 AR-1262-1	8.095	6.945	23709378	15226146	223.042	177.951	
37)	L8 AR-1262-2	8.415	7.205	43743345	12488782	220.500	182.413	
38)	L8 AR-1262-3	8.737	7.728	26257171	6907710	199.308	113.757	#
39)	L8 AR-1262-4	8.806	7.792	15776574	20963428	165.001	210.885	#
40)	L8 AR-1262-5	9.472	8.293	11555188	7670514	177.591	163.649	
41)	L9 AR-1268-1	8.737	7.728	26257171	6907710	111.492	40.301	#

Quantitation Report (Not Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070920.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 11:19
Operator : YP\AJ
Sample : AR1660ICC250
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 31 03:02:22 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 28 03:13:46 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml	
42) L9	AR-1268-2	8.806	7.792	15776574	20963428	80.775	146.308	#
43) L9	AR-1268-3	0.000	8.001	0	412782	N.D.	3.352	#
44) L9	AR-1268-4	9.472	8.293	11555188	7670514	163.331	145.239	
45) L9	AR-1268-5	9.890	8.596	2482779	969431	5.214	2.679	#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Quantitation Report (Not Reviewed)


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070920.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 11:19
Operator : YP\AJ
Sample : AR1660ICC250
Misc :
ALS Vial : 6 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1660ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 31 03:02:22 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 28 03:13:46 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.515 min
Delta R.T.: 0.003 min
Response: 39892911
Conc: 27.29 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.816 min
Delta R.T.: 0.000 min
Response: 28120307
Conc: 27.13 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.229 min
Delta R.T.: 0.004 min
Response: 27877719
Conc: 26.48 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.859 min
Delta R.T.: 0.000 min
Response: 18839232
Conc: 23.90 ng/ml



#3 AR-1016-1


R.T.: 5.667 min
Delta R.T.: 0.004 min
Response: 13581167
Conc: 267.08 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250




#3 AR-1016-1

R.T.: 4.903 min
Delta R.T.: 0.000 min
Response: 11323714
Conc: 282.47 ng/ml




#4 AR-1016-2

R.T.: 5.689 min
Delta R.T.: 0.004 min
Response: 20108471
Conc: 257.62 ng/ml



#4 AR-1016-2


R.T.: 4.922 min
Delta R.T.: 0.000 min
Response: 16375760
Conc: 283.00 ng/ml



#5 AR-1016-3


R.T.: 5.751 min
Delta R.T.: 0.004 min
Response: 12212490
Conc: 258.88 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250




#5 AR-1016-3

R.T.: 5.099 min
Delta R.T.: 0.000 min
Response: 8877034
Conc: 287.07 ng/ml




#6 AR-1016-4

R.T.: 5.848 min
Delta R.T.: 0.003 min
Response: 9803037
Conc: 266.71 ng/ml



#6 AR-1016-4


R.T.: 5.141 min
Delta R.T.: 0.000 min
Response: 7158830
Conc: 293.93 ng/ml



#7 AR-1016-5


R.T.: 6.141 min
Delta R.T.: 0.003 min
Response: 9152735
Conc: 269.90 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250




#7 AR-1016-5

R.T.: 5.355 min
Delta R.T.: 0.000 min
Response: 9323046
Conc: 288.45 ng/ml




#8 AR-1221-1

R.T.: 4.717 min
Delta R.T.: 0.000 min
Response: 1321625
Conc: 76.16 ng/ml



#8 AR-1221-1


R.T.: 4.026 min
Delta R.T.: 0.000 min
Response: 1122384
Conc: 73.63 ng/ml



#9 AR-1221-2


R.T.: 4.802 min
Delta R.T.: 0.000 min
Response: 2090814
Conc: 161.42 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250




#9 AR-1221-2

R.T.: 4.115 min
Delta R.T.: 0.000 min
Response: 1735253
Conc: 151.73 ng/ml




#10 AR-1221-3

R.T.: 4.877 min
Delta R.T.: 0.000 min
Response: 7052777
Conc: 169.06 ng/ml



#10 AR-1221-3


R.T.: 4.190 min
Delta R.T.: 0.000 min
Response: 6326573
Conc: 185.59 ng/ml



#11 AR-1232-1


R.T.: 4.877 min
Delta R.T.: -0.005 min
Response: 7052777
Conc: 211.87 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250




#11 AR-1232-1

R.T.: 4.190 min
Delta R.T.: -0.001 min
Response: 6326573
Conc: 226.06 ng/ml




#12 AR-1232-2

R.T.: 5.402 min
Delta R.T.: -0.005 min
Response: 9379837
Conc: 589.54 ng/ml



#12 AR-1232-2


R.T.: 4.922 min
Delta R.T.: -0.001 min
Response: 16375760
Conc: 583.57 ng/ml



#13 AR-1232-3


R.T.: 5.689 min
Delta R.T.: -0.005 min
Response: 20108471
Conc: 586.76 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250




#13 AR-1232-3

R.T.: 5.099 min
Delta R.T.: -0.001 min
Response: 8877034
Conc: 596.19 ng/ml




#14 AR-1232-4

R.T.: 5.848 min
Delta R.T.: -0.006 min
Response: 9803037
Conc: 624.10 ng/ml



#14 AR-1232-4


R.T.: 5.183 min
Delta R.T.: -0.001 min
Response: 8604427
Conc: 671.99 ng/ml



#15 AR-1232-5


R.T.: 5.939 min
Delta R.T.: -0.005 min
Response: 7957300
Conc: 747.84 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250




#15 AR-1232-5

R.T.: 5.355 min
Delta R.T.: -0.001 min
Response: 9323046
Conc: 657.72 ng/ml




#16 AR-1242-1

R.T.: 5.667 min
Delta R.T.: 0.003 min
Response: 13581167
Conc: 308.82 ng/ml



#16 AR-1242-1


R.T.: 4.903 min
Delta R.T.: 0.000 min
Response: 11323714
Conc: 331.14 ng/ml



#17 AR-1242-2


R.T.: 5.689 min
Delta R.T.: 0.003 min
Response: 20108471
Conc: 312.65 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250




#17 AR-1242-2

R.T.: 4.922 min
Delta R.T.: 0.001 min
Response: 16375760
Conc: 332.45 ng/ml




#18 AR-1242-3

R.T.: 5.751 min
Delta R.T.: 0.003 min
Response: 12212490
Conc: 297.58 ng/ml



#18 AR-1242-3


R.T.: 5.099 min
Delta R.T.: 0.000 min
Response: 8877034
Conc: 333.17 ng/ml



#19 AR-1242-4


R.T.: 5.848 min
Delta R.T.: 0.003 min
Response: 9803037
Conc: 321.17 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250




#19 AR-1242-4

R.T.: 5.183 min
Delta R.T.: 0.001 min
Response: 8604427
Conc: 326.19 ng/ml




#20 AR-1242-5

R.T.: 6.579 min
Delta R.T.: 0.004 min
Response: 4761571
Conc: 130.42 ng/ml



#20 AR-1242-5


R.T.: 5.708 min
Delta R.T.: 0.001 min
Response: 8766943
Conc: 283.19 ng/ml



#21 AR-1248-1


R.T.: 5.667 min
Delta R.T.: 0.002 min
Response: 13581167
Conc: 374.27 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250




#21 AR-1248-1

R.T.: 4.903 min
Delta R.T.: 0.000 min
Response: 11323714
Conc: 418.16 ng/ml




#22 AR-1248-2

R.T.: 5.939 min
Delta R.T.: 0.002 min
Response: 7957300
Conc: 177.07 ng/ml



#22 AR-1248-2


R.T.: 5.141 min
Delta R.T.: 0.001 min
Response: 7158830
Conc: 194.94 ng/ml



#23 AR-1248-3


R.T.: 6.141 min
Delta R.T.: 0.000 min
Response: 9152735
Conc: 180.75 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250




#23 AR-1248-3

R.T.: 5.183 min
Delta R.T.: 0.000 min
Response: 8604427
Conc: 222.57 ng/ml




#24 AR-1248-4

R.T.: 6.539 min
Delta R.T.: 0.000 min
Response: 4049391
Conc: 63.13 ng/ml



#24 AR-1248-4


R.T.: 5.355 min
Delta R.T.: 0.000 min
Response: 9323046
Conc: 206.53 ng/ml



#25 AR-1248-5


R.T.: 6.579 min
Delta R.T.: 0.002 min
Response: 4761571
Conc: 78.26 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250




#25 AR-1248-5

R.T.: 5.749 min
Delta R.T.: 0.000 min
Response: 2416322
Conc: 56.62 ng/ml




#26 AR-1254-1

R.T.: 6.515 min
Delta R.T.: 0.000 min
Response: 10069870
Conc: 156.39 ng/ml



#26 AR-1254-1


R.T.: 5.708 min
Delta R.T.: 0.000 min
Response: 8766943
Conc: 132.80 ng/ml



#27 AR-1254-2


R.T.: 6.731 min
Delta R.T.: 0.000 min
Response: 10106718
Conc: 99.98 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250




#27 AR-1254-2

R.T.: 5.856 min
Delta R.T.: 0.000 min
Response: 6407649
Conc: 112.22 ng/ml




#28 AR-1254-3

R.T.: 7.094 min
Delta R.T.: 0.000 min
Response: 7072472
Conc: 70.92 ng/ml



#28 AR-1254-3


R.T.: 6.276 min
Delta R.T.: 0.016 min
Response: 11091003
Conc: 127.72 ng/ml



#29 AR-1254-4


R.T.: 7.379 min
Delta R.T.: 0.002 min
Response: 4686448
Conc: 54.12 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250




#29 AR-1254-4

R.T.: 6.489 min
Delta R.T.: 0.000 min
Response: 1684678
Conc: 30.03 ng/ml




#30 AR-1254-5

R.T.: 7.792 min
Delta R.T.: -0.001 min
Response: 26750553
Conc: 332.92 ng/ml



#30 AR-1254-5


R.T.: 6.906 min
Delta R.T.: 0.000 min
Response: 17034582
Conc: 236.10 ng/ml



#31 AR-1260-1


R.T.: 7.260 min
Delta R.T.: 0.004 min
Response: 18182257
Conc: 270.76 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250




#31 AR-1260-1

R.T.: 6.391 min
Delta R.T.: -0.001 min
Response: 14056672
Conc: 274.51 ng/ml




#32 AR-1260-2

R.T.: 7.513 min
Delta R.T.: 0.004 min
Response: 26780079
Conc: 278.56 ng/ml



#32 AR-1260-2


R.T.: 6.579 min
Delta R.T.: 0.000 min
Response: 17295947
Conc: 271.12 ng/ml



#33 AR-1260-3


R.T.: 7.871 min
Delta R.T.: 0.003 min
Response: 23122222
Conc: 289.32 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250




#33 AR-1260-3

R.T.: 6.733 min
Delta R.T.: 0.000 min
Response: 14684885
Conc: 269.51 ng/ml




#34 AR-1260-4

R.T.: 8.095 min
Delta R.T.: 0.004 min
Response: 23709378
Conc: 301.12 ng/ml



#34 AR-1260-4


R.T.: 7.205 min
Delta R.T.: 0.000 min
Response: 12488782
Conc: 261.13 ng/ml



#35 AR-1260-5


R.T.: 8.415 min
Delta R.T.: 0.004 min
Response: 43743345
Conc: 275.31 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250




#35 AR-1260-5

R.T.: 7.446 min
Delta R.T.: 0.000 min
Response: 31815782
Conc: 255.53 ng/ml




#36 AR-1262-1

R.T.: 8.095 min
Delta R.T.: 0.003 min
Response: 23709378
Conc: 223.04 ng/ml



#36 AR-1262-1


R.T.: 6.945 min
Delta R.T.: 0.000 min
Response: 15226146
Conc: 177.95 ng/ml



#37 AR-1262-2


R.T.: 8.415 min
Delta R.T.: 0.003 min
Response: 43743345
Conc: 220.50 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250




#37 AR-1262-2

R.T.: 7.205 min
Delta R.T.: 0.000 min
Response: 12488782
Conc: 182.41 ng/ml




#38 AR-1262-3

R.T.: 8.737 min
Delta R.T.: 0.008 min
Response: 26257171
Conc: 199.31 ng/ml



#38 AR-1262-3


R.T.: 7.728 min
Delta R.T.: 0.000 min
Response: 6907710
Conc: 113.76 ng/ml



#39 AR-1262-4


R.T.: 8.806 min
Delta R.T.: -0.011 min
Response: 15776574
Conc: 165.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250




#39 AR-1262-4

R.T.: 7.792 min
Delta R.T.: -0.001 min
Response: 20963428
Conc: 210.89 ng/ml




#40 AR-1262-5

R.T.: 9.472 min
Delta R.T.: 0.003 min
Response: 11555188
Conc: 177.59 ng/ml



#40 AR-1262-5


R.T.: 8.293 min
Delta R.T.: 0.000 min
Response: 7670514
Conc: 163.65 ng/ml



#41 AR-1268-1


R.T.: 8.737 min
Delta R.T.: 0.008 min
Response: 26257171
Conc: 111.49 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250




#41 AR-1268-1

R.T.: 7.728 min
Delta R.T.: 0.000 min
Response: 6907710
Conc: 40.30 ng/ml




#42 AR-1268-2

R.T.: 8.806 min
Delta R.T.: -0.016 min
Response: 15776574
Conc: 80.77 ng/ml



#42 AR-1268-2


R.T.: 7.792 min
Delta R.T.: -0.002 min
Response: 20963428
Conc: 146.31 ng/ml



#43 AR-1268-3


R.T.: 0.000 min
Exp R.T. : 9.053 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250




#43 AR-1268-3

R.T.: 8.001 min
Delta R.T.: 0.000 min
Response: 412782
Conc: 3.35 ng/ml




#44 AR-1268-4

R.T.: 9.472 min
Delta R.T.: 0.000 min
Response: 11555188
Conc: 163.33 ng/ml



#44 AR-1268-4


R.T.: 8.293 min
Delta R.T.: 0.000 min
Response: 7670514
Conc: 145.24 ng/ml



#45 AR-1268-5

R.T.: 9.890 min
Delta R.T.: 0.002 min
Response: 2482779
Conc: 5.21 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250



#45 AR-1268-5

R.T.: 8.596 min
Delta R.T.: 0.000 min
Response: 969431
Conc: 2.68 ng/ml

Quantitation Report (Not Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070921.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 11:35
Operator : YP\AJ
Sample : AR1660ICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 31 03:04:09 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 28 03:13:46 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml	
	System Monitoring Compounds							
1)	SA Tetrachlo...	4.513	3.816	6533335	5359624	4.470	5.170	
2)	SA Decachlor...	10.230	8.858	4733739	3413055	4.497	4.330	
	Target Compounds							
3)	L1 AR-1016-1	5.666	4.903	2617394	2265971	51.472	56.525	
4)	L1 AR-1016-2	5.688	4.922	4198554	3321854	53.791	57.407	
5)	L1 AR-1016-3	5.751	5.099	2631097	1626306	55.774	52.592	
6)	L1 AR-1016-4	5.848	5.140	1614372	1232754	43.922	50.614	
7)	L1 AR-1016-5	6.140	5.356	1323106	1866333	39.016	57.743	#
8)	L2 AR-1221-1	4.717	4.024	421830	114713	24.308	7.525	#
9)	L2 AR-1221-2	4.801	4.115	218742	388578	16.887	33.978	#
10)	L2 AR-1221-3	4.877	4.189	1349629	1162778	32.352	34.110	
11)	L3 AR-1232-1	4.877	4.189	1349629	1162778	40.543	41.548	
12)	L3 AR-1232-2	5.401	4.922	1794886	3321854	112.813	118.379	
13)	L3 AR-1232-3	5.688	5.099	4198554	1626306	122.513	109.224	
14)	L3 AR-1232-4	5.848	5.183	1614372	1568895	102.777	122.528	
15)	L3 AR-1232-5	5.935	5.356	1663453	1866333	156.334	131.666	
16)	L4 AR-1242-1	5.666	4.903	2617394	2265971	59.517	66.265	
17)	L4 AR-1242-2	5.688	4.922	4198554	3321854	65.280	67.438	
18)	L4 AR-1242-3	5.751	5.099	2631097	1626306	64.111	61.038	
19)	L4 AR-1242-4	5.848	5.183	1614372	1568895	52.890	59.477	
20)	L4 AR-1242-5	6.582	5.708	1738801	1466013	47.625	47.354	
21)	L5 AR-1248-1	5.666	4.903	2617394	2265971	72.129	83.677	
22)	L5 AR-1248-2	5.935	5.140	1663453	1232754	37.016	33.569	
23)	L5 AR-1248-3	6.140	5.183	1323106	1568895	26.129	40.583	#
24)	L5 AR-1248-4	6.539	5.356	1621071	1866333	25.271	41.345	#
25)	L5 AR-1248-5	6.582	5.748	1738801	371096	28.577	8.695	#
26)	L6 AR-1254-1	6.515	5.708	3489366	1466013	54.191	22.206	#
27)	L6 AR-1254-2	6.731	5.857	4772204	1146779	47.207	20.084	#
28)	L6 AR-1254-3	7.095	6.275	7925059	3163207	79.467	36.425	#
29)	L6 AR-1254-4	7.380	6.490	1640237	349900	18.941	6.237	#
30)	L6 AR-1254-5	7.791	6.906	5692306	2994420	70.843	41.502	#
31)	L7 AR-1260-1	7.258	6.391	4772844	2989801	71.074	58.387	
32)	L7 AR-1260-2	7.512	6.579	5123964	3357287	53.299	52.627	
33)	L7 AR-1260-3	7.871	6.732	5463057	2827385	68.358	51.891	
34)	L7 AR-1260-4	8.095	7.205	9380427	2746307	119.136	57.423	#
35)	L7 AR-1260-5	8.414	7.445	11051960	7013440	69.559	56.328	
36)	L8 AR-1262-1	8.095	6.945	9380427	2613596	88.245	30.546	#
37)	L8 AR-1262-2	8.414	7.205	11051960	2746307	55.710	40.113	#
38)	L8 AR-1262-3	8.737	7.729	4423235	1556480	33.575	25.632	
39)	L8 AR-1262-4	8.806	7.792	1951193	3771123	20.407	37.936	#
40)	L8 AR-1262-5	9.470	8.292	1729860	1260427	26.586	26.891	
41)	L9 AR-1268-1	8.737	7.729	4423235	1556480	18.782	9.081	#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070921.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 11:35
Operator : YP\AJ
Sample : AR1660ICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 31 03:04:09 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 28 03:13:46 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.806	7.792	1951193	3771123	9.990	26.319 #
43) L9	AR-1268-3	0.000	8.026	0	152620	N.D.	1.239 #
44) L9	AR-1268-4	9.470	8.292	1729860	1260427	24.451	23.866

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Quantitation Report (Not Reviewed)


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070921.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 11:35
Operator : YP\AJ
Sample : AR1660ICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1660ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 31 03:04:09 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 28 03:13:46 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.513 min
Delta R.T.: 0.001 min
Response: 6533335
Conc: 4.47 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050




#1 Tetrachloro-m-xylene

R.T.: 3.816 min
Delta R.T.: 0.000 min
Response: 5359624
Conc: 5.17 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.230 min
Delta R.T.: 0.005 min
Response: 4733739
Conc: 4.50 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.858 min
Delta R.T.: 0.000 min
Response: 3413055
Conc: 4.33 ng/ml



#3 AR-1016-1


R.T.: 5.666 min
Delta R.T.: 0.002 min
Response: 2617394
Conc: 51.47 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050




#3 AR-1016-1

R.T.: 4.903 min
Delta R.T.: 0.000 min
Response: 2265971
Conc: 56.52 ng/ml




#4 AR-1016-2

R.T.: 5.688 min
Delta R.T.: 0.002 min
Response: 4198554
Conc: 53.79 ng/ml



#4 AR-1016-2


R.T.: 4.922 min
Delta R.T.: 0.000 min
Response: 3321854
Conc: 57.41 ng/ml



#5 AR-1016-3


R.T.: 5.751 min
Delta R.T.: 0.003 min
Response: 2631097
Conc: 55.77 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050




#5 AR-1016-3

R.T.: 5.099 min
Delta R.T.: 0.000 min
Response: 1626306
Conc: 52.59 ng/ml




#6 AR-1016-4

R.T.: 5.848 min
Delta R.T.: 0.003 min
Response: 1614372
Conc: 43.92 ng/ml



#6 AR-1016-4


R.T.: 5.140 min
Delta R.T.: 0.000 min
Response: 1232754
Conc: 50.61 ng/ml



#7 AR-1016-5


R.T.: 6.140 min
Delta R.T.: 0.002 min
Response: 1323106
Conc: 39.02 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050




#7 AR-1016-5

R.T.: 5.356 min
Delta R.T.: 0.000 min
Response: 1866333
Conc: 57.74 ng/ml




#8 AR-1221-1

R.T.: 4.717 min
Delta R.T.: 0.000 min
Response: 421830
Conc: 24.31 ng/ml



#8 AR-1221-1


R.T.: 4.024 min
Delta R.T.: -0.004 min
Response: 114713
Conc: 7.53 ng/ml



#9 AR-1221-2


R.T.: 4.801 min
Delta R.T.: -0.001 min
Response: 218742
Conc: 16.89 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050




#9 AR-1221-2

R.T.: 4.115 min
Delta R.T.: 0.000 min
Response: 388578
Conc: 33.98 ng/ml




#10 AR-1221-3

R.T.: 4.877 min
Delta R.T.: 0.000 min
Response: 1349629
Conc: 32.35 ng/ml



#10 AR-1221-3


R.T.: 4.189 min
Delta R.T.: 0.000 min
Response: 1162778
Conc: 34.11 ng/ml



#11 AR-1232-1


R.T.: 4.877 min
Delta R.T.: -0.005 min
Response: 1349629
Conc: 40.54 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050




#11 AR-1232-1

R.T.: 4.189 min
Delta R.T.: -0.001 min
Response: 1162778
Conc: 41.55 ng/ml




#12 AR-1232-2

R.T.: 5.401 min
Delta R.T.: -0.007 min
Response: 1794886
Conc: 112.81 ng/ml



#12 AR-1232-2


R.T.: 4.922 min
Delta R.T.: -0.001 min
Response: 3321854
Conc: 118.38 ng/ml



#13 AR-1232-3


R.T.: 5.688 min
Delta R.T.: -0.007 min
Response: 4198554
Conc: 122.51 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050




#13 AR-1232-3

R.T.: 5.099 min
Delta R.T.: -0.002 min
Response: 1626306
Conc: 109.22 ng/ml




#14 AR-1232-4

R.T.: 5.848 min
Delta R.T.: -0.006 min
Response: 1614372
Conc: 102.78 ng/ml



#14 AR-1232-4


R.T.: 5.183 min
Delta R.T.: -0.002 min
Response: 1568895
Conc: 122.53 ng/ml



#15 AR-1232-5


R.T.: 5.935 min
Delta R.T.: -0.009 min
Response: 1663453
Conc: 156.33 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050




#15 AR-1232-5

R.T.: 5.356 min
Delta R.T.: 0.000 min
Response: 1866333
Conc: 131.67 ng/ml




#16 AR-1242-1

R.T.: 5.666 min
Delta R.T.: 0.002 min
Response: 2617394
Conc: 59.52 ng/ml



#16 AR-1242-1


R.T.: 4.903 min
Delta R.T.: 0.000 min
Response: 2265971
Conc: 66.26 ng/ml



#17 AR-1242-2


R.T.: 5.688 min
Delta R.T.: 0.002 min
Response: 4198554
Conc: 65.28 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050




#17 AR-1242-2

R.T.: 4.922 min
Delta R.T.: 0.000 min
Response: 3321854
Conc: 67.44 ng/ml




#18 AR-1242-3

R.T.: 5.751 min
Delta R.T.: 0.003 min
Response: 2631097
Conc: 64.11 ng/ml



#18 AR-1242-3


R.T.: 5.099 min
Delta R.T.: 0.000 min
Response: 1626306
Conc: 61.04 ng/ml



#19 AR-1242-4


R.T.: 5.848 min
Delta R.T.: 0.003 min
Response: 1614372
Conc: 52.89 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050




#19 AR-1242-4

R.T.: 5.183 min
Delta R.T.: 0.000 min
Response: 1568895
Conc: 59.48 ng/ml




#20 AR-1242-5

R.T.: 6.582 min
Delta R.T.: 0.006 min
Response: 1738801
Conc: 47.62 ng/ml



#20 AR-1242-5


R.T.: 5.708 min
Delta R.T.: 0.001 min
Response: 1466013
Conc: 47.35 ng/ml



#21 AR-1248-1


R.T.: 5.666 min
Delta R.T.: 0.000 min
Response: 2617394
Conc: 72.13 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050




#21 AR-1248-1

R.T.: 4.903 min
Delta R.T.: 0.000 min
Response: 2265971
Conc: 83.68 ng/ml




#22 AR-1248-2

R.T.: 5.935 min
Delta R.T.: -0.002 min
Response: 1663453
Conc: 37.02 ng/ml



#22 AR-1248-2


R.T.: 5.140 min
Delta R.T.: 0.000 min
Response: 1232754
Conc: 33.57 ng/ml



#23 AR-1248-3


R.T.: 6.140 min
Delta R.T.: 0.000 min
Response: 1323106
Conc: 26.13 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050




#23 AR-1248-3

R.T.: 5.183 min
Delta R.T.: 0.000 min
Response: 1568895
Conc: 40.58 ng/ml




#24 AR-1248-4

R.T.: 6.539 min
Delta R.T.: 0.000 min
Response: 1621071
Conc: 25.27 ng/ml



#24 AR-1248-4


R.T.: 5.356 min
Delta R.T.: 0.001 min
Response: 1866333
Conc: 41.34 ng/ml



#25 AR-1248-5


R.T.: 6.582 min
Delta R.T.: 0.004 min
Response: 1738801
Conc: 28.58 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050




#25 AR-1248-5

R.T.: 5.748 min
Delta R.T.: 0.000 min
Response: 371096
Conc: 8.70 ng/ml




#26 AR-1254-1

R.T.: 6.515 min
Delta R.T.: 0.000 min
Response: 3489366
Conc: 54.19 ng/ml



#26 AR-1254-1


R.T.: 5.708 min
Delta R.T.: 0.000 min
Response: 1466013
Conc: 22.21 ng/ml



#27 AR-1254-2


R.T.: 6.731 min
Delta R.T.: 0.000 min
Response: 4772204
Conc: 47.21 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050




#27 AR-1254-2

R.T.: 5.857 min
Delta R.T.: 0.000 min
Response: 1146779
Conc: 20.08 ng/ml




#28 AR-1254-3

R.T.: 7.095 min
Delta R.T.: 0.000 min
Response: 7925059
Conc: 79.47 ng/ml



#28 AR-1254-3


R.T.: 6.275 min
Delta R.T.: 0.015 min
Response: 3163207
Conc: 36.43 ng/ml



#29 AR-1254-4


R.T.: 7.380 min
Delta R.T.: 0.003 min
Response: 1640237
Conc: 18.94 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050




#29 AR-1254-4

R.T.: 6.490 min
Delta R.T.: 0.001 min
Response: 349900
Conc: 6.24 ng/ml




#30 AR-1254-5

R.T.: 7.791 min
Delta R.T.: -0.002 min
Response: 5692306
Conc: 70.84 ng/ml



#30 AR-1254-5


R.T.: 6.906 min
Delta R.T.: -0.001 min
Response: 2994420
Conc: 41.50 ng/ml



#31 AR-1260-1


R.T.: 7.258 min
Delta R.T.: 0.003 min
Response: 4772844
Conc: 71.07 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050




#31 AR-1260-1

R.T.: 6.391 min
Delta R.T.: 0.000 min
Response: 2989801
Conc: 58.39 ng/ml




#32 AR-1260-2

R.T.: 7.512 min
Delta R.T.: 0.002 min
Response: 5123964
Conc: 53.30 ng/ml



#32 AR-1260-2


R.T.: 6.579 min
Delta R.T.: 0.000 min
Response: 3357287
Conc: 52.63 ng/ml



#33 AR-1260-3


R.T.: 7.871 min
Delta R.T.: 0.003 min
Response: 5463057
Conc: 68.36 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050




#33 AR-1260-3

R.T.: 6.732 min
Delta R.T.: -0.001 min
Response: 2827385
Conc: 51.89 ng/ml




#34 AR-1260-4

R.T.: 8.095 min
Delta R.T.: 0.004 min
Response: 9380427
Conc: 119.14 ng/ml



#34 AR-1260-4


R.T.: 7.205 min
Delta R.T.: 0.000 min
Response: 2746307
Conc: 57.42 ng/ml



#35 AR-1260-5


R.T.: 8.414 min
Delta R.T.: 0.002 min
Response: 11051960
Conc: 69.56 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050




#35 AR-1260-5

R.T.: 7.445 min
Delta R.T.: 0.000 min
Response: 7013440
Conc: 56.33 ng/ml




#36 AR-1262-1

R.T.: 8.095 min
Delta R.T.: 0.003 min
Response: 9380427
Conc: 88.24 ng/ml



#36 AR-1262-1


R.T.: 6.945 min
Delta R.T.: 0.000 min
Response: 2613596
Conc: 30.55 ng/ml



#37 AR-1262-2


R.T.: 8.414 min
Delta R.T.: 0.001 min
Response: 11051960
Conc: 55.71 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050




#37 AR-1262-2

R.T.: 7.205 min
Delta R.T.: 0.000 min
Response: 2746307
Conc: 40.11 ng/ml




#38 AR-1262-3

R.T.: 8.737 min
Delta R.T.: 0.008 min
Response: 4423235
Conc: 33.58 ng/ml



#38 AR-1262-3


R.T.: 7.729 min
Delta R.T.: 0.000 min
Response: 1556480
Conc: 25.63 ng/ml



#39 AR-1262-4


R.T.: 8.806 min
Delta R.T.: -0.011 min
Response: 1951193
Conc: 20.41 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050




#39 AR-1262-4

R.T.: 7.792 min
Delta R.T.: 0.000 min
Response: 3771123
Conc: 37.94 ng/ml




#40 AR-1262-5

R.T.: 9.470 min
Delta R.T.: 0.001 min
Response: 1729860
Conc: 26.59 ng/ml



#40 AR-1262-5


R.T.: 8.292 min
Delta R.T.: -0.002 min
Response: 1260427
Conc: 26.89 ng/ml



#41 AR-1268-1


R.T.: 8.737 min
Delta R.T.: 0.008 min
Response: 4423235
Conc: 18.78 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050




#41 AR-1268-1

R.T.: 7.729 min
Delta R.T.: 0.001 min
Response: 1556480
Conc: 9.08 ng/ml




#42 AR-1268-2

R.T.: 8.806 min
Delta R.T.: -0.016 min
Response: 1951193
Conc: 9.99 ng/ml



#42 AR-1268-2


R.T.: 7.792 min
Delta R.T.: -0.001 min
Response: 3771123
Conc: 26.32 ng/ml



#43 AR-1268-3


R.T.: 0.000 min
Exp R.T. : 9.053 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050




#43 AR-1268-3

R.T.: 8.026 min
Delta R.T.: 0.026 min
Response: 152620
Conc: 1.24 ng/ml



#44 AR-1268-4

R.T.: 9.470 min
Delta R.T.: -0.002 min
Response: 1729860
Conc: 24.45 ng/ml



#44 AR-1268-4

R.T.: 8.292 min
Delta R.T.: -0.001 min
Response: 1260427
Conc: 23.87 ng/ml

Quantitation Report (Not Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070923.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 12:11
Operator : YP\AJ
Sample : AR1232ICC500
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1232ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 12:39:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 12:38:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.520	3.817	74435949	51438566	50.000	50.000
2)	SA Decachlor...	10.236	8.860	99828492	40517792	110.430	50.000 #
	Target Compounds						
11)	L3 AR-1232-1	4.883	4.191	16644355	13993352	500.000	500.000
12)	L3 AR-1232-2	5.408	4.923	7955169	14030567	500.000	500.000
13)	L3 AR-1232-3	5.694	5.100	17135200	7444793	500.000	500.000
14)	L3 AR-1232-4	5.854	5.185	7853790	6402183	500.000	500.000
15)	L3 AR-1232-5	5.943	5.357	5320177	7087391	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Quantitation Report (Not Reviewed)


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070923.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 12:11
Operator : YP\AJ
Sample : AR1232ICC500
Misc :
ALS Vial : 9 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1232ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 12:39:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 12:38:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.520 min
Delta R.T.: 0.000 min
Response: 74435949
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1232ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.817 min
Delta R.T.: 0.000 min
Response: 51438566
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.236 min
Delta R.T.: 0.001 min
Response: 99828492
Conc: 110.43 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.860 min
Delta R.T.: 0.000 min
Response: 40517792
Conc: 50.00 ng/ml



#11 AR-1232-1


R.T.: 4.883 min
Delta R.T.: 0.000 min
Response: 16644355
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1232ICC500




#11 AR-1232-1

R.T.: 4.191 min
Delta R.T.: 0.000 min
Response: 13993352
Conc: 500.00 ng/ml




#12 AR-1232-2

R.T.: 5.408 min
Delta R.T.: 0.000 min
Response: 7955169
Conc: 500.00 ng/ml



#12 AR-1232-2


R.T.: 4.923 min
Delta R.T.: 0.000 min
Response: 14030567
Conc: 500.00 ng/ml



#13 AR-1232-3


R.T.: 5.694 min
Delta R.T.: 0.000 min
Response: 17135200
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1232ICC500




#13 AR-1232-3

R.T.: 5.100 min
Delta R.T.: 0.000 min
Response: 7444793
Conc: 500.00 ng/ml




#14 AR-1232-4

R.T.: 5.854 min
Delta R.T.: 0.000 min
Response: 7853790
Conc: 500.00 ng/ml



#14 AR-1232-4


R.T.: 5.185 min
Delta R.T.: 0.000 min
Response: 6402183
Conc: 500.00 ng/ml



#15 AR-1232-5

R.T.: 5.943 min
Delta R.T.: 0.000 min
Response: 5320177
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1232ICC500



#15 AR-1232-5

R.T.: 5.357 min
Delta R.T.: 0.000 min
Response: 7087391
Conc: 500.00 ng/ml

Quantitation Report (Not Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070928.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 13:33
Operator : YP\AJ
Sample : AR1242ICC050
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 13:57:16 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 13:24:34 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.512	3.815	6633492	5532966	4.273	4.939
2)	SA Decachlor...	10.226	8.857	5282830	3387334	4.676	4.164
	Target Compounds						
16)	L4 AR-1242-1	5.664	4.902	2665542	1773341	60.984	48.354
17)	L4 AR-1242-2	5.687	4.921	3354217	3201456	49.372	61.016
18)	L4 AR-1242-3	5.750	5.098	2977988	1401755	74.143	49.661 #
19)	L4 AR-1242-4	5.844	5.181	1598185	1807177	49.578	66.685 #
20)	L4 AR-1242-5	6.576	5.707	2292388	1930099	60.410	59.828

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Quantitation Report (Not Reviewed)


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070928.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 13:33
Operator : YP\AJ
Sample : AR1242ICC050
Misc :
ALS Vial : 14 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1242ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 13:57:16 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 13:24:34 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.512 min
Delta R.T.: 0.001 min
Response: 6633492
Conc: 4.27 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC050




#1 Tetrachloro-m-xylene

R.T.: 3.815 min
Delta R.T.: 0.000 min
Response: 5532966
Conc: 4.94 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.226 min
Delta R.T.: 0.000 min
Response: 5282830
Conc: 4.68 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.857 min
Delta R.T.: -0.001 min
Response: 3387334
Conc: 4.16 ng/ml



#16 AR-1242-1


R.T.: 5.664 min
Delta R.T.: 0.000 min
Response: 2665542
Conc: 60.98 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC050




#16 AR-1242-1

R.T.: 4.902 min
Delta R.T.: 0.000 min
Response: 1773341
Conc: 48.35 ng/ml




#17 AR-1242-2

R.T.: 5.687 min
Delta R.T.: 0.001 min
Response: 3354217
Conc: 49.37 ng/ml



#17 AR-1242-2


R.T.: 4.921 min
Delta R.T.: 0.000 min
Response: 3201456
Conc: 61.02 ng/ml



#18 AR-1242-3


R.T.: 5.750 min
Delta R.T.: 0.002 min
Response: 2977988
Conc: 74.14 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC050




#18 AR-1242-3

R.T.: 5.098 min
Delta R.T.: 0.000 min
Response: 1401755
Conc: 49.66 ng/ml




#19 AR-1242-4

R.T.: 5.844 min
Delta R.T.: 0.000 min
Response: 1598185
Conc: 49.58 ng/ml



#19 AR-1242-4


R.T.: 5.181 min
Delta R.T.: 0.000 min
Response: 1807177
Conc: 66.69 ng/ml



#20 AR-1242-5

R.T.: 6.576 min
Delta R.T.: 0.000 min
Response: 2292388
Conc: 60.41 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC050



#20 AR-1242-5

R.T.: 5.707 min
Delta R.T.: 0.000 min
Response: 1930099
Conc: 59.83 ng/ml

Quantitation Report (Not Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070933.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 14:55
Operator : YP\AJ
Sample : AR1248ICC050
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 15:16:04 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 15:04:54 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.513	3.815	7324663	5848649	4.791	5.521
2)	SA Decachlor...	10.227	8.859	5549589	3661642	4.859	4.550
	Target Compounds						
21)	L5 AR-1248-1	5.665	4.902	1857345	1382826	51.489	51.338
22)	L5 AR-1248-2	5.936	5.140	2633031	1982278	58.287	55.076
23)	L5 AR-1248-3	6.139	5.182	2192198	2128069	43.549	56.472 #
24)	L5 AR-1248-4	6.538	5.354	3237397	2437650	49.085	55.103
25)	L5 AR-1248-5	6.578	5.748	3690931	2234402	59.064	52.978


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070933.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 14:55
Operator : YP\AJ
Sample : AR1248ICC050
Misc :
ALS Vial : 19 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1248ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 15:16:04 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 15:04:54 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.513 min
Delta R.T.: 0.000 min
Response: 7324663
Conc: 4.79 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC050




#1 Tetrachloro-m-xylene

R.T.: 3.815 min
Delta R.T.: 0.000 min
Response: 5848649
Conc: 5.52 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.227 min
Delta R.T.: 0.000 min
Response: 5549589
Conc: 4.86 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.859 min
Delta R.T.: 0.000 min
Response: 3661642
Conc: 4.55 ng/ml



#21 AR-1248-1


R.T.: 5.665 min
Delta R.T.: 0.000 min
Response: 1857345
Conc: 51.49 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC050




#21 AR-1248-1

R.T.: 4.902 min
Delta R.T.: 0.000 min
Response: 1382826
Conc: 51.34 ng/ml




#22 AR-1248-2

R.T.: 5.936 min
Delta R.T.: 0.000 min
Response: 2633031
Conc: 58.29 ng/ml



#22 AR-1248-2


R.T.: 5.140 min
Delta R.T.: 0.000 min
Response: 1982278
Conc: 55.08 ng/ml



#23 AR-1248-3


R.T.: 6.139 min
Delta R.T.: 0.000 min
Response: 2192198
Conc: 43.55 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC050




#23 AR-1248-3

R.T.: 5.182 min
Delta R.T.: 0.000 min
Response: 2128069
Conc: 56.47 ng/ml




#24 AR-1248-4

R.T.: 6.538 min
Delta R.T.: -0.001 min
Response: 3237397
Conc: 49.09 ng/ml



#24 AR-1248-4


R.T.: 5.354 min
Delta R.T.: 0.000 min
Response: 2437650
Conc: 55.10 ng/ml



#25 AR-1248-5

R.T.: 6.578 min
Delta R.T.: 0.000 min
Response: 3690931
Conc: 59.06 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC050



#25 AR-1248-5

R.T.: 5.748 min
Delta R.T.: 0.000 min
Response: 2234402
Conc: 52.98 ng/ml

Quantitation Report (Not Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070934.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 15:11
Operator : YP\AJ
Sample : AR1254ICC1000
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 16:19:31 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 16:07:11 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.514	3.816	150.1E6	104.3E6	95.275	94.048
2)	SA Decachlor...	10.231	8.860	109.5E6	78348381	93.754	86.018
	Target Compounds						
26)	L6 AR-1254-1	6.516	5.709	60157506	58928147	918.830	891.736
27)	L6 AR-1254-2	6.732	5.857	92838442	50145967	911.948	882.952
28)	L6 AR-1254-3	7.095	6.261	92334650	77448562	928.765	882.456
29)	L6 AR-1254-4	7.379	6.490	84238161	49998215	970.458	893.353
30)	L6 AR-1254-5	7.793	6.908	87302185	65979008	1079.742	918.596

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Quantitation Report (Not Reviewed)


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070934.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 15:11
Operator : YP\AJ
Sample : AR1254ICC1000
Misc :
ALS Vial : 20 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 16:19:31 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 16:07:11 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.514 min
Delta R.T.: 0.000 min
Response: 150128836
Conc: 95.27 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000




#1 Tetrachloro-m-xylene

R.T.: 3.816 min
Delta R.T.: 0.000 min
Response: 104259560
Conc: 94.05 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.231 min
Delta R.T.: 0.001 min
Response: 109543241
Conc: 93.75 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.860 min
Delta R.T.: 0.000 min
Response: 78348381
Conc: 86.02 ng/ml



#26 AR-1254-1


R.T.: 6.516 min
Delta R.T.: 0.000 min
Response: 60157506
Conc: 918.83 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000




#26 AR-1254-1

R.T.: 5.709 min
Delta R.T.: 0.000 min
Response: 58928147
Conc: 891.74 ng/ml




#27 AR-1254-2

R.T.: 6.732 min
Delta R.T.: 0.000 min
Response: 92838442
Conc: 911.95 ng/ml



#27 AR-1254-2


R.T.: 5.857 min
Delta R.T.: 0.000 min
Response: 50145967
Conc: 882.95 ng/ml



#28 AR-1254-3


R.T.: 7.095 min
Delta R.T.: 0.000 min
Response: 92334650
Conc: 928.77 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000




#28 AR-1254-3

R.T.: 6.261 min
Delta R.T.: 0.001 min
Response: 77448562
Conc: 882.46 ng/ml




#29 AR-1254-4

R.T.: 7.379 min
Delta R.T.: 0.002 min
Response: 84238161
Conc: 970.46 ng/ml



#29 AR-1254-4


R.T.: 6.490 min
Delta R.T.: 0.000 min
Response: 49998215
Conc: 893.35 ng/ml



#30 AR-1254-5

R.T.: 7.793 min
Delta R.T.: 0.000 min
Response: 87302185
Conc: 1079.74 ng/m

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000



#30 AR-1254-5

R.T.: 6.908 min
Delta R.T.: 0.000 min
Response: 65979008
Conc: 918.60 ng/ml

Quantitation Report (Not Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070935.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 15:27
Operator : YP\AJ
Sample : AR1254ICC750
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 16:19:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 16:07:11 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.514	3.816	116.3E6	81865229	73.835	73.847
2)	SA Decachlor...	10.228	8.859	85851924	59713645	73.477	65.559
	Target Compounds						
26)	L6 AR-1254-1	6.515	5.708	47300573	46349687	722.456	701.391
27)	L6 AR-1254-2	6.731	5.856	72785411	40339655	714.968	710.286
28)	L6 AR-1254-3	7.094	6.260	73541603	62731680	739.732	714.771
29)	L6 AR-1254-4	7.377	6.490	65835410	40961211	758.451	731.883
30)	L6 AR-1254-5	7.793	6.907	68599534	53519032	848.430	745.122

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Quantitation Report (Not Reviewed)


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070935.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 15:27
Operator : YP\AJ
Sample : AR1254ICC750
Misc :
ALS Vial : 21 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1254ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 16:19:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 16:07:11 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.514 min
Delta R.T.: 0.000 min
Response: 116345667
Conc: 73.84 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750




#1 Tetrachloro-m-xylene

R.T.: 3.816 min
Delta R.T.: 0.000 min
Response: 81865229
Conc: 73.85 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.228 min
Delta R.T.: -0.002 min
Response: 85851924
Conc: 73.48 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.859 min
Delta R.T.: 0.000 min
Response: 59713645
Conc: 65.56 ng/ml



#26 AR-1254-1


R.T.: 6.515 min
Delta R.T.: 0.000 min
Response: 47300573
Conc: 722.46 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750




#26 AR-1254-1

R.T.: 5.708 min
Delta R.T.: 0.000 min
Response: 46349687
Conc: 701.39 ng/ml




#27 AR-1254-2

R.T.: 6.731 min
Delta R.T.: 0.000 min
Response: 72785411
Conc: 714.97 ng/ml



#27 AR-1254-2


R.T.: 5.856 min
Delta R.T.: 0.000 min
Response: 40339655
Conc: 710.29 ng/ml



#28 AR-1254-3


R.T.: 7.094 min
Delta R.T.: 0.000 min
Response: 73541603
Conc: 739.73 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750




#28 AR-1254-3

R.T.: 6.260 min
Delta R.T.: 0.000 min
Response: 62731680
Conc: 714.77 ng/ml




#29 AR-1254-4

R.T.: 7.377 min
Delta R.T.: 0.000 min
Response: 65835410
Conc: 758.45 ng/ml



#29 AR-1254-4


R.T.: 6.490 min
Delta R.T.: 0.000 min
Response: 40961211
Conc: 731.88 ng/ml



#30 AR-1254-5

R.T.: 7.793 min
Delta R.T.: 0.000 min
Response: 68599534
Conc: 848.43 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750



#30 AR-1254-5

R.T.: 6.907 min
Delta R.T.: 0.000 min
Response: 53519032
Conc: 745.12 ng/ml

Quantitation Report (Not Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070937.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 16:00
Operator : YP\AJ
Sample : AR1254ICC250
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 16:20:08 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 16:07:11 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.514	3.816	40672603	29879101	25.812	26.953
2)	SA Decachlor...	10.230	8.859	29858259	21257096	25.554	23.338
	Target Compounds						
26)	L6 AR-1254-1	6.516	5.708	16701604	18446961	255.096	279.151
27)	L6 AR-1254-2	6.732	5.856	26389026	16008677	259.218	281.875
28)	L6 AR-1254-3	7.095	6.260	25960285	24772594	261.126	282.261
29)	L6 AR-1254-4	7.377	6.489	22197209	16043587	255.721	286.662
30)	L6 AR-1254-5	7.793	6.907	23415043	20606411	289.594	286.894


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070937.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 16:00
Operator : YP\AJ
Sample : AR1254ICC250
Misc :
ALS Vial : 23 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1254ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 16:20:08 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 16:07:11 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.514 min
Delta R.T.: 0.000 min
Response: 40672603
Conc: 25.81 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.816 min
Delta R.T.: 0.000 min
Response: 29879101
Conc: 26.95 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.230 min
Delta R.T.: 0.000 min
Response: 29858259
Conc: 25.55 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.859 min
Delta R.T.: 0.000 min
Response: 21257096
Conc: 23.34 ng/ml



#26 AR-1254-1


R.T.: 6.516 min
Delta R.T.: 0.000 min
Response: 16701604
Conc: 255.10 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250




#26 AR-1254-1

R.T.: 5.708 min
Delta R.T.: 0.000 min
Response: 18446961
Conc: 279.15 ng/ml




#27 AR-1254-2

R.T.: 6.732 min
Delta R.T.: 0.000 min
Response: 26389026
Conc: 259.22 ng/ml



#27 AR-1254-2


R.T.: 5.856 min
Delta R.T.: 0.000 min
Response: 16008677
Conc: 281.87 ng/ml



#28 AR-1254-3


R.T.: 7.095 min
Delta R.T.: 0.000 min
Response: 25960285
Conc: 261.13 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250




#28 AR-1254-3

R.T.: 6.260 min
Delta R.T.: 0.000 min
Response: 24772594
Conc: 282.26 ng/ml




#29 AR-1254-4

R.T.: 7.377 min
Delta R.T.: 0.000 min
Response: 22197209
Conc: 255.72 ng/ml



#29 AR-1254-4


R.T.: 6.489 min
Delta R.T.: 0.000 min
Response: 16043587
Conc: 286.66 ng/ml



#30 AR-1254-5

R.T.: 7.793 min
Delta R.T.: 0.000 min
Response: 23415043
Conc: 289.59 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250



#30 AR-1254-5

R.T.: 6.907 min
Delta R.T.: 0.000 min
Response: 20606411
Conc: 286.89 ng/ml

Quantitation Report (Not Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070940.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 16:48
Operator : YP\AJ
Sample : AR1268ICC1000
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 17:58:03 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 17:56:10 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.516	3.817	152.3E6	105.1E6	94.769	94.009
2)	SA Decachlor...	10.232	8.860	195.7E6	148.5E6	93.377	94.968
	Target Compounds						
41)	L9 AR-1268-1	8.730	7.729	226.6E6	165.2E6	945.404	945.364
42)	L9 AR-1268-2	8.824	7.795	190.4E6	137.8E6	950.797	945.407
43)	L9 AR-1268-3	9.056	8.001	163.6E6	120.1E6	940.749	965.766
44)	L9 AR-1268-4	9.474	8.295	70377008	51096977	960.945	957.859
45)	L9 AR-1268-5	9.891	8.596	459.6E6	357.1E6	952.414	962.556

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Quantitation Report (Not Reviewed)


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070940.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 16:48
Operator : YP\AJ
Sample : AR1268ICC1000
Misc :
ALS Vial : 26 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1268ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 17:58:03 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 17:56:10 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.516 min
Delta R.T.: 0.003 min
Response: 152348529
Conc: 94.77 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC1000




#1 Tetrachloro-m-xylene

R.T.: 3.817 min
Delta R.T.: 0.000 min
Response: 105128984
Conc: 94.01 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.232 min
Delta R.T.: 0.003 min
Response: 195698435
Conc: 93.38 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.860 min
Delta R.T.: 0.000 min
Response: 148482089
Conc: 94.97 ng/ml



#41 AR-1268-1


R.T.: 8.730 min
Delta R.T.: 0.001 min
Response: 226575551
Conc: 945.40 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC1000




#41 AR-1268-1

R.T.: 7.729 min
Delta R.T.: 0.002 min
Response: 165156091
Conc: 945.36 ng/ml




#42 AR-1268-2

R.T.: 8.824 min
Delta R.T.: 0.003 min
Response: 190380276
Conc: 950.80 ng/ml



#42 AR-1268-2


R.T.: 7.795 min
Delta R.T.: 0.001 min
Response: 137801505
Conc: 945.41 ng/ml



#43 AR-1268-3


R.T.: 9.056 min
Delta R.T.: 0.003 min
Response: 163648333
Conc: 940.75 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC1000




#43 AR-1268-3

R.T.: 8.001 min
Delta R.T.: 0.001 min
Response: 120132746
Conc: 965.77 ng/ml




#44 AR-1268-4

R.T.: 9.474 min
Delta R.T.: 0.002 min
Response: 70377008
Conc: 960.95 ng/ml



#44 AR-1268-4


R.T.: 8.295 min
Delta R.T.: 0.002 min
Response: 51096977
Conc: 957.86 ng/ml



#45 AR-1268-5

R.T.: 9.891 min
Delta R.T.: 0.003 min
Response: 459553484
Conc: 952.41 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC1000



#45 AR-1268-5

R.T.: 8.596 min
Delta R.T.: 0.000 min
Response: 357060145
Conc: 962.56 ng/ml

Quantitation Report (Not Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070941.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 17:05
Operator : YP\AJ
Sample : AR1268ICC750
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 17:58:14 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 17:56:10 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.514	3.817	119.6E6	83914927	74.397	75.039
2)	SA Decachlor...	10.230	8.861	154.1E6	118.4E6	73.551	75.751
	Target Compounds						
41)	L9 AR-1268-1	8.730	7.729	178.4E6	132.1E6	744.578	755.945
42)	L9 AR-1268-2	8.823	7.795	149.2E6	110.8E6	745.140	760.189
43)	L9 AR-1268-3	9.055	8.000	128.8E6	95630424	740.439	768.788
44)	L9 AR-1268-4	9.473	8.294	55249435	40878954	754.389	766.313
45)	L9 AR-1268-5	9.890	8.596	361.2E6	282.2E6	748.577	760.668


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070941.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 17:05
Operator : YP\AJ
Sample : AR1268ICC750
Misc :
ALS Vial : 27 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1268ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 17:58:14 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 17:56:10 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.514 min
Delta R.T.: 0.001 min
Response: 119597931
Conc: 74.40 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC750




#1 Tetrachloro-m-xylene

R.T.: 3.817 min
Delta R.T.: 0.000 min
Response: 83914927
Conc: 75.04 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.230 min
Delta R.T.: 0.000 min
Response: 154146813
Conc: 73.55 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.861 min
Delta R.T.: 0.000 min
Response: 118436920
Conc: 75.75 ng/ml



#41 AR-1268-1


R.T.: 8.730 min
Delta R.T.: 0.001 min
Response: 178445590
Conc: 744.58 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC750




#41 AR-1268-1

R.T.: 7.729 min
Delta R.T.: 0.002 min
Response: 132064453
Conc: 755.95 ng/ml




#42 AR-1268-2

R.T.: 8.823 min
Delta R.T.: 0.002 min
Response: 149201176
Conc: 745.14 ng/ml



#42 AR-1268-2


R.T.: 7.795 min
Delta R.T.: 0.001 min
Response: 110804324
Conc: 760.19 ng/ml



#43 AR-1268-3


R.T.: 9.055 min
Delta R.T.: 0.002 min
Response: 128803338
Conc: 740.44 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC750




#43 AR-1268-3

R.T.: 8.000 min
Delta R.T.: 0.000 min
Response: 95630424
Conc: 768.79 ng/ml




#44 AR-1268-4

R.T.: 9.473 min
Delta R.T.: 0.001 min
Response: 55249435
Conc: 754.39 ng/ml



#44 AR-1268-4


R.T.: 8.294 min
Delta R.T.: 0.000 min
Response: 40878954
Conc: 766.31 ng/ml



#45 AR-1268-5

R.T.: 9.890 min
Delta R.T.: 0.002 min
Response: 361198889
Conc: 748.58 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC750



#45 AR-1268-5

R.T.: 8.596 min
Delta R.T.: 0.000 min
Response: 282169730
Conc: 760.67 ng/ml

Quantitation Report (Not Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070942.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 17:21
Operator : YP\AJ
Sample : AR1268ICC500
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 17:58:28 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 17:56:10 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.513	3.816	80378727	55914176	50.000	50.000
2)	SA Decachlor...	10.230	8.860	104.8E6	78175161	50.000	50.000
	Target Compounds						
41)	L9 AR-1268-1	8.729	7.728	119.8E6	87398242	500.000	500.273
42)	L9 AR-1268-2	8.821	7.793	100.1E6	72879434	500.000	500.000
43)	L9 AR-1268-3	9.053	8.000	86977678	62195601	500.000	500.000
44)	L9 AR-1268-4	9.472	8.293	36618642	26672497	500.000	500.000
45)	L9 AR-1268-5	9.888	8.596	241.3E6	185.5E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Quantitation Report (Not Reviewed)


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070942.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 17:21
Operator : YP\AJ
Sample : AR1268ICC500
Misc :
ALS Vial : 28 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1268ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 17:58:28 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 17:56:10 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.513 min
Delta R.T.: 0.000 min
Response: 80378727
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.816 min
Delta R.T.: 0.000 min
Response: 55914176
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.230 min
Delta R.T.: 0.000 min
Response: 104789095
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.860 min
Delta R.T.: 0.000 min
Response: 78175161
Conc: 50.00 ng/ml



#41 AR-1268-1


R.T.: 8.729 min
Delta R.T.: 0.000 min
Response: 119830016
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC500




#41 AR-1268-1

R.T.: 7.728 min
Delta R.T.: 0.000 min
Response: 87398242
Conc: 500.27 ng/ml




#42 AR-1268-2

R.T.: 8.821 min
Delta R.T.: 0.000 min
Response: 100116173
Conc: 500.00 ng/ml



#42 AR-1268-2


R.T.: 7.793 min
Delta R.T.: 0.000 min
Response: 72879434
Conc: 500.00 ng/ml



#43 AR-1268-3


R.T.: 9.053 min
Delta R.T.: 0.000 min
Response: 86977678
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC500




#43 AR-1268-3

R.T.: 8.000 min
Delta R.T.: 0.000 min
Response: 62195601
Conc: 500.00 ng/ml




#44 AR-1268-4

R.T.: 9.472 min
Delta R.T.: 0.000 min
Response: 36618642
Conc: 500.00 ng/ml



#44 AR-1268-4


R.T.: 8.293 min
Delta R.T.: 0.000 min
Response: 26672497
Conc: 500.00 ng/ml



#45 AR-1268-5

R.T.: 9.888 min
Delta R.T.: 0.000 min
Response: 241257148
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC500



#45 AR-1268-5

R.T.: 8.596 min
Delta R.T.: 0.000 min
Response: 185474929
Conc: 500.00 ng/ml

Quantitation Report (Not Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070943.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 17:37
Operator : YP\AJ
Sample : AR1268ICC250
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 17:58:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 17:56:10 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.515	3.817	41125313	29588552	25.582	26.459
2)	SA Decachlor...	10.232	8.859	54285349	40518249	25.902	25.915
	Target Compounds						
41)	L9 AR-1268-1	8.730	7.729	61973048	44387893	258.587	254.079
42)	L9 AR-1268-2	8.824	7.794	51237231	36694564	255.889	251.748
43)	L9 AR-1268-3	9.055	8.000	44311644	31811936	254.730	255.741
44)	L9 AR-1268-4	9.473	8.295	18700738	13922493	255.345	260.990
45)	L9 AR-1268-5	9.890	8.596	124.6E6	91636186	258.260	247.031

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Quantitation Report (Not Reviewed)


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070943.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 17:37
Operator : YP\AJ
Sample : AR1268ICC250
Misc :
ALS Vial : 29 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1268ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 17:58:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 17:56:10 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.515 min
Delta R.T.: 0.002 min
Response: 41125313
Conc: 25.58 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.817 min
Delta R.T.: 0.000 min
Response: 29588552
Conc: 26.46 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.232 min
Delta R.T.: 0.003 min
Response: 54285349
Conc: 25.90 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.859 min
Delta R.T.: 0.000 min
Response: 40518249
Conc: 25.92 ng/ml



#41 AR-1268-1


R.T.: 8.730 min
Delta R.T.: 0.002 min
Response: 61973048
Conc: 258.59 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250




#41 AR-1268-1

R.T.: 7.729 min
Delta R.T.: 0.001 min
Response: 44387893
Conc: 254.08 ng/ml




#42 AR-1268-2

R.T.: 8.824 min
Delta R.T.: 0.002 min
Response: 51237231
Conc: 255.89 ng/ml



#42 AR-1268-2


R.T.: 7.794 min
Delta R.T.: 0.000 min
Response: 36694564
Conc: 251.75 ng/ml



#43 AR-1268-3


R.T.: 9.055 min
Delta R.T.: 0.002 min
Response: 44311644
Conc: 254.73 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250




#43 AR-1268-3

R.T.: 8.000 min
Delta R.T.: 0.000 min
Response: 31811936
Conc: 255.74 ng/ml




#44 AR-1268-4

R.T.: 9.473 min
Delta R.T.: 0.002 min
Response: 18700738
Conc: 255.34 ng/ml



#44 AR-1268-4


R.T.: 8.295 min
Delta R.T.: 0.002 min
Response: 13922493
Conc: 260.99 ng/ml



#45 AR-1268-5

R.T.: 9.890 min
Delta R.T.: 0.002 min
Response: 124614186
Conc: 258.26 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250



#45 AR-1268-5

R.T.: 8.596 min
Delta R.T.: 0.000 min
Response: 91636186
Conc: 247.03 ng/ml

Quantitation Report (Not Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070944.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 17:53
Operator : YP\AJ
Sample : AR1268ICC050
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 18:13:54 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 17:56:10 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.515	3.816	6795391	6304440	4.206	5.529 #
2)	SA Decachlor...	10.230	8.859	10035177	7161586	4.762	4.510
	Target Compounds						
41)	L9 AR-1268-1	8.729	7.728	11274140	8298716	46.621	47.221
42)	L9 AR-1268-2	8.822	7.794	9103995	6916833	45.210	47.131
43)	L9 AR-1268-3	9.054	7.999	8398993	5823381	48.183	46.078
44)	L9 AR-1268-4	9.471	8.293	3082639	2842824	41.712	52.149 #
45)	L9 AR-1268-5	9.889	8.596	22937830	16928167	47.050	45.599

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Quantitation Report (Not Reviewed)


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070944.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 17:53
Operator : YP\AJ
Sample : AR1268ICC050
Misc :
ALS Vial : 30 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1268ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 18:13:54 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 27 17:56:10 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.515 min
Delta R.T.: 0.002 min
Response: 6795391
Conc: 4.21 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC050




#1 Tetrachloro-m-xylene

R.T.: 3.816 min
Delta R.T.: 0.000 min
Response: 6304440
Conc: 5.53 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.230 min
Delta R.T.: 0.000 min
Response: 10035177
Conc: 4.76 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.859 min
Delta R.T.: -0.001 min
Response: 7161586
Conc: 4.51 ng/ml



#41 AR-1268-1


R.T.: 8.729 min
Delta R.T.: 0.000 min
Response: 11274140
Conc: 46.62 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC050




#41 AR-1268-1

R.T.: 7.728 min
Delta R.T.: 0.000 min
Response: 8298716
Conc: 47.22 ng/ml




#42 AR-1268-2

R.T.: 8.822 min
Delta R.T.: 0.000 min
Response: 9103995
Conc: 45.21 ng/ml



#42 AR-1268-2


R.T.: 7.794 min
Delta R.T.: 0.000 min
Response: 6916833
Conc: 47.13 ng/ml



#43 AR-1268-3


R.T.: 9.054 min
Delta R.T.: 0.002 min
Response: 8398993
Conc: 48.18 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC050




#43 AR-1268-3

R.T.: 7.999 min
Delta R.T.: 0.000 min
Response: 5823381
Conc: 46.08 ng/ml




#44 AR-1268-4

R.T.: 9.471 min
Delta R.T.: 0.000 min
Response: 3082639
Conc: 41.71 ng/ml



#44 AR-1268-4


R.T.: 8.293 min
Delta R.T.: 0.000 min
Response: 2842824
Conc: 52.15 ng/ml



#45 AR-1268-5

R.T.: 9.889 min
Delta R.T.: 0.000 min
Response: 22937830
Conc: 47.05 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC050



#45 AR-1268-5

R.T.: 8.596 min
Delta R.T.: 0.000 min
Response: 16928167
Conc: 45.60 ng/ml

Quantitation Report (Not Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070948.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 19:47
Operator : YP\AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP032725AR1254

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 28 01:39:10 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 28 01:37:11 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.513	3.816	80813408	55809150	53.470	50.687
2)	SA Decachlor...	10.228	8.859	60442284	44113864	53.599	53.797
	Target Compounds						
26)	L6 AR-1254-1	6.514	5.708	34026950	33372474	528.449	505.507
27)	L6 AR-1254-2	6.730	5.856	52171420	28060451	516.086	491.442
28)	L6 AR-1254-3	7.093	6.260	52059367	42608413	522.014	490.647
29)	L6 AR-1254-4	7.375	6.489	45346150	27407702	523.647	488.547
30)	L6 AR-1254-5	7.792	6.907	48183080	36336253	599.658	503.617

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Quantitation Report (Not Reviewed)


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032725\
Data File : PP070948.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 19:47
Operator : YP\AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 34 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
ICVPP032725AR1254

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 28 01:39:10 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 28 01:37:11 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.513 min
Delta R.T.: -0.001 min
Response: 80813408
Conc: 53.47 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP032725AR1254




#1 Tetrachloro-m-xylene

R.T.: 3.816 min
Delta R.T.: 0.000 min
Response: 55809150
Conc: 50.69 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.228 min
Delta R.T.: -0.002 min
Response: 60442284
Conc: 53.60 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.859 min
Delta R.T.: 0.000 min
Response: 44113864
Conc: 53.80 ng/ml



#26 AR-1254-1


R.T.: 6.514 min
Delta R.T.: -0.001 min
Response: 34026950
Conc: 528.45 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP032725AR1254




#26 AR-1254-1

R.T.: 5.708 min
Delta R.T.: 0.000 min
Response: 33372474
Conc: 505.51 ng/ml




#27 AR-1254-2

R.T.: 6.730 min
Delta R.T.: -0.002 min
Response: 52171420
Conc: 516.09 ng/ml



#27 AR-1254-2


R.T.: 5.856 min
Delta R.T.: 0.000 min
Response: 28060451
Conc: 491.44 ng/ml



#28 AR-1254-3


R.T.: 7.093 min
Delta R.T.: -0.001 min
Response: 52059367
Conc: 522.01 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP032725AR1254




#28 AR-1254-3

R.T.: 6.260 min
Delta R.T.: 0.000 min
Response: 42608413
Conc: 490.65 ng/ml




#29 AR-1254-4

R.T.: 7.375 min
Delta R.T.: -0.001 min
Response: 45346150
Conc: 523.65 ng/ml



#29 AR-1254-4


R.T.: 6.489 min
Delta R.T.: 0.000 min
Response: 27407702
Conc: 488.55 ng/ml



#30 AR-1254-5

R.T.: 7.792 min
Delta R.T.: -0.001 min
Response: 48183080
Conc: 599.66 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP032725AR1254



#30 AR-1254-5

R.T.: 6.907 min
Delta R.T.: 0.000 min
Response: 36336253
Conc: 503.62 ng/ml

Quantitation Report (Not Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040825\
Data File : PP071136.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Apr 2025 09:37
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 10:57:00 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 28 03:13:46 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml	
	System Monitoring Compounds							
1)	SA Tetrachlo...	4.524	3.817	75031562	49693990	51.336	47.937	
2)	SA Decachlor...	10.255	8.862	53747529	34514110	51.061	43.789	
	Target Compounds							
3)	L1 AR-1016-1	5.679	4.904	25640835	19684563	504.232	491.032	
4)	L1 AR-1016-2	5.679	4.923	25640835	27978272	328.503	483.509	#
5)	L1 AR-1016-3	5.763	5.101	23060044	16163902	488.823	522.709	
6)	L1 AR-1016-4	5.861	5.142	19323906	13212574	525.739	542.479	
7)	L1 AR-1016-5	6.154	5.357	18307730	17134433	539.858	530.130	
8)	L2 AR-1221-1	4.723	4.026	2456666	2180640	141.565	143.053	
9)	L2 AR-1221-2	4.813	4.115	3663977	3206249	282.869	280.357	
10)	L2 AR-1221-3	4.888	4.191	13293050	11445681	318.646	335.759	
11)	L3 AR-1232-1	4.888	4.191	13293050	11445681	399.326	408.969	
12)	L3 AR-1232-2	5.414	4.923	18281252	27978272	1149.017	997.047	
13)	L3 AR-1232-3	5.701	5.101	38706585	16163902	1129.447	1085.584	
14)	L3 AR-1232-4	5.861	5.185	19323906	15635898	1230.228	1221.138	
15)	L3 AR-1232-5	5.951	5.357	15905498	17134433	1494.828	1208.797	
16)	L4 AR-1242-1	5.679	4.904	25640835	19684563	583.049	575.642	
17)	L4 AR-1242-2	5.679	4.923	25640835	27978272	398.668	567.998	#
18)	L4 AR-1242-3	5.763	5.101	23060044	16163902	561.899	606.656	
19)	L4 AR-1242-4	5.861	5.185	19323906	15635898	633.086	592.757	
20)	L4 AR-1242-5	6.592	5.710	6855876	16691995	187.779	539.177	#
21)	L5 AR-1248-1	5.679	4.904	25640835	19684563	706.602	726.905	
22)	L5 AR-1248-2	5.951	5.142	15905498	13212574	353.939	359.795	
23)	L5 AR-1248-3	6.154	5.185	18307730	15635898	361.545	404.453	
24)	L5 AR-1248-4	6.528	5.357	18184937	17134433	283.486	379.576	#
25)	L5 AR-1248-5	6.592	5.750	6855876	4488455	112.675	105.168	
26)	L6 AR-1254-1	6.528	5.710	18184937	16691995	282.418	252.841	
27)	L6 AR-1254-2	6.745	5.858	16493060	11838446	163.151	207.335	#
28)	L6 AR-1254-3	7.109	6.244	8745572	4764413	87.694	54.863	#
29)	L6 AR-1254-4	7.392	6.491	6515054	3701399	75.234	65.978	
30)	L6 AR-1254-5	7.807	6.908	49428901	33433766	615.162	463.388	
31)	L7 AR-1260-1	7.273	6.393	34786874	28041460	518.026	547.617	
32)	L7 AR-1260-2	7.527	6.582	51604417	34095263	536.784	534.460	
33)	L7 AR-1260-3	7.885	6.735	40227656	29796945	503.360	546.865	
34)	L7 AR-1260-4	8.076	7.207	7015115	23421548	89.096	489.721	#
35)	L7 AR-1260-5	8.431	7.448	80513512	57532352	506.737	462.069	
36)	L8 AR-1262-1	8.076	6.947	7015115	28743067	65.993	335.927	#
37)	L8 AR-1262-2	8.431	7.207	80513512	23421548	405.849	342.098	
38)	L8 AR-1262-3	8.756	7.730	49500185	13610298	375.736	224.136	#
39)	L8 AR-1262-4	8.835	7.794	29791065	37704999	311.574	379.300	
40)	L8 AR-1262-5	9.493	8.296	21754078	14122672	334.337	301.305	
41)	L9 AR-1268-1	8.756	7.730	49500185	13610298	210.185	79.406	#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040825\
Data File : PP071136.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Apr 2025 09:37
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 10:57:00 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 28 03:13:46 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.835	7.794	29791065	37704999	152.528	263.151 #
43) L9	AR-1268-3	9.071	8.001	324667	608932	1.900	4.945 #
44) L9	AR-1268-4	9.493	8.296	21754078	14122672	307.490	267.408
45) L9	AR-1268-5	9.911	8.598	5287089	3490210	11.103	9.645

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Quantitation Report (Not Reviewed)


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040825\
Data File : PP071136.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Apr 2025 09:37
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 10:57:00 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 28 03:13:46 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.524 min
Delta R.T.: 0.012 min
Response: 75031562
Conc: 51.34 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.817 min
Delta R.T.: 0.000 min
Response: 49693990
Conc: 47.94 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.255 min
Delta R.T.: 0.030 min
Response: 53747529
Conc: 51.06 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.862 min
Delta R.T.: 0.003 min
Response: 34514110
Conc: 43.79 ng/ml



#3 AR-1016-1


R.T.: 5.679 min
Delta R.T.: 0.016 min
Response: 25640835
Conc: 504.23 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.904 min
Delta R.T.: 0.001 min
Response: 19684563
Conc: 491.03 ng/ml




#4 AR-1016-2

R.T.: 5.679 min
Delta R.T.: -0.006 min
Response: 25640835
Conc: 328.50 ng/ml



#4 AR-1016-2


R.T.: 4.923 min
Delta R.T.: 0.001 min
Response: 27978272
Conc: 483.51 ng/ml



#5 AR-1016-3


R.T.: 5.763 min
Delta R.T.: 0.016 min
Response: 23060044
Conc: 488.82 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 5.101 min
Delta R.T.: 0.001 min
Response: 16163902
Conc: 522.71 ng/ml




#6 AR-1016-4

R.T.: 5.861 min
Delta R.T.: 0.015 min
Response: 19323906
Conc: 525.74 ng/ml



#6 AR-1016-4


R.T.: 5.142 min
Delta R.T.: 0.001 min
Response: 13212574
Conc: 542.48 ng/ml



#7 AR-1016-5


R.T.: 6.154 min
Delta R.T.: 0.016 min
Response: 18307730
Conc: 539.86 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.357 min
Delta R.T.: 0.002 min
Response: 17134433
Conc: 530.13 ng/ml




#8 AR-1221-1

R.T.: 4.723 min
Delta R.T.: 0.007 min
Response: 2456666
Conc: 141.56 ng/ml



#8 AR-1221-1


R.T.: 4.026 min
Delta R.T.: 0.000 min
Response: 2180640
Conc: 143.05 ng/ml



#9 AR-1221-2


R.T.: 4.813 min
Delta R.T.: 0.010 min
Response: 3663977
Conc: 282.87 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#9 AR-1221-2

R.T.: 4.115 min
Delta R.T.: 0.001 min
Response: 3206249
Conc: 280.36 ng/ml




#10 AR-1221-3

R.T.: 4.888 min
Delta R.T.: 0.010 min
Response: 13293050
Conc: 318.65 ng/ml



#10 AR-1221-3


R.T.: 4.191 min
Delta R.T.: 0.000 min
Response: 11445681
Conc: 335.76 ng/ml



#11 AR-1232-1


R.T.: 4.888 min
Delta R.T.: 0.005 min
Response: 13293050
Conc: 399.33 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#11 AR-1232-1

R.T.: 4.191 min
Delta R.T.: 0.000 min
Response: 11445681
Conc: 408.97 ng/ml




#12 AR-1232-2

R.T.: 5.414 min
Delta R.T.: 0.006 min
Response: 18281252
Conc: 1149.02 ng/ml



#12 AR-1232-2


R.T.: 4.923 min
Delta R.T.: 0.000 min
Response: 27978272
Conc: 997.05 ng/ml



#13 AR-1232-3


R.T.: 5.701 min
Delta R.T.: 0.007 min
Response: 38706585
Conc: 1129.45 ng/m

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#13 AR-1232-3

R.T.: 5.101 min
Delta R.T.: 0.000 min
Response: 16163902
Conc: 1085.58 ng/ml




#14 AR-1232-4

R.T.: 5.861 min
Delta R.T.: 0.007 min
Response: 19323906
Conc: 1230.23 ng/ml



#14 AR-1232-4


R.T.: 5.185 min
Delta R.T.: 0.000 min
Response: 15635898
Conc: 1221.14 ng/ml



#15 AR-1232-5


R.T.: 5.951 min
Delta R.T.: 0.008 min
Response: 15905498
Conc: 1494.83 ng/m

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#15 AR-1232-5

R.T.: 5.357 min
Delta R.T.: 0.000 min
Response: 17134433
Conc: 1208.80 ng/ml




#16 AR-1242-1

R.T.: 5.679 min
Delta R.T.: 0.015 min
Response: 25640835
Conc: 583.05 ng/ml



#16 AR-1242-1


R.T.: 4.904 min
Delta R.T.: 0.002 min
Response: 19684563
Conc: 575.64 ng/ml



#17 AR-1242-2


R.T.: 5.679 min
Delta R.T.: -0.007 min
Response: 25640835
Conc: 398.67 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#17 AR-1242-2

R.T.: 4.923 min
Delta R.T.: 0.002 min
Response: 27978272
Conc: 568.00 ng/ml




#18 AR-1242-3

R.T.: 5.763 min
Delta R.T.: 0.015 min
Response: 23060044
Conc: 561.90 ng/ml



#18 AR-1242-3


R.T.: 5.101 min
Delta R.T.: 0.002 min
Response: 16163902
Conc: 606.66 ng/ml



#19 AR-1242-4


R.T.: 5.861 min
Delta R.T.: 0.015 min
Response: 19323906
Conc: 633.09 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#19 AR-1242-4

R.T.: 5.185 min
Delta R.T.: 0.003 min
Response: 15635898
Conc: 592.76 ng/ml




#20 AR-1242-5

R.T.: 6.592 min
Delta R.T.: 0.017 min
Response: 6855876
Conc: 187.78 ng/ml



#20 AR-1242-5


R.T.: 5.710 min
Delta R.T.: 0.003 min
Response: 16691995
Conc: 539.18 ng/ml



#21 AR-1248-1


R.T.: 5.679 min
Delta R.T.: 0.014 min
Response: 25640835
Conc: 706.60 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#21 AR-1248-1

R.T.: 4.904 min
Delta R.T.: 0.001 min
Response: 19684563
Conc: 726.90 ng/ml




#22 AR-1248-2

R.T.: 5.951 min
Delta R.T.: 0.014 min
Response: 15905498
Conc: 353.94 ng/ml



#22 AR-1248-2


R.T.: 5.142 min
Delta R.T.: 0.002 min
Response: 13212574
Conc: 359.80 ng/ml



#23 AR-1248-3


R.T.: 6.154 min
Delta R.T.: 0.014 min
Response: 18307730
Conc: 361.54 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#23 AR-1248-3

R.T.: 5.185 min
Delta R.T.: 0.002 min
Response: 15635898
Conc: 404.45 ng/ml




#24 AR-1248-4

R.T.: 6.528 min
Delta R.T.: -0.010 min
Response: 18184937
Conc: 283.49 ng/ml



#24 AR-1248-4


R.T.: 5.357 min
Delta R.T.: 0.002 min
Response: 17134433
Conc: 379.58 ng/ml



#25 AR-1248-5


R.T.: 6.592 min
Delta R.T.: 0.015 min
Response: 6855876
Conc: 112.68 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#25 AR-1248-5

R.T.: 5.750 min
Delta R.T.: 0.002 min
Response: 4488455
Conc: 105.17 ng/ml




#26 AR-1254-1

R.T.: 6.528 min
Delta R.T.: 0.013 min
Response: 18184937
Conc: 282.42 ng/ml



#26 AR-1254-1


R.T.: 5.710 min
Delta R.T.: 0.002 min
Response: 16691995
Conc: 252.84 ng/ml



#27 AR-1254-2


R.T.: 6.745 min
Delta R.T.: 0.013 min
Response: 16493060
Conc: 163.15 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#27 AR-1254-2

R.T.: 5.858 min
Delta R.T.: 0.001 min
Response: 11838446
Conc: 207.34 ng/ml




#28 AR-1254-3

R.T.: 7.109 min
Delta R.T.: 0.014 min
Response: 8745572
Conc: 87.69 ng/ml



#28 AR-1254-3


R.T.: 6.244 min
Delta R.T.: -0.016 min
Response: 4764413
Conc: 54.86 ng/ml



#29 AR-1254-4


R.T.: 7.392 min
Delta R.T.: 0.015 min
Response: 6515054
Conc: 75.23 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#29 AR-1254-4

R.T.: 6.491 min
Delta R.T.: 0.002 min
Response: 3701399
Conc: 65.98 ng/ml




#30 AR-1254-5

R.T.: 7.807 min
Delta R.T.: 0.014 min
Response: 49428901
Conc: 615.16 ng/ml



#30 AR-1254-5


R.T.: 6.908 min
Delta R.T.: 0.001 min
Response: 33433766
Conc: 463.39 ng/ml



#31 AR-1260-1


R.T.: 7.273 min
Delta R.T.: 0.017 min
Response: 34786874
Conc: 518.03 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#31 AR-1260-1

R.T.: 6.393 min
Delta R.T.: 0.001 min
Response: 28041460
Conc: 547.62 ng/ml




#32 AR-1260-2

R.T.: 7.527 min
Delta R.T.: 0.017 min
Response: 51604417
Conc: 536.78 ng/ml



#32 AR-1260-2


R.T.: 6.582 min
Delta R.T.: 0.002 min
Response: 34095263
Conc: 534.46 ng/ml



#33 AR-1260-3


R.T.: 7.885 min
Delta R.T.: 0.018 min
Response: 40227656
Conc: 503.36 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#33 AR-1260-3

R.T.: 6.735 min
Delta R.T.: 0.001 min
Response: 29796945
Conc: 546.87 ng/ml




#34 AR-1260-4

R.T.: 8.076 min
Delta R.T.: -0.016 min
Response: 7015115
Conc: 89.10 ng/ml



#34 AR-1260-4


R.T.: 7.207 min
Delta R.T.: 0.002 min
Response: 23421548
Conc: 489.72 ng/ml



#35 AR-1260-5


R.T.: 8.431 min
Delta R.T.: 0.020 min
Response: 80513512
Conc: 506.74 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#35 AR-1260-5

R.T.: 7.448 min
Delta R.T.: 0.002 min
Response: 57532352
Conc: 462.07 ng/ml




#36 AR-1262-1

R.T.: 8.076 min
Delta R.T.: -0.017 min
Response: 7015115
Conc: 65.99 ng/ml



#36 AR-1262-1


R.T.: 6.947 min
Delta R.T.: 0.002 min
Response: 28743067
Conc: 335.93 ng/ml



#37 AR-1262-2


R.T.: 8.431 min
Delta R.T.: 0.019 min
Response: 80513512
Conc: 405.85 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#37 AR-1262-2

R.T.: 7.207 min
Delta R.T.: 0.003 min
Response: 23421548
Conc: 342.10 ng/ml




#38 AR-1262-3

R.T.: 8.756 min
Delta R.T.: 0.027 min
Response: 49500185
Conc: 375.74 ng/ml



#38 AR-1262-3


R.T.: 7.730 min
Delta R.T.: 0.002 min
Response: 13610298
Conc: 224.14 ng/ml



#39 AR-1262-4


R.T.: 8.835 min
Delta R.T.: 0.019 min
Response: 29791065
Conc: 311.57 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#39 AR-1262-4

R.T.: 7.794 min
Delta R.T.: 0.000 min
Response: 37704999
Conc: 379.30 ng/ml




#40 AR-1262-5

R.T.: 9.493 min
Delta R.T.: 0.024 min
Response: 21754078
Conc: 334.34 ng/ml



#40 AR-1262-5


R.T.: 8.296 min
Delta R.T.: 0.002 min
Response: 14122672
Conc: 301.31 ng/ml



#41 AR-1268-1


R.T.: 8.756 min
Delta R.T.: 0.027 min
Response: 49500185
Conc: 210.19 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#41 AR-1268-1

R.T.: 7.730 min
Delta R.T.: 0.002 min
Response: 13610298
Conc: 79.41 ng/ml




#42 AR-1268-2

R.T.: 8.835 min
Delta R.T.: 0.014 min
Response: 29791065
Conc: 152.53 ng/ml



#42 AR-1268-2


R.T.: 7.794 min
Delta R.T.: 0.000 min
Response: 37704999
Conc: 263.15 ng/ml



#43 AR-1268-3


R.T.: 9.071 min
Delta R.T.: 0.019 min
Response: 324667
Conc: 1.90 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#43 AR-1268-3

R.T.: 8.001 min
Delta R.T.: 0.001 min
Response: 608932
Conc: 4.94 ng/ml




#44 AR-1268-4

R.T.: 9.493 min
Delta R.T.: 0.021 min
Response: 21754078
Conc: 307.49 ng/ml



#44 AR-1268-4


R.T.: 8.296 min
Delta R.T.: 0.003 min
Response: 14122672
Conc: 267.41 ng/ml



#45 AR-1268-5

R.T.: 9.911 min
Delta R.T.: 0.023 min
Response: 5287089
Conc: 11.10 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500



#45 AR-1268-5

R.T.: 8.598 min
Delta R.T.: 0.002 min
Response: 3490210
Conc: 9.64 ng/ml