

Surrogate Summary

SDG No.: Q1730

Client: Nobis Group

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PO109971.D	PIBLK-PO109971.D	Tetrachloro-m-xylene	1	20	18.5	93		60	140
		Decachlorobiphenyl	1	20	20.0	100		60	140
		Tetrachloro-m-xylene	2	20	18.6	93		60	140
		Decachlorobiphenyl	2	20	21.3	107		60	140
I.BLK-PO110267.D	PIBLK-PO110267.D	Tetrachloro-m-xylene	1	20	19.5	97		60	140
		Decachlorobiphenyl	1	20	18.4	92		60	140
		Tetrachloro-m-xylene	2	20	18.8	94		60	140
		Decachlorobiphenyl	2	20	15.9	79		60	140
Q1730-01	OU4-VSL-15-040325	Tetrachloro-m-xylene	1	20	20.2	101		44	130
		Decachlorobiphenyl	1	20	17.6	88		60	125
		Tetrachloro-m-xylene	2	20	19.4	97		44	130
		Decachlorobiphenyl	2	20	15.9	79		60	125
Q1730-03	OU4-VSL-16-040325	Tetrachloro-m-xylene	1	20	22.4	112		44	130
		Decachlorobiphenyl	1	20	19.4	97		60	125
		Tetrachloro-m-xylene	2	20	21.5	108		44	130
		Decachlorobiphenyl	2	20	16.8	84		60	125
Q1730-03MS	OU4-VSL-16-040325MS	Tetrachloro-m-xylene	1	20	22.8	114		44	130
		Decachlorobiphenyl	1	20	20.9	104		60	125
		Tetrachloro-m-xylene	2	20	21.8	109		44	130
		Decachlorobiphenyl	2	20	18.0	90		60	125
Q1730-03MSD	OU4-VSL-16-040325MSD	Tetrachloro-m-xylene	1	20	21.6	108		44	130
		Decachlorobiphenyl	1	20	19.8	99		60	125
		Tetrachloro-m-xylene	2	20	20.8	104		44	130
		Decachlorobiphenyl	2	20	17.3	86		60	125
I.BLK-PO110282.D	PIBLK-PO110282.D	Tetrachloro-m-xylene	1	20	19.3	97		60	140
		Decachlorobiphenyl	1	20	17.2	86		60	140
		Tetrachloro-m-xylene	2	20	18.8	94		60	140
		Decachlorobiphenyl	2	20	16.2	81		60	140
PB167469BL	PB167469BL	Tetrachloro-m-xylene	1	20	19.0	95		44	130
		Decachlorobiphenyl	1	20	17.3	86		60	125
		Tetrachloro-m-xylene	2	20	18.3	91		44	130
		Decachlorobiphenyl	2	20	16.0	80		60	125
PB167469BS	PB167469BS	Tetrachloro-m-xylene	1	20	19.9	99		44	130
		Decachlorobiphenyl	1	20	19.1	95		60	125
		Tetrachloro-m-xylene	2	20	18.9	95		44	130
		Decachlorobiphenyl	2	20	17.3	86		60	125
I.BLK-PO110290.D	PIBLK-PO110290.D	Tetrachloro-m-xylene	1	20	19.5	98		60	140
		Decachlorobiphenyl	1	20	17.7	89		60	140
		Tetrachloro-m-xylene	2	20	18.9	94		60	140
		Decachlorobiphenyl	2	20	16.2	81		60	140
I.BLK-PO110296.D	PIBLK-PO110296.D	Tetrachloro-m-xylene	1	20	18.7	93		60	140

Surrogate Summary

SDG No.: Q1730

Client: Nobis Group

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PO110296.D	PIBLK-PO110296.D	Decachlorobiphenyl	1	20	17.8	89		60	140
		Tetrachloro-m-xylene	2	20	17.7	89		60	140
		Decachlorobiphenyl	2	20	15.4	77		60	140
Q1730-05	OU4-VSL-17-040325	Tetrachloro-m-xylene	1	20	23.2	116		44	130
		Decachlorobiphenyl	1	20	21.1	105		60	125
		Tetrachloro-m-xylene	2	20	22.1	110		44	130
		Decachlorobiphenyl	2	20	18.1	91		60	125
Q1730-07	OU4-PCS-TC-21-040325	Tetrachloro-m-xylene	1	20	20.3	102		44	130
		Decachlorobiphenyl	1	20	17.1	85		60	125
		Tetrachloro-m-xylene	2	20	19.2	96		44	130
		Decachlorobiphenyl	2	20	14.1	71		60	125
Q1730-09	OU4-PCS-TC-22-040325	Tetrachloro-m-xylene	1	20	20.7	104		44	130
		Decachlorobiphenyl	1	20	17.9	90		60	125
		Tetrachloro-m-xylene	2	20	19.7	98		44	130
		Decachlorobiphenyl	2	20	14.8	74		60	125
Q1730-11	OU4-PCS-TC-23-040325	Tetrachloro-m-xylene	1	20	20.9	104		44	130
		Decachlorobiphenyl	1	20	18.0	90		60	125
		Tetrachloro-m-xylene	2	20	19.9	99		44	130
		Decachlorobiphenyl	2	20	14.8	74		60	125
Q1730-13	OU4-PCS-TC-24-040325	Tetrachloro-m-xylene	1	20	20.5	102		44	130
		Decachlorobiphenyl	1	20	18.9	94		60	125
		Tetrachloro-m-xylene	2	20	19.4	97		44	130
		Decachlorobiphenyl	2	20	15.7	78		60	125
I.BLK-PO110306.D	PIBLK-PO110306.D	Tetrachloro-m-xylene	1	20	18.2	91		60	140
		Decachlorobiphenyl	1	20	18.4	92		60	140
		Tetrachloro-m-xylene	2	20	17.4	87		60	140
		Decachlorobiphenyl	2	20	15.3	77		60	140
I.BLK-PP070916.D	PIBLK-PP070916.D	Tetrachloro-m-xylene	1	20	18.8	94		60	140
		Decachlorobiphenyl	1	20	19.4	97		60	140
		Tetrachloro-m-xylene	2	20	18.1	91		60	140
		Decachlorobiphenyl	2	20	23.4	117		60	140
I.BLK-PP071140.D	PIBLK-PP071140.D	Tetrachloro-m-xylene	1	20	22.8	114		60	140
		Decachlorobiphenyl	1	20	23.9	119		60	140
		Tetrachloro-m-xylene	2	20	23.1	116		60	140
		Decachlorobiphenyl	2	20	21.5	108		60	140
Q1730-15	OU4-PCS-TC-25-040325	Tetrachloro-m-xylene	1	20	24.7	123		44	130
		Decachlorobiphenyl	1	20	23.3	116		60	125
		Tetrachloro-m-xylene	2	20	24.8	124		44	130
		Decachlorobiphenyl	2	20	20.5	102		60	125
Q1730-17	OU4-PCS-TC-26-040325	Tetrachloro-m-xylene	1	20	25.1	125		44	130
		Decachlorobiphenyl	1	20	24.9	124		60	125

Surrogate Summary

SDG No.: Q1730

Client: Nobis Group

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
Q1730-17	OU4-PCS-TC-26-040325	Tetrachloro-m-xylene	2	20	24.8	124		44	130
		Decachlorobiphenyl	2	20	22.0	110		60	125
Q1730-19	OU4-CF-15-040325	Tetrachloro-m-xylene	1	20	25.8	129		44	130
		Decachlorobiphenyl	1	20	24.1	121		60	125
		Tetrachloro-m-xylene	2	20	27.5	138	*	44	130
		Decachlorobiphenyl	2	20	21.7	108		60	125
I.BLK-PP071155.D	PIBLK-PP071155.D	Tetrachloro-m-xylene	1	20	20.6	103		60	140
		Decachlorobiphenyl	1	20	21.0	105		60	140
		Tetrachloro-m-xylene	2	20	21.1	105		60	140
		Decachlorobiphenyl	2	20	18.4	92		60	140