

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: NOBI03

Lab Code: CHEM **Case No.:** Q1730 **SAS No.:** Q1730 **SDG NO.:** Q1730

Instrument ID: ECD_L **Date(s) Analyzed:** 03/11/2025 03/11/2025

GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	6.24	6.14	6.34	25951700
		2	6.44	6.34	6.54	16397600
		3	7.06	6.96	7.16	82030600
		4	7.15	7.05	7.25	62943200
		5	7.93	7.83	8.03	45040400

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: NOBI03

Lab Code: CHEM **Case No.:** Q1730 **SAS No.:** Q1730 **SDG NO.:** Q1730

Instrument ID: ECD_L **Date(s) Analyzed:** 03/11/2025 03/11/2025

GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	5.00	4.90	5.10	26020300
		2	5.32	5.22	5.42	25489100
		3	5.68	5.58	5.78	28236400
		4	6.60	6.50	6.70	96540400
		5	7.04	6.94	7.14	93337600