

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1730 SAS No.: Q1730 SDG NO.: Q1730

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No. (PEM): PEM - PL094567.D Date Analyzed: 03/11/2025

Lab Sample No.(PEM): PEM Time Analyzed: 10:08

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.052	8.950	9.150	20.540	20.000	2.7
Tetrachloro-m-xylene	3.536	3.490	3.590	19.600	20.000	-2.0
alpha-BHC	3.992	3.940	4.040	10.230	10.000	2.3
beta-BHC	4.524	4.470	4.570	10.180	10.000	1.8
gamma-BHC (Lindane)	4.325	4.270	4.380	10.310	10.000	3.1
Endrin	6.572	6.500	6.640	43.430	50.000	-13.1
4,4'-DDT	7.021	6.950	7.090	87.030	100.000	-13.0
Methoxychlor	7.498	7.430	7.570	214.630	250.000	-14.1

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No. (PEM): PEM - PL094567.D Date Analyzed: 03/11/2025

Lab Sample No.(PEM): PEM Time Analyzed: 10:08

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.905	7.800	8.010	19.590	20.000	-2.1
Tetrachloro-m-xylene	2.771	2.720	2.820	19.320	20.000	-3.4
alpha-BHC	3.273	3.220	3.320	9.150	10.000	-8.5
beta-BHC	3.903	3.850	3.950	10.130	10.000	1.3
gamma-BHC (Lindane)	3.603	3.550	3.650	9.160	10.000	-8.4
Endrin	5.632	5.560	5.700	44.410	50.000	-11.2
4,4'-DDT	6.030	5.960	6.100	97.400	100.000	-2.6
Methoxychlor	6.605	6.530	6.680	224.380	250.000	-10.2