

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP031225

Review By	anahy	Review On	3/13/2025 11:41:15 AM
Supervise By	Jagrut	Supervise On	3/17/2025 4:44:19 PM
SubDirectory	BP031225	HP Acquire Method	BNA_P
		HP Processing Method	bp031225

STD. NAME	STD REF.#
Tune/Reschk	SP6717
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12653,10ul/1000ul sample
ICV/I.BLK	SP6686
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP024159.D	12 Mar 2025 15:36		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BP024160.D	12 Mar 2025 16:17		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BP024161.D	12 Mar 2025 16:57	Compound #32,54,65 removed from 5ppm	RC/JU	Ok,M
4	SSTDICC010	SSTDICC010	BP024162.D	12 Mar 2025 17:38		RC/JU	Ok
5	SSTDICC020	SSTDICC020	BP024163.D	12 Mar 2025 18:19		RC/JU	Ok
6	SSTDICCC040	SSTDICCC040	BP024164.D	12 Mar 2025 19:00	This calibration is good for 8270E and 8270E DOD methods	RC/JU	Ok
7	SSTDICC050	SSTDICC050	BP024165.D	12 Mar 2025 19:41		RC/JU	Ok,M
8	SSTDICC060	SSTDICC060	BP024166.D	12 Mar 2025 20:21	Compound #69 Removed from 60PPM	RC/JU	Ok,M
9	SSTDICC080	SSTDICC080	BP024167.D	12 Mar 2025 21:02	Compound #32,54,69 Removed from 80PPM	RC/JU	Ok,M
10	SSTDICV040	ICVBP031225	BP024168.D	12 Mar 2025 21:43		RC/JU	Ok,M
11	PB167078BL	PB167078BL	BP024169.D	12 Mar 2025 22:24		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP040825

Review By	anahy	Review On	4/9/2025 10:38:32 AM
Supervise By	Jagrut	Supervise On	4/9/2025 4:51:33 PM
SubDirectory	BP040825	HP Acquire Method	BNA_P
		HP Processing Method	bp031225

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12658,10ul/1000ul sample
ICV/I.BLK	SP6686
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP024210.D	08 Apr 2025 13:02		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BP024211.D	08 Apr 2025 13:43		RC/JU	Ok,M
3	PB167474BL	PB167474BL	BP024212.D	08 Apr 2025 14:24	Use only for Q1730	RC/JU	Ok
4	PB167474BS	PB167474BS	BP024213.D	08 Apr 2025 15:04	Use only for Q1730	RC/JU	Ok
5	Q1730-01	OU4-VSL-15-040325	BP024214.D	08 Apr 2025 15:45		RC/JU	Ok
6	Q1730-01MS	OU4-VSL-15-040325M	BP024215.D	08 Apr 2025 16:26	Internal Standard Fail, CCC fail low for com# 69	RC/JU	Not Ok
7	Q1730-01MSD	OU4-VSL-15-040325M	BP024216.D	08 Apr 2025 17:07	CCC fail low side for com#69	RC/JU	Not Ok
8	Q1730-03	OU4-VSL-16-040325	BP024217.D	08 Apr 2025 17:47		RC/JU	Ok
9	Q1730-05	OU4-VSL-17-040325	BP024218.D	08 Apr 2025 18:28		RC/JU	Ok
10	Q1730-07	OU4-PCS-TC-21-040325	BP024219.D	08 Apr 2025 19:09		RC/JU	Ok
11	Q1730-09	OU4-PCS-TC-22-040325	BP024220.D	08 Apr 2025 19:50		RC/JU	Ok
12	Q1730-11	OU4-PCS-TC-23-040325	BP024221.D	08 Apr 2025 20:31		RC/JU	Ok
13	Q1730-13	OU4-PCS-TC-24-040325	BP024222.D	08 Apr 2025 21:11		RC/JU	Ok
14	Q1730-15	OU4-PCS-TC-25-040325	BP024223.D	08 Apr 2025 21:52		RC/JU	Ok
15	Q1730-17	OU4-PCS-TC-26-040325	BP024224.D	08 Apr 2025 22:33		RC/JU	Ok
16	Q1730-19	OU4-CF-15-040325	BP024225.D	08 Apr 2025 23:13		RC/JU	Ok
17	SSTDCCC040	SSTDCCC040EC	BP024226.D	08 Apr 2025 23:54		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP040925

Review By	anahy	Review On	4/10/2025 10:38:27 AM
Supervise By	Jagrut	Supervise On	4/10/2025 3:33:16 PM
SubDirectory	BP040925	HP Acquire Method	BNA_P
		HP Processing Method	bp031225

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12658,10ul/1000ul sample
ICV/I.BLK	SP6686
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP024227.D	09 Apr 2025 08:42		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BP024228.D	09 Apr 2025 10:08		RC/JU	Ok,M
3	PB167467TB	PB167467TB	BP024229.D	09 Apr 2025 10:49		RC/JU	Ok
4	PB167474BS	PB167474BS	BP024230.D	09 Apr 2025 11:29	not used.	RC/JU	Not Ok
5	Q1730-01MS	OU4-VSL-15-040325M	BP024231.D	09 Apr 2025 12:15		RC/JU	Ok
6	Q1730-01MSD	OU4-VSL-15-040325M	BP024232.D	09 Apr 2025 12:56		RC/JU	Ok
7	Q1737-02	RT3069	BP024233.D	09 Apr 2025 13:37		RC/JU	Ok
8	Q1712-04	Z-05A	BP024234.D	09 Apr 2025 14:18		RC/JU	Ok
9	Q1712-04MS	Z-05AMS	BP024235.D	09 Apr 2025 14:59		RC/JU	Ok
10	Q1712-04MSD	Z-05AMSD	BP024236.D	09 Apr 2025 15:39		RC/JU	Ok
11	Q1712-08	TT-7	BP024237.D	09 Apr 2025 16:20		RC/JU	Ok
12	Q1732-04	TT-8	BP024238.D	09 Apr 2025 17:01		RC/JU	Ok
13	Q1746-07	EB-2025-4-7	BP024239.D	09 Apr 2025 17:41		RC/JU	Ok
14	Q1746-01	B-149-SB01	BP024240.D	09 Apr 2025 18:22		RC/JU	Ok
15	Q1747-01	ARS20-0008	BP024241.D	09 Apr 2025 19:03		RC/JU	Ok
16	Q1745-01	IB-6A-WC	BP024242.D	09 Apr 2025 19:43		RC/JU	Ok,M
17	SSTDCCC040	SSTDCCC040EC	BP024243.D	09 Apr 2025 20:24		RC/JU	Ok,M

M : Manual Integration