

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: NOBI03

Lab Code: CHEM Case No.: Q1730 SAS No.: Q1730 SDG No.: Q1730

Instrument ID: MSVOA_Y Calibration Date/Time: 04/04/2025 17:44

Lab File ID: VY021792.D Init. Calib. Date(s): 03/27/2025 03/27/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 10:08 12:02

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
Methylene Chloride	50.000	53.51	-7.019997	20