

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY032725

Review By	Mahesh Dadoda	Review On	3/28/2025 3:29:57 PM				
Supervise By	Semsettin Yesilyurt	Supervise On	3/28/2025 3:30:48 PM				
SubDirectory	VY032725	HP Acquire Method	MSVOA_Y	HP Processing Method	82y032725s.m		
STD. NAME		STD REF.#					
Tune/Reschk		VP133499					
Initial Calibration Stds		VP133500,VP133501,VP133502,VP133503,VP133505,VP133508					
CCC		VP133510					
Internal Standard/PEM		VP131783					
ICV/I.BLK		VP133509					
Surrogate Standard							
MS/MSD Standard							
LCS Standard							

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY021655.D	27 Mar 2025 09:23		SY/MD	Ok
2	VSTDICC005	VSTDICC005	VY021656.D	27 Mar 2025 10:08		SY/MD	Ok,M
3	VSTDICC010	VSTDICC010	VY021657.D	27 Mar 2025 10:31		SY/MD	Ok,M
4	VSTDICC020	VSTDICC020	VY021658.D	27 Mar 2025 10:54	Comp.#20 is on Linear Regression	SY/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VY021659.D	27 Mar 2025 11:16		SY/MD	Ok,M
6	VSTDICC100	VSTDICC100	VY021660.D	27 Mar 2025 11:39		SY/MD	Ok,M
7	VSTDICC150	VSTDICC150	VY021661.D	27 Mar 2025 12:02		SY/MD	Ok,M
8	VIBLK	VIBLK	VY021662.D	27 Mar 2025 12:35		SY/MD	Ok
9	VSTDICV050	ICVVY032725	VY021663.D	27 Mar 2025 13:15		SY/MD	Ok,M
10	VY0327SBL01	VY0327SBL01	VY021664.D	27 Mar 2025 13:59		SY/MD	Ok
11	VY0327SBS01	VY0327SBS01	VY021665.D	27 Mar 2025 14:38		SY/MD	Ok,M
12	VY0327SBSD01	VY0327SBSD01	VY021666.D	27 Mar 2025 15:00		SY/MD	Ok,M
13	Q1661-02	351	VY021667.D	27 Mar 2025 15:24	vial-B Not purge	SY/MD	Not Ok
14	Q1645-03RE	TP-4-VOCRE	VY021668.D	27 Mar 2025 15:48	vial-B Internal Standard Fail	SY/MD	Confirms
15	Q1662-03	TP-6-VOC	VY021669.D	27 Mar 2025 16:11	vial-A NOT PURGE	SY/MD	Not Ok
16	Q1662-07	TP-7-VOC	VY021670.D	27 Mar 2025 16:35	vial-A Internal Standard Fail	SY/MD	ReRun
17	Q1661-02	351	VY021671.D	27 Mar 2025 16:58	vial-A NOT PURGE	SY/MD	Not Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY032725

Review By	Mahesh Dadoda	Review On	3/28/2025 3:29:57 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	3/28/2025 3:30:48 PM		
SubDirectory	VY032725	HP Acquire Method	MSVOA_Y	HP Processing Method	82y032725s.m
STD. NAME		STD REF.#			
Tune/Reschk		VP133499			
Initial Calibration Stds		VP133500,VP133501,VP133502,VP133503,VP133505,VP133508			
CCC		VP133510			
Internal Standard/PEM		VP131783			
ICV/I.BLK		VP133509			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	Q1650-02	STOCK-PILE-BIN2	VY021672.D	27 Mar 2025 17:21	vial-A NOT PURGE	SY/MD	Not Ok
19	Q1654-01	RT3407	VY021673.D	27 Mar 2025 17:45	vial-A Internal Standard Fail	SY/MD	Ok
20	Q1656-01	VNJ239	VY021674.D	27 Mar 2025 18:08	vial-A NOT PURGE	SY/MD	Not Ok
21	Q1648-09	TP-7	VY021675.D	27 Mar 2025 18:32	vial-A	SY/MD	Ok
22	Q1648-07	TP-6	VY021676.D	27 Mar 2025 18:55	vial-A	SY/MD	Ok
23	Q1648-05	TP-3	VY021677.D	27 Mar 2025 19:19	vial-A	SY/MD	Ok
24	Q1648-01	TP-4	VY021678.D	27 Mar 2025 19:42	vial-A	SY/MD	Ok
25	Q1648-03	TP-5	VY021679.D	27 Mar 2025 20:05	vial-A	SY/MD	Ok
26	VSTDCCC050	VSTDCCC050EC	VY021680.D	27 Mar 2025 20:28		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY040425

Review By	Maresh Dadoda	Review On	4/7/2025 12:13:49 PM
Supervise By	Semsettin Yesilyurt	Supervise On	4/7/2025 1:12:39 PM
SubDirectory	VY040425	HP Acquire Method	MSVOA_Y
		HP Processing Method	82y032725s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133593
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133594,VP133595 VP131783

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY021775.D	04 Apr 2025 08:52		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY021776.D	04 Apr 2025 09:22		SY/MD	Ok,M
3	VY0404SBL01	VY0404SBL01	VY021777.D	04 Apr 2025 09:57		SY/MD	Ok
4	VY0404SBS01	VY0404SBS01	VY021778.D	04 Apr 2025 10:26		SY/MD	Ok,M
5	VY0404SBSD01	VY0404SBSD01	VY021779.D	04 Apr 2025 10:49		SY/MD	Ok,M
6	Q1721-01	AU-05-040325	VY021780.D	04 Apr 2025 12:22	vial-A	SY/MD	Ok
7	Q1730-01	OU4-VSL-15-040325	VY021781.D	04 Apr 2025 13:27	vial-A	SY/MD	Ok
8	Q1730-03	OU4-VSL-16-040325	VY021782.D	04 Apr 2025 13:50	vial-A	SY/MD	Ok
9	Q1730-05	OU4-VSL-17-040325	VY021783.D	04 Apr 2025 14:13	vial-A	SY/MD	Ok
10	Q1730-07	OU4-PCS-TC-21-040325	VY021784.D	04 Apr 2025 14:37	vial-A	SY/MD	Ok
11	Q1730-09	OU4-PCS-TC-22-040325	VY021785.D	04 Apr 2025 15:00	vial-A	SY/MD	Ok
12	Q1730-11	OU4-PCS-TC-23-040325	VY021786.D	04 Apr 2025 15:24	vial-A	SY/MD	Ok,M
13	Q1730-13	OU4-PCS-TC-24-040325	VY021787.D	04 Apr 2025 15:47	vial-A	SY/MD	Ok
14	Q1730-15	OU4-PCS-TC-25-040325	VY021788.D	04 Apr 2025 16:11	vial-A	SY/MD	Ok
15	Q1730-17	OU4-PCS-TC-26-040325	VY021789.D	04 Apr 2025 16:34	Vial-A Internal Standard Fail	SY/MD	ReRun
16	Q1730-19	OU4-CF-15-040325	VY021790.D	04 Apr 2025 16:57	vial-A Internal Standard Fail	SY/MD	ReRun
17	VIBLK	VIBLK	VY021791.D	04 Apr 2025 17:21		SY/MD	Ok
18	VSTDCCC050	VSTDCCC050EC	VY021792.D	04 Apr 2025 17:44		SY/MD	Ok,M

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY040425

Review By	Mahesh Dadoda	Review On	4/7/2025 12:13:49 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	4/7/2025 1:12:39 PM		
SubDirectory	VY040425	HP Acquire Method	MSVOA_Y	HP Processing Method	82y032725s.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP133593			
CCC		VP133594,VP133595			
Internal Standard/PEM		VP131783			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY040725

Review By	Maresh Dadoda	Review On	4/8/2025 2:06:29 PM
Supervise By	Semsettin Yesilyurt	Supervise On	4/8/2025 2:07:32 PM
SubDirectory	VY040725	HP Acquire Method	MSVOA_Y HP Processing Method 82y032725s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133613
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133614,VP133615 VP131783

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY021793.D	07 Apr 2025 08:38		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY021794.D	07 Apr 2025 09:10		SY/MD	Ok,M
3	VY0407SBL01	VY0407SBL01	VY021795.D	07 Apr 2025 13:35		SY/MD	Ok
4	VY0407SBS01	VY0407SBS01	VY021796.D	07 Apr 2025 14:07		SY/MD	Ok,M
5	VY0407SBSD01	VY0407SBSD01	VY021797.D	07 Apr 2025 14:30		SY/MD	Ok,M
6	Q1730-17	OU4-PCS-TC-26-04032	VY021798.D	07 Apr 2025 15:10	vial-B	SY/MD	Ok
7	Q1730-19	OU4-CF-15-040325	VY021799.D	07 Apr 2025 15:32	vial-B	SY/MD	Ok
8	Q1733-01	ETGI-328	VY021800.D	07 Apr 2025 15:56	vial-A	SY/MD	Ok
9	Q1735-01	50660-50661-50662-56	VY021801.D	07 Apr 2025 16:19	vial-A	SY/MD	Ok
10	Q1737-01	RT3069	VY021802.D	07 Apr 2025 16:42	vial-A Internal Standard Fail	SY/MD	ReRun
11	Q1738-01	OK-02-040425	VY021803.D	07 Apr 2025 17:06	vial-A Internal Standard Fail	SY/MD	ReRun
12	VSTDCCC050	VSTDCCC050EC	VY021804.D	07 Apr 2025 17:29		SY/MD	Ok,M

M : Manual Integration