

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: NOBI03
Lab Code: CHEM Case No.: Q1730 SAS No.: Q1730 SDG NO.: Q1730
Lab File ID: VY021794.D Date Analyzed: 04/07/2025
Instrument ID: MSVOA_Y Time Analyzed: 09:10
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	224832	7.71	336981	8.62	298811	11.42
UPPER LIMIT	449664	8.207	673962	9.116	597622	11.92
LOWER LIMIT	112416	7.207	168491	8.116	149406	10.92
EPA SAMPLE NO.						
OU4-PCS-TC-26-040325	178880	7.71	307549	8.62	258255	11.42
OU4-CF-15-040325	155604	7.71	277750	8.62	243254	11.42
VY0407SBL01	188077	7.71	313333	8.62	265951	11.42
VY0407SBS01	202553	7.71	317911	8.62	287816	11.42

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.