

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: NOBI03
 Lab Code: CHEM Case No.: Q1730 SAS No.: Q1730 SDG NO.: Q1730
 Lab File ID: VY021776.D Date Analyzed: 04/04/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:22
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	160179	13.347				
UPPER LIMIT	320358	13.847				
LOWER LIMIT	80089.5	12.847				
EPA SAMPLE NO.						
OU4-VSL-15-040325	228826	13.35				
OU4-VSL-16-040325	230456	13.35				
OU4-VSL-17-040325	238203	13.35				
OU4-PCS-TC-21-040325	236951	13.35				
OU4-PCS-TC-22-040325	235341	13.35				
OU4-PCS-TC-23-040325	250886	13.35				
OU4-PCS-TC-24-040325	247734	13.35				
OU4-PCS-TC-25-040325	247639	13.35				
VY0404SBL01	199436	13.35				
VY0404SBS01	152811	13.35				
VY0404SBSD01	151992	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.