

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: NOBI03
 Lab Code: CHEM Case No.: Q1730 SAS No.: Q1730 SDG NO.: Q1730
 Lab File ID: VY021794.D Date Analyzed: 04/07/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:10
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	153287	13.346				
UPPER LIMIT	306574	13.846				
LOWER LIMIT	76643.5	12.846				
EPA SAMPLE NO.						
OU4-PCS-TC-26-040325	113263	13.35				
OU4-CF-15-040325	107208	13.35				
VY0407SBL01	117542	13.35				
VY0407SBS01	149046	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.