



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

JACOBS Engineering Group, Inc.

Project Name: Former Schlumberger STC PTC Site D3868221

Project # N/A

Chemtech Project # Q1731

Test Name: Metals Group4,Dissolved ICP-Group2

A. Number of Samples and Date of Receipt:

7 Water samples were received on 04/03/2025.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested: Alkalinity, Anions Group1, Dissolved ICP-Group2, Dissolved Metals Group3, Metals Group4, SVOC-SIMGroup1, TDS and VOCMS Group3. This data package contains results for Metals Group4,Dissolved ICP-Group2.

C. Analytical Techniques:

The analysis of Dissolved ICP-Group2,Metals Group4 was based on method 6020B and digestion based on method 3010 (waters).

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Blank Spike met requirements for all samples.

The Duplicate (MW-18B-56-040225DUP) analysis met criteria for all samples except for Manganese due to sample matrix interference.

The Matrix Spike (MW-18B-56-040225MS) analysis met criteria for all samples except for Arsenic and Potassium due to Chemical Interference during Digestion Process.

The Matrix Spike Duplicate (MW-18B-56-040225MSD) analysis met criteria for all samples except for Arsenic due to Chemical Interference during Digestion Process.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

The Serial Dilution met the acceptable requirements.

E. Additional Comments: Sample Q1731-05 analyse as Total Metal and Q1731-06 analyse as Dissolve Metal.

Collision cell is being used to remove potential interferences. The analytes Na, Mg, Al, K, V, Cr, Mn, Fe, Co, Ni, Cu, Zn, As are being analyzed with collision cell and analytes Be, B, Ca, Ti, Se, Sr, Zr, Mo, Ag, Cd, Sn, Sb, Ba, Tl, Pb, U are being analyzed with Non-Collision Cell. Helium gas is used for the Collision Cell analysis.



I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____