

Duplicate Sample Summary

Client: JACOBS Engineering Group, Inc. Project: Former Schlumberger STC PTC Site D3868221 Client ID: MOD-FOR-LRSAMSD	SDG No.: Q1731 Sample ID: Q1716-01 Percent Solids for Spike Sample: 0
---	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Fluoride	mg/L	+/-15	2.40		2.40		1	0		04/04/2025
Chloride	mg/L	+/-15	410	OR	409	OR	1	0		04/04/2025
Sulfate	mg/L	+/-15	68.2	OR	68.8	OR	1	1		04/04/2025
Bromide	mg/L	+/-15	10.4		10.8		1	4		04/04/2025
Nitrate	mg/L	+/-15	2.70		2.80		1	4		04/04/2025
Orthophosphate as P	mg/L	+/-15	4.90		5.10		1	4		04/04/2025
Nitrite	mg/L	+/-15	2.90		3.10		1	7		04/04/2025

Duplicate Sample Summary

Client:	JACOBS Engineering Group, Inc.	SDG No.:	Q1731
Project:	Former Schlumberger STC PTC Site D3868221	Sample ID:	Q1731-05
Client ID:	EB01-040325DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TDS	mg/L	+/-5	1.00	J	1.00	J	1	0		04/04/2025