

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1731 SAS No.: Q1731 SDG NO.: Q1731

EPA Sample No.: SSTDICCC0.4 Date Analyzed: 03/10/2025

Lab File ID: BN036559.D Time Analyzed: 12:54

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	2207	7.724	5091	10.51	3026	14.37
UPPER LIMIT	4414	8.224	10182	11.009	6052	14.866
LOWER LIMIT	1103.5	7.224	2545.5	10.009	1513	13.866
EPA SAMPLE NO.						
01 RMW-03B-90-040325DL	1423	7.70	3367	10.49	2092	14.33
02 PB167468BL	2702	7.69	6445	10.48	3394	14.33
03 PB167468BS	1969	7.70	4826	10.48	2640	14.33
04 PB167468BSD	1816	7.70	4421	10.48	2443	14.33
05 RMW-01B-82-040325	1599	7.70	3869	10.48	2334	14.33
06 RMW-04B-91-040325	1693	7.70	4270	10.48	2557	14.33
07 RMW-01B-82-040325-FD	1722	7.70	4117	10.49	2520	14.33
08 RMW-03B-90-040325	1670	7.70	4185	10.48	2430	14.33
09 EB01-040325	1624	7.70	4101	10.48	2430	14.33

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.