

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1731 SAS No.: Q1731 SDG NO.: Q1731

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 04/04/2025

Lab File ID: BN036837.D Time Analyzed: 13:42

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	2261	7.688	5761	10.48	3391	14.33
UPPER LIMIT	4522	8.188	11522	10.977	6782	14.834
LOWER LIMIT	1130.5	7.188	2880.5	9.977	1695.5	13.834
EPA SAMPLE NO.						
01 RMW-01B-82-040325	1599	7.70	3869	10.48	2334	14.33
02 PB167468BS	1969	7.70	4826	10.48	2640	14.33
03 RMW-04B-91-040325	1693	7.70	4270	10.48	2557	14.33
04 PB167468BSD	1816	7.70	4421	10.48	2443	14.33
05 RMW-01B-82-040325-FD	1722	7.70	4117	10.49	2520	14.33
06 RMW-03B-90-040325	1670	7.70	4185	10.48	2430	14.33
07 EB01-040325	1624	7.70	4101	10.48	2430	14.33

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.