

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1731 SAS No.: Q1731 SDG NO.: Q1731

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 04/07/2025

Lab File ID: BN036847.D Time Analyzed: 09:10

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	1549	7.688	3923	10.48	2335	14.33
UPPER LIMIT	3098	8.188	7846	10.977	4670	14.834
LOWER LIMIT	774.5	7.188	1961.5	9.977	1167.5	13.834
EPA SAMPLE NO.						
01 PB167468BL	1482	7.70	3397	10.49	1996	14.35
02 RMW-03B-90-040325DL	1423	7.70	3367	10.49	2092	14.33

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.