284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

DATA PACKAGE

VOLATILE ORGANICS

PROJECT NAME : FORMER SCHLUMBERGER STC PTC SITE D3868221

JACOBS ENGINEERING GROUP, INC.

412 Mt. Kemble Ave

Downtown Building

Morristown, NJ - 07960

Phone No: 9732670555

ORDER ID : Q1731

ATTENTION : John Ynfante

Laboratory Certification ID # 20012

Alliance TECHNICAL GROUP

Table Of Contents for Q1731

1) VOLATILES DATA 2

2) Signature Page 4

3) Case Narrative 5

4) Qualifier Page 7

5) Conformance/Non Conformance 8

6) QA Checklist 10

7) Chronicle 11

8) Hit Summary 12

9) QC Data Summary For VOCMS Group3 13

- 9.1) Deuterated Monitoring Compound Summary 14
- 9.2) LCS/LCSD Summary 15
- 9.3) Method Blank Summary 18
- 9.4) GS/MS Tune Summary 20
- 9.5) Internal Standard Area and RT Summary 23

10) Sample Data 27

- 10.1) RMW-01B-82-040325 28
- 10.2) RMW-04B-91-040325 36
- 10.3) RMW-01B-82-040325-FD 43
- 10.4) RMW-03B-90-040325 50
- 10.5) EB01-040325 60
- 10.6) TB01-040325 68

11) Calibration Data Summary 75

- 11.1) Initial Calibration Data 76
  - 11.1.1) VX040225 76
- 11.2) Continued Calibration Data 445
  - 11.2.1) VX045586.D 445
  - 11.2.2) VX045614.D 504

12) QC Sample Data 563

- 12.1) Tune Raw Data 564
- 12.2) Method Blank Data 567
- 12.3) LCS Data 581
- 12.4) LCSD Data 687

13) Manual Integration 740

14) Analytical Runlogs 744

15) Standard Prep Logs 755

16) Shipping Document 857
16.1) Chain Of Custody 858
16.2) Lab Certificate 859
16.3) Internal COC 860

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Cover Page

Order ID : Q1731

Project ID : Former Schlumberger STC PTC Site D3868221

Client : JACOBS Engineering Group, Inc.

Lab Sample Number	Client Sample Number
Q1731-01	RMW-01B-82-040325
Q1731-02	RMW-04B-91-040325
Q1731-03	RMW-01B-82-040325-FD
Q1731-04	RMW-03B-90-040325
Q1731-05	EB01-040325
Q1731-06	EB01-040325
Q1731-07	TB01-040325

I certify that the data package is in compliance with the terms and conditions of the contract,both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 4/12/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

JACOBS Engineering Group, Inc.
Project Name: Former Schlumberger STC PTC Site D3868221
Project # N/A
Chemtech Project # Q1731
Test Name: VOCMS Group3

A. Number of Samples and Date of Receipt:
7 Water samples were received on 04/03/2025.

B. Parameters
According to the Chain of Custody document, the following analyses were requested: Alkalinity, Anions Group1, Dissolved ICP-Group2, Dissolved Metals Group3, Metals Group4, SVOC-SIMGroup1, TDS and VOCMS Group3. This data package contains results for VOCMS Group3.

C. Analytical Techniques:
The analysis perform ed on instrument MSVOA_X were done using GC colu mn DB-624UI 20m 0.18mm 1.0 um. Cat#121-1324UIThe analysis of VOCMS Group3 was based on method 8260D.

D. QA/ QC Samples:
The Holding Times were met for all analysis.
The Surrogate recoveries met the acceptable criteria.
The Internal Standards Areas met the acceptable requirements.
The Retention Times were acceptable for all samples.

The RPD met criteria.
The Blank Spike met requirements for all samples.
The Blank Spike Duplicate met requirements for all samples.
The Blank analysis did not indicate the presence of lab contamination.
The Initial Calibration met the requirements.
The Continuous Calibration met the requirements.
The Tuning criteria met requirements.

Samples RMW-03B-90-040325 was diluted due to past history of this sample containing high amounts of compounds cis-1,2-Dichloroethene and Trichloroethene.

E. Additional Comments:
Samples for MS/MSD for VOC analysis were not provided with this set of samples. The Blank Spike Duplicate is reported with the data.

284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:
Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

284 Sheffield Street, Mountainside NJ 07092 Tel. 908-789-8900 Fax: 908-789-8922

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following " Results Qualifiers" are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. "10 U". This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This is flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as "12 B".
E	Indicates the analyte 's concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a "P".
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: Q1731 MATRIX: Water

METHOD: 8260D

		NA	NO	YES
1.	Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)			✓
2.	GC/MS Tuning Specifications BFB Meet Criteria (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)			✓
3.	GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 Series.			✓
4.	GC/MS Calibration - Initial Calibration performed before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series.			✓
5.	GC/MS Calibration Requirements. The Initial Calibration met the requirements. The Continuous Calibration met the requirements.			✓
6.	Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
7.	Surrogate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable ranges.			✓
8.	Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.			✓
9.	Internal Standard Area/Retention Time Shift Meet Criteria Comments:			✓
10.	Analysis Holding Time Met If not met, list number of days exceeded for each sample:			✓

ADDITIONAL COMMENTS:

Samples RMW-03B-90-040325 was diluted due to past history of this sample containing high amounts of compounds cis-1,2-Dichloroethene and Trichloroethene.

Samples for MS/MSD for VOC analysis were not provided with this set of samples. The Blank Spike Duplicate is reported with the data.

GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY (CONTINUED)

NA NO YES

Trip Blank was not provided with this set of samples.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is >20% for the Initial Calibration curve for SW-846 analysis.

_____ _____

QA REVIEW Date

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q1731

	Completed
For thorough review, the report must have the following:	
GENERAL:	
Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)	✓
Check chain-of-custody for proper relinquish/return of samples	✓
Is the chain of custody signed and complete	✓
Check internal chain-of-custody for proper relinquish/return of samples /sample extracts	✓
Collect information for each project id from server. Were all requirements followed	✓
COVER PAGE:	
Do numbers of samples correspond to the number of samples in the Chain of Custody on login page	✓
Do lab numbers and client Ids on cover page agree with the Chain of Custody	✓
CHAIN OF CUSTODY:	
Do requested analyses on Chain of Custody agree with form I results	✓
Do requested analyses on Chain of Custody agree with the log-in page	✓
Were the correct method log-in for analysis according to the Analytical Request and Chain of Castody	✓
Were the samples received within hold time	✓
Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle	✓
ANALYTICAL:	
Was method requirement followed?	✓
Was client requirement followed?	✓
Does the case narrative summarize all QC failure?	✓
All runlogs and manual integration are reviewed for requirements	✓
All manual calculations and /or hand notations verified	✓

QA Review Signature: SOHIL JODHANI **Date:** 04/12/2025

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

LAB CHRONICLE

OrderID:	Q1731	**OrderDate:**	4/4/2025 10:52:00 AM
Client:	JACOBS Engineering Group, Inc.	**Project:**	Former Schlumberger STC PTC Site D3868221
Contact:	John Ynfante	**Location:**	L31,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1731-01	**RMW-01B-82-040325**	**Water**			**04/03/25**			**04/03/25**
			VOCMS Group3	8260-Low			04/07/25	
Q1731-02	**RMW-04B-91-040325**	**Water**			**04/03/25**			**04/03/25**
			VOCMS Group3	8260-Low			04/07/25	
Q1731-03	**RMW-01B-82-040325-FD**	**Water**			**04/03/25**			**04/03/25**
			VOCMS Group3	8260-Low			04/07/25	
Q1731-04	**RMW-03B-90-040325**	**Water**			**04/03/25**			**04/03/25**
			VOCMS Group3	8260-Low			04/07/25	
Q1731-05	**EB01-040325**	**Water**			**04/03/25**			**04/03/25**
			VOCMS Group3	8260-Low			04/04/25	
Q1731-07	**TB01-040325**	**Water**			**04/03/25**			**04/03/25**
			VOCMS Group3	8260-Low			04/04/25	

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q1731

Client: JACOBS Engineering Group, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID:	**RMW-01B-82-040325**							
Q1731-01	RMW-01B-82-0403	Water	Trichloroethene	0.36	J	0.090	1.00	ug/L
			Total Voc :	0.36				
			Total Concentration:	0.36				
Client ID:	**RMW-03B-90-040325**							
Q1731-04	RMW-03B-90-0403	Water	1,1-Dichloroethene	37.7	J	9.20	40.0	ug/L
Q1731-04	RMW-03B-90-0403	Water	cis-1,2-Dichloroethene	1200		7.60	40.0	ug/L
Q1731-04	RMW-03B-90-0403	Water	Trichloroethene	3400		3.70	40.0	ug/L
Q1731-04	RMW-03B-90-0403	Water	Tetrachloroethene	30.8	J	9.20	40.0	ug/L
			Total Voc :	4670				
			Total Concentration:	4670				

QC SUMMARY

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Surrogate Summary

SDG No.: Q1731

Client: JACOBS Engineering Group, Inc.

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits Low	Limits High
Q1731-01	RMW-01B-82-040325	1,2-Dichloroethane-d4	50	54.6	109	70 (74)	130 (125)
		Dibromofluoromethane	50	52.4	105	70 (75)	130 (124)
		Toluene-d8	50	51.0	102	70 (86)	130 (113)
		4-Bromofluorobenzene	50	49.3	99	70 (77)	130 (121)
Q1731-02	RMW-04B-91-040325	1,2-Dichloroethane-d4	50	55.6	111	70 (74)	130 (125)
		Dibromofluoromethane	50	52.0	104	70 (75)	130 (124)
		Toluene-d8	50	51.4	103	70 (86)	130 (113)
		4-Bromofluorobenzene	50	50.0	100	70 (77)	130 (121)
Q1731-03	RMW-01B-82-040325-FD	1,2-Dichloroethane-d4	50	55.3	111	70 (74)	130 (125)
		Dibromofluoromethane	50	52.2	104	70 (75)	130 (124)
		Toluene-d8	50	51.8	104	70 (86)	130 (113)
		4-Bromofluorobenzene	50	50.4	101	70 (77)	130 (121)
Q1731-04	RMW-03B-90-040325	1,2-Dichloroethane-d4	50	53.6	107	70 (74)	130 (125)
		Dibromofluoromethane	50	51.9	104	70 (75)	130 (124)
		Toluene-d8	50	51.5	103	70 (86)	130 (113)
		4-Bromofluorobenzene	50	52.1	104	70 (77)	130 (121)
Q1731-05	EB01-040325	1,2-Dichloroethane-d4	50	55.3	111	70 (74)	130 (125)
		Dibromofluoromethane	50	51.6	103	70 (75)	130 (124)
		Toluene-d8	50	50.8	102	70 (86)	130 (113)
		4-Bromofluorobenzene	50	51.5	103	70 (77)	130 (121)
Q1731-07	TB01-040325	1,2-Dichloroethane-d4	50	54.5	109	70 (74)	130 (125)
		Dibromofluoromethane	50	51.0	102	70 (75)	130 (124)
		Toluene-d8	50	50.5	101	70 (86)	130 (113)
		4-Bromofluorobenzene	50	48.4	97	70 (77)	130 (121)
VX0404WBL01	VX0404WBL01	1,2-Dichloroethane-d4	50	54.8	110	70 (74)	130 (125)
		Dibromofluoromethane	50	52.0	104	70 (75)	130 (124)
		Toluene-d8	50	50.8	102	70 (86)	130 (113)
		4-Bromofluorobenzene	50	49.5	99	70 (77)	130 (121)
VX0404WBS01	VX0404WBS01	1,2-Dichloroethane-d4	50	52.9	106	70 (74)	130 (125)
		Dibromofluoromethane	50	53.3	107	70 (75)	130 (124)
		Toluene-d8	50	53.2	106	70 (86)	130 (113)
		4-Bromofluorobenzene	50	53.2	106	70 (77)	130 (121)
VX0404WBSD0	VX0404WBSD01	1,2-Dichloroethane-d4	50	53.3	107	70 (74)	130 (125)
		Dibromofluoromethane	50	52.6	105	70 (75)	130 (124)
		Toluene-d8	50	53.1	106	70 (86)	130 (113)
		4-Bromofluorobenzene	50	53.9	108	70 (77)	130 (121)
VX0407WBL01	VX0407WBL01	1,2-Dichloroethane-d4	50	56.0	112	70 (74)	130 (125)
		Dibromofluoromethane	50	52.1	104	70 (75)	130 (124)
		Toluene-d8	50	51.0	102	70 (86)	130 (113)
		4-Bromofluorobenzene	50	51.6	103	70 (77)	130 (121)
VX0407WBS01	VX0407WBS01	1,2-Dichloroethane-d4	50	50.8	102	70 (74)	130 (125)
		Dibromofluoromethane	50	50.0	100	70 (75)	130 (124)
		Toluene-d8	50	49.8	99	70 (86)	130 (113)
		4-Bromofluorobenzene	50	50.4	101	70 (77)	130 (121)

( ) = LABORATORY INHOUSE LIMIT

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1731

Client: JACOBS Engineering Group, Inc.

Analytical Method: SW8260-Low **Datafile :** VX045589.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Limits Low	Limits High	RPD
VX0404WBS01	Vinyl chloride	20	19.6	ug/L	98			70 (65)	130 (117)	
	1,1-Dichloroethene	20	19.4	ug/L	97			70 (74)	130 (110)	
	1,1-Dichloroethane	20	19.5	ug/L	98			70 (78)	130 (112)	
	cis-1,2-Dichloroethene	20	19.2	ug/L	96			70 (77)	130 (110)	
	1,1,1-Trichloroethane	20	19.7	ug/L	99			70 (80)	130 (108)	
	Benzene	20	19.9	ug/L	100			70 (82)	130 (109)	
	1,2-Dichloroethane	20	20.4	ug/L	102			70 (80)	130 (115)	
	Trichloroethene	20	19.2	ug/L	96			70 (77)	130 (113)	
	1,1,2-Trichloroethane	20	20.5	ug/L	103			70 (83)	130 (112)	
	Tetrachloroethene	20	19.6	ug/L	98			70 (67)	130 (123)	

( ) = LABORATORY INHOUSE LIMIT

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1731

Client: JACOBS Engineering Group, Inc.

Analytical Method: SW8260-Low **Datafile :** VX045590.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Limits Low	Limits High	Limits RPD
VX0404WBSD01	Vinyl chloride	20	19.2	ug/L	96	2		70 (65)	130 (117)	20 (20)
	1,1-Dichloroethene	20	19.0	ug/L	95	2		70 (74)	130 (110)	20 (20)
	1,1-Dichloroethane	20	19.3	ug/L	97	1		70 (78)	130 (112)	20 (20)
	cis-1,2-Dichloroethene	20	19.2	ug/L	96	0		70 (77)	130 (110)	20 (20)
	1,1,1-Trichloroethane	20	19.6	ug/L	98	1		70 (80)	130 (108)	20 (20)
	Benzene	20	19.6	ug/L	98	2		70 (82)	130 (109)	20 (20)
	1,2-Dichloroethane	20	20.7	ug/L	104	2		70 (80)	130 (115)	20 (20)
	Trichloroethene	20	19.7	ug/L	99	3		70 (77)	130 (113)	20 (20)
	1,1,2-Trichloroethane	20	20.6	ug/L	103	0		70 (83)	130 (112)	20 (20)
	Tetrachloroethene	20	19.4	ug/L	97	1		70 (67)	130 (123)	20 (20)

( ) = LABORATORY INHOUSE LIMIT

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1731

Client: JACOBS Engineering Group, Inc.

Analytical Method: SW8260-Low **Datafile :** VX045617.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Limits Low	Limits High	RPD
VX0407WBS01	Vinyl chloride	20	18.0	ug/L	90			70 (65)	130 (117)	
	1,1-Dichloroethene	20	17.5	ug/L	88			70 (74)	130 (110)	
	1,1-Dichloroethane	20	17.8	ug/L	89			70 (78)	130 (112)	
	cis-1,2-Dichloroethene	20	17.5	ug/L	88			70 (77)	130 (110)	
	1,1,1-Trichloroethane	20	18.0	ug/L	90			70 (80)	130 (108)	
	Benzene	20	18.0	ug/L	90			70 (82)	130 (109)	
	1,2-Dichloroethane	20	18.7	ug/L	94			70 (80)	130 (115)	
	Trichloroethene	20	18.0	ug/L	90			70 (77)	130 (113)	
	1,1,2-Trichloroethane	20	18.2	ug/L	91			70 (83)	130 (112)	
	Tetrachloroethene	20	19.3	ug/L	97			70 (67)	130 (123)	

( ) = LABORATORY INHOUSE LIMIT

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VX0404WBL01

Lab Name: CHEMTECH Contract: JACO05

Lab Code: CHEM Case No.: Q1731 SAS No.: Q1731 SDG NO.: Q1731

Lab File ID: VX045588.D Lab Sample ID: VX0404WBL01

Date Analyzed: 04/04/2025 Time Analyzed: 10:51

GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

Instrument ID: MSVOA_X

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VX0404WBS01	VX0404WBS01	VX045589.D	04/04/2025
VX0404WBSD01	VX0404WBSD01	VX045590.D	04/04/2025
TB01-040325	Q1731-07	VX045599.D	04/04/2025
EB01-040325	Q1731-05	VX045600.D	04/04/2025

COMMENTS:

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VX0407WBL01

Lab Name: CHEMTECH Contract: JACO05

Lab Code: CHEM Case No.: Q1731 SAS No.: Q1731 SDG NO.: Q1731

Lab File ID: VX045616.D Lab Sample ID: VX0407WBL01

Date Analyzed: 04/07/2025 Time Analyzed: 10:11

GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

Instrument ID: MSVOA_X

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VX0407WBS01	VX0407WBS01	VX045617.D	04/07/2025
RMW-03B-90-040325	Q1731-04	VX045619.D	04/07/2025
RMW-01B-82-040325	Q1731-01	VX045620.D	04/07/2025
RMW-04B-91-040325	Q1731-02	VX045621.D	04/07/2025
RMW-01B-82-040325-FD	Q1731-03	VX045622.D	04/07/2025

COMMENTS:

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: JACO05

Lab Code: CHEM Case No.: Q1731 SAS No.: Q1731 SDG NO.: Q1731

Lab File ID: VX045524.D BFB Injection Date: 04/01/2025

Instrument ID: MSVOA_X BFB Injection Time: 16:15

GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	23.6
75	30.0 - 60.0% of mass 95	58.2
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.3
173	Less than 2.0% of mass 174	0.8 (1.2) 1
174	50.0 - 100.0% of mass 95	66.8
175	5.0 - 9.0% of mass 174	5.2 (7.8) 1
176	95.0 - 101.0% of mass 174	65.3 (97.8) 1
177	5.0 - 9.0% of mass 176	4.4 (6.8) 2

1-Value is % mass 174 2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDICC001	VSTDICC001	VX045525.D	04/01/2025	17:06
VSTDICC005	VSTDICC005	VX045526.D	04/01/2025	17:29
VSTDICC020	VSTDICC020	VX045527.D	04/01/2025	17:52
VSTDICCC050	VSTDICCC050	VX045528.D	04/01/2025	18:15
VSTDICC100	VSTDICC100	VX045529.D	04/01/2025	18:38
VSTDICC150	VSTDICC150	VX045530.D	04/01/2025	19:02

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: JACO05

Lab Code: CHEM Case No.: Q1731 SAS No.: Q1731 SDG NO.: Q1731

Lab File ID: VX045585.D BFB Injection Date: 04/04/2025

Instrument ID: MSVOA_X BFB Injection Time: 09:31

GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	25
75	30.0 - 60.0% of mass 95	57.7
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	7.2
173	Less than 2.0% of mass 174	0.9 (1.4) 1
174	50.0 - 100.0% of mass 95	68.1
175	5.0 - 9.0% of mass 174	5.7 (8.3) 1
176	95.0 - 101.0% of mass 174	65.8 (96.6) 1
177	5.0 - 9.0% of mass 176	4.9 (7.5) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX045586.D	04/04/2025	10:00
VX0404WBL01	VX0404WBL01	VX045588.D	04/04/2025	10:51
VX0404WBS01	VX0404WBS01	VX045589.D	04/04/2025	11:14
VX0404WBSD01	VX0404WBSD01	VX045590.D	04/04/2025	11:40
TB01-040325	Q1731-07	VX045599.D	04/04/2025	15:10
EB01-040325	Q1731-05	VX045600.D	04/04/2025	15:33

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: JACO05

Lab Code: CHEM Case No.: Q1731 SAS No.: Q1731 SDG NO.: Q1731

Lab File ID: VX045613.D BFB Injection Date: 04/07/2025

Instrument ID: MSVOA_X BFB Injection Time: 08:34

GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	23.4
75	30.0 - 60.0% of mass 95	58.2
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	7.1
173	Less than 2.0% of mass 174	0.8 (1.2) 1
174	50.0 - 100.0% of mass 95	67.9
175	5.0 - 9.0% of mass 174	5.4 (8) 1
176	95.0 - 101.0% of mass 174	67.3 (99.2) 1
177	5.0 - 9.0% of mass 176	4.1 (6.1) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX045614.D	04/07/2025	09:17
VX0407WBL01	VX0407WBL01	VX045616.D	04/07/2025	10:11
VX0407WBS01	VX0407WBS01	VX045617.D	04/07/2025	10:34
RMW-03B-90-040325	Q1731-04	VX045619.D	04/07/2025	11:23
RMW-01B-82-040325	Q1731-01	VX045620.D	04/07/2025	11:46
RMW-04B-91-040325	Q1731-02	VX045621.D	04/07/2025	12:09
RMW-01B-82-040325-FD	Q1731-03	VX045622.D	04/07/2025	12:33

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05

Lab Code: CHEM Case No.: Q1731 SAS No.: Q1731 SDG NO.: Q1731

Lab File ID: VX045586.D Date Analyzed: 04/04/2025

Instrument ID: MSVOA_X Time Analyzed: 10:00

GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	96707	5.54	163287	6.75	142048	10.05
UPPER LIMIT	193414	6.038	326574	7.251	284096	10.549
LOWER LIMIT	48353.5	5.038	81643.5	6.251	71024	9.549
EPA SAMPLE NO.						
EB01-040325	64123	5.54	127534	6.76	119214	10.06
TB01-040325	62156	5.55	122509	6.76	111172	10.05
VX0404WBL01	66301	5.54	128686	6.76	118182	10.06
VX0404WBS01	95127	5.55	168453	6.76	151744	10.05
VX0404WBSD01	92597	5.54	161617	6.76	145135	10.05

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05

Lab Code: CHEM Case No.: Q1731 SAS No.: Q1731 SDG NO.: Q1731

Lab File ID: VX045586.D Date Analyzed: 04/04/2025

Instrument ID: MSVOA_X Time Analyzed: 10:00

GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	67379	12.018				
UPPER LIMIT	134758	12.518				
LOWER LIMIT	33689.5	11.518				
EPA SAMPLE NO.						
EB01-040325	49367	12.02				
TB01-040325	43047	12.02				
VX0404WBL01	46025	12.02				
VX0404WBS01	66584	12.02				
VX0404WBSD01	65264	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05

Lab Code: CHEM Case No.: Q1731 SAS No.: Q1731 SDG NO.: Q1731

Lab File ID: VX045614.D Date Analyzed: 04/07/2025

Instrument ID: MSVOA_X Time Analyzed: 09:17

GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	104243	5.54	177492	6.75	157116	10.05
UPPER LIMIT	208486	6.038	354984	7.251	314232	10.549
LOWER LIMIT	52121.5	5.038	88746	6.251	78558	9.549
EPA SAMPLE NO.						
RMW-01B-82-040325	63367	5.54	122757	6.76	113697	10.05
RMW-04B-91-040325	61435	5.55	120707	6.76	111174	10.06
RMW-01B-82-040325-FD	62969	5.55	123324	6.76	115726	10.06
RMW-03B-90-040325	64989	5.55	126519	6.76	118849	10.05
VX0407WBL01	67505	5.54	132975	6.76	123601	10.05
VX0407WBS01	89694	5.54	158717	6.76	138981	10.05

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05

Lab Code: CHEM Case No.: Q1731 SAS No.: Q1731 SDG NO.: Q1731

Lab File ID: VX045614.D Date Analyzed: 04/07/2025

Instrument ID: MSVOA_X Time Analyzed: 09:17

GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	73594	12.018				
UPPER LIMIT	147188	12.518				
LOWER LIMIT	36797	11.518				
EPA SAMPLE NO.						
RMW-01B-82-040325	42386	12.02				
RMW-04B-91-040325	42372	12.02				
RMW-01B-82-040325-FD	47624	12.02				
RMW-03B-90-040325	48503	12.02				
VX0407WBL01	49926	12.02				
VX0407WBS01	64861	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

SAMPLE DATA

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	04/03/25
Project:	Former Schlumberger STC PTC Site D3868221	Date Received:	04/03/25
Client Sample ID:	RMW-01B-82-040325	SDG No.:	Q1731
Lab Sample ID:	Q1731-01	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group3
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045620.D	1		04/07/25 11:46	VX040725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	0.26	U	0.26	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.23	U	0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.19	U	0.19	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.20	U	0.20	1.00	ug/L
71-43-2	Benzene	0.15	U	0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.22	U	0.22	1.00	ug/L
79-01-6	Trichloroethene	0.36	J	0.090	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	1.00	ug/L
127-18-4	Tetrachloroethene	0.23	U	0.23	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	54.6		70 (74) - 130 (125)	109%	SPK: 50
1868-53-7	Dibromofluoromethane	52.4		70 (75) - 130 (124)	105%	SPK: 50
2037-26-5	Toluene-d8	51.1		70 (86) - 130 (113)	102%	SPK: 50
460-00-4	4-Bromofluorobenzene	49.3		70 (77) - 130 (121)	99%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	63400	5.544			
540-36-3	1,4-Difluorobenzene	123000	6.757			
3114-55-4	Chlorobenzene-d5	114000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	42400	12.018			

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements
J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040725\
Data File : VX045620.D
Acq On : 07 Apr 2025 11:46
Operator : JC/MD
Sample : Q1731-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Quant Time: Apr 07 14:39:29 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
RMW-01B-82-040325

Manual Integrations APPROVED

Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/08/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	5.544	168	63367	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	122757	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	113697	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	42386	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63235	54.569	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 109.140%			
35) Dibromofluoromethane	5.379	113	45610	52.367	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 104.740%			
50) Toluene-d8	8.647	98	155197	51.052	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 102.100%			
62) 4-Bromofluorobenzene	11.079	95	54545	49.259	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 98.520%			
Target Compounds						Qvalue
16) Acetone	2.386	43	1012	2.127	ug/l	100
44) Trichloroethene	7.153	130	309m	0.362	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040725\
Data File : VX045620.D
Acq On : 07 Apr 2025 11:46
Operator : JC/MD
Sample : Q1731-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1


Quant Time: Apr 07 14:39:29 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
RMW-01B-82-040325

Manual Integrations APPROVED

Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/08/2025






#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 7
Delta R.T. -0.006 min
Lab File: VX045620.D
Acq: 07 Apr 2025 11:46

Tgt Ion:168 Resp: 6336


Ion	Ratio	Lower	Upper
168	100		
99	71.5	52.3	78.5



Instrument :
MSVOA_X
ClientSampleId :
RMW-01B-82-040325

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/08/2025




#16
Acetone
Concen: 2.127 ug/l
RT: 2.386 min Scan# 213
Delta R.T. 0.000 min
Lab File: VX045620.D
Acq: 07 Apr 2025 11:46

Tgt Ion: 43 Resp: 1012

Ion	Ratio	Lower	Upper
43	100		
58	27.7	22.3	33.5






#33
1,2-Dichloroethane-d4
Concen: 54.569 ug/l
RT: 5.952 min Scan# 7
Delta R.T. 0.000 min
Lab File: VX045620.D
Acq: 07 Apr 2025 11:46

Tgt Ion: 65 Resp: 6323


Ion	Ratio	Lower	Upper
65	100		
67	50.4	0.0	99.0



Instrument :
MSVOA_X
ClientSampleId :
RMW-01B-82-040325

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/08/2025




#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.757 min Scan# 930
Delta R.T. 0.000 min
Lab File: VX045620.D
Acq: 07 Apr 2025 11:46

Tgt Ion:114 Resp: 122757

Ion	Ratio	Lower	Upper
114	100		
63	24.0	0.0	46.8
88	17.5	0.0	35.4






#35
Dibromofluoromethane
Concen: 52.367 ug/l
RT: 5.379 min Scan# 7
Delta R.T. 0.000 min
Lab File: VX045620.D
Acq: 07 Apr 2025 11:46

Tgt Ion:113 Resp: 4561


Ion	Ratio	Lower	Upper
113	100		
111	102.2	81.8	122.6
192	16.3	13.8	20.6



Instrument :
MSVOA_X
ClientSampleId :
RMW-01B-82-040325

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/08/2025




#44
Trichloroethene
Concen: 0.362 ug/l m
RT: 7.153 min Scan# 995
Delta R.T. 0.031 min
Lab File: VX045620.D
Acq: 07 Apr 2025 11:46

Tgt Ion:130 Resp: 309

Ion	Ratio	Lower	Upper
130	100		
95	98.4	0.0	202.0






#50
Toluene-d8
Concen: 51.052 ug/l
RT: 8.647 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045620.D
Acq: 07 Apr 2025 11:46

Tgt Ion: 98 Resp: 15519


Ion	Ratio	Lower	Upper
98	100		
100	64.1	52.2	78.4



Instrument :
MSVOA_X
ClientSampleId :
RMW-01B-82-040325

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/08/2025




#62
4-Bromofluorobenzene
Concen: 49.259 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX045620.D
Acq: 07 Apr 2025 11:46

Tgt Ion: 95 Resp: 54545

Ion	Ratio	Lower	Upper
95	100		
174	68.7	0.0	135.8
176	65.9	0.0	131.4






#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX045620.D
Acq: 07 Apr 2025 11:46

Tgt Ion:117 Resp: 11369


Ion	Ratio	Lower	Upper
117	100		
82	63.2	49.2	73.8
119	33.1	25.1	37.7



Instrument :
MSVOA_X
ClientSampleId :
RMW-01B-82-040325

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/08/2025



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. 0.000 min
Lab File: VX045620.D
Acq: 07 Apr 2025 11:46

Tgt Ion:152 Resp: 42386

Ion	Ratio	Lower	Upper
152	100		
115	65.3	46.9	140.7
150	158.3	0.0	349.4



Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	04/03/25
Project:	Former Schlumberger STC PTC Site D3868221	Date Received:	04/03/25
Client Sample ID:	RMW-04B-91-040325	SDG No.:	Q1731
Lab Sample ID:	Q1731-02	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group3
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045621.D	1		04/07/25 12:09	VX040725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	0.26	U	0.26	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.23	U	0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.19	U	0.19	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.20	U	0.20	1.00	ug/L
71-43-2	Benzene	0.15	U	0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.22	U	0.22	1.00	ug/L
79-01-6	Trichloroethene	0.090	U	0.090	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	1.00	ug/L
127-18-4	Tetrachloroethene	0.23	U	0.23	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	55.6		70 (74) - 130 (125)	111%	SPK: 50
1868-53-7	Dibromofluoromethane	52.1		70 (75) - 130 (124)	104%	SPK: 50
2037-26-5	Toluene-d8	51.4		70 (86) - 130 (113)	103%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.1		70 (77) - 130 (121)	100%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	61400	5.55			
540-36-3	1,4-Difluorobenzene	121000	6.757			
3114-55-4	Chlorobenzene-d5	111000	10.055			
3855-82-1	1,4-Dichlorobenzene-d4	42400	12.024			

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements
J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040725\
Data File : VX045621.D
Acq On : 07 Apr 2025 12:09
Operator : JC/MD
Sample : Q1731-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RMW-04B-91-040325

Quant Time: Apr 07 14:39:39 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	5.550	168	61435	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	120707	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	111174	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	42372	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	62477	55.610	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 111.220%			
35) Dibromofluoromethane	5.385	113	44578	52.052	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 104.100%			
50) Toluene-d8	8.647	98	153555	51.369	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 102.740%			
62) 4-Bromofluorobenzene	11.079	95	54498	50.052	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 100.100%			


Target Compounds Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040725\
Data File : VX045621.D
Acq On : 07 Apr 2025 12:09
Operator : JC/MD
Sample : Q1731-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RMW-04B-91-040325

Quant Time: Apr 07 14:39:39 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration







#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 7
Delta R.T. -0.000 min
Lab File: VX045621.D
Acq: 07 Apr 2025 12:09

Instrument :
MSVOA_X
ClientSampleId :
RMW-04B-91-040325

Tgt Ion:168 Resp: 61435

Ion	Ratio	Lower	Upper
168	100		
99	68.9	52.3	78.5







#33
1,2-Dichloroethane-d4
Concen: 55.610 ug/l
RT: 5.952 min Scan# 798
Delta R.T. -0.000 min
Lab File: VX045621.D
Acq: 07 Apr 2025 12:09

Tgt Ion: 65 Resp: 62477

Ion	Ratio	Lower	Upper
65	100		
67	49.5	0.0	99.0







#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.757 min Scan# 9
Delta R.T. -0.000 min
Lab File: VX045621.D
Acq: 07 Apr 2025 12:09

Instrument :
MSVOA_X
ClientSampleId :
RMW-04B-91-040325

Tgt Ion:114 Resp: 120707

Ion	Ratio	Lower	Upper
114	100		
63	24.0	0.0	46.8
88	16.4	0.0	35.4







#35
Dibromofluoromethane
Concen: 52.052 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.006 min
Lab File: VX045621.D
Acq: 07 Apr 2025 12:09

Tgt Ion:113 Resp: 44578

Ion	Ratio	Lower	Upper
113	100		
111	104.1	81.8	122.6
192	17.0	13.8	20.6







#50
Toluene-d8
Concen: 51.369 ug/l
RT: 8.647 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045621.D
Acq: 07 Apr 2025 12:09

Instrument :
MSVOA_X
ClientSampleId :
RMW-04B-91-040325

Tgt Ion: 98 Resp: 153555

Ion	Ratio	Lower	Upper
98	100		
100	65.3	52.2	78.4







#62
4-Bromofluorobenzene
Concen: 50.052 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX045621.D
Acq: 07 Apr 2025 12:09

Tgt Ion: 95 Resp: 54498

Ion	Ratio	Lower	Upper
95	100		
174	68.5	0.0	135.8
176	64.1	0.0	131.4







#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 14
Delta R.T. 0.006 min
Lab File: VX045621.D
Acq: 07 Apr 2025 12:09

Instrument :
MSVOA_X
ClientSampleId :
RMW-04B-91-040325

Tgt Ion:117 Resp: 111174

Ion	Ratio	Lower	Upper
117	100		
82	60.5	49.2	73.8
119	32.8	25.1	37.7






#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.006 min
Lab File: VX045621.D
Acq: 07 Apr 2025 12:09

Tgt Ion:152 Resp: 42372

Ion	Ratio	Lower	Upper
152	100		
115	65.7	46.9	140.7
150	160.0	0.0	349.4



Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	04/03/25
Project:	Former Schlumberger STC PTC Site D3868221	Date Received:	04/03/25
Client Sample ID:	RMW-01B-82-040325-FD	SDG No.:	Q1731
Lab Sample ID:	Q1731-03	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group3
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045622.D	1		04/07/25 12:33	VX040725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	0.26	U	0.26	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.23	U	0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.19	U	0.19	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.20	U	0.20	1.00	ug/L
71-43-2	Benzene	0.15	U	0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.22	U	0.22	1.00	ug/L
79-01-6	Trichloroethene	0.090	U	0.090	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	1.00	ug/L
127-18-4	Tetrachloroethene	0.23	U	0.23	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	55.3		70 (74) - 130 (125)	111%	SPK: 50
1868-53-7	Dibromofluoromethane	52.2		70 (75) - 130 (124)	104%	SPK: 50
2037-26-5	Toluene-d8	51.8		70 (86) - 130 (113)	104%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.4		70 (77) - 130 (121)	101%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	63000	5.55			
540-36-3	1,4-Difluorobenzene	123000	6.757			
3114-55-4	Chlorobenzene-d5	116000	10.055			
3855-82-1	1,4-Dichlorobenzene-d4	47600	12.018			

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040725\
Data File : VX045622.D
Acq On : 07 Apr 2025 12:33
Operator : JC/MD
Sample : Q1731-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RMW-01B-82-040325-FD

Quant Time: Apr 07 14:39:48 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	5.550	168	62969	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	123324	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	115726	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	47624	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63709	55.325	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 110.660%			
35) Dibromofluoromethane	5.379	113	45692	52.220	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 104.440%			
50) Toluene-d8	8.647	98	158109	51.770	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 103.540%			
62) 4-Bromofluorobenzene	11.079	95	56045	50.380	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 100.760%			


Target Compounds Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040725\
Data File : VX045622.D
Acq On : 07 Apr 2025 12:33
Operator : JC/MD
Sample : Q1731-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RMW-01B-82-040325-FD

Quant Time: Apr 07 14:39:48 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration







#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 7
Delta R.T. -0.000 min
Lab File: VX045622.D
Acq: 07 Apr 2025 12:33

Instrument :
MSVOA_X
ClientSampleId :
RMW-01B-82-040325-FD

Tgt Ion:168 Resp: 62969

Ion	Ratio	Lower	Upper
168	100		
99	67.3	52.3	78.5







#33
1,2-Dichloroethane-d4
Concen: 55.325 ug/l
RT: 5.952 min Scan# 798
Delta R.T. -0.000 min
Lab File: VX045622.D
Acq: 07 Apr 2025 12:33

Tgt Ion: 65 Resp: 63709

Ion	Ratio	Lower	Upper
65	100		
67	49.9	0.0	99.0







#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.757 min Scan# 9
Delta R.T. -0.000 min
Lab File: VX045622.D
Acq: 07 Apr 2025 12:33

Instrument :
MSVOA_X
ClientSampleId :
RMW-01B-82-040325-FD

Tgt Ion:114 Resp: 123324

Ion	Ratio	Lower	Upper
114	100		
63	23.4	0.0	46.8
88	16.9	0.0	35.4







#35
Dibromofluoromethane
Concen: 52.220 ug/l
RT: 5.379 min Scan# 704
Delta R.T. -0.000 min
Lab File: VX045622.D
Acq: 07 Apr 2025 12:33

Tgt Ion:113 Resp: 45692

Ion	Ratio	Lower	Upper
113	100		
111	103.4	81.8	122.6
192	16.9	13.8	20.6







#50
Toluene-d8
Concen: 51.770 ug/l
RT: 8.647 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045622.D
Acq: 07 Apr 2025 12:33

Instrument :
MSVOA_X
ClientSampleId :
RMW-01B-82-040325-FD

Tgt Ion: 98 Resp: 158109

Ion	Ratio	Lower	Upper
98	100		
100	64.7	52.2	78.4







#62
4-Bromofluorobenzene
Concen: 50.380 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX045622.D
Acq: 07 Apr 2025 12:33

Tgt Ion: 95 Resp: 56045

Ion	Ratio	Lower	Upper
95	100		
174	70.0	0.0	135.8
176	66.8	0.0	131.4







#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 14
Delta R.T. 0.006 min
Lab File: VX045622.D
Acq: 07 Apr 2025 12:33

Instrument :
MSVOA_X
ClientSampleId :
RMW-01B-82-040325-FD

Tgt Ion:117 Resp: 115726

Ion	Ratio	Lower	Upper
117	100		
82	59.4	49.2	73.8
119	30.5	25.1	37.7






#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VX045622.D
Acq: 07 Apr 2025 12:33

Tgt Ion:152 Resp: 47624

Ion	Ratio	Lower	Upper
152	100		
115	66.0	46.9	140.7
150	157.7	0.0	349.4



Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	04/03/25
Project:	Former Schlumberger STC PTC Site D3868221	Date Received:	04/03/25
Client Sample ID:	RMW-03B-90-040325	SDG No.:	Q1731
Lab Sample ID:	Q1731-04	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group3
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045619.D	40		04/07/25 11:23	VX040725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	10.4	U	10.4	40.0	ug/L
75-35-4	1,1-Dichloroethene	37.7	J	9.20	40.0	ug/L
75-34-3	1,1-Dichloroethane	9.20	U	9.20	40.0	ug/L
156-59-2	cis-1,2-Dichloroethene	1200		7.60	40.0	ug/L
71-55-6	1,1,1-Trichloroethane	8.00	U	8.00	40.0	ug/L
71-43-2	Benzene	6.00	U	6.00	40.0	ug/L
107-06-2	1,2-Dichloroethane	8.80	U	8.80	40.0	ug/L
79-01-6	Trichloroethene	3400		3.70	40.0	ug/L
79-00-5	1,1,2-Trichloroethane	8.40	U	8.40	40.0	ug/L
127-18-4	Tetrachloroethene	30.8	J	9.20	40.0	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	53.6		70 (74) - 130 (125)	107%	SPK: 50
1868-53-7	Dibromofluoromethane	51.9		70 (75) - 130 (124)	104%	SPK: 50
2037-26-5	Toluene-d8	51.5		70 (86) - 130 (113)	103%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.1		70 (77) - 130 (121)	104%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	65000	5.55			
540-36-3	1,4-Difluorobenzene	127000	6.757			
3114-55-4	Chlorobenzene-d5	119000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	48500	12.018			

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements
J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040725\
Data File : VX045619.D
Acq On : 07 Apr 2025 11:23
Operator : JC/MD
Sample : Q1731-04 40X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RMW-03B-90-040325

Quant Time: Apr 07 14:39:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	5.550	168	64989	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	126519	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	118849	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	48503	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63693	53.592	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 107.180%			
35) Dibromofluoromethane	5.385	113	46602	51.915	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 103.840%			
50) Toluene-d8	8.647	98	161473	51.537	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 103.080%			
62) 4-Bromofluorobenzene	11.079	95	59465	52.105	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 104.200%			
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.319	101	52944	66.308	ug/l	98
12) 1,1-Dichloroethene	2.306	96	735	0.942	ug/l #	44
27) cis-1,2-Dichloroethene	4.483	96	29406	30.282	ug/l	92
44) Trichloroethene	7.123	130	75209	85.483	ug/l	95
64) Tetrachloroethene	9.275	164	647	0.771	ug/l #	74


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040725\
Data File : VX045619.D
Acq On : 07 Apr 2025 11:23
Operator : JC/MD
Sample : Q1731-04 40X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RMW-03B-90-040325

Quant Time: Apr 07 14:39:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration







#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 7
Delta R.T. -0.000 min
Lab File: VX045619.D
Acq: 07 Apr 2025 11:23

Instrument :
MSVOA_X
ClientSampleId :
RMW-03B-90-040325

Tgt Ion:168 Resp: 64989

Ion	Ratio	Lower	Upper
168	100		
99	64.7	52.3	78.5







#9
1,1,2-Trichlorotrifluoroethane
Concen: 66.308 ug/l
RT: 2.319 min Scan# 202
Delta R.T. -0.006 min
Lab File: VX045619.D
Acq: 07 Apr 2025 11:23

Tgt Ion:101 Resp: 52944

Ion	Ratio	Lower	Upper
101	100		
85	47.6	38.9	58.3
151	70.6	57.5	86.3







#12
1,1-Dichloroethene
Concen: 0.942 ug/l
RT: 2.306 min Scan# 2
Delta R.T. -0.006 min
Lab File: VX045619.D
Acq: 07 Apr 2025 11:23

Instrument :
MSVOA_X
ClientSampleId :
RMW-03B-90-040325

Tgt Ion: 96 Resp: 735

Ion	Ratio	Lower	Upper
96	100		
61	100.0	145.7	218.5#
98	23.2	51.6	77.4#







#27
cis-1,2-Dichloroethene
Concen: 30.282 ug/l
RT: 4.483 min Scan# 557
Delta R.T. 0.000 min
Lab File: VX045619.D
Acq: 07 Apr 2025 11:23

Tgt Ion: 96 Resp: 29406

Ion	Ratio	Lower	Upper
96	100		
61	143.1	0.0	313.2
98	63.8	0.0	126.4







#33
1,2-Dichloroethane-d4
Concen: 53.592 ug/l
RT: 5.952 min Scan# 7
Delta R.T. 0.000 min
Lab File: VX045619.D
Acq: 07 Apr 2025 11:23

Instrument :
MSVOA_X
ClientSampleId :
RMW-03B-90-040325

Tgt Ion: 65 Resp: 63693

Ion	Ratio	Lower	Upper
65	100		
67	50.1	0.0	99.0







#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.757 min Scan# 930
Delta R.T. 0.000 min
Lab File: VX045619.D
Acq: 07 Apr 2025 11:23

Tgt Ion:114 Resp: 126519

Ion	Ratio	Lower	Upper
114	100		
63	24.9	0.0	46.8
88	17.1	0.0	35.4







#35
Dibromofluoromethane
Concen: 51.915 ug/l
RT: 5.385 min Scan# 7
Delta R.T. 0.006 min
Lab File: VX045619.D
Acq: 07 Apr 2025 11:23

Instrument :
MSVOA_X
ClientSampleId :
RMW-03B-90-040325

Tgt Ion:113 Resp: 46602

Ion	Ratio	Lower	Upper
113	100		
111	104.2	81.8	122.6
192	16.5	13.8	20.6







#44
Trichloroethene
Concen: 85.483 ug/l
RT: 7.123 min Scan# 990
Delta R.T. 0.000 min
Lab File: VX045619.D
Acq: 07 Apr 2025 11:23

Tgt Ion:130 Resp: 75209

Ion	Ratio	Lower	Upper
130	100		
95	105.7	0.0	202.0







#50
Toluene-d8
Concen: 51.537 ug/l
RT: 8.647 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045619.D
Acq: 07 Apr 2025 11:23

Instrument :
MSVOA_X
ClientSampleId :
RMW-03B-90-040325

Tgt Ion: 98 Resp: 161473

Ion	Ratio	Lower	Upper
98	100		
100	63.3	52.2	78.4







#62
4-Bromofluorobenzene
Concen: 52.105 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX045619.D
Acq: 07 Apr 2025 11:23

Tgt Ion: 95 Resp: 59465

Ion	Ratio	Lower	Upper
95	100		
174	66.8	0.0	135.8
176	64.1	0.0	131.4







#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX045619.D
Acq: 07 Apr 2025 11:23

Instrument :
MSVOA_X
ClientSampleId :
RMW-03B-90-040325

Tgt Ion:117 Resp: 118849

Ion	Ratio	Lower	Upper
117	100		
82	63.2	49.2	73.8
119	32.3	25.1	37.7







#64
Tetrachloroethene
Concen: 0.771 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. 0.006 min
Lab File: VX045619.D
Acq: 07 Apr 2025 11:23

Tgt Ion:164 Resp: 647

Ion	Ratio	Lower	Upper
164	100		
166	136.8	103.8	155.8
129	146.4	84.3	126.5#
131	137.4	80.9	121.3#






#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045619.D
Acq: 07 Apr 2025 11:23

Instrument :
MSVOA_X
ClientSampleId :
RMW-03B-90-040325

Tgt Ion:152 Resp: 48503

Ion	Ratio	Lower	Upper
152	100		
115	64.9	46.9	140.7
150	154.8	0.0	349.4



Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	04/03/25
Project:	Former Schlumberger STC PTC Site D3868221	Date Received:	04/03/25
Client Sample ID:	EB01-040325	SDG No.:	Q1731
Lab Sample ID:	Q1731-05	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group3
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045600.D	1		04/04/25 15:33	VX040425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	0.26	U	0.26	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.23	U	0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.19	U	0.19	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.20	U	0.20	1.00	ug/L
71-43-2	Benzene	0.15	U	0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.22	U	0.22	1.00	ug/L
79-01-6	Trichloroethene	0.090	U	0.090	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	1.00	ug/L
127-18-4	Tetrachloroethene	0.23	U	0.23	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	55.3		70 (74) - 130 (125)	111%	SPK: 50
1868-53-7	Dibromofluoromethane	51.6		70 (75) - 130 (124)	103%	SPK: 50
2037-26-5	Toluene-d8	50.8		70 (86) - 130 (113)	102%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.5		70 (77) - 130 (121)	103%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	64100	5.544			
540-36-3	1,4-Difluorobenzene	128000	6.757			
3114-55-4	Chlorobenzene-d5	119000	10.055			
3855-82-1	1,4-Dichlorobenzene-d4	49400	12.018			

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements
J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040425\
Data File : VX045600.D
Acq On : 04 Apr 2025 15:33
Operator : JC/MD
Sample : Q1731-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB01-040325

Quant Time: Apr 04 23:12:58 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	5.544	168	64123	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	127534	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	119214	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	49367	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	64832	55.287	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 110.580%			
35) Dibromofluoromethane	5.379	113	46677	51.585	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 103.180%			
50) Toluene-d8	8.647	98	160288	50.751	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 101.500%			
62) 4-Bromofluorobenzene	11.079	95	59219	51.476	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 102.960%			
Target Compounds						Qvalue
16) Acetone	2.380	43	14652	30.429	ug/l	96
20) Methylene Chloride	2.788	84	1138	1.262	ug/l #	78


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040425\
Data File : VX045600.D
Acq On : 04 Apr 2025 15:33
Operator : JC/MD
Sample : Q1731-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB01-040325

Quant Time: Apr 04 23:12:58 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration







#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 7
Delta R.T. -0.006 min
Lab File: VX045600.D
Acq: 04 Apr 2025 15:33

Instrument :
MSVOA_X
ClientSampleId :
EB01-040325

Tgt Ion:168 Resp: 64123

Ion	Ratio	Lower	Upper
168	100		
99	70.8	52.3	78.5







#16
Acetone
Concen: 30.429 ug/l
RT: 2.380 min Scan# 212
Delta R.T. -0.006 min
Lab File: VX045600.D
Acq: 04 Apr 2025 15:33

Tgt Ion: 43 Resp: 14652

Ion	Ratio	Lower	Upper
43	100		
58	30.1	22.3	33.5







#20
Methylene Chloride
Concen: 1.262 ug/l
RT: 2.788 min Scan# 2
Delta R.T. 0.006 min
Lab File: VX045600.D
Acq: 04 Apr 2025 15:33

Instrument :
MSVOA_X
ClientSampleId :
EB01-040325

Tgt Ion: 84 Resp: 1138

Ion	Ratio	Lower	Upper
84	100		
49	164.4	107.5	161.3#
51	36.1	33.0	49.6
86	81.5	50.1	75.1#







#33
1,2-Dichloroethane-d4
Concen: 55.287 ug/l
RT: 5.952 min Scan# 798
Delta R.T. -0.000 min
Lab File: VX045600.D
Acq: 04 Apr 2025 15:33

Tgt Ion: 65 Resp: 64832

Ion	Ratio	Lower	Upper
65	100		
67	48.3	0.0	99.0







#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.757 min Scan# 9
Delta R.T. -0.000 min
Lab File: VX045600.D
Acq: 04 Apr 2025 15:33

Instrument :
MSVOA_X
ClientSampleId :
EB01-040325

Tgt Ion:114 Resp: 127534

Ion	Ratio	Lower	Upper
114	100		
63	23.9	0.0	46.8
88	16.8	0.0	35.4







#35
Dibromofluoromethane
Concen: 51.585 ug/l
RT: 5.379 min Scan# 704
Delta R.T. -0.000 min
Lab File: VX045600.D
Acq: 04 Apr 2025 15:33

Tgt Ion:113 Resp: 46677

Ion	Ratio	Lower	Upper
113	100		
111	101.8	81.8	122.6
192	16.3	13.8	20.6







#50
Toluene-d8
Concen: 50.751 ug/l
RT: 8.647 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045600.D
Acq: 04 Apr 2025 15:33

Instrument :
MSVOA_X
ClientSampleId :
EB01-040325

Tgt Ion: 98 Resp: 160288

Ion	Ratio	Lower	Upper
98	100		
100	64.1	52.2	78.4







#62
4-Bromofluorobenzene
Concen: 51.476 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX045600.D
Acq: 04 Apr 2025 15:33

Tgt Ion: 95 Resp: 59219

Ion	Ratio	Lower	Upper
95	100		
174	66.9	0.0	135.8
176	63.6	0.0	131.4







#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 14
Delta R.T. 0.006 min
Lab File: VX045600.D
Acq: 04 Apr 2025 15:33

Instrument :
MSVOA_X
ClientSampleId :
EB01-040325

Tgt Ion:117 Resp: 119214

Ion	Ratio	Lower	Upper
117	100		
82	59.0	49.2	73.8
119	32.2	25.1	37.7






#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VX045600.D
Acq: 04 Apr 2025 15:33

Tgt Ion:152 Resp: 49367

Ion	Ratio	Lower	Upper
152	100		
115	65.0	46.9	140.7
150	157.8	0.0	349.4



Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	04/03/25
Project:	Former Schlumberger STC PTC Site D3868221	Date Received:	04/03/25
Client Sample ID:	TB01-040325	SDG No.:	Q1731
Lab Sample ID:	Q1731-07	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group3
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045599.D	1		04/04/25 15:10	VX040425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	0.26	U	0.26	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.23	U	0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.19	U	0.19	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.20	U	0.20	1.00	ug/L
71-43-2	Benzene	0.15	U	0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.22	U	0.22	1.00	ug/L
79-01-6	Trichloroethene	0.090	U	0.090	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	1.00	ug/L
127-18-4	Tetrachloroethene	0.23	U	0.23	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	54.5		70 (74) - 130 (125)	109%	SPK: 50
1868-53-7	Dibromofluoromethane	51.0		70 (75) - 130 (124)	102%	SPK: 50
2037-26-5	Toluene-d8	50.6		70 (86) - 130 (113)	101%	SPK: 50
460-00-4	4-Bromofluorobenzene	48.4		70 (77) - 130 (121)	97%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	62200	5.55			
540-36-3	1,4-Difluorobenzene	123000	6.757			
3114-55-4	Chlorobenzene-d5	111000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	43000	12.018			

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements
J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040425\
Data File : VX045599.D
Acq On : 04 Apr 2025 15:10
Operator : JC/MD
Sample : Q1731-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB01-040325

Quant Time: Apr 04 23:12:48 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	5.550	168	62156	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	122509	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	111172	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	43047	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	61983	54.530	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 109.060%			
35) Dibromofluoromethane	5.379	113	44298	50.964	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 101.920%			
50) Toluene-d8	8.647	98	153363	50.550	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 101.100%			
62) 4-Bromofluorobenzene	11.079	95	53504	48.416	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 96.840%			
Target Compounds						Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040425\
Data File : VX045599.D
Acq On : 04 Apr 2025 15:10
Operator : JC/MD
Sample : Q1731-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB01-040325

Quant Time: Apr 04 23:12:48 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration







#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 7
Delta R.T. -0.000 min
Lab File: VX045599.D
Acq: 04 Apr 2025 15:10

Instrument :
MSVOA_X
ClientSampleId :
TB01-040325

Tgt Ion:168 Resp: 62156

Ion	Ratio	Lower	Upper
168	100		
99	70.8	52.3	78.5







#33
1,2-Dichloroethane-d4
Concen: 54.530 ug/l
RT: 5.952 min Scan# 798
Delta R.T. -0.000 min
Lab File: VX045599.D
Acq: 04 Apr 2025 15:10

Tgt Ion: 65 Resp: 61983

Ion	Ratio	Lower	Upper
65	100		
67	50.0	0.0	99.0







#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.757 min Scan# 9
Delta R.T. -0.000 min
Lab File: VX045599.D
Acq: 04 Apr 2025 15:10

Instrument :
MSVOA_X
ClientSampleId :
TB01-040325

Tgt Ion:114 Resp: 122509

Ion	Ratio	Lower	Upper
114	100		
63	25.2	0.0	46.8
88	16.9	0.0	35.4







#35
Dibromofluoromethane
Concen: 50.964 ug/l
RT: 5.379 min Scan# 704
Delta R.T. -0.000 min
Lab File: VX045599.D
Acq: 04 Apr 2025 15:10

Tgt Ion:113 Resp: 44298

Ion	Ratio	Lower	Upper
113	100		
111	103.4	81.8	122.6
192	17.1	13.8	20.6







#50
Toluene-d8
Concen: 50.550 ug/l
RT: 8.647 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045599.D
Acq: 04 Apr 2025 15:10

Instrument :
MSVOA_X
ClientSampleId :
TB01-040325

Tgt Ion: 98 Resp: 153363

Ion	Ratio	Lower	Upper
98	100		
100	64.6	52.2	78.4







#62
4-Bromofluorobenzene
Concen: 48.416 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX045599.D
Acq: 04 Apr 2025 15:10

Tgt Ion: 95 Resp: 53504

Ion	Ratio	Lower	Upper
95	100		
174	68.4	0.0	135.8
176	65.1	0.0	131.4







#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 14
Delta R.T. -0.000 min
Lab File: VX045599.D
Acq: 04 Apr 2025 15:10

Instrument :
MSVOA_X
ClientSampleId :
TB01-040325

Tgt Ion:117 Resp: 111172

Ion	Ratio	Lower	Upper
117	100		
82	61.2	49.2	73.8
119	32.1	25.1	37.7






#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VX045599.D
Acq: 04 Apr 2025 15:10

Tgt Ion:152 Resp: 43047

Ion	Ratio	Lower	Upper
152	100		
115	63.1	46.9	140.7
150	156.6	0.0	349.4



CALIBRATION SUMMARY

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: JACO05

Lab Code: CHEM Case No.: Q1731 SAS No.: Q1731 SDG No.: Q1731

Instrument ID: MSVOA_X Calibration Date(s): 04/01/2025 04/01/2025

Heated Purge: (Y/N) N Calibration Time(s): 17:06 19:02

GC Column: DB-624UI ID: 0.18 (mm)

LAB FILE ID: RRF001 = VX045525.D RRF005 = VX045526.D RRF020 = VX045527.D
RRF050 = VX045528.D RRF100 = VX045529.D RRF150 = VX045530.D

COMPOUND	RRF001	RRF005	RRF020	RRF050	RRF100	RRF150	$\overline{RRF}$	% RSD
Vinyl Chloride	0.671	0.662	0.701	0.716	0.738	0.730	0.703	4.4
1,1-Dichloroethene	0.563	0.588	0.612	0.604	0.618	0.616	0.600	3.5
1,1-Dichloroethane	1.211	1.240	1.318	1.269	1.283	1.292	1.269	3
cis-1,2-Dichloroethene	0.762	0.696	0.760	0.751	0.754	0.760	0.747	3.4
1,1,1-Trichloroethane	1.028	1.052	1.120	1.109	1.153	1.167	1.105	5
Benzene	1.414	1.416	1.519	1.481	1.483	1.465	1.463	2.8
1,2-Dichloroethane	0.533	0.585	0.649	0.622	0.620	0.617	0.604	6.7
Trichloroethene	0.349	0.322	0.356	0.351	0.351	0.356	0.348	3.7
1,1,2-Trichloroethane	0.337	0.346	0.376	0.359	0.358	0.340	0.353	4.1
Tetrachloroethene	0.346	0.373	0.371	0.347	0.333	0.347	0.353	4.5
1,2-Dichloroethane-d4		0.962	0.900	0.868	0.904	0.937	0.914	3.9
Dibromofluoromethane		0.372	0.342	0.345	0.353	0.362	0.355	3.5
Toluene-d8		1.257	1.233	1.214	1.230	1.257	1.238	1.5
4-Bromofluorobenzene		0.413	0.448	0.453	0.481	0.460	0.451	5.5

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

Response Factor Report MSVOA_X2

Method Path : Z:\voasrv\HPCHEM1\MSVOA_X\Method\
Method File : 82X040225W.M
Title : SW846 8260
Last Update : Wed Apr 02 03:11:43 2025
Response Via : Initial Calibration

Calibration Files
1 =VX045525.D 5 =VX045526.D 20 =VX045527.D 50 =VX045528.D 100 =VX045529.D 150 =VX045530.D

		Compound	1	5	20	50	100	150	Avg	%RSD
1)	I	Pentafluorobenzene	----------ISTD----------							
2)	T	Dichlorodifluo...	0.615	0.695	0.700	0.820	0.837	0.820	0.748	12.13
3)	P	Chloromethane	0.764	0.734	0.777	0.784	0.815	0.734	0.768	4.05
4)	C	Vinyl Chloride	0.671	0.662	0.701	0.716	0.738	0.730	0.703	4.43#
5)	T	Bromomethane		0.342	0.327	0.330	0.341	0.327	0.333	2.20
6)	T	Chloroethane	0.378	0.397	0.390	0.398	0.355	0.319	0.373	8.27
7)	T	Trichlorofluor...	1.051	0.999	1.075	1.086	1.089	0.989	1.048	4.22
8)	T	Diethyl Ether	0.345	0.340	0.356	0.355	0.358	0.357	0.352	2.11
9)	T	1,1,2-Trichlor...	0.575	0.599	0.635	0.621	0.621	0.634	0.614	3.74
10)	T	Methyl Iodide		0.705	0.789	0.799	0.788	0.750	0.766	5.08
11)	T	Tert butyl alc...		0.111	0.124	0.127	0.128	0.124	0.123	5.44
12)	CM	1,1-Dichloroet...	0.563	0.588	0.612	0.604	0.618	0.616	0.600	3.51#
13)	T	Acrolein		0.176	0.161	0.167	0.170	0.172	0.169	3.41
14)	T	Allyl chloride	1.092	1.068	1.142	1.172	1.199	1.160	1.139	4.36
15)	T	Acrylonitrile	0.356	0.382	0.413	0.407	0.394	0.376	0.388	5.38
16)	T	Acetone	0.400	0.369	0.387	0.378	0.365	0.355	0.375	4.28
17)	T	Carbon Disulfide	1.334	1.327	1.434	1.553	1.604	1.642	1.483	9.24
18)	T	Methyl Acetate	0.901	0.863	0.864	0.857	0.855	0.846	0.864	2.22
19)	T	Methyl tert-bu...	1.915	1.964	2.169	2.118	2.217	2.216	2.100	6.20
20)	T	Methylene Chlo...	0.692	0.695	0.726	0.709	0.705	0.690	0.703	1.96
21)	T	trans-1,2-Dich...	0.574	0.594	0.636	0.624	0.621	0.630	0.613	3.94
22)	T	Diisopropyl ether	2.003	2.120	2.329	2.293	2.359	2.357	2.243	6.59
23)	T	Vinyl Acetate	1.542	1.693	2.005	2.087	2.147	2.156	1.938	13.34
24)	P	1,1-Dichloroet...	1.211	1.240	1.318	1.269	1.283	1.292	1.269	3.01
25)	T	2-Butanone	0.474	0.537	0.582	0.586	0.571	0.548	0.550	7.55
26)	T	2,2-Dichloropr...	0.676	0.757	0.802	0.847	0.903	0.933	0.820	11.67
27)	T	cis-1,2-Dichlo...	0.762	0.696	0.760	0.751	0.754	0.760	0.747	3.43
28)	T	Bromochloromet...	0.671	0.608	0.617	0.596	0.610	0.576	0.613	5.18
29)	T	Tetrahydrofuran	0.339	0.341	0.371	0.371	0.364	0.352	0.356	4.05
30)	C	Chloroform	1.244	1.309	1.348	1.315	1.309	1.309	1.306	2.60#
31)	T	Cyclohexane		1.044	1.148	1.123	1.145	1.149	1.122	4.00
32)	T	1,1,1-Trichlor...	1.028	1.052	1.120	1.109	1.153	1.167	1.105	4.96
33)	S	1,2-Dichloroet...		0.962	0.900	0.868	0.904	0.937	0.914	3.94
34)	I	1,4-Difluorobenzene	----------ISTD----------							
35)	S	Dibromofluorom...		0.372	0.342	0.345	0.353	0.362	0.355	3.48
36)	T	1,1-Dichloropr...	0.477	0.443	0.474	0.485	0.496	0.499	0.479	4.26
37)	T	Ethyl Acetate	0.582	0.565	0.609	0.623	0.629	0.614	0.604	4.15
38)	T	Carbon Tetrach...	0.450	0.497	0.521	0.536	0.545	0.556	0.518	7.49
39)	T	Methylcyclohexane	0.519	0.527	0.596	0.622	0.628	0.632	0.587	8.76
40)	TM	Benzene	1.414	1.416	1.519	1.481	1.483	1.465	1.463	2.81
41)	T	Methacrylonitrile	0.252	0.276	0.338	0.351	0.350	0.343	0.318	13.52
42)	TM	1,2-Dichloroet...	0.533	0.585	0.649	0.622	0.620	0.617	0.604	6.72
43)	T	Isopropyl Acetate	0.735	0.831	0.952	0.978	1.002	0.991	0.915	11.77
44)	TM	Trichloroethene	0.349	0.322	0.356	0.351	0.351	0.356	0.348	3.67
45)	C	1,2-Dichloropr...	0.309	0.365	0.388	0.376	0.379	0.374	0.365	7.83#
46)	T	Dibromomethane	0.237	0.287	0.297	0.292	0.286	0.282	0.280	7.76
47)	T	Bromodichlorom...	0.521	0.519	0.576	0.571	0.587	0.583	0.560	5.56
48)	T	Methyl methacr...	0.405	0.416	0.486	0.508	0.518	0.502	0.472	10.44
49)	T	1,4-Dioxane	0.007	0.010	0.009	0.009	0.009	0.008	0.009	13.93
50)	S	Toluene-d8		1.257	1.233	1.214	1.230	1.257	1.238	1.50
51)	T	4-Methyl-2-Pen...	0.499	0.578	0.661	0.663	0.647	0.589	0.606	10.57
52)	CM	Toluene	0.817	0.866	0.939	0.910	0.905	0.875	0.885	4.79#
53)	T	t-1,3-Dichloro...	0.316	0.400	0.465	0.508	0.555	0.558	0.467	20.27
54)	T	cis-1,3-Dichlo...	0.371	0.474	0.546	0.568	0.597	0.600	0.526	16.88
55)	T	1,1,2-Trichlor...	0.337	0.346	0.376	0.359	0.358	0.340	0.353	4.11
56)	T	Ethyl methacry...	0.407	0.482	0.562	0.600	0.629	0.595	0.546	15.54

Method Path : Z:\voasrv\HPCHEM1\MSVOA_X\Method\
Method File : 82X040225W.M

57)	T	1,3-Dichloropr...	0.555	0.614	0.652	0.633	0.630	0.601	0.614	5.49
58)	T	2-Chloroethyl ...	0.221	0.257	0.286	0.298	0.301	0.293	0.276	11.26
59)	T	2-Hexanone	0.363	0.429	0.488	0.492	0.484	0.439	0.449	11.14
60)	T	Dibromochlorom...	0.313	0.352	0.404	0.412	0.421	0.405	0.385	11.06
61)	T	1,2-Dibromoethane	0.294	0.351	0.380	0.373	0.380	0.364	0.357	9.12
62)	S	4-Bromofluorob...		0.413	0.448	0.453	0.481	0.460	0.451	5.45
63)	I	Chlorobenzene-d5	----------ISTD----------							
64)	T	Tetrachloroethene	0.346	0.373	0.371	0.347	0.333	0.347	0.353	4.47
65)	PM	Chlorobenzene	0.951	1.054	1.123	1.084	1.086	1.112	1.068	5.83
66)	T	1,1,1,2-Tetrac...	0.368	0.344	0.372	0.373	0.379	0.390	0.371	4.18
67)	C	Ethyl Benzene	1.608	1.819	2.002	2.007	1.993	2.053	1.914	8.89#
68)	T	m/p-Xylenes	0.594	0.669	0.732	0.728	0.723	0.729	0.696	7.92
69)	T	o-Xylene	0.562	0.676	0.725	0.719	0.716	0.714	0.686	9.17
70)	T	Styrene	0.897	1.043	1.202	1.216	1.230	1.202	1.132	11.82
71)	P	Bromoform	0.220	0.252	0.272	0.296	0.307	0.315	0.277	13.07
72)	I	1,4-Dichlorobenzen...	----------ISTD----------							
73)	T	Isopropylbenzene	3.581	3.850	4.224	4.181	4.043	4.151	4.005	6.17
74)	T	N-amyl acetate	1.442	1.682	1.960	2.058	2.132	2.221	1.916	15.51
75)	P	1,1,2,2-Tetrac...	1.457	1.428	1.461	1.384	1.354	1.358	1.407	3.43
76)	T	1,2,3-Trichlor...	1.221	1.257	1.251	1.225	1.192	1.171	1.220	2.71
77)	T	Bromobenzene	0.877	0.927	0.953	0.938	0.925	0.932	0.925	2.77
78)	T	n-propylbenzene	3.734	4.346	5.012	4.896	4.819	4.872	4.613	10.59
79)	T	2-Chlorotoluene	2.894	2.873	3.130	2.966	2.915	2.918	2.949	3.19
80)	T	1,3,5-Trimethy...	2.931	3.119	3.571	3.511	3.367	3.371	3.312	7.35
81)	T	trans-1,4-Dich...		0.274	0.323	0.371	0.401	0.428	0.359	17.16
82)	T	4-Chlorotoluene	2.961	3.156	3.470	3.449	3.350	3.343	3.288	5.93
83)	T	tert-Butylbenzene	2.894	3.124	3.437	3.427	3.387	3.405	3.279	6.78
84)	T	1,2,4-Trimethy...	2.906	3.129	3.529	3.523	3.438	3.420	3.324	7.57
85)	T	sec-Butylbenzene	3.195	3.828	4.316	4.300	4.232	4.292	4.027	11.12
86)	T	p-Isopropyltol...	2.734	3.157	3.515	3.524	3.506	3.484	3.320	9.63
87)	T	1,3-Dichlorobe...	1.658	1.605	1.726	1.699	1.714	1.706	1.684	2.70
88)	T	1,4-Dichlorobe...	1.671	1.724	1.768	1.703	1.674	1.706	1.708	2.09
89)	T	n-Butylbenzene	2.117	2.486	2.974	3.160	3.256	3.283	2.879	16.50
90)	T	Hexachloroethane	0.493	0.493	0.556	0.597	0.623	0.661	0.571	12.14
91)	T	1,2-Dichlorobe...	1.644	1.645	1.750	1.678	1.665	1.667	1.675	2.34
92)	T	1,2-Dibromo-3-...	0.196	0.260	0.312	0.316	0.333	0.349	0.294	19.37
93)	T	1,2,4-Trichlor...	0.727	0.844	0.948	0.947	1.045	1.083	0.932	14.07
94)	T	Hexachlorobuta...	0.379	0.389	0.401	0.404	0.415	0.422	0.402	4.03
95)	T	Naphthalene	2.544	3.070	3.630	3.722	3.867	3.982	3.469	15.93
96)	T	1,2,3-Trichlor...	0.782	0.916	0.999	1.000	1.063	1.097	0.976	11.63

(#) = Out of Range

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045525.D
Acq On : 01 Apr 2025 17:06
Operator : JC/MD
Sample : VSTDICC001
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 3 Sample Multiplier: 1

Quant Time: Apr 02 02:50:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 02:44:48 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) Pentafluorobenzene	5.544	168	96737	50.000	ug/l		0.00
34) 1,4-Difluorobenzene	6.757	114	175266	50.000	ug/l		0.00
63) Chlorobenzene-d5	10.049	117	153312	50.000	ug/l		0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	61114	50.000	ug/l		0.00
System Monitoring Compounds							
33) 1,2-Dichloroethane-d4	0.000	65	0d	0.000	ug/l		
Spiked Amount 50.000	Range 74 - 125		Recovery =	0.000%#			
35) Dibromofluoromethane	0.000	113	0d	0.000	ug/l		
Spiked Amount 50.000	Range 75 - 124		Recovery =	0.000%#			
50) Toluene-d8	0.000	98	0d	0.000	ug/l		
Spiked Amount 50.000	Range 86 - 113		Recovery =	0.000%#			
62) 4-Bromofluorobenzene	0.000	95	0d	0.000	ug/l		
Spiked Amount 50.000	Range 77 - 121		Recovery =	0.000%#			
Target Compounds							Qvalue
2) Dichlorodifluoromethane	1.167	85	1189	0.822	ug/l		94
3) Chloromethane	1.307	50	1479	0.995	ug/l	#	86
4) Vinyl Chloride	1.374	62	1299	0.955	ug/l		98
6) Chloroethane	1.673	64	732	1.014	ug/l	#	67
7) Trichlorofluoromethane	1.880	101	2034	1.003	ug/l		93
8) Diethyl Ether	2.130	74	668	0.981	ug/l		83
9) 1,1,2-Trichlorotrifluo...	2.325	101	1113	0.936	ug/l		94
12) 1,1-Dichloroethene	2.313	96	1090	0.939	ug/l		95
14) Allyl chloride	2.660	41	2113	0.959	ug/l		93
15) Acrylonitrile	3.075	53	3446	4.592	ug/l		95
16) Acetone	2.380	43	3867	5.323	ug/l		98
17) Carbon Disulfide	2.508	76	2580	0.899	ug/l	#	95
18) Methyl Acetate	2.703	43	1744	1.043	ug/l		91
19) Methyl tert-butyl Ether	3.117	73	3705	0.912	ug/l		92
20) Methylene Chloride	2.782	84	1338	0.984	ug/l		91
21) trans-1,2-Dichloroethene	3.087	96	1111	0.936	ug/l	#	93
22) Diisopropyl ether	3.758	45	3875	0.893	ug/l	#	59
23) Vinyl Acetate	3.727	43	14921	3.979	ug/l		95
24) 1,1-Dichloroethane	3.611	63	2343	0.954	ug/l	#	90
25) 2-Butanone	4.581	43	4590	4.317	ug/l		92
26) 2,2-Dichloropropane	4.471	77	1307	0.824	ug/l		85
27) cis-1,2-Dichloroethene	4.495	96	1475	1.020	ug/l		80
28) Bromochloromethane	4.891	49	1298	1.094	ug/l	#	97
29) Tetrahydrofuran	5.032	42	3276	4.752	ug/l		97
30) Chloroform	5.093	83	2406	0.953	ug/l		84
32) 1,1,1-Trichloroethane	5.385	97	1989	0.931	ug/l	#	53
36) 1,1-Dichloropropene	5.696	75	1671	0.995	ug/l	#	89
37) Ethyl Acetate	4.751	43	2040	0.964	ug/l	#	76
38) Carbon Tetrachloride	5.672	117	1578	0.870	ug/l	#	91
39) Methylcyclohexane	7.379	83	1818	0.883	ug/l	#	89
40) Benzene	6.038	78	4957	0.967	ug/l		97
41) Methacrylonitrile	4.952	41	883m	0.784	ug/l		
42) 1,2-Dichloroethane	6.092	62	1867	0.882	ug/l		86
43) Isopropyl Acetate	6.361	43	2578	0.804	ug/l	#	86
44) Trichloroethene	7.123	130	1224	1.004	ug/l		87
45) 1,2-Dichloropropane	7.434	63	1082	0.846	ug/l		90

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045525.D
Acq On : 01 Apr 2025 17:06
Operator : JC/MD
Sample : VSTDICC001
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED
Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025

Quant Time: Apr 02 02:50:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 02:44:48 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
46) Dibromomethane	7.592	93	831	0.846	ug/l		90
47) Bromodichloromethane	7.824	83	1827	0.931	ug/l	#	77
48) Methyl methacrylate	7.720	41	1419	0.857	ug/l		92
49) 1,4-Dioxane	7.671	88	458m	15.288	ug/l		
51) 4-Methyl-2-Pentanone	8.574	43	8746	4.114	ug/l		100
52) Toluene	8.720	92	2863	0.923	ug/l		96
53) t-1,3-Dichloropropene	8.988	75	1109	2.850	ug/l		93
54) cis-1,3-Dichloropropene	8.379	75	1300	0.705	ug/l		90
55) 1,1,2-Trichloroethane	9.159	97	1183	0.957	ug/l	#	76
56) Ethyl methacrylate	9.129	69	1425m	0.745	ug/l		
57) 1,3-Dichloropropane	9.311	76	1946	0.904	ug/l		91
58) 2-Chloroethyl Vinyl ether	8.245	63	3880	4.008	ug/l		93
59) 2-Hexanone	9.439	43	6360	4.040	ug/l		85
60) Dibromochloromethane	9.519	129	1097	0.814	ug/l		99
61) 1,2-Dibromoethane	9.610	107	1032	0.825	ug/l		96
64) Tetrachloroethene	9.269	164	1062	0.981	ug/l		92
65) Chlorobenzene	10.080	112	2916	0.890	ug/l		95
66) 1,1,1,2-Tetrachloroethane	10.159	131	1128	0.992	ug/l	#	64
67) Ethyl Benzene	10.195	91	4930	0.840	ug/l		90
68) m/p-Xylenes	10.305	106	3645	1.708	ug/l		97
69) o-Xylene	10.640	106	1724	0.820	ug/l		83
70) Styrene	10.659	104	2750	0.793	ug/l		94
71) Bromoform	10.805	173	675	0.795	ug/l	#	94
73) Isopropylbenzene	10.964	105	4377	0.894	ug/l		95
74) N-amyl acetate	10.854	43	1762	0.752	ug/l	#	93
75) 1,1,2,2-Tetrachloroethane	11.214	83	1781	1.036	ug/l		94
76) 1,2,3-Trichloropropane	11.244	75	1493m	1.002	ug/l		
77) Bromobenzene	11.201	156	1072	0.948	ug/l		95
78) n-propylbenzene	11.305	91	4564	0.809	ug/l		97
79) 2-Chlorotoluene	11.366	91	3537	0.981	ug/l		100
80) 1,3,5-Trimethylbenzene	11.451	105	3582	0.885	ug/l		100
82) 4-Chlorotoluene	11.457	91	3619	0.900	ug/l		96
83) tert-Butylbenzene	11.713	119	3537	0.883	ug/l		94
84) 1,2,4-Trimethylbenzene	11.750	105	3552	0.874	ug/l		99
85) sec-Butylbenzene	11.890	105	3905	0.793	ug/l		98
86) p-Isopropyltoluene	12.012	119	3342	0.824	ug/l		94
87) 1,3-Dichlorobenzene	11.969	146	2026	0.984	ug/l		97
88) 1,4-Dichlorobenzene	12.043	146	2042m	0.978	ug/l		
89) n-Butylbenzene	12.335	91	2587	0.735	ug/l		99
90) Hexachloroethane	12.543	117	602	0.863	ug/l		95
91) 1,2-Dichlorobenzene	12.335	146	2009	0.981	ug/l		93
92) 1,2-Dibromo-3-Chloropr...	12.945	75	239	0.665	ug/l	#	73
93) 1,2,4-Trichlorobenzene	13.591	180	888	0.779	ug/l		94
94) Hexachlorobutadiene	13.725	225	463	0.943	ug/l		90
95) Naphthalene	13.780	128	3109	0.733	ug/l		99
96) 1,2,3-Trichlorobenzene	13.969	180	956	0.801	ug/l		88

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045525.D
Acq On : 01 Apr 2025 17:06
Operator : JC/MD
Sample : VSTDICC001
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 3 Sample Multiplier: 1


Quant Time: Apr 02 02:50:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 02:44:48 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025






#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 7
Delta R.T. -0.006 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion:168 Resp: 9673


Ion	Ratio	Lower	Upper
168	100		
99	69.3	52.3	78.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#2
Dichlorodifluoromethane
Concen: 0.822 ug/l
RT: 1.167 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion: 85 Resp: 1189

Ion	Ratio	Lower	Upper
85	100		
87	35.5	16.1	48.2






#3
Chloromethane
Concen: 0.995 ug/l
RT: 1.307 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion: 50 Resp: 147


Ion	Ratio	Lower	Upper
50	100		
52	24.4	25.9	38.9



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#4
Vinyl Chloride
Concen: 0.955 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion: 62 Resp: 1299

Ion	Ratio	Lower	Upper
62	100		
64	32.9	25.5	38.3






#6
Chloroethane
Concen: 1.014 ug/l
RT: 1.673 min Scan# 9
Delta R.T. 0.006 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion: 64 Resp: 73


Ion	Ratio	Lower	Upper
64	100		
66	51.4	26.3	39.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#7
Trichlorofluoromethane
Concen: 1.003 ug/l
RT: 1.880 min Scan# 130
Delta R.T. 0.006 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion:101 Resp: 2034

Ion	Ratio	Lower	Upper
101	100		
103	58.3	51.1	76.7





#8
Diethyl Ether
Concen: 0.981 ug/l
RT: 2.130 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06


Tgt Ion: 74 Resp: 66


Ion	Ratio	Lower	Upper
74	100		
45	124.1	53.0	159.1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025







#9
1,1,2-Trichlorotrifluoroethane
Concen: 0.936 ug/l
RT: 2.325 min Scan# 203
Delta R.T. 0.000 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion:101 Resp: 1113

Ion	Ratio	Lower	Upper
101	100		
85	48.7	38.9	58.3
151	80.3	57.5	86.3






#12
1,1-Dichloroethene
Concen: 0.939 ug/l
RT: 2.313 min Scan# 2(
Delta R.T. 0.000 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion: 96 Resp: 109(


Ion	Ratio	Lower	Upper
96	100		
61	187.2	145.7	218.5
98	58.3	51.6	77.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#14
Allyl chloride
Concen: 0.959 ug/l
RT: 2.660 min Scan# 258
Delta R.T. 0.000 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion: 41 Resp: 2113

Ion	Ratio	Lower	Upper
41	100		
39	82.6	60.2	90.2
76	28.3	24.6	36.8






#15
Acrylonitrile
Concen: 4.592 ug/l
RT: 3.075 min Scan# 3
Delta R.T. 0.012 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion: 53 Resp: 344


Ion	Ratio	Lower	Upper
53	100		
52	89.4	67.4	101.2
51	40.1	29.9	44.9



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#16
Acetone
Concen: 5.323 ug/l
RT: 2.380 min Scan# 212
Delta R.T. -0.006 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion: 43 Resp: 3867

Ion	Ratio	Lower	Upper
43	100		
58	29.2	22.3	33.5





#17
Carbon Disulfide
Concen: 0.899 ug/l
RT: 2.508 min Scan# 2
Delta R.T. 0.000 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06


Tgt Ion: 76 Resp: 258


Ion	Ratio	Lower	Upper
76	100		
78	10.7	7.0	10.6

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025







#18
Methyl Acetate
Concen: 1.043 ug/l
RT: 2.703 min Scan# 265
Delta R.T. 0.000 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion: 43 Resp: 1744

Ion	Ratio	Lower	Upper
43	100		
74	17.8	17.5	26.3






#19
Methyl tert-butyl Ether
Concen: 0.912 ug/l
RT: 3.117 min Scan# 3
Delta R.T. 0.006 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion: 73 Resp: 370


Ion	Ratio	Lower	Upper
73	100		
57	18.1	17.6	26.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#20
Methylene Chloride
Concen: 0.984 ug/l
RT: 2.782 min Scan# 278
Delta R.T. 0.000 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion: 84 Resp: 1338

Ion	Ratio	Lower	Upper
84	100		
49	121.7	107.5	161.3
51	33.8	33.0	49.6
86	65.7	50.1	75.1






#21
trans-1,2-Dichloroethene
Concen: 0.936 ug/l
RT: 3.087 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion: 96 Resp: 111


Ion	Ratio	Lower	Upper
96	100		
61	152.6	121.0	181.6
98	79.7	49.7	74.5#



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#22
Diisopropyl ether
Concen: 0.893 ug/l
RT: 3.758 min Scan# 438
Delta R.T. -0.006 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion: 45 Resp: 3875

Ion	Ratio	Lower	Upper
45	100		
43	123.5	59.3	88.9#
87	28.2	20.4	30.6
59	8.0	9.0	13.6#





#23
Vinyl Acetate
Concen: 3.979 ug/l
RT: 3.727 min Scan# 4
Delta R.T. 0.006 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06


Tgt Ion: 43 Resp: 1492


Ion	Ratio	Lower	Upper
43	100		
86	7.7	7.5	11.3

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025







#24
1,1-Dichloroethane
Concen: 0.954 ug/l
RT: 3.611 min Scan# 414
Delta R.T. 0.006 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion: 63 Resp: 2343

Ion	Ratio	Lower	Upper
63	100		
98	7.4	3.4	10.1
100	11.6	2.2	6.6#






#25
2-Butanone
Concen: 4.317 ug/l
RT: 4.581 min Scan# 5
Delta R.T. 0.025 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion: 43 Resp: 459


Ion	Ratio	Lower	Upper
43	100		
72	27.2	18.6	28.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#26
2,2-Dichloropropane
Concen: 0.824 ug/l
RT: 4.471 min Scan# 555
Delta R.T. 0.000 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion: 77 Resp: 1307

Ion	Ratio	Lower	Upper
77	100		
97	13.9	10.6	31.8






#27
cis-1,2-Dichloroethene
Concen: 1.020 ug/l
RT: 4.495 min Scan# 5
Delta R.T. 0.012 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion: 96 Resp: 147


Ion	Ratio	Lower	Upper
96	100		
61	127.7	0.0	313.2
98	52.4	0.0	126.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#28
Bromochloromethane
Concen: 1.094 ug/l
RT: 4.891 min Scan# 624
Delta R.T. -0.006 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion: 49 Resp: 1298

Ion	Ratio	Lower	Upper
49	100		
129	0.0	0.0	3.8
130	66.0	54.6	81.8





#29
Tetrahydrofuran
Concen: 4.752 ug/l
RT: 5.032 min Scan# 64
Delta R.T. 0.031 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06


Tgt Ion: 42 Resp: 327


Ion	Ratio	Lower	Upper
42	100		
72	42.3	33.4	50.0
71	36.0	31.4	47.0

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025







#30
Chloroform
Concen: 0.953 ug/l
RT: 5.093 min Scan# 657
Delta R.T. 0.006 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion: 83 Resp: 2406

Ion	Ratio	Lower	Upper
83	100		
85	53.7	52.8	79.2






#32
1,1,1-Trichloroethane
Concen: 0.931 ug/l
RT: 5.385 min Scan# 7
Delta R.T. 0.006 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion: 97 Resp: 198


Ion	Ratio	Lower	Upper
97	100		
99	0.0	51.7	77.5#
61	52.8	41.0	61.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#33
1,2-Dichloroethane-d4
Concen: 0.000 ug/l
Expected RT: 5.95 min

Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion: 65

Sig	Exp Ratio
65	100
67	49.5



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.757 min Scan# 9
Delta R.T. 0.000 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06


Tgt Ion:114 Resp: 17526


Ion	Ratio	Lower	Upper
114	100		
63	24.3	0.0	46.8
88	18.1	0.0	35.4

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025







#35
Dibromofluoromethane
Concen: 0.000 ug/l
Expected RT: 5.38 min

Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion: 113

Sig	Exp Ratio
113	100
111	102.2
192	17.2






#36
1,1-Dichloropropene
Concen: 0.995 ug/l
RT: 5.696 min Scan# 7
Delta R.T. 0.012 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion: 75 Resp: 167


Ion	Ratio	Lower	Upper
75	100		
110	27.6	16.9	50.6
77	25.0	25.0	37.4#



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#37
Ethyl Acetate
Concen: 0.964 ug/l
RT: 4.751 min Scan# 601
Delta R.T. 0.037 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion: 43 Resp: 2040

Ion	Ratio	Lower	Upper
43	100		
61	3.3	9.9	14.9#
70	0.0	7.5	11.3#






#38
Carbon Tetrachloride
Concen: 0.870 ug/l
RT: 5.672 min Scan# 7
Delta R.T. 0.000 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion:117 Resp: 157


Ion	Ratio	Lower	Upper
117	100		
119	102.7	76.9	115.3
121	39.6	24.2	36.2#



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#39
Methylcyclohexane
Concen: 0.883 ug/l
RT: 7.379 min Scan# 1032
Delta R.T. 0.006 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion: 83 Resp: 1818

Ion	Ratio	Lower	Upper
83	100		
55	77.0	64.9	97.3
98	62.1	37.4	56.0#






#40
Benzene
Concen: 0.967 ug/l
RT: 6.038 min Scan# 8
Delta R.T. 0.006 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion: 78 Resp: 495


Ion	Ratio	Lower	Upper
78	100		
77	25.5	19.0	28.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#41
Methacrylonitrile
Concen: 0.784 ug/l m
RT: 4.952 min Scan# 634
Delta R.T. 0.031 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion: 41 Resp: 883

Ion	Ratio	Lower	Upper
41	100		
39	0.0	47.0	70.6#
67	0.0	53.2	79.8#
52	0.0	25.0	37.6#





#42
1,2-Dichloroethane
Concen: 0.882 ug/l
RT: 6.092 min Scan# 8
Delta R.T. 0.012 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06


Tgt Ion: 62 Resp: 186


Ion	Ratio	Lower	Upper
62	100		
98	2.6	0.0	15.2

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025







#43
Isopropyl Acetate
Concen: 0.804 ug/l
RT: 6.361 min Scan# 865
Delta R.T. 0.025 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion: 43 Resp: 2578

Ion	Ratio	Lower	Upper
43	100		
61	10.6	14.1	21.1#
87	6.9	9.2	13.8#






#44
Trichloroethene
Concen: 1.004 ug/l
RT: 7.123 min Scan# 9
Delta R.T. 0.000 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion:130 Resp: 122


Ion	Ratio	Lower	Upper
130	100		
95	88.0	0.0	202.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#45
1,2-Dichloropropane
Concen: 0.846 ug/l
RT: 7.434 min Scan# 1041
Delta R.T. 0.006 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion: 63 Resp: 1082

Ion	Ratio	Lower	Upper
63	100		
65	26.6	25.8	38.6






#46
Dibromomethane
Concen: 0.846 ug/l
RT: 7.592 min Scan# 1
Delta R.T. 0.018 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion: 93 Resp: 83


Ion	Ratio	Lower	Upper
93	100		
95	92.2	64.8	97.2
174	88.4	65.4	98.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#47
Bromodichloromethane
Concen: 0.931 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. 0.006 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion: 83 Resp: 1827

Ion	Ratio	Lower	Upper
83	100		
85	42.7	49.7	74.5#
127	8.5	6.1	9.1






#48
Methyl methacrylate
Concen: 0.857 ug/l
RT: 7.720 min Scan# 1
Delta R.T. 0.031 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion: 41 Resp: 141


Ion	Ratio	Lower	Upper
41	100		
69	64.7	58.2	87.2
39	72.7	53.6	80.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#49
1,4-Dioxane
Concen: 15.288 ug/l m
RT: 7.671 min Scan# 1080
Delta R.T. 0.012 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion: 88 Resp: 458

Ion	Ratio	Lower	Upper
88	100		
43	30.3	32.2	48.4#
58	72.3	58.6	88.0





#50
Toluene-d8
Concen: 0.000 ug/l
Expected RT: 8.65 min

Lab File: VX045525.D
Acq: 01 Apr 2025 17:06


Tgt Ion: 98


Sig	Exp Ratio
98	100
100	65.3


Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001


Manual Integrations APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025











#51
4-Methyl-2-Pentanone
Concen: 4.114 ug/l
RT: 8.574 min Scan# 1228
Delta R.T. 0.000 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion: 43 Resp: 8746

Ion	Ratio	Lower	Upper
43	100		
58	35.5	28.5	42.7






#52
Toluene
Concen: 0.923 ug/l
RT: 8.720 min Scan# 1
Delta R.T. 0.006 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion: 92 Resp: 286


Ion	Ratio	Lower	Upper
92	100		
91	167.2	138.7	208.1



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#53
t-1,3-Dichloropropene
Concen: 2.850 ug/l
RT: 8.988 min Scan# 1296
Delta R.T. 0.012 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion: 75 Resp: 1109

Ion	Ratio	Lower	Upper
75	100		
77	35.6	25.4	38.0






#54
cis-1,3-Dichloropropene
Concen: 0.705 ug/l
RT: 8.379 min Scan# 1
Delta R.T. 0.013 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion: 75 Resp: 130


Ion	Ratio	Lower	Upper
75	100		
77	25.6	24.6	36.8
39	52.4	48.4	72.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#55
1,1,2-Trichloroethane
Concen: 0.957 ug/l
RT: 9.159 min Scan# 1324
Delta R.T. 0.012 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion: 97 Resp: 1183

Ion	Ratio	Lower	Upper
97	100		
83	53.9	71.8	107.8#
85	48.5	47.0	70.6
99	51.7	48.6	73.0






#56
Ethyl methacrylate
Concen: 0.745 ug/l m
RT: 9.129 min Scan# 1
Delta R.T. 0.013 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion: 69 Resp: 142


Ion	Ratio	Lower	Upper
69	100		
41	77.3	61.3	91.9
39	56.9	40.2	60.2



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#57
1,3-Dichloropropane
Concen: 0.904 ug/l
RT: 9.311 min Scan# 1349
Delta R.T. 0.006 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion: 76 Resp: 1946

Ion	Ratio	Lower	Upper
76	100		
78	27.5	26.0	39.0






#58
2-Chloroethyl Vinyl ether
Concen: 4.008 ug/l
RT: 8.245 min Scan# 1
Delta R.T. 0.006 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion: 63 Resp: 388


Ion	Ratio	Lower	Upper
63	100		
106	22.6	21.0	31.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#59
2-Hexanone
Concen: 4.040 ug/l
RT: 9.439 min Scan# 1370
Delta R.T. 0.012 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion: 43 Resp: 6360

Ion	Ratio	Lower	Upper
43	100		
58	39.5	24.9	74.6






#60
Dibromochloromethane
Concen: 0.814 ug/l
RT: 9.519 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion:129 Resp: 109


Ion	Ratio	Lower	Upper
129	100		
127	80.0	39.4	118.2



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#61
1,2-Dibromoethane
Concen: 0.825 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. 0.006 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion:107 Resp: 1032

Ion	Ratio	Lower	Upper
107	100		
109	98.1	75.7	113.5





#62
4-Bromofluorobenzene
Concen: 0.000 ug/l
Expected RT: 11.08 min

Lab File: VX045525.D
Acq: 01 Apr 2025 17:06


Tgt Ion: 95


Sig	Exp Ratio
95	100
174	67.9
176	65.7


Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001


Manual Integrations
APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025











#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. 0.000 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion:117 Resp: 153312

Ion	Ratio	Lower	Upper
117	100		
82	60.5	49.2	73.8
119	31.0	25.1	37.7






#64
Tetrachloroethene
Concen: 0.981 ug/l
RT: 9.269 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion:164 Resp: 106


Ion	Ratio	Lower	Upper
164	100		
166	137.7	103.8	155.8
129	90.1	84.3	126.5
131	103.7	80.9	121.3



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#65
Chlorobenzene
Concen: 0.890 ug/l
RT: 10.080 min Scan# 1475
Delta R.T. 0.006 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion:112 Resp: 2916

Ion	Ratio	Lower	Upper
112	100		
114	33.5	24.7	37.1






#66
1,1,1,2-Tetrachloroethane
Concen: 0.992 ug/l
RT: 10.159 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion:131 Resp: 112


Ion	Ratio	Lower	Upper
131	100		
133	91.6	47.5	142.5
119	0.0	32.9	98.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#67
Ethyl Benzene
Concen: 0.840 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.006 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion: 91 Resp: 4930

Ion	Ratio	Lower	Upper
91	100		
106	35.2	23.7	35.5






#68
m/p-Xylenes
Concen: 1.708 ug/l
RT: 10.305 min Scan# 1
Delta R.T. 0.006 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion:106 Resp: 364


Ion	Ratio	Lower	Upper
106	100		
91	221.2	172.6	258.8



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#69
o-Xylene
Concen: 0.820 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. 0.000 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion:106 Resp: 1724

Ion	Ratio	Lower	Upper
106	100		
91	254.4	113.6	340.6






#70
Styrene
Concen: 0.793 ug/l
RT: 10.659 min Scan# 1
Delta R.T. 0.006 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion:104 Resp: 275


Ion	Ratio	Lower	Upper
104	100		
78	50.1	45.4	68.2
103	53.2	43.8	65.8



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#71
Bromoform
Concen: 0.795 ug/l
RT: 10.805 min Scan# 1594
Delta R.T. 0.006 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion:173 Resp: 675

Ion	Ratio	Lower	Upper
173	100		
175	42.7	23.4	70.2
254	0.0	0.0	0.0






#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion:152 Resp: 6111


Ion	Ratio	Lower	Upper
152	100		
115	63.5	46.9	140.7
150	155.9	0.0	349.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#73
Isopropylbenzene
Concen: 0.894 ug/l
RT: 10.964 min Scan# 1620
Delta R.T. 0.006 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion:105 Resp: 4377

Ion	Ratio	Lower	Upper
105	100		
120	27.6	12.7	38.0






#74
N-amyl acetate
Concen: 0.752 ug/l
RT: 10.854 min Scan# 1
Delta R.T. 0.012 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion: 43 Resp: 176


Ion	Ratio	Lower	Upper
43	100		
70	34.4	30.1	45.1
55	17.3	18.3	27.5
61	19.8	17.5	26.3



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#75
1,1,2,2-Tetrachloroethane
Concen: 1.036 ug/l
RT: 11.214 min Scan# 1661
Delta R.T. 0.006 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion: 83 Resp: 1781

Ion	Ratio	Lower	Upper
83	100		
131	5.3	4.7	14.1
85	61.1	32.4	97.2






#76
1,2,3-Trichloropropane
Concen: 1.002 ug/l m
RT: 11.244 min Scan# 1
Delta R.T. 0.006 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion: 75 Resp: 149


Ion	Ratio	Lower	Upper
75	100		
77	42.1	20.0	59.9



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#77
Bromobenzene
Concen: 0.948 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. 0.006 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion:156 Resp: 1072

Ion	Ratio	Lower	Upper
156	100		
77	176.5	92.6	277.8
158	98.4	48.1	144.4






#78
n-propylbenzene
Concen: 0.809 ug/l
RT: 11.305 min Scan# 1
Delta R.T. 0.006 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion: 91 Resp: 456


Ion	Ratio	Lower	Upper
91	100		
120	22.8	10.8	32.3



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#79
2-Chlorotoluene
Concen: 0.981 ug/l
RT: 11.366 min Scan# 1686
Delta R.T. 0.006 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion: 91 Resp: 3537

Ion	Ratio	Lower	Upper
91	100		
126	31.8	15.8	47.4






#80
1,3,5-Trimethylbenzene
Concen: 0.885 ug/l
RT: 11.451 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion:105 Resp: 358


Ion	Ratio	Lower	Upper
105	100		
120	45.8	23.0	69.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#82
4-Chlorotoluene
Concen: 0.900 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.006 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion: 91 Resp: 3619

Ion	Ratio	Lower	Upper
91	100		
126	24.7	13.3	39.9






#83
tert-Butylbenzene
Concen: 0.883 ug/l
RT: 11.713 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion:119 Resp: 353


Ion	Ratio	Lower	Upper
119	100		
91	70.3	32.4	97.2
134	25.2	11.5	34.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#84
1,2,4-Trimethylbenzene
Concen: 0.874 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion:105 Resp: 3552

Ion	Ratio	Lower	Upper
105	100		
120	42.0	21.4	64.3





#85
sec-Butylbenzene
Concen: 0.793 ug/l
RT: 11.890 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06


Tgt Ion:105 Resp: 390


Ion	Ratio	Lower	Upper
105	100		
134	18.3	9.6	28.6

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025







#86
p-Isopropyltoluene
Concen: 0.824 ug/l
RT: 12.012 min Scan# 1792
Delta R.T. 0.006 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion:119 Resp: 3342

Ion	Ratio	Lower	Upper
119	100		
134	26.6	12.6	37.8
91	31.9	13.7	41.1






#87
1,3-Dichlorobenzene
Concen: 0.984 ug/l
RT: 11.969 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion:146 Resp: 202


Ion	Ratio	Lower	Upper
146	100		
111	42.2	22.2	66.6
148	61.2	31.9	95.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#88
1,4-Dichlorobenzene
Concen: 0.978 ug/l m
RT: 12.043 min Scan# 1797
Delta R.T. 0.006 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion:146 Resp: 2042

Ion	Ratio	Lower	Upper
146	100		
111	41.8	20.9	62.8
148	60.7	31.5	94.5






#89
n-Butylbenzene
Concen: 0.735 ug/l
RT: 12.335 min Scan# 1
Delta R.T. 0.006 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion: 91 Resp: 258


Ion	Ratio	Lower	Upper
91	100		
92	53.2	26.9	80.6
134	24.8	11.9	35.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#90
Hexachloroethane
Concen: 0.863 ug/l
RT: 12.543 min Scan# 1879
Delta R.T. 0.006 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion:117 Resp: 602

Ion	Ratio	Lower	Upper
117	100		
201	66.3	31.2	93.6






#91
1,2-Dichlorobenzene
Concen: 0.981 ug/l
RT: 12.335 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion:146 Resp: 200


Ion	Ratio	Lower	Upper
146	100		
111	42.0	22.9	68.8
148	58.2	31.9	95.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#92
1,2-Dibromo-3-Chloropropane
Concen: 0.665 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. 0.006 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion: 75 Resp: 239

Ion	Ratio	Lower	Upper
75	100		
155	105.4	34.9	104.7#
157	103.8	44.9	134.7






#93
1,2,4-Trichlorobenzene
Concen: 0.779 ug/l
RT: 13.591 min Scan# 2
Delta R.T. 0.006 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion:180 Resp: 88


Ion	Ratio	Lower	Upper
180	100		
182	87.6	47.9	143.6
145	35.8	17.9	53.8



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#94
Hexachlorobutadiene
Concen: 0.943 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. 0.006 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion:225 Resp: 463

Ion	Ratio	Lower	Upper
225	100		
223	54.9	31.1	93.3
227	56.4	32.2	96.6






#95
Naphthalene
Concen: 0.733 ug/l
RT: 13.780 min Scan# 2
Delta R.T. 0.006 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion:128 Resp: 310


Ion	Ratio	Lower	Upper
128	100		
127	13.0	10.3	15.5
129	11.7	8.6	12.8



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025



#96
1,2,3-Trichlorobenzene
Concen: 0.801 ug/l
RT: 13.969 min Scan# 2113
Delta R.T. 0.012 min
Lab File: VX045525.D
Acq: 01 Apr 2025 17:06

Tgt Ion:180 Resp: 956

Ion	Ratio	Lower	Upper
180	100		
182	80.5	47.3	141.8
145	40.4	18.2	54.6



Quantitation Report (QT Reviewed)

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045526.D
Acq On : 01 Apr 2025 17:29
Operator : JC/MD
Sample : VSTDICC005
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Quant Time: Apr 02 02:51:06 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 02:44:48 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED
Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	5.544	168	93234	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	166137	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	143937	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	63909	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	8969	5.260	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	10.520%#		
35) Dibromofluoromethane	5.391	113	6181	5.244	ug/l	0.01
Spiked Amount 50.000	Range 75 - 124		Recovery =	10.480%#		
50) Toluene-d8	8.647	98	20890	5.077	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	10.160%#		
62) 4-Bromofluorobenzene	11.079	95	6867	4.582	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	9.160%#		
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.166	85	6482	4.649	ug/l	91
3) Chloromethane	1.307	50	6846	4.779	ug/l	99
4) Vinyl Chloride	1.374	62	6169	4.706	ug/l	100
5) Bromomethane	1.593	94	3185	5.124	ug/l	90
6) Chloroethane	1.666	64	3703	5.324	ug/l	96
7) Trichlorofluoromethane	1.874	101	9316	4.766	ug/l	94
8) Diethyl Ether	2.136	74	3171	4.832	ug/l	96
9) 1,1,2-Trichlorotrifluo...	2.325	101	5589	4.879	ug/l	99
10) Methyl Iodide	2.447	142	6574	4.601	ug/l	97
11) Tert butyl alcohol	2.983	59	5186	22.633	ug/l	99
12) 1,1-Dichloroethene	2.313	96	5485	4.901	ug/l	90
13) Acrolein	2.239	56	8202	25.993	ug/l	97
14) Allyl chloride	2.654	41	9960	4.690	ug/l	95
15) Acrylonitrile	3.069	53	17811	24.624	ug/l	97
16) Acetone	2.386	43	17188	24.550	ug/l	99
17) Carbon Disulfide	2.502	76	12376	4.477	ug/l #	93
18) Methyl Acetate	2.703	43	8042	4.989	ug/l	96
19) Methyl tert-butyl Ether	3.111	73	18314	4.677	ug/l	100
20) Methylene Chloride	2.782	84	6481	4.945	ug/l	96
21) trans-1,2-Dichloroethene	3.087	96	5534	4.840	ug/l	95
22) Diisopropyl ether	3.764	45	19765	4.725	ug/l #	82
23) Vinyl Acetate	3.721	43	78907	21.831	ug/l	98
24) 1,1-Dichloroethane	3.605	63	11562	4.887	ug/l	98
25) 2-Butanone	4.574	43	25026	24.419	ug/l	99
26) 2,2-Dichloropropane	4.477	77	7054	4.615	ug/l	95
27) cis-1,2-Dichloroethene	4.483	96	6485	4.655	ug/l	95
28) Bromochloromethane	4.898	49	5666	4.956	ug/l	95
29) Tetrahydrofuran	5.013	42	15913	23.950	ug/l	100
30) Chloroform	5.087	83	12200	5.012	ug/l	99
31) Cyclohexane	5.465	56	9731	4.653	ug/l	97
32) 1,1,1-Trichloroethane	5.379	97	9811	4.763	ug/l	99
36) 1,1-Dichloropropene	5.690	75	7354	4.621	ug/l	98
37) Ethyl Acetate	4.721	43	9384	4.679	ug/l #	91
38) Carbon Tetrachloride	5.666	117	8264	4.804	ug/l	99
39) Methylcyclohexane	7.379	83	8757	4.489	ug/l	95
40) Benzene	6.038	78	23522	4.839	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045526.D
Acq On : 01 Apr 2025 17:29
Operator : JC/MD
Sample : VSTDICC005
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Quant Time: Apr 02 02:51:06 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 02:44:48 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
41) Methacrylonitrile	4.922	41	4593	4.302	ug/l		92
42) 1,2-Dichloroethane	6.092	62	9713	4.838	ug/l		100
43) Isopropyl Acetate	6.348	43	13803	4.541	ug/l		97
44) Trichloroethene	7.129	130	5355	4.635	ug/l		99
45) 1,2-Dichloropropane	7.434	63	6060	4.997	ug/l		99
46) Dibromomethane	7.580	93	4767	5.120	ug/l		99
47) Bromodichloromethane	7.818	83	8619	4.636	ug/l	#	97
48) Methyl methacrylate	7.696	41	6908	4.401	ug/l		98
49) 1,4-Dioxane	7.665	88	3205	112.864	ug/l		95
51) 4-Methyl-2-Pentanone	8.574	43	48050	23.844	ug/l		97
52) Toluene	8.714	92	14391	4.893	ug/l		100
53) t-1,3-Dichloropropene	8.982	75	6643	5.834	ug/l		97
54) cis-1,3-Dichloropropene	8.366	75	7873	4.506	ug/l		95
55) 1,1,2-Trichloroethane	9.153	97	5741	4.898	ug/l		90
56) Ethyl methacrylate	9.116	69	8012	4.417	ug/l		95
57) 1,3-Dichloropropane	9.305	76	10208	5.002	ug/l		99
58) 2-Chloroethyl Vinyl ether	8.244	63	21382	23.302	ug/l		99
59) 2-Hexanone	9.433	43	35628	23.875	ug/l		99
60) Dibromochloromethane	9.519	129	5855	4.582	ug/l		99
61) 1,2-Dibromoethane	9.610	107	5837	4.920	ug/l		98
64) Tetrachloroethene	9.275	164	5373	5.287	ug/l		93
65) Chlorobenzene	10.073	112	15178	4.935	ug/l		96
66) 1,1,1,2-Tetrachloroethane	10.159	131	4945	4.631	ug/l		97
67) Ethyl Benzene	10.195	91	26180	4.753	ug/l		97
68) m/p-Xylenes	10.299	106	19268	9.617	ug/l		99
69) o-Xylene	10.640	106	9735	4.932	ug/l		99
70) Styrene	10.653	104	15017	4.610	ug/l		96
71) Bromoform	10.799	173	3634	4.558	ug/l	#	95
73) Isopropylbenzene	10.964	105	24604	4.806	ug/l		100
74) N-amyl acetate	10.848	43	10750	4.390	ug/l		99
75) 1,1,2,2-Tetrachloroethane	11.213	83	9127	5.075	ug/l		98
76) 1,2,3-Trichloropropane	11.238	75	8032m	5.152	ug/l		
77) Bromobenzene	11.195	156	5927	5.011	ug/l		97
78) n-propylbenzene	11.305	91	27777	4.711	ug/l		98
79) 2-Chlorotoluene	11.360	91	18358	4.870	ug/l		99
80) 1,3,5-Trimethylbenzene	11.451	105	19936	4.710	ug/l		98
81) trans-1,4-Dichloro-2-b...	11.018	75	1752	3.814	ug/l		85
82) 4-Chlorotoluene	11.457	91	20171	4.799	ug/l		99
83) tert-Butylbenzene	11.713	119	19962	4.763	ug/l		98
84) 1,2,4-Trimethylbenzene	11.750	105	19998	4.707	ug/l		99
85) sec-Butylbenzene	11.890	105	24465	4.753	ug/l		100
86) p-Isopropyltoluene	12.006	119	20178	4.755	ug/l		99
87) 1,3-Dichlorobenzene	11.969	146	10255	4.763	ug/l		98
88) 1,4-Dichlorobenzene	12.036	146	11016	5.047	ug/l		93
89) n-Butylbenzene	12.329	91	15887	4.317	ug/l		98
90) Hexachloroethane	12.536	117	3151	4.320	ug/l		98
91) 1,2-Dichlorobenzene	12.335	146	10510	4.910	ug/l		98
92) 1,2-Dibromo-3-Chloropr...	12.945	75	1661	4.417	ug/l		92
93) 1,2,4-Trichlorobenzene	13.591	180	5391	4.524	ug/l		95
94) Hexachlorobutadiene	13.719	225	2483	4.836	ug/l		98
95) Naphthalene	13.774	128	19618	4.424	ug/l		97
96) 1,2,3-Trichlorobenzene	13.957	180	5854	4.692	ug/l		97

ICAL Form

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045526.D
Acq On : 01 Apr 2025 17:29
Operator : JC/MD
Sample : VSTDICC005
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Quant Time: Apr 02 02:51:06 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 02:44:48 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045526.D
Acq On : 01 Apr 2025 17:29
Operator : JC/MD
Sample : VSTDICC005
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1


Quant Time: Apr 02 02:51:06 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 02:44:48 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025






#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 7
Delta R.T. -0.006 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion:168 Resp: 9323


Ion	Ratio	Lower	Upper
168	100		
99	68.8	52.3	78.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#2
Dichlorodifluoromethane
Concen: 4.649 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 85 Resp: 6482

Ion	Ratio	Lower	Upper
85	100		
87	27.3	16.1	48.2






#3
Chloromethane
Concen: 4.779 ug/l
RT: 1.307 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 50 Resp: 684


Ion	Ratio	Lower	Upper
50	100		
52	32.8	25.9	38.9



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#4
Vinyl Chloride
Concen: 4.706 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 62 Resp: 6169

Ion	Ratio	Lower	Upper
62	100		
64	31.9	25.5	38.3






#5
Bromomethane
Concen: 5.124 ug/l
RT: 1.593 min Scan# 8
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 94 Resp: 318


Ion	Ratio	Lower	Upper
94	100		
96	83.6	74.8	112.2



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#6
Chloroethane
Concen: 5.324 ug/l
RT: 1.666 min Scan# 95
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 64 Resp: 3703

Ion	Ratio	Lower	Upper
64	100		
66	30.9	26.3	39.5






#7
Trichlorofluoromethane
Concen: 4.766 ug/l
RT: 1.874 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion:101 Resp: 931


Ion	Ratio	Lower	Upper
101	100		
103	68.2	51.1	76.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#8
Diethyl Ether
Concen: 4.832 ug/l
RT: 2.136 min Scan# 172
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 74 Resp: 3171

Ion	Ratio	Lower	Upper
74	100		
45	110.2	53.0	159.1






#9
1,1,2-Trichlorotrifluoroethane
Concen: 4.879 ug/l
RT: 2.325 min Scan# 2
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion:101 Resp: 558


Ion	Ratio	Lower	Upper
101	100		
85	47.8	38.9	58.3
151	72.4	57.5	86.3



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#10
Methyl Iodide
Concen: 4.601 ug/l
RT: 2.447 min Scan# 223
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion:142 Resp: 6574

Ion	Ratio	Lower	Upper
142	100		
127	48.4	40.2	60.4
141	13.7	11.7	17.5






#11
Tert butyl alcohol
Concen: 22.633 ug/l
RT: 2.983 min Scan# 3
Delta R.T. 0.006 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 59 Resp: 518


Ion	Ratio	Lower	Upper
59	100		
57	11.2	8.6	12.8



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#12
1,1-Dichloroethene
Concen: 4.901 ug/l
RT: 2.313 min Scan# 201
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 96 Resp: 5485

Ion	Ratio	Lower	Upper
96	100		
61	164.9	145.7	218.5
98	60.8	51.6	77.4






#13
Acrolein
Concen: 25.993 ug/l
RT: 2.239 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 56 Resp: 820


Ion	Ratio	Lower	Upper
56	100		
55	67.9	56.2	84.2



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#14
Allyl chloride
Concen: 4.690 ug/l
RT: 2.654 min Scan# 257
Delta R.T. -0.006 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 41 Resp: 9960

Ion	Ratio	Lower	Upper
41	100		
39	79.8	60.2	90.2
76	28.8	24.6	36.8





#15
Acrylonitrile
Concen: 24.624 ug/l
RT: 3.069 min Scan# 3
Delta R.T. 0.006 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29


Tgt Ion: 53 Resp: 1781


Ion	Ratio	Lower	Upper
53	100		
52	82.1	67.4	101.2
51	39.3	29.9	44.9

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025







#16
Acetone
Concen: 24.550 ug/l
RT: 2.386 min Scan# 213
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 43 Resp: 17188

Ion	Ratio	Lower	Upper
43	100		
58	28.4	22.3	33.5






#17
Carbon Disulfide
Concen: 4.477 ug/l
RT: 2.502 min Scan# 2
Delta R.T. -0.006 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 76 Resp: 1237


Ion	Ratio	Lower	Upper
76	100		
78	11.4	7.0	10.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#18
Methyl Acetate
Concen: 4.989 ug/l
RT: 2.703 min Scan# 265
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 43 Resp: 8042

Ion	Ratio	Lower	Upper
43	100		
74	20.2	17.5	26.3






#19
Methyl tert-butyl Ether
Concen: 4.677 ug/l
RT: 3.111 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 73 Resp: 1831


Ion	Ratio	Lower	Upper
73	100		
57	21.9	17.6	26.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#20
Methylene Chloride
Concen: 4.945 ug/l
RT: 2.782 min Scan# 278
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 84 Resp: 6481

Ion	Ratio	Lower	Upper
84	100		
49	139.6	107.5	161.3
51	40.9	33.0	49.6
86	59.1	50.1	75.1






#21
trans-1,2-Dichloroethene
Concen: 4.840 ug/l
RT: 3.087 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 96 Resp: 553


Ion	Ratio	Lower	Upper
96	100		
61	144.4	121.0	181.6
98	58.5	49.7	74.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#22
Diisopropyl ether
Concen: 4.725 ug/l
RT: 3.764 min Scan# 439
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 45 Resp: 19765

Ion	Ratio	Lower	Upper
45	100		
43	94.7	59.3	88.9#
87	27.8	20.4	30.6
59	10.3	9.0	13.6





#23
Vinyl Acetate
Concen: 21.831 ug/l
RT: 3.721 min Scan# 4
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29


Tgt Ion: 43 Resp: 7890


Ion	Ratio	Lower	Upper
43	100		
86	8.7	7.5	11.3

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025







#24
1,1-Dichloroethane
Concen: 4.887 ug/l
RT: 3.605 min Scan# 413
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 63 Resp: 11562

Ion	Ratio	Lower	Upper
63	100		
98	6.3	3.4	10.1
100	3.6	2.2	6.6






#25
2-Butanone
Concen: 24.419 ug/l
RT: 4.574 min Scan# 5
Delta R.T. 0.018 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 43 Resp: 2502


Ion	Ratio	Lower	Upper
43	100		
72	23.8	18.6	28.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#26
2,2-Dichloropropane
Concen: 4.615 ug/l
RT: 4.477 min Scan# 556
Delta R.T. 0.006 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 77 Resp: 7054

Ion	Ratio	Lower	Upper
77	100		
97	23.4	10.6	31.8






#27
cis-1,2-Dichloroethene
Concen: 4.655 ug/l
RT: 4.483 min Scan# 5
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 96 Resp: 648


Ion	Ratio	Lower	Upper
96	100		
61	162.4	0.0	313.2
98	68.5	0.0	126.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#28
Bromochloromethane
Concen: 4.956 ug/l
RT: 4.898 min Scan# 625
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 49 Resp: 5666

Ion	Ratio	Lower	Upper
49	100		
129	1.2	0.0	3.8
130	72.2	54.6	81.8





#29
Tetrahydrofuran
Concen: 23.950 ug/l
RT: 5.013 min Scan# 64
Delta R.T. 0.012 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29


Tgt Ion: 42 Resp: 1591


Ion	Ratio	Lower	Upper
42	100		
72	41.9	33.4	50.0
71	39.2	31.4	47.0

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025







#30
Chloroform
Concen: 5.012 ug/l
RT: 5.087 min Scan# 656
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 83 Resp: 12200

Ion	Ratio	Lower	Upper
83	100		
85	65.0	52.8	79.2






#31
Cyclohexane
Concen: 4.653 ug/l
RT: 5.465 min Scan# 7
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 56 Resp: 973


Ion	Ratio	Lower	Upper
56	100		
69	31.5	23.6	35.4
84	82.2	67.7	101.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#32
1,1,1-Trichloroethane
Concen: 4.763 ug/l
RT: 5.379 min Scan# 704
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 97 Resp: 9811

Ion	Ratio	Lower	Upper
97	100		
99	63.2	51.7	77.5
61	52.0	41.0	61.4






#33
1,2-Dichloroethane-d4
Concen: 5.260 ug/l
RT: 5.958 min Scan# 7
Delta R.T. 0.006 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 65 Resp: 896


Ion	Ratio	Lower	Upper
65	100		
67	46.7	0.0	99.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.757 min Scan# 930
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion:114 Resp: 166137

Ion	Ratio	Lower	Upper
114	100		
63	24.4	0.0	46.8
88	16.1	0.0	35.4






#35
Dibromofluoromethane
Concen: 5.244 ug/l
RT: 5.391 min Scan# 7
Delta R.T. 0.012 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion:113 Resp: 618


Ion	Ratio	Lower	Upper
113	100		
111	98.8	81.8	122.6
192	14.4	13.8	20.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#36
1,1-Dichloropropene
Concen: 4.621 ug/l
RT: 5.690 min Scan# 755
Delta R.T. 0.006 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 75 Resp: 7354

Ion	Ratio	Lower	Upper
75	100		
110	33.3	16.9	50.6
77	29.8	25.0	37.4






#37
Ethyl Acetate
Concen: 4.679 ug/l
RT: 4.721 min Scan# 5
Delta R.T. 0.006 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 43 Resp: 938


Ion	Ratio	Lower	Upper
43	100		
61	6.3	9.9	14.9
70	8.9	7.5	11.3



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#38
Carbon Tetrachloride
Concen: 4.804 ug/l
RT: 5.666 min Scan# 751
Delta R.T. -0.006 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion:117 Resp: 8264

Ion	Ratio	Lower	Upper
117	100		
119	96.4	76.9	115.3
121	31.7	24.2	36.2






#39
Methylcyclohexane
Concen: 4.489 ug/l
RT: 7.379 min Scan# 1
Delta R.T. 0.006 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 83 Resp: 875


Ion	Ratio	Lower	Upper
83	100		
55	84.0	64.9	97.3
98	52.1	37.4	56.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#40
Benzene
Concen: 4.839 ug/l
RT: 6.038 min Scan# 812
Delta R.T. 0.006 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 78 Resp: 23522

Ion	Ratio	Lower	Upper
78	100		
77	23.6	19.0	28.6






#41
Methacrylonitrile
Concen: 4.302 ug/l
RT: 4.922 min Scan# 6
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 41 Resp: 459


Ion	Ratio	Lower	Upper
41	100		
39	61.4	47.0	70.6
67	75.2	53.2	79.8
52	37.4	25.0	37.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#42
1,2-Dichloroethane
Concen: 4.838 ug/l
RT: 6.092 min Scan# 821
Delta R.T. 0.012 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 62 Resp: 9713

Ion	Ratio	Lower	Upper
62	100		
98	7.7	0.0	15.2






#43
Isopropyl Acetate
Concen: 4.541 ug/l
RT: 6.348 min Scan# 8
Delta R.T. 0.012 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 43 Resp: 1380


Ion	Ratio	Lower	Upper
43	100		
61	15.8	14.1	21.1
87	11.3	9.2	13.8



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#44
Trichloroethene
Concen: 4.635 ug/l
RT: 7.129 min Scan# 991
Delta R.T. 0.006 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion:130 Resp: 5355

Ion	Ratio	Lower	Upper
130	100		
95	102.2	0.0	202.0






#45
1,2-Dichloropropane
Concen: 4.997 ug/l
RT: 7.434 min Scan# 1
Delta R.T. 0.006 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 63 Resp: 606


Ion	Ratio	Lower	Upper
63	100		
65	31.5	25.8	38.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#46
Dibromomethane
Concen: 5.120 ug/l
RT: 7.580 min Scan# 1065
Delta R.T. 0.006 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 93 Resp: 4767

Ion	Ratio	Lower	Upper
93	100		
95	81.0	64.8	97.2
174	83.1	65.4	98.0






#47
Bromodichloromethane
Concen: 4.636 ug/l
RT: 7.818 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 83 Resp: 861


Ion	Ratio	Lower	Upper
83	100		
85	64.1	49.7	74.5
127	6.0	6.1	9.1



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#48
Methyl methacrylate
Concen: 4.401 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. 0.006 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 41 Resp: 6908

Ion	Ratio	Lower	Upper
41	100		
69	74.6	58.2	87.2
39	68.3	53.6	80.4






#49
1,4-Dioxane
Concen: 112.864 ug/l
RT: 7.665 min Scan# 1
Delta R.T. 0.006 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 88 Resp: 320


Ion	Ratio	Lower	Upper
88	100		
43	32.4	32.2	48.4
58	72.7	58.6	88.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#50
Toluene-d8
Concen: 5.077 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 98 Resp: 20890

Ion	Ratio	Lower	Upper
98	100		
100	63.5	52.2	78.4






#51
4-Methyl-2-Pentanone
Concen: 23.844 ug/l
RT: 8.574 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 43 Resp: 4805


Ion	Ratio	Lower	Upper
43	100		
58	37.3	28.5	42.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#52
Toluene
Concen: 4.893 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 92 Resp: 14391

Ion	Ratio	Lower	Upper
92	100		
91	173.1	138.7	208.1






#53
t-1,3-Dichloropropene
Concen: 5.834 ug/l
RT: 8.982 min Scan# 1
Delta R.T. 0.006 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 75 Resp: 664


Ion	Ratio	Lower	Upper
75	100		
77	30.1	25.4	38.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#54
cis-1,3-Dichloropropene
Concen: 4.506 ug/l
RT: 8.366 min Scan# 1194
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 75 Resp: 7873

Ion	Ratio	Lower	Upper
75	100		
77	28.6	24.6	36.8
39	64.4	48.4	72.6






#55
1,1,2-Trichloroethane
Concen: 4.898 ug/l
RT: 9.153 min Scan# 1
Delta R.T. 0.006 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 97 Resp: 574


Ion	Ratio	Lower	Upper
97	100		
83	96.8	71.8	107.8
85	52.9	47.0	70.6
99	72.3	48.6	73.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#56
Ethyl methacrylate
Concen: 4.417 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 69 Resp: 8012

Ion	Ratio	Lower	Upper
69	100		
41	75.8	61.3	91.9
39	58.5	40.2	60.2






#57
1,3-Dichloropropane
Concen: 5.002 ug/l
RT: 9.305 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 76 Resp: 1020


Ion	Ratio	Lower	Upper
76	100		
78	33.1	26.0	39.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#58
2-Chloroethyl Vinyl ether
Concen: 23.302 ug/l
RT: 8.244 min Scan# 1174
Delta R.T. 0.006 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 63 Resp: 21382

Ion	Ratio	Lower	Upper
63	100		
106	25.8	21.0	31.6






#59
2-Hexanone
Concen: 23.875 ug/l
RT: 9.433 min Scan# 1
Delta R.T. 0.006 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 43 Resp: 3562


Ion	Ratio	Lower	Upper
43	100		
58	48.7	24.9	74.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#60
Dibromochloromethane
Concen: 4.582 ug/l
RT: 9.519 min Scan# 1383
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion:129 Resp: 5855

Ion	Ratio	Lower	Upper
129	100		
127	77.5	39.4	118.2






#61
1,2-Dibromoethane
Concen: 4.920 ug/l
RT: 9.610 min Scan# 1
Delta R.T. 0.006 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion:107 Resp: 583


Ion	Ratio	Lower	Upper
107	100		
109	93.0	75.7	113.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#62
4-Bromofluorobenzene
Concen: 4.582 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 95 Resp: 6867

Ion	Ratio	Lower	Upper
95	100		
174	73.1	0.0	135.8
176	73.3	0.0	131.4






#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion:117 Resp: 14393


Ion	Ratio	Lower	Upper
117	100		
82	62.3	49.2	73.8
119	32.6	25.1	37.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#64
Tetrachloroethene
Concen: 5.287 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. 0.006 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion:164 Resp: 5373

Ion	Ratio	Lower	Upper
164	100		
166	127.6	103.8	155.8
129	98.6	84.3	126.5
131	87.3	80.9	121.3





#65
Chlorobenzene
Concen: 4.935 ug/l
RT: 10.073 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29


Tgt Ion:112 Resp: 1517


Ion	Ratio	Lower	Upper
112	100		
114	28.5	24.7	37.1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025







#66
1,1,1,2-Tetrachloroethane
Concen: 4.631 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion:131 Resp: 4945

Ion	Ratio	Lower	Upper
131	100		
133	95.6	47.5	142.5
119	70.0	32.9	98.6





#67
Ethyl Benzene
Concen: 4.753 ug/l
RT: 10.195 min Scan# 14
Delta R.T. 0.006 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29


Tgt Ion: 91 Resp: 2618


Ion	Ratio	Lower	Upper
91	100		
106	31.4	23.7	35.5

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025







#68
m/p-Xylenes
Concen: 9.617 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion:106 Resp: 19268

Ion	Ratio	Lower	Upper
106	100		
91	213.6	172.6	258.8






#69
o-Xylene
Concen: 4.932 ug/l
RT: 10.640 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion:106 Resp: 973


Ion	Ratio	Lower	Upper
106	100		
91	225.8	113.6	340.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#70
Styrene
Concen: 4.610 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion:104 Resp: 15017

Ion	Ratio	Lower	Upper
104	100		
78	59.8	45.4	68.2
103	57.1	43.8	65.8





#71
Bromoform
Concen: 4.558 ug/l
RT: 10.799 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29


Tgt Ion:173 Resp: 363


Ion	Ratio	Lower	Upper
173	100		
175	43.7	23.4	70.2
254	0.0	0.0	0.0

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025







#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion:152 Resp: 63909

Ion	Ratio	Lower	Upper
152	100		
115	66.5	46.9	140.7
150	157.5	0.0	349.4





#73
Isopropylbenzene
Concen: 4.806 ug/l
RT: 10.964 min Scan# 1
Delta R.T. 0.006 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29


Tgt Ion:105 Resp: 2460


Ion	Ratio	Lower	Upper
105	100		
120	25.5	12.7	38.0

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025







#74
N-amyl acetate
Concen: 4.390 ug/l
RT: 10.848 min Scan# 1601
Delta R.T. 0.006 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 43 Resp: 10750

Ion	Ratio	Lower	Upper
43	100		
70	37.8	30.1	45.1
55	23.0	18.3	27.5
61	20.1	17.5	26.3






#75
1,1,2,2-Tetrachloroethane
Concen: 5.075 ug/l
RT: 11.213 min Scan# 1
Delta R.T. 0.006 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 83 Resp: 912


Ion	Ratio	Lower	Upper
83	100		
131	10.2	4.7	14.1
85	63.5	32.4	97.2



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#76
1,2,3-Trichloropropane
Concen: 5.152 ug/l m
RT: 11.238 min Scan# 1665
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 75 Resp: 8032

Ion	Ratio	Lower	Upper
75	100		
77	40.2	20.0	59.9






#77
Bromobenzene
Concen: 5.011 ug/l
RT: 11.195 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion:156 Resp: 592


Ion	Ratio	Lower	Upper
156	100		
77	178.8	92.6	277.8
158	97.0	48.1	144.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#78
n-propylbenzene
Concen: 4.711 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. 0.006 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 91 Resp: 27777

Ion	Ratio	Lower	Upper
91	100		
120	20.6	10.8	32.3






#79
2-Chlorotoluene
Concen: 4.870 ug/l
RT: 11.360 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 91 Resp: 1835


Ion	Ratio	Lower	Upper
91	100		
126	31.3	15.8	47.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#80
1,3,5-Trimethylbenzene
Concen: 4.710 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion:105 Resp: 19936

Ion	Ratio	Lower	Upper
105	100		
120	47.1	23.0	69.0





#81
trans-1,4-Dichloro-2-butene
Concen: 3.814 ug/l
RT: 11.018 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29


Tgt Ion: 75 Resp: 175


Ion	Ratio	Lower	Upper
75	100		
53	149.1	101.8	152.6
89	43.1	37.2	55.8

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025







#82
4-Chlorotoluene
Concen: 4.799 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.006 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 91 Resp: 20171

Ion	Ratio	Lower	Upper
91	100		
126	27.2	13.3	39.9






#83
tert-Butylbenzene
Concen: 4.763 ug/l
RT: 11.713 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion:119 Resp: 1996


Ion	Ratio	Lower	Upper
119	100		
91	65.7	32.4	97.2
134	24.6	11.5	34.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#84
1,2,4-Trimethylbenzene
Concen: 4.707 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion:105 Resp: 19998

Ion	Ratio	Lower	Upper
105	100		
120	43.5	21.4	64.3






#85
sec-Butylbenzene
Concen: 4.753 ug/l
RT: 11.890 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion:105 Resp: 2446


Ion	Ratio	Lower	Upper
105	100		
134	19.3	9.6	28.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#86
p-Isopropyltoluene
Concen: 4.755 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion:119 Resp: 20178

Ion	Ratio	Lower	Upper
119	100		
134	25.1	12.6	37.8
91	26.4	13.7	41.1






#87
1,3-Dichlorobenzene
Concen: 4.763 ug/l
RT: 11.969 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion:146 Resp: 1025


Ion	Ratio	Lower	Upper
146	100		
111	42.9	22.2	66.6
148	65.7	31.9	95.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#88
1,4-Dichlorobenzene
Concen: 5.047 ug/l
RT: 12.036 min Scan# 1796
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion:146 Resp: 11016

Ion	Ratio	Lower	Upper
146	100		
111	51.4	20.9	62.8
148	64.2	31.5	94.5






#89
n-Butylbenzene
Concen: 4.317 ug/l
RT: 12.329 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 91 Resp: 1588


Ion	Ratio	Lower	Upper
91	100		
92	52.1	26.9	80.6
134	24.4	11.9	35.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#90
Hexachloroethane
Concen: 4.320 ug/l
RT: 12.536 min Scan# 1878
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion:117 Resp: 3151

Ion	Ratio	Lower	Upper
117	100		
201	60.6	31.2	93.6






#91
1,2-Dichlorobenzene
Concen: 4.910 ug/l
RT: 12.335 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion:146 Resp: 1051


Ion	Ratio	Lower	Upper
146	100		
111	45.2	22.9	68.8
148	66.3	31.9	95.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#92
1,2-Dibromo-3-Chloropropane
Concen: 4.417 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. 0.006 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion: 75 Resp: 1661

Ion	Ratio	Lower	Upper
75	100		
155	67.3	34.9	104.7
157	100.4	44.9	134.7






#93
1,2,4-Trichlorobenzene
Concen: 4.524 ug/l
RT: 13.591 min Scan# 2
Delta R.T. 0.006 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion:180 Resp: 539


Ion	Ratio	Lower	Upper
180	100		
182	91.7	47.9	143.6
145	30.7	17.9	53.8



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#94
Hexachlorobutadiene
Concen: 4.836 ug/l
RT: 13.719 min Scan# 2072
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion:225 Resp: 2483

Ion	Ratio	Lower	Upper
225	100		
223	64.4	31.1	93.3
227	62.8	32.2	96.6






#95
Naphthalene
Concen: 4.424 ug/l
RT: 13.774 min Scan# 2
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion:128 Resp: 1961


Ion	Ratio	Lower	Upper
128	100		
127	11.5	10.3	15.5
129	10.2	8.6	12.8



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025



#96
1,2,3-Trichlorobenzene
Concen: 4.692 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. 0.000 min
Lab File: VX045526.D
Acq: 01 Apr 2025 17:29

Tgt Ion:180 Resp: 5854

Ion	Ratio	Lower	Upper
180	100		
182	90.7	47.3	141.8
145	35.2	18.2	54.6



Quantitation Report (QT Reviewed)

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045527.D
Acq On : 01 Apr 2025 17:52
Operator : JC/MD
Sample : VSTDICC020
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Quant Time: Apr 02 02:52:02 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 02:44:48 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED
Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	5.550	168	90228	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	160754	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	144087	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	65341	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	32500	19.697	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 39.400%#			
35) Dibromofluoromethane	5.385	113	21963	19.256	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 38.520%#			
50) Toluene-d8	8.647	98	79303	19.921	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 39.840%#			
62) 4-Bromofluorobenzene	11.079	95	28795	19.858	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 39.720%#			
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.166	85	25265	18.724	ug/l	97
3) Chloromethane	1.301	50	28034	20.223	ug/l	98
4) Vinyl Chloride	1.374	62	25285	19.931	ug/l	97
5) Bromomethane	1.593	94	11815	19.642	ug/l	99
6) Chloroethane	1.666	64	14071	20.905	ug/l	98
7) Trichlorofluoromethane	1.874	101	38815	20.518	ug/l	98
8) Diethyl Ether	2.136	74	12852	20.235	ug/l	99
9) 1,1,2-Trichlorotrifluo...	2.319	101	22917	20.673	ug/l	95
10) Methyl Iodide	2.441	142	28478	20.596	ug/l	99
11) Tert butyl alcohol	2.983	59	22453	101.253	ug/l	100
12) 1,1-Dichloroethene	2.306	96	22103	20.406	ug/l	96
13) Acrolein	2.239	56	28991	94.938	ug/l	99
14) Allyl chloride	2.660	41	41233	20.063	ug/l	99
15) Acrylonitrile	3.062	53	74487	106.412	ug/l	98
16) Acetone	2.386	43	69788	103.000	ug/l	99
17) Carbon Disulfide	2.502	76	51771	19.352	ug/l	98
18) Methyl Acetate	2.703	43	31190	19.994	ug/l	99
19) Methyl tert-butyl Ether	3.117	73	78284	20.660	ug/l	96
20) Methylene Chloride	2.782	84	26212	20.664	ug/l	96
21) trans-1,2-Dichloroethene	3.087	96	22968	20.755	ug/l	92
22) Diisopropyl ether	3.757	45	84046	20.760	ug/l #	81
23) Vinyl Acetate	3.721	43	361881	103.458	ug/l	99
24) 1,1-Dichloroethane	3.605	63	47560	20.772	ug/l	99
25) 2-Butanone	4.556	43	105062	105.931	ug/l	99
26) 2,2-Dichloropropane	4.471	77	28956	19.574	ug/l	99
27) cis-1,2-Dichloroethene	4.483	96	27425	20.342	ug/l	98
28) Bromochloromethane	4.897	49	22285	20.141	ug/l	100
29) Tetrahydrofuran	5.007	42	66994	104.187	ug/l	99
30) Chloroform	5.093	83	48639	20.646	ug/l	100
31) Cyclohexane	5.458	56	41450	20.478	ug/l	99
32) 1,1,1-Trichloroethane	5.373	97	40407	20.270	ug/l	99
36) 1,1-Dichloropropene	5.684	75	30504	19.808	ug/l	98
37) Ethyl Acetate	4.715	43	39163	20.180	ug/l	98
38) Carbon Tetrachloride	5.672	117	33533	20.148	ug/l	94
39) Methylcyclohexane	7.373	83	38303	20.291	ug/l	97
40) Benzene	6.031	78	97667	20.766	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045527.D
Acq On : 01 Apr 2025 17:52
Operator : JC/MD
Sample : VSTDICC020
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Quant Time: Apr 02 02:52:02 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 02:44:48 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
41) Methacrylonitrile	4.922	41	21740	21.046	ug/l		97
42) 1,2-Dichloroethane	6.086	62	41733	21.484	ug/l		100
43) Isopropyl Acetate	6.342	43	61222	20.814	ug/l		99
44) Trichloroethene	7.123	130	22918	20.501	ug/l		97
45) 1,2-Dichloropropane	7.428	63	24926	21.241	ug/l		90
46) Dibromomethane	7.580	93	19104	21.207	ug/l		99
47) Bromodichloromethane	7.818	83	37036	20.587	ug/l		99
48) Methyl methacrylate	7.696	41	31247	20.575	ug/l		99
49) 1,4-Dioxane	7.659	88	12104	440.515	ug/l		96
51) 4-Methyl-2-Pentanone	8.574	43	212645	109.055	ug/l		99
52) Toluene	8.714	92	60357	21.208	ug/l		99
53) t-1,3-Dichloropropene	8.976	75	29870	18.764	ug/l		96
54) cis-1,3-Dichloropropene	8.366	75	35099	20.762	ug/l		95
55) 1,1,2-Trichloroethane	9.147	97	24191	21.331	ug/l		96
56) Ethyl methacrylate	9.116	69	36155	20.600	ug/l		99
57) 1,3-Dichloropropane	9.305	76	41900	21.217	ug/l		100
58) 2-Chloroethyl Vinyl ether	8.238	63	91923	103.534	ug/l		100
59) 2-Hexanone	9.427	43	156999	108.730	ug/l		99
60) Dibromochloromethane	9.519	129	25992	21.023	ug/l		98
61) 1,2-Dibromoethane	9.604	107	24456	21.303	ug/l		99
64) Tetrachloroethene	9.269	164	21379	21.015	ug/l		98
65) Chlorobenzene	10.079	112	64727	21.023	ug/l		99
66) 1,1,1,2-Tetrachloroethane	10.159	131	21447	20.064	ug/l		98
67) Ethyl Benzene	10.189	91	115368	20.921	ug/l		99
68) m/p-Xylenes	10.299	106	84420	42.092	ug/l		99
69) o-Xylene	10.640	106	41803	21.158	ug/l		100
70) Styrene	10.653	104	69290	21.249	ug/l		99
71) Bromoform	10.799	173	15659	19.619	ug/l	#	99
73) Isopropylbenzene	10.957	105	110396	21.093	ug/l		100
74) N-amyl acetate	10.842	43	51238	20.466	ug/l		99
75) 1,1,2,2-Tetrachloroethane	11.207	83	38177	20.764	ug/l		99
76) 1,2,3-Trichloropropane	11.238	75	32690m	20.510	ug/l		
77) Bromobenzene	11.195	156	24902	20.592	ug/l		100
78) n-propylbenzene	11.299	91	130994	21.729	ug/l		98
79) 2-Chlorotoluene	11.360	91	81817	21.229	ug/l		99
80) 1,3,5-Trimethylbenzene	11.451	105	93328	21.564	ug/l		99
81) trans-1,4-Dichloro-2-b...	11.018	75	8435	17.961	ug/l		88
82) 4-Chlorotoluene	11.451	91	90704	21.108	ug/l		99
83) tert-Butylbenzene	11.713	119	89828	20.964	ug/l		98
84) 1,2,4-Trimethylbenzene	11.750	105	92242	21.235	ug/l		100
85) sec-Butylbenzene	11.890	105	112800	21.433	ug/l		100
86) p-Isopropyltoluene	12.006	119	91866	21.173	ug/l		99
87) 1,3-Dichlorobenzene	11.969	146	45105	20.490	ug/l		100
88) 1,4-Dichlorobenzene	12.036	146	46208	20.706	ug/l		99
89) n-Butylbenzene	12.329	91	77739	20.662	ug/l		100
90) Hexachloroethane	12.536	117	14544	19.505	ug/l		100
91) 1,2-Dichlorobenzene	12.335	146	45732	20.896	ug/l		99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	8153	21.203	ug/l		100
93) 1,2,4-Trichlorobenzene	13.585	180	24787	20.347	ug/l		98
94) Hexachlorobutadiene	13.725	225	10488	19.979	ug/l		97
95) Naphthalene	13.774	128	94875	20.928	ug/l		100
96) 1,2,3-Trichlorobenzene	13.957	180	26113	20.471	ug/l		98

ICAL Form

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045527.D
Acq On : 01 Apr 2025 17:52
Operator : JC/MD
Sample : VSTDICC020
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025

Quant Time: Apr 02 02:52:02 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 02:44:48 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045527.D
Acq On : 01 Apr 2025 17:52
Operator : JC/MD
Sample : VSTDICC020
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1


Quant Time: Apr 02 02:52:02 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 02:44:48 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025






#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 7
Delta R.T. -0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion:168 Resp: 9022


Ion	Ratio	Lower	Upper
168	100		
99	69.1	52.3	78.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#2
Dichlorodifluoromethane
Concen: 18.724 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 85 Resp: 25265

Ion	Ratio	Lower	Upper
85	100		
87	33.7	16.1	48.2





#3
Chloromethane
Concen: 20.223 ug/l
RT: 1.301 min Scan# 3
Delta R.T. -0.006 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52


Tgt Ion: 50 Resp: 2803


Ion	Ratio	Lower	Upper
50	100		
52	31.4	25.9	38.9

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025







#4
Vinyl Chloride
Concen: 19.931 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 62 Resp: 25285

Ion	Ratio	Lower	Upper
62	100		
64	33.5	25.5	38.3






#5
Bromomethane
Concen: 19.642 ug/l
RT: 1.593 min Scan# 8
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 94 Resp: 1181


Ion	Ratio	Lower	Upper
94	100		
96	92.7	74.8	112.2



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#6
Chloroethane
Concen: 20.905 ug/l
RT: 1.666 min Scan# 95
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 64 Resp: 14071

Ion	Ratio	Lower	Upper
64	100		
66	31.9	26.3	39.5






#7
Trichlorofluoromethane
Concen: 20.518 ug/l
RT: 1.874 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion:101 Resp: 3881


Ion	Ratio	Lower	Upper
101	100		
103	65.2	51.1	76.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#8
Diethyl Ether
Concen: 20.235 ug/l
RT: 2.136 min Scan# 172
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 74 Resp: 12852

Ion	Ratio	Lower	Upper
74	100		
45	107.6	53.0	159.1






#9
1,1,2-Trichlorotrifluoroethane
Concen: 20.673 ug/l
RT: 2.319 min Scan# 2
Delta R.T. -0.006 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion:101 Resp: 2291


Ion	Ratio	Lower	Upper
101	100		
85	46.2	38.9	58.3
151	66.8	57.5	86.3



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#10
Methyl Iodide
Concen: 20.596 ug/l
RT: 2.441 min Scan# 222
Delta R.T. -0.006 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion:142 Resp: 28478

Ion	Ratio	Lower	Upper
142	100		
127	51.5	40.2	60.4
141	14.8	11.7	17.5






#11
Tert butyl alcohol
Concen: 101.253 ug/l
RT: 2.983 min Scan# 3
Delta R.T. 0.006 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 59 Resp: 2245


Ion	Ratio	Lower	Upper
59	100		
57	10.8	8.6	12.8



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#12
1,1-Dichloroethene
Concen: 20.406 ug/l
RT: 2.306 min Scan# 200
Delta R.T. -0.006 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 96 Resp: 22103

Ion	Ratio	Lower	Upper
96	100		
61	175.0	145.7	218.5
98	64.2	51.6	77.4






#13
Acrolein
Concen: 94.938 ug/l
RT: 2.239 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 56 Resp: 2899


Ion	Ratio	Lower	Upper
56	100		
55	71.2	56.2	84.2



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#14
Allyl chloride
Concen: 20.063 ug/l
RT: 2.660 min Scan# 258
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 41 Resp: 41233

Ion	Ratio	Lower	Upper
41	100		
39	74.2	60.2	90.2
76	31.0	24.6	36.8






#15
Acrylonitrile
Concen: 106.412 ug/l
RT: 3.062 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 53 Resp: 7448


Ion	Ratio	Lower	Upper
53	100		
52	81.8	67.4	101.2
51	37.6	29.9	44.9



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#16
Acetone
Concen: 103.000 ug/l
RT: 2.386 min Scan# 213
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 43 Resp: 69788

Ion	Ratio	Lower	Upper
43	100		
58	28.6	22.3	33.5





#17
Carbon Disulfide
Concen: 19.352 ug/l
RT: 2.502 min Scan# 2
Delta R.T. -0.006 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52


Tgt Ion: 76 Resp: 5177


Ion	Ratio	Lower	Upper
76	100		
78	9.5	7.0	10.6

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025







#18
Methyl Acetate
Concen: 19.994 ug/l
RT: 2.703 min Scan# 265
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 43 Resp: 31190

Ion	Ratio	Lower	Upper
43	100		
74	21.3	17.5	26.3






#19
Methyl tert-butyl Ether
Concen: 20.660 ug/l
RT: 3.117 min Scan# 3
Delta R.T. 0.006 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 73 Resp: 7828


Ion	Ratio	Lower	Upper
73	100		
57	20.0	17.6	26.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#20
Methylene Chloride
Concen: 20.664 ug/l
RT: 2.782 min Scan# 278
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 84 Resp: 26212

Ion	Ratio	Lower	Upper
84	100		
49	139.9	107.5	161.3
51	43.0	33.0	49.6
86	61.1	50.1	75.1






#21
trans-1,2-Dichloroethene
Concen: 20.755 ug/l
RT: 3.087 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 96 Resp: 2296


Ion	Ratio	Lower	Upper
96	100		
61	139.8	121.0	181.6
98	56.7	49.7	74.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#22
Diisopropyl ether
Concen: 20.760 ug/l
RT: 3.757 min Scan# 438
Delta R.T. -0.006 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 45 Resp: 84046

Ion	Ratio	Lower	Upper
45	100		
43	98.0	59.3	88.9#
87	25.9	20.4	30.6
59	10.3	9.0	13.6





#23
Vinyl Acetate
Concen: 103.458 ug/l
RT: 3.721 min Scan# 4
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52


Tgt Ion: 43 Resp: 36188


Ion	Ratio	Lower	Upper
43	100		
86	9.6	7.5	11.3

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025







#24
1,1-Dichloroethane
Concen: 20.772 ug/l
RT: 3.605 min Scan# 413
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 63 Resp: 47560

Ion	Ratio	Lower	Upper
63	100		
98	6.5	3.4	10.1
100	3.8	2.2	6.6





#25
2-Butanone
Concen: 105.931 ug/l
RT: 4.556 min Scan# 5
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52


Tgt Ion: 43 Resp: 10506


Ion	Ratio	Lower	Upper
43	100		
72	22.9	18.6	28.0

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025







#26
2,2-Dichloropropane
Concen: 19.574 ug/l
RT: 4.471 min Scan# 555
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 77 Resp: 28956

Ion	Ratio	Lower	Upper
77	100		
97	21.8	10.6	31.8






#27
cis-1,2-Dichloroethene
Concen: 20.342 ug/l
RT: 4.483 min Scan# 5
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 96 Resp: 2742


Ion	Ratio	Lower	Upper
96	100		
61	155.2	0.0	313.2
98	65.5	0.0	126.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#28
Bromochloromethane
Concen: 20.141 ug/l
RT: 4.897 min Scan# 625
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 49 Resp: 22285

Ion	Ratio	Lower	Upper
49	100		
129	1.8	0.0	3.8
130	68.0	54.6	81.8






#29
Tetrahydrofuran
Concen: 104.187 ug/l
RT: 5.007 min Scan# 64
Delta R.T. 0.006 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 42 Resp: 6699


Ion	Ratio	Lower	Upper
42	100		
72	42.1	33.4	50.0
71	38.2	31.4	47.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#30
Chloroform
Concen: 20.646 ug/l
RT: 5.093 min Scan# 657
Delta R.T. 0.006 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 83 Resp: 48639

Ion	Ratio	Lower	Upper
83	100		
85	65.8	52.8	79.2






#31
Cyclohexane
Concen: 20.478 ug/l
RT: 5.458 min Scan# 7
Delta R.T. -0.006 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 56 Resp: 4145


Ion	Ratio	Lower	Upper
56	100		
69	30.0	23.6	35.4
84	85.5	67.7	101.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#32
1,1,1-Trichloroethane
Concen: 20.270 ug/l
RT: 5.373 min Scan# 703
Delta R.T. -0.006 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 97 Resp: 40407

Ion	Ratio	Lower	Upper
97	100		
99	63.3	51.7	77.5
61	51.3	41.0	61.4






#33
1,2-Dichloroethane-d4
Concen: 19.697 ug/l
RT: 5.952 min Scan# 7
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 65 Resp: 3250


Ion	Ratio	Lower	Upper
65	100		
67	51.5	0.0	99.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.757 min Scan# 930
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion:114 Resp: 160754

Ion	Ratio	Lower	Upper
114	100		
63	24.2	0.0	46.8
88	16.7	0.0	35.4






#35
Dibromofluoromethane
Concen: 19.256 ug/l
RT: 5.385 min Scan# 7
Delta R.T. 0.006 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion:113 Resp: 2196


Ion	Ratio	Lower	Upper
113	100		
111	103.8	81.8	122.6
192	17.0	13.8	20.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#36
1,1-Dichloropropene
Concen: 19.808 ug/l
RT: 5.684 min Scan# 754
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 75 Resp: 30504

Ion	Ratio	Lower	Upper
75	100		
110	34.0	16.9	50.6
77	32.7	25.0	37.4





#37
Ethyl Acetate
Concen: 20.180 ug/l
RT: 4.715 min Scan# 5
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52


Tgt Ion: 43 Resp: 3916


Ion	Ratio	Lower	Upper
43	100		
61	13.1	9.9	14.9
70	10.1	7.5	11.3

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025







#38
Carbon Tetrachloride
Concen: 20.148 ug/l
RT: 5.672 min Scan# 752
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion:117 Resp: 33533

Ion	Ratio	Lower	Upper
117	100		
119	88.5	76.9	115.3
121	30.3	24.2	36.2






#39
Methylcyclohexane
Concen: 20.291 ug/l
RT: 7.373 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 83 Resp: 3830


Ion	Ratio	Lower	Upper
83	100		
55	82.7	64.9	97.3
98	49.2	37.4	56.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#40
Benzene
Concen: 20.766 ug/l
RT: 6.031 min Scan# 811
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 78 Resp: 97667

Ion	Ratio	Lower	Upper
78	100		
77	24.3	19.0	28.6






#41
Methacrylonitrile
Concen: 21.046 ug/l
RT: 4.922 min Scan# 6
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 41 Resp: 2174


Ion	Ratio	Lower	Upper
41	100		
39	60.3	47.0	70.6
67	69.9	53.2	79.8
52	33.1	25.0	37.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#42
1,2-Dichloroethane
Concen: 21.484 ug/l
RT: 6.086 min Scan# 820
Delta R.T. 0.006 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 62 Resp: 41733

Ion	Ratio	Lower	Upper
62	100		
98	7.7	0.0	15.2





#43
Isopropyl Acetate
Concen: 20.814 ug/l
RT: 6.342 min Scan# 8
Delta R.T. 0.006 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52


Tgt Ion: 43 Resp: 6122


Ion	Ratio	Lower	Upper
43	100		
61	17.3	14.1	21.1
87	11.2	9.2	13.8

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025







#44
Trichloroethene
Concen: 20.501 ug/l
RT: 7.123 min Scan# 990
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion:130 Resp: 22918

Ion	Ratio	Lower	Upper
130	100		
95	104.4	0.0	202.0






#45
1,2-Dichloropropane
Concen: 21.241 ug/l
RT: 7.428 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 63 Resp: 2492


Ion	Ratio	Lower	Upper
63	100		
65	26.9	25.8	38.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#46
Dibromomethane
Concen: 21.207 ug/l
RT: 7.580 min Scan# 1065
Delta R.T. 0.006 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 93 Resp: 19104

Ion	Ratio	Lower	Upper
93	100		
95	79.2	64.8	97.2
174	81.1	65.4	98.0






#47
Bromodichloromethane
Concen: 20.587 ug/l
RT: 7.818 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 83 Resp: 3703


Ion	Ratio	Lower	Upper
83	100		
85	61.7	49.7	74.5
127	8.7	6.1	9.1



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#48
Methyl methacrylate
Concen: 20.575 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. 0.006 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 41 Resp: 31247

Ion	Ratio	Lower	Upper
41	100		
69	73.4	58.2	87.2
39	68.7	53.6	80.4





#49
1,4-Dioxane
Concen: 440.515 ug/l
RT: 7.659 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52


Tgt Ion: 88 Resp: 1210


Ion	Ratio	Lower	Upper
88	100		
43	43.5	32.2	48.4
58	76.4	58.6	88.0

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025







#50
Toluene-d8
Concen: 19.921 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 98 Resp: 79303

Ion	Ratio	Lower	Upper
98	100		
100	64.7	52.2	78.4






#51
4-Methyl-2-Pentanone
Concen: 109.055 ug/l
RT: 8.574 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 43 Resp: 21264


Ion	Ratio	Lower	Upper
43	100		
58	35.2	28.5	42.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#52
Toluene
Concen: 21.208 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 92 Resp: 60357

Ion	Ratio	Lower	Upper
92	100		
91	171.6	138.7	208.1






#53
t-1,3-Dichloropropene
Concen: 18.764 ug/l
RT: 8.976 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 75 Resp: 2987


Ion	Ratio	Lower	Upper
75	100		
77	29.5	25.4	38.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#54
cis-1,3-Dichloropropene
Concen: 20.762 ug/l
RT: 8.366 min Scan# 1194
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 75 Resp: 35099

Ion	Ratio	Lower	Upper
75	100		
77	32.1	24.6	36.8
39	65.0	48.4	72.6






#55
1,1,2-Trichloroethane
Concen: 21.331 ug/l
RT: 9.147 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 97 Resp: 2419


Ion	Ratio	Lower	Upper
97	100		
83	93.4	71.8	107.8
85	55.9	47.0	70.6
99	64.9	48.6	73.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#56
Ethyl methacrylate
Concen: 20.600 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 69 Resp: 36155

Ion	Ratio	Lower	Upper
69	100		
41	77.4	61.3	91.9
39	51.5	40.2	60.2






#57
1,3-Dichloropropane
Concen: 21.217 ug/l
RT: 9.305 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 76 Resp: 4190


Ion	Ratio	Lower	Upper
76	100		
78	32.5	26.0	39.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#58
2-Chloroethyl Vinyl ether
Concen: 103.534 ug/l
RT: 8.238 min Scan# 1173
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 63 Resp: 91923

Ion	Ratio	Lower	Upper
63	100		
106	26.1	21.0	31.6






#59
2-Hexanone
Concen: 108.730 ug/l
RT: 9.427 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 43 Resp: 15699


Ion	Ratio	Lower	Upper
43	100		
58	49.1	24.9	74.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#60
Dibromochloromethane
Concen: 21.023 ug/l
RT: 9.519 min Scan# 1383
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion:129 Resp: 25992

Ion	Ratio	Lower	Upper
129	100		
127	77.3	39.4	118.2






#61
1,2-Dibromoethane
Concen: 21.303 ug/l
RT: 9.604 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion:107 Resp: 2445


Ion	Ratio	Lower	Upper
107	100		
109	93.8	75.7	113.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#62
4-Bromofluorobenzene
Concen: 19.858 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 95 Resp: 28795

Ion	Ratio	Lower	Upper
95	100		
174	71.1	0.0	135.8
176	65.8	0.0	131.4






#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion:117 Resp: 14408


Ion	Ratio	Lower	Upper
117	100		
82	61.5	49.2	73.8
119	31.5	25.1	37.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#64
Tetrachloroethene
Concen: 21.015 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. -0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion:164 Resp: 21379

Ion	Ratio	Lower	Upper
164	100		
166	132.2	103.8	155.8
129	104.9	84.3	126.5
131	98.1	80.9	121.3






#65
Chlorobenzene
Concen: 21.023 ug/l
RT: 10.079 min Scan# 14
Delta R.T. 0.006 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion:112 Resp: 6472


Ion	Ratio	Lower	Upper
112	100		
114	31.6	24.7	37.1



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#66
1,1,1,2-Tetrachloroethane
Concen: 20.064 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion:131 Resp: 21447

Ion	Ratio	Lower	Upper
131	100		
133	92.8	47.5	142.5
119	66.3	32.9	98.6






#67
Ethyl Benzene
Concen: 20.921 ug/l
RT: 10.189 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 91 Resp: 11536


Ion	Ratio	Lower	Upper
91	100		
106	29.0	23.7	35.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#68
m/p-Xylenes
Concen: 42.092 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion:106 Resp: 84420

Ion	Ratio	Lower	Upper
106	100		
91	214.4	172.6	258.8






#69
o-Xylene
Concen: 21.158 ug/l
RT: 10.640 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion:106 Resp: 4180


Ion	Ratio	Lower	Upper
106	100		
91	226.9	113.6	340.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#70
Styrene
Concen: 21.249 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion:104 Resp: 69290

Ion	Ratio	Lower	Upper
104	100		
78	57.0	45.4	68.2
103	56.5	43.8	65.8





#71
Bromoform
Concen: 19.619 ug/l
RT: 10.799 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52


Tgt Ion:173 Resp: 1565


Ion	Ratio	Lower	Upper
173	100		
175	47.5	23.4	70.2
254	0.0	0.0	0.0

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025







#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion:152 Resp: 65341

Ion	Ratio	Lower	Upper
152	100		
115	76.8	46.9	140.7
150	166.0	0.0	349.4






#73
Isopropylbenzene
Concen: 21.093 ug/l
RT: 10.957 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion:105 Resp: 11039


Ion	Ratio	Lower	Upper
105	100		
120	25.5	12.7	38.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#74
N-amyl acetate
Concen: 20.466 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 43 Resp: 51238

Ion	Ratio	Lower	Upper
43	100		
70	38.1	30.1	45.1
55	22.5	18.3	27.5
61	22.1	17.5	26.3






#75
1,1,2,2-Tetrachloroethane
Concen: 20.764 ug/l
RT: 11.207 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 83 Resp: 3817


Ion	Ratio	Lower	Upper
83	100		
131	9.0	4.7	14.1
85	63.7	32.4	97.2



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#76
1,2,3-Trichloropropane
Concen: 20.510 ug/l m
RT: 11.238 min Scan# 1665
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 75 Resp: 32690

Ion	Ratio	Lower	Upper
75	100		
77	41.2	20.0	59.9





#77
Bromobenzene
Concen: 20.592 ug/l
RT: 11.195 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52


Tgt Ion:156 Resp: 2490


Ion	Ratio	Lower	Upper
156	100		
77	185.0	92.6	277.8
158	96.7	48.1	144.4

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025







#78
n-propylbenzene
Concen: 21.729 ug/l
RT: 11.299 min Scan# 1675
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 91 Resp: 130994

Ion	Ratio	Lower	Upper
91	100		
120	20.5	10.8	32.3






#79
2-Chlorotoluene
Concen: 21.229 ug/l
RT: 11.360 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 91 Resp: 8181


Ion	Ratio	Lower	Upper
91	100		
126	30.9	15.8	47.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#80
1,3,5-Trimethylbenzene
Concen: 21.564 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion:105 Resp: 93328

Ion	Ratio	Lower	Upper
105	100		
120	46.5	23.0	69.0






#81
trans-1,4-Dichloro-2-butene
Concen: 17.961 ug/l
RT: 11.018 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 75 Resp: 843


Ion	Ratio	Lower	Upper
75	100		
53	144.7	101.8	152.6
89	45.5	37.2	55.8



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#82
4-Chlorotoluene
Concen: 21.108 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 91 Resp: 90704

Ion	Ratio	Lower	Upper
91	100		
126	26.2	13.3	39.9






#83
tert-Butylbenzene
Concen: 20.964 ug/l
RT: 11.713 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion:119 Resp: 8982


Ion	Ratio	Lower	Upper
119	100		
91	65.7	32.4	97.2
134	24.2	11.5	34.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#84
1,2,4-Trimethylbenzene
Concen: 21.235 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion:105 Resp: 92242

Ion	Ratio	Lower	Upper
105	100		
120	43.1	21.4	64.3






#85
sec-Butylbenzene
Concen: 21.433 ug/l
RT: 11.890 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion:105 Resp: 11280


Ion	Ratio	Lower	Upper
105	100		
134	19.3	9.6	28.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#86
p-Isopropyltoluene
Concen: 21.173 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion:119 Resp: 91866

Ion	Ratio	Lower	Upper
119	100		
134	25.7	12.6	37.8
91	27.1	13.7	41.1






#87
1,3-Dichlorobenzene
Concen: 20.490 ug/l
RT: 11.969 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion:146 Resp: 4510


Ion	Ratio	Lower	Upper
146	100		
111	43.9	22.2	66.6
148	63.6	31.9	95.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#88
1,4-Dichlorobenzene
Concen: 20.706 ug/l
RT: 12.036 min Scan# 1796
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion:146 Resp: 46208

Ion	Ratio	Lower	Upper
146	100		
111	43.6	20.9	62.8
148	62.6	31.5	94.5






#89
n-Butylbenzene
Concen: 20.662 ug/l
RT: 12.329 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 91 Resp: 7773


Ion	Ratio	Lower	Upper
91	100		
92	53.5	26.9	80.6
134	23.4	11.9	35.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#90
Hexachloroethane
Concen: 19.505 ug/l
RT: 12.536 min Scan# 1878
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion:117 Resp: 14544

Ion	Ratio	Lower	Upper
117	100		
201	62.2	31.2	93.6






#91
1,2-Dichlorobenzene
Concen: 20.896 ug/l
RT: 12.335 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion:146 Resp: 4573


Ion	Ratio	Lower	Upper
146	100		
111	45.0	22.9	68.8
148	63.6	31.9	95.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#92
1,2-Dibromo-3-Chloropropane
Concen: 21.203 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion: 75 Resp: 8153

Ion	Ratio	Lower	Upper
75	100		
155	70.0	34.9	104.7
157	89.1	44.9	134.7






#93
1,2,4-Trichlorobenzene
Concen: 20.347 ug/l
RT: 13.585 min Scan# 2
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion:180 Resp: 2478


Ion	Ratio	Lower	Upper
180	100		
182	96.6	47.9	143.6
145	33.6	17.9	53.8



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#94
Hexachlorobutadiene
Concen: 19.979 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. 0.006 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion:225 Resp: 10488

Ion	Ratio	Lower	Upper
225	100		
223	63.6	31.1	93.3
227	61.6	32.2	96.6






#95
Naphthalene
Concen: 20.928 ug/l
RT: 13.774 min Scan# 2
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion:128 Resp: 9487


Ion	Ratio	Lower	Upper
128	100		
127	13.2	10.3	15.5
129	10.7	8.6	12.8



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025



#96
1,2,3-Trichlorobenzene
Concen: 20.471 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. 0.000 min
Lab File: VX045527.D
Acq: 01 Apr 2025 17:52

Tgt Ion:180 Resp: 26113

Ion	Ratio	Lower	Upper
180	100		
182	97.4	47.3	141.8
145	36.7	18.2	54.6



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045528.D
Acq On : 01 Apr 2025 18:15
Operator : JC/MD
Sample : VSTDICCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED
Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025

Quant Time: Apr 02 02:53:01 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 02:44:48 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	5.550	168	95658	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	168287	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	150201	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	68786	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	83054	47.477	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 94.960%			
35) Dibromofluoromethane	5.379	113	58108	48.667	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 97.340%			
50) Toluene-d8	8.647	98	204338	49.031	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 98.060%			
62) 4-Bromofluorobenzene	11.079	95	76190	50.190	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 100.380%			
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.166	85	78444	54.834	ug/l	100
3) Chloromethane	1.307	50	75032	51.054	ug/l	100
4) Vinyl Chloride	1.374	62	68516	50.942	ug/l	100
5) Bromomethane	1.593	94	31584	49.526	ug/l	100
6) Chloroethane	1.666	64	38112	53.409	ug/l	100
7) Trichlorofluoromethane	1.874	101	103888	51.798	ug/l	100
8) Diethyl Ether	2.136	74	33953	50.424	ug/l	100
9) 1,1,2-Trichlorotrifluo...	2.325	101	59431	50.569	ug/l	100
10) Methyl Iodide	2.447	142	76404	52.121	ug/l	100
11) Tert butyl alcohol	2.977	59	60610	257.809	ug/l	100
12) 1,1-Dichloroethene	2.313	96	57738	50.278	ug/l	100
13) Acrolein	2.239	56	79989	247.074	ug/l	100
14) Allyl chloride	2.660	41	112085	51.441	ug/l	100
15) Acrylonitrile	3.062	53	194445	262.015	ug/l	100
16) Acetone	2.386	43	180563	251.366	ug/l	100
17) Carbon Disulfide	2.508	76	148601	52.393	ug/l	100
18) Methyl Acetate	2.703	43	82000	49.581	ug/l	100
19) Methyl tert-butyl Ether	3.111	73	202582	50.429	ug/l	100
20) Methylene Chloride	2.782	84	67855	50.457	ug/l	100
21) trans-1,2-Dichloroethene	3.087	96	59696	50.882	ug/l	100
22) Diisopropyl ether	3.764	45	219374	51.111	ug/l	100
23) Vinyl Acetate	3.721	43	998064	269.139	ug/l	100
24) 1,1-Dichloroethane	3.605	63	121348	49.992	ug/l	100
25) 2-Butanone	4.556	43	280076	266.363	ug/l	100
26) 2,2-Dichloropropane	4.471	77	81054	51.683	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	71871	50.282	ug/l	100
28) Bromochloromethane	4.897	49	57046	48.632	ug/l	100
29) Tetrahydrofuran	5.001	42	177402	260.230	ug/l	100
30) Chloroform	5.086	83	125797	50.366	ug/l	100
31) Cyclohexane	5.464	56	107407	50.052	ug/l	100
32) 1,1,1-Trichloroethane	5.379	97	106062	50.184	ug/l	100
36) 1,1-Dichloropropene	5.684	75	81607	50.620	ug/l	100
37) Ethyl Acetate	4.715	43	104896	51.632	ug/l	100
38) Carbon Tetrachloride	5.672	117	90247	51.797	ug/l	100
39) Methylcyclohexane	7.373	83	104678	52.970	ug/l	100
40) Benzene	6.031	78	249161	50.605	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045528.D
Acq On : 01 Apr 2025 18:15
Operator : JC/MD
Sample : VSTDICCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED
Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025

Quant Time: Apr 02 02:53:01 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 02:44:48 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
41) Methacrylonitrile	4.922	41	59068	54.623	ug/l		100
42) 1,2-Dichloroethane	6.080	62	104601	51.437	ug/l		100
43) Isopropyl Acetate	6.336	43	164588	53.452	ug/l		100
44) Trichloroethene	7.123	130	59042	50.452	ug/l		100
45) 1,2-Dichloropropane	7.427	63	63316	51.540	ug/l		100
46) Dibromomethane	7.574	93	49125	52.092	ug/l		100
47) Bromodichloromethane	7.818	83	96176	51.068	ug/l		100
48) Methyl methacrylate	7.690	41	85483	53.768	ug/l		100
49) 1,4-Dioxane	7.659	88	31199	1084.635	ug/l		100
51) 4-Methyl-2-Pentanone	8.574	43	558029	273.375	ug/l		100
52) Toluene	8.714	92	153143	51.402	ug/l		100
53) t-1,3-Dichloropropene	8.976	75	85462	47.317	ug/l		100
54) cis-1,3-Dichloropropene	8.366	75	95524	53.975	ug/l		100
55) 1,1,2-Trichloroethane	9.147	97	60452	50.919	ug/l		100
56) Ethyl methacrylate	9.116	69	100998	54.970	ug/l		100
57) 1,3-Dichloropropane	9.305	76	106580	51.554	ug/l		100
58) 2-Chloroethyl Vinyl ether	8.238	63	250428	269.434	ug/l		100
59) 2-Hexanone	9.427	43	413824	273.765	ug/l		100
60) Dibromochloromethane	9.519	129	69258	53.511	ug/l		100
61) 1,2-Dibromoethane	9.604	107	62812	52.264	ug/l		100
64) Tetrachloroethene	9.269	164	52155	49.181	ug/l		100
65) Chlorobenzene	10.073	112	162829	50.733	ug/l		100
66) 1,1,1,2-Tetrachloroethane	10.159	131	55974	50.233	ug/l		100
67) Ethyl Benzene	10.189	91	301481	52.447	ug/l		100
68) m/p-Xylenes	10.299	106	218702	104.606	ug/l		100
69) o-Xylene	10.640	106	108028	52.452	ug/l		100
70) Styrene	10.653	104	182620	53.723	ug/l		100
71) Bromoform	10.799	173	44397	53.361	ug/l	#	100
73) Isopropylbenzene	10.957	105	287628	52.203	ug/l		100
74) N-amyl acetate	10.842	43	141589	53.721	ug/l		100
75) 1,1,2,2-Tetrachloroethane	11.207	83	95200	49.185	ug/l		100
76) 1,2,3-Trichloropropane	11.238	75	84279m	50.230	ug/l		
77) Bromobenzene	11.195	156	64527	50.686	ug/l		100
78) n-propylbenzene	11.299	91	336796	53.069	ug/l		100
79) 2-Chlorotoluene	11.360	91	203994	50.279	ug/l		100
80) 1,3,5-Trimethylbenzene	11.451	105	241536	53.013	ug/l		100
81) trans-1,4-Dichloro-2-b...	11.018	75	25488	51.556	ug/l		100
82) 4-Chlorotoluene	11.451	91	237213	52.439	ug/l		100
83) tert-Butylbenzene	11.713	119	235715	52.256	ug/l		100
84) 1,2,4-Trimethylbenzene	11.750	105	242301	52.986	ug/l		100
85) sec-Butylbenzene	11.890	105	295799	53.391	ug/l		100
86) p-Isopropyltoluene	12.006	119	242416	53.073	ug/l		100
87) 1,3-Dichlorobenzene	11.969	146	116890	50.441	ug/l		100
88) 1,4-Dichlorobenzene	12.036	146	117164	49.871	ug/l		100
89) n-Butylbenzene	12.329	91	217345	54.874	ug/l		100
90) Hexachloroethane	12.536	117	41095	52.352	ug/l		100
91) 1,2-Dichlorobenzene	12.335	146	115426	50.099	ug/l		100
92) 1,2-Dibromo-3-Chloropr...	12.939	75	21712	53.638	ug/l		100
93) 1,2,4-Trichlorobenzene	13.585	180	65145	50.797	ug/l		100
94) Hexachlorobutadiene	13.719	225	27795	50.296	ug/l		100
95) Naphthalene	13.774	128	256009	53.643	ug/l		100
96) 1,2,3-Trichlorobenzene	13.957	180	68781	51.219	ug/l		100

ICAL Form

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045528.D
Acq On : 01 Apr 2025 18:15
Operator : JC/MD
Sample : VSTDICCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025

Quant Time: Apr 02 02:53:01 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 02:44:48 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045528.D
Acq On : 01 Apr 2025 18:15
Operator : JC/MD
Sample : VSTDICCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1


Quant Time: Apr 02 02:53:01 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 02:44:48 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025






#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 7
Delta R.T. -0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion:168 Resp: 9565


Ion	Ratio	Lower	Upper
168	100		
99	65.4	52.3	78.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#2
Dichlorodifluoromethane
Concen: 54.834 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion: 85 Resp: 78444

Ion	Ratio	Lower	Upper
85	100		
87	32.1	16.1	48.2






#3
Chloromethane
Concen: 51.054 ug/l
RT: 1.307 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion: 50 Resp: 7503


Ion	Ratio	Lower	Upper
50	100		
52	32.4	25.9	38.9



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#4
Vinyl Chloride
Concen: 50.942 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion: 62 Resp: 68516

Ion	Ratio	Lower	Upper
62	100		
64	31.9	25.5	38.3






#5
Bromomethane
Concen: 49.526 ug/l
RT: 1.593 min Scan# 8
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion: 94 Resp: 3158


Ion	Ratio	Lower	Upper
94	100		
96	93.5	74.8	112.2



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#6
Chloroethane
Concen: 53.409 ug/l
RT: 1.666 min Scan# 95
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion: 64 Resp: 38112

Ion	Ratio	Lower	Upper
64	100		
66	32.9	26.3	39.5






#7
Trichlorofluoromethane
Concen: 51.798 ug/l
RT: 1.874 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion:101 Resp: 10388


Ion	Ratio	Lower	Upper
101	100		
103	63.9	51.1	76.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#8
Diethyl Ether
Concen: 50.424 ug/l
RT: 2.136 min Scan# 172
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion: 74 Resp: 33953

Ion	Ratio	Lower	Upper
74	100		
45	106.1	53.0	159.1






#9
1,1,2-Trichlorotrifluoroethane
Concen: 50.569 ug/l
RT: 2.325 min Scan# 2
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion:101 Resp: 5943


Ion	Ratio	Lower	Upper
101	100		
85	48.6	38.9	58.3
151	71.9	57.5	86.3



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#10
Methyl Iodide
Concen: 52.121 ug/l
RT: 2.447 min Scan# 223
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion:142 Resp: 76404

Ion	Ratio	Lower	Upper
142	100		
127	50.3	40.2	60.4
141	14.6	11.7	17.5





#11
Tert butyl alcohol
Concen: 257.809 ug/l
RT: 2.977 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15


Tgt Ion: 59 Resp: 6061


Ion	Ratio	Lower	Upper
59	100		
57	10.7	8.6	12.8

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations
APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025







#12
1,1-Dichloroethene
Concen: 50.278 ug/l
RT: 2.313 min Scan# 201
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion: 96 Resp: 57738

Ion	Ratio	Lower	Upper
96	100		
61	182.1	145.7	218.5
98	64.5	51.6	77.4






#13
Acrolein
Concen: 247.074 ug/l
RT: 2.239 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion: 56 Resp: 7998


Ion	Ratio	Lower	Upper
56	100		
55	70.2	56.2	84.2



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#14
Allyl chloride
Concen: 51.441 ug/l
RT: 2.660 min Scan# 258
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion: 41 Resp: 112085

Ion	Ratio	Lower	Upper
41	100		
39	75.2	60.2	90.2
76	30.7	24.6	36.8






#15
Acrylonitrile
Concen: 262.015 ug/l
RT: 3.062 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion: 53 Resp: 19444


Ion	Ratio	Lower	Upper
53	100		
52	84.3	67.4	101.2
51	37.4	29.9	44.9



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#16
Acetone
Concen: 251.366 ug/l
RT: 2.386 min Scan# 213
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion: 43 Resp: 180563

Ion	Ratio	Lower	Upper
43	100		
58	27.9	22.3	33.5





#17
Carbon Disulfide
Concen: 52.393 ug/l
RT: 2.508 min Scan# 2
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15


Tgt Ion: 76 Resp: 14860


Ion	Ratio	Lower	Upper
76	100		
78	8.8	7.0	10.6

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025







#18
Methyl Acetate
Concen: 49.581 ug/l
RT: 2.703 min Scan# 265
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion: 43 Resp: 82000

Ion	Ratio	Lower	Upper
43	100		
74	21.9	17.5	26.3






#19
Methyl tert-butyl Ether
Concen: 50.429 ug/l
RT: 3.111 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion: 73 Resp: 20258


Ion	Ratio	Lower	Upper
73	100		
57	22.0	17.6	26.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#20
Methylene Chloride
Concen: 50.457 ug/l
RT: 2.782 min Scan# 278
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion: 84 Resp: 67855

Ion	Ratio	Lower	Upper
84	100		
49	134.4	107.5	161.3
51	41.3	33.0	49.6
86	62.6	50.1	75.1






#21
trans-1,2-Dichloroethene
Concen: 50.882 ug/l
RT: 3.087 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion: 96 Resp: 5969


Ion	Ratio	Lower	Upper
96	100		
61	151.3	121.0	181.6
98	62.1	49.7	74.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#22
Diisopropyl ether
Concen: 51.111 ug/l
RT: 3.764 min Scan# 439
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion: 45 Resp: 219374

Ion	Ratio	Lower	Upper
45	100		
43	74.1	59.3	88.9
87	25.5	20.4	30.6
59	11.3	9.0	13.6





#23
Vinyl Acetate
Concen: 269.139 ug/l
RT: 3.721 min Scan# 4
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15


Tgt Ion: 43 Resp: 99806


Ion	Ratio	Lower	Upper
43	100		
86	9.4	7.5	11.3

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations
APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025







#24
1,1-Dichloroethane
Concen: 49.992 ug/l
RT: 3.605 min Scan# 413
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion: 63 Resp: 121348

Ion	Ratio	Lower	Upper
63	100		
98	6.7	3.4	10.1
100	4.4	2.2	6.6





#25
2-Butanone
Concen: 266.363 ug/l
RT: 4.556 min Scan# 5
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15


Tgt Ion: 43 Resp: 28007


Ion	Ratio	Lower	Upper
43	100		
72	23.3	18.6	28.0

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations
APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025







#26
2,2-Dichloropropane
Concen: 51.683 ug/l
RT: 4.471 min Scan# 555
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion: 77 Resp: 81054

Ion	Ratio	Lower	Upper
77	100		
97	21.2	10.6	31.8






#27
cis-1,2-Dichloroethene
Concen: 50.282 ug/l
RT: 4.483 min Scan# 5
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion: 96 Resp: 7187


Ion	Ratio	Lower	Upper
96	100		
61	156.6	0.0	313.2
98	63.2	0.0	126.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#28
Bromochloromethane
Concen: 48.632 ug/l
RT: 4.897 min Scan# 625
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion: 49 Resp: 57046

Ion	Ratio	Lower	Upper
49	100		
129	1.9	0.0	3.8
130	68.2	54.6	81.8





#29
Tetrahydrofuran
Concen: 260.230 ug/l
RT: 5.001 min Scan# 64
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15


Tgt Ion: 42 Resp: 17740


Ion	Ratio	Lower	Upper
42	100		
72	41.7	33.4	50.0
71	39.2	31.4	47.0

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations
APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025







#30
Chloroform
Concen: 50.366 ug/l
RT: 5.086 min Scan# 656
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion: 83 Resp: 125797

Ion	Ratio	Lower	Upper
83	100		
85	66.0	52.8	79.2






#31
Cyclohexane
Concen: 50.052 ug/l
RT: 5.464 min Scan# 7
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion: 56 Resp: 10740


Ion	Ratio	Lower	Upper
56	100		
69	29.5	23.6	35.4
84	84.6	67.7	101.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#32
1,1,1-Trichloroethane
Concen: 50.184 ug/l
RT: 5.379 min Scan# 704
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion: 97 Resp: 106062

Ion	Ratio	Lower	Upper
97	100		
99	64.6	51.7	77.5
61	51.2	41.0	61.4






#33
1,2-Dichloroethane-d4
Concen: 47.477 ug/l
RT: 5.952 min Scan# 7
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion: 65 Resp: 8305


Ion	Ratio	Lower	Upper
65	100		
67	49.5	0.0	99.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.757 min Scan# 930
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion:114 Resp: 168287

Ion	Ratio	Lower	Upper
114	100		
63	23.4	0.0	46.8
88	17.7	0.0	35.4






#35
Dibromofluoromethane
Concen: 48.667 ug/l
RT: 5.379 min Scan# 7
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion:113 Resp: 5810


Ion	Ratio	Lower	Upper
113	100		
111	102.2	81.8	122.6
192	17.2	13.8	20.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#36
1,1-Dichloropropene
Concen: 50.620 ug/l
RT: 5.684 min Scan# 754
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion: 75 Resp: 81607

Ion	Ratio	Lower	Upper
75	100		
110	33.7	16.9	50.6
77	31.2	25.0	37.4






#37
Ethyl Acetate
Concen: 51.632 ug/l
RT: 4.715 min Scan# 5
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion: 43 Resp: 10489


Ion	Ratio	Lower	Upper
43	100		
61	12.4	9.9	14.9
70	9.4	7.5	11.3



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#38
Carbon Tetrachloride
Concen: 51.797 ug/l
RT: 5.672 min Scan# 752
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion:117 Resp: 90247

Ion	Ratio	Lower	Upper
117	100		
119	96.1	76.9	115.3
121	30.2	24.2	36.2






#39
Methylcyclohexane
Concen: 52.970 ug/l
RT: 7.373 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion: 83 Resp: 10467


Ion	Ratio	Lower	Upper
83	100		
55	81.1	64.9	97.3
98	46.7	37.4	56.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#40
Benzene
Concen: 50.605 ug/l
RT: 6.031 min Scan# 811
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion: 78 Resp: 249161

Ion	Ratio	Lower	Upper
78	100		
77	23.8	19.0	28.6






#41
Methacrylonitrile
Concen: 54.623 ug/l
RT: 4.922 min Scan# 6
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion: 41 Resp: 5906


Ion	Ratio	Lower	Upper
41	100		
39	58.8	47.0	70.6
67	66.5	53.2	79.8
52	31.3	25.0	37.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#42
1,2-Dichloroethane
Concen: 51.437 ug/l
RT: 6.080 min Scan# 819
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion: 62 Resp: 104601

Ion	Ratio	Lower	Upper
62	100		
98	7.6	0.0	15.2





#43
Isopropyl Acetate
Concen: 53.452 ug/l
RT: 6.336 min Scan# 8
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15


Tgt Ion: 43 Resp: 16458


Ion	Ratio	Lower	Upper
43	100		
61	17.6	14.1	21.1
87	11.5	9.2	13.8

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations
APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025







#44
Trichloroethene
Concen: 50.452 ug/l
RT: 7.123 min Scan# 990
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion:130 Resp: 59042

Ion	Ratio	Lower	Upper
130	100		
95	101.0	0.0	202.0






#45
1,2-Dichloropropane
Concen: 51.540 ug/l
RT: 7.427 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion: 63 Resp: 6331


Ion	Ratio	Lower	Upper
63	100		
65	32.2	25.8	38.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#46
Dibromomethane
Concen: 52.092 ug/l
RT: 7.574 min Scan# 1064
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion: 93 Resp: 49125

Ion	Ratio	Lower	Upper
93	100		
95	81.0	64.8	97.2
174	81.7	65.4	98.0






#47
Bromodichloromethane
Concen: 51.068 ug/l
RT: 7.818 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion: 83 Resp: 9617


Ion	Ratio	Lower	Upper
83	100		
85	62.1	49.7	74.5
127	7.6	6.1	9.1



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#48
Methyl methacrylate
Concen: 53.768 ug/l
RT: 7.690 min Scan# 1083
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion: 41 Resp: 85483

Ion	Ratio	Lower	Upper
41	100		
69	72.7	58.2	87.2
39	67.0	53.6	80.4






#49
1,4-Dioxane
Concen: 1084.635 ug/l
RT: 7.659 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion: 88 Resp: 3119


Ion	Ratio	Lower	Upper
88	100		
43	40.3	32.2	48.4
58	73.3	58.6	88.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#50
Toluene-d8
Concen: 49.031 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion: 98 Resp: 204338

Ion	Ratio	Lower	Upper
98	100		
100	65.3	52.2	78.4






#51
4-Methyl-2-Pentanone
Concen: 273.375 ug/l
RT: 8.574 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion: 43 Resp: 55802


Ion	Ratio	Lower	Upper
43	100		
58	35.6	28.5	42.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#52
Toluene
Concen: 51.402 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion: 92 Resp: 153143

Ion	Ratio	Lower	Upper
92	100		
91	173.4	138.7	208.1






#53
t-1,3-Dichloropropene
Concen: 47.317 ug/l
RT: 8.976 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion: 75 Resp: 8546


Ion	Ratio	Lower	Upper
75	100		
77	31.7	25.4	38.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#54
cis-1,3-Dichloropropene
Concen: 53.975 ug/l
RT: 8.366 min Scan# 1194
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion: 75 Resp: 95524

Ion	Ratio	Lower	Upper
75	100		
77	30.7	24.6	36.8
39	60.5	48.4	72.6






#55
1,1,2-Trichloroethane
Concen: 50.919 ug/l
RT: 9.147 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion: 97 Resp: 6045


Ion	Ratio	Lower	Upper
97	100		
83	89.8	71.8	107.8
85	58.8	47.0	70.6
99	60.8	48.6	73.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#56
Ethyl methacrylate
Concen: 54.970 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion: 69 Resp: 100998

Ion	Ratio	Lower	Upper
69	100		
41	76.6	61.3	91.9
39	50.2	40.2	60.2






#57
1,3-Dichloropropane
Concen: 51.554 ug/l
RT: 9.305 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion: 76 Resp: 10658


Ion	Ratio	Lower	Upper
76	100		
78	32.5	26.0	39.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#58
2-Chloroethyl Vinyl ether
Concen: 269.434 ug/l
RT: 8.238 min Scan# 1173
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion: 63 Resp: 250428

Ion	Ratio	Lower	Upper
63	100		
106	26.3	21.0	31.6





#59
2-Hexanone
Concen: 273.765 ug/l
RT: 9.427 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15


Tgt Ion: 43 Resp: 41382


Ion	Ratio	Lower	Upper
43	100		
58	49.7	24.9	74.6

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025







#60
Dibromochloromethane
Concen: 53.511 ug/l
RT: 9.519 min Scan# 1383
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion:129 Resp: 69258

Ion	Ratio	Lower	Upper
129	100		
127	78.8	39.4	118.2






#61
1,2-Dibromoethane
Concen: 52.264 ug/l
RT: 9.604 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion:107 Resp: 6281


Ion	Ratio	Lower	Upper
107	100		
109	94.6	75.7	113.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#62
4-Bromofluorobenzene
Concen: 50.190 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion: 95 Resp: 76190

Ion	Ratio	Lower	Upper
95	100		
174	67.9	0.0	135.8
176	65.7	0.0	131.4






#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion:117 Resp: 15020


Ion	Ratio	Lower	Upper
117	100		
82	61.5	49.2	73.8
119	31.4	25.1	37.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#64
Tetrachloroethene
Concen: 49.181 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. -0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion:164 Resp: 52155

Ion	Ratio	Lower	Upper
164	100		
166	129.8	103.8	155.8
129	105.4	84.3	126.5
131	101.1	80.9	121.3






#65
Chlorobenzene
Concen: 50.733 ug/l
RT: 10.073 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion:112 Resp: 16282


Ion	Ratio	Lower	Upper
112	100		
114	30.9	24.7	37.1



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#66
1,1,1,2-Tetrachloroethane
Concen: 50.233 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion:131 Resp: 55974

Ion	Ratio	Lower	Upper
131	100		
133	95.0	47.5	142.5
119	65.7	32.9	98.6






#67
Ethyl Benzene
Concen: 52.447 ug/l
RT: 10.189 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion: 91 Resp: 30148


Ion	Ratio	Lower	Upper
91	100		
106	29.6	23.7	35.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#68
m/p-Xylenes
Concen: 104.606 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion:106 Resp: 218702

Ion	Ratio	Lower	Upper
106	100		
91	215.7	172.6	258.8






#69
o-Xylene
Concen: 52.452 ug/l
RT: 10.640 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion:106 Resp: 10802


Ion	Ratio	Lower	Upper
106	100		
91	227.1	113.6	340.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#70
Styrene
Concen: 53.723 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion:104 Resp: 182620

Ion	Ratio	Lower	Upper
104	100		
78	56.8	45.4	68.2
103	54.8	43.8	65.8






#71
Bromoform
Concen: 53.361 ug/l
RT: 10.799 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion:173 Resp: 4439


Ion	Ratio	Lower	Upper
173	100		
175	46.8	23.4	70.2
254	0.0	0.0	0.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion:152 Resp: 68786

Ion	Ratio	Lower	Upper
152	100		
115	93.8	46.9	140.7
150	174.7	0.0	349.4






#73
Isopropylbenzene
Concen: 52.203 ug/l
RT: 10.957 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion:105 Resp: 28762


Ion	Ratio	Lower	Upper
105	100		
120	25.3	12.7	38.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#74
N-amyl acetate
Concen: 53.721 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion: 43 Resp: 141589

Ion	Ratio	Lower	Upper
43	100		
70	37.6	30.1	45.1
55	22.9	18.3	27.5
61	21.9	17.5	26.3





#75
1,1,2,2-Tetrachloroethane
Concen: 49.185 ug/l
RT: 11.207 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15


Tgt Ion: 83 Resp: 9520


Ion	Ratio	Lower	Upper
83	100		
131	9.4	4.7	14.1
85	64.8	32.4	97.2

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations
APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025







#76
1,2,3-Trichloropropane
Concen: 50.230 ug/l m
RT: 11.238 min Scan# 1665
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion: 75 Resp: 84279

Ion	Ratio	Lower	Upper
75	100		
77	39.9	20.0	59.9






#77
Bromobenzene
Concen: 50.686 ug/l
RT: 11.195 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion:156 Resp: 6452


Ion	Ratio	Lower	Upper
156	100		
77	185.2	92.6	277.8
158	96.3	48.1	144.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#78
n-propylbenzene
Concen: 53.069 ug/l
RT: 11.299 min Scan# 1675
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion: 91 Resp: 336796

Ion	Ratio	Lower	Upper
91	100		
120	21.5	10.8	32.3





#79
2-Chlorotoluene
Concen: 50.279 ug/l
RT: 11.360 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15


Tgt Ion: 91 Resp: 20399


Ion	Ratio	Lower	Upper
91	100		
126	31.6	15.8	47.4

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025







#80
1,3,5-Trimethylbenzene
Concen: 53.013 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion:105 Resp: 241536

Ion	Ratio	Lower	Upper
105	100		
120	46.0	23.0	69.0






#81
trans-1,4-Dichloro-2-butene
Concen: 51.556 ug/l
RT: 11.018 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion: 75 Resp: 2548


Ion	Ratio	Lower	Upper
75	100		
53	127.2	101.8	152.6
89	46.5	37.2	55.8



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#82
4-Chlorotoluene
Concen: 52.439 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion: 91 Resp: 237213

Ion	Ratio	Lower	Upper
91	100		
126	26.6	13.3	39.9






#83
tert-Butylbenzene
Concen: 52.256 ug/l
RT: 11.713 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion:119 Resp: 23571


Ion	Ratio	Lower	Upper
119	100		
91	64.8	32.4	97.2
134	23.0	11.5	34.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#84
1,2,4-Trimethylbenzene
Concen: 52.986 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion:105 Resp: 242301

Ion	Ratio	Lower	Upper
105	100		
120	42.9	21.4	64.3






#85
sec-Butylbenzene
Concen: 53.391 ug/l
RT: 11.890 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion:105 Resp: 29579


Ion	Ratio	Lower	Upper
105	100		
134	19.1	9.6	28.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#86
p-Isopropyltoluene
Concen: 53.073 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion:119 Resp: 242416

Ion	Ratio	Lower	Upper
119	100		
134	25.2	12.6	37.8
91	27.4	13.7	41.1






#87
1,3-Dichlorobenzene
Concen: 50.441 ug/l
RT: 11.969 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion:146 Resp: 11689


Ion	Ratio	Lower	Upper
146	100		
111	44.4	22.2	66.6
148	63.8	31.9	95.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#88
1,4-Dichlorobenzene
Concen: 49.871 ug/l
RT: 12.036 min Scan# 1796
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion:146 Resp: 117164

Ion	Ratio	Lower	Upper
146	100		
111	41.9	20.9	62.8
148	63.0	31.5	94.5






#89
n-Butylbenzene
Concen: 54.874 ug/l
RT: 12.329 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion: 91 Resp: 21734


Ion	Ratio	Lower	Upper
91	100		
92	53.7	26.9	80.6
134	23.8	11.9	35.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#90
Hexachloroethane
Concen: 52.352 ug/l
RT: 12.536 min Scan# 1878
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion:117 Resp: 41095

Ion	Ratio	Lower	Upper
117	100		
201	62.4	31.2	93.6






#91
1,2-Dichlorobenzene
Concen: 50.099 ug/l
RT: 12.335 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion:146 Resp: 11542


Ion	Ratio	Lower	Upper
146	100		
111	45.9	22.9	68.8
148	63.7	31.9	95.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#92
1,2-Dibromo-3-Chloropropane
Concen: 53.638 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion: 75 Resp: 21712

Ion	Ratio	Lower	Upper
75	100		
155	69.8	34.9	104.7
157	89.8	44.9	134.7






#93
1,2,4-Trichlorobenzene
Concen: 50.797 ug/l
RT: 13.585 min Scan# 2
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion:180 Resp: 6514


Ion	Ratio	Lower	Upper
180	100		
182	95.7	47.9	143.6
145	35.9	17.9	53.8



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#94
Hexachlorobutadiene
Concen: 50.296 ug/l
RT: 13.719 min Scan# 2072
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion:225 Resp: 27795

Ion	Ratio	Lower	Upper
225	100		
223	62.2	31.1	93.3
227	64.4	32.2	96.6






#95
Naphthalene
Concen: 53.643 ug/l
RT: 13.774 min Scan# 2
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion:128 Resp: 25600


Ion	Ratio	Lower	Upper
128	100		
127	12.9	10.3	15.5
129	10.7	8.6	12.8



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025



#96
1,2,3-Trichlorobenzene
Concen: 51.219 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. 0.000 min
Lab File: VX045528.D
Acq: 01 Apr 2025 18:15

Tgt Ion:180 Resp: 68781

Ion	Ratio	Lower	Upper
180	100		
182	94.5	47.3	141.8
145	36.4	18.2	54.6



Quantitation Report (QT Reviewed)

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045529.D
Acq On : 01 Apr 2025 18:38
Operator : JC/MD
Sample : VSTDICC100
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Quant Time: Apr 02 02:54:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 02:44:48 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED
Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	5.544	168	86107	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	152724	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	136369	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	64612	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	155757	98.914	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 197.820%#			
35) Dibromofluoromethane	5.379	113	107908	99.585	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 199.160%#			
50) Toluene-d8	8.647	98	375599	99.309	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 198.620%#			
62) 4-Bromofluorobenzene	11.079	95	146944	106.664	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 213.320%#			
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.166	85	144108	111.909	ug/l	99
3) Chloromethane	1.307	50	140360	106.098	ug/l	100
4) Vinyl Chloride	1.374	62	127174	105.042	ug/l	99
5) Bromomethane	1.593	94	58706	102.266	ug/l	98
6) Chloroethane	1.660	64	61099	95.120	ug/l	98
7) Trichlorofluoromethane	1.867	101	187555	103.887	ug/l	98
8) Diethyl Ether	2.136	74	61683	101.767	ug/l	99
9) 1,1,2-Trichlorotrifluo...	2.319	101	106989	101.133	ug/l	97
10) Methyl Iodide	2.441	142	135732	102.863	ug/l	100
11) Tert butyl alcohol	2.989	59	110044	519.999	ug/l	99
12) 1,1-Dichloroethene	2.306	96	106404	102.934	ug/l	99
13) Acrolein	2.233	56	146268	501.914	ug/l	99
14) Allyl chloride	2.654	41	206484	105.276	ug/l	99
15) Acrylonitrile	3.062	53	338970	507.426	ug/l	99
16) Acetone	2.386	43	314285	486.054	ug/l	100
17) Carbon Disulfide	2.502	76	276285	108.215	ug/l	100
18) Methyl Acetate	2.703	43	147272	98.925	ug/l	99
19) Methyl tert-butyl Ether	3.111	73	381719	105.561	ug/l	99
20) Methylene Chloride	2.782	84	121435	100.316	ug/l	98
21) trans-1,2-Dichloroethene	3.081	96	106896	101.220	ug/l	98
22) Diisopropyl ether	3.757	45	406185	105.132	ug/l	95
23) Vinyl Acetate	3.721	43	1848817	553.853	ug/l	99
24) 1,1-Dichloroethane	3.605	63	220934	101.114	ug/l	100
25) 2-Butanone	4.556	43	491249	519.018	ug/l	100
26) 2,2-Dichloropropane	4.465	77	155592	110.215	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	129785	100.871	ug/l	98
28) Bromochloromethane	4.891	49	105093	99.530	ug/l	100
29) Tetrahydrofuran	5.001	42	313121	510.264	ug/l	99
30) Chloroform	5.086	83	225402	100.255	ug/l	99
31) Cyclohexane	5.458	56	197136	102.056	ug/l	98
32) 1,1,1-Trichloroethane	5.379	97	198511	104.346	ug/l	99
36) 1,1-Dichloropropene	5.684	75	151448	103.515	ug/l	99
37) Ethyl Acetate	4.715	43	192130	104.207	ug/l	99
38) Carbon Tetrachloride	5.672	117	166465	105.278	ug/l	99
39) Methylcyclohexane	7.373	83	191752	106.921	ug/l	99
40) Benzene	6.031	78	452946	101.369	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045529.D
Acq On : 01 Apr 2025 18:38
Operator : JC/MD
Sample : VSTDICC100
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Quant Time: Apr 02 02:54:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 02:44:48 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
41) Methacrylonitrile	4.922	41	107037	109.068	ug/l		99
42) 1,2-Dichloroethane	6.080	62	189528	102.696	ug/l		100
43) Isopropyl Acetate	6.336	43	305974	109.494	ug/l		100
44) Trichloroethene	7.123	130	107332	101.062	ug/l		97
45) 1,2-Dichloropropane	7.421	63	115731	103.806	ug/l		98
46) Dibromomethane	7.574	93	87381	102.101	ug/l		97
47) Bromodichloromethane	7.818	83	179265	104.886	ug/l		98
48) Methyl methacrylate	7.690	41	158241	109.675	ug/l		99
49) 1,4-Dioxane	7.665	88	52624	2015.906	ug/l		97
51) 4-Methyl-2-Pentanone	8.574	43	988881	533.814	ug/l		99
52) Toluene	8.714	92	276323	102.199	ug/l		99
53) t-1,3-Dichloropropene	8.976	75	169413	100.645	ug/l		100
54) cis-1,3-Dichloropropene	8.366	75	182328	113.521	ug/l		99
55) 1,1,2-Trichloroethane	9.147	97	109208	101.361	ug/l		98
56) Ethyl methacrylate	9.116	69	192034	115.168	ug/l		99
57) 1,3-Dichloropropane	9.305	76	192473	102.588	ug/l		100
58) 2-Chloroethyl Vinyl ether	8.238	63	460297	545.696	ug/l		100
59) 2-Hexanone	9.427	43	738772	538.538	ug/l		99
60) Dibromochloromethane	9.519	129	128700	109.570	ug/l		98
61) 1,2-Dibromoethane	9.604	107	115919	106.282	ug/l		99
64) Tetrachloroethene	9.269	164	90817	94.325	ug/l		99
65) Chlorobenzene	10.073	112	296300	101.682	ug/l		98
66) 1,1,1,2-Tetrachloroethane	10.159	131	103437	102.243	ug/l		99
67) Ethyl Benzene	10.189	91	543568	104.152	ug/l		99
68) m/p-Xylenes	10.299	106	394241	207.694	ug/l		100
69) o-Xylene	10.640	106	195381	104.487	ug/l		99
70) Styrene	10.653	104	335333	108.654	ug/l		100
71) Bromoform	10.799	173	83632	110.714	ug/l	#	97
73) Isopropylbenzene	10.957	105	522436	100.944	ug/l		100
74) N-amyl acetate	10.842	43	275462	111.267	ug/l		99
75) 1,1,2,2-Tetrachloroethane	11.213	83	174977	96.243	ug/l		99
76) 1,2,3-Trichloropropane	11.238	75	154048m	97.744	ug/l		
77) Bromobenzene	11.195	156	119561	99.983	ug/l		99
78) n-propylbenzene	11.299	91	622679	104.454	ug/l		100
79) 2-Chlorotoluene	11.360	91	376682	98.840	ug/l		99
80) 1,3,5-Trimethylbenzene	11.451	105	435137	101.675	ug/l		100
81) trans-1,4-Dichloro-2-b...	11.018	75	51873	111.704	ug/l		95
82) 4-Chlorotoluene	11.451	91	432884	101.877	ug/l		99
83) tert-Butylbenzene	11.713	119	437649	103.292	ug/l		99
84) 1,2,4-Trimethylbenzene	11.750	105	444211	103.414	ug/l		99
85) sec-Butylbenzene	11.890	105	546857	105.082	ug/l		100
86) p-Isopropyltoluene	12.006	119	453109	105.609	ug/l		100
87) 1,3-Dichlorobenzene	11.969	146	221481	101.749	ug/l		99
88) 1,4-Dichlorobenzene	12.036	146	216367	98.047	ug/l		98
89) n-Butylbenzene	12.329	91	420694	113.075	ug/l		99
90) Hexachloroethane	12.536	117	80489	109.160	ug/l		100
91) 1,2-Dichlorobenzene	12.335	146	215157	99.419	ug/l		100
92) 1,2-Dibromo-3-Chloropr...	12.939	75	43047	113.214	ug/l		99
93) 1,2,4-Trichlorobenzene	13.585	180	135046	112.106	ug/l		99
94) Hexachlorobutadiene	13.725	225	53636	103.326	ug/l		98
95) Naphthalene	13.774	128	499719	111.473	ug/l		100
96) 1,2,3-Trichlorobenzene	13.957	180	137328	108.869	ug/l		100

ICAL Form

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045529.D
Acq On : 01 Apr 2025 18:38
Operator : JC/MD
Sample : VSTDICC100
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025

Quant Time: Apr 02 02:54:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 02:44:48 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045529.D
Acq On : 01 Apr 2025 18:38
Operator : JC/MD
Sample : VSTDICC100
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1


Quant Time: Apr 02 02:54:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 02:44:48 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025






#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 7
Delta R.T. -0.006 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion:168 Resp: 8610


Ion	Ratio	Lower	Upper
168	100		
99	65.3	52.3	78.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#2
Dichlorodifluoromethane
Concen: 111.909 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 85 Resp: 144108

Ion	Ratio	Lower	Upper
85	100		
87	32.4	16.1	48.2






#3
Chloromethane
Concen: 106.098 ug/l
RT: 1.307 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 50 Resp: 14036


Ion	Ratio	Lower	Upper
50	100		
52	32.4	25.9	38.9



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#4
Vinyl Chloride
Concen: 105.042 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 62 Resp: 127174

Ion	Ratio	Lower	Upper
62	100		
64	31.2	25.5	38.3






#5
Bromomethane
Concen: 102.266 ug/l
RT: 1.593 min Scan# 8
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 94 Resp: 5870


Ion	Ratio	Lower	Upper
94	100		
96	91.9	74.8	112.2



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#6
Chloroethane
Concen: 95.120 ug/l
RT: 1.660 min Scan# 94
Delta R.T. -0.006 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 64 Resp: 61099

Ion	Ratio	Lower	Upper
64	100		
66	31.7	26.3	39.5






#7
Trichlorofluoromethane
Concen: 103.887 ug/l
RT: 1.867 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion:101 Resp: 18755


Ion	Ratio	Lower	Upper
101	100		
103	62.5	51.1	76.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#8
Diethyl Ether
Concen: 101.767 ug/l
RT: 2.136 min Scan# 172
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 74 Resp: 61683

Ion	Ratio	Lower	Upper
74	100		
45	106.7	53.0	159.1






#9
1,1,2-Trichlorotrifluoroethane
Concen: 101.133 ug/l
RT: 2.319 min Scan# 2
Delta R.T. -0.006 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion:101 Resp: 10698


Ion	Ratio	Lower	Upper
101	100		
85	47.3	38.9	58.3
151	69.0	57.5	86.3



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#10
Methyl Iodide
Concen: 102.863 ug/l
RT: 2.441 min Scan# 222
Delta R.T. -0.006 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion:142 Resp: 135732

Ion	Ratio	Lower	Upper
142	100		
127	50.6	40.2	60.4
141	14.5	11.7	17.5






#11
Tert butyl alcohol
Concen: 519.999 ug/l
RT: 2.989 min Scan# 3
Delta R.T. 0.012 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 59 Resp: 11004


Ion	Ratio	Lower	Upper
59	100		
57	10.9	8.6	12.8



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#12
1,1-Dichloroethene
Concen: 102.934 ug/l
RT: 2.306 min Scan# 200
Delta R.T. -0.006 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 96 Resp: 106404

Ion	Ratio	Lower	Upper
96	100		
61	183.3	145.7	218.5
98	63.3	51.6	77.4






#13
Acrolein
Concen: 501.914 ug/l
RT: 2.233 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 56 Resp: 14626


Ion	Ratio	Lower	Upper
56	100		
55	69.5	56.2	84.2



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#14
Allyl chloride
Concen: 105.276 ug/l
RT: 2.654 min Scan# 257
Delta R.T. -0.006 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 41 Resp: 206484

Ion	Ratio	Lower	Upper
41	100		
39	75.8	60.2	90.2
76	31.3	24.6	36.8






#15
Acrylonitrile
Concen: 507.426 ug/l
RT: 3.062 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 53 Resp: 33897


Ion	Ratio	Lower	Upper
53	100		
52	83.9	67.4	101.2
51	38.0	29.9	44.9



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#16
Acetone
Concen: 486.054 ug/l
RT: 2.386 min Scan# 213
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 43 Resp: 314285

Ion	Ratio	Lower	Upper
43	100		
58	27.8	22.3	33.5






#17
Carbon Disulfide
Concen: 108.215 ug/l
RT: 2.502 min Scan# 2
Delta R.T. -0.006 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 76 Resp: 27628


Ion	Ratio	Lower	Upper
76	100		
78	8.9	7.0	10.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#18
Methyl Acetate
Concen: 98.925 ug/l
RT: 2.703 min Scan# 265
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 43 Resp: 147272

Ion	Ratio	Lower	Upper
43	100		
74	22.2	17.5	26.3






#19
Methyl tert-butyl Ether
Concen: 105.561 ug/l
RT: 3.111 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 73 Resp: 38171


Ion	Ratio	Lower	Upper
73	100		
57	22.4	17.6	26.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#20
Methylene Chloride
Concen: 100.316 ug/l
RT: 2.782 min Scan# 278
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 84 Resp: 121435

Ion	Ratio	Lower	Upper
84	100		
49	137.2	107.5	161.3
51	42.1	33.0	49.6
86	63.5	50.1	75.1






#21
trans-1,2-Dichloroethene
Concen: 101.220 ug/l
RT: 3.081 min Scan# 3
Delta R.T. -0.006 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 96 Resp: 10689


Ion	Ratio	Lower	Upper
96	100		
61	153.1	121.0	181.6
98	63.7	49.7	74.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#22
Diisopropyl ether
Concen: 105.132 ug/l
RT: 3.757 min Scan# 438
Delta R.T. -0.006 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 45 Resp: 406185

Ion	Ratio	Lower	Upper
45	100		
43	80.1	59.3	88.9
87	24.7	20.4	30.6
59	10.6	9.0	13.6






#23
Vinyl Acetate
Concen: 553.853 ug/l
RT: 3.721 min Scan# 4
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 43 Resp: 184881


Ion	Ratio	Lower	Upper
43	100		
86	9.6	7.5	11.3



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#24
1,1-Dichloroethane
Concen: 101.114 ug/l
RT: 3.605 min Scan# 413
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 63 Resp: 220934

Ion	Ratio	Lower	Upper
63	100		
98	6.6	3.4	10.1
100	4.2	2.2	6.6






#25
2-Butanone
Concen: 519.018 ug/l
RT: 4.556 min Scan# 5
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 43 Resp: 49124


Ion	Ratio	Lower	Upper
43	100		
72	23.5	18.6	28.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#26
2,2-Dichloropropane
Concen: 110.215 ug/l
RT: 4.465 min Scan# 554
Delta R.T. -0.006 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 77 Resp: 155592

Ion	Ratio	Lower	Upper
77	100		
97	21.2	10.6	31.8






#27
cis-1,2-Dichloroethene
Concen: 100.871 ug/l
RT: 4.483 min Scan# 5
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 96 Resp: 12978


Ion	Ratio	Lower	Upper
96	100		
61	159.4	0.0	313.2
98	64.7	0.0	126.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#28
Bromochloromethane
Concen: 99.530 ug/l
RT: 4.891 min Scan# 624
Delta R.T. -0.006 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 49 Resp: 105093

Ion	Ratio	Lower	Upper
49	100		
129	1.8	0.0	3.8
130	68.4	54.6	81.8





#29
Tetrahydrofuran
Concen: 510.264 ug/l
RT: 5.001 min Scan# 64
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38


Tgt Ion: 42 Resp: 31312


Ion	Ratio	Lower	Upper
42	100		
72	42.3	33.4	50.0
71	39.1	31.4	47.0

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025







#30
Chloroform
Concen: 100.255 ug/l
RT: 5.086 min Scan# 656
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 83 Resp: 225402

Ion	Ratio	Lower	Upper
83	100		
85	65.1	52.8	79.2






#31
Cyclohexane
Concen: 102.056 ug/l
RT: 5.458 min Scan# 7
Delta R.T. -0.006 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 56 Resp: 19713


Ion	Ratio	Lower	Upper
56	100		
69	29.8	23.6	35.4
84	82.1	67.7	101.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#32
1,1,1-Trichloroethane
Concen: 104.346 ug/l
RT: 5.379 min Scan# 704
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 97 Resp: 198511

Ion	Ratio	Lower	Upper
97	100		
99	64.1	51.7	77.5
61	50.2	41.0	61.4






#33
1,2-Dichloroethane-d4
Concen: 98.914 ug/l
RT: 5.946 min Scan# 7
Delta R.T. -0.006 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 65 Resp: 15575


Ion	Ratio	Lower	Upper
65	100		
67	49.3	0.0	99.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.757 min Scan# 930
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion:114 Resp: 152724

Ion	Ratio	Lower	Upper
114	100		
63	24.4	0.0	46.8
88	16.6	0.0	35.4






#35
Dibromofluoromethane
Concen: 99.585 ug/l
RT: 5.379 min Scan# 7
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion:113 Resp: 10790


Ion	Ratio	Lower	Upper
113	100		
111	102.5	81.8	122.6
192	16.4	13.8	20.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#36
1,1-Dichloropropene
Concen: 103.515 ug/l
RT: 5.684 min Scan# 754
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 75 Resp: 151448

Ion	Ratio	Lower	Upper
75	100		
110	32.8	16.9	50.6
77	30.4	25.0	37.4






#37
Ethyl Acetate
Concen: 104.207 ug/l
RT: 4.715 min Scan# 5
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 43 Resp: 19213


Ion	Ratio	Lower	Upper
43	100		
61	12.9	9.9	14.9
70	9.8	7.5	11.3



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#38
Carbon Tetrachloride
Concen: 105.278 ug/l
RT: 5.672 min Scan# 752
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion:117 Resp: 166465

Ion	Ratio	Lower	Upper
117	100		
119	95.5	76.9	115.3
121	30.9	24.2	36.2






#39
Methylcyclohexane
Concen: 106.921 ug/l
RT: 7.373 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 83 Resp: 19175


Ion	Ratio	Lower	Upper
83	100		
55	80.4	64.9	97.3
98	47.6	37.4	56.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#40
Benzene
Concen: 101.369 ug/l
RT: 6.031 min Scan# 811
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 78 Resp: 452946

Ion	Ratio	Lower	Upper
78	100		
77	23.9	19.0	28.6






#41
Methacrylonitrile
Concen: 109.068 ug/l
RT: 4.922 min Scan# 6
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 41 Resp: 10703


Ion	Ratio	Lower	Upper
41	100		
39	57.9	47.0	70.6
67	66.2	53.2	79.8
52	30.7	25.0	37.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#42
1,2-Dichloroethane
Concen: 102.696 ug/l
RT: 6.080 min Scan# 819
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 62 Resp: 189528

Ion	Ratio	Lower	Upper
62	100		
98	7.4	0.0	15.2






#43
Isopropyl Acetate
Concen: 109.494 ug/l
RT: 6.336 min Scan# 8
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 43 Resp: 30597


Ion	Ratio	Lower	Upper
43	100		
61	17.6	14.1	21.1
87	11.6	9.2	13.8



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#44
Trichloroethene
Concen: 101.062 ug/l
RT: 7.123 min Scan# 990
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion:130 Resp: 107332

Ion	Ratio	Lower	Upper
130	100		
95	104.4	0.0	202.0






#45
1,2-Dichloropropane
Concen: 103.806 ug/l
RT: 7.421 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 63 Resp: 11573


Ion	Ratio	Lower	Upper
63	100		
65	30.8	25.8	38.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#46
Dibromomethane
Concen: 102.101 ug/l
RT: 7.574 min Scan# 1064
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 93 Resp: 87381

Ion	Ratio	Lower	Upper
93	100		
95	83.4	64.8	97.2
174	84.4	65.4	98.0






#47
Bromodichloromethane
Concen: 104.886 ug/l
RT: 7.818 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 83 Resp: 17926


Ion	Ratio	Lower	Upper
83	100		
85	63.8	49.7	74.5
127	8.2	6.1	9.1



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#48
Methyl methacrylate
Concen: 109.675 ug/l
RT: 7.690 min Scan# 1083
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 41 Resp: 158241

Ion	Ratio	Lower	Upper
41	100		
69	73.5	58.2	87.2
39	65.9	53.6	80.4





#49
1,4-Dioxane
Concen: 2015.906 ug/l
RT: 7.665 min Scan# 1
Delta R.T. 0.006 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38


Tgt Ion: 88 Resp: 5262


Ion	Ratio	Lower	Upper
88	100		
43	42.4	32.2	48.4
58	75.5	58.6	88.0

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025







#50
Toluene-d8
Concen: 99.309 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 98 Resp: 375599

Ion	Ratio	Lower	Upper
98	100		
100	65.7	52.2	78.4






#51
4-Methyl-2-Pentanone
Concen: 533.814 ug/l
RT: 8.574 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 43 Resp: 98888


Ion	Ratio	Lower	Upper
43	100		
58	36.1	28.5	42.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#52
Toluene
Concen: 102.199 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 92 Resp: 276323

Ion	Ratio	Lower	Upper
92	100		
91	172.5	138.7	208.1






#53
t-1,3-Dichloropropene
Concen: 100.645 ug/l
RT: 8.976 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 75 Resp: 16941


Ion	Ratio	Lower	Upper
75	100		
77	31.7	25.4	38.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#54
cis-1,3-Dichloropropene
Concen: 113.521 ug/l
RT: 8.366 min Scan# 1194
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 75 Resp: 182328

Ion	Ratio	Lower	Upper
75	100		
77	31.6	24.6	36.8
39	61.4	48.4	72.6






#55
1,1,2-Trichloroethane
Concen: 101.361 ug/l
RT: 9.147 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 97 Resp: 10920


Ion	Ratio	Lower	Upper
97	100		
83	90.8	71.8	107.8
85	58.2	47.0	70.6
99	63.0	48.6	73.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#56
Ethyl methacrylate
Concen: 115.168 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 69 Resp: 192034

Ion	Ratio	Lower	Upper
69	100		
41	76.5	61.3	91.9
39	49.0	40.2	60.2






#57
1,3-Dichloropropane
Concen: 102.588 ug/l
RT: 9.305 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 76 Resp: 19247


Ion	Ratio	Lower	Upper
76	100		
78	32.4	26.0	39.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#58
2-Chloroethyl Vinyl ether
Concen: 545.696 ug/l
RT: 8.238 min Scan# 1173
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 63 Resp: 460297

Ion	Ratio	Lower	Upper
63	100		
106	26.4	21.0	31.6






#59
2-Hexanone
Concen: 538.538 ug/l
RT: 9.427 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 43 Resp: 73877


Ion	Ratio	Lower	Upper
43	100		
58	50.3	24.9	74.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#60
Dibromochloromethane
Concen: 109.570 ug/l
RT: 9.519 min Scan# 1383
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion:129 Resp: 128700

Ion	Ratio	Lower	Upper
129	100		
127	76.8	39.4	118.2






#61
1,2-Dibromoethane
Concen: 106.282 ug/l
RT: 9.604 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion:107 Resp: 11591


Ion	Ratio	Lower	Upper
107	100		
109	93.7	75.7	113.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#62
4-Bromofluorobenzene
Concen: 106.664 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 95 Resp: 146944

Ion	Ratio	Lower	Upper
95	100		
174	67.8	0.0	135.8
176	64.7	0.0	131.4





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 14
Delta R.T. 0.006 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38


Tgt Ion:117 Resp: 13636


Ion	Ratio	Lower	Upper
117	100		
82	57.2	49.2	73.8
119	31.2	25.1	37.7

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025







#64
Tetrachloroethene
Concen: 94.325 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. -0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion:164 Resp: 90817

Ion	Ratio	Lower	Upper
164	100		
166	129.3	103.8	155.8
129	105.9	84.3	126.5
131	102.5	80.9	121.3






#65
Chlorobenzene
Concen: 101.682 ug/l
RT: 10.073 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion:112 Resp: 29630


Ion	Ratio	Lower	Upper
112	100		
114	31.9	24.7	37.1



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#66
1,1,1,2-Tetrachloroethane
Concen: 102.243 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion:131 Resp: 103437

Ion	Ratio	Lower	Upper
131	100		
133	96.1	47.5	142.5
119	65.0	32.9	98.6






#67
Ethyl Benzene
Concen: 104.152 ug/l
RT: 10.189 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 91 Resp: 54356


Ion	Ratio	Lower	Upper
91	100		
106	29.9	23.7	35.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#68
m/p-Xylenes
Concen: 207.694 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion:106 Resp: 394241

Ion	Ratio	Lower	Upper
106	100		
91	216.1	172.6	258.8






#69
o-Xylene
Concen: 104.487 ug/l
RT: 10.640 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion:106 Resp: 19538


Ion	Ratio	Lower	Upper
106	100		
91	228.0	113.6	340.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#70
Styrene
Concen: 108.654 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion:104 Resp: 335333

Ion	Ratio	Lower	Upper
104	100		
78	57.0	45.4	68.2
103	54.7	43.8	65.8






#71
Bromoform
Concen: 110.714 ug/l
RT: 10.799 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion:173 Resp: 8363


Ion	Ratio	Lower	Upper
173	100		
175	49.0	23.4	70.2
254	0.0	0.0	0.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion:152 Resp: 64612

Ion	Ratio	Lower	Upper
152	100		
115	121.0	46.9	140.7
150	188.4	0.0	349.4






#73
Isopropylbenzene
Concen: 100.944 ug/l
RT: 10.957 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion:105 Resp: 52243


Ion	Ratio	Lower	Upper
105	100		
120	25.2	12.7	38.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#74
N-amyl acetate
Concen: 111.267 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 43 Resp: 275462

Ion	Ratio	Lower	Upper
43	100		
70	38.1	30.1	45.1
55	22.4	18.3	27.5
61	22.1	17.5	26.3





#75
1,1,2,2-Tetrachloroethane
Concen: 96.243 ug/l
RT: 11.213 min Scan# 1
Delta R.T. 0.006 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38


Tgt Ion: 83 Resp: 17497


Ion	Ratio	Lower	Upper
83	100		
131	9.8	4.7	14.1
85	64.3	32.4	97.2

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025







#76
1,2,3-Trichloropropane
Concen: 97.744 ug/l m
RT: 11.238 min Scan# 1665
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 75 Resp: 154048

Ion	Ratio	Lower	Upper
75	100		
77	41.8	20.0	59.9






#77
Bromobenzene
Concen: 99.983 ug/l
RT: 11.195 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion:156 Resp: 11956


Ion	Ratio	Lower	Upper
156	100		
77	183.4	92.6	277.8
158	95.9	48.1	144.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#78
n-propylbenzene
Concen: 104.454 ug/l
RT: 11.299 min Scan# 1675
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 91 Resp: 622679

Ion	Ratio	Lower	Upper
91	100		
120	21.5	10.8	32.3





#79
2-Chlorotoluene
Concen: 98.840 ug/l
RT: 11.360 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38


Tgt Ion: 91 Resp: 37668


Ion	Ratio	Lower	Upper
91	100		
126	31.2	15.8	47.4

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025







#80
1,3,5-Trimethylbenzene
Concen: 101.675 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion:105 Resp: 435137

Ion	Ratio	Lower	Upper
105	100		
120	45.9	23.0	69.0





#81
trans-1,4-Dichloro-2-butene
Concen: 111.704 ug/l
RT: 11.018 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38


Tgt Ion: 75 Resp: 5187


Ion	Ratio	Lower	Upper
75	100		
53	119.5	101.8	152.6
89	46.9	37.2	55.8

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025







#82
4-Chlorotoluene
Concen: 101.877 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 91 Resp: 432884

Ion	Ratio	Lower	Upper
91	100		
126	26.9	13.3	39.9






#83
tert-Butylbenzene
Concen: 103.292 ug/l
RT: 11.713 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion:119 Resp: 43764


Ion	Ratio	Lower	Upper
119	100		
91	65.8	32.4	97.2
134	22.4	11.5	34.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#84
1,2,4-Trimethylbenzene
Concen: 103.414 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion:105 Resp: 444211

Ion	Ratio	Lower	Upper
105	100		
120	42.4	21.4	64.3






#85
sec-Butylbenzene
Concen: 105.082 ug/l
RT: 11.890 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion:105 Resp: 54685


Ion	Ratio	Lower	Upper
105	100		
134	19.3	9.6	28.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#86
p-Isopropyltoluene
Concen: 105.609 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion:119 Resp: 453109

Ion	Ratio	Lower	Upper
119	100		
134	25.1	12.6	37.8
91	27.4	13.7	41.1






#87
1,3-Dichlorobenzene
Concen: 101.749 ug/l
RT: 11.969 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion:146 Resp: 22148


Ion	Ratio	Lower	Upper
146	100		
111	44.1	22.2	66.6
148	63.1	31.9	95.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#88
1,4-Dichlorobenzene
Concen: 98.047 ug/l
RT: 12.036 min Scan# 1796
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion:146 Resp: 216367

Ion	Ratio	Lower	Upper
146	100		
111	43.2	20.9	62.8
148	64.0	31.5	94.5






#89
n-Butylbenzene
Concen: 113.075 ug/l
RT: 12.329 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 91 Resp: 42069


Ion	Ratio	Lower	Upper
91	100		
92	53.3	26.9	80.6
134	23.1	11.9	35.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#90
Hexachloroethane
Concen: 109.160 ug/l
RT: 12.536 min Scan# 1878
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion:117 Resp: 80489

Ion	Ratio	Lower	Upper
117	100		
201	62.1	31.2	93.6






#91
1,2-Dichlorobenzene
Concen: 99.419 ug/l
RT: 12.335 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion:146 Resp: 21515


Ion	Ratio	Lower	Upper
146	100		
111	46.6	22.9	68.8
148	63.7	31.9	95.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#92
1,2-Dibromo-3-Chloropropane
Concen: 113.214 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion: 75 Resp: 43047

Ion	Ratio	Lower	Upper
75	100		
155	70.2	34.9	104.7
157	90.6	44.9	134.7






#93
1,2,4-Trichlorobenzene
Concen: 112.106 ug/l
RT: 13.585 min Scan# 2
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion:180 Resp: 13504


Ion	Ratio	Lower	Upper
180	100		
182	95.1	47.9	143.6
145	34.6	17.9	53.8



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#94
Hexachlorobutadiene
Concen: 103.326 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. 0.006 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion:225 Resp: 53636

Ion	Ratio	Lower	Upper
225	100		
223	64.2	31.1	93.3
227	62.9	32.2	96.6






#95
Naphthalene
Concen: 111.473 ug/l
RT: 13.774 min Scan# 2
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion:128 Resp: 49971


Ion	Ratio	Lower	Upper
128	100		
127	13.0	10.3	15.5
129	11.0	8.6	12.8



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025



#96
1,2,3-Trichlorobenzene
Concen: 108.869 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. 0.000 min
Lab File: VX045529.D
Acq: 01 Apr 2025 18:38

Tgt Ion:180 Resp: 137328

Ion	Ratio	Lower	Upper
180	100		
182	94.5	47.3	141.8
145	36.0	18.2	54.6



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045530.D
Acq On : 01 Apr 2025 19:02
Operator : JC/MD
Sample : VSTDICC150
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED
Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025

Quant Time: Apr 02 02:54:59 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 02:44:48 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	5.544	168	90033	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	160037	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	133342	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	61380	50.000	ug/l	# 0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	252998	153.661	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 307.320%#			
35) Dibromofluoromethane	5.379	113	173598	152.887	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 305.780%#			
50) Toluene-d8	8.647	98	603264	152.216	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 304.440%#			
62) 4-Bromofluorobenzene	11.079	95	220921	153.035	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 306.060%#			
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.167	85	221436	164.460	ug/l	100
3) Chloromethane	1.307	50	198321	143.373	ug/l	98
4) Vinyl Chloride	1.374	62	197096	155.697	ug/l	100
5) Bromomethane	1.593	94	88225	146.986	ug/l	97
6) Chloroethane	1.660	64	86251	128.421	ug/l	97
7) Trichlorofluoromethane	1.868	101	267098	141.495	ug/l	100
8) Diethyl Ether	2.136	74	96467	152.215	ug/l	100
9) 1,1,2-Trichlorotrifluo...	2.319	101	171118	154.698	ug/l	97
10) Methyl Iodide	2.441	142	202594	146.838	ug/l	99
11) Tert butyl alcohol	2.989	59	167769	758.202	ug/l	100
12) 1,1-Dichloroethene	2.307	96	166365	153.921	ug/l	97
13) Acrolein	2.239	56	232818	764.070	ug/l	100
14) Allyl chloride	2.654	41	313268	152.755	ug/l	99
15) Acrylonitrile	3.069	53	507965	727.247	ug/l	100
16) Acetone	2.386	43	479564	709.323	ug/l	99
17) Carbon Disulfide	2.502	76	443495	166.133	ug/l	99
18) Methyl Acetate	2.703	43	228558	146.832	ug/l	99
19) Methyl tert-butyl Ether	3.117	73	598532	158.301	ug/l	99
20) Methylene Chloride	2.782	84	186382	147.254	ug/l	99
21) trans-1,2-Dichloroethene	3.081	96	170283	154.210	ug/l	98
22) Diisopropyl ether	3.764	45	636728	157.617	ug/l	94
23) Vinyl Acetate	3.721	43	2911326	834.120	ug/l	100
24) 1,1-Dichloroethane	3.605	63	349040	152.778	ug/l	99
25) 2-Butanone	4.556	43	740096	747.834	ug/l	99
26) 2,2-Dichloropropane	4.465	77	252063	170.765	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	205254	152.571	ug/l	98
28) Bromochloromethane	4.891	49	155600	140.938	ug/l	97
29) Tetrahydrofuran	5.001	42	475589	741.226	ug/l	100
30) Chloroform	5.087	83	353681	150.451	ug/l	98
31) Cyclohexane	5.458	56	310219	153.595	ug/l	97
32) 1,1,1-Trichloroethane	5.373	97	315136	158.426	ug/l	98
36) 1,1-Dichloropropene	5.684	75	239776	156.399	ug/l	98
37) Ethyl Acetate	4.715	43	294559	152.461	ug/l	99
38) Carbon Tetrachloride	5.672	117	266770	161.005	ug/l	99
39) Methylcyclohexane	7.373	83	303251	161.366	ug/l	96
40) Benzene	6.031	78	703311	150.208	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045530.D
Acq On : 01 Apr 2025 19:02
Operator : JC/MD
Sample : VSTDICC150
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Quant Time: Apr 02 02:54:59 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 02:44:48 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
41) Methacrylonitrile	4.922	41	164699	160.155	ug/l		98
42) 1,2-Dichloroethane	6.080	62	296168	153.146	ug/l		100
43) Isopropyl Acetate	6.336	43	475823	162.494	ug/l		99
44) Trichloroethene	7.117	130	170940	153.599	ug/l		95
45) 1,2-Dichloropropane	7.428	63	179457	153.610	ug/l		98
46) Dibromomethane	7.574	93	135415	150.996	ug/l		98
47) Bromodichloromethane	7.818	83	279881	156.272	ug/l		98
48) Methyl methacrylate	7.690	41	240818	159.281	ug/l		99
49) 1,4-Dioxane	7.659	88	74929	2739.196	ug/l		99
51) 4-Methyl-2-Pentanone	8.574	43	1414771	728.817	ug/l		99
52) Toluene	8.714	92	420015	148.245	ug/l		100
53) t-1,3-Dichloropropene	8.976	75	267671	150.589	ug/l		98
54) cis-1,3-Dichloropropene	8.360	75	287973	171.104	ug/l		98
55) 1,1,2-Trichloroethane	9.147	97	163426	144.752	ug/l		98
56) Ethyl methacrylate	9.116	69	285895	163.624	ug/l		98
57) 1,3-Dichloropropane	9.305	76	288430	146.708	ug/l		100
58) 2-Chloroethyl Vinyl ether	8.238	63	703881	796.340	ug/l		99
59) 2-Hexanone	9.433	43	1053971	733.198	ug/l		100
60) Dibromochloromethane	9.519	129	194342	157.895	ug/l		98
61) 1,2-Dibromoethane	9.610	107	174571	152.745	ug/l		99
64) Tetrachloroethene	9.269	164	138926	147.568	ug/l		98
65) Chlorobenzene	10.080	112	444631	156.049	ug/l		98
66) 1,1,1,2-Tetrachloroethane	10.159	131	156069	157.769	ug/l		99
67) Ethyl Benzene	10.189	91	821130	160.907	ug/l		99
68) m/p-Xylenes	10.299	106	583228	314.231	ug/l		99
69) o-Xylene	10.640	106	285670	156.241	ug/l		100
70) Styrene	10.653	104	480722	159.299	ug/l		99
71) Bromoform	10.799	173	126115	170.743	ug/l	#	98
73) Isopropylbenzene	10.957	105	764406	155.475	ug/l		100
74) N-amyl acetate	10.842	43	408934	173.877	ug/l		99
75) 1,1,2,2-Tetrachloroethane	11.213	83	249980	144.737	ug/l		100
76) 1,2,3-Trichloropropane	11.238	75	215694m	144.064	ug/l		
77) Bromobenzene	11.195	156	171573	151.033	ug/l		99
78) n-propylbenzene	11.305	91	897084	158.408	ug/l		100
79) 2-Chlorotoluene	11.360	91	537263	148.399	ug/l		99
80) 1,3,5-Trimethylbenzene	11.451	105	620819	152.700	ug/l		100
81) trans-1,4-Dichloro-2-b...	11.018	75	78805	178.636	ug/l		93
82) 4-Chlorotoluene	11.451	91	615594	152.505	ug/l		99
83) tert-Butylbenzene	11.713	119	627039	155.783	ug/l		100
84) 1,2,4-Trimethylbenzene	11.750	105	629704	154.317	ug/l		100
85) sec-Butylbenzene	11.890	105	790355	159.869	ug/l		100
86) p-Isopropyltoluene	12.006	119	641521	157.397	ug/l		100
87) 1,3-Dichlorobenzene	11.969	146	314077	151.885	ug/l		100
88) 1,4-Dichlorobenzene	12.037	146	314197	149.876	ug/l		99
89) n-Butylbenzene	12.329	91	604445	171.019	ug/l		100
90) Hexachloroethane	12.536	117	121761	173.829	ug/l		99
91) 1,2-Dichlorobenzene	12.335	146	307031	149.342	ug/l		100
92) 1,2-Dibromo-3-Chloropr...	12.939	75	64317	178.062	ug/l		99
93) 1,2,4-Trichlorobenzene	13.585	180	199365	174.213	ug/l		99
94) Hexachlorobutadiene	13.725	225	77792	157.752	ug/l		99
95) Naphthalene	13.774	128	733300	172.191	ug/l		100
96) 1,2,3-Trichlorobenzene	13.957	180	201989	168.562	ug/l		99

ICAL Form

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045530.D
Acq On : 01 Apr 2025 19:02
Operator : JC/MD
Sample : VSTDICC150
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025

Quant Time: Apr 02 02:54:59 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 02:44:48 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045530.D
Acq On : 01 Apr 2025 19:02
Operator : JC/MD
Sample : VSTDICC150
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1


Quant Time: Apr 02 02:54:59 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 02:44:48 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025






#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 7
Delta R.T. -0.006 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion:168 Resp: 9003


Ion	Ratio	Lower	Upper
168	100		
99	70.3	52.3	78.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#2
Dichlorodifluoromethane
Concen: 164.460 ug/l
RT: 1.167 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 85 Resp: 221436

Ion	Ratio	Lower	Upper
85	100		
87	32.0	16.1	48.2






#3
Chloromethane
Concen: 143.373 ug/l
RT: 1.307 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 50 Resp: 19832


Ion	Ratio	Lower	Upper
50	100		
52	33.6	25.9	38.9



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#4
Vinyl Chloride
Concen: 155.697 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 62 Resp: 197096

Ion	Ratio	Lower	Upper
62	100		
64	32.1	25.5	38.3






#5
Bromomethane
Concen: 146.986 ug/l
RT: 1.593 min Scan# 8
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 94 Resp: 8822


Ion	Ratio	Lower	Upper
94	100		
96	96.1	74.8	112.2



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#6
Chloroethane
Concen: 128.421 ug/l
RT: 1.660 min Scan# 94
Delta R.T. -0.006 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 64 Resp: 86251

Ion	Ratio	Lower	Upper
64	100		
66	31.4	26.3	39.5






#7
Trichlorofluoromethane
Concen: 141.495 ug/l
RT: 1.868 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion:101 Resp: 26709


Ion	Ratio	Lower	Upper
101	100		
103	63.8	51.1	76.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#8
Diethyl Ether
Concen: 152.215 ug/l
RT: 2.136 min Scan# 172
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 74 Resp: 96467

Ion	Ratio	Lower	Upper
74	100		
45	105.8	53.0	159.1





#9
1,1,2-Trichlorotrifluoroethane
Concen: 154.698 ug/l
RT: 2.319 min Scan# 2
Delta R.T. -0.006 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02


Tgt Ion:101 Resp: 17111


Ion	Ratio	Lower	Upper
101	100		
85	46.9	38.9	58.3
151	69.3	57.5	86.3

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations
APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025







#10
Methyl Iodide
Concen: 146.838 ug/l
RT: 2.441 min Scan# 222
Delta R.T. -0.006 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion:142 Resp: 202594

Ion	Ratio	Lower	Upper
142	100		
127	51.2	40.2	60.4
141	14.3	11.7	17.5






#11
Tert butyl alcohol
Concen: 758.202 ug/l
RT: 2.989 min Scan# 3
Delta R.T. 0.012 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 59 Resp: 16776


Ion	Ratio	Lower	Upper
59	100		
57	10.7	8.6	12.8



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#12
1,1-Dichloroethene
Concen: 153.921 ug/l
RT: 2.307 min Scan# 200
Delta R.T. -0.006 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 96 Resp: 166365

Ion	Ratio	Lower	Upper
96	100		
61	187.0	145.7	218.5
98	65.0	51.6	77.4






#13
Acrolein
Concen: 764.070 ug/l
RT: 2.239 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 56 Resp: 23281


Ion	Ratio	Lower	Upper
56	100		
55	70.2	56.2	84.2



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#14
Allyl chloride
Concen: 152.755 ug/l
RT: 2.654 min Scan# 257
Delta R.T. -0.006 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 41 Resp: 313268

Ion	Ratio	Lower	Upper
41	100		
39	75.0	60.2	90.2
76	32.2	24.6	36.8






#15
Acrylonitrile
Concen: 727.247 ug/l
RT: 3.069 min Scan# 3
Delta R.T. 0.006 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 53 Resp: 50796


Ion	Ratio	Lower	Upper
53	100		
52	83.8	67.4	101.2
51	37.7	29.9	44.9



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#16
Acetone
Concen: 709.323 ug/l
RT: 2.386 min Scan# 213
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 43 Resp: 479564

Ion	Ratio	Lower	Upper
43	100		
58	27.2	22.3	33.5






#17
Carbon Disulfide
Concen: 166.133 ug/l
RT: 2.502 min Scan# 2
Delta R.T. -0.006 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 76 Resp: 44349


Ion	Ratio	Lower	Upper
76	100		
78	9.1	7.0	10.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#18
Methyl Acetate
Concen: 146.832 ug/l
RT: 2.703 min Scan# 265
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 43 Resp: 228558

Ion	Ratio	Lower	Upper
43	100		
74	22.4	17.5	26.3





#19
Methyl tert-butyl Ether
Concen: 158.301 ug/l
RT: 3.117 min Scan# 3
Delta R.T. 0.006 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02


Tgt Ion: 73 Resp: 59853


Ion	Ratio	Lower	Upper
73	100		
57	21.7	17.6	26.4

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations
APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025







#20
Methylene Chloride
Concen: 147.254 ug/l
RT: 2.782 min Scan# 278
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 84 Resp: 186382

Ion	Ratio	Lower	Upper
84	100		
49	135.1	107.5	161.3
51	40.8	33.0	49.6
86	63.4	50.1	75.1






#21
trans-1,2-Dichloroethene
Concen: 154.210 ug/l
RT: 3.081 min Scan# 3
Delta R.T. -0.006 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 96 Resp: 17028


Ion	Ratio	Lower	Upper
96	100		
61	153.7	121.0	181.6
98	62.7	49.7	74.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#22
Diisopropyl ether
Concen: 157.617 ug/l
RT: 3.764 min Scan# 439
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 45 Resp: 636728

Ion	Ratio	Lower	Upper
45	100		
43	67.0	59.3	88.9
87	25.5	20.4	30.6
59	10.6	9.0	13.6






#23
Vinyl Acetate
Concen: 834.120 ug/l
RT: 3.721 min Scan# 4
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 43 Resp: 291132


Ion	Ratio	Lower	Upper
43	100		
86	9.6	7.5	11.3



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#24
1,1-Dichloroethane
Concen: 152.778 ug/l
RT: 3.605 min Scan# 413
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 63 Resp: 349040

Ion	Ratio	Lower	Upper
63	100		
98	6.4	3.4	10.1
100	3.9	2.2	6.6





#25
2-Butanone
Concen: 747.834 ug/l
RT: 4.556 min Scan# 5
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02


Tgt Ion: 43 Resp: 74009


Ion	Ratio	Lower	Upper
43	100		
72	23.6	18.6	28.0

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations
APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025







#26
2,2-Dichloropropane
Concen: 170.765 ug/l
RT: 4.465 min Scan# 554
Delta R.T. -0.006 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 77 Resp: 252063

Ion	Ratio	Lower	Upper
77	100		
97	21.2	10.6	31.8






#27
cis-1,2-Dichloroethene
Concen: 152.571 ug/l
RT: 4.483 min Scan# 5
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 96 Resp: 20525


Ion	Ratio	Lower	Upper
96	100		
61	158.2	0.0	313.2
98	64.9	0.0	126.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#28
Bromochloromethane
Concen: 140.938 ug/l
RT: 4.891 min Scan# 624
Delta R.T. -0.006 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 49 Resp: 155600

Ion	Ratio	Lower	Upper
49	100		
129	1.8	0.0	3.8
130	70.7	54.6	81.8






#29
Tetrahydrofuran
Concen: 741.226 ug/l
RT: 5.001 min Scan# 64
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 42 Resp: 47558


Ion	Ratio	Lower	Upper
42	100		
72	42.3	33.4	50.0
71	39.2	31.4	47.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#30
Chloroform
Concen: 150.451 ug/l
RT: 5.087 min Scan# 656
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 83 Resp: 353681

Ion	Ratio	Lower	Upper
83	100		
85	64.2	52.8	79.2






#31
Cyclohexane
Concen: 153.595 ug/l
RT: 5.458 min Scan# 7
Delta R.T. -0.006 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 56 Resp: 31021


Ion	Ratio	Lower	Upper
56	100		
69	30.7	23.6	35.4
84	81.9	67.7	101.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#32
1,1,1-Trichloroethane
Concen: 158.426 ug/l
RT: 5.373 min Scan# 703
Delta R.T. -0.006 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 97 Resp: 315136

Ion	Ratio	Lower	Upper
97	100		
99	63.7	51.7	77.5
61	49.1	41.0	61.4






#33
1,2-Dichloroethane-d4
Concen: 153.661 ug/l
RT: 5.952 min Scan# 7
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 65 Resp: 25299


Ion	Ratio	Lower	Upper
65	100		
67	49.4	0.0	99.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.757 min Scan# 930
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion:114 Resp: 160037

Ion	Ratio	Lower	Upper
114	100		
63	24.3	0.0	46.8
88	17.6	0.0	35.4






#35
Dibromofluoromethane
Concen: 152.887 ug/l
RT: 5.379 min Scan# 7
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion:113 Resp: 17359


Ion	Ratio	Lower	Upper
113	100		
111	103.4	81.8	122.6
192	17.3	13.8	20.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#36
1,1-Dichloropropene
Concen: 156.399 ug/l
RT: 5.684 min Scan# 754
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 75 Resp: 239776

Ion	Ratio	Lower	Upper
75	100		
110	32.4	16.9	50.6
77	30.6	25.0	37.4






#37
Ethyl Acetate
Concen: 152.461 ug/l
RT: 4.715 min Scan# 5
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 43 Resp: 29455


Ion	Ratio	Lower	Upper
43	100		
61	12.5	9.9	14.9
70	10.0	7.5	11.3



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#38
Carbon Tetrachloride
Concen: 161.005 ug/l
RT: 5.672 min Scan# 752
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion:117 Resp: 266770

Ion	Ratio	Lower	Upper
117	100		
119	95.9	76.9	115.3
121	31.2	24.2	36.2






#39
Methylcyclohexane
Concen: 161.366 ug/l
RT: 7.373 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 83 Resp: 30325


Ion	Ratio	Lower	Upper
83	100		
55	85.4	64.9	97.3
98	49.0	37.4	56.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#40
Benzene
Concen: 150.208 ug/l
RT: 6.031 min Scan# 811
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 78 Resp: 703311

Ion	Ratio	Lower	Upper
78	100		
77	24.0	19.0	28.6






#41
Methacrylonitrile
Concen: 160.155 ug/l
RT: 4.922 min Scan# 6
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 41 Resp: 16469


Ion	Ratio	Lower	Upper
41	100		
39	57.4	47.0	70.6
67	65.5	53.2	79.8
52	29.9	25.0	37.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#42
1,2-Dichloroethane
Concen: 153.146 ug/l
RT: 6.080 min Scan# 819
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 62 Resp: 296168

Ion	Ratio	Lower	Upper
62	100		
98	7.6	0.0	15.2






#43
Isopropyl Acetate
Concen: 162.494 ug/l
RT: 6.336 min Scan# 8
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 43 Resp: 47582


Ion	Ratio	Lower	Upper
43	100		
61	18.1	14.1	21.1
87	11.8	9.2	13.8



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#44
Trichloroethene
Concen: 153.599 ug/l
RT: 7.117 min Scan# 989
Delta R.T. -0.006 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion:130 Resp: 170940

Ion	Ratio	Lower	Upper
130	100		
95	106.1	0.0	202.0





#45
1,2-Dichloropropane
Concen: 153.610 ug/l
RT: 7.428 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02


Tgt Ion: 63 Resp: 17945

Ion	Ratio	Lower	Upper
63	100		
65	31.3	25.8	38.6

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025



#46
Dibromomethane
Concen: 150.996 ug/l
RT: 7.574 min Scan# 1064
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 93 Resp: 135415


Ion	Ratio	Lower	Upper
93	100		
95	82.4	64.8	97.2
174	84.1	65.4	98.0



#47
Bromodichloromethane
Concen: 156.272 ug/l
RT: 7.818 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 83 Resp: 27988


Ion	Ratio	Lower	Upper
83	100		
85	63.6	49.7	74.5
127	8.1	6.1	9.1



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#48
Methyl methacrylate
Concen: 159.281 ug/l
RT: 7.690 min Scan# 1083
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 41 Resp: 240818

Ion	Ratio	Lower	Upper
41	100		
69	73.0	58.2	87.2
39	64.9	53.6	80.4






#49
1,4-Dioxane
Concen: 2739.196 ug/l
RT: 7.659 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 88 Resp: 7492


Ion	Ratio	Lower	Upper
88	100		
43	40.8	32.2	48.4
58	74.2	58.6	88.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#50
Toluene-d8
Concen: 152.216 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 98 Resp: 603264

Ion	Ratio	Lower	Upper
98	100		
100	65.5	52.2	78.4






#51
4-Methyl-2-Pentanone
Concen: 728.817 ug/l
RT: 8.574 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 43 Resp: 141477


Ion	Ratio	Lower	Upper
43	100		
58	36.2	28.5	42.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#52
Toluene
Concen: 148.245 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 92 Resp: 420015

Ion	Ratio	Lower	Upper
92	100		
91	173.3	138.7	208.1






#53
t-1,3-Dichloropropene
Concen: 150.589 ug/l
RT: 8.976 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 75 Resp: 26767


Ion	Ratio	Lower	Upper
75	100		
77	30.6	25.4	38.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#54
cis-1,3-Dichloropropene
Concen: 171.104 ug/l
RT: 8.360 min Scan# 1193
Delta R.T. -0.006 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 75 Resp: 287973

Ion	Ratio	Lower	Upper
75	100		
77	30.6	24.6	36.8
39	62.9	48.4	72.6






#55
1,1,2-Trichloroethane
Concen: 144.752 ug/l
RT: 9.147 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 97 Resp: 16342


Ion	Ratio	Lower	Upper
97	100		
83	88.2	71.8	107.8
85	55.9	47.0	70.6
99	62.2	48.6	73.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#56
Ethyl methacrylate
Concen: 163.624 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 69 Resp: 285895

Ion	Ratio	Lower	Upper
69	100		
41	76.2	61.3	91.9
39	48.0	40.2	60.2






#57
1,3-Dichloropropane
Concen: 146.708 ug/l
RT: 9.305 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 76 Resp: 28843


Ion	Ratio	Lower	Upper
76	100		
78	32.2	26.0	39.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#58
2-Chloroethyl Vinyl ether
Concen: 796.340 ug/l
RT: 8.238 min Scan# 1173
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 63 Resp: 703881

Ion	Ratio	Lower	Upper
63	100		
106	26.6	21.0	31.6





#59
2-Hexanone
Concen: 733.198 ug/l
RT: 9.433 min Scan# 1
Delta R.T. 0.006 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02


Tgt Ion: 43 Resp: 105397


Ion	Ratio	Lower	Upper
43	100		
58	49.8	24.9	74.6

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025







#60
Dibromochloromethane
Concen: 157.895 ug/l
RT: 9.519 min Scan# 1383
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion:129 Resp: 194342

Ion	Ratio	Lower	Upper
129	100		
127	77.4	39.4	118.2






#61
1,2-Dibromoethane
Concen: 152.745 ug/l
RT: 9.610 min Scan# 1
Delta R.T. 0.006 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion:107 Resp: 17457


Ion	Ratio	Lower	Upper
107	100		
109	93.9	75.7	113.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#62
4-Bromofluorobenzene
Concen: 153.035 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 95 Resp: 220921

Ion	Ratio	Lower	Upper
95	100		
174	66.9	0.0	135.8
176	64.3	0.0	131.4





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 14
Delta R.T. 0.006 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02


Tgt Ion:117 Resp: 13334


Ion	Ratio	Lower	Upper
117	100		
82	59.5	49.2	73.8
119	32.7	25.1	37.7

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025







#64
Tetrachloroethene
Concen: 147.568 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. -0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion:164 Resp: 138926

Ion	Ratio	Lower	Upper
164	100		
166	128.9	103.8	155.8
129	107.9	84.3	126.5
131	103.4	80.9	121.3






#65
Chlorobenzene
Concen: 156.049 ug/l
RT: 10.080 min Scan# 14
Delta R.T. 0.006 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion:112 Resp: 44463


Ion	Ratio	Lower	Upper
112	100		
114	31.7	24.7	37.1



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#66
1,1,1,2-Tetrachloroethane
Concen: 157.769 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion:131 Resp: 156069

Ion	Ratio	Lower	Upper
131	100		
133	94.6	47.5	142.5
119	64.4	32.9	98.6






#67
Ethyl Benzene
Concen: 160.907 ug/l
RT: 10.189 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 91 Resp: 82113


Ion	Ratio	Lower	Upper
91	100		
106	29.1	23.7	35.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#68
m/p-Xylenes
Concen: 314.231 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion:106 Resp: 583228

Ion	Ratio	Lower	Upper
106	100		
91	216.7	172.6	258.8






#69
o-Xylene
Concen: 156.241 ug/l
RT: 10.640 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion:106 Resp: 28567


Ion	Ratio	Lower	Upper
106	100		
91	227.6	113.6	340.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#70
Styrene
Concen: 159.299 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion:104 Resp: 480722

Ion	Ratio	Lower	Upper
104	100		
78	57.9	45.4	68.2
103	55.4	43.8	65.8





#71
Bromoform
Concen: 170.743 ug/l
RT: 10.799 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02


Tgt Ion:173 Resp: 12611

Ion	Ratio	Lower	Upper
173	100		
175	48.0	23.4	70.2
254	0.1	0.0	0.0#

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion:152 Resp: 61380


Ion	Ratio	Lower	Upper
152	100		
115	151.4	46.9	140.7#
150	208.6	0.0	349.4



#73
Isopropylbenzene
Concen: 155.475 ug/l
RT: 10.957 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion:105 Resp: 76440


Ion	Ratio	Lower	Upper
105	100		
120	25.3	12.7	38.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#74
N-amyl acetate
Concen: 173.877 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 43 Resp: 408934

Ion	Ratio	Lower	Upper
43	100		
70	38.3	30.1	45.1
55	22.0	18.3	27.5
61	22.3	17.5	26.3






#75
1,1,2,2-Tetrachloroethane
Concen: 144.737 ug/l
RT: 11.213 min Scan# 1
Delta R.T. 0.006 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 83 Resp: 24998


Ion	Ratio	Lower	Upper
83	100		
131	9.7	4.7	14.1
85	64.8	32.4	97.2



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#76
1,2,3-Trichloropropane
Concen: 144.064 ug/l m
RT: 11.238 min Scan# 1665
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 75 Resp: 215694

Ion	Ratio	Lower	Upper
75	100		
77	42.3	20.0	59.9





#77
Bromobenzene
Concen: 151.033 ug/l
RT: 11.195 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02


Tgt Ion:156 Resp: 17157


Ion	Ratio	Lower	Upper
156	100		
77	185.8	92.6	277.8
158	95.3	48.1	144.4

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025







#78
n-propylbenzene
Concen: 158.408 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. 0.006 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 91 Resp: 897084

Ion	Ratio	Lower	Upper
91	100		
120	21.5	10.8	32.3





#79
2-Chlorotoluene
Concen: 148.399 ug/l
RT: 11.360 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02


Tgt Ion: 91 Resp: 53726


Ion	Ratio	Lower	Upper
91	100		
126	30.9	15.8	47.4

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations
APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025







#80
1,3,5-Trimethylbenzene
Concen: 152.700 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion:105 Resp: 620819

Ion	Ratio	Lower	Upper
105	100		
120	46.1	23.0	69.0





#81
trans-1,4-Dichloro-2-butene
Concen: 178.636 ug/l
RT: 11.018 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02


Tgt Ion: 75 Resp: 7880

Ion	Ratio	Lower	Upper
75	100		
53	116.0	101.8	152.6
89	46.8	37.2	55.8

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025



#82
4-Chlorotoluene
Concen: 152.505 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 91 Resp: 615594


Ion	Ratio	Lower	Upper
91	100		
126	26.9	13.3	39.9



#83
tert-Butylbenzene
Concen: 155.783 ug/l
RT: 11.713 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion:119 Resp: 62703


Ion	Ratio	Lower	Upper
119	100		
91	65.2	32.4	97.2
134	22.7	11.5	34.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#84
1,2,4-Trimethylbenzene
Concen: 154.317 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion:105 Resp: 629704

Ion	Ratio	Lower	Upper
105	100		
120	42.7	21.4	64.3






#85
sec-Butylbenzene
Concen: 159.869 ug/l
RT: 11.890 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion:105 Resp: 79035


Ion	Ratio	Lower	Upper
105	100		
134	18.9	9.6	28.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#86
p-Isopropyltoluene
Concen: 157.397 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion:119 Resp: 641521

Ion	Ratio	Lower	Upper
119	100		
134	25.3	12.6	37.8
91	27.6	13.7	41.1





#87
1,3-Dichlorobenzene
Concen: 151.885 ug/l
RT: 11.969 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02


Tgt Ion:146 Resp: 31407


Ion	Ratio	Lower	Upper
146	100		
111	44.6	22.2	66.6
148	63.6	31.9	95.7

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations
APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025







#88
1,4-Dichlorobenzene
Concen: 149.876 ug/l
RT: 12.037 min Scan# 1796
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion:146 Resp: 314197

Ion	Ratio	Lower	Upper
146	100		
111	42.3	20.9	62.8
148	63.5	31.5	94.5






#89
n-Butylbenzene
Concen: 171.019 ug/l
RT: 12.329 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 91 Resp: 60444


Ion	Ratio	Lower	Upper
91	100		
92	53.9	26.9	80.6
134	23.6	11.9	35.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#90
Hexachloroethane
Concen: 173.829 ug/l
RT: 12.536 min Scan# 1878
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion:117 Resp: 121761

Ion	Ratio	Lower	Upper
117	100		
201	61.9	31.2	93.6






#91
1,2-Dichlorobenzene
Concen: 149.342 ug/l
RT: 12.335 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion:146 Resp: 30703


Ion	Ratio	Lower	Upper
146	100		
111	46.3	22.9	68.8
148	63.8	31.9	95.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#92
1,2-Dibromo-3-Chloropropane
Concen: 178.062 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion: 75 Resp: 64317

Ion	Ratio	Lower	Upper
75	100		
155	69.7	34.9	104.7
157	91.3	44.9	134.7






#93
1,2,4-Trichlorobenzene
Concen: 174.213 ug/l
RT: 13.585 min Scan# 2
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion:180 Resp: 19936


Ion	Ratio	Lower	Upper
180	100		
182	95.4	47.9	143.6
145	34.6	17.9	53.8



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#94
Hexachlorobutadiene
Concen: 157.752 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. 0.006 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion:225 Resp: 77792

Ion	Ratio	Lower	Upper
225	100		
223	63.4	31.1	93.3
227	63.8	32.2	96.6






#95
Naphthalene
Concen: 172.191 ug/l
RT: 13.774 min Scan# 2
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion:128 Resp: 73330


Ion	Ratio	Lower	Upper
128	100		
127	12.8	10.3	15.5
129	10.7	8.6	12.8



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025



#96
1,2,3-Trichlorobenzene
Concen: 168.562 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. 0.000 min
Lab File: VX045530.D
Acq: 01 Apr 2025 19:02

Tgt Ion:180 Resp: 201989

Ion	Ratio	Lower	Upper
180	100		
182	94.0	47.3	141.8
145	35.8	18.2	54.6



Quantitation Report (QT Reviewed)

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045532.D
Acq On : 01 Apr 2025 19:48
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Manual Integrations APPROVED
Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025

Quant Time: Apr 02 03:14:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	5.544	168	96972	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	174855	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	152417	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	67633	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	89251	50.329	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 100.660%			
35) Dibromofluoromethane	5.379	113	61107	49.256	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 98.520%			
50) Toluene-d8	8.647	98	217200	50.160	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 100.320%			
62) 4-Bromofluorobenzene	11.079	95	80376	50.959	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 101.920%			
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.166	85	76070	52.454	ug/l	97
3) Chloromethane	1.301	50	72033	48.349	ug/l	99
4) Vinyl Chloride	1.374	62	66559	48.816	ug/l	99
5) Bromomethane	1.593	94	30379	46.991	ug/l	99
6) Chloroethane	1.672	64	37317	51.586	ug/l	100
7) Trichlorofluoromethane	1.880	101	102379	50.354	ug/l	96
8) Diethyl Ether	2.130	74	33603	49.228	ug/l	99
9) 1,1,2-Trichlorotrifluo...	2.325	101	59007	49.528	ug/l	98
10) Methyl Iodide	2.447	142	72261	48.626	ug/l	99
11) Tert butyl alcohol	2.971	59	58774	246.612	ug/l	99
12) 1,1-Dichloroethene	2.313	96	57888	49.726	ug/l	99
13) Acrolein	2.233	56	78156	238.141	ug/l	99
14) Allyl chloride	2.660	41	111636	50.541	ug/l	99
15) Acrylonitrile	3.062	53	186837	248.351	ug/l	100
16) Acetone	2.380	43	176153	241.904	ug/l	99
17) Carbon Disulfide	2.508	76	149677	52.057	ug/l	100
18) Methyl Acetate	2.703	43	81720	48.742	ug/l	100
19) Methyl tert-butyl Ether	3.111	73	207487	50.950	ug/l	100
20) Methylene Chloride	2.782	84	66875	49.055	ug/l	99
21) trans-1,2-Dichloroethene	3.087	96	59229	49.800	ug/l	98
22) Diisopropyl ether	3.757	45	221449	50.895	ug/l	92
23) Vinyl Acetate	3.715	43	1005423	267.449	ug/l	99
24) 1,1-Dichloroethane	3.605	63	122240	49.677	ug/l	99
25) 2-Butanone	4.550	43	271686	254.882	ug/l	97
26) 2,2-Dichloropropane	4.471	77	82704	52.020	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	71072	49.049	ug/l	98
28) Bromochloromethane	4.891	49	57557	48.403	ug/l	99
29) Tetrahydrofuran	5.001	42	172405	249.473	ug/l	100
30) Chloroform	5.086	83	125621	49.614	ug/l	98
31) Cyclohexane	5.458	56	107862	49.583	ug/l	97
32) 1,1,1-Trichloroethane	5.379	97	107830	50.330	ug/l	99
36) 1,1-Dichloropropene	5.684	75	83901	50.088	ug/l	99
37) Ethyl Acetate	4.708	43	105095	49.787	ug/l	98
38) Carbon Tetrachloride	5.672	117	91723	50.667	ug/l	100
39) Methylcyclohexane	7.373	83	102919	50.124	ug/l	99
40) Benzene	6.031	78	251758	49.212	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045532.D
Acq On : 01 Apr 2025 19:48
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Quant Time: Apr 02 03:14:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Manual Integrations APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025

Compound	R.T.	QIon	Response	Conc Units		Dev(Min)
41) Methacrylonitrile	4.916	41	59140	53.098 ug/l		99
42) 1,2-Dichloroethane	6.080	62	103270	48.875 ug/l		100
43) Isopropyl Acetate	6.336	43	164894	51.539 ug/l		100
44) Trichloroethene	7.123	130	57931	47.643 ug/l		92
45) 1,2-Dichloropropane	7.421	63	63210	49.521 ug/l		95
46) Dibromomethane	7.574	93	47564	48.542 ug/l		97
47) Bromodichloromethane	7.818	83	96418	49.273 ug/l		96
48) Methyl methacrylate	7.690	41	85084	51.507 ug/l		99
49) 1,4-Dioxane	7.659	88	28552	955.327 ug/l		99
51) 4-Methyl-2-Pentanone	8.568	43	540822	254.994 ug/l		100
52) Toluene	8.714	92	151664	48.994 ug/l		100
53) t-1,3-Dichloropropene	8.976	75	87744	46.783 ug/l		98
54) cis-1,3-Dichloropropene	8.360	75	99074	53.878 ug/l		98
55) 1,1,2-Trichloroethane	9.147	97	60452	49.007 ug/l		97
56) Ethyl methacrylate	9.116	69	102525	53.705 ug/l		100
57) 1,3-Dichloropropane	9.305	76	107121	49.869 ug/l		100
58) 2-Chloroethyl Vinyl ether	8.238	63	251285	260.201 ug/l		99
59) 2-Hexanone	9.427	43	395336	251.710 ug/l		100
60) Dibromochloromethane	9.519	129	68562	50.983 ug/l		99
61) 1,2-Dibromoethane	9.604	107	63301	50.693 ug/l		100
64) Tetrachloroethene	9.269	164	50834	47.239 ug/l		99
65) Chlorobenzene	10.073	112	163518	50.207 ug/l		98
66) 1,1,1,2-Tetrachloroethane	10.159	131	57096	50.495 ug/l		99
67) Ethyl Benzene	10.189	91	300747	51.558 ug/l		100
68) m/p-Xylenes	10.299	106	217943	102.728 ug/l		99
69) o-Xylene	10.640	106	107947	51.650 ug/l		100
70) Styrene	10.653	104	181524	52.624 ug/l		100
71) Bromoform	10.799	173	44489	52.694 ug/l	#	98
73) Isopropylbenzene	10.957	105	286010	52.794 ug/l		100
74) N-amyl acetate	10.842	43	140941	54.387 ug/l		99
75) 1,1,2,2-Tetrachloroethane	11.207	83	92505	48.608 ug/l		99
76) 1,2,3-Trichloropropane	11.238	75	82060m	49.741 ug/l		
77) Bromobenzene	11.195	156	63244	50.525 ug/l		98
78) n-propylbenzene	11.299	91	336753	53.967 ug/l		100
79) 2-Chlorotoluene	11.360	91	203677	51.057 ug/l		99
80) 1,3,5-Trimethylbenzene	11.445	105	239822	53.534 ug/l		100
81) trans-1,4-Dichloro-2-b...	11.018	75	25511	52.482 ug/l		98
82) 4-Chlorotoluene	11.451	91	233159	52.422 ug/l		100
83) tert-Butylbenzene	11.713	119	236276	53.274 ug/l		100
84) 1,2,4-Trimethylbenzene	11.750	105	238257	52.990 ug/l		100
85) sec-Butylbenzene	11.890	105	293633	53.903 ug/l		100
86) p-Isopropyltoluene	12.006	119	244031	54.337 ug/l		99
87) 1,3-Dichlorobenzene	11.963	146	115113	50.521 ug/l		100
88) 1,4-Dichlorobenzene	12.036	146	113458	49.117 ug/l		97
89) n-Butylbenzene	12.329	91	214931	55.189 ug/l		99
90) Hexachloroethane	12.536	117	41729	54.066 ug/l		98
91) 1,2-Dichlorobenzene	12.329	146	113126	49.938 ug/l		100
92) 1,2-Dibromo-3-Chloropr...	12.939	75	21184	53.226 ug/l		100
93) 1,2,4-Trichlorobenzene	13.585	180	65488	51.935 ug/l		98
94) Hexachlorobutadiene	13.725	225	27147	49.961 ug/l		99
95) Naphthalene	13.774	128	247600	52.765 ug/l		100
96) 1,2,3-Trichlorobenzene	13.957	180	66542	50.396 ug/l		100

ICAL Form

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045532.D
Acq On : 01 Apr 2025 19:48
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Manual Integrations APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025

Quant Time: Apr 02 03:14:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045532.D
Acq On : 01 Apr 2025 19:48
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1


Quant Time: Apr 02 03:14:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Manual Integrations APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025






#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 7
Delta R.T. -0.006 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion:168 Resp: 9697


Ion	Ratio	Lower	Upper
168	100		
99	71.1	52.3	78.5



Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#2
Dichlorodifluoromethane
Concen: 52.454 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion: 85 Resp: 76070

Ion	Ratio	Lower	Upper
85	100		
87	33.6	16.1	48.2






#3
Chloromethane
Concen: 48.349 ug/l
RT: 1.301 min Scan# 3
Delta R.T. -0.006 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion: 50 Resp: 7203


Ion	Ratio	Lower	Upper
50	100		
52	31.6	25.9	38.9



Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#4
Vinyl Chloride
Concen: 48.816 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion: 62 Resp: 66559

Ion	Ratio	Lower	Upper
62	100		
64	32.5	25.5	38.3






#5
Bromomethane
Concen: 46.991 ug/l
RT: 1.593 min Scan# 8
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion: 94 Resp: 3037


Ion	Ratio	Lower	Upper
94	100		
96	94.7	74.8	112.2



Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#6
Chloroethane
Concen: 51.586 ug/l
RT: 1.672 min Scan# 96
Delta R.T. 0.006 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion: 64 Resp: 37317

Ion	Ratio	Lower	Upper
64	100		
66	33.0	26.3	39.5






#7
Trichlorofluoromethane
Concen: 50.354 ug/l
RT: 1.880 min Scan# 1
Delta R.T. 0.006 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion:101 Resp: 10237


Ion	Ratio	Lower	Upper
101	100		
103	67.2	51.1	76.7



Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#8
Diethyl Ether
Concen: 49.228 ug/l
RT: 2.130 min Scan# 171
Delta R.T. -0.006 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion: 74 Resp: 33603

Ion	Ratio	Lower	Upper
74	100		
45	105.2	53.0	159.1






#9
1,1,2-Trichlorotrifluoroethane
Concen: 49.528 ug/l
RT: 2.325 min Scan# 2
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion:101 Resp: 5900


Ion	Ratio	Lower	Upper
101	100		
85	47.5	38.9	58.3
151	70.1	57.5	86.3



Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#10
Methyl Iodide
Concen: 48.626 ug/l
RT: 2.447 min Scan# 223
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion:142 Resp: 72261

Ion	Ratio	Lower	Upper
142	100		
127	50.6	40.2	60.4
141	14.1	11.7	17.5






#11
Tert butyl alcohol
Concen: 246.612 ug/l
RT: 2.971 min Scan# 3
Delta R.T. -0.006 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion: 59 Resp: 5877


Ion	Ratio	Lower	Upper
59	100		
57	10.9	8.6	12.8



Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#12
1,1-Dichloroethene
Concen: 49.726 ug/l
RT: 2.313 min Scan# 201
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion: 96 Resp: 57888

Ion	Ratio	Lower	Upper
96	100		
61	182.3	145.7	218.5
98	62.8	51.6	77.4






#13
Acrolein
Concen: 238.141 ug/l
RT: 2.233 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion: 56 Resp: 7815


Ion	Ratio	Lower	Upper
56	100		
55	71.0	56.2	84.2



Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#14
Allyl chloride
Concen: 50.541 ug/l
RT: 2.660 min Scan# 258
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion: 41 Resp: 111636

Ion	Ratio	Lower	Upper
41	100		
39	76.2	60.2	90.2
76	31.6	24.6	36.8






#15
Acrylonitrile
Concen: 248.351 ug/l
RT: 3.062 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion: 53 Resp: 18683


Ion	Ratio	Lower	Upper
53	100		
52	84.1	67.4	101.2
51	37.6	29.9	44.9



Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#16
Acetone
Concen: 241.904 ug/l
RT: 2.380 min Scan# 212
Delta R.T. -0.006 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion: 43 Resp: 176153

Ion	Ratio	Lower	Upper
43	100		
58	27.2	22.3	33.5






#17
Carbon Disulfide
Concen: 52.057 ug/l
RT: 2.508 min Scan# 2
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion: 76 Resp: 14967


Ion	Ratio	Lower	Upper
76	100		
78	8.7	7.0	10.6



Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#18
Methyl Acetate
Concen: 48.742 ug/l
RT: 2.703 min Scan# 265
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion: 43 Resp: 81720

Ion	Ratio	Lower	Upper
43	100		
74	21.8	17.5	26.3






#19
Methyl tert-butyl Ether
Concen: 50.950 ug/l
RT: 3.111 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion: 73 Resp: 20748


Ion	Ratio	Lower	Upper
73	100		
57	22.2	17.6	26.4



Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#20
Methylene Chloride
Concen: 49.055 ug/l
RT: 2.782 min Scan# 278
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion: 84 Resp: 66875

Ion	Ratio	Lower	Upper
84	100		
49	135.9	107.5	161.3
51	39.9	33.0	49.6
86	62.1	50.1	75.1






#21
trans-1,2-Dichloroethene
Concen: 49.800 ug/l
RT: 3.087 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion: 96 Resp: 5922


Ion	Ratio	Lower	Upper
96	100		
61	148.2	121.0	181.6
98	63.6	49.7	74.5



Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#22
Diisopropyl ether
Concen: 50.895 ug/l
RT: 3.757 min Scan# 438
Delta R.T. -0.006 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion: 45 Resp: 221449

Ion	Ratio	Lower	Upper
45	100		
43	83.3	59.3	88.9
87	26.0	20.4	30.6
59	11.4	9.0	13.6





#23
Vinyl Acetate
Concen: 267.449 ug/l
RT: 3.715 min Scan# 4
Delta R.T. -0.006 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48


Tgt Ion: 43 Resp: 100542


Ion	Ratio	Lower	Upper
43	100		
86	9.1	7.5	11.3

Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Manual Integrations
APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025







#24
1,1-Dichloroethane
Concen: 49.677 ug/l
RT: 3.605 min Scan# 413
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion: 63 Resp: 122240

Ion	Ratio	Lower	Upper
63	100		
98	7.3	3.4	10.1
100	4.2	2.2	6.6





#25
2-Butanone
Concen: 254.882 ug/l
RT: 4.550 min Scan# 5
Delta R.T. -0.006 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48


Tgt Ion: 43 Resp: 27168


Ion	Ratio	Lower	Upper
43	100		
72	22.0	18.6	28.0

Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Manual Integrations
APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025







#26
2,2-Dichloropropane
Concen: 52.020 ug/l
RT: 4.471 min Scan# 555
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion: 77 Resp: 82704

Ion	Ratio	Lower	Upper
77	100		
97	21.2	10.6	31.8






#27
cis-1,2-Dichloroethene
Concen: 49.049 ug/l
RT: 4.483 min Scan# 5
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion: 96 Resp: 7107


Ion	Ratio	Lower	Upper
96	100		
61	158.6	0.0	313.2
98	65.1	0.0	126.4



Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#28
Bromochloromethane
Concen: 48.403 ug/l
RT: 4.891 min Scan# 624
Delta R.T. -0.006 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion: 49 Resp: 57557

Ion	Ratio	Lower	Upper
49	100		
129	1.8	0.0	3.8
130	69.3	54.6	81.8





#29
Tetrahydrofuran
Concen: 249.473 ug/l
RT: 5.001 min Scan# 64
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48


Tgt Ion: 42 Resp: 17240


Ion	Ratio	Lower	Upper
42	100		
72	41.7	33.4	50.0
71	39.0	31.4	47.0

Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Manual Integrations
APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025







#30
Chloroform
Concen: 49.614 ug/l
RT: 5.086 min Scan# 656
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion: 83 Resp: 125621

Ion	Ratio	Lower	Upper
83	100		
85	64.2	52.8	79.2






#31
Cyclohexane
Concen: 49.583 ug/l
RT: 5.458 min Scan# 7
Delta R.T. -0.006 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion: 56 Resp: 10786


Ion	Ratio	Lower	Upper
56	100		
69	31.7	23.6	35.4
84	82.7	67.7	101.5



Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#32
1,1,1-Trichloroethane
Concen: 50.330 ug/l
RT: 5.379 min Scan# 704
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion: 97 Resp: 107830

Ion	Ratio	Lower	Upper
97	100		
99	64.9	51.7	77.5
61	50.5	41.0	61.4






#33
1,2-Dichloroethane-d4
Concen: 50.329 ug/l
RT: 5.952 min Scan# 7
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion: 65 Resp: 8925


Ion	Ratio	Lower	Upper
65	100		
67	49.8	0.0	99.0



Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.757 min Scan# 930
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion:114 Resp: 174855

Ion	Ratio	Lower	Upper
114	100		
63	24.5	0.0	46.8
88	18.1	0.0	35.4






#35
Dibromofluoromethane
Concen: 49.256 ug/l
RT: 5.379 min Scan# 7
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion:113 Resp: 6110


Ion	Ratio	Lower	Upper
113	100		
111	103.5	81.8	122.6
192	17.2	13.8	20.6



Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#36
1,1-Dichloropropene
Concen: 50.088 ug/l
RT: 5.684 min Scan# 754
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion: 75 Resp: 83901

Ion	Ratio	Lower	Upper
75	100		
110	32.5	16.9	50.6
77	31.1	25.0	37.4






#37
Ethyl Acetate
Concen: 49.787 ug/l
RT: 4.708 min Scan# 5
Delta R.T. -0.006 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion: 43 Resp: 10509


Ion	Ratio	Lower	Upper
43	100		
61	12.9	9.9	14.9
70	10.2	7.5	11.3



Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#38
Carbon Tetrachloride
Concen: 50.667 ug/l
RT: 5.672 min Scan# 752
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion:117 Resp: 91723

Ion	Ratio	Lower	Upper
117	100		
119	95.9	76.9	115.3
121	30.6	24.2	36.2





#39
Methylcyclohexane
Concen: 50.124 ug/l
RT: 7.373 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48


Tgt Ion: 83 Resp: 10291


Ion	Ratio	Lower	Upper
83	100		
55	82.1	64.9	97.3
98	45.7	37.4	56.0

Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Manual Integrations
APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025







#40
Benzene
Concen: 49.212 ug/l
RT: 6.031 min Scan# 811
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion: 78 Resp: 251758

Ion	Ratio	Lower	Upper
78	100		
77	24.0	19.0	28.6






#41
Methacrylonitrile
Concen: 53.098 ug/l
RT: 4.916 min Scan# 6
Delta R.T. -0.006 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion: 41 Resp: 5914


Ion	Ratio	Lower	Upper
41	100		
39	58.5	47.0	70.6
67	65.5	53.2	79.8
52	30.6	25.0	37.6



Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#42
1,2-Dichloroethane
Concen: 48.875 ug/l
RT: 6.080 min Scan# 819
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion: 62 Resp: 103270

Ion	Ratio	Lower	Upper
62	100		
98	7.6	0.0	15.2





#43
Isopropyl Acetate
Concen: 51.539 ug/l
RT: 6.336 min Scan# 8
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48


Tgt Ion: 43 Resp: 16489


Ion	Ratio	Lower	Upper
43	100		
61	17.7	14.1	21.1
87	11.6	9.2	13.8

Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Manual Integrations
APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025







#44
Trichloroethene
Concen: 47.643 ug/l
RT: 7.123 min Scan# 990
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion:130 Resp: 57931

Ion	Ratio	Lower	Upper
130	100		
95	108.6	0.0	202.0






#45
1,2-Dichloropropane
Concen: 49.521 ug/l
RT: 7.421 min Scan# 1(
Delta R.T. -0.006 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion: 63 Resp: 6321(


Ion	Ratio	Lower	Upper
63	100		
65	29.5	25.8	38.6



Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#46
Dibromomethane
Concen: 48.542 ug/l
RT: 7.574 min Scan# 1064
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion: 93 Resp: 47564

Ion	Ratio	Lower	Upper
93	100		
95	83.6	64.8	97.2
174	84.5	65.4	98.0






#47
Bromodichloromethane
Concen: 49.273 ug/l
RT: 7.818 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion: 83 Resp: 9641


Ion	Ratio	Lower	Upper
83	100		
85	65.2	49.7	74.5
127	8.1	6.1	9.1



Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#48
Methyl methacrylate
Concen: 51.507 ug/l
RT: 7.690 min Scan# 1083
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion: 41 Resp: 85084

Ion	Ratio	Lower	Upper
41	100		
69	74.5	58.2	87.2
39	67.3	53.6	80.4





#49
1,4-Dioxane
Concen: 955.327 ug/l
RT: 7.659 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48


Tgt Ion: 88 Resp: 2855


Ion	Ratio	Lower	Upper
88	100		
43	40.3	32.2	48.4
58	72.2	58.6	88.0

Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Manual Integrations
APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025







#50
Toluene-d8
Concen: 50.160 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion: 98 Resp: 217200

Ion	Ratio	Lower	Upper
98	100		
100	64.9	52.2	78.4






#51
4-Methyl-2-Pentanone
Concen: 254.994 ug/l
RT: 8.568 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion: 43 Resp: 54082


Ion	Ratio	Lower	Upper
43	100		
58	35.5	28.5	42.7



Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#52
Toluene
Concen: 48.994 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion: 92 Resp: 151664

Ion	Ratio	Lower	Upper
92	100		
91	174.0	138.7	208.1






#53
t-1,3-Dichloropropene
Concen: 46.783 ug/l
RT: 8.976 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion: 75 Resp: 8774


Ion	Ratio	Lower	Upper
75	100		
77	30.8	25.4	38.0



Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#54
cis-1,3-Dichloropropene
Concen: 53.878 ug/l
RT: 8.360 min Scan# 1193
Delta R.T. -0.006 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion: 75 Resp: 99074

Ion	Ratio	Lower	Upper
75	100		
77	31.3	24.6	36.8
39	62.0	48.4	72.6






#55
1,1,2-Trichloroethane
Concen: 49.007 ug/l
RT: 9.147 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion: 97 Resp: 6045


Ion	Ratio	Lower	Upper
97	100		
83	86.6	71.8	107.8
85	55.7	47.0	70.6
99	59.4	48.6	73.0



Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#56
Ethyl methacrylate
Concen: 53.705 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion: 69 Resp: 102525

Ion	Ratio	Lower	Upper
69	100		
41	76.7	61.3	91.9
39	50.7	40.2	60.2






#57
1,3-Dichloropropane
Concen: 49.869 ug/l
RT: 9.305 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion: 76 Resp: 10712


Ion	Ratio	Lower	Upper
76	100		
78	32.5	26.0	39.0



Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#58
2-Chloroethyl Vinyl ether
Concen: 260.201 ug/l
RT: 8.238 min Scan# 1173
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion: 63 Resp: 251285

Ion	Ratio	Lower	Upper
63	100		
106	26.6	21.0	31.6






#59
2-Hexanone
Concen: 251.710 ug/l
RT: 9.427 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion: 43 Resp: 39533


Ion	Ratio	Lower	Upper
43	100		
58	49.8	24.9	74.6



Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#60
Dibromochloromethane
Concen: 50.983 ug/l
RT: 9.519 min Scan# 1383
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion:129 Resp: 68562

Ion	Ratio	Lower	Upper
129	100		
127	78.2	39.4	118.2






#61
1,2-Dibromoethane
Concen: 50.693 ug/l
RT: 9.604 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion:107 Resp: 6330


Ion	Ratio	Lower	Upper
107	100		
109	94.2	75.7	113.5



Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#62
4-Bromofluorobenzene
Concen: 50.959 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion: 95 Resp: 80376

Ion	Ratio	Lower	Upper
95	100		
174	66.4	0.0	135.8
176	65.2	0.0	131.4






#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion:117 Resp: 15241


Ion	Ratio	Lower	Upper
117	100		
82	61.4	49.2	73.8
119	31.5	25.1	37.7



Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#64
Tetrachloroethene
Concen: 47.239 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. -0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion:164 Resp: 50834

Ion	Ratio	Lower	Upper
164	100		
166	128.0	103.8	155.8
129	104.3	84.3	126.5
131	102.6	80.9	121.3






#65
Chlorobenzene
Concen: 50.207 ug/l
RT: 10.073 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion:112 Resp: 16351


Ion	Ratio	Lower	Upper
112	100		
114	32.0	24.7	37.1



Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#66
1,1,1,2-Tetrachloroethane
Concen: 50.495 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion:131 Resp: 57096

Ion	Ratio	Lower	Upper
131	100		
133	93.6	47.5	142.5
119	65.0	32.9	98.6






#67
Ethyl Benzene
Concen: 51.558 ug/l
RT: 10.189 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion: 91 Resp: 30074


Ion	Ratio	Lower	Upper
91	100		
106	29.6	23.7	35.5



Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#68
m/p-Xylenes
Concen: 102.728 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion:106 Resp: 217943

Ion	Ratio	Lower	Upper
106	100		
91	217.6	172.6	258.8






#69
o-Xylene
Concen: 51.650 ug/l
RT: 10.640 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion:106 Resp: 10794


Ion	Ratio	Lower	Upper
106	100		
91	226.8	113.6	340.6



Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#70
Styrene
Concen: 52.624 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion:104 Resp: 181524

Ion	Ratio	Lower	Upper
104	100		
78	57.0	45.4	68.2
103	55.2	43.8	65.8






#71
Bromoform
Concen: 52.694 ug/l
RT: 10.799 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion:173 Resp: 4448


Ion	Ratio	Lower	Upper
173	100		
175	47.8	23.4	70.2
254	0.0	0.0	0.0



Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion:152 Resp: 67633

Ion	Ratio	Lower	Upper
152	100		
115	94.2	46.9	140.7
150	175.4	0.0	349.4






#73
Isopropylbenzene
Concen: 52.794 ug/l
RT: 10.957 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion:105 Resp: 28601


Ion	Ratio	Lower	Upper
105	100		
120	25.5	12.7	38.0



Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#74
N-amyl acetate
Concen: 54.387 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion: 43 Resp: 140941

Ion	Ratio	Lower	Upper
43	100		
70	38.2	30.1	45.1
55	23.3	18.3	27.5
61	21.7	17.5	26.3





#75
1,1,2,2-Tetrachloroethane
Concen: 48.608 ug/l
RT: 11.207 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48


Tgt Ion: 83 Resp: 9250

Ion	Ratio	Lower	Upper
83	100		
131	9.5	4.7	14.1
85	63.9	32.4	97.2

Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025



#76
1,2,3-Trichloropropane
Concen: 49.741 ug/l m
RT: 11.238 min Scan# 1665
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion: 75 Resp: 82060


Ion	Ratio	Lower	Upper
75	100		
77	40.0	20.0	59.9



#77
Bromobenzene
Concen: 50.525 ug/l
RT: 11.195 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion:156 Resp: 6324


Ion	Ratio	Lower	Upper
156	100		
77	188.0	92.6	277.8
158	97.5	48.1	144.4



Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#78
n-propylbenzene
Concen: 53.967 ug/l
RT: 11.299 min Scan# 1675
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion: 91 Resp: 336753

Ion	Ratio	Lower	Upper
91	100		
120	21.4	10.8	32.3






#79
2-Chlorotoluene
Concen: 51.057 ug/l
RT: 11.360 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion: 91 Resp: 20367


Ion	Ratio	Lower	Upper
91	100		
126	31.2	15.8	47.4



Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#80
1,3,5-Trimethylbenzene
Concen: 53.534 ug/l
RT: 11.445 min Scan# 1699
Delta R.T. -0.006 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion:105 Resp: 239822

Ion	Ratio	Lower	Upper
105	100		
120	45.8	23.0	69.0






#81
trans-1,4-Dichloro-2-butene
Concen: 52.482 ug/l
RT: 11.018 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion: 75 Resp: 2551


Ion	Ratio	Lower	Upper
75	100		
53	124.4	101.8	152.6
89	46.4	37.2	55.8



Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#82
4-Chlorotoluene
Concen: 52.422 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion: 91 Resp: 233159

Ion	Ratio	Lower	Upper
91	100		
126	26.8	13.3	39.9






#83
tert-Butylbenzene
Concen: 53.274 ug/l
RT: 11.713 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion:119 Resp: 23627


Ion	Ratio	Lower	Upper
119	100		
91	64.8	32.4	97.2
134	22.6	11.5	34.5



Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#84
1,2,4-Trimethylbenzene
Concen: 52.990 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion:105 Resp: 238257

Ion	Ratio	Lower	Upper
105	100		
120	42.8	21.4	64.3






#85
sec-Butylbenzene
Concen: 53.903 ug/l
RT: 11.890 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion:105 Resp: 29363


Ion	Ratio	Lower	Upper
105	100		
134	18.9	9.6	28.6



Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#86
p-Isopropyltoluene
Concen: 54.337 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion:119 Resp: 244031

Ion	Ratio	Lower	Upper
119	100		
134	24.8	12.6	37.8
91	27.0	13.7	41.1






#87
1,3-Dichlorobenzene
Concen: 50.521 ug/l
RT: 11.963 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion:146 Resp: 11511


Ion	Ratio	Lower	Upper
146	100		
111	44.3	22.2	66.6
148	63.9	31.9	95.7



Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#88
1,4-Dichlorobenzene
Concen: 49.117 ug/l
RT: 12.036 min Scan# 1796
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion:146 Resp: 113458

Ion	Ratio	Lower	Upper
146	100		
111	44.1	20.9	62.8
148	64.5	31.5	94.5





#89
n-Butylbenzene
Concen: 55.189 ug/l
RT: 12.329 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48


Tgt Ion: 91 Resp: 21493

Ion	Ratio	Lower	Upper
91	100		
92	53.1	26.9	80.6
134	23.0	11.9	35.7

Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025



#90
Hexachloroethane
Concen: 54.066 ug/l
RT: 12.536 min Scan# 1878
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion:117 Resp: 41729


Ion	Ratio	Lower	Upper
117	100		
201	61.0	31.2	93.6



#91
1,2-Dichlorobenzene
Concen: 49.938 ug/l
RT: 12.329 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion:146 Resp: 11312


Ion	Ratio	Lower	Upper
146	100		
111	45.9	22.9	68.8
148	64.1	31.9	95.5



Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#92
1,2-Dibromo-3-Chloropropane
Concen: 53.226 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion: 75 Resp: 21184

Ion	Ratio	Lower	Upper
75	100		
155	69.4	34.9	104.7
157	90.2	44.9	134.7






#93
1,2,4-Trichlorobenzene
Concen: 51.935 ug/l
RT: 13.585 min Scan# 2
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion:180 Resp: 6548


Ion	Ratio	Lower	Upper
180	100		
182	94.5	47.9	143.6
145	34.1	17.9	53.8



Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#94
Hexachlorobutadiene
Concen: 49.961 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. 0.006 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion:225 Resp: 27147

Ion	Ratio	Lower	Upper
225	100		
223	62.9	31.1	93.3
227	63.0	32.2	96.6






#95
Naphthalene
Concen: 52.765 ug/l
RT: 13.774 min Scan# 2
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion:128 Resp: 24760


Ion	Ratio	Lower	Upper
128	100		
127	12.9	10.3	15.5
129	10.7	8.6	12.8



Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Manual Integrations APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025



#96
1,2,3-Trichlorobenzene
Concen: 50.396 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. 0.000 min
Lab File: VX045532.D
Acq: 01 Apr 2025 19:48

Tgt Ion:180 Resp: 66542

Ion	Ratio	Lower	Upper
180	100		
182	94.8	47.3	141.8
145	36.4	18.2	54.6



Evaluate Continuing Calibration Report

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045532.D
Acq On : 01 Apr 2025 19:48
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Quant Time: Apr 02 03:14:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1	I	Pentafluorobenzene	1.000	1.000	0.0	101	0.00
2	T	Dichlorodifluoromethane	0.748	0.784	-4.8	97	0.00
3	P	Chloromethane	0.768	0.743	3.3	96	0.00
4	C	Vinyl Chloride	0.703	0.686	2.4#	97	0.00
5	T	Bromomethane	0.333	0.313	6.0	96	0.00
6	T	Chloroethane	0.373	0.385	-3.2	98	0.00
7	T	Trichlorofluoromethane	1.048	1.056	-0.8	99	0.00
8	T	Diethyl Ether	0.352	0.347	1.4	99	0.00
9	T	1,1,2-Trichlorotrifluoroeth	0.614	0.608	1.0	99	0.00
10	T	Methyl Iodide	0.766	0.745	2.7	95	0.00
11	T	Tert butyl alcohol	0.123	0.121	1.6	97	0.00
12	CM	1,1-Dichloroethene	0.600	0.597	0.5#	100	0.00
13	T	Acrolein	0.169	0.161	4.7	98	0.00
14	T	Allyl chloride	1.139	1.151	-1.1	100	0.00
15	T	Acrylonitrile	0.388	0.385	0.8	96	0.00
16	T	Acetone	0.375	0.363	3.2	98	0.00
17	T	Carbon Disulfide	1.483	1.544	-4.1	101	0.00
18	T	Methyl Acetate	0.864	0.843	2.4	100	0.00
19	T	Methyl tert-butyl Ether	2.100	2.140	-1.9	102	0.00
20	T	Methylene Chloride	0.703	0.690	1.8	99	0.00
21	T	trans-1,2-Dichloroethene	0.613	0.611	0.3	99	0.00
22	T	Diisopropyl ether	2.243	2.284	-1.8	101	0.00
23	T	Vinyl Acetate	1.938	2.074	-7.0	101	0.00
24	P	1,1-Dichloroethane	1.269	1.261	0.6	101	0.00
25	T	2-Butanone	0.550	0.560	-1.8	97	0.00
26	T	2,2-Dichloropropane	0.820	0.853	-4.0	102	0.00
27	T	cis-1,2-Dichloroethene	0.747	0.733	1.9	99	0.00
28	T	Bromochloromethane	0.613	0.594	3.1	101	0.00
29	T	Tetrahydrofuran	0.356	0.356	0.0	97	0.00
30	C	Chloroform	1.306	1.295	0.8#	100	0.00
31	T	Cyclohexane	1.122	1.112	0.9	100	0.00
32	T	1,1,1-Trichloroethane	1.105	1.112	-0.6	102	0.00
33	S	1,2-Dichloroethane-d4	0.914	0.920	-0.7	107	0.00
34	I	1,4-Difluorobenzene	1.000	1.000	0.0	104	0.00
35	S	Dibromofluoromethane	0.355	0.349	1.7	105	0.00
36	T	1,1-Dichloropropene	0.479	0.480	-0.2	103	0.00
37	T	Ethyl Acetate	0.604	0.601	0.5	100	0.00
38	T	Carbon Tetrachloride	0.518	0.525	-1.4	102	0.00
39	T	Methylcyclohexane	0.587	0.589	-0.3	98	0.00
40	TM	Benzene	1.463	1.440	1.6	101	0.00
41	T	Methacrylonitrile	0.318	0.338	-6.3	100	0.00
42	TM	1,2-Dichloroethane	0.604	0.591	2.2	99	0.00
43	T	Isopropyl Acetate	0.915	0.943	-3.1	100	0.00
44	TM	Trichloroethene	0.348	0.331	4.9	98	0.00
45	C	1,2-Dichloropropane	0.365	0.361	1.1#	100	0.00
46	T	Dibromomethane	0.280	0.272	2.9	97	0.00
47	T	Bromodichloromethane	0.560	0.551	1.6	100	0.00
48	T	Methyl methacrylate	0.472	0.487	-3.2	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045532.D
Acq On : 01 Apr 2025 19:48
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Quant Time: Apr 02 03:14:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49	T	1,4-Dioxane	0.009	0.008	11.1	92	0.00
50	S	Toluene-d8	1.238	1.242	-0.3	106	0.00
51	T	4-Methyl-2-Pentanone	0.606	0.619	-2.1	97	0.00
52	CM	Toluene	0.885	0.867	2.0#	99	0.00
53	T	t-1,3-Dichloropropene	0.467	0.502	-7.5	103	0.00
54	T	cis-1,3-Dichloropropene	0.526	0.567	-7.8	104	0.00
55	T	1,1,2-Trichloroethane	0.353	0.346	2.0	100	0.00
56	T	Ethyl methacrylate	0.546	0.586	-7.3	102	0.00
57	T	1,3-Dichloropropane	0.614	0.613	0.2	101	0.00
58	T	2-Chloroethyl Vinyl ether	0.276	0.287	-4.0	100	0.00
59	T	2-Hexanone	0.449	0.452	-0.7	96	0.00
60	T	Dibromochloromethane	0.385	0.392	-1.8	99	0.00
61	T	1,2-Dibromoethane	0.357	0.362	-1.4	101	0.00
62	S	4-Bromofluorobenzene	0.451	0.460	-2.0	105	0.00
63	I	Chlorobenzene-d5	1.000	1.000	0.0	101	0.00
64	T	Tetrachloroethene	0.353	0.334	5.4	97	0.00
65	PM	Chlorobenzene	1.068	1.073	-0.5	100	0.00
66	T	1,1,1,2-Tetrachloroethane	0.371	0.375	-1.1	102	0.00
67	C	Ethyl Benzene	1.914	1.973	-3.1#	100	0.00
68	T	m/p-Xylenes	0.696	0.715	-2.7	100	0.00
69	T	o-Xylene	0.686	0.708	-3.2	100	0.00
70	T	Styrene	1.132	1.191	-5.2	99	0.00
71	P	Bromoform	0.277	0.292	-5.4	100	0.00
72	I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
73	T	Isopropylbenzene	4.005	4.229	-5.6	99	0.00
74	T	N-amyl acetate	1.916	2.084	-8.8	100	0.00
75	P	1,1,2,2-Tetrachloroethane	1.407	1.368	2.8	97	0.00
76	T	1,2,3-Trichloropropane	1.220	1.213	0.6	97	0.00
77	T	Bromobenzene	0.925	0.935	-1.1	98	0.00
78	T	n-propylbenzene	4.613	4.979	-7.9	100	0.00
79	T	2-Chlorotoluene	2.949	3.012	-2.1	100	0.00
80	T	1,3,5-Trimethylbenzene	3.312	3.546	-7.1	99	0.00
81	T	trans-1,4-Dichloro-2-butene	0.359	0.377	-5.0	100	0.00
82	T	4-Chlorotoluene	3.288	3.447	-4.8	98	0.00
83	T	tert-Butylbenzene	3.279	3.494	-6.6	100	0.00
84	T	1,2,4-Trimethylbenzene	3.324	3.523	-6.0	98	0.00
85	T	sec-Butylbenzene	4.027	4.342	-7.8	99	0.00
86	T	p-Isopropyltoluene	3.320	3.608	-8.7	101	0.00
87	T	1,3-Dichlorobenzene	1.684	1.702	-1.1	98	0.00
88	T	1,4-Dichlorobenzene	1.708	1.678	1.8	97	0.00
89	T	n-Butylbenzene	2.879	3.178	-10.4	99	0.00
90	T	Hexachloroethane	0.571	0.617	-8.1	102	0.00
91	T	1,2-Dichlorobenzene	1.675	1.673	0.1	98	0.00
92	T	1,2-Dibromo-3-Chloropropane	0.294	0.313	-6.5	98	0.00
93	T	1,2,4-Trichlorobenzene	0.932	0.968	-3.9	101	0.00
94	T	Hexachlorobutadiene	0.402	0.401	0.2	98	0.00
95	T	Naphthalene	3.469	3.661	-5.5	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045532.D
Acq On : 01 Apr 2025 19:48
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Quant Time: Apr 02 03:14:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.976	0.984	-0.8	97	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6

Evaluate Continuing Calibration Report

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045532.D
Acq On : 01 Apr 2025 19:48
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Quant Time: Apr 02 03:14:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	Pentafluorobenzene	50.000	50.000	0.0	101	0.00
2	T	Dichlorodifluoromethane	50.000	52.454	-4.9	97	0.00
3	P	Chloromethane	50.000	48.349	3.3	96	0.00
4	C	Vinyl Chloride	50.000	48.816	2.4#	97	0.00
5	T	Bromomethane	50.000	46.991	6.0	96	0.00
6	T	Chloroethane	50.000	51.586	-3.2	98	0.00
7	T	Trichlorofluoromethane	50.000	50.354	-0.7	99	0.00
8	T	Diethyl Ether	50.000	49.228	1.5	99	0.00
9	T	1,1,2-Trichlorotrifluoroeth	50.000	49.528	0.9	99	0.00
10	T	Methyl Iodide	50.000	48.626	2.7	95	0.00
11	T	Tert butyl alcohol	250.000	246.612	1.4	97	0.00
12	CM	1,1-Dichloroethene	50.000	49.726	0.5#	100	0.00
13	T	Acrolein	250.000	238.141	4.7	98	0.00
14	T	Allyl chloride	50.000	50.541	-1.1	100	0.00
15	T	Acrylonitrile	250.000	248.351	0.7	96	0.00
16	T	Acetone	250.000	241.904	3.2	98	0.00
17	T	Carbon Disulfide	50.000	52.057	-4.1	101	0.00
18	T	Methyl Acetate	50.000	48.742	2.5	100	0.00
19	T	Methyl tert-butyl Ether	50.000	50.950	-1.9	102	0.00
20	T	Methylene Chloride	50.000	49.055	1.9	99	0.00
21	T	trans-1,2-Dichloroethene	50.000	49.800	0.4	99	0.00
22	T	Diisopropyl ether	50.000	50.895	-1.8	101	0.00
23	T	Vinyl Acetate	250.000	267.449	-7.0	101	0.00
24	P	1,1-Dichloroethane	50.000	49.677	0.6	101	0.00
25	T	2-Butanone	250.000	254.882	-2.0	97	0.00
26	T	2,2-Dichloropropane	50.000	52.020	-4.0	102	0.00
27	T	cis-1,2-Dichloroethene	50.000	49.049	1.9	99	0.00
28	T	Bromochloromethane	50.000	48.403	3.2	101	0.00
29	T	Tetrahydrofuran	250.000	249.473	0.2	97	0.00
30	C	Chloroform	50.000	49.614	0.8#	100	0.00
31	T	Cyclohexane	50.000	49.583	0.8	100	0.00
32	T	1,1,1-Trichloroethane	50.000	50.330	-0.7	102	0.00
33	S	1,2-Dichloroethane-d4	50.000	50.329	-0.7	107	0.00
34	I	1,4-Difluorobenzene	50.000	50.000	0.0	104	0.00
35	S	Dibromofluoromethane	50.000	49.256	1.5	105	0.00
36	T	1,1-Dichloropropene	50.000	50.088	-0.2	103	0.00
37	T	Ethyl Acetate	50.000	49.787	0.4	100	0.00
38	T	Carbon Tetrachloride	50.000	50.667	-1.3	102	0.00
39	T	Methylcyclohexane	50.000	50.124	-0.2	98	0.00
40	TM	Benzene	50.000	49.212	1.6	101	0.00
41	T	Methacrylonitrile	50.000	53.098	-6.2	100	0.00
42	TM	1,2-Dichloroethane	50.000	48.875	2.3	99	0.00
43	T	Isopropyl Acetate	50.000	51.539	-3.1	100	0.00
44	TM	Trichloroethene	50.000	47.643	4.7	98	0.00
45	C	1,2-Dichloropropane	50.000	49.521	1.0#	100	0.00
46	T	Dibromomethane	50.000	48.542	2.9	97	0.00
47	T	Bromodichloromethane	50.000	49.273	1.5	100	0.00
48	T	Methyl methacrylate	50.000	51.507	-3.0	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045532.D
Acq On : 01 Apr 2025 19:48
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Quant Time: Apr 02 03:14:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49	T	1,4-Dioxane	1000.000	955.327	4.5	92	0.00
50	S	Toluene-d8	50.000	50.160	-0.3	106	0.00
51	T	4-Methyl-2-Pentanone	250.000	254.994	-2.0	97	0.00
52	CM	Toluene	50.000	48.994	2.0#	99	0.00
53	T	t-1,3-Dichloropropene	50.000	46.783	6.4	103	0.00
54	T	cis-1,3-Dichloropropene	50.000	53.878	-7.8	104	0.00
55	T	1,1,2-Trichloroethane	50.000	49.007	2.0	100	0.00
56	T	Ethyl methacrylate	50.000	53.705	-7.4	102	0.00
57	T	1,3-Dichloropropane	50.000	49.869	0.3	101	0.00
58	T	2-Chloroethyl Vinyl ether	250.000	260.201	-4.1	100	0.00
59	T	2-Hexanone	250.000	251.710	-0.7	96	0.00
60	T	Dibromochloromethane	50.000	50.983	-2.0	99	0.00
61	T	1,2-Dibromoethane	50.000	50.693	-1.4	101	0.00
62	S	4-Bromofluorobenzene	50.000	50.959	-1.9	105	0.00
63	I	Chlorobenzene-d5	50.000	50.000	0.0	101	0.00
64	T	Tetrachloroethene	50.000	47.239	5.5	97	0.00
65	PM	Chlorobenzene	50.000	50.207	-0.4	100	0.00
66	T	1,1,1,2-Tetrachloroethane	50.000	50.495	-1.0	102	0.00
67	C	Ethyl Benzene	50.000	51.558	-3.1#	100	0.00
68	T	m/p-Xylenes	100.000	102.728	-2.7	100	0.00
69	T	o-Xylene	50.000	51.650	-3.3	100	0.00
70	T	Styrene	50.000	52.624	-5.2	99	0.00
71	P	Bromoform	50.000	52.694	-5.4	100	0.00
72	I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	98	0.00
73	T	Isopropylbenzene	50.000	52.794	-5.6	99	0.00
74	T	N-amyl acetate	50.000	54.387	-8.8	100	0.00
75	P	1,1,2,2-Tetrachloroethane	50.000	48.608	2.8	97	0.00
76	T	1,2,3-Trichloropropane	50.000	49.741	0.5	97	0.00
77	T	Bromobenzene	50.000	50.525	-1.0	98	0.00
78	T	n-propylbenzene	50.000	53.967	-7.9	100	0.00
79	T	2-Chlorotoluene	50.000	51.057	-2.1	100	0.00
80	T	1,3,5-Trimethylbenzene	50.000	53.534	-7.1	99	0.00
81	T	trans-1,4-Dichloro-2-butene	50.000	52.482	-5.0	100	0.00
82	T	4-Chlorotoluene	50.000	52.422	-4.8	98	0.00
83	T	tert-Butylbenzene	50.000	53.274	-6.5	100	0.00
84	T	1,2,4-Trimethylbenzene	50.000	52.990	-6.0	98	0.00
85	T	sec-Butylbenzene	50.000	53.903	-7.8	99	0.00
86	T	p-Isopropyltoluene	50.000	54.337	-8.7	101	0.00
87	T	1,3-Dichlorobenzene	50.000	50.521	-1.0	98	0.00
88	T	1,4-Dichlorobenzene	50.000	49.117	1.8	97	0.00
89	T	n-Butylbenzene	50.000	55.189	-10.4	99	0.00
90	T	Hexachloroethane	50.000	54.066	-8.1	102	0.00
91	T	1,2-Dichlorobenzene	50.000	49.938	0.1	98	0.00
92	T	1,2-Dibromo-3-Chloropropane	50.000	53.226	-6.5	98	0.00
93	T	1,2,4-Trichlorobenzene	50.000	51.935	-3.9	101	0.00
94	T	Hexachlorobutadiene	50.000	49.961	0.1	98	0.00
95	T	Naphthalene	50.000	52.765	-5.5	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045532.D
Acq On : 01 Apr 2025 19:48
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Quant Time: Apr 02 03:14:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	50.396	-0.8	97	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: JACO05

Lab Code: CHEM Case No.: Q1731 SAS No.: Q1731 SDG No.: Q1731

Instrument ID: MSVOA_X Calibration Date/Time: 04/04/2025 10:00

Lab File ID: VX045586.D Init. Calib. Date(s): 04/01/2025 04/01/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 17:06 19:02

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.703	0.731		3.98	20
1,1-Dichloroethene	0.600	0.599		-0.17	20
1,1-Dichloroethane	1.269	1.270	0.1	0.08	20
cis-1,2-Dichloroethene	0.747	0.729		-2.41	20
1,1,1-Trichloroethane	1.105	1.107		0.18	20
Benzene	1.463	1.507		3.01	20
1,2-Dichloroethane	0.604	0.645		6.79	20
Trichloroethene	0.348	0.357		2.59	20
1,1,2-Trichloroethane	0.353	0.370		4.82	20
Tetrachloroethene	0.353	0.376		6.52	20
1,2-Dichloroethane-d4	0.914	0.776		-15.1	20
Dibromofluoromethane	0.355	0.311		-12.39	20
Toluene-d8	1.238	1.079		-12.84	20
4-Bromofluorobenzene	0.451	0.411		-8.87	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040425\
Data File : VX045586.D
Acq On : 04 Apr 2025 10:00
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Apr 04 23:10:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations APPROVED

Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	5.538	168	96707	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	6.751	114	163287	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	142048	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	67379	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	75034	42.428	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 84.860%			
35) Dibromofluoromethane	5.379	113	50835	43.879	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 87.760%			
50) Toluene-d8	8.641	98	176247	43.586	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 87.180%			
62) 4-Bromofluorobenzene	11.079	95	67111	45.563	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 91.120%			
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.166	85	81543	56.382	ug/l	98
3) Chloromethane	1.307	50	77087	51.883	ug/l	98
4) Vinyl Chloride	1.374	62	70654	51.962	ug/l	99
5) Bromomethane	1.593	94	31732	49.218	ug/l	99
6) Chloroethane	1.666	64	37965	52.626	ug/l	96
7) Trichlorofluoromethane	1.874	101	106831	52.688	ug/l	98
8) Diethyl Ether	2.130	74	33835	49.704	ug/l	96
9) 1,1,2-Trichlorotrifluo...	2.319	101	63673	53.591	ug/l	97
10) Methyl Iodide	2.441	142	74481	50.258	ug/l	99
11) Tert butyl alcohol	2.977	59	56849	239.188	ug/l	99
12) 1,1-Dichloroethene	2.306	96	57913	49.883	ug/l	97
13) Acrolein	2.233	56	96410	294.566	ug/l	100
14) Allyl chloride	2.654	41	113707	51.619	ug/l	98
15) Acrylonitrile	3.062	53	186427	248.485	ug/l	99
16) Acetone	2.380	43	178014	245.130	ug/l	97
17) Carbon Disulfide	2.502	76	145106	50.605	ug/l	100
18) Methyl Acetate	2.703	43	82725	49.477	ug/l	99
19) Methyl tert-butyl Ether	3.111	73	203434	50.091	ug/l	97
20) Methylene Chloride	2.782	84	67127	49.375	ug/l	98
21) trans-1,2-Dichloroethene	3.081	96	58900	49.659	ug/l	96
22) Diisopropyl ether	3.757	45	220705	50.863	ug/l	97
23) Vinyl Acetate	3.715	43	975461	260.190	ug/l	100
24) 1,1-Dichloroethane	3.599	63	122843	50.059	ug/l	99
25) 2-Butanone	4.550	43	265772	250.017	ug/l	98
26) 2,2-Dichloropropane	4.465	77	91259	57.558	ug/l	99
27) cis-1,2-Dichloroethene	4.477	96	70467	48.765	ug/l	98
28) Bromochloromethane	4.885	49	58888	49.658	ug/l	99
29) Tetrahydrofuran	5.001	42	168431	244.391	ug/l	100
30) Chloroform	5.080	83	127207	50.378	ug/l	96
31) Cyclohexane	5.458	56	109881	50.650	ug/l	97
32) 1,1,1-Trichloroethane	5.373	97	107072	50.113	ug/l	99
36) 1,1-Dichloropropene	5.684	75	81866	52.336	ug/l	100
37) Ethyl Acetate	4.708	43	101136	51.305	ug/l	99
38) Carbon Tetrachloride	5.666	117	93252	55.161	ug/l	99
39) Methylcyclohexane	7.373	83	107689	56.163	ug/l	100
40) Benzene	6.025	78	246115	51.517	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040425\
Data File : VX045586.D
Acq On : 04 Apr 2025 10:00
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Apr 04 23:10:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations APPROVED

Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
41) Methacrylonitrile	4.910	41	58687	56.425	ug/l		97
42) 1,2-Dichloroethane	6.074	62	105249	53.340	ug/l		100
43) Isopropyl Acetate	6.330	43	157563	52.737	ug/l		99
44) Trichloroethene	7.117	130	58260	51.308	ug/l		94
45) 1,2-Dichloropropane	7.421	63	62596	52.514	ug/l		98
46) Dibromomethane	7.574	93	47591	52.011	ug/l		98
47) Bromodichloromethane	7.818	83	97528	53.371	ug/l		100
48) Methyl methacrylate	7.690	41	83087	53.861	ug/l		99
49) 1,4-Dioxane	7.659	88	27137	972.308	ug/l		96
51) 4-Methyl-2-Pentanone	8.568	43	531642	268.424	ug/l		99
52) Toluene	8.714	92	151389	52.369	ug/l		99
53) t-1,3-Dichloropropene	8.976	75	86126	49.056	ug/l		100
54) cis-1,3-Dichloropropene	8.360	75	96025	55.919	ug/l		98
55) 1,1,2-Trichloroethane	9.147	97	60484	52.506	ug/l		96
56) Ethyl methacrylate	9.116	69	95283	53.447	ug/l		99
57) 1,3-Dichloropropane	9.305	76	105500	52.594	ug/l		99
58) 2-Chloroethyl Vinyl ether	8.238	63	231389	256.573	ug/l		100
59) 2-Hexanone	9.427	43	391210	266.730	ug/l		99
60) Dibromochloromethane	9.519	129	69085	55.012	ug/l		97
61) 1,2-Dibromoethane	9.604	107	62072	53.230	ug/l		99
64) Tetrachloroethene	9.269	164	53413	53.258	ug/l		97
65) Chlorobenzene	10.073	112	160997	53.041	ug/l		97
66) 1,1,1,2-Tetrachloroethane	10.159	131	55317	52.492	ug/l		99
67) Ethyl Benzene	10.189	91	297368	54.700	ug/l		100
68) m/p-Xylenes	10.299	106	216775	109.636	ug/l		98
69) o-Xylene	10.640	106	105932	54.386	ug/l		99
70) Styrene	10.653	104	179659	55.886	ug/l		99
71) Bromoform	10.799	173	43659	55.486	ug/l	#	98
73) Isopropylbenzene	10.957	105	288495	53.453	ug/l		100
74) N-amyl acetate	10.842	43	134287	52.015	ug/l		100
75) 1,1,2,2-Tetrachloroethane	11.207	83	91045	48.021	ug/l		99
76) 1,2,3-Trichloropropane	11.238	75	81348m	49.496	ug/l		
77) Bromobenzene	11.195	156	63908	51.248	ug/l		100
78) n-propylbenzene	11.299	91	337878	54.351	ug/l		100
79) 2-Chlorotoluene	11.360	91	203932	51.314	ug/l		99
80) 1,3,5-Trimethylbenzene	11.451	105	239716	53.712	ug/l		100
81) trans-1,4-Dichloro-2-b...	11.018	75	25247	52.135	ug/l		99
82) 4-Chlorotoluene	11.451	91	235414	53.128	ug/l		100
83) tert-Butylbenzene	11.713	119	234285	53.024	ug/l		99
84) 1,2,4-Trimethylbenzene	11.750	105	241455	53.903	ug/l		99
85) sec-Butylbenzene	11.890	105	298315	54.969	ug/l		100
86) p-Isopropyltoluene	12.006	119	246346	55.060	ug/l		99
87) 1,3-Dichlorobenzene	11.963	146	116982	51.535	ug/l		100
88) 1,4-Dichlorobenzene	12.036	146	116831	50.768	ug/l		99
89) n-Butylbenzene	12.329	91	223310	57.557	ug/l		99
90) Hexachloroethane	12.536	117	42554	55.342	ug/l		99
91) 1,2-Dichlorobenzene	12.335	146	115345	51.109	ug/l		100
92) 1,2-Dibromo-3-Chloropr...	12.939	75	21298	53.714	ug/l		99
93) 1,2,4-Trichlorobenzene	13.585	180	68119	54.225	ug/l		99
94) Hexachlorobutadiene	13.725	225	29467	54.435	ug/l		97
95) Naphthalene	13.774	128	241025	51.558	ug/l		99
96) 1,2,3-Trichlorobenzene	13.957	180	68237	51.875	ug/l		99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040425\
Data File : VX045586.D
Acq On : 04 Apr 2025 10:00
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Apr 04 23:10:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations APPROVED

Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040425\
Data File : VX045586.D
Acq On : 04 Apr 2025 10:00
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1


Quant Time: Apr 04 23:10:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations APPROVED

Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025






#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.538 min Scan# 7
Delta R.T. -0.012 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion:168 Resp: 9670


Ion	Ratio	Lower	Upper
168	100		
99	70.0	52.3	78.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#2
Dichlorodifluoromethane
Concen: 56.382 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion: 85 Resp: 81543

Ion	Ratio	Lower	Upper
85	100		
87	33.0	16.1	48.2






#3
Chloromethane
Concen: 51.883 ug/l
RT: 1.307 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion: 50 Resp: 7708


Ion	Ratio	Lower	Upper
50	100		
52	33.6	25.9	38.9



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#4
Vinyl Chloride
Concen: 51.962 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion: 62 Resp: 70654

Ion	Ratio	Lower	Upper
62	100		
64	32.2	25.5	38.3






#5
Bromomethane
Concen: 49.218 ug/l
RT: 1.593 min Scan# 8
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion: 94 Resp: 3173


Ion	Ratio	Lower	Upper
94	100		
96	92.6	74.8	112.2



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#6
Chloroethane
Concen: 52.626 ug/l
RT: 1.666 min Scan# 95
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion: 64 Resp: 37965

Ion	Ratio	Lower	Upper
64	100		
66	30.4	26.3	39.5





#7
Trichlorofluoromethane
Concen: 52.688 ug/l
RT: 1.874 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00


Tgt Ion:101 Resp: 10683


Ion	Ratio	Lower	Upper
101	100		
103	65.4	51.1	76.7

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025







#8
Diethyl Ether
Concen: 49.704 ug/l
RT: 2.130 min Scan# 171
Delta R.T. -0.006 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion: 74 Resp: 33835

Ion	Ratio	Lower	Upper
74	100		
45	109.8	53.0	159.1





#9
1,1,2-Trichlorotrifluoroethane
Concen: 53.591 ug/l
RT: 2.319 min Scan# 2
Delta R.T. -0.006 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00


Tgt Ion:101 Resp: 6367


Ion	Ratio	Lower	Upper
101	100		
85	48.0	38.9	58.3
151	68.8	57.5	86.3

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025







#10
Methyl Iodide
Concen: 50.258 ug/l
RT: 2.441 min Scan# 222
Delta R.T. -0.006 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion:142 Resp: 74481

Ion	Ratio	Lower	Upper
142	100		
127	50.8	40.2	60.4
141	14.7	11.7	17.5





#11
Tert butyl alcohol
Concen: 239.188 ug/l
RT: 2.977 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00


Tgt Ion: 59 Resp: 5684


Ion	Ratio	Lower	Upper
59	100		
57	11.2	8.6	12.8

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025







#12
1,1-Dichloroethene
Concen: 49.883 ug/l
RT: 2.306 min Scan# 200
Delta R.T. -0.006 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion: 96 Resp: 57913

Ion	Ratio	Lower	Upper
96	100		
61	187.1	145.7	218.5
98	62.1	51.6	77.4






#13
Acrolein
Concen: 294.566 ug/l
RT: 2.233 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion: 56 Resp: 9641


Ion	Ratio	Lower	Upper
56	100		
55	70.0	56.2	84.2



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#14
Allyl chloride
Concen: 51.619 ug/l
RT: 2.654 min Scan# 257
Delta R.T. -0.006 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion: 41 Resp: 113707

Ion	Ratio	Lower	Upper
41	100		
39	77.6	60.2	90.2
76	31.1	24.6	36.8






#15
Acrylonitrile
Concen: 248.485 ug/l
RT: 3.062 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion: 53 Resp: 18642


Ion	Ratio	Lower	Upper
53	100		
52	83.1	67.4	101.2
51	37.6	29.9	44.9



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#16
Acetone
Concen: 245.130 ug/l
RT: 2.380 min Scan# 212
Delta R.T. -0.006 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion: 43 Resp: 178014

Ion	Ratio	Lower	Upper
43	100		
58	26.3	22.3	33.5






#17
Carbon Disulfide
Concen: 50.605 ug/l
RT: 2.502 min Scan# 2
Delta R.T. -0.006 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion: 76 Resp: 14510


Ion	Ratio	Lower	Upper
76	100		
78	9.0	7.0	10.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#18
Methyl Acetate
Concen: 49.477 ug/l
RT: 2.703 min Scan# 265
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion: 43 Resp: 82725

Ion	Ratio	Lower	Upper
43	100		
74	22.3	17.5	26.3






#19
Methyl tert-butyl Ether
Concen: 50.091 ug/l
RT: 3.111 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion: 73 Resp: 20343


Ion	Ratio	Lower	Upper
73	100		
57	23.2	17.6	26.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#20
Methylene Chloride
Concen: 49.375 ug/l
RT: 2.782 min Scan# 278
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion: 84 Resp: 67127

Ion	Ratio	Lower	Upper
84	100		
49	138.8	107.5	161.3
51	41.6	33.0	49.6
86	63.0	50.1	75.1






#21
trans-1,2-Dichloroethene
Concen: 49.659 ug/l
RT: 3.081 min Scan# 3
Delta R.T. -0.006 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion: 96 Resp: 5890


Ion	Ratio	Lower	Upper
96	100		
61	157.7	121.0	181.6
98	63.9	49.7	74.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#22
Diisopropyl ether
Concen: 50.863 ug/l
RT: 3.757 min Scan# 438
Delta R.T. -0.006 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion: 45 Resp: 220705

Ion	Ratio	Lower	Upper
45	100		
43	76.4	59.3	88.9
87	24.0	20.4	30.6
59	11.1	9.0	13.6





#23
Vinyl Acetate
Concen: 260.190 ug/l
RT: 3.715 min Scan# 4
Delta R.T. -0.006 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00


Tgt Ion: 43 Resp: 97546


Ion	Ratio	Lower	Upper
43	100		
86	9.3	7.5	11.3

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025







#24
1,1-Dichloroethane
Concen: 50.059 ug/l
RT: 3.599 min Scan# 412
Delta R.T. -0.006 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion: 63 Resp: 122843

Ion	Ratio	Lower	Upper
63	100		
98	6.2	3.4	10.1
100	4.1	2.2	6.6






#25
2-Butanone
Concen: 250.017 ug/l
RT: 4.550 min Scan# 5
Delta R.T. -0.006 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion: 43 Resp: 26577


Ion	Ratio	Lower	Upper
43	100		
72	22.6	18.6	28.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#26
2,2-Dichloropropane
Concen: 57.558 ug/l
RT: 4.465 min Scan# 554
Delta R.T. -0.006 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion: 77 Resp: 91259

Ion	Ratio	Lower	Upper
77	100		
97	20.8	10.6	31.8





#27
cis-1,2-Dichloroethene
Concen: 48.765 ug/l
RT: 4.477 min Scan# 5
Delta R.T. -0.006 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00


Tgt Ion: 96 Resp: 7046


Ion	Ratio	Lower	Upper
96	100		
61	160.3	0.0	313.2
98	64.3	0.0	126.4

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025







#28
Bromochloromethane
Concen: 49.658 ug/l
RT: 4.885 min Scan# 623
Delta R.T. -0.012 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion: 49 Resp: 58888

Ion	Ratio	Lower	Upper
49	100		
129	1.8	0.0	3.8
130	67.2	54.6	81.8





#29
Tetrahydrofuran
Concen: 244.391 ug/l
RT: 5.001 min Scan# 64
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00


Tgt Ion: 42 Resp: 16843


Ion	Ratio	Lower	Upper
42	100		
72	41.9	33.4	50.0
71	38.8	31.4	47.0

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025







#30
Chloroform
Concen: 50.378 ug/l
RT: 5.080 min Scan# 655
Delta R.T. -0.006 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion: 83 Resp: 127207

Ion	Ratio	Lower	Upper
83	100		
85	62.4	52.8	79.2






#31
Cyclohexane
Concen: 50.650 ug/l
RT: 5.458 min Scan# 7
Delta R.T. -0.006 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion: 56 Resp: 10988


Ion	Ratio	Lower	Upper
56	100		
69	29.9	23.6	35.4
84	80.7	67.7	101.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#32
1,1,1-Trichloroethane
Concen: 50.113 ug/l
RT: 5.373 min Scan# 703
Delta R.T. -0.006 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion: 97 Resp: 107072

Ion	Ratio	Lower	Upper
97	100		
99	64.6	51.7	77.5
61	52.1	41.0	61.4






#33
1,2-Dichloroethane-d4
Concen: 42.428 ug/l
RT: 5.946 min Scan# 7
Delta R.T. -0.006 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion: 65 Resp: 7503


Ion	Ratio	Lower	Upper
65	100		
67	48.9	0.0	99.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.751 min Scan# 929
Delta R.T. -0.006 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion:114 Resp: 163287

Ion	Ratio	Lower	Upper
114	100		
63	23.3	0.0	46.8
88	16.9	0.0	35.4






#35
Dibromofluoromethane
Concen: 43.879 ug/l
RT: 5.379 min Scan# 7
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion:113 Resp: 5083


Ion	Ratio	Lower	Upper
113	100		
111	102.8	81.8	122.6
192	16.2	13.8	20.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#36
1,1-Dichloropropene
Concen: 52.336 ug/l
RT: 5.684 min Scan# 754
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion: 75 Resp: 81866

Ion	Ratio	Lower	Upper
75	100		
110	33.3	16.9	50.6
77	31.0	25.0	37.4






#37
Ethyl Acetate
Concen: 51.305 ug/l
RT: 4.708 min Scan# 5
Delta R.T. -0.006 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion: 43 Resp: 10113


Ion	Ratio	Lower	Upper
43	100		
61	12.9	9.9	14.9
70	9.6	7.5	11.3



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#38
Carbon Tetrachloride
Concen: 55.161 ug/l
RT: 5.666 min Scan# 751
Delta R.T. -0.006 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion:117 Resp: 93252

Ion	Ratio	Lower	Upper
117	100		
119	94.6	76.9	115.3
121	30.3	24.2	36.2





#39
Methylcyclohexane
Concen: 56.163 ug/l
RT: 7.373 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00


Tgt Ion: 83 Resp: 10768


Ion	Ratio	Lower	Upper
83	100		
55	81.0	64.9	97.3
98	46.8	37.4	56.0

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025







#40
Benzene
Concen: 51.517 ug/l
RT: 6.025 min Scan# 810
Delta R.T. -0.006 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion: 78 Resp: 246115

Ion	Ratio	Lower	Upper
78	100		
77	24.1	19.0	28.6






#41
Methacrylonitrile
Concen: 56.425 ug/l
RT: 4.910 min Scan# 6
Delta R.T. -0.012 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion: 41 Resp: 5868


Ion	Ratio	Lower	Upper
41	100		
39	56.4	47.0	70.6
67	64.0	53.2	79.8
52	30.4	25.0	37.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#42
1,2-Dichloroethane
Concen: 53.340 ug/l
RT: 6.074 min Scan# 818
Delta R.T. -0.006 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion: 62 Resp: 105249

Ion	Ratio	Lower	Upper
62	100		
98	7.6	0.0	15.2





#43
Isopropyl Acetate
Concen: 52.737 ug/l
RT: 6.330 min Scan# 8
Delta R.T. -0.006 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00


Tgt Ion: 43 Resp: 15756


Ion	Ratio	Lower	Upper
43	100		
61	17.1	14.1	21.1
87	11.8	9.2	13.8

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025







#44
Trichloroethene
Concen: 51.308 ug/l
RT: 7.117 min Scan# 989
Delta R.T. -0.006 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion:130 Resp: 58260

Ion	Ratio	Lower	Upper
130	100		
95	107.1	0.0	202.0






#45
1,2-Dichloropropane
Concen: 52.514 ug/l
RT: 7.421 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion: 63 Resp: 6259


Ion	Ratio	Lower	Upper
63	100		
65	31.3	25.8	38.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#46
Dibromomethane
Concen: 52.011 ug/l
RT: 7.574 min Scan# 1064
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion: 93 Resp: 47591

Ion	Ratio	Lower	Upper
93	100		
95	82.8	64.8	97.2
174	83.4	65.4	98.0






#47
Bromodichloromethane
Concen: 53.371 ug/l
RT: 7.818 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion: 83 Resp: 9752


Ion	Ratio	Lower	Upper
83	100		
85	61.9	49.7	74.5
127	7.9	6.1	9.1



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#48
Methyl methacrylate
Concen: 53.861 ug/l
RT: 7.690 min Scan# 1083
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion: 41 Resp: 83087

Ion	Ratio	Lower	Upper
41	100		
69	72.1	58.2	87.2
39	65.7	53.6	80.4





#49
1,4-Dioxane
Concen: 972.308 ug/l
RT: 7.659 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00


Tgt Ion: 88 Resp: 2713


Ion	Ratio	Lower	Upper
88	100		
43	43.6	32.2	48.4
58	76.3	58.6	88.0

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025







#50
Toluene-d8
Concen: 43.586 ug/l
RT: 8.641 min Scan# 1239
Delta R.T. -0.006 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion: 98 Resp: 176247

Ion	Ratio	Lower	Upper
98	100		
100	65.0	52.2	78.4






#51
4-Methyl-2-Pentanone
Concen: 268.424 ug/l
RT: 8.568 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion: 43 Resp: 53164


Ion	Ratio	Lower	Upper
43	100		
58	35.2	28.5	42.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#52
Toluene
Concen: 52.369 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion: 92 Resp: 151389

Ion	Ratio	Lower	Upper
92	100		
91	172.1	138.7	208.1






#53
t-1,3-Dichloropropene
Concen: 49.056 ug/l
RT: 8.976 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion: 75 Resp: 8612


Ion	Ratio	Lower	Upper
75	100		
77	31.7	25.4	38.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#54
cis-1,3-Dichloropropene
Concen: 55.919 ug/l
RT: 8.360 min Scan# 1193
Delta R.T. -0.006 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion: 75 Resp: 96025

Ion	Ratio	Lower	Upper
75	100		
77	30.9	24.6	36.8
39	63.2	48.4	72.6






#55
1,1,2-Trichloroethane
Concen: 52.506 ug/l
RT: 9.147 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion: 97 Resp: 6048


Ion	Ratio	Lower	Upper
97	100		
83	85.2	71.8	107.8
85	54.6	47.0	70.6
99	61.6	48.6	73.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#56
Ethyl methacrylate
Concen: 53.447 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion: 69 Resp: 95283

Ion	Ratio	Lower	Upper
69	100		
41	77.7	61.3	91.9
39	50.0	40.2	60.2






#57
1,3-Dichloropropane
Concen: 52.594 ug/l
RT: 9.305 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion: 76 Resp: 10550


Ion	Ratio	Lower	Upper
76	100		
78	31.7	26.0	39.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#58
2-Chloroethyl Vinyl ether
Concen: 256.573 ug/l
RT: 8.238 min Scan# 1173
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion: 63 Resp: 231389

Ion	Ratio	Lower	Upper
63	100		
106	26.5	21.0	31.6






#59
2-Hexanone
Concen: 266.730 ug/l
RT: 9.427 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion: 43 Resp: 39121


Ion	Ratio	Lower	Upper
43	100		
58	49.0	24.9	74.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#60
Dibromochloromethane
Concen: 55.012 ug/l
RT: 9.519 min Scan# 1383
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion:129 Resp: 69085

Ion	Ratio	Lower	Upper
129	100		
127	76.1	39.4	118.2






#61
1,2-Dibromoethane
Concen: 53.230 ug/l
RT: 9.604 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion:107 Resp: 6207


Ion	Ratio	Lower	Upper
107	100		
109	93.9	75.7	113.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#62
4-Bromofluorobenzene
Concen: 45.563 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion: 95 Resp: 67111

Ion	Ratio	Lower	Upper
95	100		
174	66.8	0.0	135.8
176	64.5	0.0	131.4






#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion:117 Resp: 14204


Ion	Ratio	Lower	Upper
117	100		
82	61.8	49.2	73.8
119	32.1	25.1	37.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#64
Tetrachloroethene
Concen: 53.258 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. -0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion:164 Resp: 53413

Ion	Ratio	Lower	Upper
164	100		
166	125.6	103.8	155.8
129	102.4	84.3	126.5
131	100.0	80.9	121.3






#65
Chlorobenzene
Concen: 53.041 ug/l
RT: 10.073 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion:112 Resp: 16099


Ion	Ratio	Lower	Upper
112	100		
114	32.3	24.7	37.1



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#66
1,1,1,2-Tetrachloroethane
Concen: 52.492 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion:131 Resp: 55317

Ion	Ratio	Lower	Upper
131	100		
133	96.5	47.5	142.5
119	64.9	32.9	98.6






#67
Ethyl Benzene
Concen: 54.700 ug/l
RT: 10.189 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion: 91 Resp: 29736


Ion	Ratio	Lower	Upper
91	100		
106	29.4	23.7	35.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#68
m/p-Xylenes
Concen: 109.636 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion:106 Resp: 216775

Ion	Ratio	Lower	Upper
106	100		
91	218.4	172.6	258.8





#69
o-Xylene
Concen: 54.386 ug/l
RT: 10.640 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00


Tgt Ion:106 Resp: 10593


Ion	Ratio	Lower	Upper
106	100		
91	228.1	113.6	340.6

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025







#70
Styrene
Concen: 55.886 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion:104 Resp: 179659

Ion	Ratio	Lower	Upper
104	100		
78	57.9	45.4	68.2
103	54.6	43.8	65.8






#71
Bromoform
Concen: 55.486 ug/l
RT: 10.799 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion:173 Resp: 4365


Ion	Ratio	Lower	Upper
173	100		
175	48.2	23.4	70.2
254	0.0	0.0	0.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion:152 Resp: 67379

Ion	Ratio	Lower	Upper
152	100		
115	95.7	46.9	140.7
150	173.4	0.0	349.4






#73
Isopropylbenzene
Concen: 53.453 ug/l
RT: 10.957 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion:105 Resp: 28849


Ion	Ratio	Lower	Upper
105	100		
120	25.4	12.7	38.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#74
N-amyl acetate
Concen: 52.015 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion: 43 Resp: 134287

Ion	Ratio	Lower	Upper
43	100		
70	37.7	30.1	45.1
55	23.0	18.3	27.5
61	21.3	17.5	26.3






#75
1,1,2,2-Tetrachloroethane
Concen: 48.021 ug/l
RT: 11.207 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion: 83 Resp: 9104


Ion	Ratio	Lower	Upper
83	100		
131	9.7	4.7	14.1
85	64.4	32.4	97.2



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#76
1,2,3-Trichloropropane
Concen: 49.496 ug/l m
RT: 11.238 min Scan# 1665
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion: 75 Resp: 81348

Ion	Ratio	Lower	Upper
75	100		
77	39.7	20.0	59.9






#77
Bromobenzene
Concen: 51.248 ug/l
RT: 11.195 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion:156 Resp: 6390


Ion	Ratio	Lower	Upper
156	100		
77	185.3	92.6	277.8
158	97.2	48.1	144.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#78
n-propylbenzene
Concen: 54.351 ug/l
RT: 11.299 min Scan# 1675
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion: 91 Resp: 337878

Ion	Ratio	Lower	Upper
91	100		
120	21.3	10.8	32.3





#79
2-Chlorotoluene
Concen: 51.314 ug/l
RT: 11.360 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00


Tgt Ion: 91 Resp: 20393


Ion	Ratio	Lower	Upper
91	100		
126	31.2	15.8	47.4

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025







#80
1,3,5-Trimethylbenzene
Concen: 53.712 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion:105 Resp: 239716

Ion	Ratio	Lower	Upper
105	100		
120	46.2	23.0	69.0





#81
trans-1,4-Dichloro-2-butene
Concen: 52.135 ug/l
RT: 11.018 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00


Tgt Ion: 75 Resp: 2524


Ion	Ratio	Lower	Upper
75	100		
53	128.6	101.8	152.6
89	47.2	37.2	55.8

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025







#82
4-Chlorotoluene
Concen: 53.128 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion: 91 Resp: 235414

Ion	Ratio	Lower	Upper
91	100		
126	26.8	13.3	39.9






#83
tert-Butylbenzene
Concen: 53.024 ug/l
RT: 11.713 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion:119 Resp: 23428


Ion	Ratio	Lower	Upper
119	100		
91	65.5	32.4	97.2
134	22.8	11.5	34.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#84
1,2,4-Trimethylbenzene
Concen: 53.903 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion:105 Resp: 241455

Ion	Ratio	Lower	Upper
105	100		
120	42.2	21.4	64.3





#85
sec-Butylbenzene
Concen: 54.969 ug/l
RT: 11.890 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00


Tgt Ion:105 Resp: 29831


Ion	Ratio	Lower	Upper
105	100		
134	18.9	9.6	28.6

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025







#86
p-Isopropyltoluene
Concen: 55.060 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion:119 Resp: 246346

Ion	Ratio	Lower	Upper
119	100		
134	25.4	12.6	37.8
91	27.9	13.7	41.1





#87
1,3-Dichlorobenzene
Concen: 51.535 ug/l
RT: 11.963 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00


Tgt Ion:146 Resp: 11698


Ion	Ratio	Lower	Upper
146	100		
111	44.3	22.2	66.6
148	63.6	31.9	95.7

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025







#88
1,4-Dichlorobenzene
Concen: 50.768 ug/l
RT: 12.036 min Scan# 1796
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion:146 Resp: 116831

Ion	Ratio	Lower	Upper
146	100		
111	42.3	20.9	62.8
148	64.5	31.5	94.5





#89
n-Butylbenzene
Concen: 57.557 ug/l
RT: 12.329 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00


Tgt Ion: 91 Resp: 22331


Ion	Ratio	Lower	Upper
91	100		
92	53.0	26.9	80.6
134	23.5	11.9	35.7

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025







#90
Hexachloroethane
Concen: 55.342 ug/l
RT: 12.536 min Scan# 1878
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion:117 Resp: 42554

Ion	Ratio	Lower	Upper
117	100		
201	63.1	31.2	93.6






#91
1,2-Dichlorobenzene
Concen: 51.109 ug/l
RT: 12.335 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion:146 Resp: 11534


Ion	Ratio	Lower	Upper
146	100		
111	46.1	22.9	68.8
148	63.7	31.9	95.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#92
1,2-Dibromo-3-Chloropropane
Concen: 53.714 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion: 75 Resp: 21298

Ion	Ratio	Lower	Upper
75	100		
155	69.8	34.9	104.7
157	87.7	44.9	134.7






#93
1,2,4-Trichlorobenzene
Concen: 54.225 ug/l
RT: 13.585 min Scan# 2
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion:180 Resp: 6811


Ion	Ratio	Lower	Upper
180	100		
182	94.8	47.9	143.6
145	34.2	17.9	53.8



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#94
Hexachlorobutadiene
Concen: 54.435 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. 0.006 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion:225 Resp: 29467

Ion	Ratio	Lower	Upper
225	100		
223	64.2	31.1	93.3
227	61.2	32.2	96.6






#95
Naphthalene
Concen: 51.558 ug/l
RT: 13.774 min Scan# 2
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion:128 Resp: 24102


Ion	Ratio	Lower	Upper
128	100		
127	12.7	10.3	15.5
129	11.1	8.6	12.8



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025



#96
1,2,3-Trichlorobenzene
Concen: 51.875 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. 0.000 min
Lab File: VX045586.D
Acq: 04 Apr 2025 10:00

Tgt Ion:180 Resp: 68237

Ion	Ratio	Lower	Upper
180	100		
182	95.6	47.3	141.8
145	36.7	18.2	54.6



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040425\
Data File : VX045586.D
Acq On : 04 Apr 2025 10:00
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Apr 04 23:10:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1	I	Pentafluorobenzene	1.000	1.000	0.0	101	-0.01
2	T	Dichlorodifluoromethane	0.748	0.843	-12.7	104	0.00
3	P	Chloromethane	0.768	0.797	-3.8	103	0.00
4	C	Vinyl Chloride	0.703	0.731	-4.0#	103	0.00
5	T	Bromomethane	0.333	0.328	1.5	100	0.00
6	T	Chloroethane	0.373	0.393	-5.4	100	0.00
7	T	Trichlorofluoromethane	1.048	1.105	-5.4	103	0.00
8	T	Diethyl Ether	0.352	0.350	0.6	100	0.00
9	T	1,1,2-Trichlorotrifluoroeth	0.614	0.658	-7.2	107	0.00
10	T	Methyl Iodide	0.766	0.770	-0.5	97	0.00
11	T	Tert butyl alcohol	0.123	0.118	4.1	94	0.00
12	CM	1,1-Dichloroethene	0.600	0.599	0.2#	100	0.00
13	T	Acrolein	0.169	0.199	-17.8	121	0.00
14	T	Allyl chloride	1.139	1.176	-3.2	101	0.00
15	T	Acrylonitrile	0.388	0.386	0.5	96	0.00
16	T	Acetone	0.375	0.368	1.9	99	0.00
17	T	Carbon Disulfide	1.483	1.500	-1.1	98	0.00
18	T	Methyl Acetate	0.864	0.855	1.0	101	0.00
19	T	Methyl tert-butyl Ether	2.100	2.104	-0.2	100	0.00
20	T	Methylene Chloride	0.703	0.694	1.3	99	0.00
21	T	trans-1,2-Dichloroethene	0.613	0.609	0.7	99	0.00
22	T	Diisopropyl ether	2.243	2.282	-1.7	101	0.00
23	T	Vinyl Acetate	1.938	2.017	-4.1	98	0.00
24	P	1,1-Dichloroethane	1.269	1.270	-0.1	101	0.00
25	T	2-Butanone	0.550	0.550	0.0	95	0.00
26	T	2,2-Dichloropropane	0.820	0.944	-15.1	113	0.00
27	T	cis-1,2-Dichloroethene	0.747	0.729	2.4	98	0.00
28	T	Bromochloromethane	0.613	0.609	0.7	103	-0.01
29	T	Tetrahydrofuran	0.356	0.348	2.2	95	0.00
30	C	Chloroform	1.306	1.315	-0.7#	101	0.00
31	T	Cyclohexane	1.122	1.136	-1.2	102	0.00
32	T	1,1,1-Trichloroethane	1.105	1.107	-0.2	101	0.00
33	S	1,2-Dichloroethane-d4	0.914	0.776	15.1	90	0.00
34	I	1,4-Difluorobenzene	1.000	1.000	0.0	97	0.00
35	S	Dibromofluoromethane	0.355	0.311	12.4	87	0.00
36	T	1,1-Dichloropropene	0.479	0.501	-4.6	100	0.00
37	T	Ethyl Acetate	0.604	0.619	-2.5	96	0.00
38	T	Carbon Tetrachloride	0.518	0.571	-10.2	103	0.00
39	T	Methylcyclohexane	0.587	0.660	-12.4	103	0.00
40	TM	Benzene	1.463	1.507	-3.0	99	0.00
41	T	Methacrylonitrile	0.318	0.359	-12.9	99	-0.01
42	TM	1,2-Dichloroethane	0.604	0.645	-6.8	101	0.00
43	T	Isopropyl Acetate	0.915	0.965	-5.5	96	0.00
44	TM	Trichloroethene	0.348	0.357	-2.6	99	0.00
45	C	1,2-Dichloropropane	0.365	0.383	-4.9#	99	0.00
46	T	Dibromomethane	0.280	0.291	-3.9	97	0.00
47	T	Bromodichloromethane	0.560	0.597	-6.6	101	0.00
48	T	Methyl methacrylate	0.472	0.509	-7.8	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040425\
Data File : VX045586.D
Acq On : 04 Apr 2025 10:00
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Apr 04 23:10:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49	T	1,4-Dioxane	0.009	0.008	11.1	87	0.00
50	S	Toluene-d8	1.238	1.079	12.8	86	0.00
51	T	4-Methyl-2-Pentanone	0.606	0.651	-7.4	95	0.00
52	CM	Toluene	0.885	0.927	-4.7#	99	0.00
53	T	t-1,3-Dichloropropene	0.467	0.527	-12.8	101	0.00
54	T	cis-1,3-Dichloropropene	0.526	0.588	-11.8	101	0.00
55	T	1,1,2-Trichloroethane	0.353	0.370	-4.8	100	0.00
56	T	Ethyl methacrylate	0.546	0.584	-7.0	94	0.00
57	T	1,3-Dichloropropane	0.614	0.646	-5.2	99	0.00
58	T	2-Chloroethyl Vinyl ether	0.276	0.283	-2.5	92	0.00
59	T	2-Hexanone	0.449	0.479	-6.7	95	0.00
60	T	Dibromochloromethane	0.385	0.423	-9.9	100	0.00
61	T	1,2-Dibromoethane	0.357	0.380	-6.4	99	0.00
62	S	4-Bromofluorobenzene	0.451	0.411	8.9	88	0.00
63	I	Chlorobenzene-d5	1.000	1.000	0.0	95	0.00
64	T	Tetrachloroethene	0.353	0.376	-6.5	102	0.00
65	PM	Chlorobenzene	1.068	1.133	-6.1	99	0.00
66	T	1,1,1,2-Tetrachloroethane	0.371	0.389	-4.9	99	0.00
67	C	Ethyl Benzene	1.914	2.093	-9.4#	99	0.00
68	T	m/p-Xylenes	0.696	0.763	-9.6	99	0.00
69	T	o-Xylene	0.686	0.746	-8.7	98	0.00
70	T	Styrene	1.132	1.265	-11.7	98	0.00
71	P	Bromoform	0.277	0.307	-10.8	98	0.00
72	I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
73	T	Isopropylbenzene	4.005	4.282	-6.9	100	0.00
74	T	N-amyl acetate	1.916	1.993	-4.0	95	0.00
75	P	1,1,2,2-Tetrachloroethane	1.407	1.351	4.0	96	0.00
76	T	1,2,3-Trichloropropane	1.220	1.207	1.1	97	0.00
77	T	Bromobenzene	0.925	0.948	-2.5	99	0.00
78	T	n-propylbenzene	4.613	5.015	-8.7	100	0.00
79	T	2-Chlorotoluene	2.949	3.027	-2.6	100	0.00
80	T	1,3,5-Trimethylbenzene	3.312	3.558	-7.4	99	0.00
81	T	trans-1,4-Dichloro-2-butene	0.359	0.375	-4.5	99	0.00
82	T	4-Chlorotoluene	3.288	3.494	-6.3	99	0.00
83	T	tert-Butylbenzene	3.279	3.477	-6.0	99	0.00
84	T	1,2,4-Trimethylbenzene	3.324	3.584	-7.8	100	0.00
85	T	sec-Butylbenzene	4.027	4.427	-9.9	101	0.00
86	T	p-Isopropyltoluene	3.320	3.656	-10.1	102	0.00
87	T	1,3-Dichlorobenzene	1.684	1.736	-3.1	100	0.00
88	T	1,4-Dichlorobenzene	1.708	1.734	-1.5	100	0.00
89	T	n-Butylbenzene	2.879	3.314	-15.1	103	0.00
90	T	Hexachloroethane	0.571	0.632	-10.7	104	0.00
91	T	1,2-Dichlorobenzene	1.675	1.712	-2.2	100	0.00
92	T	1,2-Dibromo-3-Chloropropane	0.294	0.316	-7.5	98	0.00
93	T	1,2,4-Trichlorobenzene	0.932	1.011	-8.5	105	0.00
94	T	Hexachlorobutadiene	0.402	0.437	-8.7	106	0.00
95	T	Naphthalene	3.469	3.577	-3.1	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040425\
Data File : VX045586.D
Acq On : 04 Apr 2025 10:00
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Apr 04 23:10:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.976	1.013	-3.8	99	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040425\
Data File : VX045586.D
Acq On : 04 Apr 2025 10:00
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Apr 04 23:10:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	Pentafluorobenzene	50.000	50.000	0.0	101	-0.01
2	T	Dichlorodifluoromethane	50.000	56.382	-12.8	104	0.00
3	P	Chloromethane	50.000	51.883	-3.8	103	0.00
4	C	Vinyl Chloride	50.000	51.962	-3.9#	103	0.00
5	T	Bromomethane	50.000	49.218	1.6	100	0.00
6	T	Chloroethane	50.000	52.626	-5.3	100	0.00
7	T	Trichlorofluoromethane	50.000	52.688	-5.4	103	0.00
8	T	Diethyl Ether	50.000	49.704	0.6	100	0.00
9	T	1,1,2-Trichlorotrifluoroeth	50.000	53.591	-7.2	107	0.00
10	T	Methyl Iodide	50.000	50.258	-0.5	97	0.00
11	T	Tert butyl alcohol	250.000	239.188	4.3	94	0.00
12	CM	1,1-Dichloroethene	50.000	49.883	0.2#	100	0.00
13	T	Acrolein	250.000	294.566	-17.8	121	0.00
14	T	Allyl chloride	50.000	51.619	-3.2	101	0.00
15	T	Acrylonitrile	250.000	248.485	0.6	96	0.00
16	T	Acetone	250.000	245.130	1.9	99	0.00
17	T	Carbon Disulfide	50.000	50.605	-1.2	98	0.00
18	T	Methyl Acetate	50.000	49.477	1.0	101	0.00
19	T	Methyl tert-butyl Ether	50.000	50.091	-0.2	100	0.00
20	T	Methylene Chloride	50.000	49.375	1.3	99	0.00
21	T	trans-1,2-Dichloroethene	50.000	49.659	0.7	99	0.00
22	T	Diisopropyl ether	50.000	50.863	-1.7	101	0.00
23	T	Vinyl Acetate	250.000	260.190	-4.1	98	0.00
24	P	1,1-Dichloroethane	50.000	50.059	-0.1	101	0.00
25	T	2-Butanone	250.000	250.017	-0.0	95	0.00
26	T	2,2-Dichloropropane	50.000	57.558	-15.1	113	0.00
27	T	cis-1,2-Dichloroethene	50.000	48.765	2.5	98	0.00
28	T	Bromochloromethane	50.000	49.658	0.7	103	-0.01
29	T	Tetrahydrofuran	250.000	244.391	2.2	95	0.00
30	C	Chloroform	50.000	50.378	-0.8#	101	0.00
31	T	Cyclohexane	50.000	50.650	-1.3	102	0.00
32	T	1,1,1-Trichloroethane	50.000	50.113	-0.2	101	0.00
33	S	1,2-Dichloroethane-d4	50.000	42.428	15.1	90	0.00
34	I	1,4-Difluorobenzene	50.000	50.000	0.0	97	0.00
35	S	Dibromofluoromethane	50.000	43.879	12.2	87	0.00
36	T	1,1-Dichloropropene	50.000	52.336	-4.7	100	0.00
37	T	Ethyl Acetate	50.000	51.305	-2.6	96	0.00
38	T	Carbon Tetrachloride	50.000	55.161	-10.3	103	0.00
39	T	Methylcyclohexane	50.000	56.163	-12.3	103	0.00
40	TM	Benzene	50.000	51.517	-3.0	99	0.00
41	T	Methacrylonitrile	50.000	56.425	-12.8	99	-0.01
42	TM	1,2-Dichloroethane	50.000	53.340	-6.7	101	0.00
43	T	Isopropyl Acetate	50.000	52.737	-5.5	96	0.00
44	TM	Trichloroethene	50.000	51.308	-2.6	99	0.00
45	C	1,2-Dichloropropane	50.000	52.514	-5.0#	99	0.00
46	T	Dibromomethane	50.000	52.011	-4.0	97	0.00
47	T	Bromodichloromethane	50.000	53.371	-6.7	101	0.00
48	T	Methyl methacrylate	50.000	53.861	-7.7	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040425\
Data File : VX045586.D
Acq On : 04 Apr 2025 10:00
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Apr 04 23:10:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49	T	1,4-Dioxane	1000.000	972.308	2.8	87	0.00
50	S	Toluene-d8	50.000	43.586	12.8	86	0.00
51	T	4-Methyl-2-Pentanone	250.000	268.424	-7.4	95	0.00
52	CM	Toluene	50.000	52.369	-4.7#	99	0.00
53	T	t-1,3-Dichloropropene	50.000	49.056	1.9	101	0.00
54	T	cis-1,3-Dichloropropene	50.000	55.919	-11.8	101	0.00
55	T	1,1,2-Trichloroethane	50.000	52.506	-5.0	100	0.00
56	T	Ethyl methacrylate	50.000	53.447	-6.9	94	0.00
57	T	1,3-Dichloropropane	50.000	52.594	-5.2	99	0.00
58	T	2-Chloroethyl Vinyl ether	250.000	256.573	-2.6	92	0.00
59	T	2-Hexanone	250.000	266.730	-6.7	95	0.00
60	T	Dibromochloromethane	50.000	55.012	-10.0	100	0.00
61	T	1,2-Dibromoethane	50.000	53.230	-6.5	99	0.00
62	S	4-Bromofluorobenzene	50.000	45.563	8.9	88	0.00
63	I	Chlorobenzene-d5	50.000	50.000	0.0	95	0.00
64	T	Tetrachloroethene	50.000	53.258	-6.5	102	0.00
65	PM	Chlorobenzene	50.000	53.041	-6.1	99	0.00
66	T	1,1,1,2-Tetrachloroethane	50.000	52.492	-5.0	99	0.00
67	C	Ethyl Benzene	50.000	54.700	-9.4#	99	0.00
68	T	m/p-Xylenes	100.000	109.636	-9.6	99	0.00
69	T	o-Xylene	50.000	54.386	-8.8	98	0.00
70	T	Styrene	50.000	55.886	-11.8	98	0.00
71	P	Bromoform	50.000	55.486	-11.0	98	0.00
72	I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	98	0.00
73	T	Isopropylbenzene	50.000	53.453	-6.9	100	0.00
74	T	N-amyl acetate	50.000	52.015	-4.0	95	0.00
75	P	1,1,2,2-Tetrachloroethane	50.000	48.021	4.0	96	0.00
76	T	1,2,3-Trichloropropane	50.000	49.496	1.0	97	0.00
77	T	Bromobenzene	50.000	51.248	-2.5	99	0.00
78	T	n-propylbenzene	50.000	54.351	-8.7	100	0.00
79	T	2-Chlorotoluene	50.000	51.314	-2.6	100	0.00
80	T	1,3,5-Trimethylbenzene	50.000	53.712	-7.4	99	0.00
81	T	trans-1,4-Dichloro-2-butene	50.000	52.135	-4.3	99	0.00
82	T	4-Chlorotoluene	50.000	53.128	-6.3	99	0.00
83	T	tert-Butylbenzene	50.000	53.024	-6.0	99	0.00
84	T	1,2,4-Trimethylbenzene	50.000	53.903	-7.8	100	0.00
85	T	sec-Butylbenzene	50.000	54.969	-9.9	101	0.00
86	T	p-Isopropyltoluene	50.000	55.060	-10.1	102	0.00
87	T	1,3-Dichlorobenzene	50.000	51.535	-3.1	100	0.00
88	T	1,4-Dichlorobenzene	50.000	50.768	-1.5	100	0.00
89	T	n-Butylbenzene	50.000	57.557	-15.1	103	0.00
90	T	Hexachloroethane	50.000	55.342	-10.7	104	0.00
91	T	1,2-Dichlorobenzene	50.000	51.109	-2.2	100	0.00
92	T	1,2-Dibromo-3-Chloropropane	50.000	53.714	-7.4	98	0.00
93	T	1,2,4-Trichlorobenzene	50.000	54.225	-8.5	105	0.00
94	T	Hexachlorobutadiene	50.000	54.435	-8.9	106	0.00
95	T	Naphthalene	50.000	51.558	-3.1	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040425\
Data File : VX045586.D
Acq On : 04 Apr 2025 10:00
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Apr 04 23:10:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	51.875	-3.8	99	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: JACO05

Lab Code: CHEM Case No.: Q1731 SAS No.: Q1731 SDG No.: Q1731

Instrument ID: MSVOA_X Calibration Date/Time: 04/07/2025 09:17

Lab File ID: VX045614.D Init. Calib. Date(s): 04/01/2025 04/01/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 17:06 19:02

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.703	0.707		0.57	20
1,1-Dichloroethene	0.600	0.585		-2.5	20
1,1-Dichloroethane	1.269	1.210	0.1	-4.65	20
cis-1,2-Dichloroethene	0.747	0.708		-5.22	20
1,1,1-Trichloroethane	1.105	1.091		-1.27	20
Benzene	1.463	1.436		-1.85	20
1,2-Dichloroethane	0.604	0.605		0.17	20
Trichloroethene	0.348	0.346		-0.57	20
1,1,2-Trichloroethane	0.353	0.349		-1.13	20
Tetrachloroethene	0.353	0.360		1.98	20
1,2-Dichloroethane-d4	0.914	0.904		-1.09	20
Dibromofluoromethane	0.355	0.366		3.1	20
Toluene-d8	1.238	1.255		1.37	20
4-Bromofluorobenzene	0.451	0.480		6.43	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040725\
Data File : VX045614.D
Acq On : 07 Apr 2025 09:17
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Apr 08 01:33:45 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations APPROVED

Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	5.538	168	104243	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	6.751	114	177492	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	157116	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	73594	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	94254	49.443	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 98.880%			
35) Dibromofluoromethane	5.379	113	64970	51.592	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 103.180%			
50) Toluene-d8	8.641	98	222665	50.658	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 101.320%			
62) 4-Bromofluorobenzene	11.079	95	85194	53.211	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 106.420%			
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.166	85	86584	55.540	ug/l	100
3) Chloromethane	1.307	50	79882	49.877	ug/l	99
4) Vinyl Chloride	1.374	62	73660	50.256	ug/l	99
5) Bromomethane	1.593	94	32649	46.980	ug/l	99
6) Chloroethane	1.660	64	39253	50.478	ug/l	98
7) Trichlorofluoromethane	1.867	101	111225	50.889	ug/l	100
8) Diethyl Ether	2.130	74	36037	49.111	ug/l	99
9) 1,1,2-Trichlorotrifluo...	2.319	101	66695	52.076	ug/l	98
10) Methyl Iodide	2.441	142	79791	49.948	ug/l	97
11) Tert butyl alcohol	2.977	59	64541	251.921	ug/l	99
12) 1,1-Dichloroethene	2.306	96	61015	48.756	ug/l	99
13) Acrolein	2.233	56	58537	165.921	ug/l	99
14) Allyl chloride	2.654	41	118904	50.076	ug/l	99
15) Acrylonitrile	3.062	53	196504	242.982	ug/l	99
16) Acetone	2.380	43	235381	300.693	ug/l	98
17) Carbon Disulfide	2.502	76	153475	49.655	ug/l	100
18) Methyl Acetate	2.703	43	90670	50.309	ug/l	100
19) Methyl tert-butyl Ether	3.111	73	216183	49.382	ug/l	99
20) Methylene Chloride	2.782	84	69933	47.720	ug/l	95
21) trans-1,2-Dichloroethene	3.081	96	61608	48.187	ug/l	97
22) Diisopropyl ether	3.757	45	226624	48.452	ug/l	99
23) Vinyl Acetate	3.715	43	1018875	252.123	ug/l	99
24) 1,1-Dichloroethane	3.599	63	126177	47.700	ug/l	100
25) 2-Butanone	4.550	43	301049	262.730	ug/l	98
26) 2,2-Dichloropropane	4.458	77	98828	57.826	ug/l	99
27) cis-1,2-Dichloroethene	4.477	96	73761	47.355	ug/l	97
28) Bromochloromethane	4.885	49	59900	46.860	ug/l	97
29) Tetrahydrofuran	4.995	42	175733	236.552	ug/l	99
30) Chloroform	5.080	83	131279	48.232	ug/l	99
31) Cyclohexane	5.458	56	112556	48.132	ug/l	95
32) 1,1,1-Trichloroethane	5.373	97	113722	49.377	ug/l	99
36) 1,1-Dichloropropene	5.678	75	85301	50.168	ug/l	98
37) Ethyl Acetate	4.708	43	106911	49.894	ug/l	99
38) Carbon Tetrachloride	5.666	117	95420	51.926	ug/l	99
39) Methylcyclohexane	7.373	83	111471	53.483	ug/l	98
40) Benzene	6.025	78	254875	49.081	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040725\
Data File : VX045614.D
Acq On : 07 Apr 2025 09:17
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Apr 08 01:33:45 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations APPROVED

Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
41) Methacrylonitrile	4.910	41	59241	52.399	ug/l		99
42) 1,2-Dichloroethane	6.080	62	107343	50.048	ug/l		100
43) Isopropyl Acetate	6.330	43	166101	51.145	ug/l		99
44) Trichloroethene	7.117	130	61500	49.827	ug/l		98
45) 1,2-Dichloropropane	7.421	63	64374	49.683	ug/l		99
46) Dibromomethane	7.574	93	49916	50.186	ug/l		98
47) Bromodichloromethane	7.812	83	103188	51.949	ug/l		98
48) Methyl methacrylate	7.690	41	87167	51.984	ug/l		99
49) 1,4-Dioxane	7.653	88	33084	1090.518	ug/l		99
51) 4-Methyl-2-Pentanone	8.568	43	554818	257.706	ug/l		100
52) Toluene	8.714	92	156451	49.789	ug/l		98
53) t-1,3-Dichloropropene	8.976	75	92990	48.743	ug/l		99
54) cis-1,3-Dichloropropene	8.360	75	102648	54.992	ug/l		98
55) 1,1,2-Trichloroethane	9.147	97	62014	49.526	ug/l		97
56) Ethyl methacrylate	9.116	69	102307	52.794	ug/l		99
57) 1,3-Dichloropropane	9.305	76	109205	50.084	ug/l		100
58) 2-Chloroethyl Vinyl ether	8.238	63	259683	264.902	ug/l		100
59) 2-Hexanone	9.427	43	413888	259.607	ug/l		99
60) Dibromochloromethane	9.519	129	72717	53.270	ug/l		99
61) 1,2-Dibromoethane	9.604	107	64982	51.266	ug/l		99
64) Tetrachloroethene	9.269	164	56640	51.060	ug/l		98
65) Chlorobenzene	10.073	112	167704	49.952	ug/l		99
66) 1,1,1,2-Tetrachloroethane	10.159	131	58489	50.179	ug/l		100
67) Ethyl Benzene	10.189	91	306928	51.044	ug/l		99
68) m/p-Xylenes	10.299	106	225755	103.227	ug/l		100
69) o-Xylene	10.640	106	110931	51.491	ug/l		100
70) Styrene	10.653	104	188342	52.968	ug/l		100
71) Bromoform	10.799	173	46614	53.560	ug/l	#	97
73) Isopropylbenzene	10.957	105	297453	50.459	ug/l		100
74) N-amyl acetate	10.841	43	141548	50.197	ug/l		100
75) 1,1,2,2-Tetrachloroethane	11.207	83	94897	45.826	ug/l		99
76) 1,2,3-Trichloropropane	11.238	75	84561m	47.106	ug/l		
77) Bromobenzene	11.195	156	67447	49.519	ug/l		99
78) n-propylbenzene	11.299	91	352421	51.903	ug/l		100
79) 2-Chlorotoluene	11.360	91	214755	49.474	ug/l		99
80) 1,3,5-Trimethylbenzene	11.451	105	251615	51.617	ug/l		100
81) trans-1,4-Dichloro-2-b...	11.018	75	27832	52.619	ug/l		100
82) 4-Chlorotoluene	11.451	91	245372	50.699	ug/l		99
83) tert-Butylbenzene	11.713	119	244934	50.753	ug/l		99
84) 1,2,4-Trimethylbenzene	11.750	105	253240	51.760	ug/l		100
85) sec-Butylbenzene	11.890	105	313700	52.923	ug/l		100
86) p-Isopropyltoluene	12.006	119	258241	52.844	ug/l		100
87) 1,3-Dichlorobenzene	11.969	146	123697	49.891	ug/l		99
88) 1,4-Dichlorobenzene	12.036	146	121216	48.225	ug/l		98
89) n-Butylbenzene	12.329	91	237301	55.998	ug/l		99
90) Hexachloroethane	12.536	117	45082	53.679	ug/l		100
91) 1,2-Dichlorobenzene	12.335	146	121280	49.201	ug/l		99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	22906	52.891	ug/l		99
93) 1,2,4-Trichlorobenzene	13.585	180	72928	53.151	ug/l		97
94) Hexachlorobutadiene	13.719	225	32018	54.153	ug/l		97
95) Naphthalene	13.774	128	258746	50.674	ug/l		99
96) 1,2,3-Trichlorobenzene	13.963	180	74671	51.972	ug/l		98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040725\
Data File : VX045614.D
Acq On : 07 Apr 2025 09:17
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Apr 08 01:33:45 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations APPROVED

Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040725\
Data File : VX045614.D
Acq On : 07 Apr 2025 09:17
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1


Quant Time: Apr 08 01:33:45 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations APPROVED

Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025






#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.538 min Scan# 7
Delta R.T. -0.012 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion:168 Resp: 10424


Ion	Ratio	Lower	Upper
168	100		
99	69.9	52.3	78.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#2
Dichlorodifluoromethane
Concen: 55.540 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion: 85 Resp: 86584

Ion	Ratio	Lower	Upper
85	100		
87	32.4	16.1	48.2






#3
Chloromethane
Concen: 49.877 ug/l
RT: 1.307 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion: 50 Resp: 7988


Ion	Ratio	Lower	Upper
50	100		
52	32.1	25.9	38.9



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#4
Vinyl Chloride
Concen: 50.256 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion: 62 Resp: 73660

Ion	Ratio	Lower	Upper
62	100		
64	31.4	25.5	38.3






#5
Bromomethane
Concen: 46.980 ug/l
RT: 1.593 min Scan# 8
Delta R.T. 0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion: 94 Resp: 3264


Ion	Ratio	Lower	Upper
94	100		
96	92.1	74.8	112.2



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#6
Chloroethane
Concen: 50.478 ug/l
RT: 1.660 min Scan# 94
Delta R.T. -0.006 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion: 64 Resp: 39253

Ion	Ratio	Lower	Upper
64	100		
66	34.0	26.3	39.5





#7
Trichlorofluoromethane
Concen: 50.889 ug/l
RT: 1.867 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17


Tgt Ion:101 Resp: 11122


Ion	Ratio	Lower	Upper
101	100		
103	63.9	51.1	76.7

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025







#8
Diethyl Ether
Concen: 49.111 ug/l
RT: 2.130 min Scan# 171
Delta R.T. -0.006 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion: 74 Resp: 36037

Ion	Ratio	Lower	Upper
74	100		
45	104.6	53.0	159.1





#9
1,1,2-Trichlorotrifluoroethane
Concen: 52.076 ug/l
RT: 2.319 min Scan# 2
Delta R.T. -0.006 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17


Tgt Ion:101 Resp: 6669


Ion	Ratio	Lower	Upper
101	100		
85	47.4	38.9	58.3
151	70.5	57.5	86.3

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025







#10
Methyl Iodide
Concen: 49.948 ug/l
RT: 2.441 min Scan# 222
Delta R.T. -0.006 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion:142 Resp: 79791

Ion	Ratio	Lower	Upper
142	100		
127	52.4	40.2	60.4
141	14.0	11.7	17.5





#11
Tert butyl alcohol
Concen: 251.921 ug/l
RT: 2.977 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17


Tgt Ion: 59 Resp: 6454


Ion	Ratio	Lower	Upper
59	100		
57	11.0	8.6	12.8

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025







#12
1,1-Dichloroethene
Concen: 48.756 ug/l
RT: 2.306 min Scan# 200
Delta R.T. -0.006 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion: 96 Resp: 61015

Ion	Ratio	Lower	Upper
96	100		
61	180.7	145.7	218.5
98	65.4	51.6	77.4






#13
Acrolein
Concen: 165.921 ug/l
RT: 2.233 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion: 56 Resp: 5853


Ion	Ratio	Lower	Upper
56	100		
55	70.7	56.2	84.2



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#14
Allyl chloride
Concen: 50.076 ug/l
RT: 2.654 min Scan# 257
Delta R.T. -0.006 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion: 41 Resp: 118904

Ion	Ratio	Lower	Upper
41	100		
39	75.8	60.2	90.2
76	31.5	24.6	36.8






#15
Acrylonitrile
Concen: 242.982 ug/l
RT: 3.062 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion: 53 Resp: 19650


Ion	Ratio	Lower	Upper
53	100		
52	83.2	67.4	101.2
51	37.6	29.9	44.9



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#16
Acetone
Concen: 300.693 ug/l
RT: 2.380 min Scan# 212
Delta R.T. -0.006 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion: 43 Resp: 235381

Ion	Ratio	Lower	Upper
43	100		
58	26.9	22.3	33.5





#17
Carbon Disulfide
Concen: 49.655 ug/l
RT: 2.502 min Scan# 2
Delta R.T. -0.006 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17


Tgt Ion: 76 Resp: 15347


Ion	Ratio	Lower	Upper
76	100		
78	8.8	7.0	10.6

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025







#18
Methyl Acetate
Concen: 50.309 ug/l
RT: 2.703 min Scan# 265
Delta R.T. -0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion: 43 Resp: 90670

Ion	Ratio	Lower	Upper
43	100		
74	22.0	17.5	26.3






#19
Methyl tert-butyl Ether
Concen: 49.382 ug/l
RT: 3.111 min Scan# 3
Delta R.T. -0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion: 73 Resp: 21618


Ion	Ratio	Lower	Upper
73	100		
57	21.6	17.6	26.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#20
Methylene Chloride
Concen: 47.720 ug/l
RT: 2.782 min Scan# 278
Delta R.T. -0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion: 84 Resp: 69933

Ion	Ratio	Lower	Upper
84	100		
49	142.6	107.5	161.3
51	42.2	33.0	49.6
86	64.7	50.1	75.1





#21
trans-1,2-Dichloroethene
Concen: 48.187 ug/l
RT: 3.081 min Scan# 3
Delta R.T. -0.006 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17


Tgt Ion: 96 Resp: 6160


Ion	Ratio	Lower	Upper
96	100		
61	154.8	121.0	181.6
98	64.7	49.7	74.5

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025







#22
Diisopropyl ether
Concen: 48.452 ug/l
RT: 3.757 min Scan# 438
Delta R.T. -0.006 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion: 45 Resp: 226624

Ion	Ratio	Lower	Upper
45	100		
43	74.8	59.3	88.9
87	25.6	20.4	30.6
59	10.7	9.0	13.6





#23
Vinyl Acetate
Concen: 252.123 ug/l
RT: 3.715 min Scan# 4
Delta R.T. -0.006 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17


Tgt Ion: 43 Resp: 101887


Ion	Ratio	Lower	Upper
43	100		
86	9.8	7.5	11.3

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025







#24
1,1-Dichloroethane
Concen: 47.700 ug/l
RT: 3.599 min Scan# 412
Delta R.T. -0.006 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion: 63 Resp: 126177

Ion	Ratio	Lower	Upper
63	100		
98	6.7	3.4	10.1
100	4.1	2.2	6.6





#25
2-Butanone
Concen: 262.730 ug/l
RT: 4.550 min Scan# 5
Delta R.T. -0.006 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17


Tgt Ion: 43 Resp: 30104


Ion	Ratio	Lower	Upper
43	100		
72	24.2	18.6	28.0

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025







#26
2,2-Dichloropropane
Concen: 57.826 ug/l
RT: 4.458 min Scan# 553
Delta R.T. -0.012 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion: 77 Resp: 98828

Ion	Ratio	Lower	Upper
77	100		
97	21.5	10.6	31.8





#27
cis-1,2-Dichloroethene
Concen: 47.355 ug/l
RT: 4.477 min Scan# 5
Delta R.T. -0.006 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17


Tgt Ion: 96 Resp: 7376


Ion	Ratio	Lower	Upper
96	100		
61	161.0	0.0	313.2
98	65.6	0.0	126.4

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025







#28
Bromochloromethane
Concen: 46.860 ug/l
RT: 4.885 min Scan# 623
Delta R.T. -0.012 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion: 49 Resp: 59900

Ion	Ratio	Lower	Upper
49	100		
129	1.9	0.0	3.8
130	70.4	54.6	81.8





#29
Tetrahydrofuran
Concen: 236.552 ug/l
RT: 4.995 min Scan# 64
Delta R.T. -0.006 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17


Tgt Ion: 42 Resp: 17573


Ion	Ratio	Lower	Upper
42	100		
72	42.3	33.4	50.0
71	39.5	31.4	47.0

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025







#30
Chloroform
Concen: 48.232 ug/l
RT: 5.080 min Scan# 655
Delta R.T. -0.006 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion: 83 Resp: 131279

Ion	Ratio	Lower	Upper
83	100		
85	65.1	52.8	79.2






#31
Cyclohexane
Concen: 48.132 ug/l
RT: 5.458 min Scan# 7
Delta R.T. -0.006 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion: 56 Resp: 11255


Ion	Ratio	Lower	Upper
56	100		
69	30.4	23.6	35.4
84	79.2	67.7	101.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#32
1,1,1-Trichloroethane
Concen: 49.377 ug/l
RT: 5.373 min Scan# 703
Delta R.T. -0.006 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion: 97 Resp: 113722

Ion	Ratio	Lower	Upper
97	100		
99	63.8	51.7	77.5
61	51.1	41.0	61.4






#33
1,2-Dichloroethane-d4
Concen: 49.443 ug/l
RT: 5.946 min Scan# 7
Delta R.T. -0.006 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion: 65 Resp: 9425


Ion	Ratio	Lower	Upper
65	100		
67	48.4	0.0	99.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.751 min Scan# 929
Delta R.T. -0.006 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion:114 Resp: 177492

Ion	Ratio	Lower	Upper
114	100		
63	23.7	0.0	46.8
88	16.6	0.0	35.4






#35
Dibromofluoromethane
Concen: 51.592 ug/l
RT: 5.379 min Scan# 7(
Delta R.T. -0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion:113 Resp: 6497(


Ion	Ratio	Lower	Upper
113	100		
111	102.8	81.8	122.6
192	17.2	13.8	20.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#36
1,1-Dichloropropene
Concen: 50.168 ug/l
RT: 5.678 min Scan# 753
Delta R.T. -0.006 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion: 75 Resp: 85301

Ion	Ratio	Lower	Upper
75	100		
110	32.4	16.9	50.6
77	30.7	25.0	37.4






#37
Ethyl Acetate
Concen: 49.894 ug/l
RT: 4.708 min Scan# 5
Delta R.T. -0.006 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion: 43 Resp: 10691


Ion	Ratio	Lower	Upper
43	100		
61	12.1	9.9	14.9
70	9.8	7.5	11.3



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#38
Carbon Tetrachloride
Concen: 51.926 ug/l
RT: 5.666 min Scan# 751
Delta R.T. -0.006 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion:117 Resp: 95420

Ion	Ratio	Lower	Upper
117	100		
119	97.5	76.9	115.3
121	30.3	24.2	36.2






#39
Methylcyclohexane
Concen: 53.483 ug/l
RT: 7.373 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion: 83 Resp: 11147


Ion	Ratio	Lower	Upper
83	100		
55	82.3	64.9	97.3
98	48.6	37.4	56.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#40
Benzene
Concen: 49.081 ug/l
RT: 6.025 min Scan# 810
Delta R.T. -0.006 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion: 78 Resp: 254875

Ion	Ratio	Lower	Upper
78	100		
77	23.0	19.0	28.6






#41
Methacrylonitrile
Concen: 52.399 ug/l
RT: 4.910 min Scan# 6
Delta R.T. -0.012 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion: 41 Resp: 5924


Ion	Ratio	Lower	Upper
41	100		
39	60.2	47.0	70.6
67	67.2	53.2	79.8
52	32.2	25.0	37.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#42
1,2-Dichloroethane
Concen: 50.048 ug/l
RT: 6.080 min Scan# 819
Delta R.T. 0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion: 62 Resp: 107343

Ion	Ratio	Lower	Upper
62	100		
98	7.6	0.0	15.2





#43
Isopropyl Acetate
Concen: 51.145 ug/l
RT: 6.330 min Scan# 8
Delta R.T. -0.006 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17


Tgt Ion: 43 Resp: 16610


Ion	Ratio	Lower	Upper
43	100		
61	17.8	14.1	21.1
87	11.8	9.2	13.8

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025







#44
Trichloroethene
Concen: 49.827 ug/l
RT: 7.117 min Scan# 989
Delta R.T. -0.006 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion:130 Resp: 61500

Ion	Ratio	Lower	Upper
130	100		
95	103.0	0.0	202.0






#45
1,2-Dichloropropane
Concen: 49.683 ug/l
RT: 7.421 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion: 63 Resp: 6437


Ion	Ratio	Lower	Upper
63	100		
65	31.5	25.8	38.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#46
Dibromomethane
Concen: 50.186 ug/l
RT: 7.574 min Scan# 1064
Delta R.T. 0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion: 93 Resp: 49916

Ion	Ratio	Lower	Upper
93	100		
95	82.0	64.8	97.2
174	84.8	65.4	98.0






#47
Bromodichloromethane
Concen: 51.949 ug/l
RT: 7.812 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion: 83 Resp: 10318


Ion	Ratio	Lower	Upper
83	100		
85	63.3	49.7	74.5
127	8.2	6.1	9.1



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#48
Methyl methacrylate
Concen: 51.984 ug/l
RT: 7.690 min Scan# 1083
Delta R.T. 0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion: 41 Resp: 87167

Ion	Ratio	Lower	Upper
41	100		
69	72.5	58.2	87.2
39	65.7	53.6	80.4






#49
1,4-Dioxane
Concen: 1090.518 ug/l
RT: 7.653 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion: 88 Resp: 3308


Ion	Ratio	Lower	Upper
88	100		
43	41.2	32.2	48.4
58	72.5	58.6	88.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#50
Toluene-d8
Concen: 50.658 ug/l
RT: 8.641 min Scan# 1239
Delta R.T. -0.006 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion: 98 Resp: 222665

Ion	Ratio	Lower	Upper
98	100		
100	65.2	52.2	78.4






#51
4-Methyl-2-Pentanone
Concen: 257.706 ug/l
RT: 8.568 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion: 43 Resp: 55481


Ion	Ratio	Lower	Upper
43	100		
58	35.7	28.5	42.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#52
Toluene
Concen: 49.789 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. -0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion: 92 Resp: 156451

Ion	Ratio	Lower	Upper
92	100		
91	171.0	138.7	208.1






#53
t-1,3-Dichloropropene
Concen: 48.743 ug/l
RT: 8.976 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion: 75 Resp: 9299


Ion	Ratio	Lower	Upper
75	100		
77	32.2	25.4	38.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#54
cis-1,3-Dichloropropene
Concen: 54.992 ug/l
RT: 8.360 min Scan# 1193
Delta R.T. -0.006 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion: 75 Resp: 102648

Ion	Ratio	Lower	Upper
75	100		
77	30.1	24.6	36.8
39	62.1	48.4	72.6






#55
1,1,2-Trichloroethane
Concen: 49.526 ug/l
RT: 9.147 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion: 97 Resp: 6201


Ion	Ratio	Lower	Upper
97	100		
83	85.6	71.8	107.8
85	58.2	47.0	70.6
99	61.9	48.6	73.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#56
Ethyl methacrylate
Concen: 52.794 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. 0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion: 69 Resp: 102307

Ion	Ratio	Lower	Upper
69	100		
41	77.3	61.3	91.9
39	49.2	40.2	60.2






#57
1,3-Dichloropropane
Concen: 50.084 ug/l
RT: 9.305 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion: 76 Resp: 10920


Ion	Ratio	Lower	Upper
76	100		
78	32.4	26.0	39.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#58
2-Chloroethyl Vinyl ether
Concen: 264.902 ug/l
RT: 8.238 min Scan# 1173
Delta R.T. 0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion: 63 Resp: 259683

Ion	Ratio	Lower	Upper
63	100		
106	26.4	21.0	31.6






#59
2-Hexanone
Concen: 259.607 ug/l
RT: 9.427 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion: 43 Resp: 41388


Ion	Ratio	Lower	Upper
43	100		
58	49.3	24.9	74.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#60
Dibromochloromethane
Concen: 53.270 ug/l
RT: 9.519 min Scan# 1383
Delta R.T. -0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion:129 Resp: 72717

Ion	Ratio	Lower	Upper
129	100		
127	78.3	39.4	118.2






#61
1,2-Dibromoethane
Concen: 51.266 ug/l
RT: 9.604 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion:107 Resp: 6498


Ion	Ratio	Lower	Upper
107	100		
109	93.4	75.7	113.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#62
4-Bromofluorobenzene
Concen: 53.211 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion: 95 Resp: 85194

Ion	Ratio	Lower	Upper
95	100		
174	67.2	0.0	135.8
176	65.9	0.0	131.4






#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion:117 Resp: 15711(


Ion	Ratio	Lower	Upper
117	100		
82	61.6	49.2	73.8
119	31.7	25.1	37.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#64
Tetrachloroethene
Concen: 51.060 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. -0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion:164 Resp: 56640

Ion	Ratio	Lower	Upper
164	100		
166	125.9	103.8	155.8
129	104.2	84.3	126.5
131	98.7	80.9	121.3






#65
Chlorobenzene
Concen: 49.952 ug/l
RT: 10.073 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion:112 Resp: 16770


Ion	Ratio	Lower	Upper
112	100		
114	31.5	24.7	37.1



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#66
1,1,1,2-Tetrachloroethane
Concen: 50.179 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. 0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion:131 Resp: 58489

Ion	Ratio	Lower	Upper
131	100		
133	95.5	47.5	142.5
119	65.6	32.9	98.6





#67
Ethyl Benzene
Concen: 51.044 ug/l
RT: 10.189 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17


Tgt Ion: 91 Resp: 30692


Ion	Ratio	Lower	Upper
91	100		
106	30.2	23.7	35.5

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025







#68
m/p-Xylenes
Concen: 103.227 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. -0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion:106 Resp: 225755

Ion	Ratio	Lower	Upper
106	100		
91	215.7	172.6	258.8






#69
o-Xylene
Concen: 51.491 ug/l
RT: 10.640 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion:106 Resp: 11093


Ion	Ratio	Lower	Upper
106	100		
91	227.0	113.6	340.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#70
Styrene
Concen: 52.968 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. 0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion:104 Resp: 188342

Ion	Ratio	Lower	Upper
104	100		
78	57.4	45.4	68.2
103	54.7	43.8	65.8






#71
Bromoform
Concen: 53.560 ug/l
RT: 10.799 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17


Tgt Ion:173 Resp: 4661

Ion	Ratio	Lower	Upper
173	100		
175	48.7	23.4	70.2
254	0.0	0.0	0.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Manual Integrations APPROVED
Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion:152 Resp: 73594

Ion	Ratio	Lower	Upper
152	100		
115	95.4	46.9	140.7
150	174.3	0.0	349.4






#73
Isopropylbenzene
Concen: 50.459 ug/l
RT: 10.957 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion:105 Resp: 29745


Ion	Ratio	Lower	Upper
105	100		
120	25.2	12.7	38.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#74
N-amyl acetate
Concen: 50.197 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion: 43 Resp: 141548

Ion	Ratio	Lower	Upper
43	100		
70	37.5	30.1	45.1
55	23.2	18.3	27.5
61	22.0	17.5	26.3






#75
1,1,2,2-Tetrachloroethane
Concen: 45.826 ug/l
RT: 11.207 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion: 83 Resp: 9489


Ion	Ratio	Lower	Upper
83	100		
131	10.0	4.7	14.1
85	64.2	32.4	97.2



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#76
1,2,3-Trichloropropane
Concen: 47.106 ug/l m
RT: 11.238 min Scan# 1665
Delta R.T. -0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion: 75 Resp: 84561

Ion	Ratio	Lower	Upper
75	100		
77	41.6	20.0	59.9






#77
Bromobenzene
Concen: 49.519 ug/l
RT: 11.195 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion:156 Resp: 6744


Ion	Ratio	Lower	Upper
156	100		
77	183.2	92.6	277.8
158	96.8	48.1	144.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#78
n-propylbenzene
Concen: 51.903 ug/l
RT: 11.299 min Scan# 1675
Delta R.T. -0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion: 91 Resp: 352421

Ion	Ratio	Lower	Upper
91	100		
120	21.4	10.8	32.3






#79
2-Chlorotoluene
Concen: 49.474 ug/l
RT: 11.360 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion: 91 Resp: 21475


Ion	Ratio	Lower	Upper
91	100		
126	31.0	15.8	47.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#80
1,3,5-Trimethylbenzene
Concen: 51.617 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion:105 Resp: 251615

Ion	Ratio	Lower	Upper
105	100		
120	46.1	23.0	69.0





#81
trans-1,4-Dichloro-2-butene
Concen: 52.619 ug/l
RT: 11.018 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17


Tgt Ion: 75 Resp: 2783


Ion	Ratio	Lower	Upper
75	100		
53	126.7	101.8	152.6
89	46.8	37.2	55.8

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025







#82
4-Chlorotoluene
Concen: 50.699 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion: 91 Resp: 245372

Ion	Ratio	Lower	Upper
91	100		
126	27.0	13.3	39.9






#83
tert-Butylbenzene
Concen: 50.753 ug/l
RT: 11.713 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17


Tgt Ion:119 Resp: 24493

Ion	Ratio	Lower	Upper
119	100		
91	65.4	32.4	97.2
134	22.9	11.5	34.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Manual Integrations APPROVED
Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#84
1,2,4-Trimethylbenzene
Concen: 51.760 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion:105 Resp: 253240

Ion	Ratio	Lower	Upper
105	100		
120	42.8	21.4	64.3






#85
sec-Butylbenzene
Concen: 52.923 ug/l
RT: 11.890 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion:105 Resp: 31370


Ion	Ratio	Lower	Upper
105	100		
134	19.2	9.6	28.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#86
p-Isopropyltoluene
Concen: 52.844 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. 0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion:119 Resp: 258241

Ion	Ratio	Lower	Upper
119	100		
134	25.3	12.6	37.8
91	27.7	13.7	41.1






#87
1,3-Dichlorobenzene
Concen: 49.891 ug/l
RT: 11.969 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion:146 Resp: 12369


Ion	Ratio	Lower	Upper
146	100		
111	44.2	22.2	66.6
148	62.4	31.9	95.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#88
1,4-Dichlorobenzene
Concen: 48.225 ug/l
RT: 12.036 min Scan# 1796
Delta R.T. 0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion:146 Resp: 121216

Ion	Ratio	Lower	Upper
146	100		
111	44.1	20.9	62.8
148	63.6	31.5	94.5





#89
n-Butylbenzene
Concen: 55.998 ug/l
RT: 12.329 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17


Tgt Ion: 91 Resp: 23730


Ion	Ratio	Lower	Upper
91	100		
92	53.2	26.9	80.6
134	23.7	11.9	35.7

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025







#90
Hexachloroethane
Concen: 53.679 ug/l
RT: 12.536 min Scan# 1878
Delta R.T. -0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion:117 Resp: 45082

Ion	Ratio	Lower	Upper
117	100		
201	62.1	31.2	93.6






#91
1,2-Dichlorobenzene
Concen: 49.201 ug/l
RT: 12.335 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion:146 Resp: 12128


Ion	Ratio	Lower	Upper
146	100		
111	45.3	22.9	68.8
148	62.6	31.9	95.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#92
1,2-Dibromo-3-Chloropropane
Concen: 52.891 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. -0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion: 75 Resp: 22906

Ion	Ratio	Lower	Upper
75	100		
155	70.6	34.9	104.7
157	90.7	44.9	134.7






#93
1,2,4-Trichlorobenzene
Concen: 53.151 ug/l
RT: 13.585 min Scan# 2
Delta R.T. 0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion:180 Resp: 7292


Ion	Ratio	Lower	Upper
180	100		
182	93.4	47.9	143.6
145	33.9	17.9	53.8



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#94
Hexachlorobutadiene
Concen: 54.153 ug/l
RT: 13.719 min Scan# 2072
Delta R.T. -0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion:225 Resp: 32018

Ion	Ratio	Lower	Upper
225	100		
223	64.2	31.1	93.3
227	62.0	32.2	96.6






#95
Naphthalene
Concen: 50.674 ug/l
RT: 13.774 min Scan# 2
Delta R.T. -0.000 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion:128 Resp: 25874


Ion	Ratio	Lower	Upper
128	100		
127	12.7	10.3	15.5
129	11.0	8.6	12.8



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025



#96
1,2,3-Trichlorobenzene
Concen: 51.972 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. 0.006 min
Lab File: VX045614.D
Acq: 07 Apr 2025 09:17

Tgt Ion:180 Resp: 74671

Ion	Ratio	Lower	Upper
180	100		
182	93.0	47.3	141.8
145	35.4	18.2	54.6



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040725\
Data File : VX045614.D
Acq On : 07 Apr 2025 09:17
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Apr 08 01:33:45 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1	I	Pentafluorobenzene	1.000	1.000	0.0	109	-0.01
2	T	Dichlorodifluoromethane	0.748	0.831	-11.1	110	0.00
3	P	Chloromethane	0.768	0.766	0.3	106	0.00
4	C	Vinyl Chloride	0.703	0.707	-0.6#	108	0.00
5	T	Bromomethane	0.333	0.313	6.0	103	0.00
6	T	Chloroethane	0.373	0.377	-1.1	103	0.00
7	T	Trichlorofluoromethane	1.048	1.067	-1.8	107	0.00
8	T	Diethyl Ether	0.352	0.346	1.7	106	0.00
9	T	1,1,2-Trichlorotrifluoroeth	0.614	0.640	-4.2	112	0.00
10	T	Methyl Iodide	0.766	0.765	0.1	104	0.00
11	T	Tert butyl alcohol	0.123	0.124	-0.8	106	0.00
12	CM	1,1-Dichloroethene	0.600	0.585	2.5#	106	0.00
13	T	Acrolein	0.169	0.112	33.7#	73	0.00
14	T	Allyl chloride	1.139	1.141	-0.2	106	0.00
15	T	Acrylonitrile	0.388	0.377	2.8	101	0.00
16	T	Acetone	0.375	0.452	-20.5	130	0.00
17	T	Carbon Disulfide	1.483	1.472	0.7	103	0.00
18	T	Methyl Acetate	0.864	0.870	-0.7	111	0.00
19	T	Methyl tert-butyl Ether	2.100	2.074	1.2	107	0.00
20	T	Methylene Chloride	0.703	0.671	4.6	103	0.00
21	T	trans-1,2-Dichloroethene	0.613	0.591	3.6	103	0.00
22	T	Diisopropyl ether	2.243	2.174	3.1	103	0.00
23	T	Vinyl Acetate	1.938	1.955	-0.9	102	0.00
24	P	1,1-Dichloroethane	1.269	1.210	4.6	104	0.00
25	T	2-Butanone	0.550	0.578	-5.1	107	0.00
26	T	2,2-Dichloropropane	0.820	0.948	-15.6	122	-0.01
27	T	cis-1,2-Dichloroethene	0.747	0.708	5.2	103	0.00
28	T	Bromochloromethane	0.613	0.575	6.2	105	-0.01
29	T	Tetrahydrofuran	0.356	0.337	5.3	99	0.00
30	C	Chloroform	1.306	1.259	3.6#	104	0.00
31	T	Cyclohexane	1.122	1.080	3.7	105	0.00
32	T	1,1,1-Trichloroethane	1.105	1.091	1.3	107	0.00
33	S	1,2-Dichloroethane-d4	0.914	0.904	1.1	113	0.00
34	I	1,4-Difluorobenzene	1.000	1.000	0.0	105	0.00
35	S	Dibromofluoromethane	0.355	0.366	-3.1	112	0.00
36	T	1,1-Dichloropropene	0.479	0.481	-0.4	105	0.00
37	T	Ethyl Acetate	0.604	0.602	0.3	102	0.00
38	T	Carbon Tetrachloride	0.518	0.538	-3.9	106	0.00
39	T	Methylcyclohexane	0.587	0.628	-7.0	106	0.00
40	TM	Benzene	1.463	1.436	1.8	102	0.00
41	T	Methacrylonitrile	0.318	0.334	-5.0	100	-0.01
42	TM	1,2-Dichloroethane	0.604	0.605	-0.2	103	0.00
43	T	Isopropyl Acetate	0.915	0.936	-2.3	101	0.00
44	TM	Trichloroethene	0.348	0.346	0.6	104	0.00
45	C	1,2-Dichloropropane	0.365	0.363	0.5#	102	0.00
46	T	Dibromomethane	0.280	0.281	-0.4	102	0.00
47	T	Bromodichloromethane	0.560	0.581	-3.7	107	0.00
48	T	Methyl methacrylate	0.472	0.491	-4.0	102	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040725\
Data File : VX045614.D
Acq On : 07 Apr 2025 09:17
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Apr 08 01:33:45 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49	T	1,4-Dioxane	0.009	0.009	0.0	106	0.00
50	S	Toluene-d8	1.238	1.255	-1.4	109	0.00
51	T	4-Methyl-2-Pentanone	0.606	0.625	-3.1	99	0.00
52	CM	Toluene	0.885	0.881	0.5#	102	0.00
53	T	t-1,3-Dichloropropene	0.467	0.524	-12.2	109	0.00
54	T	cis-1,3-Dichloropropene	0.526	0.578	-9.9	107	0.00
55	T	1,1,2-Trichloroethane	0.353	0.349	1.1	103	0.00
56	T	Ethyl methacrylate	0.546	0.576	-5.5	101	0.00
57	T	1,3-Dichloropropane	0.614	0.615	-0.2	102	0.00
58	T	2-Chloroethyl Vinyl ether	0.276	0.293	-6.2	104	0.00
59	T	2-Hexanone	0.449	0.466	-3.8	100	0.00
60	T	Dibromochloromethane	0.385	0.410	-6.5	105	0.00
61	T	1,2-Dibromoethane	0.357	0.366	-2.5	103	0.00
62	S	4-Bromofluorobenzene	0.451	0.480	-6.4	112	0.00
63	I	Chlorobenzene-d5	1.000	1.000	0.0	105	0.00
64	T	Tetrachloroethene	0.353	0.360	-2.0	109	0.00
65	PM	Chlorobenzene	1.068	1.067	0.1	103	0.00
66	T	1,1,1,2-Tetrachloroethane	0.371	0.372	-0.3	104	0.00
67	C	Ethyl Benzene	1.914	1.954	-2.1#	102	0.00
68	T	m/p-Xylenes	0.696	0.718	-3.2	103	0.00
69	T	o-Xylene	0.686	0.706	-2.9	103	0.00
70	T	Styrene	1.132	1.199	-5.9	103	0.00
71	P	Bromoform	0.277	0.297	-7.2	105	0.00
72	I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	107	0.00
73	T	Isopropylbenzene	4.005	4.042	-0.9	103	0.00
74	T	N-amyl acetate	1.916	1.923	-0.4	100	0.00
75	P	1,1,2,2-Tetrachloroethane	1.407	1.289	8.4	100	0.00
76	T	1,2,3-Trichloropropane	1.220	1.149	5.8	100	0.00
77	T	Bromobenzene	0.925	0.916	1.0	105	0.00
78	T	n-propylbenzene	4.613	4.789	-3.8	105	0.00
79	T	2-Chlorotoluene	2.949	2.918	1.1	105	0.00
80	T	1,3,5-Trimethylbenzene	3.312	3.419	-3.2	104	0.00
81	T	trans-1,4-Dichloro-2-butene	0.359	0.378	-5.3	109	0.00
82	T	4-Chlorotoluene	3.288	3.334	-1.4	103	0.00
83	T	tert-Butylbenzene	3.279	3.328	-1.5	104	0.00
84	T	1,2,4-Trimethylbenzene	3.324	3.441	-3.5	105	0.00
85	T	sec-Butylbenzene	4.027	4.263	-5.9	106	0.00
86	T	p-Isopropyltoluene	3.320	3.509	-5.7	107	0.00
87	T	1,3-Dichlorobenzene	1.684	1.681	0.2	106	0.00
88	T	1,4-Dichlorobenzene	1.708	1.647	3.6	103	0.00
89	T	n-Butylbenzene	2.879	3.224	-12.0	109	0.00
90	T	Hexachloroethane	0.571	0.613	-7.4	110	0.00
91	T	1,2-Dichlorobenzene	1.675	1.648	1.6	105	0.00
92	T	1,2-Dibromo-3-Chloropropane	0.294	0.311	-5.8	105	0.00
93	T	1,2,4-Trichlorobenzene	0.932	0.991	-6.3	112	0.00
94	T	Hexachlorobutadiene	0.402	0.435	-8.2	115	0.00
95	T	Naphthalene	3.469	3.516	-1.4	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040725\
Data File : VX045614.D
Acq On : 07 Apr 2025 09:17
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Apr 08 01:33:45 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.976	1.015	-4.0	109	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040725\
Data File : VX045614.D
Acq On : 07 Apr 2025 09:17
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Apr 08 01:33:45 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	Pentafluorobenzene	50.000	50.000	0.0	109	-0.01
2	T	Dichlorodifluoromethane	50.000	55.540	-11.1	110	0.00
3	P	Chloromethane	50.000	49.877	0.2	106	0.00
4	C	Vinyl Chloride	50.000	50.256	-0.5#	108	0.00
5	T	Bromomethane	50.000	46.980	6.0	103	0.00
6	T	Chloroethane	50.000	50.478	-1.0	103	0.00
7	T	Trichlorofluoromethane	50.000	50.889	-1.8	107	0.00
8	T	Diethyl Ether	50.000	49.111	1.8	106	0.00
9	T	1,1,2-Trichlorotrifluoroeth	50.000	52.076	-4.2	112	0.00
10	T	Methyl Iodide	50.000	49.948	0.1	104	0.00
11	T	Tert butyl alcohol	250.000	251.921	-0.8	106	0.00
12	CM	1,1-Dichloroethene	50.000	48.756	2.5#	106	0.00
13	T	Acrolein	250.000	165.921	33.6#	73	0.00
14	T	Allyl chloride	50.000	50.076	-0.2	106	0.00
15	T	Acrylonitrile	250.000	242.982	2.8	101	0.00
16	T	Acetone	250.000	300.693	-20.3	130	0.00
17	T	Carbon Disulfide	50.000	49.655	0.7	103	0.00
18	T	Methyl Acetate	50.000	50.309	-0.6	111	0.00
19	T	Methyl tert-butyl Ether	50.000	49.382	1.2	107	0.00
20	T	Methylene Chloride	50.000	47.720	4.6	103	0.00
21	T	trans-1,2-Dichloroethene	50.000	48.187	3.6	103	0.00
22	T	Diisopropyl ether	50.000	48.452	3.1	103	0.00
23	T	Vinyl Acetate	250.000	252.123	-0.8	102	0.00
24	P	1,1-Dichloroethane	50.000	47.700	4.6	104	0.00
25	T	2-Butanone	250.000	262.730	-5.1	107	0.00
26	T	2,2-Dichloropropane	50.000	57.826	-15.7	122	-0.01
27	T	cis-1,2-Dichloroethene	50.000	47.355	5.3	103	0.00
28	T	Bromochloromethane	50.000	46.860	6.3	105	-0.01
29	T	Tetrahydrofuran	250.000	236.552	5.4	99	0.00
30	C	Chloroform	50.000	48.232	3.5#	104	0.00
31	T	Cyclohexane	50.000	48.132	3.7	105	0.00
32	T	1,1,1-Trichloroethane	50.000	49.377	1.2	107	0.00
33	S	1,2-Dichloroethane-d4	50.000	49.443	1.1	113	0.00
34	I	1,4-Difluorobenzene	50.000	50.000	0.0	105	0.00
35	S	Dibromofluoromethane	50.000	51.592	-3.2	112	0.00
36	T	1,1-Dichloropropene	50.000	50.168	-0.3	105	0.00
37	T	Ethyl Acetate	50.000	49.894	0.2	102	0.00
38	T	Carbon Tetrachloride	50.000	51.926	-3.9	106	0.00
39	T	Methylcyclohexane	50.000	53.483	-7.0	106	0.00
40	TM	Benzene	50.000	49.081	1.8	102	0.00
41	T	Methacrylonitrile	50.000	52.399	-4.8	100	-0.01
42	TM	1,2-Dichloroethane	50.000	50.048	-0.1	103	0.00
43	T	Isopropyl Acetate	50.000	51.145	-2.3	101	0.00
44	TM	Trichloroethene	50.000	49.827	0.3	104	0.00
45	C	1,2-Dichloropropane	50.000	49.683	0.6#	102	0.00
46	T	Dibromomethane	50.000	50.186	-0.4	102	0.00
47	T	Bromodichloromethane	50.000	51.949	-3.9	107	0.00
48	T	Methyl methacrylate	50.000	51.984	-4.0	102	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040725\
Data File : VX045614.D
Acq On : 07 Apr 2025 09:17
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Apr 08 01:33:45 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49	T	1,4-Dioxane	1000.000	1090.518	-9.1	106	0.00
50	S	Toluene-d8	50.000	50.658	-1.3	109	0.00
51	T	4-Methyl-2-Pentanone	250.000	257.706	-3.1	99	0.00
52	CM	Toluene	50.000	49.789	0.4#	102	0.00
53	T	t-1,3-Dichloropropene	50.000	48.743	2.5	109	0.00
54	T	cis-1,3-Dichloropropene	50.000	54.992	-10.0	107	0.00
55	T	1,1,2-Trichloroethane	50.000	49.526	0.9	103	0.00
56	T	Ethyl methacrylate	50.000	52.794	-5.6	101	0.00
57	T	1,3-Dichloropropane	50.000	50.084	-0.2	102	0.00
58	T	2-Chloroethyl Vinyl ether	250.000	264.902	-6.0	104	0.00
59	T	2-Hexanone	250.000	259.607	-3.8	100	0.00
60	T	Dibromochloromethane	50.000	53.270	-6.5	105	0.00
61	T	1,2-Dibromoethane	50.000	51.266	-2.5	103	0.00
62	S	4-Bromofluorobenzene	50.000	53.211	-6.4	112	0.00
63	I	Chlorobenzene-d5	50.000	50.000	0.0	105	0.00
64	T	Tetrachloroethene	50.000	51.060	-2.1	109	0.00
65	PM	Chlorobenzene	50.000	49.952	0.1	103	0.00
66	T	1,1,1,2-Tetrachloroethane	50.000	50.179	-0.4	104	0.00
67	C	Ethyl Benzene	50.000	51.044	-2.1#	102	0.00
68	T	m/p-Xylenes	100.000	103.227	-3.2	103	0.00
69	T	o-Xylene	50.000	51.491	-3.0	103	0.00
70	T	Styrene	50.000	52.968	-5.9	103	0.00
71	P	Bromoform	50.000	53.560	-7.1	105	0.00
72	I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	107	0.00
73	T	Isopropylbenzene	50.000	50.459	-0.9	103	0.00
74	T	N-amyl acetate	50.000	50.197	-0.4	100	0.00
75	P	1,1,2,2-Tetrachloroethane	50.000	45.826	8.3	100	0.00
76	T	1,2,3-Trichloropropane	50.000	47.106	5.8	100	0.00
77	T	Bromobenzene	50.000	49.519	1.0	105	0.00
78	T	n-propylbenzene	50.000	51.903	-3.8	105	0.00
79	T	2-Chlorotoluene	50.000	49.474	1.1	105	0.00
80	T	1,3,5-Trimethylbenzene	50.000	51.617	-3.2	104	0.00
81	T	trans-1,4-Dichloro-2-butene	50.000	52.619	-5.2	109	0.00
82	T	4-Chlorotoluene	50.000	50.699	-1.4	103	0.00
83	T	tert-Butylbenzene	50.000	50.753	-1.5	104	0.00
84	T	1,2,4-Trimethylbenzene	50.000	51.760	-3.5	105	0.00
85	T	sec-Butylbenzene	50.000	52.923	-5.8	106	0.00
86	T	p-Isopropyltoluene	50.000	52.844	-5.7	107	0.00
87	T	1,3-Dichlorobenzene	50.000	49.891	0.2	106	0.00
88	T	1,4-Dichlorobenzene	50.000	48.225	3.5	103	0.00
89	T	n-Butylbenzene	50.000	55.998	-12.0	109	0.00
90	T	Hexachloroethane	50.000	53.679	-7.4	110	0.00
91	T	1,2-Dichlorobenzene	50.000	49.201	1.6	105	0.00
92	T	1,2-Dibromo-3-Chloropropane	50.000	52.891	-5.8	105	0.00
93	T	1,2,4-Trichlorobenzene	50.000	53.151	-6.3	112	0.00
94	T	Hexachlorobutadiene	50.000	54.153	-8.3	115	0.00
95	T	Naphthalene	50.000	50.674	-1.3	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040725\
Data File : VX045614.D
Acq On : 07 Apr 2025 09:17
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Apr 08 01:33:45 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	51.972	-3.9	109	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6


QC SAMPLE DATA

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045524.D
Acq On : 01 Apr 2025 16:15
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Title : SW846 8260
Last Update : Wed Apr 02 03:11:43 2025



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632


Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	23.6	7184	PASS
75	95	30	60	58.2	17768	PASS
95	95	100	100	100.0	30504	PASS
96	95	5	9	6.3	1916	PASS
173	174	0.00	2	1.2	253	PASS
174	95	50	100	66.8	20365	PASS
175	174	5	9	7.8	1596	PASS
176	174	95	101	97.8	19923	PASS
177	176	5	9	6.8	1352	PASS

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040425\
Data File : VX045585.D
Acq On : 04 Apr 2025 09:31
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Title : SW846 8260
Last Update : Wed Apr 02 03:11:43 2025



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632


Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	25.0	3902	PASS
75	95	30	60	57.7	9007	PASS
95	95	100	100	100.0	15601	PASS
96	95	5	9	7.2	1124	PASS
173	174	0.00	2	1.4	148	PASS
174	95	50	100	68.1	10629	PASS
175	174	5	9	8.3	882	PASS
176	174	95	101	96.6	10272	PASS
177	176	5	9	7.5	768	PASS

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040725\
Data File : VX045613.D
Acq On : 07 Apr 2025 08:34
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Title : SW846 8260
Last Update : Wed Apr 02 03:11:43 2025



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	23.4	7056	PASS
75	95	30	60	58.2	17581	PASS
95	95	100	100	100.0	30192	PASS
96	95	5	9	7.1	2129	PASS
173	174	0.00	2	1.2	253	PASS
174	95	50	100	67.9	20491	PASS
175	174	5	9	8.0	1630	PASS
176	174	95	101	99.2	20331	PASS
177	176	5	9	6.1	1237	PASS

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	
Project:	Former Schlumberger STC PTC Site D3868221	Date Received:	
Client Sample ID:	VX0404WBL01	SDG No.:	Q1731
Lab Sample ID:	VX0404WBL01	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group3
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045588.D	1		04/04/25 10:51	VX040425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	0.26	U	0.26	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.23	U	0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.19	U	0.19	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.20	U	0.20	1.00	ug/L
71-43-2	Benzene	0.15	U	0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.22	U	0.22	1.00	ug/L
79-01-6	Trichloroethene	0.090	U	0.090	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	1.00	ug/L
127-18-4	Tetrachloroethene	0.23	U	0.23	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	54.8		70 (74) - 130 (125)	110%	SPK: 50
1868-53-7	Dibromofluoromethane	52.0		70 (75) - 130 (124)	104%	SPK: 50
2037-26-5	Toluene-d8	50.8		70 (86) - 130 (113)	102%	SPK: 50
460-00-4	4-Bromofluorobenzene	49.5		70 (77) - 130 (121)	99%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	66300	5.543			
540-36-3	1,4-Difluorobenzene	129000	6.757			
3114-55-4	Chlorobenzene-d5	118000	10.055			
3855-82-1	1,4-Dichlorobenzene-d4	46000	12.018			

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements
J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040425\
Data File : VX045588.D
Acq On : 04 Apr 2025 10:51
Operator : JC/MD
Sample : VX0404WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0404WBL01

Quant Time: Apr 04 23:10:30 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	5.543	168	66301	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	128686	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	118182	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	46025	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	66417	54.778	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 109.560%			
35) Dibromofluoromethane	5.379	113	47516	52.042	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 104.080%			
50) Toluene-d8	8.647	98	161989	50.831	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 101.660%			
62) 4-Bromofluorobenzene	11.079	95	57504	49.538	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 99.080%			


Target Compounds Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040425\
Data File : VX045588.D
Acq On : 04 Apr 2025 10:51
Operator : JC/MD
Sample : VX0404WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0404WBL01

Quant Time: Apr 04 23:10:30 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration







#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.543 min Scan# 7
Delta R.T. -0.007 min
Lab File: VX045588.D
Acq: 04 Apr 2025 10:51

Instrument :
MSVOA_X
ClientSampleId :
VX0404WBL01

Tgt Ion:168 Resp: 66301

Ion	Ratio	Lower	Upper
168	100		
99	67.3	52.3	78.5







#33
1,2-Dichloroethane-d4
Concen: 54.778 ug/l
RT: 5.952 min Scan# 798
Delta R.T. -0.000 min
Lab File: VX045588.D
Acq: 04 Apr 2025 10:51

Tgt Ion: 65 Resp: 66417

Ion	Ratio	Lower	Upper
65	100		
67	48.8	0.0	99.0







#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.757 min Scan# 9
Delta R.T. -0.000 min
Lab File: VX045588.D
Acq: 04 Apr 2025 10:51

Instrument :
MSVOA_X
ClientSampleId :
VX0404WBL01

Tgt Ion:114 Resp: 128686

Ion	Ratio	Lower	Upper
114	100		
63	24.5	0.0	46.8
88	17.7	0.0	35.4







#35
Dibromofluoromethane
Concen: 52.042 ug/l
RT: 5.379 min Scan# 704
Delta R.T. -0.000 min
Lab File: VX045588.D
Acq: 04 Apr 2025 10:51

Tgt Ion:113 Resp: 47516

Ion	Ratio	Lower	Upper
113	100		
111	103.9	81.8	122.6
192	17.3	13.8	20.6







#50
Toluene-d8
Concen: 50.831 ug/l
RT: 8.647 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045588.D
Acq: 04 Apr 2025 10:51

Instrument :
MSVOA_X
ClientSampleId :
VX0404WBL01

Tgt Ion: 98 Resp: 161989

Ion	Ratio	Lower	Upper
98	100		
100	64.8	52.2	78.4







#62
4-Bromofluorobenzene
Concen: 49.538 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX045588.D
Acq: 04 Apr 2025 10:51

Tgt Ion: 95 Resp: 57504

Ion	Ratio	Lower	Upper
95	100		
174	68.9	0.0	135.8
176	66.6	0.0	131.4







#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 14
Delta R.T. 0.006 min
Lab File: VX045588.D
Acq: 04 Apr 2025 10:51

Instrument :
MSVOA_X
ClientSampleId :
VX0404WBL01

Tgt Ion:117 Resp: 118182

Ion	Ratio	Lower	Upper
117	100		
82	58.3	49.2	73.8
119	31.9	25.1	37.7






#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VX045588.D
Acq: 04 Apr 2025 10:51

Tgt Ion:152 Resp: 46025

Ion	Ratio	Lower	Upper
152	100		
115	64.7	46.9	140.7
150	155.8	0.0	349.4



Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	
Project:	Former Schlumberger STC PTC Site D3868221	Date Received:	
Client Sample ID:	VX0407WBL01	SDG No.:	Q1731
Lab Sample ID:	VX0407WBL01	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group3
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045616.D	1		04/07/25 10:11	VX040725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	0.26	U	0.26	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.23	U	0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.19	U	0.19	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.20	U	0.20	1.00	ug/L
71-43-2	Benzene	0.15	U	0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.22	U	0.22	1.00	ug/L
79-01-6	Trichloroethene	0.090	U	0.090	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	1.00	ug/L
127-18-4	Tetrachloroethene	0.23	U	0.23	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	56.0		70 (74) - 130 (125)	112%	SPK: 50
1868-53-7	Dibromofluoromethane	52.1		70 (75) - 130 (124)	104%	SPK: 50
2037-26-5	Toluene-d8	51.0		70 (86) - 130 (113)	102%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.6		70 (77) - 130 (121)	103%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	67500	5.544			
540-36-3	1,4-Difluorobenzene	133000	6.757			
3114-55-4	Chlorobenzene-d5	124000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	49900	12.018			

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements
J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040725\
Data File : VX045616.D
Acq On : 07 Apr 2025 10:11
Operator : JC/MD
Sample : VX0407WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0407WBL01

Quant Time: Apr 08 01:34:17 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	5.544	168	67505	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	132975	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	123601	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	49926	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	69079	55.957	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 111.920%			
35) Dibromofluoromethane	5.385	113	49178	52.125	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 104.260%			
50) Toluene-d8	8.647	98	167958	51.004	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 102.000%			
62) 4-Bromofluorobenzene	11.079	95	61925	51.626	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 103.260%			


Target Compounds Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040725\
Data File : VX045616.D
Acq On : 07 Apr 2025 10:11
Operator : JC/MD
Sample : VX0407WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0407WBL01

Quant Time: Apr 08 01:34:17 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration







#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 7
Delta R.T. -0.006 min
Lab File: VX045616.D
Acq: 07 Apr 2025 10:11

Instrument :
MSVOA_X
ClientSampleId :
VX0407WBL01

Tgt Ion:168 Resp: 67505

Ion	Ratio	Lower	Upper
168	100		
99	71.5	52.3	78.5







#33
1,2-Dichloroethane-d4
Concen: 55.957 ug/l
RT: 5.946 min Scan# 797
Delta R.T. -0.006 min
Lab File: VX045616.D
Acq: 07 Apr 2025 10:11

Tgt Ion: 65 Resp: 69079

Ion	Ratio	Lower	Upper
65	100		
67	48.6	0.0	99.0







#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.757 min Scan# 9
Delta R.T. -0.000 min
Lab File: VX045616.D
Acq: 07 Apr 2025 10:11

Instrument :
MSVOA_X
ClientSampleId :
VX0407WBL01

Tgt Ion:114 Resp: 132975

Ion	Ratio	Lower	Upper
114	100		
63	23.9	0.0	46.8
88	16.2	0.0	35.4







#35
Dibromofluoromethane
Concen: 52.125 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.006 min
Lab File: VX045616.D
Acq: 07 Apr 2025 10:11

Tgt Ion:113 Resp: 49178

Ion	Ratio	Lower	Upper
113	100		
111	105.0	81.8	122.6
192	16.4	13.8	20.6







#50
Toluene-d8
Concen: 51.004 ug/l
RT: 8.647 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045616.D
Acq: 07 Apr 2025 10:11

Instrument :
MSVOA_X
ClientSampleId :
VX0407WBL01

Tgt Ion: 98 Resp: 167958

Ion	Ratio	Lower	Upper
98	100		
100	64.8	52.2	78.4







#62
4-Bromofluorobenzene
Concen: 51.626 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX045616.D
Acq: 07 Apr 2025 10:11

Tgt Ion: 95 Resp: 61925

Ion	Ratio	Lower	Upper
95	100		
174	67.3	0.0	135.8
176	65.4	0.0	131.4







#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 14
Delta R.T. -0.000 min
Lab File: VX045616.D
Acq: 07 Apr 2025 10:11

Instrument :
MSVOA_X
ClientSampleId :
VX0407WBL01

Tgt Ion:117 Resp: 123601

Ion	Ratio	Lower	Upper
117	100		
82	63.5	49.2	73.8
119	32.0	25.1	37.7






#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VX045616.D
Acq: 07 Apr 2025 10:11

Tgt Ion:152 Resp: 49926

Ion	Ratio	Lower	Upper
152	100		
115	64.5	46.9	140.7
150	156.2	0.0	349.4



Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	
Project:	Former Schlumberger STC PTC Site D3868221	Date Received:	
Client Sample ID:	VX0404WBS01	SDG No.:	Q1731
Lab Sample ID:	VX0404WBS01	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group3
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045589.D	1		04/04/25 11:14	VX040425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	19.6		0.26	1.00	ug/L
75-35-4	1,1-Dichloroethene	19.4		0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	19.5		0.23	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	19.2		0.19	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	19.7		0.20	1.00	ug/L
71-43-2	Benzene	19.9		0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	20.4		0.22	1.00	ug/L
79-01-6	Trichloroethene	19.2		0.090	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	20.5		0.21	1.00	ug/L
127-18-4	Tetrachloroethene	19.6		0.23	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	52.9		70 (74) - 130 (125)	106%	SPK: 50
1868-53-7	Dibromofluoromethane	53.3		70 (75) - 130 (124)	107%	SPK: 50
2037-26-5	Toluene-d8	53.2		70 (86) - 130 (113)	106%	SPK: 50
460-00-4	4-Bromofluorobenzene	53.2		70 (77) - 130 (121)	106%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	95100	5.549			
540-36-3	1,4-Difluorobenzene	168000	6.757			
3114-55-4	Chlorobenzene-d5	152000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	66600	12.018			

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040425\
Data File : VX045589.D
Acq On : 04 Apr 2025 11:14
Operator : JC/MD
Sample : VX0404WBS01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Quant Time: Apr 04 23:10:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VX0404WBS01

Manual Integrations APPROVED

Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	5.549	168	95127	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	168453	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	151744	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	66584	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	91980	52.873	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 105.740%			
35) Dibromofluoromethane	5.385	113	63679	53.280	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 106.560%			
50) Toluene-d8	8.646	98	221910	53.195	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 106.400%			
62) 4-Bromofluorobenzene	11.079	95	80880	53.227	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 106.460%			
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.166	85	30356	21.338	ug/l	98
3) Chloromethane	1.306	50	28046	19.190	ug/l	97
4) Vinyl Chloride	1.373	62	26269	19.640	ug/l	99
5) Bromomethane	1.593	94	12147	19.154	ug/l	100
6) Chloroethane	1.666	64	14563	20.522	ug/l	98
7) Trichlorofluoromethane	1.873	101	40168	20.139	ug/l	95
8) Diethyl Ether	2.135	74	13256	19.796	ug/l	98
9) 1,1,2-Trichlorotrifluo...	2.318	101	23531	20.134	ug/l	98
10) Methyl Iodide	2.446	142	27472	18.845	ug/l	97
11) Tert butyl alcohol	2.971	59	22553	96.466	ug/l	99
12) 1,1-Dichloroethene	2.312	96	22114	19.364	ug/l	97
13) Acrolein	2.233	56	35297	109.636	ug/l	96
14) Allyl chloride	2.660	41	42970	19.831	ug/l	99
15) Acrylonitrile	3.062	53	73835	100.048	ug/l	99
16) Acetone	2.379	43	71844	100.574	ug/l	95
17) Carbon Disulfide	2.507	76	51789	18.361	ug/l	100
18) Methyl Acetate	2.702	43	33157	20.160	ug/l	99
19) Methyl tert-butyl Ether	3.111	73	77491	19.397	ug/l	98
20) Methylene Chloride	2.782	84	26080	19.502	ug/l	98
21) trans-1,2-Dichloroethene	3.086	96	22921	19.646	ug/l	95
22) Diisopropyl ether	3.757	45	82229	19.265	ug/l #	85
23) Vinyl Acetate	3.721	43	363855	98.665	ug/l	99
24) 1,1-Dichloroethane	3.605	63	46953	19.451	ug/l	99
25) 2-Butanone	4.556	43	106093	101.462	ug/l	99
26) 2,2-Dichloropropane	4.464	77	32865	21.073	ug/l	99
27) cis-1,2-Dichloroethene	4.483	96	27229	19.156	ug/l	99
28) Bromochloromethane	4.891	49	23795	20.399	ug/l	97
29) Tetrahydrofuran	5.007	42	66295	97.791	ug/l	99
30) Chloroform	5.086	83	49173	19.797	ug/l	94
31) Cyclohexane	5.464	56	41598	19.493	ug/l	99
32) 1,1,1-Trichloroethane	5.379	97	41502	19.747	ug/l	98
36) 1,1-Dichloropropene	5.684	75	31963	19.807	ug/l	98
37) Ethyl Acetate	4.714	43	38551	18.957	ug/l	99
38) Carbon Tetrachloride	5.671	117	35237	20.204	ug/l	95
39) Methylcyclohexane	7.372	83	39542	19.990	ug/l	95
40) Benzene	6.031	78	97905	19.865	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040425\
Data File : VX045589.D
Acq On : 04 Apr 2025 11:14
Operator : JC/MD
Sample : VX0404WBS01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Quant Time: Apr 04 23:10:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VX0404WBS01

Manual Integrations APPROVED

Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
41) Methacrylonitrile	4.922	41	22088	20.585	ug/l		98
42) 1,2-Dichloroethane	6.086	62	41452	20.364	ug/l		100
43) Isopropyl Acetate	6.342	43	59314	19.244	ug/l		99
44) Trichloroethene	7.122	130	22535	19.237	ug/l		89
45) 1,2-Dichloropropane	7.427	63	24353	19.804	ug/l		92
46) Dibromomethane	7.580	93	19189	20.328	ug/l		98
47) Bromodichloromethane	7.817	83	37575	19.932	ug/l		96
48) Methyl methacrylate	7.689	41	31993	20.104	ug/l		99
49) 1,4-Dioxane	7.659	88	11949	414.998	ug/l		98
51) 4-Methyl-2-Pentanone	8.573	43	209485	102.524	ug/l		99
52) Toluene	8.714	92	58522	19.623	ug/l		98
53) t-1,3-Dichloropropene	8.976	75	30749	18.474	ug/l		99
54) cis-1,3-Dichloropropene	8.366	75	35158	19.846	ug/l		96
55) 1,1,2-Trichloroethane	9.152	97	24341	20.482	ug/l		95
56) Ethyl methacrylate	9.116	69	36657	19.931	ug/l		98
57) 1,3-Dichloropropane	9.305	76	42216	20.400	ug/l		99
58) 2-Chloroethyl Vinyl ether	8.238	63	95516	102.664	ug/l		99
59) 2-Hexanone	9.427	43	152877	101.036	ug/l		98
60) Dibromochloromethane	9.518	129	25841	19.946	ug/l		99
61) 1,2-Dibromoethane	9.610	107	23976	19.930	ug/l		100
64) Tetrachloroethene	9.274	164	21048	19.646	ug/l		97
65) Chlorobenzene	10.079	112	63546	19.598	ug/l		98
66) 1,1,1,2-Tetrachloroethane	10.158	131	22150	19.676	ug/l		97
67) Ethyl Benzene	10.189	91	113903	19.613	ug/l		98
68) m/p-Xylenes	10.299	106	82893	39.245	ug/l		99
69) o-Xylene	10.640	106	41423	19.908	ug/l		99
70) Styrene	10.652	104	68156	19.846	ug/l		98
71) Bromoform	10.799	173	15776	18.768	ug/l	#	96
73) Isopropylbenzene	10.963	105	109126	20.461	ug/l		100
74) N-amyl acetate	10.841	43	49771	19.508	ug/l		99
75) 1,1,2,2-Tetrachloroethane	11.207	83	37577	20.056	ug/l		96
76) 1,2,3-Trichloropropane	11.237	75	31448m	19.363	ug/l		
77) Bromobenzene	11.195	156	25007	20.293	ug/l		98
78) n-propylbenzene	11.298	91	126822	20.644	ug/l		99
79) 2-Chlorotoluene	11.359	91	78774	20.058	ug/l		99
80) 1,3,5-Trimethylbenzene	11.451	105	91980	20.856	ug/l		99
81) trans-1,4-Dichloro-2-b...	11.018	75	8789	18.366	ug/l		95
82) 4-Chlorotoluene	11.451	91	90453	20.657	ug/l		100
83) tert-Butylbenzene	11.713	119	88794	20.336	ug/l		97
84) 1,2,4-Trimethylbenzene	11.750	105	91248	20.614	ug/l		99
85) sec-Butylbenzene	11.890	105	111812	20.849	ug/l		100
86) p-Isopropyltoluene	12.006	119	90695	20.513	ug/l		99
87) 1,3-Dichlorobenzene	11.969	146	44092	19.656	ug/l		99
88) 1,4-Dichlorobenzene	12.036	146	44331	19.494	ug/l		98
89) n-Butylbenzene	12.329	91	77639	20.250	ug/l		99
90) Hexachloroethane	12.536	117	15371	20.229	ug/l		99
91) 1,2-Dichlorobenzene	12.335	146	44325	19.875	ug/l		99
92) 1,2-Dibromo-3-Chloropr...	12.938	75	7932	20.243	ug/l		97
93) 1,2,4-Trichlorobenzene	13.585	180	23660	19.059	ug/l		97
94) Hexachlorobutadiene	13.725	225	10408	19.457	ug/l		97
95) Naphthalene	13.774	128	87543	18.950	ug/l		100
96) 1,2,3-Trichlorobenzene	13.956	180	24870	19.132	ug/l		97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040425\
Data File : VX045589.D
Acq On : 04 Apr 2025 11:14
Operator : JC/MD
Sample : VX0404WBS01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Quant Time: Apr 04 23:10:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VX0404WBS01

Manual Integrations APPROVED

Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040425\
Data File : VX045589.D
Acq On : 04 Apr 2025 11:14
Operator : JC/MD
Sample : VX0404WBS01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1


Quant Time: Apr 04 23:10:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VX0404WBS01

Manual Integrations APPROVED

Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025






#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.549 min Scan# 7
Delta R.T. -0.001 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion:168 Resp: 9512


Ion	Ratio	Lower	Upper
168	100		
99	68.6	52.3	78.5



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBS01

Manual Integrations APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#2
Dichlorodifluoromethane
Concen: 21.338 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 85 Resp: 30356

Ion	Ratio	Lower	Upper
85	100		
87	30.9	16.1	48.2






#3
Chloromethane
Concen: 19.190 ug/l
RT: 1.306 min Scan# 3
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 50 Resp: 2804


Ion	Ratio	Lower	Upper
50	100		
52	34.1	25.9	38.9



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#4
Vinyl Chloride
Concen: 19.640 ug/l
RT: 1.373 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 62 Resp: 26269

Ion	Ratio	Lower	Upper
62	100		
64	32.3	25.5	38.3






#5
Bromomethane
Concen: 19.154 ug/l
RT: 1.593 min Scan# 8
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 94 Resp: 1214


Ion	Ratio	Lower	Upper
94	100		
96	93.8	74.8	112.2



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#6
Chloroethane
Concen: 20.522 ug/l
RT: 1.666 min Scan# 95
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 64 Resp: 14563

Ion	Ratio	Lower	Upper
64	100		
66	31.6	26.3	39.5






#7
Trichlorofluoromethane
Concen: 20.139 ug/l
RT: 1.873 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion:101 Resp: 4016


Ion	Ratio	Lower	Upper
101	100		
103	68.1	51.1	76.7



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#8
Diethyl Ether
Concen: 19.796 ug/l
RT: 2.135 min Scan# 172
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 74 Resp: 13256

Ion	Ratio	Lower	Upper
74	100		
45	108.1	53.0	159.1





#9
1,1,2-Trichlorotrifluoroethane
Concen: 20.134 ug/l
RT: 2.318 min Scan# 2
Delta R.T. -0.006 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14


Tgt Ion:101 Resp: 2353


Ion	Ratio	Lower	Upper
101	100		
85	48.6	38.9	58.3
151	69.7	57.5	86.3

Instrument :
MSVOA_X
ClientSampleId :
VX0404WBS01

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025







#10
Methyl Iodide
Concen: 18.845 ug/l
RT: 2.446 min Scan# 223
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion:142 Resp: 27472

Ion	Ratio	Lower	Upper
142	100		
127	53.0	40.2	60.4
141	14.4	11.7	17.5






#11
Tert butyl alcohol
Concen: 96.466 ug/l
RT: 2.971 min Scan# 3
Delta R.T. -0.006 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 59 Resp: 2255


Ion	Ratio	Lower	Upper
59	100		
57	10.9	8.6	12.8



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#12
1,1-Dichloroethene
Concen: 19.364 ug/l
RT: 2.312 min Scan# 201
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 96 Resp: 22114

Ion	Ratio	Lower	Upper
96	100		
61	185.2	145.7	218.5
98	61.5	51.6	77.4





#13
Acrolein
Concen: 109.636 ug/l
RT: 2.233 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14


Tgt Ion: 56 Resp: 3529


Ion	Ratio	Lower	Upper
56	100		
55	73.5	56.2	84.2

Instrument :
MSVOA_X
ClientSampleId :
VX0404WBS01

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025







#14
Allyl chloride
Concen: 19.831 ug/l
RT: 2.660 min Scan# 258
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 41 Resp: 42970

Ion	Ratio	Lower	Upper
41	100		
39	74.3	60.2	90.2
76	30.2	24.6	36.8





#15
Acrylonitrile
Concen: 100.048 ug/l
RT: 3.062 min Scan# 3
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14


Tgt Ion: 53 Resp: 7383


Ion	Ratio	Lower	Upper
53	100		
52	82.8	67.4	101.2
51	38.0	29.9	44.9

Instrument :
MSVOA_X
ClientSampleId :
VX0404WBS01

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025







#16
Acetone
Concen: 100.574 ug/l
RT: 2.379 min Scan# 212
Delta R.T. -0.006 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 43 Resp: 71844

Ion	Ratio	Lower	Upper
43	100		
58	25.1	22.3	33.5






#17
Carbon Disulfide
Concen: 18.361 ug/l
RT: 2.507 min Scan# 2
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 76 Resp: 5178


Ion	Ratio	Lower	Upper
76	100		
78	8.6	7.0	10.6



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#18
Methyl Acetate
Concen: 20.160 ug/l
RT: 2.702 min Scan# 265
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 43 Resp: 33157

Ion	Ratio	Lower	Upper
43	100		
74	21.4	17.5	26.3






#19
Methyl tert-butyl Ether
Concen: 19.397 ug/l
RT: 3.111 min Scan# 3
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 73 Resp: 7749


Ion	Ratio	Lower	Upper
73	100		
57	21.0	17.6	26.4



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#20
Methylene Chloride
Concen: 19.502 ug/l
RT: 2.782 min Scan# 278
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 84 Resp: 26080

Ion	Ratio	Lower	Upper
84	100		
49	136.8	107.5	161.3
51	40.7	33.0	49.6
86	64.6	50.1	75.1






#21
trans-1,2-Dichloroethene
Concen: 19.646 ug/l
RT: 3.086 min Scan# 3
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 96 Resp: 2292


Ion	Ratio	Lower	Upper
96	100		
61	158.2	121.0	181.6
98	64.0	49.7	74.5



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#22
Diisopropyl ether
Concen: 19.265 ug/l
RT: 3.757 min Scan# 438
Delta R.T. -0.006 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 45 Resp: 82229

Ion	Ratio	Lower	Upper
45	100		
43	92.3	59.3	88.9#
87	24.6	20.4	30.6
59	11.1	9.0	13.6





#23
Vinyl Acetate
Concen: 98.665 ug/l
RT: 3.721 min Scan# 4
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14


Tgt Ion: 43 Resp: 36385


Ion	Ratio	Lower	Upper
43	100		
86	9.1	7.5	11.3

Instrument :
MSVOA_X
ClientSampleId :
VX0404WBS01

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025







#24
1,1-Dichloroethane
Concen: 19.451 ug/l
RT: 3.605 min Scan# 413
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 63 Resp: 46953

Ion	Ratio	Lower	Upper
63	100		
98	7.2	3.4	10.1
100	4.3	2.2	6.6






#25
2-Butanone
Concen: 101.462 ug/l
RT: 4.556 min Scan# 5
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 43 Resp: 10609


Ion	Ratio	Lower	Upper
43	100		
72	22.9	18.6	28.0



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#26
2,2-Dichloropropane
Concen: 21.073 ug/l
RT: 4.464 min Scan# 554
Delta R.T. -0.006 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 77 Resp: 32865

Ion	Ratio	Lower	Upper
77	100		
97	21.8	10.6	31.8






#27
cis-1,2-Dichloroethene
Concen: 19.156 ug/l
RT: 4.483 min Scan# 5
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 96 Resp: 2722


Ion	Ratio	Lower	Upper
96	100		
61	157.7	0.0	313.2
98	63.6	0.0	126.4



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#28
Bromochloromethane
Concen: 20.399 ug/l
RT: 4.891 min Scan# 624
Delta R.T. -0.006 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 49 Resp: 23795

Ion	Ratio	Lower	Upper
49	100		
129	0.8	0.0	3.8
130	65.5	54.6	81.8





#29
Tetrahydrofuran
Concen: 97.791 ug/l
RT: 5.007 min Scan# 64
Delta R.T. 0.006 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14


Tgt Ion: 42 Resp: 6629


Ion	Ratio	Lower	Upper
42	100		
72	42.0	33.4	50.0
71	38.4	31.4	47.0

Instrument :
MSVOA_X
ClientSampleId :
VX0404WBS01

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025







#30
Chloroform
Concen: 19.797 ug/l
RT: 5.086 min Scan# 656
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 83 Resp: 49173

Ion	Ratio	Lower	Upper
83	100		
85	61.3	52.8	79.2






#31
Cyclohexane
Concen: 19.493 ug/l
RT: 5.464 min Scan# 7
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 56 Resp: 4159


Ion	Ratio	Lower	Upper
56	100		
69	28.2	23.6	35.4
84	84.4	67.7	101.5



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#32
1,1,1-Trichloroethane
Concen: 19.747 ug/l
RT: 5.379 min Scan# 704
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 97 Resp: 41502

Ion	Ratio	Lower	Upper
97	100		
99	63.4	51.7	77.5
61	52.7	41.0	61.4






#33
1,2-Dichloroethane-d4
Concen: 52.873 ug/l
RT: 5.952 min Scan# 7
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 65 Resp: 9198


Ion	Ratio	Lower	Upper
65	100		
67	49.8	0.0	99.0



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.757 min Scan# 930
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion:114 Resp: 168453

Ion	Ratio	Lower	Upper
114	100		
63	23.2	0.0	46.8
88	16.5	0.0	35.4






#35
Dibromofluoromethane
Concen: 53.280 ug/l
RT: 5.385 min Scan# 7
Delta R.T. 0.006 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion:113 Resp: 6367


Ion	Ratio	Lower	Upper
113	100		
111	101.9	81.8	122.6
192	16.3	13.8	20.6



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#36
1,1-Dichloropropene
Concen: 19.807 ug/l
RT: 5.684 min Scan# 754
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 75 Resp: 31963

Ion	Ratio	Lower	Upper
75	100		
110	32.8	16.9	50.6
77	29.7	25.0	37.4






#37
Ethyl Acetate
Concen: 18.957 ug/l
RT: 4.714 min Scan# 5
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 43 Resp: 3855


Ion	Ratio	Lower	Upper
43	100		
61	12.5	9.9	14.9
70	10.0	7.5	11.3



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#38
Carbon Tetrachloride
Concen: 20.204 ug/l
RT: 5.671 min Scan# 752
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion:117 Resp: 35237

Ion	Ratio	Lower	Upper
117	100		
119	90.6	76.9	115.3
121	32.2	24.2	36.2






#39
Methylcyclohexane
Concen: 19.990 ug/l
RT: 7.372 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 83 Resp: 3954


Ion	Ratio	Lower	Upper
83	100		
55	84.2	64.9	97.3
98	51.4	37.4	56.0



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBS01

Manual Integrations APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#40
Benzene
Concen: 19.865 ug/l
RT: 6.031 min Scan# 811
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 78 Resp: 97905

Ion	Ratio	Lower	Upper
78	100		
77	24.0	19.0	28.6






#41
Methacrylonitrile
Concen: 20.585 ug/l
RT: 4.922 min Scan# 6
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 41 Resp: 2208


Ion	Ratio	Lower	Upper
41	100		
39	62.5	47.0	70.6
67	66.4	53.2	79.8
52	31.3	25.0	37.6



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#42
1,2-Dichloroethane
Concen: 20.364 ug/l
RT: 6.086 min Scan# 820
Delta R.T. 0.006 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 62 Resp: 41452

Ion	Ratio	Lower	Upper
62	100		
98	7.5	0.0	15.2






#43
Isopropyl Acetate
Concen: 19.244 ug/l
RT: 6.342 min Scan# 8
Delta R.T. 0.006 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 43 Resp: 5931


Ion	Ratio	Lower	Upper
43	100		
61	18.1	14.1	21.1
87	11.3	9.2	13.8



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#44
Trichloroethene
Concen: 19.237 ug/l
RT: 7.122 min Scan# 990
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion:130 Resp: 22535

Ion	Ratio	Lower	Upper
130	100		
95	111.8	0.0	202.0






#45
1,2-Dichloropropane
Concen: 19.804 ug/l
RT: 7.427 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 63 Resp: 2435


Ion	Ratio	Lower	Upper
63	100		
65	27.7	25.8	38.6



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#46
Dibromomethane
Concen: 20.328 ug/l
RT: 7.580 min Scan# 1065
Delta R.T. 0.006 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 93 Resp: 19189

Ion	Ratio	Lower	Upper
93	100		
95	80.6	64.8	97.2
174	84.3	65.4	98.0






#47
Bromodichloromethane
Concen: 19.932 ug/l
RT: 7.817 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 83 Resp: 3757


Ion	Ratio	Lower	Upper
83	100		
85	65.1	49.7	74.5
127	8.1	6.1	9.1



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#48
Methyl methacrylate
Concen: 20.104 ug/l
RT: 7.689 min Scan# 1083
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 41 Resp: 31993

Ion	Ratio	Lower	Upper
41	100		
69	71.6	58.2	87.2
39	65.8	53.6	80.4






#49
1,4-Dioxane
Concen: 414.998 ug/l
RT: 7.659 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 88 Resp: 1194


Ion	Ratio	Lower	Upper
88	100		
43	42.4	32.2	48.4
58	72.6	58.6	88.0



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#50
Toluene-d8
Concen: 53.195 ug/l
RT: 8.646 min Scan# 1240
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 98 Resp: 221910

Ion	Ratio	Lower	Upper
98	100		
100	63.8	52.2	78.4






#51
4-Methyl-2-Pentanone
Concen: 102.524 ug/l
RT: 8.573 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 43 Resp: 20948


Ion	Ratio	Lower	Upper
43	100		
58	35.0	28.5	42.7



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#52
Toluene
Concen: 19.623 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 92 Resp: 58522

Ion	Ratio	Lower	Upper
92	100		
91	176.2	138.7	208.1






#53
t-1,3-Dichloropropene
Concen: 18.474 ug/l
RT: 8.976 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 75 Resp: 3074


Ion	Ratio	Lower	Upper
75	100		
77	31.2	25.4	38.0



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#54
cis-1,3-Dichloropropene
Concen: 19.846 ug/l
RT: 8.366 min Scan# 1194
Delta R.T. 0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 75 Resp: 35158

Ion	Ratio	Lower	Upper
75	100		
77	32.9	24.6	36.8
39	63.6	48.4	72.6






#55
1,1,2-Trichloroethane
Concen: 20.482 ug/l
RT: 9.152 min Scan# 1
Delta R.T. 0.006 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 97 Resp: 2434


Ion	Ratio	Lower	Upper
97	100		
83	83.0	71.8	107.8
85	59.2	47.0	70.6
99	64.8	48.6	73.0



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#56
Ethyl methacrylate
Concen: 19.931 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 69 Resp: 36657

Ion	Ratio	Lower	Upper
69	100		
41	78.8	61.3	91.9
39	50.6	40.2	60.2






#57
1,3-Dichloropropane
Concen: 20.400 ug/l
RT: 9.305 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 76 Resp: 4221


Ion	Ratio	Lower	Upper
76	100		
78	31.9	26.0	39.0



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#58
2-Chloroethyl Vinyl ether
Concen: 102.664 ug/l
RT: 8.238 min Scan# 1173
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 63 Resp: 95516

Ion	Ratio	Lower	Upper
63	100		
106	25.9	21.0	31.6






#59
2-Hexanone
Concen: 101.036 ug/l
RT: 9.427 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 43 Resp: 15287


Ion	Ratio	Lower	Upper
43	100		
58	48.7	24.9	74.6



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#60
Dibromochloromethane
Concen: 19.946 ug/l
RT: 9.518 min Scan# 1383
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion:129 Resp: 25841

Ion	Ratio	Lower	Upper
129	100		
127	79.5	39.4	118.2






#61
1,2-Dibromoethane
Concen: 19.930 ug/l
RT: 9.610 min Scan# 1
Delta R.T. 0.006 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion:107 Resp: 2397


Ion	Ratio	Lower	Upper
107	100		
109	94.6	75.7	113.5



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#62
4-Bromofluorobenzene
Concen: 53.227 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 95 Resp: 80880

Ion	Ratio	Lower	Upper
95	100		
174	68.0	0.0	135.8
176	65.0	0.0	131.4






#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 14
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion:117 Resp: 15174


Ion	Ratio	Lower	Upper
117	100		
82	61.1	49.2	73.8
119	31.0	25.1	37.7



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#64
Tetrachloroethene
Concen: 19.646 ug/l
RT: 9.274 min Scan# 1343
Delta R.T. 0.005 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion:164 Resp: 21048

Ion	Ratio	Lower	Upper
164	100		
166	125.8	103.8	155.8
129	104.6	84.3	126.5
131	97.6	80.9	121.3






#65
Chlorobenzene
Concen: 19.598 ug/l
RT: 10.079 min Scan# 14
Delta R.T. 0.006 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion:112 Resp: 6354


Ion	Ratio	Lower	Upper
112	100		
114	32.1	24.7	37.1



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#66
1,1,1,2-Tetrachloroethane
Concen: 19.676 ug/l
RT: 10.158 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion:131 Resp: 22150

Ion	Ratio	Lower	Upper
131	100		
133	92.3	47.5	142.5
119	62.5	32.9	98.6






#67
Ethyl Benzene
Concen: 19.613 ug/l
RT: 10.189 min Scan# 14
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 91 Resp: 11390


Ion	Ratio	Lower	Upper
91	100		
106	28.4	23.7	35.5



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#68
m/p-Xylenes
Concen: 39.245 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion:106 Resp: 82893

Ion	Ratio	Lower	Upper
106	100		
91	218.1	172.6	258.8






#69
o-Xylene
Concen: 19.908 ug/l
RT: 10.640 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion:106 Resp: 4142


Ion	Ratio	Lower	Upper
106	100		
91	229.0	113.6	340.6



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#70
Styrene
Concen: 19.846 ug/l
RT: 10.652 min Scan# 1569
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion:104 Resp: 68156

Ion	Ratio	Lower	Upper
104	100		
78	58.8	45.4	68.2
103	55.7	43.8	65.8






#71
Bromoform
Concen: 18.768 ug/l
RT: 10.799 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion:173 Resp: 1577


Ion	Ratio	Lower	Upper
173	100		
175	49.5	23.4	70.2
254	0.0	0.0	0.0



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion:152 Resp: 66584

Ion	Ratio	Lower	Upper
152	100		
115	77.8	46.9	140.7
150	162.7	0.0	349.4






#73
Isopropylbenzene
Concen: 20.461 ug/l
RT: 10.963 min Scan# 1
Delta R.T. 0.006 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion:105 Resp: 10912


Ion	Ratio	Lower	Upper
105	100		
120	25.5	12.7	38.0



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#74
N-amyl acetate
Concen: 19.508 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 43 Resp: 49771

Ion	Ratio	Lower	Upper
43	100		
70	37.5	30.1	45.1
55	23.5	18.3	27.5
61	21.1	17.5	26.3






#75
1,1,2,2-Tetrachloroethane
Concen: 20.056 ug/l
RT: 11.207 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 83 Resp: 3757


Ion	Ratio	Lower	Upper
83	100		
131	8.9	4.7	14.1
85	61.4	32.4	97.2



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#76
1,2,3-Trichloropropane
Concen: 19.363 ug/l m
RT: 11.237 min Scan# 1665
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 75 Resp: 31448

Ion	Ratio	Lower	Upper
75	100		
77	41.4	20.0	59.9






#77
Bromobenzene
Concen: 20.293 ug/l
RT: 11.195 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion:156 Resp: 2500


Ion	Ratio	Lower	Upper
156	100		
77	182.0	92.6	277.8
158	94.5	48.1	144.4



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#78
n-propylbenzene
Concen: 20.644 ug/l
RT: 11.298 min Scan# 1675
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 91 Resp: 126822

Ion	Ratio	Lower	Upper
91	100		
120	20.8	10.8	32.3





#79
2-Chlorotoluene
Concen: 20.058 ug/l
RT: 11.359 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14


Tgt Ion: 91 Resp: 7877


Ion	Ratio	Lower	Upper
91	100		
126	31.0	15.8	47.4

Instrument :
MSVOA_X
ClientSampleId :
VX0404WBS01

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025







#80
1,3,5-Trimethylbenzene
Concen: 20.856 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion:105 Resp: 91980

Ion	Ratio	Lower	Upper
105	100		
120	45.5	23.0	69.0





#81
trans-1,4-Dichloro-2-butene
Concen: 18.366 ug/l
RT: 11.018 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14


Tgt Ion: 75 Resp: 878


Ion	Ratio	Lower	Upper
75	100		
53	135.6	101.8	152.6
89	46.6	37.2	55.8

Instrument :
MSVOA_X
ClientSampleId :
VX0404WBS01

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025







#82
4-Chlorotoluene
Concen: 20.657 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 91 Resp: 90453

Ion	Ratio	Lower	Upper
91	100		
126	26.6	13.3	39.9






#83
tert-Butylbenzene
Concen: 20.336 ug/l
RT: 11.713 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion:119 Resp: 8879


Ion	Ratio	Lower	Upper
119	100		
91	68.3	32.4	97.2
134	22.9	11.5	34.5



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBS01

Manual Integrations APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#84
1,2,4-Trimethylbenzene
Concen: 20.614 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion:105 Resp: 91248

Ion	Ratio	Lower	Upper
105	100		
120	42.1	21.4	64.3






#85
sec-Butylbenzene
Concen: 20.849 ug/l
RT: 11.890 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion:105 Resp: 11181


Ion	Ratio	Lower	Upper
105	100		
134	19.1	9.6	28.6



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBS01

Manual Integrations APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#86
p-Isopropyltoluene
Concen: 20.513 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion:119 Resp: 90695

Ion	Ratio	Lower	Upper
119	100		
134	25.4	12.6	37.8
91	27.7	13.7	41.1






#87
1,3-Dichlorobenzene
Concen: 19.656 ug/l
RT: 11.969 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion:146 Resp: 4409


Ion	Ratio	Lower	Upper
146	100		
111	43.8	22.2	66.6
148	63.4	31.9	95.7



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBS01

Manual Integrations APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#88
1,4-Dichlorobenzene
Concen: 19.494 ug/l
RT: 12.036 min Scan# 1796
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion:146 Resp: 44331

Ion	Ratio	Lower	Upper
146	100		
111	43.7	20.9	62.8
148	63.9	31.5	94.5






#89
n-Butylbenzene
Concen: 20.250 ug/l
RT: 12.329 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 91 Resp: 7763


Ion	Ratio	Lower	Upper
91	100		
92	53.6	26.9	80.6
134	23.1	11.9	35.7



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#90
Hexachloroethane
Concen: 20.229 ug/l
RT: 12.536 min Scan# 1878
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion:117 Resp: 15371

Ion	Ratio	Lower	Upper
117	100		
201	61.3	31.2	93.6






#91
1,2-Dichlorobenzene
Concen: 19.875 ug/l
RT: 12.335 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion:146 Resp: 4432


Ion	Ratio	Lower	Upper
146	100		
111	44.5	22.9	68.8
148	63.5	31.9	95.5



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#92
1,2-Dibromo-3-Chloropropane
Concen: 20.243 ug/l
RT: 12.938 min Scan# 1944
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion: 75 Resp: 7932

Ion	Ratio	Lower	Upper
75	100		
155	66.9	34.9	104.7
157	87.5	44.9	134.7






#93
1,2,4-Trichlorobenzene
Concen: 19.059 ug/l
RT: 13.585 min Scan# 2
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion:180 Resp: 2366


Ion	Ratio	Lower	Upper
180	100		
182	93.8	47.9	143.6
145	32.8	17.9	53.8



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBS01

Manual Integrations APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#94
Hexachlorobutadiene
Concen: 19.457 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. 0.006 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion:225 Resp: 10408

Ion	Ratio	Lower	Upper
225	100		
223	64.9	31.1	93.3
227	65.7	32.2	96.6






#95
Naphthalene
Concen: 18.950 ug/l
RT: 13.774 min Scan# 2
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion:128 Resp: 8754


Ion	Ratio	Lower	Upper
128	100		
127	12.9	10.3	15.5
129	11.1	8.6	12.8



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBS01

Manual Integrations APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025



#96
1,2,3-Trichlorobenzene
Concen: 19.132 ug/l
RT: 13.956 min Scan# 2111
Delta R.T. -0.000 min
Lab File: VX045589.D
Acq: 04 Apr 2025 11:14

Tgt Ion:180 Resp: 24870

Ion	Ratio	Lower	Upper
180	100		
182	91.6	47.3	141.8
145	34.2	18.2	54.6



Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	
Project:	Former Schlumberger STC PTC Site D3868221	Date Received:	
Client Sample ID:	VX0407WBS01	SDG No.:	Q1731
Lab Sample ID:	VX0407WBS01	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group3
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045617.D	1		04/07/25 10:34	VX040725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	18.0		0.26	1.00	ug/L
75-35-4	1,1-Dichloroethene	17.5		0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	17.8		0.23	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	17.5		0.19	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	18.0		0.20	1.00	ug/L
71-43-2	Benzene	18.0		0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	18.7		0.22	1.00	ug/L
79-01-6	Trichloroethene	18.0		0.090	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	18.2		0.21	1.00	ug/L
127-18-4	Tetrachloroethene	19.3		0.23	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	50.8		70 (74) - 130 (125)	102%	SPK: 50
1868-53-7	Dibromofluoromethane	50.0		70 (75) - 130 (124)	100%	SPK: 50
2037-26-5	Toluene-d8	49.7		70 (86) - 130 (113)	99%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.4		70 (77) - 130 (121)	101%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	89700	5.544			
540-36-3	1,4-Difluorobenzene	159000	6.757			
3114-55-4	Chlorobenzene-d5	139000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	64900	12.018			

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements
J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040725\
Data File : VX045617.D
Acq On : 07 Apr 2025 10:34
Operator : JC/MD
Sample : VX0407WBS01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Quant Time: Apr 08 01:34:29 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VX0407WBS01

Manual Integrations APPROVED

Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	5.544	168	89694	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	158717	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	138981	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	64861	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	83331	50.803	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 101.600%			
35) Dibromofluoromethane	5.379	113	56342	50.033	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 100.060%			
50) Toluene-d8	8.647	98	195538	49.749	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 99.500%			
62) 4-Bromofluorobenzene	11.079	95	72106	50.364	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 100.720%			
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.166	85	26539	19.785	ug/l	96
3) Chloromethane	1.307	50	24082	17.476	ug/l	98
4) Vinyl Chloride	1.374	62	22749	18.039	ug/l	96
5) Bromomethane	1.593	94	10531	17.611	ug/l	95
6) Chloroethane	1.666	64	13064	19.525	ug/l	95
7) Trichlorofluoromethane	1.874	101	35064	18.645	ug/l	98
8) Diethyl Ether	2.130	74	10899	17.262	ug/l	92
9) 1,1,2-Trichlorotrifluo...	2.319	101	20363	18.479	ug/l	99
10) Methyl Iodide	2.447	142	23748	17.277	ug/l	98
11) Tert butyl alcohol	2.977	59	19430	88.142	ug/l	99
12) 1,1-Dichloroethene	2.313	96	18864	17.519	ug/l	96
13) Acrolein	2.233	56	32886	108.334	ug/l	98
14) Allyl chloride	2.654	41	37227	18.221	ug/l	99
15) Acrylonitrile	3.062	53	64059	92.059	ug/l	98
16) Acetone	2.380	43	60898	90.415	ug/l	96
17) Carbon Disulfide	2.502	76	44530	16.744	ug/l	100
18) Methyl Acetate	2.703	43	29937	19.305	ug/l	99
19) Methyl tert-butyl Ether	3.117	73	66392	17.626	ug/l	99
20) Methylene Chloride	2.782	84	22301	17.686	ug/l	95
21) trans-1,2-Dichloroethene	3.087	96	19342	17.583	ug/l	96
22) Diisopropyl ether	3.764	45	70517	17.522	ug/l	90
23) Vinyl Acetate	3.721	43	307639	88.474	ug/l	98
24) 1,1-Dichloroethane	3.605	63	40491	17.790	ug/l	99
25) 2-Butanone	4.556	43	89883	91.166	ug/l	100
26) 2,2-Dichloropropane	4.465	77	29308	19.930	ug/l	99
27) cis-1,2-Dichloroethene	4.483	96	23391	17.453	ug/l	96
28) Bromochloromethane	4.897	49	22119	20.110	ug/l	99
29) Tetrahydrofuran	5.007	42	57774	90.384	ug/l	99
30) Chloroform	5.093	83	42943	18.336	ug/l	99
31) Cyclohexane	5.458	56	34529	17.161	ug/l	96
32) 1,1,1-Trichloroethane	5.379	97	35642	17.986	ug/l	99
36) 1,1-Dichloropropene	5.684	75	27361	17.995	ug/l	99
37) Ethyl Acetate	4.721	43	32485	16.954	ug/l	98
38) Carbon Tetrachloride	5.672	117	30822	18.757	ug/l	97
39) Methylcyclohexane	7.373	83	32996	17.704	ug/l	97
40) Benzene	6.031	78	83405	17.961	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040725\
Data File : VX045617.D
Acq On : 07 Apr 2025 10:34
Operator : JC/MD
Sample : VX0407WBS01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Quant Time: Apr 08 01:34:29 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VX0407WBS01

Manual Integrations APPROVED

Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
41) Methacrylonitrile	4.922	41	18930	18.724	ug/l		99
42) 1,2-Dichloroethane	6.086	62	35769	18.650	ug/l		100
43) Isopropyl Acetate	6.342	43	52567	18.101	ug/l		99
44) Trichloroethene	7.123	130	19895	18.025	ug/l		98
45) 1,2-Dichloropropane	7.428	63	21087	18.200	ug/l		100
46) Dibromomethane	7.580	93	16073	18.071	ug/l		97
47) Bromodichloromethane	7.818	83	32271	18.168	ug/l		97
48) Methyl methacrylate	7.696	41	27006	18.011	ug/l		99
49) 1,4-Dioxane	7.659	88	10577	389.882	ug/l		98
51) 4-Methyl-2-Pentanone	8.574	43	181149	94.095	ug/l		98
52) Toluene	8.714	92	50130	17.841	ug/l		99
53) t-1,3-Dichloropropene	8.976	75	26146	16.895	ug/l		97
54) cis-1,3-Dichloropropene	8.366	75	30988	18.565	ug/l		99
55) 1,1,2-Trichloroethane	9.153	97	20414	18.232	ug/l		92
56) Ethyl methacrylate	9.116	69	30029	17.329	ug/l		96
57) 1,3-Dichloropropane	9.305	76	35623	18.270	ug/l		100
58) 2-Chloroethyl Vinyl ether	8.238	63	69612	79.411	ug/l		100
59) 2-Hexanone	9.427	43	130780	91.734	ug/l		100
60) Dibromochloromethane	9.519	129	22685	18.584	ug/l		99
61) 1,2-Dibromoethane	9.610	107	20690	18.254	ug/l		99
64) Tetrachloroethene	9.269	164	18965	19.327	ug/l		93
65) Chlorobenzene	10.079	112	54159	18.237	ug/l		98
66) 1,1,1,2-Tetrachloroethane	10.159	131	18533	17.975	ug/l		100
67) Ethyl Benzene	10.189	91	96615	18.164	ug/l		98
68) m/p-Xylenes	10.299	106	70049	36.210	ug/l		98
69) o-Xylene	10.640	106	34794	18.258	ug/l		98
70) Styrene	10.653	104	57575	18.305	ug/l		99
71) Bromoform	10.799	173	14155	18.386	ug/l	#	98
73) Isopropylbenzene	10.963	105	92499	17.804	ug/l		99
74) N-amyl acetate	10.842	43	42430	17.073	ug/l		99
75) 1,1,2,2-Tetrachloroethane	11.207	83	32065	17.569	ug/l		99
76) 1,2,3-Trichloropropane	11.238	75	27650m	17.477	ug/l		
77) Bromobenzene	11.195	156	21356	17.790	ug/l		98
78) n-propylbenzene	11.305	91	105538	17.636	ug/l		100
79) 2-Chlorotoluene	11.360	91	67251	17.579	ug/l		99
80) 1,3,5-Trimethylbenzene	11.451	105	76216	17.740	ug/l		99
81) trans-1,4-Dichloro-2-b...	11.018	75	7803	16.739	ug/l		96
82) 4-Chlorotoluene	11.451	91	75042	17.593	ug/l		100
83) tert-Butylbenzene	11.713	119	74797	17.585	ug/l		97
84) 1,2,4-Trimethylbenzene	11.750	105	76620	17.769	ug/l		99
85) sec-Butylbenzene	11.890	105	92241	17.657	ug/l		100
86) p-Isopropyltoluene	12.006	119	75592	17.551	ug/l		99
87) 1,3-Dichlorobenzene	11.969	146	37807	17.302	ug/l		99
88) 1,4-Dichlorobenzene	12.043	146	37519	16.936	ug/l		97
89) n-Butylbenzene	12.329	91	64350	17.230	ug/l		98
90) Hexachloroethane	12.536	117	13254	17.906	ug/l		98
91) 1,2-Dichlorobenzene	12.335	146	38002	17.492	ug/l		99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	6802	17.821	ug/l		99
93) 1,2,4-Trichlorobenzene	13.585	180	19889	16.447	ug/l		98
94) Hexachlorobutadiene	13.719	225	9339	17.922	ug/l		96
95) Naphthalene	13.774	128	73854	16.411	ug/l		99
96) 1,2,3-Trichlorobenzene	13.957	180	21009	16.591	ug/l		99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040725\
Data File : VX045617.D
Acq On : 07 Apr 2025 10:34
Operator : JC/MD
Sample : VX0407WBS01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Quant Time: Apr 08 01:34:29 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VX0407WBS01

Manual Integrations APPROVED

Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040725\
Data File : VX045617.D
Acq On : 07 Apr 2025 10:34
Operator : JC/MD
Sample : VX0407WBS01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1


Quant Time: Apr 08 01:34:29 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VX0407WBS01

Manual Integrations APPROVED

Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025






#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 7
Delta R.T. -0.006 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion:168 Resp: 8969


Ion	Ratio	Lower	Upper
168	100		
99	71.3	52.3	78.5



Instrument :
MSVOA_X
ClientSampleId :
VX0407WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#2
Dichlorodifluoromethane
Concen: 19.785 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion: 85 Resp: 26539

Ion	Ratio	Lower	Upper
85	100		
87	29.7	16.1	48.2






#3
Chloromethane
Concen: 17.476 ug/l
RT: 1.307 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion: 50 Resp: 2408


Ion	Ratio	Lower	Upper
50	100		
52	33.3	25.9	38.9



Instrument :
MSVOA_X
ClientSampleId :
VX0407WBS01

Manual Integrations APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#4
Vinyl Chloride
Concen: 18.039 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion: 62 Resp: 22749

Ion	Ratio	Lower	Upper
62	100		
64	29.4	25.5	38.3






#5
Bromomethane
Concen: 17.611 ug/l
RT: 1.593 min Scan# 8
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion: 94 Resp: 1053


Ion	Ratio	Lower	Upper
94	100		
96	88.5	74.8	112.2



Instrument :
MSVOA_X
ClientSampleId :
VX0407WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#6
Chloroethane
Concen: 19.525 ug/l
RT: 1.666 min Scan# 95
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion: 64 Resp: 13064

Ion	Ratio	Lower	Upper
64	100		
66	35.8	26.3	39.5






#7
Trichlorofluoromethane
Concen: 18.645 ug/l
RT: 1.874 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion:101 Resp: 3506


Ion	Ratio	Lower	Upper
101	100		
103	65.7	51.1	76.7



Instrument :
MSVOA_X
ClientSampleId :
VX0407WBS01

Manual Integrations APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#8
Diethyl Ether
Concen: 17.262 ug/l
RT: 2.130 min Scan# 171
Delta R.T. -0.006 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion: 74 Resp: 10899

Ion	Ratio	Lower	Upper
74	100		
45	114.0	53.0	159.1






#9
1,1,2-Trichlorotrifluoroethane
Concen: 18.479 ug/l
RT: 2.319 min Scan# 2
Delta R.T. -0.006 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion:101 Resp: 2036


Ion	Ratio	Lower	Upper
101	100		
85	48.4	38.9	58.3
151	70.8	57.5	86.3



Instrument :
MSVOA_X
ClientSampleId :
VX0407WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#10
Methyl Iodide
Concen: 17.277 ug/l
RT: 2.447 min Scan# 223
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion:142 Resp: 23748

Ion	Ratio	Lower	Upper
142	100		
127	51.6	40.2	60.4
141	13.9	11.7	17.5






#11
Tert butyl alcohol
Concen: 88.142 ug/l
RT: 2.977 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion: 59 Resp: 1943


Ion	Ratio	Lower	Upper
59	100		
57	11.0	8.6	12.8



Instrument :
MSVOA_X
ClientSampleId :
VX0407WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#12
1,1-Dichloroethene
Concen: 17.519 ug/l
RT: 2.313 min Scan# 201
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion: 96 Resp: 18864

Ion	Ratio	Lower	Upper
96	100		
61	176.8	145.7	218.5
98	66.7	51.6	77.4






#13
Acrolein
Concen: 108.334 ug/l
RT: 2.233 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion: 56 Resp: 3288


Ion	Ratio	Lower	Upper
56	100		
55	71.7	56.2	84.2



Instrument :
MSVOA_X
ClientSampleId :
VX0407WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#14
Allyl chloride
Concen: 18.221 ug/l
RT: 2.654 min Scan# 257
Delta R.T. -0.006 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion: 41 Resp: 37227

Ion	Ratio	Lower	Upper
41	100		
39	76.6	60.2	90.2
76	30.3	24.6	36.8






#15
Acrylonitrile
Concen: 92.059 ug/l
RT: 3.062 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion: 53 Resp: 6405


Ion	Ratio	Lower	Upper
53	100		
52	82.0	67.4	101.2
51	37.3	29.9	44.9



Instrument :
MSVOA_X
ClientSampleId :
VX0407WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#16
Acetone
Concen: 90.415 ug/l
RT: 2.380 min Scan# 212
Delta R.T. -0.006 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion: 43 Resp: 60898

Ion	Ratio	Lower	Upper
43	100		
58	26.0	22.3	33.5





#17
Carbon Disulfide
Concen: 16.744 ug/l
RT: 2.502 min Scan# 2
Delta R.T. -0.006 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34


Tgt Ion: 76 Resp: 4453


Ion	Ratio	Lower	Upper
76	100		
78	8.7	7.0	10.6

Instrument :
MSVOA_X
ClientSampleId :
VX0407WBS01

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025







#18
Methyl Acetate
Concen: 19.305 ug/l
RT: 2.703 min Scan# 265
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion: 43 Resp: 29937

Ion	Ratio	Lower	Upper
43	100		
74	21.5	17.5	26.3






#19
Methyl tert-butyl Ether
Concen: 17.626 ug/l
RT: 3.117 min Scan# 3
Delta R.T. 0.006 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion: 73 Resp: 6639


Ion	Ratio	Lower	Upper
73	100		
57	22.6	17.6	26.4



Instrument :
MSVOA_X
ClientSampleId :
VX0407WBS01

Manual Integrations APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#20
Methylene Chloride
Concen: 17.686 ug/l
RT: 2.782 min Scan# 278
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion: 84 Resp: 22301

Ion	Ratio	Lower	Upper
84	100		
49	144.4	107.5	161.3
51	42.3	33.0	49.6
86	62.5	50.1	75.1






#21
trans-1,2-Dichloroethene
Concen: 17.583 ug/l
RT: 3.087 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion: 96 Resp: 1934


Ion	Ratio	Lower	Upper
96	100		
61	154.5	121.0	181.6
98	67.0	49.7	74.5



Instrument :
MSVOA_X
ClientSampleId :
VX0407WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#22
Diisopropyl ether
Concen: 17.522 ug/l
RT: 3.764 min Scan# 439
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion: 45 Resp: 70517

Ion	Ratio	Lower	Upper
45	100		
43	83.7	59.3	88.9
87	29.6	20.4	30.6
59	13.3	9.0	13.6





#23
Vinyl Acetate
Concen: 88.474 ug/l
RT: 3.721 min Scan# 4
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34


Tgt Ion: 43 Resp: 30763


Ion	Ratio	Lower	Upper
43	100		
86	10.2	7.5	11.3

Instrument :
MSVOA_X
ClientSampleId :
VX0407WBS01

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025







#24
1,1-Dichloroethane
Concen: 17.790 ug/l
RT: 3.605 min Scan# 413
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion: 63 Resp: 40491

Ion	Ratio	Lower	Upper
63	100		
98	6.6	3.4	10.1
100	3.9	2.2	6.6






#25
2-Butanone
Concen: 91.166 ug/l
RT: 4.556 min Scan# 5
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion: 43 Resp: 8988


Ion	Ratio	Lower	Upper
43	100		
72	23.2	18.6	28.0



Instrument :
MSVOA_X
ClientSampleId :
VX0407WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#26
2,2-Dichloropropane
Concen: 19.930 ug/l
RT: 4.465 min Scan# 554
Delta R.T. -0.006 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion: 77 Resp: 29308

Ion	Ratio	Lower	Upper
77	100		
97	20.9	10.6	31.8






#27
cis-1,2-Dichloroethene
Concen: 17.453 ug/l
RT: 4.483 min Scan# 5
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion: 96 Resp: 2339


Ion	Ratio	Lower	Upper
96	100		
61	162.1	0.0	313.2
98	65.8	0.0	126.4



Instrument :
MSVOA_X
ClientSampleId :
VX0407WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#28
Bromochloromethane
Concen: 20.110 ug/l
RT: 4.897 min Scan# 625
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion: 49 Resp: 22119

Ion	Ratio	Lower	Upper
49	100		
129	1.5	0.0	3.8
130	67.8	54.6	81.8





#29
Tetrahydrofuran
Concen: 90.384 ug/l
RT: 5.007 min Scan# 64
Delta R.T. 0.006 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34


Tgt Ion: 42 Resp: 5777


Ion	Ratio	Lower	Upper
42	100		
72	42.4	33.4	50.0
71	39.0	31.4	47.0

Instrument :
MSVOA_X
ClientSampleId :
VX0407WBS01

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025







#30
Chloroform
Concen: 18.336 ug/l
RT: 5.093 min Scan# 657
Delta R.T. 0.006 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion: 83 Resp: 42943

Ion	Ratio	Lower	Upper
83	100		
85	66.5	52.8	79.2






#31
Cyclohexane
Concen: 17.161 ug/l
RT: 5.458 min Scan# 7
Delta R.T. -0.006 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion: 56 Resp: 3452


Ion	Ratio	Lower	Upper
56	100		
69	25.0	23.6	35.4
84	82.5	67.7	101.5



Instrument :
MSVOA_X
ClientSampleId :
VX0407WBS01

Manual Integrations APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#32
1,1,1-Trichloroethane
Concen: 17.986 ug/l
RT: 5.379 min Scan# 704
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion: 97 Resp: 35642

Ion	Ratio	Lower	Upper
97	100		
99	64.0	51.7	77.5
61	52.5	41.0	61.4






#33
1,2-Dichloroethane-d4
Concen: 50.803 ug/l
RT: 5.952 min Scan# 7
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion: 65 Resp: 8333


Ion	Ratio	Lower	Upper
65	100		
67	49.6	0.0	99.0



Instrument :
MSVOA_X
ClientSampleId :
VX0407WBS01

Manual Integrations APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.757 min Scan# 930
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion:114 Resp: 158717

Ion	Ratio	Lower	Upper
114	100		
63	23.9	0.0	46.8
88	17.8	0.0	35.4






#35
Dibromofluoromethane
Concen: 50.033 ug/l
RT: 5.379 min Scan# 7
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion:113 Resp: 5634


Ion	Ratio	Lower	Upper
113	100		
111	103.7	81.8	122.6
192	17.1	13.8	20.6



Instrument :
MSVOA_X
ClientSampleId :
VX0407WBS01

Manual Integrations APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#36
1,1-Dichloropropene
Concen: 17.995 ug/l
RT: 5.684 min Scan# 754
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion: 75 Resp: 27361

Ion	Ratio	Lower	Upper
75	100		
110	33.2	16.9	50.6
77	31.1	25.0	37.4






#37
Ethyl Acetate
Concen: 16.954 ug/l
RT: 4.721 min Scan# 5
Delta R.T. 0.006 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion: 43 Resp: 3248


Ion	Ratio	Lower	Upper
43	100		
61	13.6	9.9	14.9
70	9.4	7.5	11.3



Instrument :
MSVOA_X
ClientSampleId :
VX0407WBS01

Manual Integrations APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#38
Carbon Tetrachloride
Concen: 18.757 ug/l
RT: 5.672 min Scan# 752
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion:117 Resp: 30822

Ion	Ratio	Lower	Upper
117	100		
119	93.2	76.9	115.3
121	29.2	24.2	36.2






#39
Methylcyclohexane
Concen: 17.704 ug/l
RT: 7.373 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion: 83 Resp: 3299


Ion	Ratio	Lower	Upper
83	100		
55	83.2	64.9	97.3
98	49.9	37.4	56.0



Instrument :
MSVOA_X
ClientSampleId :
VX0407WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#40
Benzene
Concen: 17.961 ug/l
RT: 6.031 min Scan# 811
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion: 78 Resp: 83405

Ion	Ratio	Lower	Upper
78	100		
77	24.3	19.0	28.6






#41
Methacrylonitrile
Concen: 18.724 ug/l
RT: 4.922 min Scan# 6
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34


Tgt Ion: 41 Resp: 1893

Ion	Ratio	Lower	Upper
41	100		
39	58.6	47.0	70.6
67	65.7	53.2	79.8
52	32.7	25.0	37.6



Instrument :
MSVOA_X
ClientSampleId :
VX0407WBS01


Manual Integrations
APPROVED
Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#42
1,2-Dichloroethane
Concen: 18.650 ug/l
RT: 6.086 min Scan# 820
Delta R.T. 0.006 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion: 62 Resp: 35769

Ion	Ratio	Lower	Upper
62	100		
98	7.7	0.0	15.2





#43
Isopropyl Acetate
Concen: 18.101 ug/l
RT: 6.342 min Scan# 8
Delta R.T. 0.006 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34


Tgt Ion: 43 Resp: 5256


Ion	Ratio	Lower	Upper
43	100		
61	17.1	14.1	21.1
87	11.2	9.2	13.8

Instrument :
MSVOA_X
ClientSampleId :
VX0407WBS01

Manual Integrations APPROVED

Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025







#44
Trichloroethene
Concen: 18.025 ug/l
RT: 7.123 min Scan# 990
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion:130 Resp: 19895

Ion	Ratio	Lower	Upper
130	100		
95	103.1	0.0	202.0






#45
1,2-Dichloropropane
Concen: 18.200 ug/l
RT: 7.428 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion: 63 Resp: 2108


Ion	Ratio	Lower	Upper
63	100		
65	32.3	25.8	38.6



Instrument :
MSVOA_X
ClientSampleId :
VX0407WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#46
Dibromomethane
Concen: 18.071 ug/l
RT: 7.580 min Scan# 1065
Delta R.T. 0.006 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion: 93 Resp: 16073

Ion	Ratio	Lower	Upper
93	100		
95	84.6	64.8	97.2
174	83.2	65.4	98.0






#47
Bromodichloromethane
Concen: 18.168 ug/l
RT: 7.818 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion: 83 Resp: 3227


Ion	Ratio	Lower	Upper
83	100		
85	64.7	49.7	74.5
127	8.6	6.1	9.1



Instrument :
MSVOA_X
ClientSampleId :
VX0407WBS01

Manual Integrations APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#48
Methyl methacrylate
Concen: 18.011 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. 0.006 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion: 41 Resp: 27006

Ion	Ratio	Lower	Upper
41	100		
69	71.7	58.2	87.2
39	65.7	53.6	80.4






#49
1,4-Dioxane
Concen: 389.882 ug/l
RT: 7.659 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion: 88 Resp: 1057


Ion	Ratio	Lower	Upper
88	100		
43	39.8	32.2	48.4
58	76.0	58.6	88.0



Instrument :
MSVOA_X
ClientSampleId :
VX0407WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#50
Toluene-d8
Concen: 49.749 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion: 98 Resp: 195538

Ion	Ratio	Lower	Upper
98	100		
100	64.2	52.2	78.4






#51
4-Methyl-2-Pentanone
Concen: 94.095 ug/l
RT: 8.574 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion: 43 Resp: 18114


Ion	Ratio	Lower	Upper
43	100		
58	34.7	28.5	42.7



Instrument :
MSVOA_X
ClientSampleId :
VX0407WBS01

Manual Integrations APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#52
Toluene
Concen: 17.841 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion: 92 Resp: 50130

Ion	Ratio	Lower	Upper
92	100		
91	172.3	138.7	208.1






#53
t-1,3-Dichloropropene
Concen: 16.895 ug/l
RT: 8.976 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion: 75 Resp: 2614


Ion	Ratio	Lower	Upper
75	100		
77	30.2	25.4	38.0



Instrument :
MSVOA_X
ClientSampleId :
VX0407WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#54
cis-1,3-Dichloropropene
Concen: 18.565 ug/l
RT: 8.366 min Scan# 1194
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion: 75 Resp: 30988

Ion	Ratio	Lower	Upper
75	100		
77	31.9	24.6	36.8
39	60.8	48.4	72.6






#55
1,1,2-Trichloroethane
Concen: 18.232 ug/l
RT: 9.153 min Scan# 1
Delta R.T. 0.006 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion: 97 Resp: 2041


Ion	Ratio	Lower	Upper
97	100		
83	82.2	71.8	107.8
85	53.5	47.0	70.6
99	67.4	48.6	73.0



Instrument :
MSVOA_X
ClientSampleId :
VX0407WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#56
Ethyl methacrylate
Concen: 17.329 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion: 69 Resp: 30029

Ion	Ratio	Lower	Upper
69	100		
41	81.7	61.3	91.9
39	50.1	40.2	60.2






#57
1,3-Dichloropropane
Concen: 18.270 ug/l
RT: 9.305 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion: 76 Resp: 3562


Ion	Ratio	Lower	Upper
76	100		
78	32.5	26.0	39.0



Instrument :
MSVOA_X
ClientSampleId :
VX0407WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#58
2-Chloroethyl Vinyl ether
Concen: 79.411 ug/l
RT: 8.238 min Scan# 1173
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion: 63 Resp: 69612

Ion	Ratio	Lower	Upper
63	100		
106	26.3	21.0	31.6






#59
2-Hexanone
Concen: 91.734 ug/l
RT: 9.427 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion: 43 Resp: 13078


Ion	Ratio	Lower	Upper
43	100		
58	50.0	24.9	74.6



Instrument :
MSVOA_X
ClientSampleId :
VX0407WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#60
Dibromochloromethane
Concen: 18.584 ug/l
RT: 9.519 min Scan# 1383
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion:129 Resp: 22685

Ion	Ratio	Lower	Upper
129	100		
127	77.7	39.4	118.2






#61
1,2-Dibromoethane
Concen: 18.254 ug/l
RT: 9.610 min Scan# 1
Delta R.T. 0.006 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion:107 Resp: 2069


Ion	Ratio	Lower	Upper
107	100		
109	95.2	75.7	113.5



Instrument :
MSVOA_X
ClientSampleId :
VX0407WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#62
4-Bromofluorobenzene
Concen: 50.364 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion: 95 Resp: 72106

Ion	Ratio	Lower	Upper
95	100		
174	68.7	0.0	135.8
176	65.8	0.0	131.4






#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion:117 Resp: 13898


Ion	Ratio	Lower	Upper
117	100		
82	62.2	49.2	73.8
119	31.6	25.1	37.7



Instrument :
MSVOA_X
ClientSampleId :
VX0407WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#64
Tetrachloroethene
Concen: 19.327 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. -0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion:164 Resp: 18965

Ion	Ratio	Lower	Upper
164	100		
166	126.2	103.8	155.8
129	97.9	84.3	126.5
131	89.8	80.9	121.3






#65
Chlorobenzene
Concen: 18.237 ug/l
RT: 10.079 min Scan# 14
Delta R.T. 0.006 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion:112 Resp: 5415


Ion	Ratio	Lower	Upper
112	100		
114	32.0	24.7	37.1



Instrument :
MSVOA_X
ClientSampleId :
VX0407WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#66
1,1,1,2-Tetrachloroethane
Concen: 17.975 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion:131 Resp: 18533

Ion	Ratio	Lower	Upper
131	100		
133	95.5	47.5	142.5
119	66.0	32.9	98.6





#67
Ethyl Benzene
Concen: 18.164 ug/l
RT: 10.189 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34


Tgt Ion: 91 Resp: 9661


Ion	Ratio	Lower	Upper
91	100		
106	28.3	23.7	35.5

Instrument :
MSVOA_X
ClientSampleId :
VX0407WBS01

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025







#68
m/p-Xylenes
Concen: 36.210 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion:106 Resp: 70049

Ion	Ratio	Lower	Upper
106	100		
91	218.3	172.6	258.8






#69
o-Xylene
Concen: 18.258 ug/l
RT: 10.640 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion:106 Resp: 3479


Ion	Ratio	Lower	Upper
106	100		
91	229.7	113.6	340.6



Instrument :
MSVOA_X
ClientSampleId :
VX0407WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#70
Styrene
Concen: 18.305 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion:104 Resp: 57575

Ion	Ratio	Lower	Upper
104	100		
78	58.2	45.4	68.2
103	55.0	43.8	65.8






#71
Bromoform
Concen: 18.386 ug/l
RT: 10.799 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion:173 Resp: 1415


Ion	Ratio	Lower	Upper
173	100		
175	48.1	23.4	70.2
254	0.0	0.0	0.0



Instrument :
MSVOA_X
ClientSampleId :
VX0407WBS01

Manual Integrations APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion:152 Resp: 64861

Ion	Ratio	Lower	Upper
152	100		
115	73.8	46.9	140.7
150	160.2	0.0	349.4





#73
Isopropylbenzene
Concen: 17.804 ug/l
RT: 10.963 min Scan# 1
Delta R.T. 0.006 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34


Tgt Ion:105 Resp: 9249


Ion	Ratio	Lower	Upper
105	100		
120	25.7	12.7	38.0

Instrument :
MSVOA_X
ClientSampleId :
VX0407WBS01

Manual Integrations APPROVED

Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025







#74
N-amyl acetate
Concen: 17.073 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion: 43 Resp: 42430

Ion	Ratio	Lower	Upper
43	100		
70	38.3	30.1	45.1
55	23.9	18.3	27.5
61	21.4	17.5	26.3






#75
1,1,2,2-Tetrachloroethane
Concen: 17.569 ug/l
RT: 11.207 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion: 83 Resp: 3206


Ion	Ratio	Lower	Upper
83	100		
131	9.8	4.7	14.1
85	63.7	32.4	97.2



Instrument :
MSVOA_X
ClientSampleId :
VX0407WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#76
1,2,3-Trichloropropane
Concen: 17.477 ug/l m
RT: 11.238 min Scan# 1665
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion: 75 Resp: 27650

Ion	Ratio	Lower	Upper
75	100		
77	41.1	20.0	59.9






#77
Bromobenzene
Concen: 17.790 ug/l
RT: 11.195 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion:156 Resp: 2135


Ion	Ratio	Lower	Upper
156	100		
77	181.1	92.6	277.8
158	96.5	48.1	144.4



Instrument :
MSVOA_X
ClientSampleId :
VX0407WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#78
n-propylbenzene
Concen: 17.636 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. 0.006 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion: 91 Resp: 105538

Ion	Ratio	Lower	Upper
91	100		
120	21.6	10.8	32.3





#79
2-Chlorotoluene
Concen: 17.579 ug/l
RT: 11.360 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34


Tgt Ion: 91 Resp: 6725


Ion	Ratio	Lower	Upper
91	100		
126	31.2	15.8	47.4

Instrument :
MSVOA_X
ClientSampleId :
VX0407WBS01

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025







#80
1,3,5-Trimethylbenzene
Concen: 17.740 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion:105 Resp: 76216

Ion	Ratio	Lower	Upper
105	100		
120	46.8	23.0	69.0





#81
trans-1,4-Dichloro-2-butene
Concen: 16.739 ug/l
RT: 11.018 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34


Tgt Ion: 75 Resp: 780


Ion	Ratio	Lower	Upper
75	100		
53	132.8	101.8	152.6
89	46.2	37.2	55.8

Instrument :
MSVOA_X
ClientSampleId :
VX0407WBS01

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025







#82
4-Chlorotoluene
Concen: 17.593 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion: 91 Resp: 75042

Ion	Ratio	Lower	Upper
91	100		
126	26.7	13.3	39.9






#83
tert-Butylbenzene
Concen: 17.585 ug/l
RT: 11.713 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion:119 Resp: 7479


Ion	Ratio	Lower	Upper
119	100		
91	68.0	32.4	97.2
134	23.3	11.5	34.5



Instrument :
MSVOA_X
ClientSampleId :
VX0407WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#84
1,2,4-Trimethylbenzene
Concen: 17.769 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion:105 Resp: 76620

Ion	Ratio	Lower	Upper
105	100		
120	43.5	21.4	64.3






#85
sec-Butylbenzene
Concen: 17.657 ug/l
RT: 11.890 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion:105 Resp: 9224


Ion	Ratio	Lower	Upper
105	100		
134	19.1	9.6	28.6



Instrument :
MSVOA_X
ClientSampleId :
VX0407WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#86
p-Isopropyltoluene
Concen: 17.551 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion:119 Resp: 75592

Ion	Ratio	Lower	Upper
119	100		
134	25.5	12.6	37.8
91	28.5	13.7	41.1






#87
1,3-Dichlorobenzene
Concen: 17.302 ug/l
RT: 11.969 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion:146 Resp: 3780


Ion	Ratio	Lower	Upper
146	100		
111	44.2	22.2	66.6
148	64.6	31.9	95.7



Instrument :
MSVOA_X
ClientSampleId :
VX0407WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#88
1,4-Dichlorobenzene
Concen: 16.936 ug/l
RT: 12.043 min Scan# 1797
Delta R.T. 0.006 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion:146 Resp: 37519

Ion	Ratio	Lower	Upper
146	100		
111	46.5	20.9	62.8
148	63.8	31.5	94.5





#89
n-Butylbenzene
Concen: 17.230 ug/l
RT: 12.329 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34


Tgt Ion: 91 Resp: 6435


Ion	Ratio	Lower	Upper
91	100		
92	52.0	26.9	80.6
134	23.1	11.9	35.7

Instrument :
MSVOA_X
ClientSampleId :
VX0407WBS01

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025







#90
Hexachloroethane
Concen: 17.906 ug/l
RT: 12.536 min Scan# 1878
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion:117 Resp: 13254

Ion	Ratio	Lower	Upper
117	100		
201	61.2	31.2	93.6






#91
1,2-Dichlorobenzene
Concen: 17.492 ug/l
RT: 12.335 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion:146 Resp: 3800


Ion	Ratio	Lower	Upper
146	100		
111	44.9	22.9	68.8
148	64.3	31.9	95.5



Instrument :
MSVOA_X
ClientSampleId :
VX0407WBS01

Manual Integrations APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#92
1,2-Dibromo-3-Chloropropane
Concen: 17.821 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion: 75 Resp: 6802

Ion	Ratio	Lower	Upper
75	100		
155	70.9	34.9	104.7
157	90.6	44.9	134.7






#93
1,2,4-Trichlorobenzene
Concen: 16.447 ug/l
RT: 13.585 min Scan# 2
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion:180 Resp: 1988


Ion	Ratio	Lower	Upper
180	100		
182	94.3	47.9	143.6
145	33.1	17.9	53.8



Instrument :
MSVOA_X
ClientSampleId :
VX0407WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025




#94
Hexachlorobutadiene
Concen: 17.922 ug/l
RT: 13.719 min Scan# 2072
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion:225 Resp: 9339

Ion	Ratio	Lower	Upper
225	100		
223	63.8	31.1	93.3
227	59.9	32.2	96.6






#95
Naphthalene
Concen: 16.411 ug/l
RT: 13.774 min Scan# 2
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion:128 Resp: 7385


Ion	Ratio	Lower	Upper
128	100		
127	13.1	10.3	15.5
129	10.2	8.6	12.8



Instrument :
MSVOA_X
ClientSampleId :
VX0407WBS01

Manual Integrations APPROVED


Reviewed By :John Carlone 04/08/2025
Supervised By :Mahesh Dadoda 04/08/2025



#96
1,2,3-Trichlorobenzene
Concen: 16.591 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. 0.000 min
Lab File: VX045617.D
Acq: 07 Apr 2025 10:34

Tgt Ion:180 Resp: 21009

Ion	Ratio	Lower	Upper
180	100		
182	95.7	47.3	141.8
145	36.5	18.2	54.6



Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	
Project:	Former Schlumberger STC PTC Site D3868221	Date Received:	
Client Sample ID:	VX0404WBSD01	SDG No.:	Q1731
Lab Sample ID:	VX0404WBSD01	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group3
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045590.D	1		04/04/25 11:40	VX040425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	19.2		0.26	1.00	ug/L
75-35-4	1,1-Dichloroethene	19.0		0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	19.3		0.23	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	19.2		0.19	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	19.6		0.20	1.00	ug/L
71-43-2	Benzene	19.6		0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	20.7		0.22	1.00	ug/L
79-01-6	Trichloroethene	19.7		0.090	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	20.6		0.21	1.00	ug/L
127-18-4	Tetrachloroethene	19.4		0.23	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	53.3		70 (74) - 130 (125)	107%	SPK: 50
1868-53-7	Dibromofluoromethane	52.6		70 (75) - 130 (124)	105%	SPK: 50
2037-26-5	Toluene-d8	53.1		70 (86) - 130 (113)	106%	SPK: 50
460-00-4	4-Bromofluorobenzene	53.9		70 (77) - 130 (121)	108%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	92600	5.544			
540-36-3	1,4-Difluorobenzene	162000	6.757			
3114-55-4	Chlorobenzene-d5	145000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	65300	12.018			

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040425\
Data File : VX045590.D
Acq On : 04 Apr 2025 11:40
Operator : JC/MD
Sample : VX0404WBSD01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Quant Time: Apr 04 23:11:01 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VX0404WBSD01

Manual Integrations APPROVED

Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	5.544	168	92597	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	161617	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	145135	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	65264	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	90323	53.340	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 106.680%			
35) Dibromofluoromethane	5.379	113	60347	52.628	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 105.260%			
50) Toluene-d8	8.647	98	212584	53.115	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 106.220%			
62) 4-Bromofluorobenzene	11.079	95	78624	53.931	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 107.860%			
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.166	85	29372	21.210	ug/l	95
3) Chloromethane	1.307	50	26184	18.405	ug/l	97
4) Vinyl Chloride	1.374	62	25017	19.215	ug/l	99
5) Bromomethane	1.593	94	11656	18.882	ug/l	98
6) Chloroethane	1.672	64	13788	19.961	ug/l	99
7) Trichlorofluoromethane	1.874	101	39175	20.178	ug/l	98
8) Diethyl Ether	2.130	74	12792	19.625	ug/l	95
9) 1,1,2-Trichlorotrifluo...	2.319	101	23185	20.380	ug/l	97
10) Methyl Iodide	2.447	142	26079	18.378	ug/l	97
11) Tert butyl alcohol	2.971	59	22169	97.415	ug/l	98
12) 1,1-Dichloroethene	2.313	96	21129	19.007	ug/l	98
13) Acrolein	2.233	56	35155	112.178	ug/l	98
14) Allyl chloride	2.654	41	40403	19.156	ug/l	97
15) Acrylonitrile	3.062	53	73801	102.734	ug/l	98
16) Acetone	2.380	43	69871	100.485	ug/l	97
17) Carbon Disulfide	2.502	76	51231	18.660	ug/l	100
18) Methyl Acetate	2.703	43	33174	20.722	ug/l	100
19) Methyl tert-butyl Ether	3.111	73	76533	19.681	ug/l	97
20) Methylene Chloride	2.782	84	25529	19.611	ug/l	95
21) trans-1,2-Dichloroethene	3.087	96	21549	18.975	ug/l	99
22) Diisopropyl ether	3.757	45	82058	19.750	ug/l #	85
23) Vinyl Acetate	3.715	43	362593	101.009	ug/l	100
24) 1,1-Dichloroethane	3.605	63	45273	19.268	ug/l	99
25) 2-Butanone	4.556	43	105817	103.963	ug/l	98
26) 2,2-Dichloropropane	4.471	77	31149	20.518	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	26520	19.167	ug/l	98
28) Bromochloromethane	4.891	49	23299	20.519	ug/l	100
29) Tetrahydrofuran	5.001	42	66297	100.466	ug/l	100
30) Chloroform	5.086	83	48123	19.904	ug/l	98
31) Cyclohexane	5.458	56	40081	19.295	ug/l	98
32) 1,1,1-Trichloroethane	5.373	97	40071	19.587	ug/l	98
36) 1,1-Dichloropropene	5.690	75	30276	19.555	ug/l	99
37) Ethyl Acetate	4.715	43	40149	20.578	ug/l	100
38) Carbon Tetrachloride	5.666	117	34178	20.426	ug/l	93
39) Methylcyclohexane	7.373	83	39127	20.617	ug/l	95
40) Benzene	6.031	78	92456	19.553	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040425\
Data File : VX045590.D
Acq On : 04 Apr 2025 11:40
Operator : JC/MD
Sample : VX0404WBSD01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Quant Time: Apr 04 23:11:01 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VX0404WBSD01

Manual Integrations APPROVED

Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
41) Methacrylonitrile	4.922	41	22686	22.037	ug/l		98
42) 1,2-Dichloroethane	6.086	62	40492	20.733	ug/l		100
43) Isopropyl Acetate	6.342	43	60398	20.424	ug/l		99
44) Trichloroethene	7.123	130	22106	19.669	ug/l		99
45) 1,2-Dichloropropane	7.427	63	24079	20.409	ug/l		100
46) Dibromomethane	7.580	93	18760	20.714	ug/l		99
47) Bromodichloromethane	7.818	83	36921	20.413	ug/l		95
48) Methyl methacrylate	7.690	41	30841	20.199	ug/l		99
49) 1,4-Dioxane	7.659	88	11859	429.294	ug/l		98
51) 4-Methyl-2-Pentanone	8.574	43	207996	106.101	ug/l		100
52) Toluene	8.714	92	57346	20.042	ug/l		99
53) t-1,3-Dichloropropene	8.976	75	30852	19.215	ug/l		99
54) cis-1,3-Dichloropropene	8.366	75	35055	20.625	ug/l		98
55) 1,1,2-Trichloroethane	9.147	97	23535	20.642	ug/l		96
56) Ethyl methacrylate	9.116	69	36239	20.538	ug/l		98
57) 1,3-Dichloropropane	9.305	76	41392	20.848	ug/l		100
58) 2-Chloroethyl Vinyl ether	8.238	63	94302	105.646	ug/l		100
59) 2-Hexanone	9.427	43	153410	105.677	ug/l		100
60) Dibromochloromethane	9.519	129	25740	20.708	ug/l		97
61) 1,2-Dibromoethane	9.604	107	23938	20.740	ug/l		99
64) Tetrachloroethene	9.269	164	19869	19.390	ug/l		98
65) Chlorobenzene	10.073	112	62313	20.093	ug/l		99
66) 1,1,1,2-Tetrachloroethane	10.159	131	21067	19.566	ug/l		99
67) Ethyl Benzene	10.189	91	111795	20.127	ug/l		99
68) m/p-Xylenes	10.299	106	81202	40.195	ug/l		99
69) o-Xylene	10.640	106	40016	20.108	ug/l		98
70) Styrene	10.652	104	66944	20.381	ug/l		99
71) Bromoform	10.799	173	15669	19.490	ug/l	#	96
73) Isopropylbenzene	10.957	105	107035	20.475	ug/l		100
74) N-amyl acetate	10.841	43	49971	19.983	ug/l		99
75) 1,1,2,2-Tetrachloroethane	11.207	83	36342	19.790	ug/l		98
76) 1,2,3-Trichloropropane	11.238	75	31836m	19.998	ug/l		
77) Bromobenzene	11.195	156	23927	19.809	ug/l		100
78) n-propylbenzene	11.299	91	122355	20.320	ug/l		99
79) 2-Chlorotoluene	11.360	91	76626	19.906	ug/l		100
80) 1,3,5-Trimethylbenzene	11.451	105	89670	20.743	ug/l		99
81) trans-1,4-Dichloro-2-b...	11.018	75	8567	18.264	ug/l		93
82) 4-Chlorotoluene	11.451	91	87060	20.284	ug/l		99
83) tert-Butylbenzene	11.713	119	87480	20.440	ug/l		97
84) 1,2,4-Trimethylbenzene	11.750	105	89924	20.726	ug/l		99
85) sec-Butylbenzene	11.890	105	111357	21.184	ug/l		99
86) p-Isopropyltoluene	12.006	119	90417	20.864	ug/l		99
87) 1,3-Dichlorobenzene	11.969	146	43316	19.701	ug/l		99
88) 1,4-Dichlorobenzene	12.036	146	42823	19.211	ug/l		97
89) n-Butylbenzene	12.329	91	77794	20.701	ug/l		98
90) Hexachloroethane	12.536	117	15051	20.208	ug/l		99
91) 1,2-Dichlorobenzene	12.335	146	43307	19.811	ug/l		99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	8021	20.885	ug/l		97
93) 1,2,4-Trichlorobenzene	13.585	180	23587	19.385	ug/l		99
94) Hexachlorobutadiene	13.719	225	10565	20.149	ug/l		98
95) Naphthalene	13.774	128	87923	19.417	ug/l		100
96) 1,2,3-Trichlorobenzene	13.957	180	24006	18.841	ug/l		98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040425\
Data File : VX045590.D
Acq On : 04 Apr 2025 11:40
Operator : JC/MD
Sample : VX0404WBSD01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Quant Time: Apr 04 23:11:01 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VX0404WBSD01

Manual Integrations APPROVED

Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040425\
Data File : VX045590.D
Acq On : 04 Apr 2025 11:40
Operator : JC/MD
Sample : VX0404WBSD01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1


Quant Time: Apr 04 23:11:01 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VX0404WBSD01

Manual Integrations APPROVED

Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025






#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 7
Delta R.T. -0.006 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion:168 Resp: 9259


Ion	Ratio	Lower	Upper
168	100		
99	74.3	52.3	78.5



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBSD01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#2
Dichlorodifluoromethane
Concen: 21.210 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion: 85 Resp: 29372

Ion	Ratio	Lower	Upper
85	100		
87	35.1	16.1	48.2






#3
Chloromethane
Concen: 18.405 ug/l
RT: 1.307 min Scan# 3
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion: 50 Resp: 2618


Ion	Ratio	Lower	Upper
50	100		
52	34.3	25.9	38.9



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBSD01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#4
Vinyl Chloride
Concen: 19.215 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion: 62 Resp: 25017

Ion	Ratio	Lower	Upper
62	100		
64	32.7	25.5	38.3






#5
Bromomethane
Concen: 18.882 ug/l
RT: 1.593 min Scan# 8
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion: 94 Resp: 1165


Ion	Ratio	Lower	Upper
94	100		
96	95.2	74.8	112.2



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBSD01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#6
Chloroethane
Concen: 19.961 ug/l
RT: 1.672 min Scan# 96
Delta R.T. 0.006 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion: 64 Resp: 13788

Ion	Ratio	Lower	Upper
64	100		
66	33.6	26.3	39.5






#7
Trichlorofluoromethane
Concen: 20.178 ug/l
RT: 1.874 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion:101 Resp: 3917


Ion	Ratio	Lower	Upper
101	100		
103	62.5	51.1	76.7



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBSD01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#8
Diethyl Ether
Concen: 19.625 ug/l
RT: 2.130 min Scan# 171
Delta R.T. -0.006 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion: 74 Resp: 12792

Ion	Ratio	Lower	Upper
74	100		
45	111.2	53.0	159.1





#9
1,1,2-Trichlorotrifluoroethane
Concen: 20.380 ug/l
RT: 2.319 min Scan# 2
Delta R.T. -0.006 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40


Tgt Ion:101 Resp: 2318


Ion	Ratio	Lower	Upper
101	100		
85	46.1	38.9	58.3
151	69.9	57.5	86.3

Instrument :
MSVOA_X
ClientSampleId :
VX0404WBSD01

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025







#10
Methyl Iodide
Concen: 18.378 ug/l
RT: 2.447 min Scan# 223
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion:142 Resp: 26079

Ion	Ratio	Lower	Upper
142	100		
127	52.9	40.2	60.4
141	15.1	11.7	17.5






#11
Tert butyl alcohol
Concen: 97.415 ug/l
RT: 2.971 min Scan# 3
Delta R.T. -0.006 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion: 59 Resp: 2216


Ion	Ratio	Lower	Upper
59	100		
57	11.4	8.6	12.8



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBSD01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#12
1,1-Dichloroethene
Concen: 19.007 ug/l
RT: 2.313 min Scan# 201
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion: 96 Resp: 21129

Ion	Ratio	Lower	Upper
96	100		
61	185.3	145.7	218.5
98	62.2	51.6	77.4





#13
Acrolein
Concen: 112.178 ug/l
RT: 2.233 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40


Tgt Ion: 56 Resp: 3515


Ion	Ratio	Lower	Upper
56	100		
55	71.9	56.2	84.2

Instrument :
MSVOA_X
ClientSampleId :
VX0404WBSD01

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025







#14
Allyl chloride
Concen: 19.156 ug/l
RT: 2.654 min Scan# 257
Delta R.T. -0.006 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion: 41 Resp: 40403

Ion	Ratio	Lower	Upper
41	100		
39	78.2	60.2	90.2
76	30.9	24.6	36.8





#15
Acrylonitrile
Concen: 102.734 ug/l
RT: 3.062 min Scan# 3
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40


Tgt Ion: 53 Resp: 7380


Ion	Ratio	Lower	Upper
53	100		
52	82.2	67.4	101.2
51	37.5	29.9	44.9

Instrument :
MSVOA_X
ClientSampleId :
VX0404WBSD01

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025







#16
Acetone
Concen: 100.485 ug/l
RT: 2.380 min Scan# 212
Delta R.T. -0.006 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion: 43 Resp: 69871

Ion	Ratio	Lower	Upper
43	100		
58	26.6	22.3	33.5






#17
Carbon Disulfide
Concen: 18.660 ug/l
RT: 2.502 min Scan# 2
Delta R.T. -0.006 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion: 76 Resp: 5123


Ion	Ratio	Lower	Upper
76	100		
78	8.9	7.0	10.6



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBSD01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#18
Methyl Acetate
Concen: 20.722 ug/l
RT: 2.703 min Scan# 265
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion: 43 Resp: 33174

Ion	Ratio	Lower	Upper
43	100		
74	21.8	17.5	26.3





#19
Methyl tert-butyl Ether
Concen: 19.681 ug/l
RT: 3.111 min Scan# 3
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40


Tgt Ion: 73 Resp: 7653


Ion	Ratio	Lower	Upper
73	100		
57	20.8	17.6	26.4

Instrument :
MSVOA_X
ClientSampleId :
VX0404WBSD01

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025







#20
Methylene Chloride
Concen: 19.611 ug/l
RT: 2.782 min Scan# 278
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion: 84 Resp: 25529

Ion	Ratio	Lower	Upper
84	100		
49	140.7	107.5	161.3
51	38.2	33.0	49.6
86	65.9	50.1	75.1






#21
trans-1,2-Dichloroethene
Concen: 18.975 ug/l
RT: 3.087 min Scan# 3
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion: 96 Resp: 2154


Ion	Ratio	Lower	Upper
96	100		
61	152.1	121.0	181.6
98	62.5	49.7	74.5



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBSD01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#22
Diisopropyl ether
Concen: 19.750 ug/l
RT: 3.757 min Scan# 438
Delta R.T. -0.006 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion: 45 Resp: 82058

Ion	Ratio	Lower	Upper
45	100		
43	91.5	59.3	88.9#
87	23.7	20.4	30.6
59	10.9	9.0	13.6





#23
Vinyl Acetate
Concen: 101.009 ug/l
RT: 3.715 min Scan# 4
Delta R.T. -0.006 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40


Tgt Ion: 43 Resp: 36259


Ion	Ratio	Lower	Upper
43	100		
86	9.4	7.5	11.3

Instrument :
MSVOA_X
ClientSampleId :
VX0404WBSD01

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025







#24
1,1-Dichloroethane
Concen: 19.268 ug/l
RT: 3.605 min Scan# 413
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion: 63 Resp: 45273

Ion	Ratio	Lower	Upper
63	100		
98	6.5	3.4	10.1
100	3.9	2.2	6.6






#25
2-Butanone
Concen: 103.963 ug/l
RT: 4.556 min Scan# 5
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion: 43 Resp: 10581


Ion	Ratio	Lower	Upper
43	100		
72	22.1	18.6	28.0



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBSD01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#26
2,2-Dichloropropane
Concen: 20.518 ug/l
RT: 4.471 min Scan# 555
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion: 77 Resp: 31149

Ion	Ratio	Lower	Upper
77	100		
97	21.4	10.6	31.8






#27
cis-1,2-Dichloroethene
Concen: 19.167 ug/l
RT: 4.483 min Scan# 5
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion: 96 Resp: 2652


Ion	Ratio	Lower	Upper
96	100		
61	158.3	0.0	313.2
98	65.6	0.0	126.4



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBSD01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#28
Bromochloromethane
Concen: 20.519 ug/l
RT: 4.891 min Scan# 624
Delta R.T. -0.006 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion: 49 Resp: 23299

Ion	Ratio	Lower	Upper
49	100		
129	1.5	0.0	3.8
130	68.5	54.6	81.8






#29
Tetrahydrofuran
Concen: 100.466 ug/l
RT: 5.001 min Scan# 64
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion: 42 Resp: 6629


Ion	Ratio	Lower	Upper
42	100		
72	41.9	33.4	50.0
71	39.0	31.4	47.0



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBSD01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#30
Chloroform
Concen: 19.904 ug/l
RT: 5.086 min Scan# 656
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion: 83 Resp: 48123

Ion	Ratio	Lower	Upper
83	100		
85	67.6	52.8	79.2






#31
Cyclohexane
Concen: 19.295 ug/l
RT: 5.458 min Scan# 7
Delta R.T. -0.006 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion: 56 Resp: 4008


Ion	Ratio	Lower	Upper
56	100		
69	32.4	23.6	35.4
84	83.6	67.7	101.5



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBSD01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#32
1,1,1-Trichloroethane
Concen: 19.587 ug/l
RT: 5.373 min Scan# 703
Delta R.T. -0.006 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion: 97 Resp: 40071

Ion	Ratio	Lower	Upper
97	100		
99	63.4	51.7	77.5
61	52.9	41.0	61.4






#33
1,2-Dichloroethane-d4
Concen: 53.340 ug/l
RT: 5.946 min Scan# 7
Delta R.T. -0.006 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion: 65 Resp: 9032


Ion	Ratio	Lower	Upper
65	100		
67	48.8	0.0	99.0



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBSD01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.757 min Scan# 930
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion:114 Resp: 161617

Ion	Ratio	Lower	Upper
114	100		
63	23.7	0.0	46.8
88	17.9	0.0	35.4






#35
Dibromofluoromethane
Concen: 52.628 ug/l
RT: 5.379 min Scan# 7
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion:113 Resp: 6034


Ion	Ratio	Lower	Upper
113	100		
111	104.4	81.8	122.6
192	17.0	13.8	20.6



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBSD01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#36
1,1-Dichloropropene
Concen: 19.555 ug/l
RT: 5.690 min Scan# 755
Delta R.T. 0.006 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion: 75 Resp: 30276

Ion	Ratio	Lower	Upper
75	100		
110	33.0	16.9	50.6
77	31.2	25.0	37.4






#37
Ethyl Acetate
Concen: 20.578 ug/l
RT: 4.715 min Scan# 5
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion: 43 Resp: 4014


Ion	Ratio	Lower	Upper
43	100		
61	12.7	9.9	14.9
70	9.5	7.5	11.3



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBSD01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#38
Carbon Tetrachloride
Concen: 20.426 ug/l
RT: 5.666 min Scan# 751
Delta R.T. -0.006 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion:117 Resp: 34178

Ion	Ratio	Lower	Upper
117	100		
119	87.3	76.9	115.3
121	30.1	24.2	36.2






#39
Methylcyclohexane
Concen: 20.617 ug/l
RT: 7.373 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion: 83 Resp: 3912


Ion	Ratio	Lower	Upper
83	100		
55	86.3	64.9	97.3
98	49.0	37.4	56.0



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBSD01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#40
Benzene
Concen: 19.553 ug/l
RT: 6.031 min Scan# 811
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion: 78 Resp: 92456

Ion	Ratio	Lower	Upper
78	100		
77	22.9	19.0	28.6






#41
Methacrylonitrile
Concen: 22.037 ug/l
RT: 4.922 min Scan# 6
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion: 41 Resp: 2268


Ion	Ratio	Lower	Upper
41	100		
39	56.9	47.0	70.6
67	64.0	53.2	79.8
52	31.1	25.0	37.6



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBSD01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#42
1,2-Dichloroethane
Concen: 20.733 ug/l
RT: 6.086 min Scan# 820
Delta R.T. 0.006 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion: 62 Resp: 40492

Ion	Ratio	Lower	Upper
62	100		
98	7.5	0.0	15.2





#43
Isopropyl Acetate
Concen: 20.424 ug/l
RT: 6.342 min Scan# 8
Delta R.T. 0.006 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40


Tgt Ion: 43 Resp: 6039


Ion	Ratio	Lower	Upper
43	100		
61	16.9	14.1	21.1
87	11.4	9.2	13.8

Instrument :
MSVOA_X
ClientSampleId :
VX0404WBSD01

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025







#44
Trichloroethene
Concen: 19.669 ug/l
RT: 7.123 min Scan# 990
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion:130 Resp: 22106

Ion	Ratio	Lower	Upper
130	100		
95	101.8	0.0	202.0






#45
1,2-Dichloropropane
Concen: 20.409 ug/l
RT: 7.427 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion: 63 Resp: 2407


Ion	Ratio	Lower	Upper
63	100		
65	32.0	25.8	38.6



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBSD01

Manual Integrations APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#46
Dibromomethane
Concen: 20.714 ug/l
RT: 7.580 min Scan# 1065
Delta R.T. 0.006 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion: 93 Resp: 18760

Ion	Ratio	Lower	Upper
93	100		
95	81.8	64.8	97.2
174	81.0	65.4	98.0






#47
Bromodichloromethane
Concen: 20.413 ug/l
RT: 7.818 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion: 83 Resp: 3692


Ion	Ratio	Lower	Upper
83	100		
85	66.1	49.7	74.5
127	8.7	6.1	9.1



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBSD01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#48
Methyl methacrylate
Concen: 20.199 ug/l
RT: 7.690 min Scan# 1083
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion: 41 Resp: 30841

Ion	Ratio	Lower	Upper
41	100		
69	73.9	58.2	87.2
39	67.3	53.6	80.4






#49
1,4-Dioxane
Concen: 429.294 ug/l
RT: 7.659 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion: 88 Resp: 1185


Ion	Ratio	Lower	Upper
88	100		
43	42.6	32.2	48.4
58	71.8	58.6	88.0



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBSD01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#50
Toluene-d8
Concen: 53.115 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion: 98 Resp: 212584

Ion	Ratio	Lower	Upper
98	100		
100	65.6	52.2	78.4






#51
4-Methyl-2-Pentanone
Concen: 106.101 ug/l
RT: 8.574 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion: 43 Resp: 20799


Ion	Ratio	Lower	Upper
43	100		
58	35.4	28.5	42.7



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBSD01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#52
Toluene
Concen: 20.042 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion: 92 Resp: 57346

Ion	Ratio	Lower	Upper
92	100		
91	172.7	138.7	208.1






#53
t-1,3-Dichloropropene
Concen: 19.215 ug/l
RT: 8.976 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion: 75 Resp: 3085


Ion	Ratio	Lower	Upper
75	100		
77	32.1	25.4	38.0



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBSD01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#54
cis-1,3-Dichloropropene
Concen: 20.625 ug/l
RT: 8.366 min Scan# 1194
Delta R.T. 0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion: 75 Resp: 35055

Ion	Ratio	Lower	Upper
75	100		
77	30.4	24.6	36.8
39	62.6	48.4	72.6






#55
1,1,2-Trichloroethane
Concen: 20.642 ug/l
RT: 9.147 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion: 97 Resp: 2353


Ion	Ratio	Lower	Upper
97	100		
83	85.8	71.8	107.8
85	56.9	47.0	70.6
99	63.6	48.6	73.0



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBSD01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#56
Ethyl methacrylate
Concen: 20.538 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. 0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion: 69 Resp: 36239

Ion	Ratio	Lower	Upper
69	100		
41	79.2	61.3	91.9
39	49.3	40.2	60.2






#57
1,3-Dichloropropane
Concen: 20.848 ug/l
RT: 9.305 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion: 76 Resp: 4139


Ion	Ratio	Lower	Upper
76	100		
78	32.5	26.0	39.0



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBSD01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#58
2-Chloroethyl Vinyl ether
Concen: 105.646 ug/l
RT: 8.238 min Scan# 1173
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion: 63 Resp: 94302

Ion	Ratio	Lower	Upper
63	100		
106	26.5	21.0	31.6






#59
2-Hexanone
Concen: 105.677 ug/l
RT: 9.427 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion: 43 Resp: 15341


Ion	Ratio	Lower	Upper
43	100		
58	49.5	24.9	74.6



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBSD01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#60
Dibromochloromethane
Concen: 20.708 ug/l
RT: 9.519 min Scan# 1383
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion:129 Resp: 25740

Ion	Ratio	Lower	Upper
129	100		
127	76.5	39.4	118.2






#61
1,2-Dibromoethane
Concen: 20.740 ug/l
RT: 9.604 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion:107 Resp: 2393


Ion	Ratio	Lower	Upper
107	100		
109	94.0	75.7	113.5



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBSD01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#62
4-Bromofluorobenzene
Concen: 53.931 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion: 95 Resp: 78624

Ion	Ratio	Lower	Upper
95	100		
174	68.0	0.0	135.8
176	64.4	0.0	131.4






#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 14
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion:117 Resp: 14513


Ion	Ratio	Lower	Upper
117	100		
82	61.6	49.2	73.8
119	30.6	25.1	37.7



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBSD01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#64
Tetrachloroethene
Concen: 19.390 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion:164 Resp: 19869

Ion	Ratio	Lower	Upper
164	100		
166	125.8	103.8	155.8
129	106.3	84.3	126.5
131	98.6	80.9	121.3






#65
Chlorobenzene
Concen: 20.093 ug/l
RT: 10.073 min Scan# 14
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion:112 Resp: 6231


Ion	Ratio	Lower	Upper
112	100		
114	31.7	24.7	37.1



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBSD01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#66
1,1,1,2-Tetrachloroethane
Concen: 19.566 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion:131 Resp: 21067

Ion	Ratio	Lower	Upper
131	100		
133	94.8	47.5	142.5
119	64.0	32.9	98.6





#67
Ethyl Benzene
Concen: 20.127 ug/l
RT: 10.189 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40


Tgt Ion: 91 Resp: 11179


Ion	Ratio	Lower	Upper
91	100		
106	29.1	23.7	35.5

Instrument :
MSVOA_X
ClientSampleId :
VX0404WBSD01

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025







#68
m/p-Xylenes
Concen: 40.195 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion:106 Resp: 81202

Ion	Ratio	Lower	Upper
106	100		
91	217.9	172.6	258.8






#69
o-Xylene
Concen: 20.108 ug/l
RT: 10.640 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion:106 Resp: 4001


Ion	Ratio	Lower	Upper
106	100		
91	229.7	113.6	340.6



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBSD01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#70
Styrene
Concen: 20.381 ug/l
RT: 10.652 min Scan# 1569
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion:104 Resp: 66944

Ion	Ratio	Lower	Upper
104	100		
78	57.6	45.4	68.2
103	54.5	43.8	65.8






#71
Bromoform
Concen: 19.490 ug/l
RT: 10.799 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion:173 Resp: 1566


Ion	Ratio	Lower	Upper
173	100		
175	49.6	23.4	70.2
254	0.0	0.0	0.0



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBSD01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion:152 Resp: 65264

Ion	Ratio	Lower	Upper
152	100		
115	77.5	46.9	140.7
150	164.0	0.0	349.4






#73
Isopropylbenzene
Concen: 20.475 ug/l
RT: 10.957 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion:105 Resp: 10703


Ion	Ratio	Lower	Upper
105	100		
120	25.4	12.7	38.0



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBSD01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#74
N-amyl acetate
Concen: 19.983 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion: 43 Resp: 49971

Ion	Ratio	Lower	Upper
43	100		
70	37.0	30.1	45.1
55	23.2	18.3	27.5
61	21.0	17.5	26.3






#75
1,1,2,2-Tetrachloroethane
Concen: 19.790 ug/l
RT: 11.207 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion: 83 Resp: 3634


Ion	Ratio	Lower	Upper
83	100		
131	9.6	4.7	14.1
85	66.2	32.4	97.2



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBSD01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#76
1,2,3-Trichloropropane
Concen: 19.998 ug/l m
RT: 11.238 min Scan# 1665
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion: 75 Resp: 31836

Ion	Ratio	Lower	Upper
75	100		
77	40.4	20.0	59.9






#77
Bromobenzene
Concen: 19.809 ug/l
RT: 11.195 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion:156 Resp: 2392


Ion	Ratio	Lower	Upper
156	100		
77	185.7	92.6	277.8
158	96.7	48.1	144.4



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBSD01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#78
n-propylbenzene
Concen: 20.320 ug/l
RT: 11.299 min Scan# 1675
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion: 91 Resp: 122355

Ion	Ratio	Lower	Upper
91	100		
120	22.1	10.8	32.3





#79
2-Chlorotoluene
Concen: 19.906 ug/l
RT: 11.360 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40


Tgt Ion: 91 Resp: 7662


Ion	Ratio	Lower	Upper
91	100		
126	31.8	15.8	47.4

Instrument :
MSVOA_X
ClientSampleId :
VX0404WBSD01

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025







#80
1,3,5-Trimethylbenzene
Concen: 20.743 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion:105 Resp: 89670

Ion	Ratio	Lower	Upper
105	100		
120	46.9	23.0	69.0





#81
trans-1,4-Dichloro-2-butene
Concen: 18.264 ug/l
RT: 11.018 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40


Tgt Ion: 75 Resp: 856


Ion	Ratio	Lower	Upper
75	100		
53	138.1	101.8	152.6
89	47.8	37.2	55.8

Instrument :
MSVOA_X
ClientSampleId :
VX0404WBSD01

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025







#82
4-Chlorotoluene
Concen: 20.284 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion: 91 Resp: 87060

Ion	Ratio	Lower	Upper
91	100		
126	27.0	13.3	39.9






#83
tert-Butylbenzene
Concen: 20.440 ug/l
RT: 11.713 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion:119 Resp: 8748


Ion	Ratio	Lower	Upper
119	100		
91	67.6	32.4	97.2
134	23.1	11.5	34.5



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBSD01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#84
1,2,4-Trimethylbenzene
Concen: 20.726 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion:105 Resp: 89924

Ion	Ratio	Lower	Upper
105	100		
120	42.0	21.4	64.3






#85
sec-Butylbenzene
Concen: 21.184 ug/l
RT: 11.890 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion:105 Resp: 11135


Ion	Ratio	Lower	Upper
105	100		
134	18.7	9.6	28.6



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBSD01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#86
p-Isopropyltoluene
Concen: 20.864 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion:119 Resp: 90417

Ion	Ratio	Lower	Upper
119	100		
134	24.9	12.6	37.8
91	27.1	13.7	41.1






#87
1,3-Dichlorobenzene
Concen: 19.701 ug/l
RT: 11.969 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion:146 Resp: 4331


Ion	Ratio	Lower	Upper
146	100		
111	44.0	22.2	66.6
148	63.4	31.9	95.7



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBSD01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#88
1,4-Dichlorobenzene
Concen: 19.211 ug/l
RT: 12.036 min Scan# 1796
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion:146 Resp: 42823

Ion	Ratio	Lower	Upper
146	100		
111	44.4	20.9	62.8
148	65.2	31.5	94.5






#89
n-Butylbenzene
Concen: 20.701 ug/l
RT: 12.329 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion: 91 Resp: 7779


Ion	Ratio	Lower	Upper
91	100		
92	52.2	26.9	80.6
134	23.2	11.9	35.7



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBSD01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#90
Hexachloroethane
Concen: 20.208 ug/l
RT: 12.536 min Scan# 1878
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion:117 Resp: 15051

Ion	Ratio	Lower	Upper
117	100		
201	61.3	31.2	93.6






#91
1,2-Dichlorobenzene
Concen: 19.811 ug/l
RT: 12.335 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion:146 Resp: 4330


Ion	Ratio	Lower	Upper
146	100		
111	45.1	22.9	68.8
148	64.4	31.9	95.5



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBSD01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#92
1,2-Dibromo-3-Chloropropane
Concen: 20.885 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion: 75 Resp: 8021

Ion	Ratio	Lower	Upper
75	100		
155	70.6	34.9	104.7
157	84.7	44.9	134.7






#93
1,2,4-Trichlorobenzene
Concen: 19.385 ug/l
RT: 13.585 min Scan# 2
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion:180 Resp: 2358


Ion	Ratio	Lower	Upper
180	100		
182	96.3	47.9	143.6
145	33.8	17.9	53.8



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBSD01

Manual Integrations
APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025




#94
Hexachlorobutadiene
Concen: 20.149 ug/l
RT: 13.719 min Scan# 2072
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion:225 Resp: 10565

Ion	Ratio	Lower	Upper
225	100		
223	63.9	31.1	93.3
227	62.3	32.2	96.6






#95
Naphthalene
Concen: 19.417 ug/l
RT: 13.774 min Scan# 2
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion:128 Resp: 8792


Ion	Ratio	Lower	Upper
128	100		
127	12.7	10.3	15.5
129	10.8	8.6	12.8



Instrument :
MSVOA_X
ClientSampleId :
VX0404WBSD01

Manual Integrations APPROVED


Reviewed By :John Carlone 04/07/2025
Supervised By :Mahesh Dadoda 04/07/2025



#96
1,2,3-Trichlorobenzene
Concen: 18.841 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. -0.000 min
Lab File: VX045590.D
Acq: 04 Apr 2025 11:40

Tgt Ion:180 Resp: 24006

Ion	Ratio	Lower	Upper
180	100		
182	96.9	47.3	141.8
145	37.1	18.2	54.6



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	VX040225	Instrument	MSVOA_x

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDICC001	VX045525.D	1,2,3-Trichloropropane	Amit	4/2/2025 2:14:20 PM	MMDadoda	4/2/2025 2:16:05 PM	Peak Integrated by Software
VSTDICC001	VX045525.D	1,4-Dichlorobenzene	Amit	4/2/2025 2:14:20 PM	MMDadoda	4/2/2025 2:16:05 PM	Peak Integrated by Software
VSTDICC001	VX045525.D	1,4-Dioxane	Amit	4/2/2025 2:14:20 PM	MMDadoda	4/2/2025 2:16:05 PM	Peak Integrated by Software
VSTDICC001	VX045525.D	Ethyl methacrylate	Amit	4/2/2025 2:14:20 PM	MMDadoda	4/2/2025 2:16:05 PM	Peak Integrated by Software
VSTDICC001	VX045525.D	Methacrylonitrile	Amit	4/2/2025 2:14:20 PM	MMDadoda	4/2/2025 2:16:05 PM	Peak Integrated by Software
VSTDICC005	VX045526.D	1,2,3-Trichloropropane	Amit	4/2/2025 2:14:22 PM	MMDadoda	4/2/2025 2:16:07 PM	Peak Integrated by Software
VSTDICC020	VX045527.D	1,2,3-Trichloropropane	Amit	4/2/2025 2:14:25 PM	MMDadoda	4/2/2025 2:16:09 PM	Peak Integrated by Software
VSTDICCC050	VX045528.D	1,2,3-Trichloropropane	Amit	4/2/2025 2:14:26 PM	MMDadoda	4/2/2025 2:16:11 PM	Peak Integrated by Software
VSTDICC100	VX045529.D	1,2,3-Trichloropropane	Amit	4/2/2025 2:14:28 PM	MMDadoda	4/2/2025 2:16:15 PM	Peak Integrated by Software
VSTDICC150	VX045530.D	1,2,3-Trichloropropane	Amit	4/2/2025 2:14:30 PM	MMDadoda	4/2/2025 2:16:46 PM	Peak Integrated by Software
VSTDICV050	VX045532.D	1,2,3-Trichloropropane	Amit	4/2/2025 2:14:32 PM	MMDadoda	4/2/2025 2:17:50 PM	Peak Integrated by Software
VSTDCCC050	VX045534.D	1,2,3-Trichloropropane	JOHN	4/3/2025 10:04:10 AM	MMDadoda	4/3/2025 1:12:02 PM	Peak Integrated by Software
VSTDCCC050	VX045557.D	1,2,3-Trichloropropane	JOHN	4/3/2025 10:04:58 AM	MMDadoda	4/3/2025 1:12:08 PM	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	VX040225	Instrument	MSVOA_x

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	vx040425	Instrument	MSVOA_x

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDCCC050	VX045586.D	1,2,3-Trichloropropane	JOHN	4/7/2025 9:56:35 AM	MMDadoda	4/7/2025 11:32:09 AM	Peak Integrated by Software
VX0404WBS01	VX045589.D	1,2,3-Trichloropropane	JOHN	4/7/2025 9:56:40 AM	MMDadoda	4/7/2025 11:32:13 AM	Peak Integrated by Software
VX0404WBSD01	VX045590.D	1,2,3-Trichloropropane	JOHN	4/7/2025 9:56:45 AM	MMDadoda	4/7/2025 11:32:20 AM	Peak Integrated by Software
VSTDCCC050	VX045612.D	1,2,3-Trichloropropane	JOHN	4/7/2025 9:57:10 AM	MMDadoda	4/7/2025 11:32:29 AM	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	vx040725	Instrument	MSVOA_x

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDCCC050	VX045614.D	1,2,3-Trichloropropane	JOHN	4/8/2025 9:58:15 AM	MMDadoda	4/8/2025 2:05:07 PM	Peak Integrated by Software
VX0407WBS01	VX045617.D	1,2,3-Trichloropropane	JOHN	4/8/2025 9:58:20 AM	MMDadoda	4/8/2025 2:05:09 PM	Peak Integrated by Software
Q1731-01	VX045620.D	Trichloroethene	JOHN	4/7/2025 2:51:22 PM	MMDadoda	4/8/2025 2:05:15 PM	Peak Integrated by Software
VSTDCCC050	VX045633.D	1,2,3-Trichloropropane	JOHN	4/8/2025 9:58:32 AM	MMDadoda	4/8/2025 2:05:14 PM	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QCBatch ID # VX040225

Review By	John Carlone	Review On	4/2/2025 9:42:46 AM		
Supervise By	Amit Patel	Supervise On	4/2/2025 2:14:39 PM		
SubDirectory	VX040225	HP Acquire Method		HP Processing Method	82X040225W.M

STD. NAME	STD REF.#
Tune/Reschk	VP133559,VP133563
Initial Calibration Stds	VP133569,VP133570,VP133571,VP133572,VP133573,VP133574
CCC	VP133564,VP133565
Internal Standard/PEM	
ICV/I.BLK	VP133575
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr #	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VX045524.D	01 Apr 2025 16:15	JC/MD	Ok
2	VSTDICC001	VX045525.D	01 Apr 2025 17:06	JC/MD	Ok,M
3	VSTDICC005	VX045526.D	01 Apr 2025 17:29	JC/MD	Ok,M
4	VSTDICC020	VX045527.D	01 Apr 2025 17:52	JC/MD	Ok,M
5	VSTDICCC050	VX045528.D	01 Apr 2025 18:15	JC/MD	Ok,M
6	VSTDICC100	VX045529.D	01 Apr 2025 18:38	JC/MD	Ok,M
7	VSTDICC150	VX045530.D	01 Apr 2025 19:02	JC/MD	Ok,M
8	IBLK	VX045531.D	01 Apr 2025 19:25	JC/MD	Ok
9	VSTDICV050	VX045532.D	01 Apr 2025 19:48	JC/MD	Ok,M
10	BFB	VX045533.D	02 Apr 2025 09:30	JC/MD	Ok
11	VSTDCCC050	VX045534.D	02 Apr 2025 10:02	JC/MD	Ok,M
12	VX0402MBL01	VX045535.D	02 Apr 2025 10:30	JC/MD	Ok
13	VX0402WBL01	VX045536.D	02 Apr 2025 10:53	JC/MD	Ok
14	VX0402WBS01	VX045537.D	02 Apr 2025 11:16	JC/MD	Ok,M
15	VX0402WBSD01	VX045538.D	02 Apr 2025 11:44	JC/MD	Ok,M
16	Q1697-01	VX045539.D	02 Apr 2025 12:07	JC/MD	Dilution
17	Q1697-02	VX045540.D	02 Apr 2025 12:30	JC/MD	Dilution
18	IBLK	VX045541.D	02 Apr 2025 12:54	JC/MD	Ok
19	Q1697-01DL	VX045542.D	02 Apr 2025 13:21	JC/MD	Ok
20	Q1697-02DL	VX045543.D	02 Apr 2025 13:44	JC/MD	Ok
21	Q1697-03	VX045544.D	02 Apr 2025 14:07	JC/MD	Not Ok

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QCBatch ID # VX040225

Review By	John Carlone	Review On	4/2/2025 9:42:46 AM		
Supervise By	Amit Patel	Supervise On	4/2/2025 2:14:39 PM		
SubDirectory	VX040225	HP Acquire Method		HP Processing Method	82X040225W.M

STD. NAME	STD REF.#
Tune/Reschk	VP133559,VP133563
Initial Calibration Stds	VP133569,VP133570,VP133571,VP133572,VP133573,VP133574
CCC	VP133564,VP133565
Internal Standard/PEM	
ICV/I.BLK	VP133575
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

22	Q1697-04	VX045545.D	02 Apr 2025 14:31	JC/MD	Not Ok
23	Q1697-05	VX045546.D	02 Apr 2025 14:54	JC/MD	Not Ok
24	Q1697-07	VX045547.D	02 Apr 2025 15:18	JC/MD	Not Ok
25	IBLK	VX045548.D	02 Apr 2025 15:41	JC/MD	Ok
26	Q1697-03	VX045549.D	02 Apr 2025 16:04	JC/MD	Ok
27	Q1697-04	VX045550.D	02 Apr 2025 16:27	JC/MD	Ok
28	Q1697-05	VX045551.D	02 Apr 2025 16:51	JC/MD	Ok
29	Q1697-06	VX045552.D	02 Apr 2025 17:14	JC/MD	ReRun
30	Q1623-01	VX045553.D	02 Apr 2025 17:37	JC/MD	Ok
31	Q1623-02	VX045554.D	02 Apr 2025 18:01	JC/MD	Ok
32	Q1623-03	VX045555.D	02 Apr 2025 18:24	JC/MD	Ok
33	Q1623-04	VX045556.D	02 Apr 2025 18:47	JC/MD	Ok
34	VSTDCCC050	VX045557.D	02 Apr 2025 19:10	JC/MD	Ok,M

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QCBatch ID # VX040425

Review By	John Carlone	Review On	4/7/2025 9:58:56 AM		
Supervise By	Mahesh Dadoda	Supervise On	4/7/2025 12:14:08 PM		
SubDirectory	VX040425	HP Acquire Method		HP Processing Method	82X040225W.M

STD. NAME	STD REF.#
Tune/Reschk	VP133596
Initial Calibration Stds	
CCC	VP133597,VP133598
Internal Standard/PEM	
ICV/I.BLK	
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr #	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VX045585.D	04 Apr 2025 09:31	JC/MD	Ok
2	VSTDCCC050	VX045586.D	04 Apr 2025 10:00	JC/MD	Ok,M
3	VX0404MBL01	VX045587.D	04 Apr 2025 10:28	JC/MD	Ok
4	VX0404WBL01	VX045588.D	04 Apr 2025 10:51	JC/MD	Ok
5	VX0404WBS01	VX045589.D	04 Apr 2025 11:14	JC/MD	Ok,M
6	VX0404WBSD01	VX045590.D	04 Apr 2025 11:40	JC/MD	Ok,M
7	Q1711-02MS	VX045591.D	04 Apr 2025 12:04	JC/MD	Ok,M
8	IBLK	VX045592.D	04 Apr 2025 12:27	JC/MD	Ok
9	Q1716-01	VX045593.D	04 Apr 2025 12:50	JC/MD	Ok
10	Q1731-04	VX045594.D	04 Apr 2025 13:13	JC/MD	Not Ok
11	Q1731-01	VX045595.D	04 Apr 2025 13:37	JC/MD	Not Ok
12	Q1731-02	VX045596.D	04 Apr 2025 14:00	JC/MD	Not Ok
13	Q1731-03	VX045597.D	04 Apr 2025 14:23	JC/MD	Not Ok
14	IBLK	VX045598.D	04 Apr 2025 14:46	JC/MD	Ok
15	Q1731-07	VX045599.D	04 Apr 2025 15:10	JC/MD	Ok
16	Q1731-05	VX045600.D	04 Apr 2025 15:33	JC/MD	Ok
17	PB167448TB	VX045601.D	04 Apr 2025 15:56	JC/MD	Ok
18	PB167448ZHE#01	VX045602.D	04 Apr 2025 16:19	JC/MD	Ok
19	PB167448ZHE#02	VX045603.D	04 Apr 2025 16:42	JC/MD	Ok
20	PB167448ZHE#03	VX045604.D	04 Apr 2025 17:06	JC/MD	Ok
21	PB167448ZHE#04	VX045605.D	04 Apr 2025 17:29	JC/MD	Ok

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QCBatch ID # VX040425

Review By	John Carlone	Review On	4/7/2025 9:58:56 AM		
Supervise By	Mahesh Dadoda	Supervise On	4/7/2025 12:14:08 PM		
SubDirectory	VX040425	HP Acquire Method		HP Processing Method	82X040225W.M

STD. NAME	STD REF.#
Tune/Reschk	VP133596
Initial Calibration Stds	
CCC	VP133597,VP133598
Internal Standard/PEM	
ICV/I.BLK	
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

22	PB167448ZHE#05	VX045606.D	04 Apr 2025 17:52	JC/MD	Ok,M
23	Q1694-04	VX045607.D	04 Apr 2025 18:15	JC/MD	Ok
24	Q1695-04	VX045608.D	04 Apr 2025 18:38	JC/MD	Ok
25	Q1695-08	VX045609.D	04 Apr 2025 19:01	JC/MD	Ok
26	Q1712-04	VX045610.D	04 Apr 2025 19:24	JC/MD	Ok
27	Q1712-08	VX045611.D	04 Apr 2025 19:48	JC/MD	Ok
28	VSTDCCC050	VX045612.D	04 Apr 2025 20:11	JC/MD	Ok,M

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QCBatch ID # VX040725

Review By	John Carlone	Review On	4/8/2025 10:03:01 AM		
Supervise By	Mahesh Dadoda	Supervise On	4/8/2025 2:05:33 PM		
SubDirectory	VX040725	HP Acquire Method		HP Processing Method	82X040225W.M

STD. NAME	STD REF.#
Tune/Reschk	VP133600
Initial Calibration Stds	
CCC	VP133601,VP133602
Internal Standard/PEM	
ICV/I.BLK	
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr #	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VX045613.D	07 Apr 2025 08:34	JC/MD	Ok
2	VSTDCCC050	VX045614.D	07 Apr 2025 09:17	JC/MD	Ok,M
3	VX0407MBL01	VX045615.D	07 Apr 2025 09:47	JC/MD	Ok
4	VX0407WBL01	VX045616.D	07 Apr 2025 10:11	JC/MD	Ok
5	VX0407WBS01	VX045617.D	07 Apr 2025 10:34	JC/MD	Ok,M
6	VX0407WBSD01	VX045618.D	07 Apr 2025 11:00	JC/MD	Ok,M
7	Q1731-04	VX045619.D	07 Apr 2025 11:23	JC/MD	Ok
8	Q1731-01	VX045620.D	07 Apr 2025 11:46	JC/MD	Ok,M
9	Q1731-02	VX045621.D	07 Apr 2025 12:09	JC/MD	Ok
10	Q1731-03	VX045622.D	07 Apr 2025 12:33	JC/MD	Ok
11	Q1741-03	VX045623.D	07 Apr 2025 12:56	JC/MD	Ok
12	Q1741-04	VX045624.D	07 Apr 2025 13:19	JC/MD	Ok
13	Q1741-01	VX045625.D	07 Apr 2025 13:43	JC/MD	Not Ok
14	Q1741-02	VX045626.D	07 Apr 2025 14:06	JC/MD	Not Ok
15	IBLK	VX045627.D	07 Apr 2025 14:35	JC/MD	Ok
16	Q1741-02	VX045628.D	07 Apr 2025 14:58	JC/MD	Ok
17	Q1746-08	VX045629.D	07 Apr 2025 15:21	JC/MD	Ok
18	Q1746-07	VX045630.D	07 Apr 2025 15:45	JC/MD	ReRun
19	Q1746-05	VX045631.D	07 Apr 2025 16:08	JC/MD	Ok
20	Q1746-06	VX045632.D	07 Apr 2025 16:31	JC/MD	Ok
21	VSTDCCC050	VX045633.D	07 Apr 2025 16:55	JC/MD	Ok,M

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QCBatch ID # VX040225

Review By	John Carlone	Review On	4/2/2025 9:42:46 AM		
Supervise By	Amit Patel	Supervise On	4/2/2025 2:14:39 PM		
SubDirectory	VX040225	HP Acquire Method		HP Processing Method	82X040225W.M

STD. NAME	STD REF.#
Tune/Reschk	VP133559,VP133563
Initial Calibration Stds	VP133569,VP133570,VP133571,VP133572,VP133573,VP133574
CCC	VP133564,VP133565
Internal Standard/PEM	
ICV/I.BLK	VP133575
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr #	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX045524.D	01 Apr 2025 16:15		JC/MD	Ok
2	VSTDICC001	VSTDICC001	VX045525.D	01 Apr 2025 17:06	%D failed for Comp. #53 in 01PPB	JC/MD	Ok,M
3	VSTDICC005	VSTDICC005	VX045526.D	01 Apr 2025 17:29		JC/MD	Ok,M
4	VSTDICC020	VSTDICC020	VX045527.D	01 Apr 2025 17:52		JC/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VX045528.D	01 Apr 2025 18:15		JC/MD	Ok,M
6	VSTDICC100	VSTDICC100	VX045529.D	01 Apr 2025 18:38		JC/MD	Ok,M
7	VSTDICC150	VSTDICC150	VX045530.D	01 Apr 2025 19:02		JC/MD	Ok,M
8	IBLK	IBLK	VX045531.D	01 Apr 2025 19:25		JC/MD	Ok
9	VSTDICV050	ICVVX040225	VX045532.D	01 Apr 2025 19:48		JC/MD	Ok,M
10	BFB	BFB	VX045533.D	02 Apr 2025 09:30		JC/MD	Ok
11	VSTDCCC050	VSTDCCC050	VX045534.D	02 Apr 2025 10:02	pH#Lot#V12668	JC/MD	Ok,M
12	VX0402MBL01	VX0402MBL01	VX045535.D	02 Apr 2025 10:30		JC/MD	Ok
13	VX0402WBL01	VX0402WBL01	VX045536.D	02 Apr 2025 10:53		JC/MD	Ok
14	VX0402WBS01	VX0402WBS01	VX045537.D	02 Apr 2025 11:16		JC/MD	Ok,M
15	VX0402WBSD01	VX0402WBSD01	VX045538.D	02 Apr 2025 11:44		JC/MD	Ok,M
16	Q1697-01	MW-19B-72-040125	VX045539.D	02 Apr 2025 12:07	vial A pH<2 Need 200X	JC/MD	Dilution
17	Q1697-02	IW-01-55-040125	VX045540.D	02 Apr 2025 12:30	vial A pH<2 Need 10X	JC/MD	Dilution

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QCBatch ID # VX040225

Review By	John Carlone	Review On	4/2/2025 9:42:46 AM		
Supervise By	Amit Patel	Supervise On	4/2/2025 2:14:39 PM		
SubDirectory	VX040225	HP Acquire Method		HP Processing Method	82X040225W.M

STD. NAME	STD REF.#
Tune/Reschk	VP133559,VP133563
Initial Calibration Stds	VP133569,VP133570,VP133571,VP133572,VP133573,VP133574
CCC	VP133564,VP133565
Internal Standard/PEM	
ICV/I.BLK	VP133575
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

18	IBLK	IBLK	VX045541.D	02 Apr 2025 12:54		JC/MD	Ok
19	Q1697-01DL	MW-19B-72-040125DL	VX045542.D	02 Apr 2025 13:21	vial B pH<2	JC/MD	Ok
20	Q1697-02DL	IW-01-55-040125DL	VX045543.D	02 Apr 2025 13:44	vial B pH<2	JC/MD	Ok
21	Q1697-03	IW-02-55-040125	VX045544.D	02 Apr 2025 14:07	Need lower dilution	JC/MD	Not Ok
22	Q1697-04	IW-02-55-040125-FD	VX045545.D	02 Apr 2025 14:31	Need lower dilution	JC/MD	Not Ok
23	Q1697-05	IW-03-55-040125	VX045546.D	02 Apr 2025 14:54	Need lower dilution	JC/MD	Not Ok
24	Q1697-07	MW-19B-72-040125	VX045547.D	02 Apr 2025 15:18	Not On Login	JC/MD	Not Ok
25	IBLK	IBLK	VX045548.D	02 Apr 2025 15:41		JC/MD	Ok
26	Q1697-03	IW-02-55-040125	VX045549.D	02 Apr 2025 16:04	vial A pH<2	JC/MD	Ok
27	Q1697-04	IW-02-55-040125-FD	VX045550.D	02 Apr 2025 16:27	vial A pH<2	JC/MD	Ok
28	Q1697-05	IW-03-55-040125	VX045551.D	02 Apr 2025 16:51	vial A pH<2	JC/MD	Ok
29	Q1697-06	TB-01-040125	VX045552.D	02 Apr 2025 17:14	vial A pH<2 TB;Hit of comp.#44	JC/MD	ReRun
30	Q1623-01	Storage-Blank-SOIL-RE	VX045553.D	02 Apr 2025 17:37	vial A pH<2	JC/MD	Ok
31	Q1623-02	Storage-Blank-WATER-	VX045554.D	02 Apr 2025 18:01	vial A pH<2	JC/MD	Ok
32	Q1623-03	Storage-Blank-WATER-	VX045555.D	02 Apr 2025 18:24	vial A pH<2	JC/MD	Ok
33	Q1623-04	Storage-Blank-SAMPLE	VX045556.D	02 Apr 2025 18:47	vial A pH<2	JC/MD	Ok
34	VSTDCCC050	VSTDCCC050EC	VX045557.D	02 Apr 2025 19:10		JC/MD	Ok,M

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QCBatch ID # VX040425

Review By	John Carlone	Review On	4/7/2025 9:58:56 AM		
Supervise By	Mahesh Dadoda	Supervise On	4/7/2025 12:14:08 PM		
SubDirectory	VX040425	HP Acquire Method		HP Processing Method	82X040225W.M

STD. NAME	STD REF.#
Tune/Reschk	VP133596
Initial Calibration Stds	
CCC	VP133597,VP133598
Internal Standard/PEM	
ICV/I.BLK	
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr #	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX045585.D	04 Apr 2025 09:31		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX045586.D	04 Apr 2025 10:00	pH#Lot#V12668	JC/MD	Ok,M
3	VX0404MBL01	VX0404MBL01	VX045587.D	04 Apr 2025 10:28		JC/MD	Ok
4	VX0404WBL01	VX0404WBL01	VX045588.D	04 Apr 2025 10:51		JC/MD	Ok
5	VX0404WBS01	VX0404WBS01	VX045589.D	04 Apr 2025 11:14		JC/MD	Ok,M
6	VX0404WBSD01	VX0404WBSD01	VX045590.D	04 Apr 2025 11:40		JC/MD	Ok,M
7	Q1711-02MS	MW-18B-56-040225MS	VX045591.D	04 Apr 2025 12:04	vial A pH<2	JC/MD	Ok,M
8	IBLK	IBLK	VX045592.D	04 Apr 2025 12:27		JC/MD	Ok
9	Q1716-01	MOD-FOR-LRSA	VX045593.D	04 Apr 2025 12:50	vial A pH<2	JC/MD	Ok
10	Q1731-04	RMW-03B-90-040325	VX045594.D	04 Apr 2025 13:13	RUN@40X	JC/MD	Not Ok
11	Q1731-01	RMW-01B-82-040325	VX045595.D	04 Apr 2025 13:37	Need Straight Run	JC/MD	Not Ok
12	Q1731-02	RMW-04B-91-040325	VX045596.D	04 Apr 2025 14:00	Need Straight Run	JC/MD	Not Ok
13	Q1731-03	RMW-01B-82-040325-F	VX045597.D	04 Apr 2025 14:23	Need Straight Run	JC/MD	Not Ok
14	IBLK	IBLK	VX045598.D	04 Apr 2025 14:46		JC/MD	Ok
15	Q1731-07	TB01-040325	VX045599.D	04 Apr 2025 15:10	vial A pH<2 TB	JC/MD	Ok
16	Q1731-05	EB01-040325	VX045600.D	04 Apr 2025 15:33	vial A pH<2 EB	JC/MD	Ok
17	PB167448TB	PB167448TB	VX045601.D	04 Apr 2025 15:56		JC/MD	Ok
18	PB167448ZHE#01	PB167448ZHE#01	VX045602.D	04 Apr 2025 16:19		JC/MD	Ok

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QCBatch ID # VX040425

Review By	John Carlone	Review On	4/7/2025 9:58:56 AM		
Supervise By	Mahesh Dadoda	Supervise On	4/7/2025 12:14:08 PM		
SubDirectory	VX040425	HP Acquire Method		HP Processing Method	82X040225W.M

STD. NAME	STD REF.#
Tune/Reschk	VP133596
Initial Calibration Stds	
CCC	VP133597,VP133598
Internal Standard/PEM	
ICV/I.BLK	
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	PB167448ZHE#02	PB167448ZHE#02	VX045603.D	04 Apr 2025 16:42		JC/MD	Ok
20	PB167448ZHE#03	PB167448ZHE#03	VX045604.D	04 Apr 2025 17:06		JC/MD	Ok
21	PB167448ZHE#04	PB167448ZHE#04	VX045605.D	04 Apr 2025 17:29		JC/MD	Ok
22	PB167448ZHE#05	PB167448ZHE#05	VX045606.D	04 Apr 2025 17:52		JC/MD	Ok,M
23	Q1694-04	TP-18	VX045607.D	04 Apr 2025 18:15	vial A pH#5.0	JC/MD	Ok
24	Q1695-04	TT-2	VX045608.D	04 Apr 2025 18:38	vial A pH#5.0	JC/MD	Ok
25	Q1695-08	TT-3	VX045609.D	04 Apr 2025 19:01	vial A pH#5.0	JC/MD	Ok
26	Q1712-04	Z-05A	VX045610.D	04 Apr 2025 19:24	vial A pH#5.0	JC/MD	Ok
27	Q1712-08	TT-7	VX045611.D	04 Apr 2025 19:48	vial A pH#5.0	JC/MD	Ok
28	VSTDCCC050	VSTDCCC050EC	VX045612.D	04 Apr 2025 20:11		JC/MD	Ok,M

M : Manual Integration

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QCBatch ID # VX040725

Review By	John Carlone	Review On	4/8/2025 10:03:01 AM		
Supervise By	Mahesh Dadoda	Supervise On	4/8/2025 2:05:33 PM		
SubDirectory	VX040725	HP Acquire Method		HP Processing Method	82X040225W.M

STD. NAME	STD REF.#
Tune/Reschk	VP133600
Initial Calibration Stds	
CCC	VP133601,VP133602
Internal Standard/PEM	
ICV/I.BLK	
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr #	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX045613.D	07 Apr 2025 08:34		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX045614.D	07 Apr 2025 09:17		JC/MD	Ok,M
3	VX0407MBL01	VX0407MBL01	VX045615.D	07 Apr 2025 09:47	pH#Lot#V12668	JC/MD	Ok
4	VX0407WBL01	VX0407WBL01	VX045616.D	07 Apr 2025 10:11		JC/MD	Ok
5	VX0407WBS01	VX0407WBS01	VX045617.D	07 Apr 2025 10:34		JC/MD	Ok,M
6	VX0407WBSD01	VX0407WBSD01	VX045618.D	07 Apr 2025 11:00		JC/MD	Ok,M
7	Q1731-04	RMW-03B-90-040325	VX045619.D	07 Apr 2025 11:23	vial A pH<2	JC/MD	Ok
8	Q1731-01	RMW-01B-82-040325	VX045620.D	07 Apr 2025 11:46	vial A pH<2	JC/MD	Ok,M
9	Q1731-02	RMW-04B-91-040325	VX045621.D	07 Apr 2025 12:09	vial A pH<2	JC/MD	Ok
10	Q1731-03	RMW-01B-82-040325-F	VX045622.D	07 Apr 2025 12:33	vial A pH<2	JC/MD	Ok
11	Q1741-03	FB01-040425	VX045623.D	07 Apr 2025 12:56	vial A pH<2,FB	JC/MD	Ok
12	Q1741-04	TB01-040425	VX045624.D	07 Apr 2025 13:19	vial A pH<2,TB	JC/MD	Ok
13	Q1741-01	RMW-02B-66-040425	VX045625.D	07 Apr 2025 13:43	run 50x	JC/MD	Not Ok
14	Q1741-02	RMW-02B-74-040425	VX045626.D	07 Apr 2025 14:06	Need Straight Run	JC/MD	Not Ok
15	IBLK	IBLK	VX045627.D	07 Apr 2025 14:35		JC/MD	Ok
16	Q1741-02	RMW-02B-74-040425	VX045628.D	07 Apr 2025 14:58	vial A pH<2	JC/MD	Ok
17	Q1746-08	TB-2025-4-7	VX045629.D	07 Apr 2025 15:21	vial A pH<2 TB	JC/MD	Ok
18	Q1746-07	EB-2025-4-7	VX045630.D	07 Apr 2025 15:45	vial A pH<2 EB,Hit of com.#16	JC/MD	ReRun

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QCBatch ID # VX040725

Review By	John Carlone	Review On	4/8/2025 10:03:01 AM		
Supervise By	Mahesh Dadoda	Supervise On	4/8/2025 2:05:33 PM		
SubDirectory	VX040725	HP Acquire Method		HP Processing Method	82X040225W.M

STD. NAME	STD REF.#
Tune/Reschk	VP133600
Initial Calibration Stds	
CCC	VP133601,VP133602
Internal Standard/PEM	
ICV/I.BLK	
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	Q1746-05	B-158-GW01	VX045631.D	07 Apr 2025 16:08	vial A pH<2	JC/MD	Ok
20	Q1746-06	B-149-GW01	VX045632.D	07 Apr 2025 16:31	vial A pH<2	JC/MD	Ok
21	VSTDCCC050	VSTDCCC050EC	VX045633.D	07 Apr 2025 16:55		JC/MD	Ok,M

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Prep Standard - Chemical Standard Summary

Order ID : Q1731

Test : VOCMS Group3

Prepbatch ID :

Sequence ID/Qc Batch ID: vx040425,vx040725,

Standard ID :
VP131746,VP131767,VP132035,VP132096,VP133174,VP133342,VP133544,VP133596,VP133597,VP133598,VP133600,VP133601,VP133602,

Chemical ID :
V13391,V13457,V13460,V13465,V13466,V13706,V14154,V14289,V14431,V14435,V14501,V14502,V14523,V14524,V14613,V14614,V14615,V14630,V14631,V14632,V14633,V14719,V14720,V14726,V14744,V14753,V14804,V14805,V14842,V14883,V14896,V14897,V14898,V14899,W3112,

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
247	8260 Internal Standard, 250PPM	VP131746	11/22/2024	05/18/2025	Semsettin Yesilyurt	None	None	Mahesh Dadoda 11/23/2024

FROM 0.50000ml of V14289 + 49.50000ml of V14154 = Final Quantity: 50.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
218	BFB, 25PPM	VP131767	11/22/2024	05/18/2025	Semsettin Yesilyurt	None	None	Mahesh Dadoda 11/27/2024

FROM 0.50000ml of V13391 + 49.50000ml of V14154 = Final Quantity: 50.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1810	8260 Working Std(2-CVE)-800ppm	VP132035	12/10/2024	06/10/2025	Semsettin Yesilyurt	None	None	Mahesh Dadoda 12/12/2024

FROM 1.00000ml of V14630 + 1.00000ml of V14631 + 1.00000ml of V14632 + 1.00000ml of V14633 + 46.00000ml of V14614 = Final Quantity: 50.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
719	8260 Working STD (BCM)-First source, 400PPM	VP132096	12/12/2024	06/10/2025	Semsettin Yesilyurt	None	None	Mahesh Dadoda 12/19/2024

FROM 1.00000ml of V13465 + 1.00000ml of V13466 + 1.50000ml of V13457 + 1.50000ml of V13460 + 20.00000ml of V14614 = Final Quantity: 25.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
617	8260 Surrogate, 400PPM	VP133174	02/27/2025	08/27/2025	Semsettin Yesilyurt	None	None	Mahesh Dadoda 03/04/2025

FROM 0.40000ml of V13706 + 24.60000ml of V14613 = Final Quantity: 25.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
51	8260 Working STD (Acrolein) -first source, 800PPM	VP133342	03/18/2025	04/17/2025	Semsettin Yesilyurt	None	None	Mahesh Dadoda 03/20/2025

FROM 1.00000ml of V14896 + 1.00000ml of V14897 + 1.00000ml of V14898 + 1.00000ml of V14899 + 21.00000ml of V14883 = Final Quantity: 25.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
257	8260 Calibration Working STD Mix-First source, 160PPM	VP133544	04/01/2025	05/10/2025	Semsettin Yesilyurt	None	None	Mahesh Dadoda 04/03/2025

FROM 0.40000ml of V14842 + 1.00000ml of V14431 + 1.00000ml of V14435 + 1.00000ml of V14501 + 1.00000ml of V14502 + 1.00000ml of V14523 + 1.00000ml of V14524 + 1.00000ml of V14726 + 1.00000ml of V14744 + 1.00000ml of V14753 + 1.00000ml of V14804 + 1.00000ml of V14805 + 1.50000ml of V14719 + 1.50000ml of V14720 + 10.60000ml of V14615 = Final Quantity: 25.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
589	BFB TUNE CHECK	VP133596	04/04/2025	04/05/2025	John Carlone	None	None	Mahesh Dadoda 04/07/2025

FROM 39.98400ml of W3112 + 0.01600ml of VP131767 = Final Quantity: 40.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
620	50 PPB CCC, 8260-Water	VP133597	04/04/2025	04/05/2025	John Carlone	None	None	Mahesh Dadoda 04/07/2025

FROM 39.94450ml of W3112 + 0.00500ml of VP132096 + 0.00500ml of VP133174 + 0.00800ml of VP131746 + 0.01250ml of VP132035 + 0.01250ml of VP133342 + 0.01250ml of VP133544 = Final Quantity: 40.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
620	50 PPB CCC, 8260-Water	VP133598	04/04/2025	04/05/2025	John Carlone	None	None	Mahesh Dadoda 04/07/2025

FROM 39.94450ml of W3112 + 0.00500ml of VP132096 + 0.00500ml of VP133174 + 0.00800ml of VP131746 + 0.01250ml of VP132035 + 0.01250ml of VP133342 + 0.01250ml of VP133544 = Final Quantity: 40.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
589	BFB TUNE CHECK	VP133600	04/07/2025	04/08/2025	John Carlone	None	None	Mahesh Dadoda 04/07/2025

FROM 39.98400ml of W3112 + 0.01600ml of VP131767 = Final Quantity: 40.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
620	50 PPB CCC, 8260-Water	VP133601	04/07/2025	04/08/2025	John Carlone	None	None	Mahesh Dadoda 04/07/2025

FROM 39.94450ml of W3112 + 0.00500ml of VP132096 + 0.00500ml of VP133174 + 0.00800ml of VP131746 + 0.01250ml of VP132035 + 0.01250ml of VP133342 + 0.01250ml of VP133544 = Final Quantity: 40.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
620	50 PPB CCC, 8260-Water	VP133602	04/07/2025	04/08/2025	John Carlone	None	None	Mahesh Dadoda 04/07/2025

FROM 39.94450ml of W3112 + 0.00500ml of VP132096 + 0.00500ml of VP133174 + 0.00800ml of VP131746 + 0.01250ml of VP132035 + 0.01250ml of VP133342 + 0.01250ml of VP133544 = Final Quantity: 40.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30067 / BFB tuneing solution	A0191805	11/22/2025	11/22/2024 / SAM	01/13/2023 / SAM	V13391

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30225 / VOA Mix, bromochloromethane, 2000ug/mL, P&TM, 1mL/ampul	A0193071	06/12/2025	12/12/2024 / SAM	01/27/2023 / SAM	V13457

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30225 / VOA Mix, bromochloromethane, 2000ug/mL, P&TM, 1mL/ampul	A0193071	06/12/2025	12/12/2024 / SAM	01/27/2023 / SAM	V13460

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30225 / VOA Mix, bromochloromethane, 2000ug/mL, P&TM, 1mL/ampul	A0193071	06/12/2025	12/12/2024 / SAM	01/27/2023 / SAM	V13465

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30225 / VOA Mix, bromochloromethane, 2000ug/mL, P&TM, 1mL/ampul	A0193071	06/12/2025	12/12/2024 / SAM	01/27/2023 / SAM	V13466

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555582 / Custom Mixture, 8260 A/B Surrogate Mix [CS 5179-2]	A0196865	02/27/2026	02/27/2025 / SAM	04/12/2023 / SAM	V13706

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	22L0562016	05/18/2025	11/18/2024 / pedro	02/06/2024 / SAM	V14154

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555581 / Custom Standard, 8260 Internal Std [CS 5179-1]	A0210184	11/22/2025	11/22/2024 / SAM	04/15/2024 / SAM	V14289

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30489 / VOA Mix, 8260B Acetates Mix, P&TM, 1mL	A0209618	09/20/2025	03/20/2025 / SAM	08/15/2024 / SAM	V14431

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30489 / VOA Mix, 8260B Acetates Mix, P&TM, 1mL	A0209618	09/20/2025	03/20/2025 / SAM	08/15/2024 / SAM	V14435

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95317 / Universal VOA Mega Mix (Min order = 5)	021624	09/20/2025	03/20/2025 / SAM	09/17/2024 / SAM	V14501

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95317 / Universal VOA Mega Mix (Min order = 5)	021624	09/20/2025	03/20/2025 / SAM	09/17/2024 / SAM	V14502

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95319 / Revised Additions Mix (Min = 5)	091724	09/20/2025	03/20/2025 / SAM	09/18/2024 / SAM	V14523

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95319 / Revised Additions Mix (Min = 5)	091724	09/20/2025	03/20/2025 / SAM	09/18/2024 / SAM	V14524

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	22L0562016	08/27/2025	02/27/2025 / SAM	11/26/2024 / SAM	V14613

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	22L0562016	06/10/2025	12/10/2024 / SAM	11/26/2024 / SAM	V14614

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	22L0562016	09/19/2025	03/19/2025 / SAM	11/26/2024 / SAM	V14615

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ 2-Chloroethyl vinyl ether	120524	06/10/2025	12/10/2024 / SAM	12/06/2024 / SAM	V14630

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ 2-Chloroethyl vinyl ether	120524	06/10/2025	12/10/2024 / SAM	12/06/2024 / SAM	V14631

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ 2-Chloroethyl vinyl ether	120524	06/10/2025	12/10/2024 / SAM	12/06/2024 / SAM	V14632

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ 2-Chloroethyl vinyl ether	120524	06/10/2025	12/10/2024 / SAM	12/06/2024 / SAM	V14633

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30006 / VOA Mix, CLP method Calibration Std #1 ketones 5000uq/ml, PTM, 1ml	A02110618	09/20/2025	03/20/2025 / SAM	12/17/2024 / SAM	V14719

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30006 / VOA Mix, CLP method Calibration Std #1 ketones 5000uq/ml, PTM, 1ml	A02110618	09/20/2025	03/20/2025 / SAM	12/17/2024 / SAM	V14720

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30006 / VOA Mix, CLP method Calibration Std #1 ketones 5000uq/ml, PTM, 1ml	A02110618	07/30/2025	01/30/2025 / SAM	12/17/2024 / SAM	V14726

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30042 / VOA Mix,500 series method 502.2 Calibration Std #1 gases, 2000uq/ml, PTM, 1ml	A0216826	08/27/2025	02/27/2025 / SAM	12/17/2024 / SAM	V14744

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30042 / VOA Mix,500 series method 502.2 Calibration Std #1 gases, 2000uq/ml, PTM, 1ml	A0216826	07/30/2025	01/30/2025 / SAM	12/17/2024 / SAM	V14753

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555408 / Custom Standard, Vinyl Acetate Standard w/ Grav [CS 5066-6] TWO SEPARATE LOTS	A0220471	09/20/2025	03/20/2025 / SAM	01/08/2025 / SAM	V14804

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555408 / Custom Standard, Vinyl Acetate Standard w/ Grav [CS 5066-6] TWO SEPARATE LOTS	A0220471	09/20/2025	03/20/2025 / SAM	01/08/2025 / SAM	V14805

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30470 / VOA Stock Solution, tert-butanol std, 1mL, P&TM	A0217535	08/27/2025	02/27/2025 / SAM	01/21/2025 / SAM	V14842

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	22L0562016	10/25/2025	02/19/2025 / Jaswal	04/22/2024 / Jaswal	V14883

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	91980 / Acrolin Std (Min = 5)	031725	04/17/2025	03/18/2025 / SAM	03/18/2025 / SAM	V14896

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	91980 / Acrolin Std (Min = 5)	031725	04/17/2025	03/18/2025 / SAM	03/18/2025 / SAM	V14897

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	91980 / Acrolin Std (Min = 5)	031725	04/17/2025	03/18/2025 / SAM	03/18/2025 / SAM	V14898

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	91980 / Acrolin Std (Min = 5)	031725	04/17/2025	03/18/2025 / SAM	03/18/2025 / SAM	V14899

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / Iwona	07/03/2024 / Iwona	W3112

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis

12 Baxter

avantor™ | J.T.Baker®

Material No.: 9077-02
Batch No.: 22L0562016
Manufactured Date: 2022-10-26
Expiration Date: 2025-10-25
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay (CH_3OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.2 ppm
Titrable Acid (µeq/g)	≤ 0.3	0.2
Titrable Base (µeq/g)	≤ 0.10	0.03
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis – Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory,Research,or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

James Ethier
Jamie Ethier
Vice President Global Quality

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis

12 Baxter

avantor™ | J.T.Baker®

Material No.: 9077-02
Batch No.: 22L0562016
Manufactured Date: 2022-10-26
Expiration Date: 2025-10-25
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay (CH_3OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.2 ppm
Titrable Acid (µeq/g)	≤ 0.3	0.2
Titrable Base (µeq/g)	≤ 0.10	0.03
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis – Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory,Research,or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

James Ethier
Jamie Ethier
Vice President Global Quality

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis

12 Baxter

avantor™ | J.T.Baker®

Material No.: 9077-02
Batch No.: 22L0562016
Manufactured Date: 2022-10-26
Expiration Date: 2025-10-25
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay (CH_3OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.2 ppm
Titrable Acid (µeq/g)	≤ 0.3	0.2
Titrable Base (µeq/g)	≤ 0.10	0.03
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis – Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory,Research,or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

James Ethier
Jamie Ethier
Vice President Global Quality

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

Rec 09/17/24

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT

Part Number: 95317
Lot Number: 021624
Description: Universal VOA Megamix
69 components
Expiration Date: 021627
Recommended Storage: Freezer (0 °C)
Nominal Concentration (µg/mL): 2000
NIST Test ID#: 6UTB

Solvent(s): Lot#
Methanol EG359-USQ12

Weight(s) shown below were combined and diluted to (mL): 100.0 0.021 Flask Uncertainty

5E-05 Balance Uncertainty

Formulated By: Prashant Chauhan 021624 DATE

Reviewed By: Pedro L. Rentas 021624 DATE


	Compound	(RM#) Part Number	Lot Number	Dil. Factor	Initial Vol. (mL)	Initial Conc.(ug/mL)	Nominal Conc (µg/mL)	Purity (%)	Purity Uncertainty	Uncertainty Pipette (mL)	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50
1.	Acetonitrile	(0324)	021644	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20020	2001.3	8.1	75-05-8	40 ppm (70mg/m3/8H)	orl-rat 2460mg/kg
2.	Allyl chloride (3-Chloropropene)	(0325)	102396	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	107-05-1	1 ppm (3mg/m3/8H)	orl-rat 700mg/kg
3.	Carbon disulphide	(0060)	MKCR8561	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20023	2001.6	8.1	75-15-0	4 ppm (12mg/m3) (skin)	orl-rat 1200mg/kg
4.	cis-1,4-Dichloro-2-butene	(1196)	14718EF	NA	NA	NA	2000	95	0.2	NA	0.21058	0.21069	2001.1	8.5	1476-11-5	N/A	N/A
5.	trans-1,4-Dichloro-2-butene	(0466)	MKBP6041V	NA	NA	NA	2000	96.5	0.2	NA	0.20731	0.20748	2001.7	8.4	110-57-6	N/A	N/A
6.	Diethyl ether	(0153)	IK18CAS0000	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	60-29-7	N/A	N/A
7.	Ethyl methacrylate	(0381)	06128PX	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20230	2002.3	8.2	97-63-2	N/A	orl-rat 14800mg/kg
8.	Iodomethane	(0489)	SHBF8718V	NA	NA	NA	2000	99.5	0.2	NA	0.20106	0.20121	2001.5	8.2	74-88-4	5 ppm(28mg/m3/8H)(skin)	orl-rat 76mg/kg
9.	2-Methyl-1-propanol	(0445)	15241EB	NA	NA	NA	2000	99.5	0.2	NA	0.20106	0.20120	2001.4	8.1	78-83-1	50 ppm (150mg/m3/8H)	orl-rat 2460mg/kg
10.	Methacrylonitrile	(0442)	00427ET	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	126-98-7	1 ppm (3mg/m3/8H)(skin)	orl-rat 120mg/kg
11.	Methyl acrylate	(1075)	SHBK0679	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	96-33-3	10 ppm(35mg/m3/8H)(skin)	orl-rat 277mg/kg
12.	Methyl methacrylate	(0404)	MKBW5137V	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20041	2001.6	8.1	80-62-6	100 ppm (410mg/m3/8H)	orl-rat 7872mg/kg
13.	Nitrobenzene	(0228)	01213TV	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20220	2001.3	8.2	98-95-3	1 ppm (5mg/m3/8H)(skin)	orl-rat 780mg/kg
14.	2-Nitropropane	(0461)	14002JX	NA	NA	NA	2000	97.3	0.2	NA	0.20560	0.20577	2001.6	8.3	79-46-9	10 ppm (35mg/m3/8H)	orl-rat 720mg/kg
15.	Pentachloroethane	(0450)	HGA01	NA	NA	NA	2000	98	0.2	NA	0.20413	0.20430	2001.6	8.3	76-01-7	N/A	N/A
16.	1,1,2-Trichlorotrifluoroethane	(0474)	18930	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20225	2001.8	8.2	76-13-1	1000 ppm (7600mg/m3/8H)	orl-rat 43g/kg
17.	Bromodichloromethane	35171	101623	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1999.6	22.9	75-27-4	N/A	orl-rat 916mg/kg
18.	Dibromochloromethane	35171	101623	0.05	5.00	40002.1	2000	NA	NA	0.017	NA	NA	1999.6	23.0	124-48-1	N/A	orl-rat 848mg/kg
19.	cis-1,2-Dichloroethene	35171	101623	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.7	22.9	156-59-2	N/A	N/A
20.	trans-1,2-Dichloroethene	35171	101623	0.05	5.00	40002.4	2000	NA	NA	0.017	NA	NA	1999.6	23.0	156-60-5	N/A	orl-rat 1235mg/kg
21.	Methylene chloride	35171	101623	0.05	5.00	40002.8	2000	NA	NA	0.017	NA	NA	1999.6	22.9	75-09-2	500 ppm	orl-rat 820mg/kg
22.	1,1-Dichloroethene	32251	102023	0.10	10.00	20001.6	2000	NA	NA	0.042	NA	NA	1999.7	20.4	75-35-4	1 ppm (4mg/m3/8H)	orl-rat 200mg/kg
23.	Bromoform	95321	020724	0.10	10.00	20003.2	2000	NA	NA	0.042	NA	NA	1999.8	20.5	75-25-2	0.5 ppm (5mg/m3) (skin)	orl-rat 933mg/kg
24.	Carbon tetrachloride	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.4	56-23-5	2 ppm (12.6mg/m3/8H)	orl-rat 2350mg/kg
25.	Chloroform	95321	020724	0.10	10.00	20024.0	2000	NA	NA	0.042	NA	NA	2001.9	20.5	67-66-3	50 ppm (240mg/m3) (CL)	orl-rat 908mg/kg
26.	Dibromomethane	95321	020724	0.10	10.00	20002.9	2000	NA	NA	0.042	NA	NA	1999.8	20.5	74-95-3	N/A	orl-rat 108mg/kg
27.	1,1-Dichloroethane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.5	75-34-3	100 ppm	orl-rat 725mg/kg
28.	2,2-Dichloropropane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.4	594-20-7	N/A	N/A
29.	Tetrachloroethene	95321	020724	0.10	10.00	20201.1	2000	NA	NA	0.042	NA	NA	2019.6	20.6	127-18-4	25 ppm (170mg/m3/8H)(final)	orl-rat 2629mg/kg
30.	1,1,1-Trichloroethane	95321	020724	0.10	10.00	20003.0	2000	NA	NA	0.042	NA	NA	1999.8	20.5	71-55-6	350 ppm (1900mg/m3/8H)	orl-rat 10300mg/kg
31.	1,2-Dibromo-3-chloropropane	35161	112322	0.05	5.00	40016.5	2000	NA	NA	0.017	NA	NA	2000.3	22.9	96-12-8	0.001 ppm	orl-rat 170mg/kg
32.	1,2-Dibromoethane	35161	112322	0.05	5.00	40024.8	2000	NA	NA	0.017	NA	NA	2000.7	22.9	106-93-4	20 ppm (8H)	orl-rat 108mg/kg
33.	1,2-Dichloroethane	35161	112322	0.05	5.00	40018.0	2000	NA	NA	0.017	NA	NA	2000.4	22.9	107-06-2	50 ppm (8H)	orl-rat 670mg/kg
34.	1,2-Dichloropropane	35161	112322	0.05	5.00	40051.0	2000	NA	NA	0.017	NA	NA	2002.0	22.9	78-87-5	75 ppm (350mg/m3/8H)	orl-rat 1947mg/kg
35.	1,3-Dichloropropane	35161	112322	0.05	5.00	40005.9	2000	NA	NA	0.017	NA	NA	1999.8	22.9	142-28-9	N/A	unr-mus 3600mg/kg
36.	1,1-Dichloropropene	35161	112322	0.05	5.00	40012.1	2000	NA	NA	0.017	NA	NA	2000.1	29.7	563-58-6	N/A	N/A
37.	cis-1,3-Dichloropropene	35161	112322	0.05	5.00	40010.0	2000	NA	NA	0.017	NA	NA	2000.0	23.0	10061-01-5	N/A	N/A
38.	trans-1,3-Dichloropropene	35161	112322	0.05	5.00	40017.6	2000	NA	NA	0.017	NA	NA	2000.4	23.0	10061-02-6	N/A	N/A
39.	Hexachloro-1,3-butadiene	35161	112322	0.05	5.00	40021.9	2000	NA	NA	0.017	NA	NA	2000.6	29.7	87-68-3	0.02 ppm (0.24mg/m3/8H)	orl-rat 82mg/kg
40.	1,1,1,2-Tetrachloroethane	35161	112322	0.05	5.00	40011.9	2000	NA	NA	0.017	NA	NA	2000.1	22.9	630-20-6	N/A	orl-rat 670mg/kg
41.	1,1,2,2-Tetrachloroethane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.9	22.9	79-34-5	5 ppm (35mg/m3/8H)(skin)	orl-rat 800mg/kg
42.	1,1,2-Trichloroethane	35161	112322	0.05	5.00	40006.6	2000	NA	NA	0.017	NA	NA	1999.8	23.0	79-00-5	10 ppm (45mg/m3/8H)(skin)	orl-rat 836mg/kg
43.	Trichloroethene	35161	112322	0.05	5.00	40029.0	2000	NA	NA	0.017	NA	NA	2000.9	22.9	79-01-6	50 ppm (270mg/m3/8H)	orl-mus 2402mg/kg
44.	1,2,3-Trichloropropane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.9	22.9	96-18-4	10 ppm (60mg/m3/8H)	orl-rat 149.6mg/kg
45.	Benzene	35162	050823	0.05	5.00	40005.0	2000	NA	NA	0.017	NA	NA	1999.7	22.9	71-43-2	1 ppm	orl-rat 4894mg/kg
46.	Bromobenzene	35162	050823	0.05	5.00	40006.9	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-86-1	N/A	orl-rat 2699mg/kg
47.	n-Butyl benzene	35162	050823	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	104-51-8	N/A	N/A
48.	Ethyl benzene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	100-41-4	100 ppm (435mg/m3/8H)	orl-rat >2000mg/kg
49.	p-Isopropyl toluene	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	99-87-6	N/A	orl-rat 4750mg/kg
50.	Naphthalene	35162	050823	0.05	5.00	40006.2	2000	NA	NA	0.017	NA	NA	1999.8	22.9	91-20-3	10 ppm (50mg/m3/8H)	orl-rat 490mg/kg
51.	Styrene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	100-42-5	100 ppm	orl-rat 5000mg/kg
52.	Toluene	35162	050823	0.05	5.00	40006.2	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-88-3	200 ppm	orl-rat 5000mg/kg
53.	1,2,3-Trichlorobenzene	35162	050823	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.7	22.9	87-61-6	N/A	ipr-mus 1390mg/kg
54.	1,2,4-Trichlorobenzene	35162	050823	0.05	5.00	40006.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	120-82-1	5 ppm (CL) (40mg/m3)	orl-rat 756mg/kg
55.	1,2,4-Trimethylbenzene	35162	050823	0.05	5.00	40001.6	2000	NA	NA	0.017	NA	NA	1999.6	23.0	95-63-6	N/A	orl-rat 5g/kg
56.	1,3,5-Trimethylbenzene	35162	050823	0.05	5.00	40006.7	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-67-8	N/A	orl-rat 5000mg/kg
57.	m-Xylene	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-38-3	100 ppm (435mg/m3/8H)	orl-rat 5g/kg
58.	tert-Butyl benzene	35163	101923	0.05	5.00	40001.2	2000	NA	NA	0.017	NA	NA	1999.8	22.9	98-06-6	N/A	N/A
59.	sec-Butyl benzene	35163	101923	0.05	5.00	40002.4	2000	NA	NA	0.017	NA	NA	1999.8	22.9	135-98-8	N/A	orl-rat 2240mg/kg
60.	Chlorobenzene	35163	101923	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	108-90-7	75 ppm (350mg/m3/8H)	orl-rat 2290mg/kg
61.	2-Chlorotoluene	35163	101923	0.05	5.00	40000.3	2000	NA	NA	0.017	NA	NA	1999.5	22.9	95-49-8	50 ppm (250mg/m3/8H)	orl-rat 3900mg/kg
62.	4-Chlorotoluene	35163	101923	0.05	5.00	40003.3	2000	NA	NA	0.017	NA	NA	1999.7	22.9	106-43-4	N/A	orl-rat 2100mg/kg
63.	1,2-Dichlorobenzene	35163	101923	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	95-50-1	50 ppm (300mg/m3) (CL)	orl-rat 500mg/kg
64.	1,3-Dichlorobenzene	35163	101923	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1999.6	23.0	541-73-1	N/A	ipr-mus 1062mg/kg
65.	1,4-Dichlorobenzene	35163	101923	0.05	5.00	40001.8	2000	NA	NA	0.017	NA	NA	1999.6	22.9	106-46-7	75 ppm (450mg/m3/8H)	orl-rat 500mg/kg
66.	Isopropylbenzene	35163	101923	0.05	5.00	40000.8	2000	NA	NA	0.017	NA	NA	1999.5	22.9	98-82-8	50 ppm (245mg/m3/8H)	orl-rat 1400mg/kg
67.	n-Propylbenzene	35163	101923	0.05	5.00	40003.4	2000	NA	NA	0.017	NA	NA	1999.7	23.0	103-65-1	N/A	orl-rat 8040mg/kg
68.	o-Xylene	35163	101923	0.05	5.00	40040.8	2000	NA	NA	0.017	NA	NA	2001.5	23.0	95-47-6	100 ppm (435mg/m3/8H)	ipr-mus 1364mg/kg
69.	p-Xylene	35163	101923	0.05	5.00	40000.8	2000	NA	NA	0.017	NA	NA	1999.5	22.9	106-42-3	100 ppm (435mg/m3/8H)	orl-rat 5g/kg

- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Run 16, "P95317 L021624 (2000µg/mL in MeOH)"

Run Length: 60.00 min, 35998 points at 10 points/second.
Created: Sat, Feb 17, 2024 at 8:56:46 AM.
Sampled: Sequence "021624-GC5M1", Method "GC5-M1".
Analyzed using Method "GC5-M1".

Comments
GC5-M1 Analysis by Candice Warren
Column ID SPB-Vocol 105 meter X 0.53mm X 3.0µm film thickness
Flow rates: Total flow=290mL/min., Helium (carrier)=10mL/min.,
Helium(make-up)=10mL/min., Hydogen(make-up)=40mL/min., Air(make-up)=230mL/min.
Oven Profile: Temp. 1=35°C (Time 1=10 min.), Temp 2=200°C (Time 2=8.75 min.),
Rate = 4°C/min., Total run time=60 min. Injector temp.=200°C, FID Temp.=200°C.
FID Signal = Edaq Channel 1
Standard injection = 0.5µL, Range=3



Peak #	Name	FID RT (min.)
1	Ether	9.97
2	1,1,2-Trichloro-1,2,3-trifluoroethane	10.33
3	1,1-Dichloroethene	11.10
4	Acetonitrile	12.00
5	Iodomethane	12.31
6	Allyl chloride	12.56
7	Carbon disulfide/Methylene chloride	13.04
8	trans-1,2-Dichloroethene	14.07
9	1,1-Dichloroethane	15.74
10	2,2-Dichloropropane	17.74
11	cis-1,2-Dichloroethene	18.00
12	Methacrylonitrile/Methyl acrylate/Chloroform	18.49
13	Isobutanol/1,1,1-Trichloroethane	18.91
14	1,1-Dichloropropene	19.46
15	Carbon tetrachloride	20.79
16	Benzene/1,2-Dichloroethane	21.48
17	Trichloroethene	23.86
18	1,2-Dichloropropane	24.26
19	Methyl methacrylate	24.62
20	Bromodichloromethane	25.13
21	Dibromomethane/2-Nitropropane	25.46
22	cis-1,3-Dichloropropene	27.07
23	Toluene	28.09
24	Ethyl methacrylate/trans-1,3-Dichloropropene	28.73
25	1,1,2-Trichloroethane	29.34
26	Tetrachloroethene/1,3-Dichloropropane	30.34
27	Dibromochloromethane	31.16
28	1,2-Dibromoethane	31.84
29	Chlorobenzene	33.26
30	Ethylbenzene/1,1,1,2-Tetrachloroethane	33.40
31	m-Xylene/p-Xylene	33.85
32	o-Xylene	35.33
33	Styrene	35.50
34	Isopropylbenzene/Bromoform	35.48
35	cis-1,4-Dichloro-2-butene	36.80
36	1,1,2,2-Tetrachloroethane	37.23
37	1,2,3-Trichloropropane	37.77
38	n-Propylbenzene	37.92
39	trans-1,4-Dichloro-2-butene	38.05
40	Bromobenzene	38.14
41	1,3,5-Trimethylbenzene	38.50
42	2-Chlorotoluene	38.62
43	4-Chlorotoluene	38.77
44	tert-Butylbenzene	39.76
45	1,2,4-Trimethylbenzene	39.91
46	Pentachloroethane	40.17
47	sec-Butylbenzene	40.52
48	p-Isopropyltoluene	41.02
49	1,3-Dichlorobenzene	41.42
50	1,4-Dichlorobenzene	41.63
51	n-Butylbenzene	42.82
52	1,2-Dichlorobenzene	43.18
53	1,2-Dibromo-3-chloropropane	46.12
54	Nitrobenzene	46.58
55	1,2,4-Trichlorobenzene	49.26
56	Hexachlorobutadiene	49.72
57	Naphthalene	50.26
58	1,2,3-Trichlorobenzene	51.16

Absolute Standards Inc.

PO Box 5585
Hamden, CT 06518-0585

Phone: 203-281-2917
FAX: 203-281-2922

Safety Data Sheet (SDS) GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY **ANALYTICAL STANDARD DISSOLVED IN METHANOL**

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	**1-800-535-5053**
Address	44 Rossotto Dr. Hamden CT, 06514	Emergency Telephone International	**1-352-323-3500**
		Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225	**Highly Flammable Liquid and Vapor**	**H301, 311, 331**	**Toxic if swallowed, skin contact, inhaled**
H370	**Cause damage to organs**	**H351**	**Suspected of causing cancer**
P271	**Use in ventilated area**	**P280**	**Use gloves, eye protection/face sheild**
P302,332	**If on skin, wash with soap and water**	**P305,351,338**	**If in eyes, remove contacts, rinse with water**



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))			
Methanol	METHYL ALCOHOL	CAS#: 67-56-1	% (optional) > 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance.Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist.
Storage Conditions	Use ventilation Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge. Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol 67-56-1 TWA 200 ppm
Skin notation TWA 200 ppm
Potential for skin absorption , ingestion and inhalation.
Personal protective equipment Respiratory protection Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.

Section IX - Physical/Chemical Characteristics

Absolute Standards Inc.

PO Box 5585
Hamden, CT 06518-0585

Phone: 203-281-2917
FAX: 203-281-2922

Boiling Point	65°C	Specific Gravity (H2O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6

Solubility in Water COMPLETE

Appearance and Odor CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.

Section X. STABILITY AND REACTIVITY

Chemical stability — Stable under recommended storage conditions.
Possibility of hazardous reactions — Vapours may form explosive mixture with air.
Conditions to avoid — Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid — Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions. - Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50	15,400 mg/l - 96 h
EC50	24,500.00 mg/l - 48 h
EC100	10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)
UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol

IATA
UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

Rec 09/17/24

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT

Part Number: 95317
Lot Number: 021624
Description: Universal VOA Megamix
69 components
Expiration Date: 021627
Recommended Storage: Freezer (0 °C)
Nominal Concentration (µg/mL): 2000
NIST Test ID#: 6UTB

Solvent(s): Lot#
Methanol EG359-USQ12

5E-05 Balance Uncertainty
Weight(s) shown below were combined and diluted to (mL): 100.0 0.021 Flask Uncertainty

Formulated By: Prashant Chauhan 021624 DATE

Reviewed By: Pedro L. Rentas 021624 DATE


	Compound	(RM#) Part Number	Lot Number	Dil. Factor	Initial Vol. (mL)	Initial Conc.(ug/mL)	Nominal Conc (µg/mL)	Purity (%)	Purity Uncertainty	Uncertainty Pipette (mL)	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50
1.	Acetonitrile	(0324)	021644	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20020	2001.3	8.1	75-05-8	40 ppm (70mg/m3/8H)	orl-rat 2460mg/kg
2.	Allyl chloride (3-Chloropropene)	(0325)	102396	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	107-05-1	1 ppm (3mg/m3/8H)	orl-rat 700mg/kg
3.	Carbon disulphide	(0060)	MKCR8561	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20023	2001.6	8.1	75-15-0	4 ppm (12mg/m3) (skin)	orl-rat 1200mg/kg
4.	cis-1,4-Dichloro-2-butene	(1196)	14718EF	NA	NA	NA	2000	95	0.2	NA	0.21058	0.21069	2001.1	8.5	1476-11-5	N/A	N/A
5.	trans-1,4-Dichloro-2-butene	(0466)	MKBP6041V	NA	NA	NA	2000	96.5	0.2	NA	0.20731	0.20748	2001.7	8.4	110-57-6	N/A	N/A
6.	Diethyl ether	(0153)	IK18CAS0000	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	60-29-7	N/A	N/A
7.	Ethyl methacrylate	(0381)	06128PX	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20230	2002.3	8.2	97-63-2	N/A	orl-rat 14800mg/kg
8.	Iodomethane	(0489)	SHBF8718V	NA	NA	NA	2000	99.5	0.2	NA	0.20106	0.20121	2001.5	8.2	74-88-4	5 ppm(28mg/m3/8H)(skin)	orl-rat 76mg/kg
9.	2-Methyl-1-propanol	(0445)	15241EB	NA	NA	NA	2000	99.5	0.2	NA	0.20106	0.20120	2001.4	8.1	78-83-1	50 ppm (150mg/m3/8H)	orl-rat 2460mg/kg
10.	Methacrylonitrile	(0442)	00427ET	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	126-98-7	1 ppm (3mg/m3/8H)(skin)	orl-rat 120mg/kg
11.	Methyl acrylate	(1075)	SHBK0679	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	96-33-3	10 ppm(35mg/m3/8H)(skin)	orl-rat 277mg/kg
12.	Methyl methacrylate	(0404)	MKBW5137V	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20041	2001.6	8.1	80-62-6	100 ppm (410mg/m3/8H)	orl-rat 7872mg/kg
13.	Nitrobenzene	(0228)	01213TV	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20220	2001.3	8.2	98-95-3	1 ppm (5mg/m3/8H)(skin)	orl-rat 780mg/kg
14.	2-Nitropropane	(0461)	14002JX	NA	NA	NA	2000	97.3	0.2	NA	0.20560	0.20577	2001.6	8.3	79-46-9	10 ppm (35mg/m3/8H)	orl-rat 720mg/kg
15.	Pentachloroethane	(0450)	HGA01	NA	NA	NA	2000	98	0.2	NA	0.20413	0.20430	2001.6	8.3	76-01-7	N/A	N/A
16.	1,1,2-Trichlorotrifluoroethane	(0474)	18930	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20225	2001.8	8.2	76-13-1	1000 ppm (7600mg/m3/8H)	orl-rat 43g/kg
17.	Bromodichloromethane	35171	101623	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1999.6	22.9	75-27-4	N/A	orl-rat 916mg/kg
18.	Dibromochloromethane	35171	101623	0.05	5.00	40002.1	2000	NA	NA	0.017	NA	NA	1999.6	23.0	124-48-1	N/A	orl-rat 848mg/kg
19.	cis-1,2-Dichloroethene	35171	101623	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.7	22.9	156-59-2	N/A	N/A
20.	trans-1,2-Dichloroethene	35171	101623	0.05	5.00	40002.4	2000	NA	NA	0.017	NA	NA	1999.6	23.0	156-60-5	N/A	orl-rat 1235mg/kg
21.	Methylene chloride	35171	101623	0.05	5.00	40002.8	2000	NA	NA	0.017	NA	NA	1999.6	22.9	75-09-2	500 ppm	orl-rat 820mg/kg
22.	1,1-Dichloroethene	32251	102023	0.10	10.00	20001.6	2000	NA	NA	0.042	NA	NA	1999.7	20.4	75-35-4	1 ppm (4mg/m3/8H)	orl-rat 200mg/kg
23.	Bromoform	95321	020724	0.10	10.00	20003.2	2000	NA	NA	0.042	NA	NA	1999.8	20.5	75-25-2	0.5 ppm (5mg/m3) (skin)	orl-rat 933mg/kg
24.	Carbon tetrachloride	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.4	56-23-5	2 ppm (12.6mg/m3/8H)	orl-rat 2350mg/kg
25.	Chloroform	95321	020724	0.10	10.00	20024.0	2000	NA	NA	0.042	NA	NA	2001.9	20.5	67-66-3	50 ppm (240mg/m3) (CL)	orl-rat 908mg/kg
26.	Dibromomethane	95321	020724	0.10	10.00	20002.9	2000	NA	NA	0.042	NA	NA	1999.8	20.5	74-95-3	N/A	orl-rat 108mg/kg
27.	1,1-Dichloroethane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.5	75-34-3	100 ppm	orl-rat 725mg/kg
28.	2,2-Dichloropropane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.4	594-20-7	N/A	N/A
29.	Tetrachloroethene	95321	020724	0.10	10.00	20201.1	2000	NA	NA	0.042	NA	NA	2019.6	20.8	127-18-4	25 ppm (170mg/m3/8H)(final)	orl-rat 2629mg/kg
30.	1,1,1-Trichloroethane	95321	020724	0.10	10.00	20003.0	2000	NA	NA	0.042	NA	NA	1999.8	20.5	71-55-6	350 ppm (1900mg/m3/8H)	orl-rat 10300mg/kg
31.	1,2-Dibromo-3-chloropropane	35161	112322	0.05	5.00	40016.5	2000	NA	NA	0.017	NA	NA	2000.3	22.9	96-12-8	0.001 ppm	orl-rat 170mg/kg
32.	1,2-Dibromoethane	35161	112322	0.05	5.00	40024.8	2000	NA	NA	0.017	NA	NA	2000.7	22.9	106-93-4	20 ppm (8H)	orl-rat 108mg/kg
33.	1,2-Dichloroethane	35161	112322	0.05	5.00	40018.0	2000	NA	NA	0.017	NA	NA	2000.4	22.9	107-06-2	50 ppm (8H)	orl-rat 670mg/kg
34.	1,2-Dichloropropane	35161	112322	0.05	5.00	40051.0	2000	NA	NA	0.017	NA	NA	2002.0	22.9	78-87-5	75 ppm (350mg/m3/8H)	orl-rat 1947mg/kg
35.	1,3-Dichloropropane	35161	112322	0.05	5.00	40005.9	2000	NA	NA	0.017	NA	NA	1999.8	22.9	142-28-9	N/A	unr-mus 3600mg/kg
36.	1,1-Dichloropropene	35161	112322	0.05	5.00	40012.1	2000	NA	NA	0.017	NA	NA	2000.1	29.7	563-58-6	N/A	N/A
37.	cis-1,3-Dichloropropene	35161	112322	0.05	5.00	40010.0	2000	NA	NA	0.017	NA	NA	2000.0	23.0	10061-01-5	N/A	N/A
38.	trans-1,3-Dichloropropene	35161	112322	0.05	5.00	40017.6	2000	NA	NA	0.017	NA	NA	2000.4	23.0	10061-02-6	N/A	N/A
39.	Hexachloro-1,3-butadiene	35161	112322	0.05	5.00	40021.9	2000	NA	NA	0.017	NA	NA	2000.6	29.7	87-68-3	0.02 ppm (0.24mg/m3/8H)	orl-rat 82mg/kg
40.	1,1,1,2-Tetrachloroethane	35161	112322	0.05	5.00	40011.9	2000	NA	NA	0.017	NA	NA	2000.1	22.9	630-20-6	N/A	orl-rat 670mg/kg
41.	1,1,2,2-Tetrachloroethane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.9	22.9	79-34-5	5 ppm (35mg/m3/8H)(skin)	orl-rat 800mg/kg
42.	1,1,2-Trichloroethane	35161	112322	0.05	5.00	40006.6	2000	NA	NA	0.017	NA	NA	1999.8	23.0	79-00-5	10 ppm (45mg/m3/8H)(skin)	orl-rat 836mg/kg
43.	Trichloroethene	35161	112322	0.05	5.00	40029.0	2000	NA	NA	0.017	NA	NA	2000.9	22.9	79-01-6	50 ppm (270mg/m3/8H)	orl-mus 2402mg/kg
44.	1,2,3-Trichloropropane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.9	22.9	96-18-4	10 ppm (60mg/m3/8H)	orl-rat 149.6mg/kg
45.	Benzene	35162	050823	0.05	5.00	40005.0	2000	NA	NA	0.017	NA	NA	1999.7	22.9	71-43-2	1 ppm	orl-rat 4894mg/kg
46.	Bromobenzene	35162	050823	0.05	5.00	40006.9	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-86-1	N/A	orl-rat 2699mg/kg
47.	n-Butyl benzene	35162	050823	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	104-51-8	N/A	N/A
48.	Ethyl benzene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	100-41-4	100 ppm (435mg/m3/8H)	orl-rat >2000mg/kg
49.	p-Isopropyl toluene	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	99-87-6	N/A	orl-rat 4750mg/kg
50.	Naphthalene	35162	050823	0.05	5.00	40006.2	2000	NA	NA	0.017	NA	NA	1999.8	22.9	91-20-3	10 ppm (50mg/m3/8H)	orl-rat 490mg/kg
51.	Styrene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	100-42-5	100 ppm	orl-rat 5000mg/kg
52.	Toluene	35162	050823	0.05	5.00	40006.2	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-88-3	200 ppm	orl-rat 5000mg/kg
53.	1,2,3-Trichlorobenzene	35162	050823	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.7	22.9	87-61-6	N/A	ipr-mus 1390mg/kg
54.	1,2,4-Trichlorobenzene	35162	050823	0.05	5.00	40006.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	120-82-1	5 ppm (CL) (40mg/m3)	orl-rat 756mg/kg
55.	1,2,4-Trimethylbenzene	35162	050823	0.05	5.00	40001.6	2000	NA	NA	0.017	NA	NA	1999.6	23.0	95-63-6	N/A	orl-rat 5g/kg
56.	1,3,5-Trimethylbenzene	35162	050823	0.05	5.00	40006.7	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-67-8	N/A	orl-rat 5000mg/kg
57.	m-Xylene	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-38-3	100 ppm (435mg/m3/8H)	orl-rat 5g/kg
58.	tert-Butyl benzene	35163	101923	0.05	5.00	40001.2	2000	NA	NA	0.017	NA	NA	1999.8	22.9	98-06-6	N/A	N/A
59.	sec-Butyl benzene	35163	101923	0.05	5.00	40002.4	2000	NA	NA	0.017	NA	NA	1999.8	22.9	135-98-8	N/A	orl-rat 2240mg/kg
60.	Chlorobenzene	35163	101923	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	108-90-7	75 ppm (350mg/m3/8H)	orl-rat 2290mg/kg
61.	2-Chlorotoluene	35163	101923	0.05	5.00	40000.3	2000	NA	NA	0.017	NA	NA	1999.5	22.9	95-49-8	50 ppm (250mg/m3/8H)	orl-rat 3900mg/kg
62.	4-Chlorotoluene	35163	101923	0.05	5.00	40003.3	2000	NA	NA	0.017	NA	NA	1999.7	22.9	106-43-4	N/A	orl-rat 2100mg/kg
63.	1,2-Dichlorobenzene	35163	101923	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	95-50-1	50 ppm (300mg/m3) (CL)	orl-rat 500mg/kg
64.	1,3-Dichlorobenzene	35163	101923	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1999.6	23.0	541-73-1	N/A	ipr-mus 1062mg/kg
65.	1,4-Dichlorobenzene	35163	101923	0.05	5.00	40001.8	2000	NA	NA	0.017	NA	NA	1999.6	22.9	106-46-7	75 ppm (450mg/m3/8H)	orl-rat 500mg/kg
66.	Isopropylbenzene	35163	101923	0.05	5.00	40000.8	2000	NA	NA	0.017	NA	NA	1999.5	22.9	98-82-8	50 ppm (245mg/m3/8H)	orl-rat 1400mg/kg
67.	n-Propylbenzene	35163	101923	0.05	5.00	40003.4	2000	NA	NA	0.017	NA	NA	1999.7	23.0	103-65-1	N/A	orl-rat 8040mg/kg
68.	o-Xylene	35163	101923	0.05	5.00	40040.8	2000	NA	NA	0.017	NA	NA	2001.5	23.0	95-47-6	100 ppm (435mg/m3/8H)	ipr-mus 1364mg/kg
69.	p-Xylene	35163	101923	0.05	5.00	40000.8	2000	NA	NA	0.017	NA	NA	1999.5	22.9	106-42-3	100 ppm (435mg/m3/8H)	orl-rat 5g/kg

- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Run 16, "P95317 L021624 (2000µg/mL in MeOH)"

Run Length: 60.00 min, 35998 points at 10 points/second.
Created: Sat, Feb 17, 2024 at 8:56:46 AM.
Sampled: Sequence "021624-GC5M1", Method "GC5-M1".
Analyzed using Method "GC5-M1".

Comments
GC5-M1 Analysis by Candice Warren
Column ID SPB-Vocol 105 meter X 0.53mm X 3.0µm film thickness
Flow rates: Total flow=290mL/min., Helium (carrier)=10mL/min.,
Helium(make-up)=10mL/min., Hydogen(make-up)=40mL/min., Air(make-up)=230mL/min.
Oven Profile: Temp. 1=35°C (Time 1=10 min.), Temp 2=200°C (Time 2=8.75 min.),
Rate = 4°C/min., Total run time=60 min. Injector temp.=200°C, FID Temp.=200°C.
FID Signal = Edaq Channel 1
Standard injection = 0.5µL, Range=3



Peak #	Name	FID RT (min.)
1	Ether	9.97
2	1,1,2-Trichloro-1,2,3-trifluoroethane	10.33
3	1,1-Dichloroethene	11.10
4	Acetonitrile	12.00
5	Iodomethane	12.31
6	Allyl chloride	12.56
7	Carbon disulfide/Methylene chloride	13.04
8	trans-1,2-Dichloroethene	14.07
9	1,1-Dichloroethane	15.74
10	2,2-Dichloropropane	17.74
11	cis-1,2-Dichloroethene	18.00
12	Methacrylonitrile/Methyl acrylate/Chloroform	18.49
13	Isobutanol/1,1,1-Trichloroethane	18.91
14	1,1-Dichloropropene	20.46
15	Carbon tetrachloride	20.79
16	Benzene/1,2-Dichloroethane	21.48
17	Trichloroethene	23.86
18	1,2-Dichloropropane	24.26
19	Methyl methacrylate	24.62
20	Bromodichloromethane	25.13
21	Dibromomethane/2-Nitropropane	25.46
22	cis-1,3-Dichloropropene	27.02
23	Toluene	28.09
24	Ethyl methacrylate/trans-1,3-Dichloropropene	28.73
25	1,1,2-Trichloroethane	29.34
26	Tetrachloroethene/1,3-Dichloropropane	30.34
27	Dibromochloromethane	31.16
28	1,2-Dibromoethane	31.84
29	Chlorobenzene	33.26
30	Ethylbenzene/1,1,1,2-Tetrachloroethane	33.40
31	m-Xylene/p-Xylene	33.85
32	o-Xylene	35.33
33	Styrene	35.50
34	Isopropylbenzene/Bromoform	35.48
35	cis-1,4-Dichloro-2-butene	36.80
36	1,1,2,2-Tetrachloroethane	37.23
37	1,2,3-Trichloropropane	37.77
38	n-Propylbenzene	37.92
39	trans-1,4-Dichloro-2-butene	38.05
40	Bromobenzene	38.14
41	1,3,5-Trimethylbenzene	38.50
42	2-Chlorotoluene	38.62
43	4-Chlorotoluene	38.77
44	tert-Butylbenzene	39.76
45	1,2,4-Trimethylbenzene	39.91
46	Pentachloroethane	40.17
47	sec-Butylbenzene	40.52
48	p-Isopropyltoluene	41.02
49	1,3-Dichlorobenzene	41.42
50	1,4-Dichlorobenzene	41.63
51	n-Butylbenzene	42.02
52	1,2-Dichlorobenzene	42.18
53	1,2-Dibromo-3-chloropropane	46.12
54	Nitrobenzene	46.58
55	1,2,4-Trichlorobenzene	49.26
56	Hexachlorobutadiene	49.72
57	Naphthalene	50.26
58	1,2,3-Trichlorobenzene	51.16

Absolute Standards Inc.

PO Box 5585
Hamden, CT 06518-0585

Phone: 203-281-2917
FAX: 203-281-2922

Safety Data Sheet (SDS) GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY **ANALYTICAL STANDARD DISSOLVED IN METHANOL**

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	**1-800-535-5053**
Address	44 Rossotto Dr. Hamden CT, 06514	Emergency Telephone International	**1-352-323-3500**
		Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225	**Highly Flammable Liquid and Vapor**	**H301, 311, 331**	**Toxic if swallowed, skin contact, inhaled**
H370	**Cause damage to organs**	**H351**	**Suspected of causing cancer**
P271	**Use in ventilated area**	**P280**	**Use gloves, eye protection/face sheild**
P302,332	**If on skin, wash with soap and water**	**P305,351,338**	**If in eyes, remove contacts, rinse with water**



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))			
Methanol	METHYL ALCOHOL	CAS#: 67-56-1	% (optional) > 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance.Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist.
Storage Conditions	Use ventilation Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge. Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol 67-56-1 TWA 200 ppm
Skin notation TWA 200 ppm
Potential for skin absorption , ingestion and inhalation.
Personal protective equipment Respiratory protection Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H2O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6

Solubility in Water COMPLETE

Appearance and Odor CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.

Section X. STABILITY AND REACTIVITY

Chemical stability Stable under recommended storage conditions.
Possibility of hazardous reactions Vapours may form explosive mixture with air.
Conditions to avoid Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions. - Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50	15,400 mg/l - 96 h
EC50	24,500.00 mg/l - 48 h
EC100	10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)
UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol

IATA
UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

Exp 03/18/24

ANAB ISO 17034 Accredite
AR-1539 Certificate Numbe
https://Absolutestandards.co

CERTIFIED WEIGHT REPORT

Part Number: 91980
Lot Number: 031725
Description: Acrolein

Solvent(s): Water **Lot#** 072324Q

5 vial

V14895 to V14899

Expiration Date: 041725
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 5000
NIST Test ID#: 6UTB


5E-05 Balance Uncertainty
0.001 Flask Uncertainty


Weight(s) shown below were combined and diluted to (mL): 10.0

	031725
Formulated By: Lawrence Barry	DATE
	031725
Reviewed By: Pedro L. Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05170	5004.1	52.5	107-02-8	0.1 ppm	orl-rat 46mg/kg

Method: GC6MSD-1. **Detector:** Mass Selective Detector (Scan mode). **Column:** Vocol (60m X 0.25mm ID X 1.5µm film thickness). **Oven Profile:** Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.) Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. **Analyst:** Pedro Rentas. **NOTE:** Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately Long term storage is not recommended. Please contact our technical department if further information is required.





- **The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.**
- **Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).**
- **Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.**
- **All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.**
- **Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).**

Absolute Standards Inc.

PO Box 5585
Hamden, CT 06518-0585

Phone: 203-281-2917
FAX: 203-281-2922

Safety Data Sheet (SDS) GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY **ANALYTICAL STANDARD DISSOLVED IN WATER**

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	**1-800-535-5053**
Address	44 Rossotto Dr.	Emergency Telephone International	**1-352-323-3500**
	Hamden CT, 06514	Date Prepared/Revised	January 1, 2025

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

P271	**Use in ventilated area**	**H315**	**Causes skin and eye irritation.**
P302,332	**If on skin, wash with soap and water**	**P280**	**Use gloves, eye protection/face sheild**
		P305,351,338	**If in eyes, remove contacts, rinse with water**



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))		% (optional)
Water	CAS#: 7732-18-5	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance.Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.
Hazardous Decomposition products	Carbon oxides

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Water CAS#: 7732-18-5 TWA: 500 ppm

Personal protective equipment Respiratory protection Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.

Section IX - PHYSICAL/CHEMICAL CHARACTERISTICS

Boiling Point	100°C	Specific Gravity (H2O = 1)	1
Vapor Pressure (mm Hg)		Melting Point	

	NA		0°C
Vapor Density (AIR = 1)	NA	Evaporation rate (Butyl Acetate = 1)	NA

Solubility in Water — Completely miscible

Appearance and Odor — CLEAR, COLORLESS LIQUID WITH SLIGHT CHEMICAL ODOR.

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	NA
Conditions to avoid	NA
Materials to avoid	NA

Hazardous decomposition products - No data available

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - Rat	NA
LC50 Inhalation - Rat	NA
LD50 Dermal - Guinea pig	NA

Causes skin irritation.
Eye irritation

Section XII. ECOLOGICAL INFORMATION

LC50	NA
EC50	NA

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)
Not dangerous goods
Proper shipping name: Water

IATA
Not dangerous goods
Proper shipping name: Water

Section XV. REGULATORY INFORMATION

SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Material Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

Exp 03/18/24

ANAB ISO 17034 Accredite
AR-1539 Certificate Numbe
https://Absolutestandards.co

CERTIFIED WEIGHT REPORT

5 vial

Part Number: 91980
Lot Number: 031725
Description: Acrolein

Solvent(s): Water **Lot#** 072324Q

VI4895 to VI4899

Expiration Date: 041725
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 5000
NIST Test ID#: 6UTB

5E-05 Balance Uncertainty
0.001 Flask Uncertainty


Weight(s) shown below were combined and diluted to (mL): 10.0


Formulated By: Lawrence Barry 031725 DATE

Reviewed By: Pedro L. Rentas 031725 DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05170	5004.1	52.5	107-02-8	0.1 ppm	orl-rat 46mg/kg

Method: GC6MSD-1. **Detector:** Mass Selective Detector (Scan mode). **Column:** Vocol (60m X 0.25mm ID X 1.5µm film thickness). **Oven Profile:** Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.) Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. **Analyst:** Pedro Rentas. **NOTE:** Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately Long term storage is not recommended. Please contact our technical department if further information is required.





- **The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.**
- **Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).**
- **Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.**
- **All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.**
- **Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).**

Absolute Standards Inc.

PO Box 5585
Hamden, CT 06518-0585

Phone: 203-281-2917
FAX: 203-281-2922

Safety Data Sheet (SDS) GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY **ANALYTICAL STANDARD DISSOLVED IN WATER**

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	**1-800-535-5053**
Address	44 Rossotto Dr.	Emergency Telephone International	**1-352-323-3500**
	Hamden CT, 06514	Date Prepared/Revised	January 1, 2025

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

P271	**Use in ventilated area**	**H315**	**Causes skin and eye irritation.**
P302,332	**If on skin, wash with soap and water**	**P280**	**Use gloves, eye protection/face sheild**
		P305,351,338	**If in eyes, remove contacts, rinse with water**



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))		% (optional)
Water	CAS#: 7732-18-5	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance.Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.
Hazardous Decomposition products	Carbon oxides

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Water CAS#: 7732-18-5 TWA: 500 ppm

Personal protective equipment Respiratory protection Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.

Section IX - PHYSICAL/CHEMICAL CHARACTERISTICS

Boiling Point	100°C	Specific Gravity (H2O = 1)	1
Vapor Pressure (mm Hg)		Melting Point	

	NA		0°C
Vapor Density (AIR = 1)	NA	Evaporation rate (Butyl Acetate = 1)	NA

Solubility in Water — Completely miscible

Appearance and Odor — CLEAR, COLORLESS LIQUID WITH SLIGHT CHEMICAL ODOR.

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	NA
Conditions to avoid	NA
Materials to avoid	NA

Hazardous decomposition products - No data available

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - Rat	NA
LC50 Inhalation - Rat	NA
LD50 Dermal - Guinea pig	NA

Causes skin irritation.
Eye irritation

Section XII. ECOLOGICAL INFORMATION

LC50	NA
EC50	NA

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)
Not dangerous goods
Proper shipping name: Water

IATA
Not dangerous goods
Proper shipping name: Water

Section XV. REGULATORY INFORMATION

SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Material Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

Exp 03/18/24

ANAB ISO 17034 Accredite
AR-1539 Certificate Numbe
https://Absolutestandards.co

CERTIFIED WEIGHT REPORT

5 vial

Part Number:	91980
Lot Number:	031725
Description:	Acrolein
Expiration Date:	041725
Recommended Storage:	Refrigerate (4 °C)
Nominal Concentration (µg/mL):	5000
NIST Test ID#:	6UTB

Solvent(s): **Lot#**
Water 072324Q


V14895 to V14899


5E-05 Balance Uncertainty
Weight(s) shown below were combined and diluted to (mL): 10.0 0.001 Flask Uncertainty

Formulated By:	Lawrence Barry	031725 DATE
Reviewed By:	Pedro L. Rentas	031725 DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05170	5004.1	52.5	107-02-8	0.1 ppm	orl-rat 46mg/kg

Method: GC6MSD-1. **Detector:** Mass Selective Detector (Scan mode). **Column:** Vocol (60m X 0.25mm ID X 1.5µm film thickness). **Oven Profile:** Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.) Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. **Analyst:** Pedro Rentas. **NOTE:** Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately Long term storage is not recommended. Please contact our technical department if further information is required.





- **The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.**
- **Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).**
- **Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.**
- **All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.**
- **Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).**

Absolute Standards Inc.

PO Box 5585
Hamden, CT 06518-0585

Phone: 203-281-2917
FAX: 203-281-2922

Safety Data Sheet (SDS) GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY **ANALYTICAL STANDARD DISSOLVED IN WATER**

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	**1-800-535-5053**
Address	44 Rossotto Dr.	Emergency Telephone International	**1-352-323-3500**
	Hamden CT, 06514	Date Prepared/Revised	January 1, 2025

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

P271	**Use in ventilated area**	**H315**	**Causes skin and eye irritation.**
P302,332	**If on skin, wash with soap and water**	**P280**	**Use gloves, eye protection/face sheild**
		P305,351,338	**If in eyes, remove contacts, rinse with water**



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))		% (optional)
Water	CAS#: 7732-18-5	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance.Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.
Hazardous Decomposition products	Carbon oxides

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Water CAS#: 7732-18-5 TWA: 500 ppm

Personal protective equipment Respiratory protection Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.

Section IX - PHYSICAL/CHEMICAL CHARACTERISTICS

Boiling Point	100°C	Specific Gravity (H2O = 1)	1
Vapor Pressure (mm Hg)		Melting Point	

	NA		0°C
Vapor Density (AIR = 1)	NA	Evaporation rate (Butyl Acetate = 1)	NA

Solubility in Water — Completely miscible

Appearance and Odor — CLEAR, COLORLESS LIQUID WITH SLIGHT CHEMICAL ODOR.

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	NA
Conditions to avoid	NA
Materials to avoid	NA

Hazardous decomposition products - No data available

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - Rat	NA
LC50 Inhalation - Rat	NA
LD50 Dermal - Guinea pig	NA

Causes skin irritation.
Eye irritation

Section XII. ECOLOGICAL INFORMATION

LC50	NA
EC50	NA

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)
Not dangerous goods
Proper shipping name: Water

IATA
Not dangerous goods
Proper shipping name: Water

Section XV. REGULATORY INFORMATION

SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Material Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

Exp 03/18/24

ANAB ISO 17034 Accredite
AR-1539 Certificate Numbe
https://Absolutestandards.co

CERTIFIED WEIGHT REPORT

Part Number: 91980
Lot Number: 031725
Description: Acrolein

Expiration Date: 041725
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 5000
NIST Test ID#: 6UTB

Solvent(s): Water Lot# 072324Q

5 vial


V14895 to V14899


Formulated By:	Lawrence Barry	031725	DATE
Reviewed By:	Pedro L. Rentas	031725	DATE

5E-05 Balance Uncertainty
Weight(s) shown below were combined and diluted to (mL): 10.0 0.001 Flask Uncertainty

	Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50
1.	Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05170	5004.1	52.5	107-02-8	0.1 ppm	orl-rat 46mg/kg

Method: GC6MSD-1. **Detector:** Mass Selective Detector (Scan mode). **Column:** Vocol (60m X 0.25mm ID X 1.5µm film thickness). **Oven Profile:** Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.) Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. **Analyst:** Pedro Rentas. **NOTE:** Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately Long term storage is not recommended. Please contact our technical department if further information is required.





- **The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.**
- **Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).**
- **Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.**
- **All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.**
- **Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).**

Absolute Standards Inc.

PO Box 5585
Hamden, CT 06518-0585

Phone: 203-281-2917
FAX: 203-281-2922

Safety Data Sheet (SDS) GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY **ANALYTICAL STANDARD DISSOLVED IN WATER**

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	**1-800-535-5053**
Address	44 Rossotto Dr.	Emergency Telephone International	**1-352-323-3500**
	Hamden CT, 06514	Date Prepared/Revised	January 1, 2025

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

P271	**Use in ventilated area**	**H315**	**Causes skin and eye irritation.**
P302,332	**If on skin, wash with soap and water**	**P280**	**Use gloves, eye protection/face sheild**
		P305,351,338	**If in eyes, remove contacts, rinse with water**



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))		% (optional)
Water	CAS#: 7732-18-5	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance.Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.
Hazardous Decomposition products	Carbon oxides

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Water CAS#: 7732-18-5 TWA: 500 ppm

Personal protective equipment Respiratory protection Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.

Section IX - PHYSICAL/CHEMICAL CHARACTERISTICS

Boiling Point	100°C	Specific Gravity (H2O = 1)	1
Vapor Pressure (mm Hg)		Melting Point	

	NA		0°C
Vapor Density (AIR = 1)	NA	Evaporation rate (Butyl Acetate = 1)	NA

Solubility in Water — Completely miscible

Appearance and Odor — CLEAR, COLORLESS LIQUID WITH SLIGHT CHEMICAL ODOR.

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	NA
Conditions to avoid	NA
Materials to avoid	NA

Hazardous decomposition products - No data available

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - Rat	NA
LC50 Inhalation - Rat	NA
LD50 Dermal - Guinea pig	NA

Causes skin irritation.
Eye irritation

Section XII. ECOLOGICAL INFORMATION

LC50	NA
EC50	NA

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)
Not dangerous goods
Proper shipping name: Water

IATA
Not dangerous goods
Proper shipping name: Water

Section XV. REGULATORY INFORMATION

SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Material Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

Rec 09/16/24 2 Vials

ANAB ISO 17034 Accredited
AR-1539 Certificate Numb
https://Absolutestandards.co...

CERTIFIED WEIGHT REPORT

Part Number: 91980
Lot Number: 091424
Description: Acrolein

Solvent(s): Water **Lot#** 072324Q


Expiration Date: 101424
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 5000
NIST Test ID#: 6UTB


Weight(s) shown below were combined and diluted to (mL): 10.0 | 5E-05 Balance Uncertainty | 0.001 Flask Uncertainty

Justin Dippold	091424
Formulated By: Justin Dippold	DATE
Pedro Rentas	091424
Reviewed By: Pedro L. Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05175	5008.9	52.5	107-02-8	0.1 ppm	orl-rat 46mg/kg

Method: GC6MSD-1. **Detector:** Mass Selective Detector (Scan mode). **Column:** Vocol (60m X 0.25mm ID X 1.5µm film thickness). **Oven Profile:** Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.), Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. **Analyst:** Pedro Rentas. **NOTE:** Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately Long term storage is not recommended. Please contact our technical department if further information is required.





- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

Rec 09/16/24 2 Vials

ANAB ISO 17034 Accredited
AR-1539 Certificate Numb
https://Absolutestandards.co...

CERTIFIED WEIGHT REPORT

Part Number: 91980
Lot Number: 091424
Description: Acrolein

Expiration Date: 101424
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 5000
NIST Test ID#: 6UTB


Solvent(s): Water **Lot#** 072324Q


5E-05 Balance Uncertainty
Weight(s) shown below were combined and diluted to (mL): 10.0 0.001 Flask Uncertainty

Formulated By:	Justin Dippold	091424 DATE
Reviewed By:	Pedro L. Rentas	091424 DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05175	5008.9	52.5	107-02-8	0.1 ppm	orl-rat 46mg/kg

Method: GC6MSD-1. **Detector:** Mass Selective Detector (Scan mode). **Column:** Vocol (60m X 0.25mm ID X 1.5µm film thickness). **Oven Profile:** Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.), Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. **Analyst:** Pedro Rentas. **NOTE:** Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately Long term storage is not recommended. Please contact our technical department if further information is required.





- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

Rec 12/6/24
20 vial
V14630 to
V14649

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT

Part Number: 95318
Lot Number: 120524
Description: 2-Chloroethyl vinyl ether

Expiration Date: 120527
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 10000
NIST Test ID#: 6UTB

Solvent(s): Methanol **Lot#** EJ143-US


5E-05 Balance Uncertainty
0.001 Flask Uncertainty


Weight(s) shown below were combined and diluted to (mL): 50.0

Prashant Chauhan	120524	
Formulated By:	Prashant Chauhan	DATE
Pedro Rentas	120524	
Reviewed By:	Pedro L. Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc(µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50
1. 2-Chloroethyl vinyl ether	74	MKCD0033	10000	99	0.2	0.50536	0.50550	10002.9	40.5	110-75-8	N/A	ori-rat 250mg/kg

Method: GC6MSD-1.M. **Detector:** MSD. **Column:** (60m X 0.25mm X 1.5 µm). **Oven Profile:** Temp 1 = 35°C (Time 1=10min.), Temp 2 = 200°C (Time 2=8.75 min.), Rate = 4°C/min., Injector B Temp.= 200°C, Detector B Temp. = 220°C. **Analyst:** Candice Warren.





- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Absolute Standards Inc.
PO Box 5585
Hamden, CT 06518-0585
Phone: 203-281-2917
FAX: 203-281-2922

Safety Data Sheet (SDS) GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY **ANALYTICAL STANDARD DISSOLVED IN METHANOL**

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	**1-800-535-5053**
Address	44 Rossotto Dr. Hamden CT, 06514	Emergency Telephone International Date Prepared/Revised	**1-352-323-3500** January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225	**Highly Flammable Liquid and Vapor**	**H301, 311, 331**	**Toxic if swallowed, skin contact, inhaled**
H370	**Cause damage to organs**	**H351**	**Suspected of causing cancer**
P271	**Use in ventilated area**	**P280**	**Use gloves, eye protection/face sheild**
P302,332	**If on skin, wash with soap and water**	**P305,351,338**	**If in eyes, remove contacts, rinse with water**



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))			% (optional)
Methanol	METHYL ALCOHOL	CAS#: 67-56-1	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance.Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol 67-56-1 TWA 200 ppm
Skin notation TWA 200 ppm
Potential for skin absorption , ingestion and inhalation.
Personal protective equipment Respiratory protection Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H2O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6

Solubility in Water COMPLETE

Appearance and Odor CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.

Section X. STABILITY AND REACTIVITY

Chemical stability Stable under recommended storage conditions.
Possibility of hazardous reactions Vapours may form explosive mixture with air.
Conditions to avoid Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions. - Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50	15,400 mg/l - 96 h
EC50	24,500.00 mg/l - 48 h
EC100	10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)
UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol

IATA
UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

Rec 12/6/24
20 vial
V14630 to
V14649

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT

Part Number: 95318
Lot Number: 120524
Description: 2-Chloroethyl vinyl ether

Expiration Date: 120527
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 10000
NIST Test ID#: 6UTB


Solvent(s): Methanol **Lot#** EJ143-US


5E-05 Balance Uncertainty
Weight(s) shown below were combined and diluted to (mL): 50.0 0.001 Flask Uncertainty

Prashant Chauhan	120524	
Formulated By:	Prashant Chauhan	DATE
Pedro Rentas	120524	
Reviewed By:	Pedro L. Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc(µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50
1. 2-Chloroethyl vinyl ether	74	MKCD0033	10000	99	0.2	0.50536	0.50550	10002.9	40.5	110-75-8	N/A	ori-rat 250mg/kg

Method: GC6MSD-1.M. **Detector:** MSD. **Column:** (60m X 0.25mm X 1.5 µm). **Oven Profile:** Temp 1 = 35°C (Time 1=10min.), Temp 2 = 200°C (Time 2=8.75 min.), Rate = 4°C/min., Injector B Temp.= 200°C, Detector B Temp. = 220°C. **Analyst:** Candice Warren.





- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Absolute Standards Inc.
PO Box 5585
Hamden, CT 06518-0585
Phone: 203-281-2917
FAX: 203-281-2922

Safety Data Sheet (SDS) GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY **ANALYTICAL STANDARD DISSOLVED IN METHANOL**

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	**1-800-535-5053**
Address	44 Rossotto Dr. Hamden CT, 06514	Emergency Telephone International Date Prepared/Revised	**1-352-323-3500** January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225	**Highly Flammable Liquid and Vapor**	**H301, 311, 331**	**Toxic if swallowed, skin contact, inhaled**
H370	**Cause damage to organs**	**H351**	**Suspected of causing cancer**
P271	**Use in ventilated area**	**P280**	**Use gloves, eye protection/face sheild**
P302,332	**If on skin, wash with soap and water**	**P305,351,338**	**If in eyes, remove contacts, rinse with water**



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))			% (optional)
Methanol	METHYL ALCOHOL	CAS#: 67-56-1	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance.Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol 67-56-1 TWA 200 ppm
Skin notation TWA 200 ppm
Potential for skin absorption , ingestion and inhalation.
Personal protective equipment Respiratory protection Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H2O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6

Solubility in Water COMPLETE

Appearance and Odor CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.

Section X. STABILITY AND REACTIVITY

Chemical stability — Stable under recommended storage conditions.
Possibility of hazardous reactions — Vapours may form explosive mixture with air.
Conditions to avoid — Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid — Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions. - Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50	15,400 mg/l - 96 h
EC50	24,500.00 mg/l - 48 h
EC100	10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)
UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol

IATA
UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

Rec 12/6/24
20 vial
V14630 to
V14649

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT

Part Number: 95318
Lot Number: 120524
Description: 2-Chloroethyl vinyl ether

Expiration Date: 120527
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 10000
NIST Test ID#: 6UTB


Solvent(s):	Lot#
Methanol	EJ143-US


5E-05 Balance Uncertainty
Weight(s) shown below were combined and diluted to (mL): 50.0 0.001 Flask Uncertainty

Prashant Chauhan		120524
Formulated By:	Prashant Chauhan	DATE
Pedro Rentas		120524
Reviewed By:	Pedro L. Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc(µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50
1. 2-Chloroethyl vinyl ether	74	MKCD0033	10000	99	0.2	0.50536	0.50550	10002.9	40.5	110-75-8	N/A	ori-rat 250mg/kg

Method: GC6MSD-1.M. **Detector:** MSD. **Column:** (60m X 0.25mm X 1.5 µm). **Oven Profile:** Temp 1 = 35°C (Time 1=10min.), Temp 2 = 200°C (Time 2=8.75 min.), Rate = 4°C/min., Injector B Temp.= 200°C, Detector B Temp. = 220°C. **Analyst:** Candice Warren.





- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Absolute Standards Inc.
PO Box 5585
Hamden, CT 06518-0585
Phone: 203-281-2917
FAX: 203-281-2922

Safety Data Sheet (SDS) GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY **ANALYTICAL STANDARD DISSOLVED IN METHANOL**

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	**1-800-535-5053**
Address	44 Rossotto Dr. Hamden CT, 06514	Emergency Telephone International Date Prepared/Revised	**1-352-323-3500** January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225	**Highly Flammable Liquid and Vapor**	**H301, 311, 331**	**Toxic if swallowed, skin contact, inhaled**
H370	**Cause damage to organs**	**H351**	**Suspected of causing cancer**
P271	**Use in ventilated area**	**P280**	**Use gloves, eye protection/face sheild**
P302,332	**If on skin, wash with soap and water**	**P305,351,338**	**If in eyes, remove contacts, rinse with water**



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))			% (optional)
Methanol	METHYL ALCOHOL	CAS#: 67-56-1	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance.Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol 67-56-1 TWA 200 ppm
Skin notation TWA 200 ppm
Potential for skin absorption , ingestion and inhalation.
Personal protective equipment Respiratory protection Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H2O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6

Solubility in Water COMPLETE

Appearance and Odor CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.

Section X. STABILITY AND REACTIVITY

Chemical stability Stable under recommended storage conditions.
Possibility of hazardous reactions Vapours may form explosive mixture with air.
Conditions to avoid Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions. - Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50	15,400 mg/l - 96 h
EC50	24,500.00 mg/l - 48 h
EC100	10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)
UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol

IATA
UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

Rec 12/6/24
20 vial
V14630 to
V14649

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT

Part Number: 95318
Lot Number: 120524
Description: 2-Chloroethyl vinyl ether

Expiration Date: 120527
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 10000
NIST Test ID#: 6UTB


Solvent(s): Methanol **Lot#** EJ143-US


5E-05 Balance Uncertainty
Weight(s) shown below were combined and diluted to (mL): 50.0 0.001 Flask Uncertainty

Prashant Chauhan		120524
Formulated By:	Prashant Chauhan	DATE
Pedro Rentas		120524
Reviewed By:	Pedro L. Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc(µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50
1. 2-Chloroethyl vinyl ether	74	MKCD0033	10000	99	0.2	0.50536	0.50550	10002.9	40.5	110-75-8	N/A	ori-rat 250mg/kg

Method: GC6MSD-1.M. **Detector:** MSD. **Column:** (60m X 0.25mm X 1.5 µm). **Oven Profile:** Temp 1 = 35°C (Time 1=10min.), Temp 2 = 200°C (Time 2=8.75 min.), Rate = 4°C/min., Injector B Temp.= 200°C, Detector B Temp. = 220°C. **Analyst:** Candice Warren.





- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Absolute Standards Inc.
PO Box 5585
Hamden, CT 06518-0585
Phone: 203-281-2917
FAX: 203-281-2922

Safety Data Sheet (SDS) GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY **ANALYTICAL STANDARD DISSOLVED IN METHANOL**

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	**1-800-535-5053**
Address	44 Rossotto Dr. Hamden CT, 06514	Emergency Telephone International Date Prepared/Revised	**1-352-323-3500** January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225	**Highly Flammable Liquid and Vapor**	**H301, 311, 331**	**Toxic if swallowed, skin contact, inhaled**
H370	**Cause damage to organs**	**H351**	**Suspected of causing cancer**
P271	**Use in ventilated area**	**P280**	**Use gloves, eye protection/face sheild**
P302,332	**If on skin, wash with soap and water**	**P305,351,338**	**If in eyes, remove contacts, rinse with water**



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))			% (optional)
Methanol	METHYL ALCOHOL	CAS#: 67-56-1	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance.Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol 67-56-1 TWA 200 ppm
Skin notation TWA 200 ppm
Potential for skin absorption , ingestion and inhalation.
Personal protective equipment Respiratory protection Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H2O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6

Solubility in Water COMPLETE

Appearance and Odor CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.

Section X. STABILITY AND REACTIVITY

Chemical stability — Stable under recommended storage conditions.
Possibility of hazardous reactions — Vapours may form explosive mixture with air.
Conditions to avoid — Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid — Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions. - Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50	15,400 mg/l - 96 h
EC50	24,500.00 mg/l - 48 h
EC100	10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)
UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol

IATA
UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30067 **Lot No.:** A0191805

Description : 4-Bromofluorobenzene Standard

4-Bromofluorobenzene Standard 2,500µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : November 30, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1-Bromo-4-fluorobenzene (BFB)	460-00-4	184975	99%	2,483.9 µg/mL	+/- 139.5488

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:
105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:
hydrogen-constant pressure 11.0 psi.

Temp. Program:
40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)


Inj. Temp:
200°C

Det. Temp:
250°C

Det. Type:
FID

Split Vent:
40 ml/min

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Alicia Leathers - Operation Technician I

Date Mixed: 17-Nov-2022 **Balance Serial #** B251644995

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 21-Nov-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30225 **Lot No.:** A0193071

Description : Bromochloromethane Standard

Bromochloromethane 2000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : December 31, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Bromochloromethane	74-97-5	00008541	99%	2,018.0 µg/mL	+/- 113.3890

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:
105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:
hydrogen-constant pressure 11.0 psi.

Temp. Program:
40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)


Inj. Temp:
200°C

Det. Temp:
250°C

Det. Type:
FID

Split Vent:
40 ml/min

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Tom Suckar - Mix Technician

Date Mixed: 29-Dec-2022 **Balance Serial #** B707717271

Christie Mills - Operations Tech II - ARM QC

Date Passed: 03-Jan-2023

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30225 **Lot No.:** A0193071

Description : Bromochloromethane Standard

Bromochloromethane 2000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : December 31, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Bromochloromethane	74-97-5	00008541	99%	2,018.0 µg/mL	+/- 113.3890

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:
105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:
hydrogen-constant pressure 11.0 psi.

Temp. Program:
40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)


Inj. Temp:
200°C

Det. Temp:
250°C

Det. Type:
FID

Split Vent:
40 ml/min

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Tom Suckar - Mix Technician

Date Mixed: 29-Dec-2022 **Balance Serial #** B707717271

Christie Mills - Operations Tech II - ARM QC

Date Passed: 03-Jan-2023

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30225 **Lot No.:** A0193071

Description : Bromochloromethane Standard

Bromochloromethane 2000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : December 31, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Bromochloromethane	74-97-5	00008541	99%	2,018.0 µg/mL	+/- 113.3890

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:
105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:
hydrogen-constant pressure 11.0 psi.

Temp. Program:
40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)


Inj. Temp:
200°C

Det. Temp:
250°C

Det. Type:
FID

Split Vent:
40 ml/min

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Tom Suckar - Mix Technician

Date Mixed: 29-Dec-2022 **Balance Serial #** B707717271

Christie Mills - Operations Tech II - ARM QC

Date Passed: 03-Jan-2023

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30225 **Lot No.:** A0193071

Description : Bromochloromethane Standard

Bromochloromethane 2000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : December 31, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Bromochloromethane	74-97-5	00008541	99%	2,018.0 µg/mL	+/- 113.3890

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:
105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:
hydrogen-constant pressure 11.0 psi.

Temp. Program:
40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)


Inj. Temp:
200°C

Det. Temp:
250°C

Det. Type:
FID

Split Vent:
40 ml/min

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Tom Suckar - Mix Technician

Date Mixed: 29-Dec-2022 **Balance Serial #** B707717271

Christie Mills - Operations Tech II - ARM QC

Date Passed: 03-Jan-2023

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555582 **Lot No.:** A0196865

Description : Custom 8260A/B Surrogate Mix

Custom 8260A/B Surrogate Mix 25,000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2026 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2-Dichloroethane-d4	17060-07-0	PR-32845	99%	25,036.0 µg/mL	+/- 1,417.9179
2	1-Bromo-4-fluorobenzene (BFB)	460-00-4	184975	99%	25,132.0 µg/mL	+/- 1,423.3549
3	Dibromofluoromethane	1868-53-7	022013	99%	25,040.0 µg/mL	+/- 1,418.1445
4	Toluene-d8	2037-26-5	PR-33397	99%	25,028.0 µg/mL	+/- 1,417.4648

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Russ Bookhamer - Operations Technician I

Date Mixed: 11-Apr-2023 **Balance:** 1127510105

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30489 **Lot No.:** A0209618

Description : 8260B Acetates Mix

8260B Acetates Mix 2,000 µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : September 30, 2025 **Storage:** -20°C or colder

Handling: This product is photosensitive. **Ship:** On Ice

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Methyl acetate	79-20-9	SHBP3100	99%	2,019.3 µg/mL	+/- 69.7974
2	Vinyl acetate	108-05-4	RP231030CTH	98%	2,016.8 µg/mL	+/- 69.7112
3	Ethyl acetate	141-78-6	SHBQ9682	99%	2,010.7 µg/mL	+/- 69.4979
4	Isopropyl acetate	108-21-4	BCCG7069	99%	2,016.0 µg/mL	+/- 69.6822
5	Propyl acetate	109-60-4	P8XLN	99%	2,008.0 µg/mL	+/- 69.4057
6	Butyl acetate	123-86-4	SHBP6314	99%	2,007.3 µg/mL	+/- 69.3826
7	Amyl acetate	628-63-7	41325/1	97%	2,004.7 µg/mL	+/- 69.2905

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods . Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this

reason, we offer vinyl acetate in individual solution, and suggest that it be introduced into the working level calibration solution immediately before use. This will minimize problems and ensure more consistent results.

Quality Confirmation Test

Column:
105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:
hydrogen-constant pressure 11.0 psi.

Temp. Program:
40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)


Inj. Temp:
200°C

Det. Temp:
250°C

Det. Type:
FID

Split Vent:
40 ml/min

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler - Operations Tech I

Date Mixed: 28-Mar-2024 **Balance Serial #** B707717271

Dillan Murphy - Operations Technician I

Date Passed: 01-Apr-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30489 **Lot No.:** A0209618

Description : 8260B Acetates Mix

8260B Acetates Mix 2,000 µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : September 30, 2025 **Storage:** -20°C or colder

Handling: This product is photosensitive. **Ship:** On Ice

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Methyl acetate	79-20-9	SHBP3100	99%	2,019.3 µg/mL	+/- 69.7974
2	Vinyl acetate	108-05-4	RP231030CTH	98%	2,016.8 µg/mL	+/- 69.7112
3	Ethyl acetate	141-78-6	SHBQ9682	99%	2,010.7 µg/mL	+/- 69.4979
4	Isopropyl acetate	108-21-4	BCCG7069	99%	2,016.0 µg/mL	+/- 69.6822
5	Propyl acetate	109-60-4	P8XLN	99%	2,008.0 µg/mL	+/- 69.4057
6	Butyl acetate	123-86-4	SHBP6314	99%	2,007.3 µg/mL	+/- 69.3826
7	Amyl acetate	628-63-7	41325/1	97%	2,004.7 µg/mL	+/- 69.2905

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods . Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this

reason, we offer vinyl acetate in individual solution, and suggest that it be introduced into the working level calibration solution immediately before use. This will minimize problems and ensure more consistent results.

Quality Confirmation Test

Column:
105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:
hydrogen-constant pressure 11.0 psi.

Temp. Program:
40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)


Inj. Temp:
200°C

Det. Temp:
250°C

Det. Type:
FID

Split Vent:
40 ml/min

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler - Operations Tech I

Date Mixed: 28-Mar-2024 **Balance Serial #** B707717271

Dillan Murphy - Operations Technician I

Date Passed: 01-Apr-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ilac-MRA ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ilac-MRA ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555581 **Lot No.:** A0210184

Description : Custom 8260 Internal Standard Mix

Custom 8260 Internal Standard Mix 25,000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2027 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dichlorobenzene-d4	3855-82-1	PR-30447	99%	25,212.0 µg/mL	+/- 1,427.8857
2	1,4-Difluorobenzene	540-36-3	MKCS8657	99%	25,220.0 µg/mL	+/- 1,428.3388
3	Chlorobenzene-d5	3114-55-4	PR-31132	99%	25,116.0 µg/mL	+/- 1,422.4487
4	Pentafluorobenzene	363-72-4	MKCR9383	99%	25,180.0 µg/mL	+/- 1,426.0734

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

John Friedline - Operations Technician I

Date Mixed: 11-Apr-2024 **Balance:** 1127510105

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Rec 12/17/24

CERTIFIED REFERENCE MATERIAL

30 V.4

Certificate of Analysis

chromatographic plus

V14697-to-14726

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30006 **Lot No.:** A0210618

Description : VOA Calibration Mix #1

VOA Calibration Mix #1 5,000µg/mL, P&T Methanol/Water(90:10), 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Acetone	67-64-1	SHBQ8504	99%	5,014.8 µg/mL	+/- 173.2883
2	2-Butanone (MEK)	78-93-3	SHBQ4704	99%	5,012.4 µg/mL	+/- 173.2054
3	4-Methyl-2-pentanone (MIBK)	108-10-1	SHBP9200	99%	5,011.6 µg/mL	+/- 173.1777
4	2-Hexanone	591-78-6	MKCQ6663	99%	5,013.0 µg/mL	+/- 173.2261

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol/Water (90:10)
CAS # 67-56-1/7732-18-5
Purity 99%

Quality Confirmation Test

Column:
105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:
hydrogen-constant pressure 11.0 psi.

Temp. Program:
40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)


Inj. Temp:
200°C

Det. Temp:
250°C

Det. Type:
FID

Split Vent:
40 ml/min

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Dakota Parson - Operations Technician I

Date Mixed: 22-Apr-2024 **Balance Serial #** B707717271

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Rec 12/17/24

CERTIFIED REFERENCE MATERIAL

30 V.4

Certificate of Analysis

chromatographic plus

V14697-to-14726

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30006 **Lot No.:** A0210618

Description : VOA Calibration Mix #1

VOA Calibration Mix #1 5,000µg/mL, P&T Methanol/Water(90:10), 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Acetone	67-64-1	SHBQ8504	99%	5,014.8 µg/mL	+/- 173.2883
2	2-Butanone (MEK)	78-93-3	SHBQ4704	99%	5,012.4 µg/mL	+/- 173.2054
3	4-Methyl-2-pentanone (MIBK)	108-10-1	SHBP9200	99%	5,011.6 µg/mL	+/- 173.1777
4	2-Hexanone	591-78-6	MKCQ6663	99%	5,013.0 µg/mL	+/- 173.2261

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol/Water (90:10)
CAS # 67-56-1/7732-18-5
Purity 99%

Quality Confirmation Test

Column:
105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:
hydrogen-constant pressure 11.0 psi.

Temp. Program:
40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)


Inj. Temp:
200°C

Det. Temp:
250°C

Det. Type:
FID

Split Vent:
40 ml/min

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Dakota Parson - Operations Technician I

Date Mixed: 22-Apr-2024 **Balance Serial #** B707717271

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Rec 12/17/24

CERTIFIED REFERENCE MATERIAL

30 V.4

Certificate of Analysis

chromatographic plus

V14697-to-14726

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30006 **Lot No.:** A0210618

Description : VOA Calibration Mix #1

VOA Calibration Mix #1 5,000µg/mL, P&T Methanol/Water(90:10), 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Acetone	67-64-1	SHBQ8504	99%	5,014.8 µg/mL	+/- 173.2883
2	2-Butanone (MEK)	78-93-3	SHBQ4704	99%	5,012.4 µg/mL	+/- 173.2054
3	4-Methyl-2-pentanone (MIBK)	108-10-1	SHBP9200	99%	5,011.6 µg/mL	+/- 173.1777
4	2-Hexanone	591-78-6	MKCQ6663	99%	5,013.0 µg/mL	+/- 173.2261

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol/Water (90:10)
CAS # 67-56-1/7732-18-5
Purity 99%

Quality Confirmation Test

Column:
105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:
hydrogen-constant pressure 11.0 psi.

Temp. Program:
40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)


Inj. Temp:
200°C

Det. Temp:
250°C

Det. Type:
FID

Split Vent:
40 ml/min

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Dakota Parson - Operations Technician I

Date Mixed: 22-Apr-2024 **Balance Serial #** B707717271

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Rec 12/17/24

CERTIFIED REFERENCE MATERIAL

30 Vial

Certificate of Analysis

chromatographic plus

V14727 to
V14756

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30042 **Lot No.:** A0216826

Description : 502.2 Calibration Mix #1

502.2 Calibration Mix #1 2,000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : May 31, 2031 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Dichlorodifluoromethane (CFC-12)	75-71-8	00022922	99%	2,000.9 µg/mL	+/- 112.4144
2	Chloromethane (methyl chloride)	74-87-3	00022694	99%	2,000.7 µg/mL	+/- 112.3998
3	Vinyl chloride	75-01-4	00015559	99%	2,000.3 µg/mL	+/- 112.3779
4	Bromomethane (methyl bromide)	74-83-9	00017022	99%	2,001.8 µg/mL	+/- 112.4650
5	Chloroethane (ethyl chloride)	75-00-3	107-401039114-1	99%	2,000.1 µg/mL	+/- 112.3700
6	Trichlorofluoromethane (CFC-11)	75-69-4	MKCJ8658	99%	2,000.7 µg/mL	+/- 112.3992

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:
60m x 0.25mm x 1.4µm
Rtx-502.2 (cat.#10916)

Carrier Gas:
helium-constant flow 2.0 mL/min.

Temp. Program:
40°C (hold 6 min.) to 100°C
@ 6°C/min.


Inj. Temp:
200°C

Det. Temp:
250°C

Det. Type:
MSD

Split Vent:
Split ratio 10:1

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Tom Suckar - Mix Technician

Date Mixed: 23-Sep-2024 **Balance Serial #** B707717271

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 04-Oct-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Rec 12/17/24

CERTIFIED REFERENCE MATERIAL

30 Vial

Certificate of Analysis

chromatographic plus

V14727 to V14756

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30042 **Lot No.:** A0216826

Description : 502.2 Calibration Mix #1

502.2 Calibration Mix #1 2,000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : May 31, 2031 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Dichlorodifluoromethane (CFC-12)	75-71-8	00022922	99%	2,000.9 µg/mL	+/- 112.4144
2	Chloromethane (methyl chloride)	74-87-3	00022694	99%	2,000.7 µg/mL	+/- 112.3998
3	Vinyl chloride	75-01-4	00015559	99%	2,000.3 µg/mL	+/- 112.3779
4	Bromomethane (methyl bromide)	74-83-9	00017022	99%	2,001.8 µg/mL	+/- 112.4650
5	Chloroethane (ethyl chloride)	75-00-3	107-401039114-1	99%	2,000.1 µg/mL	+/- 112.3700
6	Trichlorofluoromethane (CFC-11)	75-69-4	MKCJ8658	99%	2,000.7 µg/mL	+/- 112.3992

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:
60m x 0.25mm x 1.4µm
Rtx-502.2 (cat.#10916)

Carrier Gas:
helium-constant flow 2.0 mL/min.

Temp. Program:
40°C (hold 6 min.) to 100°C
@ 6°C/min.


Inj. Temp:
200°C

Det. Temp:
250°C

Det. Type:
MSD

Split Vent:
Split ratio 10:1

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Tom Suckar - Mix Technician

Date Mixed: 23-Sep-2024 **Balance Serial #** B707717271

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 04-Oct-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

Certificate of Analysis

chromatographic plus

J14842 to 14846

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30470 **Lot No.:** A0217535

Description : tert-Butanol Standard

tert-Butanol Std 50,000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : October 31, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	tert-Butanol (TBA)	75-65-0	SHBQ8002-1	99%	50,007.5 µg/mL	+/- 717.6137

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:
105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:
hydrogen-constant pressure 11.0 psi.

Temp. Program:
40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)


Inj. Temp:
200°C

Det. Temp:
250°C

Det. Type:
FID

Split Vent:
40 ml/min

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Aaron Enyart - Operations Tech I

Date Mixed: 07-Oct-2024 **Balance Serial #** B251644995

Brittany Federinko - Operations Tech I

Date Passed: 09-Oct-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Verify Rec 01/08/21

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic

V14803 - V14822

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555408-SL **Lot No.:** A0220471

Description : Custom Vinyl Acetate Standard

Custom Vinyl Acetate Standard 8,000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : June 30, 2026 **Storage:** -20°C or colder

Handling: This product is photosensitive. **Ship:** On Ice

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Vinyl acetate	108-05-4	RD240423RSR	99%	8,066.0 µg/mL	+/- 278.7979

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods . Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this reason, we offer vinyl acetate in individual solution, and suggest that it be introduced into the working level calibration solution immediately before use. This will minimize problems and ensure more consistent results.

Quality Confirmation Test

Column:
105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:
hydrogen-constant pressure 11.0 psi.

Temp. Program:
40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)


Inj. Temp:
200°C

Det. Temp:
250°C

Det. Type:
FID

Split Vent:
40 ml/min

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Ethan Winiarski - Operations Tech I

Date Mixed: 24-Dec-2024 **Balance Serial #** 1127510105

Dillan Murphy - Operations Technician I

Date Passed: 02-Jan-2025

REVIEWED

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Zoviy Rec 01/08/21

CERTIFIED REFERENCE MATERIAL

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

Certificate of Analysis

chromatographic

V14803 - V14822

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555408-SL **Lot No.:** A0220471

Description : Custom Vinyl Acetate Standard

Custom Vinyl Acetate Standard 8,000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : June 30, 2026 **Storage:** -20°C or colder

Handling: This product is photosensitive. **Ship:** On Ice

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Vinyl acetate	108-05-4	RD240423RSR	99%	8,066.0 µg/mL	+/- 278.7979

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods . Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this reason, we offer vinyl acetate in individual solution, and suggest that it be introduced into the working level calibration solution immediately before use. This will minimize problems and ensure more consistent results.

Quality Confirmation Test

Column:
105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:
hydrogen-constant pressure 11.0 psi.

Temp. Program:
40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)


Inj. Temp:
200°C

Det. Temp:
250°C

Det. Type:
FID

Split Vent:
40 ml/min

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Ethan Winiarski - Operations Tech I

Date Mixed: 24-Dec-2024 **Balance Serial #** 1127510105

Dillan Murphy - Operations Technician I

Date Passed: 02-Jan-2025

REVIEWED

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis

avantor™ | J.T.Baker®

Material No.: 9077-02
Batch No.: 22L0562016
Manufactured Date: 2022-10-26
Expiration Date: 2025-10-25
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay (CH_3OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.2 ppm
Titrable Acid (µeq/g)	≤ 0.3	0.2
Titrable Base (µeq/g)	≤ 0.10	0.03
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis - Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory,Research,or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

James T Ethier
Jamie Ethier
Vice President Global Quality

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis

avantor™ | J.T.Baker®

V14883
V14884

Material No.: 9077-02
Batch No.: 22L0562016
Manufactured Date: 2022-10-26
Expiration Date: 2025-10-25
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay (CH_3OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.2 ppm
Titrable Acid (µeq/g)	≤ 0.3	0.2
Titrable Base (µeq/g)	≤ 0.10	0.03
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis – Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory,Research,or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

James T Ethier
Jamie Ethier
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700
Avantor Performance Materials, LLC
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone 610.386.1700

SHIPPING DOCUMENTS

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, NJ 07092
(908) 789-8900 · Fax (908) 789-8922
www.chemtech.net

ALLIANCE PROJECT NO.
QUOTE NO. Q1731
COC Number 2045982

CLIENT INFORMATION

REPORT TO BE SENT TO:
COMPANY: Jacobs
ADDRESS: 412 Mt Kemble Ave Suite 100
CITY Morristown STATE: NJ ZIP: 07960
ATTENTION: John Yufunte John.Yufunte@Jacobs.com
PHONE: FAX:

CLIENT PROJECT INFORMATION

PROJECT NAME: STC PTC
PROJECT NO.: D3808221 LOCATION: Princeton Junction
PROJECT MANAGER: Mary Murphy
e-mail: Mary.Murphy@Jacobs.com
PHONE: FAX:

CLIENT BILLING INFORMATION

BILL TO: Mary Murphy PO#:
ADDRESS:
CITY STATE: ZIP:
ATTENTION: PHONE:

DATA TURNAROUND INFORMATION

FAX (RUSH) Standard TAT DAYS*
HARDCOPY (DATA PACKAGE): DAYS*
EDD: DAYS*
*TO BE APPROVED BY CHEMTECH
STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

DATA DELIVERABLE INFORMATION

- ☐ Level 1 (Results Only)
- ☐ Level 2 (Results + QC)
- ☑ Level 3 (Results + QC + Raw Data)
- ☐ Level 4 (QC + Full Raw Data)
- ☐ NJ Reduced ☐ US EPA CLP
- ☐ NYS ASP A ☐ NYS ASP B
- ☐ Other
- ☐ EDD FORMAT

ANALYSIS

1: 1,4 Dioxane (8270E-SIM)
2: Site Specific VOCs (8260D-LL)
3: Total Metals (6020B)
4: Dissolved Iron (6020B)
5: Alkalinity (SM2320B)
6: TDS (SM2540C)
7: Anions (9056)
8:
9:

PRESERVATIVES

COMMENTS

← Specify Preservatives
A-HCl D-NaOH
B-HN03 E-ICE
C-H2SO4 F-OTHER

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE COMP	SAMPLE TYPE GRAB	SAMPLE COLLECTION DATE	SAMPLE COLLECTION TIME	# OF BOTTLES	PRES 1 (E)	PRES 2 (A/E)	PRES 3 (B/E)	PRES 4 (E)	PRES 5 (E)	PRES 6 (E)	PRES 7 (E)	PRES 8	PRES 9
1.	RMW-01B-82-040325	GW		X	4-3-25	1135	4	✓	✓							
2.	RMW-04B-91-040325	GW		X	4-3-25	145	4	✓	✓							
3.	RMW-01B-82-040325-FD	GW		X	4-3-25	1140	4	✓	✓							
4.	RMW-03B-90-040325	GW		X	4-3-25	1510	4	✓	✓							
5.	EB01-040325	DI		X	4-3-25	1545	8	✓	✓	✓	✓	✓	✓	✓		
6.	TB01-040325	DI		X	4-3-25	1600	2	✓								
7.																
8.																
9.																
10.																

PH. 1.9
Lot # 80A0441
Samples preserved (HNO3) on 4/4/25 @ 0715
Samples placed in SM-Fridge
Samples relinquished 4/7/25 @ 0715

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER:	DATE/TIME:	RECEIVED BY:
1.	4/3/25 1634	1. 1634 4-3-25
2.		2.
3.	4-3-25 1820	3.

Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP 2.7 °C
Comments: See workorder for list of site specific VOCs
Preserve Dissolved Iron sample upon arrival to the lab
Temp 2.7°C Adjusted Factor +1) IR Gun #1.

Page 1 of 1
CLIENT: ☐ Hand Delivered ☐ Other
Shipment Complete ☐ YES ☐ NO

Copyright © 2024
WHITE - ALLIANCE COPY FOR RETURN TO CLIENT YELLOW - ALLIANCE COPY PINK - SAMPLER COPY

284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q1731 JACO05
Client Name : JACOBS Engineering Grou
Client Contact : John Ynfante
Invoice Name : JACOBS Engineering Grou
Invoice Contact : John Ynfante

Order Date : 4/4/2025 10:52:00 AM
Project Name : Former Schlumberger STC
Receive DateTime : 4/3/2025 6:20:00 PM
Purchase Order :

Project Mgr :
Report Type : Level 4
EDD Type : CH2MHILL
Hard Copy Date :
Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q1731-01	RMW-01B-82-040325	Water	04/03/2025	11:35					
					VOCMS Group3		8260-Low	2 Bus. Days	
Q1731-02	RMW-04B-91-040325	Water	04/03/2025	14:15					
					VOCMS Group3		8260-Low	2 Bus. Days	
Q1731-03	RMW-01B-82-040325-FD	Water	04/03/2025	11:40					
					VOCMS Group3		8260-Low	2 Bus. Days	
Q1731-04	RMW-03B-90-040325	Water	04/03/2025	15:10					
					VOCMS Group3		8260-Low	2 Bus. Days	
Q1731-05	EB01-040325	Water	04/03/2025	15:45					
					VOCMS Group3		8260-Low	2 Bus. Days	
Q1731-07	TB01-040325	Water	04/03/2025	16:00					
					VOCMS Group3		8260-Low	2 Bus. Days	

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

LOGIN REPORT/SAMPLE TRANSFER

Order ID :	Q1731 JACO05	**Order Date :**	4/4/2025 10:52:00 AM	**Project Mgr :**	
Client Name :	JACOBS Engineering Grou	**Project Name :**	Former Schlumberger STC	**Report Type :**	Level 4
Client Contact :	John Ynfante	**Receive DateTime :**	4/3/2025 6:20:00 PM	**EDD Type :**	CH2MHILL
Invoice Name :	JACOBS Engineering Grou	**Purchase Order :**		**Hard Copy Date :**	
Invoice Contact :	John Ynfante			**Date Signoff :**	

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES

Stored in Ref #04 (VOA)

Relinguished By : [signature]
Date / Time : 4/4/25 1120

Received By : [signature]
Date / Time : 4-4-25 11:20

Storage Area : VOA Refridgerator Room