

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05
 Lab Code: CHEM Case No.: Q1731 SAS No.: Q1731 SDG NO.: Q1731
 Lab File ID: VX045586.D Date Analyzed: 04/04/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:00
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	96707	5.54	163287	6.75	142048	10.05
UPPER LIMIT	193414	6.038	326574	7.251	284096	10.549
LOWER LIMIT	48353.5	5.038	81643.5	6.251	71024	9.549
EPA SAMPLE NO.						
EB01-040325	64123	5.54	127534	6.76	119214	10.06
TB01-040325	62156	5.55	122509	6.76	111172	10.05
VX0404WBL01	66301	5.54	128686	6.76	118182	10.06
VX0404WBS01	95127	5.55	168453	6.76	151744	10.05
VX0404WBSD01	92597	5.54	161617	6.76	145135	10.05

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.